

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
 Data File : BP009801.D
 Acq On : 13 Apr 2022 02:02
 Operator : CG/JU
 Sample : N2342-10
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0HZ3

Quant Time: Apr 13 04:27:59 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 13 02:15:22 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	116668	20.000	ng/u1	0.00
20) Naphthalene-d8	10.775	136	520591	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.592	164	391236	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	889720	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.422	240	956928	20.000	ng/u1	#-0.01
88) Perylene-d12	23.886	264	942791	20.000	ng/u1	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	14228	5.443	ng/uL	0.00
4) Pyridine-d5	3.805	84	84131	11.673	ng/u1	0.00
7) Phenol-d5	7.116	99	79699	7.781	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	200858	32.939	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	201531	26.084	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	150542	17.606	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	129251	34.429	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	136291	33.114	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	244444	28.382	ng/u1	0.00
31) 4-Chloroaniline-d4	10.899	131	347707	28.345	ng/u1	0.00
46) Dimethylphthalate-d6	14.004	166	1013306	32.972	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	1105279	30.357	ng/u1	0.00
54) 4-Nitrophenol-d4	14.769	143	38046	6.826	ng/u1	-0.01
60) Fluorene-d10	15.587	176	890083	34.429	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.692	200	154791	29.368	ng/u1	0.00
73) Anthracene-d10	17.445	188	1535964	36.056	ng/u1	0.00
81) Pyrene-d10	19.669	212	1910758	36.561	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	1833862	37.352	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
Data File : BP009801.D
Acq On : 13 Apr 2022 02:02
Operator : CG/JU
Sample : N2342-10
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0HZ3

Quant Time: Apr 13 04:27:59 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 13 02:15:22 2022
Response via : Initial Calibration

