

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
 Data File : BP009749.D
 Acq On : 11 Apr 2022 13:42
 Operator : CG/JU
 Sample : N2266-03MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD

Quant Time: Apr 12 06:32:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/12/2022
 Supervised By :Yogesh Patel 04/13/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	138901	20.000	ng/ul	0.00
20) Naphthalene-d8	10.787	136	600723	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.610	164	409844	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.363	188	900522	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.445	240	969889	20.000	ng/ul	# 0.00
88) Perylene-d12	23.921	264	946429	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.393	96	13205m	4.243	ng/uL	0.00
4) Pyridine-d5	3.811	84	62427	7.275	ng/ul	0.00
7) Phenol-d5	7.128	99	98121	8.046	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	260790	35.922	ng/ul	0.00
11) 2-Chlorophenol-d4	7.505	132	253463	27.555	ng/ul	0.00
15) 4-Methylphenol-d8	8.681	113	195742	19.228	ng/ul	0.00
21) Nitrobenzene-d5	9.140	128	172991	39.933	ng/ul	0.00
24) 2-Nitrophenol-d4	9.863	143	182692	38.466	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	320937	32.293	ng/ul	0.00
31) 4-Chloroaniline-d4	10.916	131	378738	26.756	ng/ul	0.00
46) Dimethylphthalate-d6	14.022	166	1249836	38.822	ng/ul	0.00
49) Acenaphthylene-d8	14.304	160	1382899	36.258	ng/ul	0.00
54) 4-Nitrophenol-d4	14.787	143	54628	9.356	ng/ul	0.00
60) Fluorene-d10	15.604	176	1044190	38.556	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.716	200	235521	44.149	ng/ul	0.00
73) Anthracene-d10	17.463	188	1752780	40.652	ng/ul	0.00
81) Pyrene-d10	19.686	212	2183384	41.219	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.762	264	2085250	42.309	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.428	88	15708	4.951	ng/uL#	23
5) Pyridine	3.828	79	64718	7.228	ng/ul#	45
6) Benzaldehyde	7.110	77	258765	34.600	ng/ul	97
8) Phenol	7.157	94	95671	7.838	ng/ul#	93
10) Bis(2-Chloroethyl)ether	7.393	93	302555	31.820	ng/ul#	77
12) 2-Chlorophenol	7.540	128	235513	25.396	ng/ul	92
13) 2-Methylphenol	8.416	108	198228	20.826	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.516	45	444801	31.648	ng/ul#	84
16) Acetophenone	8.799	105	505415	31.512	ng/ul#	79
17) N-Nitroso-di-n-propyla...	8.793	70	264131	31.764	ng/ul#	72
18) 4-Methylphenol	8.746	108	184871	17.935	ng/ul	91
19) Hexachloroethane	9.069	117	119658	30.560	ng/ul#	81
22) Nitrobenzene	9.181	77	385777	34.694	ng/ul#	84
23) Isophorone	9.710	82	740667	32.326	ng/ul	97
25) 2-Nitrophenol	9.893	139	176164	34.777	ng/ul#	87
26) 2,4-Dimethylphenol	9.957	107	320982	26.702	ng/ul#	79
27) Bis(2-Chloroethoxy)met...	10.193	93	430092	31.972	ng/ul#	98
29) 2,4-Dichlorophenol	10.428	162	288486	29.941	ng/ul#	83
30) Naphthalene	10.840	128	1000415	32.381	ng/ul	97
32) 4-Chloroaniline	10.940	127	340952	24.340	ng/ul	97
33) Hexachlorobutadiene	11.134	225	215515	31.057	ng/ul	95
34) Caprolactam	11.710	113	15474m	4.856	ng/ul	
35) 4-Chloro-3-methylphenol	12.057	107	324639	27.807	ng/ul#	73
36) 2-Methylnaphthalene	12.445	142	721693	31.974	ng/ul	89

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
 Data File : BP009749.D
 Acq On : 11 Apr 2022 13:42
 Operator : CG/JU
 Sample : N2266-03MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD

Quant Time: Apr 12 06:32:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/12/2022
 Supervised By :Yogesh Patel 04/13/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	734338	32.222	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.810	216	418669	33.389	ng/ul#	92
40) Hexachlorocyclopentadiene	12.798	237	237084	27.246	ng/ul	95
41) 2,4,6-Trichlorophenol	13.045	196	272162	33.987	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.110	196	303689	35.093	ng/ul#	88
43) 1,1'-Biphenyl	13.445	154	1009614	33.762	ng/ul	92
44) 2-Chloronaphthalene	13.487	162	775296	33.729	ng/ul	93
45) 2-Nitroaniline	13.681	65	280270	41.605	ng/ul#	81
47) Dimethylphthalate	14.069	163	1101012	35.533	ng/ul#	92
48) 2,6-Dinitrotoluene	14.181	165	233173	40.912	ng/ul	95
50) Acenaphthylene	14.334	152	1293060	34.354	ng/ul	96
51) 3-Nitroaniline	14.504	138	218212	36.828	ng/ul#	82
52) Acenaphthene	14.675	153	833981	33.841	ng/ul	98
53) 2,4-Dinitrophenol	14.704	184	136628	39.659	ng/ul#	79
55) 4-Nitrophenol	14.798	109	46900	8.726	ng/ul#	57
56) Dibenzofuran	15.010	168	1242733	34.263	ng/ul	99
57) 2,4-Dinitrotoluene	14.963	165	350521	41.883	ng/ul#	84
58) 2,3,4,6-Tetrachlorophenol	15.234	232	283244	35.897	ng/ul#	86
59) Diethylphthalate	15.434	149	1147209	35.775	ng/ul	93
61) Fluorene	15.657	166	1042094	34.882	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.651	204	540028	33.996	ng/ul	95
63) 4-Nitroaniline	15.669	138	247005m	41.621	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.728	198	207683	39.165	ng/ul#	90
67) N-Nitrosodiphenylamine	15.863	169	932831	36.187	ng/ul	96
68) 4-Bromophenyl-phenylether	16.551	248	344235	35.314	ng/ul	95
69) Hexachlorobenzene	16.669	284	404846	36.119	ng/ul#	91
70) Atrazine	16.822	200	370060	34.779	ng/ul#	89
71) Pentachlorophenol	17.010	266	270659	35.674	ng/ul#	85
72) Phenanthrene	17.404	178	1782762	37.088	ng/ul	98
74) Anthracene	17.498	178	1810873	37.158	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.410	216	448599	34.456	ng/uL#	84
76) Pentachlorobenzene	14.934	250	442727	34.887	ng/uL	90
77) Carbazole	17.757	167	1656246	37.606	ng/ul#	95
78) Di-n-butylphthalate	18.316	149	2054421	37.934	ng/ul#	93
80) Fluoranthene	19.363	202	2268402	37.205	ng/ul#	96
82) Pyrene	19.716	202	2305841	37.301	ng/ul#	93
83) Butylbenzylphthalate	20.586	149	988152	38.966	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.351	252	736543	35.637	ng/ul#	96
85) Benzo(a)anthracene	21.427	228	2321847	37.762	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.357	149	1454315	38.126	ng/ul#	82
87) Chrysene	21.480	228	2240266	37.764	ng/ul	99
89) Di-n-octyl phthalate	22.310	149	2494428	37.618	ng/ul	100
90) Benzo(b)fluoranthene	23.163	252	2323984	36.959	ng/ul	98
91) Benzo(k)fluoranthene	23.215	252	2249777	39.209	ng/ul#	98
93) Benzo(a)pyrene	23.815	252	2230143	38.261	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.515	276	2339124	38.377	ng/ul	97
95) Dibenzo(a,h)anthracene	26.539	278	2037390	38.549	ng/ul#	95
96) Benzo(g,h,i)perylene	27.309	276	1898349	37.773	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
 Data File : BP009749.D
 Acq On : 11 Apr 2022 13:42
 Operator : CG/JU
 Sample : N2266-03MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD

Quant Time: Apr 12 06:32:42 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/12/2022
 Supervised By :Yogesh Patel 04/13/2022









#2
1,4-Dioxane
Concen: 4.951 ng/uL
RT: 3.428 min Scan# 57
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion: 88 Resp: 15708
Ion Ratio Lower Upper
88 100
43 33.9 6.8 10.2
58 81.3 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#5
Pyridine
Concen: 7.228 ng/uL
RT: 3.828 min Scan# 126
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Tgt Ion: 79 Resp: 64718
Ion Ratio Lower Upper
79 100
52 86.3 35.5 53.3#
51 45.2 20.2 30.4#









#6
Benzaldehyde
Concen: 34.600 ng/ul
RT: 7.110 min Scan# 683
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion: 77 Resp: 25876
Ion Ratio Lower Upper
77 100
105 95.8 77.0 115.4
106 90.6 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#8
Phenol
Concen: 7.838 ng/ul
RT: 7.157 min Scan# 692
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion: 94 Resp: 95671
Ion Ratio Lower Upper
94 100
65 33.6 19.0 28.6
66 43.6 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.820 ng/ul
RT: 7.393 min Scan# 731
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion: 93 Resp: 30255
Ion Ratio Lower Upper
93 100
63 82.5 45.0 67.4
95 33.0 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#12
2-Chlorophenol
Concen: 25.396 ng/ul
RT: 7.540 min Scan# 757
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Tgt Ion:128 Resp: 235513
Ion Ratio Lower Upper
128 100
64 50.5 45.3 67.9
130 31.8 29.0 43.6









#13
2-Methylphenol
Concen: 20.826 ng/ul
RT: 8.416 min Scan# 905
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:108 Resp: 198228
Ion Ratio Lower Upper
108 100
107 90.6 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.648 ng/ul
RT: 8.516 min Scan# 923
Delta R.T. 0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion: 45 Resp: 444801
Ion Ratio Lower Upper
45 100
77 13.5 17.1 25.7#
79 10.4 13.2 19.8#





#16
Acetophenone
Concen: 31.512 ng/ul
RT: 8.799 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion: 105 Resp: 50541
Ion Ratio Lower Upper
105 100
77 81.6 53.4 80.2
51 31.5 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#17
N-Nitroso-di-n-propylamine
Concen: 31.764 ng/ul
RT: 8.793 min Scan# 970
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion: 70 Resp: 264131
Ion Ratio Lower Upper
70 100
42 63.6 31.8 47.8#
101 10.0 15.6 23.4#
130 21.4 13.7 20.5#





#18
4-Methylphenol
Concen: 17.935 ng/ul
RT: 8.746 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:108 Resp: 18487
Ion Ratio Lower Upper
108 100
107 114.2 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#19
Hexachloroethane
Concen: 30.560 ng/ul
RT: 9.069 min Scan# 1017
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:117 Resp: 119658
Ion Ratio Lower Upper
117 100
201 104.5 106.5 159.7#
199 65.4 59.0 88.6





#22
Nitrobenzene
Concen: 34.694 ng/u1
RT: 9.181 min Scan# 1036
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion: 77 Resp: 38577
Ion Ratio Lower Upper
77 100
123 42.2 42.3 63.5
65 14.9 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#23
Isophorone
Concen: 32.326 ng/u1
RT: 9.710 min Scan# 1126
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42



Tgt Ion: 82 Resp: 740667
Ion Ratio Lower Upper
82 100
95 8.0 7.8 11.8
138 17.1 12.9 19.3









#25
2-Nitrophenol
Concen: 34.777 ng/ul
RT: 9.893 min Scan# 1156
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:139 Resp: 176164
Ion Ratio Lower Upper
139 100
65 61.8 49.4 74.2
109 35.6 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#26
2,4-Dimethylphenol
Concen: 26.702 ng/ul
RT: 9.957 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Tgt Ion:107 Resp: 320982
Ion Ratio Lower Upper
107 100
121 48.2 34.1 51.1
122 80.7 46.3 69.5









#27
Bis(2-Chloroethoxy)methane
Concen: 31.972 ng/ul
RT: 10.193 min Scan# 1208
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion: 93 Resp: 430092
Ion Ratio Lower Upper
93 100
95 33.2 26.2 39.4
123 10.5 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#29
2,4-Dichlorophenol
Concen: 29.941 ng/ul
RT: 10.428 min Scan# 1248
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:162 Resp: 288486
Ion Ratio Lower Upper
162 100
164 66.5 37.0 55.6
98 38.7 31.0 46.6





#30
Naphthalene
Concen: 32.381 ng/ul
RT: 10.840 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42



Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:128 Resp: 1000415
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 13.0 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#32
4-Chloroaniline
Concen: 24.340 ng/ul
RT: 10.940 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:127 Resp: 340952
Ion Ratio Lower Upper
127 100
129 32.4 24.6 37.0





#33
Hexachlorobutadiene
Concen: 31.057 ng/ul
RT: 11.134 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion: 225 Resp: 21551
Ion Ratio Lower Upper
225 100
223 62.3 53.4 80.0
227 64.1 53.8 80.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#34
Caprolactam
Concen: 4.856 ng/ul m
RT: 11.710 min Scan# 1466
Delta R.T. 0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion: 113 Resp: 15474
Ion Ratio Lower Upper
113 100
55 169.5 58.2 87.2#
56 136.0 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 27.807 ng/ul
RT: 12.057 min Scan# 1524
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:107 Resp: 324639
Ion Ratio Lower Upper
107 100
144 23.6 15.8 23.8
142 72.2 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#36
2-Methylnaphthalene
Concen: 31.974 ng/ul
RT: 12.445 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Tgt Ion:142 Resp: 721693
Ion Ratio Lower Upper
142 100
141 89.7 64.3 96.5









#37
1-Methylnaphthalene
Concen: 32.222 ng/ul
RT: 12.663 min Scan# 1628
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:142 Resp: 734338
Ion Ratio Lower Upper
142 100
141 91.8 69.8 104.8
116 4.0 5.6 8.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.389 ng/ul
RT: 12.810 min Scan# 1653
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:216 Resp: 418669
Ion Ratio Lower Upper
216 100
214 80.1 58.2 87.2
179 21.9 14.2 21.2#
108 17.3 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 27.246 ng/ul
RT: 12.798 min Scan# 1651
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:237 Resp: 237084
Ion Ratio Lower Upper
237 100
235 62.3 53.2 79.8
272 12.6 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#41
2,4,6-Trichlorophenol
Concen: 33.987 ng/ul
RT: 13.045 min Scan# 1693
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42



Tgt Ion:196 Resp: 272162
Ion Ratio Lower Upper
196 100
198 97.0 64.1 96.1#
200 29.6 19.9 29.9









#42
2,4,5-Trichlorophenol
Concen: 35.093 ng/ul
RT: 13.110 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:196 Resp: 303689
Ion Ratio Lower Upper
196 100
198 94.8 68.6 103.0
200 31.0 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#43
1,1'-Biphenyl
Concen: 33.762 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:154 Resp: 1009614
Ion Ratio Lower Upper
154 100
153 40.9 28.6 43.0
76 12.1 8.5 12.7





#44
2-Chloronaphthalene
Concen: 33.729 ng/ul
RT: 13.487 min Scan# 1767
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion: 162 Resp: 775290
Ion Ratio Lower Upper
162 100
164 32.6 27.5 41.3
127 40.4 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#45
2-Nitroaniline
Concen: 41.605 ng/ul
RT: 13.681 min Scan# 1801
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion: 65 Resp: 280270
Ion Ratio Lower Upper
65 100
92 65.5 77.9 116.9
138 94.8 81.5 122.3





#47
Dimethylphthalate
Concen: 35.533 ng/ul
RT: 14.069 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:163 Resp: 110101
Ion Ratio Lower Upper
163 100
194 4.1 4.1 6.1
164 9.9 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#48
2,6-Dinitrotoluene
Concen: 40.912 ng/ul
RT: 14.181 min Scan# 1886
Delta R.T. 0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:165 Resp: 233173
Ion Ratio Lower Upper
165 100
89 59.5 51.7 77.5
121 21.9 18.3 27.5





#50
Acenaphthylene
Concen: 34.354 ng/ul
RT: 14.334 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:152 Resp: 1293060
Ion Ratio Lower Upper
152 100
151 20.2 18.3 27.5
153 13.0 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#51
3-Nitroaniline
Concen: 36.828 ng/ul
RT: 14.504 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:138 Resp: 218212
Ion Ratio Lower Upper
138 100
108 12.5 15.8 23.8#
92 124.0 117.1 175.7





#52
Acenaphthene
Concen: 33.841 ng/ul
RT: 14.675 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:153 Resp: 83398
Ion Ratio Lower Upper
153 100
152 45.6 34.3 51.5
154 88.9 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#53
2,4-Dinitrophenol
Concen: 39.659 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:184 Resp: 136628
Ion Ratio Lower Upper
184 100
63 77.3 34.6 52.0#
154 73.1 57.7 86.5





#55
4-Nitrophenol
Concen: 8.726 ng/ul
RT: 14.798 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:109 Resp: 46900
Ion Ratio Lower Upper
109 100
139 111.4 61.4 92.2
65 123.0 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#56
Dibenzofuran
Concen: 34.263 ng/ul
RT: 15.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Tgt Ion:168 Resp: 1242733
Ion Ratio Lower Upper
168 100
139 38.2 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 41.883 ng/ul
RT: 14.963 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:165 Resp: 35052
Ion Ratio Lower Upper
165 100
63 47.4 29.5 44.3
182 3.1 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 35.897 ng/ul
RT: 15.234 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:232 Resp: 283244
Ion Ratio Lower Upper
232 100
131 43.5 25.4 38.2#
130 2.3 4.2 6.2#
166 30.1 28.1 42.1





#59
Diethylphthalate
Concen: 35.775 ng/ul
RT: 15.434 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:149 Resp: 1147209
Ion Ratio Lower Upper
149 100
177 22.4 22.0 33.0
150 12.5 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#61
Fluorene
Concen: 34.882 ng/ul
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:166 Resp: 1042094
Ion Ratio Lower Upper
166 100
165 97.1 70.3 105.5
167 13.8 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 33.996 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:204 Resp: 540028
Ion Ratio Lower Upper
204 100
206 32.9 27.6 41.4
141 62.6 46.6 69.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#63
4-Nitroaniline
Concen: 41.621 ng/ul m
RT: 15.669 min Scan# 2139
Delta R.T. 0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:138 Resp: 247005
Ion Ratio Lower Upper
138 100
92 55.2 41.6 62.4
108 94.3 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 39.165 ng/ul
RT: 15.728 min Scan# 2149
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:198 Resp: 207683
Ion Ratio Lower Upper
198 100
182 4.1 6.0 9.0
77 27.1 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#67
N-Nitrosodiphenylamine
Concen: 36.187 ng/ul
RT: 15.863 min Scan# 2172
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:169 Resp: 932831
Ion Ratio Lower Upper
169 100
168 64.2 48.3 72.5
167 34.8 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 35.314 ng/ul
RT: 16.551 min Scan# 2109
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:248 Resp: 344235
Ion Ratio Lower Upper
248 100
250 98.8 79.4 119.0
141 74.6 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#69
Hexachlorobenzene
Concen: 36.119 ng/ul
RT: 16.669 min Scan# 2309
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:284 Resp: 404846
Ion Ratio Lower Upper
284 100
142 33.2 19.2 28.8#
249 32.9 27.4 41.2





#70
Atrazine
Concen: 34.779 ng/ul
RT: 16.822 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:200 Resp: 370060
Ion Ratio Lower Upper
200 100
173 26.6 19.2 28.8
215 49.6 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#71
Pentachlorophenol
Concen: 35.674 ng/ul
RT: 17.010 min Scan# 2367
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:266 Resp: 270659
Ion Ratio Lower Upper
266 100
264 64.1 57.3 85.9
268 64.0 65.7 98.5





#72
Phenanthrene
Concen: 37.088 ng/u1
RT: 17.404 min Scan# 2433
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:178 Resp: 1782762
Ion Ratio Lower Upper
178 100
179 15.4 11.7 17.5
176 18.7 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#74
Anthracene
Concen: 37.158 ng/u1
RT: 17.498 min Scan# 2450
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:178 Resp: 1810873
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.1 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 34.456 ng/uL
RT: 13.410 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:216 Resp: 448599
Ion Ratio Lower Upper
216 100
214 79.7 50.4 75.6
218 49.1 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#76
Pentachlorobenzene
Concen: 34.887 ng/uL
RT: 14.934 min Scan# 2014
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:250 Resp: 442727
Ion Ratio Lower Upper
250 100
248 63.4 43.5 65.3
252 64.2 46.5 69.7





#77
Carbazole
Concen: 37.606 ng/ul
RT: 17.757 min Scan# 2494
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:167 Resp: 1656240
Ion Ratio Lower Upper
167 100
166 21.6 15.8 23.8
139 13.1 8.3 12.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#78
Di-n-butylphthalate
Concen: 37.934 ng/ul
RT: 18.316 min Scan# 2589
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:149 Resp: 2054421
Ion Ratio Lower Upper
149 100
150 9.0 9.6 14.4#
104 5.8 6.2 9.4#





#80
Fluoranthene
Concen: 37.205 ng/u1
RT: 19.363 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:202 Resp: 226840
Ion Ratio Lower Upper
202 100
101 9.6 9.6 14.4
100 7.7 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#82
Pyrene
Concen: 37.301 ng/u1
RT: 19.716 min Scan# 2827
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42



Tgt Ion:202 Resp: 2305841
Ion Ratio Lower Upper
202 100
101 10.8 10.6 15.8
100 9.4 9.8 14.6









#83
Butylbenzylphthalate
Concen: 38.966 ng/ul
RT: 20.586 min Scan# 2974
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42

Instrument :
BNA_P
ClientSampleId :
YAZ18MSD

Tgt Ion:149 Resp: 988152
Ion Ratio Lower Upper
149 100
91 73.1 71.1 106.7
206 21.7 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#84
3,3'-Dichlorobenzidine
Concen: 35.637 ng/ul
RT: 21.351 min Scan# 3105
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:252 Resp: 736543
Ion Ratio Lower Upper
252 100
254 64.7 49.8 74.6
126 10.0 4.3 6.5





#85
Benzo(a)anthracene
Concen: 37.762 ng/ul
RT: 21.427 min Scan# 3116
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:228 Resp: 232184
Ion Ratio Lower Upper
228 100
229 19.4 18.2 27.2
226 26.5 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 38.126 ng/ul
RT: 21.357 min Scan# 3106
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:149 Resp: 1454315
Ion Ratio Lower Upper
149 100
167 28.2 32.8 49.2#
279 5.1 5.6 8.4#





#87
Chrysene
Concen: 37.764 ng/ul
RT: 21.480 min Scan# 3126
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD



Tgt Ion:228 Resp: 2240260
Ion Ratio Lower Upper
228 100
226 29.1 22.7 34.1
229 19.7 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022



#89
Di-n-octyl phthalate
Concen: 37.618 ng/ul
RT: 22.310 min Scan# 3268
Delta R.T. -0.012 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42


Tgt Ion:149 Resp: 2494428





#90
Benzo(b)fluoranthene
Concen: 36.959 ng/ul
RT: 23.163 min Scan# 3411
Delta R.T. -0.006 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:252 Resp: 2323984
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.0
125 8.3 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#91
Benzo(k)fluoranthene
Concen: 39.209 ng/ul
RT: 23.215 min Scan# 3422
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:252 Resp: 2249777
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 8.0 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 38.261 ng/ul
RT: 23.815 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Instrument :
BNA_P
ClientSampleId :
YAZ18MSD




Tgt Ion:252 Resp: 223014
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 9.1 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/12/2022
Supervised By :Yogesh Patel 04/13/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 38.377 ng/ul
RT: 26.515 min Scan# 3983
Delta R.T. 0.000 min
Lab File: BP009749.D
Acq: 11 Apr 2022 13:42




Tgt Ion:276 Resp: 2339124
Ion Ratio Lower Upper
276 100
138 18.4 12.3 18.5
277 24.5 19.3 28.9





#95
 Dibenzo(a,h)anthracene
 Concen: 38.549 ng/ul
 RT: 26.539 min Scan# 3987
 Delta R.T. 0.006 min
 Lab File: BP009749.D
 Acq: 11 Apr 2022 13:42




Instrument :
 BNA_P
 ClientSampleId :
 YAZ18MSD




Tgt Ion:278 Resp: 2037390
 Ion Ratio Lower Upper
 278 100
 139 13.2 6.6 10.0
 279 23.6 18.4 27.6

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/12/2022
 Supervised By :Yogesh Patel 04/13/2022



#96
 Benzo(g,h,i)perylene
 Concen: 37.773 ng/ul
 RT: 27.309 min Scan# 4118
 Delta R.T. 0.000 min
 Lab File: BP009749.D
 Acq: 11 Apr 2022 13:42



Tgt Ion:276 Resp: 1898349
 Ion Ratio Lower Upper
 276 100
 138 17.6 11.6 17.4
 277 23.5 15.9 23.9

