

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
 Data File : BP009757.D
 Acq On : 11 Apr 2022 18:12
 Operator : CG/JU
 Sample : PB144004BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS004

Quant Time: Apr 12 01:52:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
 Supervised By :Yogesh Patel 04/13/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	90905	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.781	136	428408	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.604	164	311206	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.351	188	706301	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.433	240	790944	20.000	ng/u1	#-0.01
88) Perylene-d12	23.904	264	844738	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.387	96	12458m	6.116	ng/uL	0.00
4) Pyridine-d5	3.805	84	167463	29.819	ng/u1	0.00
7) Phenol-d5	7.122	99	248998	31.199	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.293	67	155651	32.760	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.499	132	192896	32.042	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	216079	32.433	ng/u1	0.00
21) Nitrobenzene-d5	9.128	128	101956	33.002	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.857	143	111920	33.044	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.393	165	217825	30.733	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.910	131	274359	27.178	ng/u1	0.00
46) Dimethylphthalate-d6	14.010	166	765718	31.323	ng/u1	-0.01
49) Acenaphthylene-d8	14.292	160	855295	29.532	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.781	143	149962m	33.826	ng/u1	0.00
60) Fluorene-d10	15.598	176	651120	31.662	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	137934	32.966	ng/u1	0.00
73) Anthracene-d10	17.451	188	1061064	31.377	ng/u1	-0.01
81) Pyrene-d10	19.680	212	1327842	30.739	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.745	264	1418249	32.240	ng/u1	-0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.422	88	23165	11.156	ng/uL#	7
5) Pyridine	3.822	79	173366	29.587	ng/u1#	41
6) Benzaldehyde	7.099	77	167348	34.191	ng/u1	94
8) Phenol	7.152	94	257122	32.186	ng/u1#	88
10) Bis(2-Chloroethyl)ether	7.387	93	193822	31.147	ng/u1#	75
12) 2-Chlorophenol	7.534	128	192694	31.749	ng/u1	93
13) 2-Methylphenol	8.410	108	197996	31.784	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.505	45	293213	31.877	ng/u1#	84
16) Acetophenone	8.793	105	330949	31.528	ng/u1#	81
17) N-Nitroso-di-n-propyla...	8.781	70	177525	32.621	ng/u1#	72
18) 4-Methylphenol	8.734	108	217744	32.278	ng/u1	89
19) Hexachloroethane	9.063	117	80483	31.408	ng/u1	84
22) Nitrobenzene	9.169	77	258263	32.569	ng/u1#	85
23) Isophorone	9.704	82	502109	30.728	ng/u1	97
25) 2-Nitrophenol	9.887	139	117338	32.482	ng/u1#	87
26) 2,4-Dimethylphenol	9.946	107	261085	30.455	ng/u1#	81
27) Bis(2-Chloroethoxy)met...	10.187	93	289141	30.139	ng/u1#	98
29) 2,4-Dichlorophenol	10.422	162	216194	31.464	ng/u1#	84
30) Naphthalene	10.828	128	667307	30.287	ng/u1	96
32) 4-Chloroaniline	10.934	127	267769	26.805	ng/u1	95
33) Hexachlorobutadiene	11.128	225	148359	29.979	ng/u1	94
34) Caprolactam	11.698	113	82219m	36.182	ng/u1	
35) 4-Chloro-3-methylphenol	12.051	107	268385	32.235	ng/u1#	70
36) 2-Methylnaphthalene	12.434	142	489753	30.426	ng/u1	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041122\
 Data File : BP009757.D
 Acq On : 11 Apr 2022 18:12
 Operator : CG/JU
 Sample : PB144004BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS004


Quant Time: Apr 12 01:52:46 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 07 19:16:52 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
 Supervised By :Yogesh Patel 04/13/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.651	142	498284	30.658	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.804	216	292411	30.711	ng/ul#	95
40) Hexachlorocyclopentadiene	12.793	237	175429	26.551	ng/ul	96
41) 2,4,6-Trichlorophenol	13.034	196	196186	32.265	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.104	196	217397	33.084	ng/ul#	88
43) 1,1'-Biphenyl	13.440	154	694046	30.565	ng/ul	92
44) 2-Chloronaphthalene	13.481	162	531995	30.480	ng/ul	94
45) 2-Nitroaniline	13.675	65	188780	36.906	ng/ul#	80
47) Dimethylphthalate	14.057	163	739495	31.430	ng/ul#	93
48) 2,6-Dinitrotoluene	14.169	165	153677	35.510	ng/ul	99
50) Acenaphthylene	14.322	152	888222	31.077	ng/ul	97
51) 3-Nitroaniline	14.492	138	148714	33.054	ng/ul#	85
52) Acenaphthene	14.669	153	575800	30.770	ng/ul	96
53) 2,4-Dinitrophenol	14.698	184	87527	33.459	ng/ul#	82
55) 4-Nitrophenol	14.798	109	139425	34.164	ng/ul#	57
56) Dibenzofuran	14.998	168	858590	31.175	ng/ul	96
57) 2,4-Dinitrotoluene	14.957	165	230783	36.316	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.222	232	194815	32.516	ng/ul#	86
59) Diethylphthalate	15.422	149	776544	31.891	ng/ul#	92
61) Fluorene	15.651	166	711302	31.355	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.645	204	376312	31.198	ng/ul	96
63) 4-Nitroaniline	15.657	138	174690m	38.765	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.722	198	131671	31.659	ng/ul#	89
67) N-Nitrosodiphenylamine	15.857	169	628597	31.090	ng/ul	94
68) 4-Bromophenyl-phenylether	16.539	248	234912	30.726	ng/ul	96
69) Hexachlorobenzene	16.663	284	275006	31.282	ng/ul#	92
70) Atrazine	16.810	200	261190	31.297	ng/ul	92
71) Pentachlorophenol	16.998	266	183630	30.859	ng/ul#	82
72) Phenanthrene	17.398	178	1202006	31.883	ng/ul	99
74) Anthracene	17.486	178	1207757	31.597	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.404	216	310742	30.430	ng/uL#	85
76) Pentachlorobenzene	14.922	250	301527	30.294	ng/uL	90
77) Carbazole	17.751	167	1113473	32.234	ng/ul#	95
78) Di-n-butylphthalate	18.310	149	1376081	32.396	ng/ul#	92
80) Fluoranthene	19.357	202	1529792	30.767	ng/ul#	94
82) Pyrene	19.710	202	1563141	31.007	ng/ul#	92
83) Butylbenzylphthalate	20.580	149	656366	31.738	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.345	252	537520	31.891	ng/ul#	97
85) Benzo(a)anthracene	21.416	228	1624805	32.404	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.345	149	980849	31.531	ng/ul#	81
87) Chrysene	21.474	228	1580980	32.680	ng/ul	99
89) Di-n-octyl phthalate	22.298	149	1709026	28.876	ng/ul	100
90) Benzo(b)fluoranthene	23.151	252	1723244	30.704	ng/ul	98
91) Benzo(k)fluoranthene	23.204	252	1647901	32.177	ng/ul#	97
93) Benzo(a)pyrene	23.798	252	1683201	32.354	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.504	276	1944913	35.751	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.521	278	1680949	35.634	ng/ul#	96
96) Benzo(g,h,i)perylene	27.286	276	1624929	36.225	ng/ul#	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.156 ng/uL
RT: 3.422 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion: 88 Resp: 2316
Ion Ratio Lower Upper
88 100
43 39.9 6.8 10.2
58 90.7 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#5
Pyridine
Concen: 29.587 ng/uL
RT: 3.822 min Scan# 125
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion: 79 Resp: 173366
Ion Ratio Lower Upper
79 100
52 91.2 35.5 53.3#
51 43.8 20.2 30.4#









#6
Benzaldehyde
Concen: 34.191 ng/ul
RT: 7.099 min Scan# 682
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :
BNA_P
ClientSampleId :
SLCS004

Tgt Ion: 77 Resp: 167348
Ion Ratio Lower Upper
77 100
105 92.1 77.0 115.4
106 88.7 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#8
Phenol
Concen: 32.186 ng/ul
RT: 7.152 min Scan# 691
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion: 94 Resp: 257122
Ion Ratio Lower Upper
94 100
65 34.0 19.0 28.6#
66 47.9 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.147 ng/ul
RT: 7.387 min Scan# 731
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion: 93 Resp: 19382
Ion Ratio Lower Upper
93 100
63 84.9 45.0 67.4
95 32.1 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#12
2-Chlorophenol
Concen: 31.749 ng/ul
RT: 7.534 min Scan# 756
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion:128 Resp: 192694
Ion Ratio Lower Upper
128 100
64 51.5 45.3 67.9
130 32.3 29.0 43.6









#13
2-Methylphenol
Concen: 31.784 ng/ul
RT: 8.410 min Scan# 905
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :
BNA_P
ClientSampleId :
SLCS004

Tgt Ion: 108 Resp: 197990
Ion Ratio Lower Upper
108 100
107 89.1 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.877 ng/ul
RT: 8.505 min Scan# 921
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion: 45 Resp: 293213
Ion Ratio Lower Upper
45 100
77 13.5 17.1 25.7#
79 10.6 13.2 19.8#





#16
Acetophenone
Concen: 31.528 ng/ul
RT: 8.793 min Scan# 91
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion: 105 Resp: 330949
Ion Ratio Lower Upper
105 100
77 79.2 53.4 80.2
51 31.7 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#17
N-Nitroso-di-n-propylamine
Concen: 32.621 ng/ul
RT: 8.781 min Scan# 968
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion: 70 Resp: 177525
Ion Ratio Lower Upper
70 100
42 65.4 31.8 47.8#
101 9.7 15.6 23.4#
130 19.5 13.7 20.5





#18
4-Methylphenol
Concen: 32.278 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004



Tgt Ion:108 Resp: 217744
Ion Ratio Lower Upper
108 100
107 115.7 83.6 125.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#19
Hexachloroethane
Concen: 31.408 ng/ul
RT: 9.063 min Scan# 1016
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:117 Resp: 80483
Ion Ratio Lower Upper
117 100
201 110.1 106.5 159.7
199 64.5 59.0 88.6





#22
Nitrobenzene
Concen: 32.569 ng/ul
RT: 9.169 min Scan# 1034
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion: 77 Resp: 25826
Ion Ratio Lower Upper
77 100
123 42.8 42.3 63.5
65 14.7 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#23
Isophorone
Concen: 30.728 ng/ul
RT: 9.704 min Scan# 1125
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion: 82 Resp: 502109
Ion Ratio Lower Upper
82 100
95 8.0 7.8 11.8
138 17.1 12.9 19.3









#25
2-Nitrophenol
Concen: 32.482 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :
BNA_P
ClientSampleId :
SLCS004

Tgt Ion:139 Resp: 117338
Ion Ratio Lower Upper
139 100
65 60.2 49.4 74.2
109 38.3 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#26
2,4-Dimethylphenol
Concen: 30.455 ng/ul
RT: 9.946 min Scan# 1166
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:107 Resp: 261085
Ion Ratio Lower Upper
107 100
121 47.2 34.1 51.1
122 78.5 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 30.139 ng/ul
RT: 10.187 min Scan# 1207
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion: 93 Resp: 28914
Ion Ratio Lower Upper
93 100
95 32.4 26.2 39.4
123 10.9 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#29
2,4-Dichlorophenol
Concen: 31.464 ng/ul
RT: 10.422 min Scan# 1247
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:162 Resp: 216194
Ion Ratio Lower Upper
162 100
164 65.3 37.0 55.6#
98 37.7 31.0 46.6





#30
Naphthalene
Concen: 30.287 ng/ul
RT: 10.828 min Scan# 1316
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:128 Resp: 667307
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 12.6 12.2 18.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#32
4-Chloroaniline
Concen: 26.805 ng/ul
RT: 10.934 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:127 Resp: 267769
Ion Ratio Lower Upper
127 100
129 33.5 24.6 37.0





#33
Hexachlorobutadiene
Concen: 29.979 ng/u1
RT: 11.128 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:225 Resp: 148359
Ion Ratio Lower Upper
225 100
223 62.1 53.4 80.0
227 62.5 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#34
Caprolactam
Concen: 36.182 ng/u1 m
RT: 11.698 min Scan# 1464
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:113 Resp: 82219
Ion Ratio Lower Upper
113 100
55 185.7 58.2 87.2#
56 145.8 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 32.235 ng/ul
RT: 12.051 min Scan# 1524
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:107 Resp: 26838
Ion Ratio Lower Upper
107 100
144 24.2 15.8 23.8
142 74.7 38.9 58.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#36
2-Methylnaphthalene
Concen: 30.426 ng/ul
RT: 12.434 min Scan# 1589
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion:142 Resp: 489753
Ion Ratio Lower Upper
142 100
141 88.9 64.3 96.5





#37

1-Methylnaphthalene

Concen: 30.658 ng/ul

RT: 12.651 min Scan# 1626

Delta R.T. -0.012 min

Lab File: BP009757.D

Acq: 11 Apr 2022 18:12

Instrument :

BNA_P

ClientSampleId :





SLCS004

Tgt Ion:142 Resp: 498284
 Ion Ratio Lower Upper
 142 100
 141 93.0 69.8 104.8
 116 4.8 5.6 8.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
 Supervised By :Yogesh Patel 04/13/2022



#39

1,2,4,5-Tetrachlorobenzene

Concen: 30.711 ng/ul




RT: 12.804 min Scan# 1652


Delta R.T. -0.006 min

Lab File: BP009757.D

Acq: 11 Apr 2022 18:12


Tgt Ion:216 Resp: 292411
 Ion Ratio Lower Upper
 216 100
 214 76.0 58.2 87.2
 179 21.3 14.2 21.2#
 108 16.5 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 26.551 ng/ul
RT: 12.793 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:237 Resp: 175429
Ion Ratio Lower Upper
237 100
235 62.8 53.2 79.8
272 12.2 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#41
2,4,6-Trichlorophenol
Concen: 32.265 ng/ul
RT: 13.034 min Scan# 1691
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:196 Resp: 196186
Ion Ratio Lower Upper
196 100
198 93.7 64.1 96.1
200 30.9 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 33.084 ng/ul
RT: 13.104 min Scan# 1703
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:196 Resp: 217397
Ion Ratio Lower Upper
196 100
198 94.1 68.6 103.0
200 30.2 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#43
1,1'-Biphenyl
Concen: 30.565 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:154 Resp: 694046
Ion Ratio Lower Upper
154 100
153 41.4 28.6 43.0
76 11.9 8.5 12.7





#44
2-Chloronaphthalene
Concen: 30.480 ng/ul
RT: 13.481 min Scan# 1767
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion: 162 Resp: 53199
Ion Ratio Lower Upper
162 100
164 32.3 27.5 41.3
127 41.3 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#45
2-Nitroaniline
Concen: 36.906 ng/ul
RT: 13.675 min Scan# 1800
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion: 65 Resp: 188780
Ion Ratio Lower Upper
65 100
92 64.6 77.9 116.9
138 95.2 81.5 122.3





#47
Dimethylphthalate
Concen: 31.430 ng/ul
RT: 14.057 min Scan# 1865
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:163 Resp: 73949
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 10.3 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#48
2,6-Dinitrotoluene
Concen: 35.510 ng/ul
RT: 14.169 min Scan# 1884
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion:165 Resp: 153677
Ion Ratio Lower Upper
165 100
89 64.0 51.7 77.5
121 22.6 18.3 27.5









#50
Acenaphthylene
Concen: 31.077 ng/ul
RT: 14.322 min Scan# 1910
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :
BNA_P
ClientSampleId :
SLCS004

Tgt Ion:152 Resp: 88822
Ion Ratio Lower Upper
152 100
151 20.4 18.3 27.5
153 13.6 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#51
3-Nitroaniline
Concen: 33.054 ng/ul
RT: 14.492 min Scan# 1939
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:138 Resp: 148714
Ion Ratio Lower Upper
138 100
108 12.9 15.8 23.8#
92 128.1 117.1 175.7





#52
Acenaphthene
Concen: 30.770 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:153 Resp: 575800
Ion Ratio Lower Upper
153 100
152 47.2 34.3 51.5
154 90.3 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#53
2,4-Dinitrophenol
Concen: 33.459 ng/ul
RT: 14.698 min Scan# 1974
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion:184 Resp: 87527
Ion Ratio Lower Upper
184 100
63 70.6 34.6 52.0#
154 69.9 57.7 86.5









#55
4-Nitrophenol
Concen: 34.164 ng/ul
RT: 14.798 min Scan# 1990
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :
BNA_P
ClientSampleId :
SLCS004

Tgt Ion:109 Resp: 13942
Ion Ratio Lower Upper
109 100
139 109.5 61.4 92.2
65 124.8 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#56
Dibenzofuran
Concen: 31.175 ng/ul
RT: 14.998 min Scan# 2025
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:168 Resp: 858590
Ion Ratio Lower Upper
168 100
139 39.8 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 36.316 ng/ul
RT: 14.957 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:165 Resp: 230783
Ion Ratio Lower Upper
165 100
63 46.1 29.5 44.3#
182 3.0 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 32.516 ng/ul
RT: 15.222 min Scan# 2063
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:232 Resp: 194815
Ion Ratio Lower Upper
232 100
131 45.8 25.4 38.2#
130 2.6 4.2 6.2#
166 32.0 28.1 42.1





#59
Diethylphthalate
Concen: 31.891 ng/ul
RT: 15.422 min Scan# 2097
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004



Tgt Ion:149 Resp: 776544
Ion Ratio Lower Upper
149 100
177 21.5 22.0 33.0
150 12.7 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#61
Fluorene
Concen: 31.355 ng/ul
RT: 15.651 min Scan# 2136
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:166 Resp: 711302
Ion Ratio Lower Upper
166 100
165 98.2 70.3 105.5
167 13.3 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 31.198 ng/ul
RT: 15.645 min Scan# 2135
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion: 204 Resp: 376312
Ion Ratio Lower Upper
204 100
206 32.7 27.6 41.4
141 61.4 46.6 69.8

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 04/13/2022
Supervised By : Yogesh Patel 04/13/2022




#63
4-Nitroaniline
Concen: 38.765 ng/ul m
RT: 15.657 min Scan# 2137
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion: 138 Resp: 174690
Ion Ratio Lower Upper
138 100
92 55.9 41.6 62.4
108 94.6 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 31.659 ng/ul
RT: 15.722 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004



Tgt Ion:198 Resp: 13167
Ion Ratio Lower Upper
198 100
182 4.3 6.0 9.0
77 27.7 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#67
N-Nitrosodiphenylamine
Concen: 31.090 ng/ul
RT: 15.857 min Scan# 2171
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Tgt Ion:169 Resp: 628597
Ion Ratio Lower Upper
169 100
168 66.0 48.3 72.5
167 35.2 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 30.726 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :
BNA_P
ClientSampleId :
SLCS004

Tgt Ion:248 Resp: 23491
Ion Ratio Lower Upper
248 100
250 97.1 79.4 119.0
141 78.2 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#69
Hexachlorobenzene
Concen: 31.282 ng/ul
RT: 16.663 min Scan# 2308
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:284 Resp: 275006
Ion Ratio Lower Upper
284 100
142 31.5 19.2 28.8#
249 32.3 27.4 41.2





#70
Atrazine
Concen: 31.297 ng/ul
RT: 16.810 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:200 Resp: 261190
Ion Ratio Lower Upper
200 100
173 24.9 19.2 28.8
215 47.5 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#71
Pentachlorophenol
Concen: 30.859 ng/ul
RT: 16.998 min Scan# 2365
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:266 Resp: 183630
Ion Ratio Lower Upper
266 100
264 61.8 57.3 85.9
268 61.8 65.7 98.5#





#72
Phenanthrene
Concen: 31.883 ng/u1
RT: 17.398 min Scan# 2433
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:178 Resp: 1202000
Ion Ratio Lower Upper
178 100
179 15.2 11.7 17.5
176 18.5 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#74
Anthracene
Concen: 31.597 ng/u1
RT: 17.486 min Scan# 2448
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:178 Resp: 1207757
Ion Ratio Lower Upper
178 100
179 15.6 12.3 18.5
176 18.0 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.430 ng/uL
RT: 13.404 min Scan# 1754
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:216 Resp: 310742
Ion Ratio Lower Upper
216 100
214 79.6 50.4 75.6
218 48.2 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#76
Pentachlorobenzene
Concen: 30.294 ng/uL
RT: 14.922 min Scan# 2012
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:250 Resp: 301527
Ion Ratio Lower Upper
250 100
248 63.5 43.5 65.3
252 63.5 46.5 69.7





#77
 Carbazole
 Concen: 32.234 ng/ul
 RT: 17.751 min Scan# 2493
 Delta R.T. -0.006 min
 Lab File: BP009757.D
 Acq: 11 Apr 2022 18:12




Instrument :
 BNA_P
 ClientSampleId :
 SLCS004




Tgt Ion:167 Resp: 111347
 Ion Ratio Lower Upper
 167 100
 166 21.5 15.8 23.8
 139 13.3 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
 Supervised By :Yogesh Patel 04/13/2022




#78
 Di-n-butylphthalate
 Concen: 32.396 ng/ul
 RT: 18.310 min Scan# 2588
 Delta R.T. -0.012 min
 Lab File: BP009757.D
 Acq: 11 Apr 2022 18:12




Tgt Ion:149 Resp: 1376081
 Ion Ratio Lower Upper
 149 100
 150 8.7 9.6 14.4#
 104 5.5 6.2 9.4#





#80
Fluoranthene
Concen: 30.767 ng/u1
RT: 19.357 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:202 Resp: 1529792
Ion Ratio Lower Upper
202 100
101 8.8 9.6 14.4#
100 7.1 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#82
Pyrene
Concen: 31.007 ng/u1
RT: 19.710 min Scan# 2826
Delta R.T. -0.006 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:202 Resp: 1563141
Ion Ratio Lower Upper
202 100
101 10.5 10.6 15.8#
100 8.7 9.8 14.6#





#83
Butylbenzylphthalate
Concen: 31.738 ng/ul
RT: 20.580 min Scan# 2974
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004



Tgt Ion:149 Resp: 656360
Ion Ratio Lower Upper
149 100
91 73.4 71.1 106.7
206 22.5 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#84
3,3'-Dichlorobenzidine
Concen: 31.891 ng/ul
RT: 21.345 min Scan# 3104
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:252 Resp: 537520
Ion Ratio Lower Upper
252 100
254 64.1 49.8 74.6
126 9.9 4.3 6.5





#85
Benzo(a)anthracene
Concen: 32.404 ng/ul
RT: 21.416 min Scan# 3116
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004



Tgt Ion:228 Resp: 162480
Ion Ratio Lower Upper
228 100
229 19.8 18.2 27.2
226 26.3 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 31.531 ng/ul
RT: 21.345 min Scan# 3104
Delta R.T. -0.018 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12


Tgt Ion:149 Resp: 980849
Ion Ratio Lower Upper
149 100
167 28.0 32.8 49.2#
279 4.6 5.6 8.4#





#87
Chrysene
Concen: 32.680 ng/u1
RT: 21.474 min Scan# 3126
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004



Tgt Ion:228 Resp: 1580980
Ion Ratio Lower Upper
228 100
226 29.0 22.7 34.1
229 19.4 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#89
Di-n-octyl phthalate
Concen: 28.876 ng/u1
RT: 22.298 min Scan# 3266
Delta R.T. -0.023 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Tgt Ion:149 Resp: 1709026






#90
Benzo(b)fluoranthene
Concen: 30.704 ng/ul
RT: 23.151 min Scan# 3411
Delta R.T. -0.018 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :

BNA_P

ClientSampleId :

SLCS004






Tgt Ion:252 Resp: 1723244
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 8.0 7.3 10.9


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#91
Benzo(k)fluoranthene
Concen: 32.177 ng/ul
RT: 23.204 min Scan# 3420
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion:252 Resp: 1647901
Ion Ratio Lower Upper
252 100
253 21.3 17.9 26.9
125 8.2 5.1 7.7#






#93
Benzo(a)pyrene
Concen: 32.354 ng/ul
RT: 23.798 min Scan# 31
Delta R.T. -0.018 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12

Instrument :

BNA_P

ClientSampleId :

SLCS004






Tgt Ion:252 Resp: 168320
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 8.6 6.4 9.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 35.751 ng/ul
RT: 26.504 min Scan# 3981
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Tgt Ion:276 Resp: 1944913
Ion Ratio Lower Upper
276 100
138 20.1 12.3 18.5#
277 25.2 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 35.634 ng/ul
RT: 26.521 min Scan# 3983
Delta R.T. -0.012 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12




Instrument :
BNA_P
ClientSampleId :
SLCS004




Tgt Ion:278 Resp: 1680949
Ion Ratio Lower Upper
278 100
139 12.1 6.6 10.0
279 24.1 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/13/2022
Supervised By :Yogesh Patel 04/13/2022



#96
Benzo(g,h,i)perylene
Concen: 36.225 ng/ul
RT: 27.286 min Scan# 4114
Delta R.T. -0.023 min
Lab File: BP009757.D
Acq: 11 Apr 2022 18:12



Tgt Ion:276 Resp: 1624929
Ion Ratio Lower Upper
276 100
138 17.4 11.6 17.4
277 24.3 15.9 23.9

