

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041223\
 Data File : BP014668.D
 Acq On : 12 Apr 2023 11:23
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673

Quant Time: Apr 13 00:48:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 01:21:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	111770	20.000	ng/ul	0.00
20) Naphthalene-d8	10.793	136	460691	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.628	164	271414	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.386	188	522536	20.000	ng/ul	0.00
79) Chrysene-d12	21.480	240	494864	20.000	ng/ul	0.00
88) Perylene-d12	23.980	264	581491	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	26895	9.331	ng/uL	0.00
4) Pyridine-d5	3.787	84	172697	23.381	ng/ul	0.00
7) Phenol-d5	7.140	99	203392	19.732	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	149281	22.572	ng/ul	0.00
11) 2-Chlorophenol-d4	7.505	132	160315	21.518	ng/ul	0.00
15) 4-Methylphenol-d8	8.693	113	166274	19.997	ng/ul	0.00
21) Nitrobenzene-d5	9.152	128	78346	21.066	ng/ul	0.00
24) 2-Nitrophenol-d4	9.875	143	86597	20.335	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	152898	19.714	ng/ul	0.00
31) 4-Chloroaniline-d4	10.940	131	191945	18.789	ng/ul	0.00
46) Dimethylphthalate-d6	14.034	166	421173	19.138	ng/ul	0.00
49) Acenaphthylene-d8	14.322	160	485696	19.890	ng/ul	0.00
54) 4-Nitrophenol-d4	14.851	143	51063	15.359	ng/ul	0.00
60) Fluorene-d10	15.622	176	350317	19.089	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	68040	17.696	ng/ul	0.00
73) Anthracene-d10	17.486	188	500507	19.792	ng/ul	0.00
81) Pyrene-d10	19.722	212	543156	16.340	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.815	264	622876	20.162	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	30451	9.614	ng/uL	88
5) Pyridine	3.811	79	181318	23.458	ng/ul#	92
6) Benzaldehyde	7.116	77	138166	26.966	ng/ul	98
8) Phenol	7.163	94	217677	20.357	ng/ul	91
10) Bis(2-Chloroethyl)ether	7.393	93	180104	20.959	ng/ul	94
12) 2-Chlorophenol	7.534	128	164728	21.243	ng/ul	91
13) 2-Methylphenol	8.422	108	156363	19.608	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.499	45	271786	24.089	ng/ul	98
16) Acetophenone	8.810	105	282639	20.585	ng/ul	93
17) N-Nitroso-di-n-propyla...	8.787	70	164822	22.221	ng/ul#	89
18) 4-Methylphenol	8.757	108	171896	19.628	ng/ul	98
19) Hexachloroethane	9.052	117	78896	23.286	ng/ul	88
22) Nitrobenzene	9.193	77	243286	22.771	ng/ul	97
23) Isophorone	9.716	82	407253	20.679	ng/ul	98
25) 2-Nitrophenol	9.904	139	91068	20.260	ng/ul	88
26) 2,4-Dimethylphenol	9.963	107	217899	21.658	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.199	93	234238	19.838	ng/ul	98
29) 2,4-Dichlorophenol	10.446	162	148383	19.441	ng/ul	94
30) Naphthalene	10.840	128	516787	20.139	ng/ul	99
32) 4-Chloroaniline	10.963	127	192758	18.629	ng/ul	98
33) Hexachlorobutadiene	11.116	225	118688	19.628	ng/ul	98
34) Caprolactam	11.757	113	36840	16.338	ng/ul	78
35) 4-Chloro-3-methylphenol	12.087	107	169936	18.101	ng/ul	99
36) 2-Methylnaphthalene	12.451	142	331433	19.058	ng/ul	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041223\
 Data File : BP014668.D
 Acq On : 12 Apr 2023 11:23
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673

Quant Time: Apr 13 00:48:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 01:21:43 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	332009	19.250	ng/ul	96
39) 1,2,4,5-Tetrachloroben...	12.816	216	190364	19.672	ng/ul	97
40) Hexachlorocyclopentadiene	12.787	237	114290	18.238	ng/ul	97
41) 2,4,6-Trichlorophenol	13.063	196	115932	19.360	ng/ul	98
42) 2,4,5-Trichlorophenol	13.140	196	120460	18.896	ng/ul	99
43) 1,1'-Biphenyl	13.457	154	447364	20.402	ng/ul	98
44) 2-Chloronaphthalene	13.504	162	351640	20.282	ng/ul	98
45) 2-Nitroaniline	13.722	65	116060	21.615	ng/ul	91
47) Dimethylphthalate	14.081	163	419858	19.081	ng/ul	98
48) 2,6-Dinitrotoluene	14.210	165	81782	19.024	ng/ul	91
50) Acenaphthylene	14.351	152	530095	20.384	ng/ul	100
51) 3-Nitroaniline	14.551	138	70344	18.265	ng/ul	96
52) Acenaphthene	14.692	153	368300	20.041	ng/ul	99
53) 2,4-Dinitrophenol	14.763	184	57210	20.551	ng/ul#	77
55) 4-Nitrophenol	14.869	109	56594	17.248	ng/ul	93
56) Dibenzofuran	15.028	168	501967	19.431	ng/ul	97
57) 2,4-Dinitrotoluene	15.004	165	118060	19.361	ng/ul#	83
58) 2,3,4,6-Tetrachlorophenol	15.257	232	102702	17.925	ng/ul	94
59) Diethylphthalate	15.445	149	422111	19.305	ng/ul	99
61) Fluorene	15.675	166	402639	19.229	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.669	204	210224	18.687	ng/ul	99
63) 4-Nitroaniline	15.722	138	55931	17.874	ng/ul	87
66) 4,6-Dinitro-2-methylph...	15.769	198	65860	17.675	ng/ul#	89
67) N-Nitrosodiphenylamine	15.886	169	334133	20.552	ng/ul	99
68) 4-Bromophenyl-phenylether	16.569	248	124960	19.218	ng/ul	96
69) Hexachlorobenzene	16.686	284	142985	19.144	ng/ul	95
70) Atrazine	16.845	200	118259	19.982	ng/ul	98
71) Pentachlorophenol	17.039	266	73830	17.063	ng/ul	96
72) Phenanthrene	17.433	178	577620	19.869	ng/ul	99
74) Anthracene	17.522	178	583906	19.929	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.422	216	194669	21.164	ng/uL	99
76) Pentachlorobenzene	14.945	250	184885	20.145	ng/uL	98
77) Carbazole	17.804	167	483221	19.708	ng/ul	100
78) Di-n-butylphthalate	18.328	149	649559	21.069	ng/ul	99
80) Fluoranthene	19.392	202	638134	16.237	ng/ul	97
82) Pyrene	19.751	202	665052	16.208	ng/ul	98
83) Butylbenzylphthalate	20.610	149	274391	19.414	ng/ul	93
84) 3,3'-Dichlorobenzidine	21.392	252	215683	19.707	ng/ul	98
85) Benzo(a)anthracene	21.463	228	699207	19.939	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.363	149	412571	21.347	ng/ul	98
87) Chrysene	21.516	228	671978	20.294	ng/ul	100
89) Di-n-octyl phthalate	22.316	149	725530	19.227	ng/ul	100
90) Benzo(b)fluoranthene	23.210	252	749102	19.074	ng/ul	99
91) Benzo(k)fluoranthene	23.263	252	766132	20.086	ng/ul	99
93) Benzo(a)pyrene	23.868	252	680599	20.134	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.598	276	937279	20.153	ng/ul	98
95) Dibenzo(a,h)anthracene	26.609	278	772488	19.839	ng/ul	97
96) Benzo(g,h,i)perylene	27.404	276	763862	20.150	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041223\
 Data File : BP014668.D
 Acq On : 12 Apr 2023 11:23
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673

Quant Time: Apr 13 00:48:09 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 01:21:43 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.614 ng/uL
RT: 3.393 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion: 88	Resp: 30451
Ion Ratio	Lower Upper
88	100
43	36.6 24.4 36.6
58	81.9 57.4 86.0



#5
Pyridine
Concen: 23.458 ng/uL
RT: 3.811 min Scan# 106
Delta R.T. 0.000 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion: 79	Resp: 181318
Ion Ratio	Lower Upper
79	100
52	82.0 61.4 92.2
51	43.2 28.5 42.7









#6
Benzaldehyde
Concen: 26.966 ng/ul
RT: 7.116 min Scan# 60
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion: 77 Resp: 138166
Ion Ratio Lower Upper
77 100
105 87.3 70.6 105.8
106 80.0 66.5 99.7



#8
Phenol
Concen: 20.357 ng/ul
RT: 7.163 min Scan# 676
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion: 94 Resp: 217677
Ion Ratio Lower Upper
94 100
65 34.8 25.2 37.8
66 52.7 36.2 54.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 20.959 ng/ul
 RT: 7.393 min Scan# 715
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673


Tgt Ion: 93 Resp: 180104
 Ion Ratio Lower Upper
 93 100
 63 85.8 63.2 94.8
 95 33.0 26.0 39.0



#12
 2-Chlorophenol
 Concen: 21.243 ng/ul
 RT: 7.534 min Scan# 739
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23

Tgt Ion:128 Resp: 164728
 Ion Ratio Lower Upper
 128 100
 64 60.9 41.4 62.0
 130 30.5 24.1 36.1





#13

2-Methylphenol

Concen: 19.608 ng/ul

RT: 8.422 min Scan# 890

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23

Instrument :

BNA_P

ClientSampleId :





SSTD020673

Tgt Ion:108 Resp: 156363

Ion Ratio Lower Upper

108 100

107 92.9 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 24.089 ng/ul

RT: 8.499 min Scan# 903

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23




Tgt Ion: 45 Resp: 271786


Ion Ratio Lower Upper

45 100

77 15.0 12.8 19.2

79 11.6 9.6 14.4









#16
Acetophenone
Concen: 20.585 ng/ul
RT: 8.810 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:105 Resp: 282639
Ion Ratio Lower Upper
105 100
77 90.1 67.3 100.9
51 33.0 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 22.221 ng/ul
RT: 8.787 min Scan# 952
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion: 70 Resp: 164822
Ion Ratio Lower Upper
70 100
42 62.7 41.1 61.7#
101 8.8 7.0 10.6
130 17.0 14.6 21.8









#18
4-Methylphenol
Concen: 19.628 ng/ul
RT: 8.757 min Scan# 947
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:108 Resp: 171896
Ion Ratio Lower Upper
108 100
107 113.3 89.0 133.6



#19
Hexachloroethane
Concen: 23.286 ng/ul
RT: 9.052 min Scan# 997
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:117 Resp: 78896
Ion Ratio Lower Upper
117 100
201 98.7 90.2 135.4
199 59.9 54.2 81.2









#22
Nitrobenzene
Concen: 22.771 ng/ul
RT: 9.193 min Scan# 1102
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion: 77 Resp: 243286
Ion Ratio Lower Upper
77 100
123 36.1 30.3 45.5
65 14.7 11.0 16.4



#23
Isophorone
Concen: 20.679 ng/ul
RT: 9.716 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion: 82 Resp: 407253
Ion Ratio Lower Upper
82 100
95 7.7 6.3 9.5
138 14.0 12.2 18.2









#25
2-Nitrophenol
Concen: 20.260 ng/ul
RT: 9.904 min Scan# 1142
Delta R.T. -0.012 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:139 Resp: 91068
Ion Ratio Lower Upper
139 100
65 71.2 49.7 74.5
109 46.7 31.1 46.7



#26
2,4-Dimethylphenol
Concen: 21.658 ng/ul
RT: 9.963 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:107 Resp: 217899
Ion Ratio Lower Upper
107 100
121 43.4 37.0 55.6
122 73.9 59.8 89.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.838 ng/ul
 RT: 10.199 min Scan# 1192
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673


Tgt Ion: 93 Resp: 234238
 Ion Ratio Lower Upper
 93 100
 95 31.4 25.8 38.8
 123 9.4 8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 19.441 ng/ul
 RT: 10.446 min Scan# 1234
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23

Tgt Ion: 162 Resp: 148383
 Ion Ratio Lower Upper
 162 100
 164 65.8 50.1 75.1
 98 43.9 30.6 45.8









#30
Naphthalene
Concen: 20.139 ng/ul
RT: 10.840 min Scan# 1301
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:128 Resp: 516787
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.3 11.0 16.6



#32
4-Chloroaniline
Concen: 18.629 ng/ul
RT: 10.963 min Scan# 1322
Delta R.T. -0.012 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:127 Resp: 192758
Ion Ratio Lower Upper
127 100
129 33.0 25.4 38.0









#33
Hexachlorobutadiene
Concen: 19.628 ng/ul
RT: 11.116 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:225 Resp: 118688
Ion Ratio Lower Upper
225 100
223 62.5 49.3 73.9
227 64.4 49.6 74.4



#34
Caprolactam
Concen: 16.338 ng/ul
RT: 11.757 min Scan# 1457
Delta R.T. -0.018 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:113 Resp: 36840
Ion Ratio Lower Upper
113 100
55 194.0 131.2 196.8
56 151.6 101.1 151.7









#35
4-Chloro-3-methylphenol
Concen: 18.101 ng/ul
RT: 12.087 min Scan# 1513
Delta R.T. -0.012 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:107 Resp: 169936
Ion Ratio Lower Upper
107 100
144 21.1 17.3 25.9
142 67.8 54.6 81.8



#36
2-Methylnaphthalene
Concen: 19.058 ng/ul
RT: 12.451 min Scan# 1575
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:142 Resp: 331433
Ion Ratio Lower Upper
142 100
141 90.4 71.6 107.4









#37
1-Methylnaphthalene
Concen: 19.250 ng/ul
RT: 12.669 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:142 Resp: 332009
Ion Ratio Lower Upper
142 100
141 91.4 76.6 114.8
116 4.7 3.6 5.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 19.672 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:216 Resp: 190364
Ion Ratio Lower Upper
216 100
214 78.9 61.0 91.6
179 22.9 16.6 24.8
108 18.7 13.8 20.6









#40
Hexachlorocyclopentadiene
Concen: 18.238 ng/ul
RT: 12.787 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:237 Resp: 114290
Ion Ratio Lower Upper
237 100
235 61.4 51.0 76.4
272 12.6 10.3 15.5



#41
2,4,6-Trichlorophenol
Concen: 19.360 ng/ul
RT: 13.063 min Scan# 1679
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:196 Resp: 115932
Ion Ratio Lower Upper
196 100
198 98.1 76.6 114.8
200 30.5 25.4 38.2









#42
2,4,5-Trichlorophenol
Concen: 18.896 ng/ul
RT: 13.140 min Scan# 1692
Delta R.T. -0.012 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:196 Resp: 120460
Ion Ratio Lower Upper
196 100
198 99.0 79.8 119.8
200 32.0 25.6 38.4



#43
1,1'-Biphenyl
Concen: 20.402 ng/ul
RT: 13.457 min Scan# 1746
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:154 Resp: 447364
Ion Ratio Lower Upper
154 100
153 41.0 32.4 48.6
76 15.0 10.7 16.1









#44
2-Chloronaphthalene
Concen: 20.282 ng/ul
RT: 13.504 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:162 Resp: 351640
Ion Ratio Lower Upper
162 100
164 33.3 26.1 39.1
127 43.8 33.7 50.5



#45
2-Nitroaniline
Concen: 21.615 ng/ul
RT: 13.722 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion: 65 Resp: 116060
Ion Ratio Lower Upper
65 100
92 64.3 53.7 80.5
138 77.8 71.7 107.5





#47

Dimethylphthalate

Concen: 19.081 ng/ul

RT: 14.081 min Scan# 1852

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD020673





Tgt Ion:163 Resp: 419858

Ion Ratio Lower Upper

163 100

194 4.2 3.1 4.7

164 10.6 7.8 11.8



#48

2,6-Dinitrotoluene

Concen: 19.024 ng/ul

RT: 14.210 min Scan# 1874

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23




Tgt Ion:165 Resp: 81782


Ion Ratio Lower Upper

165 100

89 75.2 52.7 79.1

121 23.6 17.9 26.9









#50
Acenaphthylene
Concen: 20.384 ng/ul
RT: 14.351 min Scan# 1898
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:152 Resp: 530095
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4
153 12.8 10.2 15.4



#51
3-Nitroaniline
Concen: 18.265 ng/ul
RT: 14.551 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:138 Resp: 70344
Ion Ratio Lower Upper
138 100
108 13.4 9.9 14.9
92 139.7 107.6 161.4









#52
Acenaphthene
Concen: 20.041 ng/ul
RT: 14.692 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:153 Resp: 368300
Ion Ratio Lower Upper
153 100
152 46.6 38.2 57.4
154 86.5 70.2 105.2



#53
2,4-Dinitrophenol
Concen: 20.551 ng/ul
RT: 14.763 min Scan# 1968
Delta R.T. -0.012 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:184 Resp: 57210
Ion Ratio Lower Upper
184 100
63 96.6 57.1 85.7#
154 81.8 55.3 82.9









#55
4-Nitrophenol
Concen: 17.248 ng/ul
RT: 14.869 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:109 Resp: 56594
Ion Ratio Lower Upper
109 100
139 91.6 77.4 116.2
65 130.7 97.3 145.9



#56
Dibenzofuran
Concen: 19.431 ng/ul
RT: 15.028 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:168 Resp: 501967
Ion Ratio Lower Upper
168 100
139 42.0 32.1 48.1









#57
2,4-Dinitrotoluene
Concen: 19.361 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:165 Resp: 118060
Ion Ratio Lower Upper
165 100
63 64.8 41.6 62.4#
182 2.9 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 17.925 ng/ul
RT: 15.257 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:232 Resp: 102702
Ion Ratio Lower Upper
232 100
131 47.6 35.0 52.4
130 2.5 1.9 2.9
166 32.9 24.1 36.1









#59
Diethylphthalate
Concen: 19.305 ng/ul
RT: 15.445 min Scan# 2084
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:149 Resp: 422111
Ion Ratio Lower Upper
149 100
177 22.0 18.2 27.2
150 12.7 9.8 14.8



#61
Fluorene
Concen: 19.229 ng/ul
RT: 15.675 min Scan# 2123
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:166 Resp: 402639
Ion Ratio Lower Upper
166 100
165 98.0 79.2 118.8
167 13.8 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 18.687 ng/ul
RT: 15.669 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:204 Resp: 210224
Ion Ratio Lower Upper
204 100
206 33.7 26.9 40.3
141 65.7 51.3 76.9



#63
4-Nitroaniline
Concen: 17.874 ng/ul
RT: 15.722 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:138 Resp: 55931
Ion Ratio Lower Upper
138 100
92 65.9 47.8 71.6
108 119.7 82.3 123.5









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.675 ng/ul
 RT: 15.769 min Scan# 2139
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673


Tgt Ion:198 Resp: 65860
 Ion Ratio Lower Upper
 198 100
 182 5.1 3.5 5.3
 77 37.2 24.4 36.6#



#67
 N-Nitrosodiphenylamine
 Concen: 20.552 ng/ul
 RT: 15.886 min Scan# 2159
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23

Tgt Ion:169 Resp: 334133
 Ion Ratio Lower Upper
 169 100
 168 66.5 54.2 81.2
 167 37.0 29.4 44.2









#68
4-Bromophenyl-phenylether
Concen: 19.218 ng/ul
RT: 16.569 min Scan# 2275
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:248 Resp: 124960
Ion Ratio Lower Upper
248 100
250 99.4 78.1 117.1
141 88.5 66.0 99.0



#69
Hexachlorobenzene
Concen: 19.144 ng/ul
RT: 16.686 min Scan# 2295
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:284 Resp: 142985
Ion Ratio Lower Upper
284 100
142 36.4 26.6 39.8
249 34.6 25.7 38.5





#70

Atrazine

Concen: 19.982 ng/ul

RT: 16.845 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD020673





Tgt Ion:200 Resp: 118259

Ion Ratio Lower Upper

200 100

173 25.1 20.6 30.8

215 48.0 37.2 55.8



#71

Pentachlorophenol

Concen: 17.063 ng/ul

RT: 17.039 min Scan# 2355

Delta R.T. -0.012 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23




Tgt Ion:266 Resp: 73830


Ion Ratio Lower Upper

266 100

264 66.0 51.0 76.4

268 68.8 51.4 77.2









#72
Phenanthrene
Concen: 19.869 ng/ul
RT: 17.433 min Scan# 2422
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:178 Resp: 577620
Ion Ratio Lower Upper
178 100
179 14.9 12.2 18.4
176 18.6 15.3 22.9



#74
Anthracene
Concen: 19.929 ng/ul
RT: 17.522 min Scan# 2437
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:178 Resp: 583906
Ion Ratio Lower Upper
178 100
179 15.3 12.2 18.4
176 18.8 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.164 ng/uL

RT: 13.422 min Scan# 1740

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD020673





Tgt Ion:216 Resp: 194669

Ion Ratio Lower Upper

216 100

214 76.3 61.0 91.6

218 47.5 38.7 58.1



#76

Pentachlorobenzene

Concen: 20.145 ng/uL

RT: 14.945 min Scan# 1999

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23




Tgt Ion:250 Resp: 184885


Ion Ratio Lower Upper

250 100

248 60.0 49.0 73.6

252 62.3 50.7 76.1









#77
 Carbazole
 Concen: 19.708 ng/ul
 RT: 17.804 min Scan# 2485
 Delta R.T. 0.000 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673


Tgt Ion:167 Resp: 483221
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.4 26.2
 139 14.1 11.3 16.9



#78
 Di-n-butylphthalate
 Concen: 21.069 ng/ul
 RT: 18.328 min Scan# 2574
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23

Tgt Ion:149 Resp: 649559
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.0
 104 6.7 4.8 7.2









#80
 Fluoranthene
 Concen: 16.237 ng/ul
 RT: 19.392 min Scan# 2755
 Delta R.T. -0.006 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020673

Tgt Ion:202 Resp: 638134




Ion	Ratio	Lower	Upper
202	100		
101	11.1	8.2	12.2
100	9.4	6.4	9.6




#82
 Pyrene
 Concen: 16.208 ng/ul
 RT: 19.751 min Scan# 2816
 Delta R.T. 0.000 min
 Lab File: BP014668.D
 Acq: 12 Apr 2023 11:23

Tgt Ion:202 Resp: 665052

Ion	Ratio	Lower	Upper
202	100		
101	12.7	9.7	14.5
100	10.7	8.0	12.0





#83

Butylbenzylphthalate

Concen: 19.414 ng/ul

RT: 20.610 min Scan# 2962

Delta R.T. 0.000 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD020673





Tgt Ion:149 Resp: 274391

Ion Ratio Lower Upper

149 100

91 84.3 61.8 92.8

206 21.6 17.8 26.6



#84

3,3'-Dichlorobenzidine

Concen: 19.707 ng/ul

RT: 21.392 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23




Tgt Ion:252 Resp: 215683


Ion Ratio Lower Upper

252 100

254 62.5 51.4 77.0

126 10.6 7.8 11.6





#85

Benzo(a)anthracene

Concen: 19.939 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD020673





Tgt Ion:228 Resp: 699207

Ion Ratio Lower Upper

228 100

229 18.7 15.6 23.4

226 26.7 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 21.347 ng/ul

RT: 21.363 min Scan# 3090

Delta R.T. 0.000 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23




Tgt Ion:149 Resp: 412571


Ion Ratio Lower Upper

149 100

167 26.2 21.9 32.9

279 4.6 3.7 5.5









#87
Chrysene
Concen: 20.294 ng/ul
RT: 21.516 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:228 Resp: 671978
Ion Ratio Lower Upper
228 100
226 29.8 23.8 35.8
229 19.7 16.1 24.1



#89
Di-n-octyl phthalate
Concen: 19.227 ng/ul
RT: 22.316 min Scan# 3252
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:149 Resp: 725530





#90

Benzo(b)fluoranthene

Concen: 19.074 ng/ul

RT: 23.210 min Scan# 3404

Delta R.T. -0.006 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD020673





Tgt Ion:252 Resp: 749102

Ion Ratio Lower Upper

252 100

253 22.8 18.0 27.0

125 8.8 6.8 10.2



#91

Benzo(k)fluoranthene

Concen: 20.086 ng/ul

RT: 23.263 min Scan# 3413

Delta R.T. 0.000 min

Lab File: BP014668.D

Acq: 12 Apr 2023 11:23




Tgt Ion:252 Resp: 766132


Ion Ratio Lower Upper

252 100

253 22.2 17.3 25.9

125 8.3 6.9 10.3









#93
Benzo(a)pyrene
Concen: 20.134 ng/ul
RT: 23.868 min Scan# 3516
Delta R.T. 0.000 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD020673


Tgt Ion:252 Resp: 680599
Ion Ratio Lower Upper
252 100
253 21.6 16.9 25.3
125 9.9 6.8 10.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 20.153 ng/ul
RT: 26.598 min Scan# 3980
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:276 Resp: 937279
Ion Ratio Lower Upper
276 100
138 20.0 14.9 22.3
277 24.5 20.3 30.5









#95
Dibenzo(a,h)anthracene
Concen: 19.839 ng/ul
RT: 26.609 min Scan# 3982
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Instrument :
BNA_P
ClientSampleId :
SSTD020673

Tgt Ion:278 Resp: 772488
Ion Ratio Lower Upper
278 100
139 14.0 10.4 15.6
279 24.8 18.9 28.3



#96
Benzo(g,h,i)perylene
Concen: 20.150 ng/ul
RT: 27.404 min Scan# 4117
Delta R.T. -0.006 min
Lab File: BP014668.D
Acq: 12 Apr 2023 11:23

Tgt Ion:276 Resp: 763862
Ion Ratio Lower Upper
276 100
138 19.1 13.0 19.4
277 24.3 18.4 27.6

