

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041223\
 Data File : BP014682.D
 Acq On : 12 Apr 2023 20:07
 Operator : CG/JU
 Sample : PB152055BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS055

Quant Time: Apr 13 00:51:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 01:21:43 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	97602	20.000	ng/u1	0.00
20) Naphthalene-d8	10.792	136	403342	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.627	164	234318	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.386	188	478793	20.000	ng/u1	0.00
79) Chrysene-d12	21.480	240	508741	20.000	ng/u1	0.00
88) Perylene-d12	23.980	264	571310	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.357	96	17765	7.058	ng/uL	0.00
4) Pyridine-d5	3.793	84	213165	33.050	ng/u1	0.00
7) Phenol-d5	7.140	99	287429	31.932	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.298	67	200833	34.774	ng/u1	0.00
11) 2-Chlorophenol-d4	7.504	132	214931	33.036	ng/u1	0.00
15) 4-Methylphenol-d8	8.692	113	219764	30.267	ng/u1	0.00
21) Nitrobenzene-d5	9.151	128	104932	32.226	ng/u1	0.00
24) 2-Nitrophenol-d4	9.875	143	116456	31.234	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.416	165	201229	29.635	ng/u1	0.00
31) 4-Chloroaniline-d4	10.945	131	188132	21.034	ng/u1	0.00
46) Dimethylphthalate-d6	14.033	166	556825	29.308	ng/u1	0.00
49) Acenaphthylene-d8	14.322	160	641386	30.424	ng/u1	0.00
54) 4-Nitrophenol-d4	14.851	143	87041	30.326	ng/u1	0.00
60) Fluorene-d10	15.621	176	477408	30.132	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.757	200	104239	29.588	ng/u1	0.00
73) Anthracene-d10	17.492	188	709423	30.616	ng/u1	0.00
81) Pyrene-d10	19.721	212	859433	25.150	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.815	264	955513	31.480	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	42049	15.203	ng/uL	95
5) Pyridine	3.810	79	251487	37.259	ng/u1	96
6) Benzaldehyde	7.116	77	77769	17.382	ng/u1	99
8) Phenol	7.169	94	329837	35.323	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.393	93	264150	35.202	ng/u1	95
12) 2-Chlorophenol	7.534	128	241563	35.674	ng/u1	91
13) 2-Methylphenol	8.428	108	234106	33.618	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.504	45	398659	40.464	ng/u1	98
16) Acetophenone	8.810	105	402252	33.550	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.787	70	229586	35.445	ng/u1	93
18) 4-Methylphenol	8.757	108	253480	33.144	ng/u1	98
19) Hexachloroethane	9.051	117	111275	37.610	ng/u1	90
22) Nitrobenzene	9.192	77	345212	36.904	ng/u1	97
23) Isophorone	9.716	82	576576	33.439	ng/u1	97
25) 2-Nitrophenol	9.910	139	132685	33.716	ng/u1	92
26) 2,4-Dimethylphenol	9.963	107	281896	32.003	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.198	93	338516	32.745	ng/u1	98
29) 2,4-Dichlorophenol	10.445	162	213125	31.893	ng/u1	92
30) Naphthalene	10.845	128	738440	32.868	ng/u1	99
32) 4-Chloroaniline	10.969	127	158129	17.455	ng/u1	97
33) Hexachlorobutadiene	11.116	225	165596	31.279	ng/u1	98
35) 4-Chloro-3-methylphenol	12.092	107	255111	31.038	ng/u1	99
36) 2-Methylnaphthalene	12.451	142	481005	31.592	ng/u1	100
37) 1-Methylnaphthalene	12.669	142	484074	32.057	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041223\
 Data File : BP014682.D
 Acq On : 12 Apr 2023 20:07
 Operator : CG/JU
 Sample : PB152055BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS055

Quant Time: Apr 13 00:51:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 11 01:21:43 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) 1,2,4,5-Tetrachloroben...	12.816	216	269076	32.208	ng/ul	98
40) Hexachlorocyclopentadiene	12.786	237	152542	28.195	ng/ul	96
41) 2,4,6-Trichlorophenol	13.063	196	164683	31.855	ng/ul	99
42) 2,4,5-Trichlorophenol	13.139	196	178591	32.450	ng/ul	94
43) 1,1'-Biphenyl	13.457	154	631754	33.372	ng/ul	99
44) 2-Chloronaphthalene	13.504	162	506576	33.843	ng/ul	99
45) 2-Nitroaniline	13.722	65	172139	37.135	ng/ul	90
47) Dimethylphthalate	14.080	163	611936	32.214	ng/ul	99
48) 2,6-Dinitrotoluene	14.210	165	125125	33.714	ng/ul	89
50) Acenaphthylene	14.351	152	754499	33.606	ng/ul	100
51) 3-Nitroaniline	14.551	138	103706	31.191	ng/ul	93
52) Acenaphthene	14.686	153	540532	34.070	ng/ul	99
53) 2,4-Dinitrophenol	14.763	184	75922	31.590	ng/ul	86
55) 4-Nitrophenol	14.863	109	98791	34.875	ng/ul	90
56) Dibenzofuran	15.027	168	725942	32.549	ng/ul	96
57) 2,4-Dinitrotoluene	15.004	165	186265	35.382	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.257	232	149749	30.274	ng/ul#	94
59) Diethylphthalate	15.445	149	616031	32.634	ng/ul	98
61) Fluorene	15.680	166	606870	33.571	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.669	204	309563	31.873	ng/ul	98
63) 4-Nitroaniline	15.721	138	95319	35.284	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.769	198	110853	32.468	ng/ul	96
67) N-Nitrosodiphenylamine	15.886	169	488552	32.796	ng/ul	98
68) 4-Bromophenyl-phenylether	16.568	248	183498	30.799	ng/ul	99
69) Hexachlorobenzene	16.686	284	213696	31.226	ng/ul	96
70) Atrazine	16.851	200	15812	2.916	ng/ul	94
71) Pentachlorophenol	17.039	266	114938	28.991	ng/ul	98
72) Phenanthrene	17.433	178	908167	34.094	ng/ul	100
74) Anthracene	17.521	178	914512	34.064	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.422	216	270726	32.122	ng/uL	98
76) Pentachlorobenzene	14.945	250	273190	32.486	ng/uL	98
77) Carbazole	17.804	167	791679	35.238	ng/ul	99
78) Di-n-butylphthalate	18.327	149	1016517	35.983	ng/ul	99
80) Fluoranthene	19.398	202	1088126	26.931	ng/ul	99
82) Pyrene	19.751	202	1158840	27.472	ng/ul	98
83) Butylbenzylphthalate	20.609	149	491633	33.836	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.398	252	269210	23.927	ng/ul	98
85) Benzo(a)anthracene	21.462	228	1203955	33.396	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.362	149	723330	36.405	ng/ul	100
87) Chrysene	21.521	228	1162215	34.142	ng/ul	99
89) Di-n-octyl phthalate	22.315	149	1312302	35.396	ng/ul	100
90) Benzo(b)fluoranthene	23.209	252	1295078	33.564	ng/ul	100
91) Benzo(k)fluoranthene	23.262	252	1265642	33.773	ng/ul	100
93) Benzo(a)pyrene	23.868	252	1148087	34.568	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.597	276	1553551	34.000	ng/ul	97
95) Dibenzo(a,h)anthracene	26.615	278	1287670	33.659	ng/ul	99
96) Benzo(g,h,i)perylene	27.409	276	1257218	33.755	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041223\
Data File : BP014682.D
Acq On : 12 Apr 2023 20:07
Operator : CG/JU
Sample : PB152055BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SLCS055

Quant Time: Apr 13 00:51:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP032823.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Apr 11 01:21:43 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 15.203 ng/uL
RT: 3.393 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:	88	Resp:	42049
Ion	Ratio	Lower	Upper
88	100		
43	33.2	24.4	36.6
58	75.8	57.4	86.0



#5
Pyridine
Concen: 37.259 ng/uL
RT: 3.810 min Scan# 106
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:	79	Resp:	251487
Ion	Ratio	Lower	Upper
79	100		
52	78.2	61.4	92.2
51	41.4	28.5	42.7









#6
Benzaldehyde
Concen: 17.382 ng/ul
RT: 7.116 min Scan# 668
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion: 77 Resp: 77769
Ion Ratio Lower Upper
77 100
105 88.9 70.6 105.8
106 84.8 66.5 99.7



#8
Phenol
Concen: 35.323 ng/ul
RT: 7.169 min Scan# 677
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion: 94 Resp: 329837
Ion Ratio Lower Upper
94 100
65 33.9 25.2 37.8
66 48.4 36.2 54.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 35.202 ng/ul
 RT: 7.393 min Scan# 715
 Delta R.T. -0.006 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07

Instrument :
 BNA_P
 ClientSampleId :
 SLCS055

Tgt Ion: 93 Resp: 264150




Ion	Ratio	Lower	Upper
93	100		
63	84.1	63.2	94.8
95	31.8	26.0	39.0




#12
 2-Chlorophenol
 Concen: 35.674 ng/ul
 RT: 7.534 min Scan# 739
 Delta R.T. -0.006 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07

Tgt Ion: 128 Resp: 241563

Ion	Ratio	Lower	Upper
128	100		
64	61.0	41.4	62.0
130	30.4	24.1	36.1





#13

2-Methylphenol

Concen: 33.618 ng/ul

RT: 8.428 min Scan# 890

Delta R.T. -0.000 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :





SLCS055

Tgt Ion:108 Resp: 234106

Ion Ratio Lower Upper

108 100

107 91.0 71.2 106.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 40.464 ng/ul

RT: 8.504 min Scan# 904

Delta R.T. -0.000 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion: 45 Resp: 398659


Ion Ratio Lower Upper

45 100

77 15.0 12.8 19.2

79 11.2 9.6 14.4









#16
Acetophenone
Concen: 33.550 ng/ul
RT: 8.810 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:105 Resp: 402252
Ion Ratio Lower Upper
105 100
77 86.4 67.3 100.9
51 32.4 23.0 34.4



#17
N-Nitroso-di-n-propylamine
Concen: 35.445 ng/ul
RT: 8.787 min Scan# 952
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion: 70 Resp: 229586
Ion Ratio Lower Upper
70 100
42 58.7 41.1 61.7
101 8.7 7.0 10.6
130 17.2 14.6 21.8









#18
4-Methylphenol
Concen: 33.144 ng/ul
RT: 8.757 min Scan# 947
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:108 Resp: 253480
Ion Ratio Lower Upper
108 100
107 113.5 89.0 133.6



#19
Hexachloroethane
Concen: 37.610 ng/ul
RT: 9.051 min Scan# 997
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:117 Resp: 111275
Ion Ratio Lower Upper
117 100
201 99.8 90.2 135.4
199 63.2 54.2 81.2









#22
Nitrobenzene
Concen: 36.904 ng/ul
RT: 9.192 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion: 77 Resp: 345212
Ion Ratio Lower Upper
77 100
123 36.1 30.3 45.5
65 14.2 11.0 16.4



#23
Isophorone
Concen: 33.439 ng/ul
RT: 9.716 min Scan# 1110
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion: 82 Resp: 576576
Ion Ratio Lower Upper
82 100
95 7.6 6.3 9.5
138 13.4 12.2 18.2









#25
2-Nitrophenol
Concen: 33.716 ng/ul
RT: 9.910 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:139 Resp: 132685
Ion Ratio Lower Upper
139 100
65 68.3 49.7 74.5
109 43.3 31.1 46.7



#26
2,4-Dimethylphenol
Concen: 32.003 ng/ul
RT: 9.963 min Scan# 1152
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:107 Resp: 281896
Ion Ratio Lower Upper
107 100
121 45.4 37.0 55.6
122 73.2 59.8 89.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.745 ng/ul
 RT: 10.198 min Scan# 1192
 Delta R.T. -0.006 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07




Instrument :
 BNA_P
 ClientSampleId :
 SLCS055


Tgt Ion: 93 Resp: 338516
 Ion Ratio Lower Upper
 93 100
 95 31.5 25.8 38.8
 123 9.9 8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 31.893 ng/ul
 RT: 10.445 min Scan# 1234
 Delta R.T. -0.006 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07

Tgt Ion:162 Resp: 213125
 Ion Ratio Lower Upper
 162 100
 164 69.4 50.1 75.1
 98 42.4 30.6 45.8









#30
Naphthalene
Concen: 32.868 ng/ul
RT: 10.845 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:128 Resp: 738440
Ion Ratio Lower Upper
128 100
129 10.7 8.9 13.3
127 13.2 11.0 16.6



#32
4-Chloroaniline
Concen: 17.455 ng/ul
RT: 10.969 min Scan# 1323
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:127 Resp: 158129
Ion Ratio Lower Upper
127 100
129 30.2 25.4 38.0





#33

Hexachlorobutadiene

Concen: 31.279 ng/ul

RT: 11.116 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:225 Resp: 165596

Ion Ratio Lower Upper

225 100

223 60.1 49.3 73.9

227 64.0 49.6 74.4



#35

4-Chloro-3-methylphenol

Concen: 31.038 ng/ul

RT: 12.092 min Scan# 1514

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:107 Resp: 255111


Ion Ratio Lower Upper

107 100

144 21.7 17.3 25.9

142 67.6 54.6 81.8









#36
2-Methylnaphthalene
Concen: 31.592 ng/ul
RT: 12.451 min Scan# 1575
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:142 Resp: 481005
Ion Ratio Lower Upper
142 100
141 89.2 71.6 107.4



#37
1-Methylnaphthalene
Concen: 32.057 ng/ul
RT: 12.669 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:142 Resp: 484074
Ion Ratio Lower Upper
142 100
141 95.1 76.6 114.8
116 4.6 3.6 5.4









#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.208 ng/ul
RT: 12.816 min Scan# 1637
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:	216	Resp:	269076
Ion	Ratio	Lower	Upper
216	100		
214	78.4	61.0	91.6
179	20.3	16.6	24.8
108	18.9	13.8	20.6



#40
Hexachlorocyclopentadiene
Concen: 28.195 ng/ul
RT: 12.786 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:	237	Resp:	152542
Ion	Ratio	Lower	Upper
237	100		
235	60.0	51.0	76.4
272	12.2	10.3	15.5





#41

2,4,6-Trichlorophenol

Concen: 31.855 ng/ul

RT: 13.063 min Scan# 1679

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:196 Resp: 164683

Ion Ratio Lower Upper

196 100

198 96.8 76.6 114.8

200 32.3 25.4 38.2



#42

2,4,5-Trichlorophenol

Concen: 32.450 ng/ul

RT: 13.139 min Scan# 1692

Delta R.T. -0.012 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:196 Resp: 178591


Ion Ratio Lower Upper

196 100

198 92.8 79.8 119.8

200 29.6 25.6 38.4









#43
1,1'-Biphenyl
Concen: 33.372 ng/ul
RT: 13.457 min Scan# 1746
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:154 Resp: 631754
Ion Ratio Lower Upper
154 100
153 40.6 32.4 48.6
76 15.5 10.7 16.1



#44
2-Chloronaphthalene
Concen: 33.843 ng/ul
RT: 13.504 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:162 Resp: 506576
Ion Ratio Lower Upper
162 100
164 32.7 26.1 39.1
127 43.2 33.7 50.5









#45
2-Nitroaniline
Concen: 37.135 ng/ul
RT: 13.722 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion: 65 Resp: 172139
Ion Ratio Lower Upper
65 100
92 60.9 53.7 80.5
138 78.1 71.7 107.5



#47
Dimethylphthalate
Concen: 32.214 ng/ul
RT: 14.080 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion: 163 Resp: 611936
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.2 7.8 11.8





#48

2,6-Dinitrotoluene

Concen: 33.714 ng/ul

RT: 14.210 min Scan# 1874

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:165 Resp: 125125

Ion Ratio Lower Upper

165 100

89 76.8 52.7 79.1

121 24.1 17.9 26.9



#50

Acenaphthylene

Concen: 33.606 ng/ul

RT: 14.351 min Scan# 1898

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:152 Resp: 754499


Ion Ratio Lower Upper

152 100

151 19.9 16.2 24.4

153 12.8 10.2 15.4









#51
3-Nitroaniline
Concen: 31.191 ng/ul
RT: 14.551 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:138 Resp: 103706
Ion Ratio Lower Upper
138 100
108 12.9 9.9 14.9
92 142.9 107.6 161.4



#52
Acenaphthene
Concen: 34.070 ng/ul
RT: 14.686 min Scan# 1955
Delta R.T. -0.012 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:153 Resp: 540532
Ion Ratio Lower Upper
153 100
152 46.8 38.2 57.4
154 88.5 70.2 105.2









#53
2,4-Dinitrophenol
Concen: 31.590 ng/ul
RT: 14.763 min Scan# 1968
Delta R.T. -0.012 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:184 Resp: 75922
Ion Ratio Lower Upper
184 100
63 85.7 57.1 85.7
154 77.2 55.3 82.9



#55
4-Nitrophenol
Concen: 34.875 ng/ul
RT: 14.863 min Scan# 1985
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:109 Resp: 98791
Ion Ratio Lower Upper
109 100
139 97.9 77.4 116.2
65 140.8 97.3 145.9









#56
Dibenzofuran
Concen: 32.549 ng/ul
RT: 15.027 min Scan# 2013
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:168 Resp: 725942
Ion Ratio Lower Upper
168 100
139 42.7 32.1 48.1



#57
2,4-Dinitrotoluene
Concen: 35.382 ng/ul
RT: 15.004 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:165 Resp: 186265
Ion Ratio Lower Upper
165 100
63 59.4 41.6 62.4
182 2.9 2.4 3.6









#58
2,3,4,6-Tetrachlorophenol
Concen: 30.274 ng/ul
RT: 15.257 min Scan# 2052
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Instrument :
BNA_P
ClientSampleId :
SLCS055

Tgt Ion:232 Resp: 149749




Ion	Ratio	Lower	Upper
232	100		
131	48.1	35.0	52.4
130	3.0	1.9	2.9
166	33.2	24.1	36.1




#59
Diethylphthalate
Concen: 32.634 ng/ul
RT: 15.445 min Scan# 2084
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:149 Resp: 616031

Ion	Ratio	Lower	Upper
149	100		
177	21.1	18.2	27.2
150	12.5	9.8	14.8









#61
Fluorene
Concen: 33.571 ng/ul
RT: 15.680 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:166 Resp: 606870
Ion Ratio Lower Upper
166 100
165 98.1 79.2 118.8
167 13.0 11.0 16.4



#62
4-Chlorophenyl-phenylether
Concen: 31.873 ng/ul
RT: 15.669 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:204 Resp: 309563
Ion Ratio Lower Upper
204 100
206 33.6 26.9 40.3
141 65.9 51.3 76.9









#63
4-Nitroaniline
Concen: 35.284 ng/ul
RT: 15.721 min Scan# 2131
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:138 Resp: 95319
Ion Ratio Lower Upper
138 100
92 64.9 47.8 71.6
108 112.3 82.3 123.5



#66
4,6-Dinitro-2-methylphenol
Concen: 32.468 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:198 Resp: 110853
Ion Ratio Lower Upper
198 100
182 3.8 3.5 5.3
77 32.9 24.4 36.6





#67

N-Nitrosodiphenylamine

Concen: 32.796 ng/ul

RT: 15.886 min Scan# 2159

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:169 Resp: 488552

Ion Ratio Lower Upper

169 100

168 66.1 54.2 81.2

167 35.3 29.4 44.2



#68

4-Bromophenyl-phenylether

Concen: 30.799 ng/ul

RT: 16.568 min Scan# 2275

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:248 Resp: 183498


Ion Ratio Lower Upper

248 100

250 96.8 78.1 117.1

141 83.2 66.0 99.0









#69
Hexachlorobenzene
Concen: 31.226 ng/ul
RT: 16.686 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:284 Resp: 213696
Ion Ratio Lower Upper
284 100
142 36.4 26.6 39.8
249 33.8 25.7 38.5



#70
Atrazine
Concen: 2.916 ng/ul
RT: 16.851 min Scan# 2323
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:200 Resp: 15812
Ion Ratio Lower Upper
200 100
173 26.5 20.6 30.8
215 40.7 37.2 55.8





#71

Pentachlorophenol

Concen: 28.991 ng/ul

RT: 17.039 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:266 Resp: 114938

Ion Ratio Lower Upper

266 100

264 60.9 51.0 76.4

268 64.0 51.4 77.2



#72

Phenanthrene

Concen: 34.094 ng/ul

RT: 17.433 min Scan# 2422

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:178 Resp: 908167


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.4

176 19.0 15.3 22.9









#74
 Anthracene
 Concen: 34.064 ng/uI
 RT: 17.521 min Scan# 24
 Delta R.T. -0.006 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07

Instrument :
 BNA_P
 ClientSampleId :
 SLCS055

Tgt Ion:178 Resp: 914512




Ion	Ratio	Lower	Upper
178	100		
179	15.5	12.2	18.4
176	18.4	14.9	22.3




#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 32.122 ng/uI
 RT: 13.422 min Scan# 1740
 Delta R.T. -0.006 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07

Tgt Ion:216 Resp: 270726

Ion	Ratio	Lower	Upper
216	100		
214	78.3	61.0	91.6
218	48.9	38.7	58.1









#76
 Pentachlorobenzene
 Concen: 32.486 ng/uL
 RT: 14.945 min Scan# 1999
 Delta R.T. -0.006 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07




Instrument :
 BNA_P
 ClientSampleId :
 SLCS055


Tgt Ion:250 Resp: 273190
 Ion Ratio Lower Upper
 250 100
 248 62.3 49.0 73.6
 252 62.0 50.7 76.1



#77
 Carbazole
 Concen: 35.238 ng/uL
 RT: 17.804 min Scan# 2485
 Delta R.T. -0.000 min
 Lab File: BP014682.D
 Acq: 12 Apr 2023 20:07

Tgt Ion:167 Resp: 791679
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.4 26.2
 139 13.4 11.3 16.9









#78
Di-n-butylphthalate
Concen: 35.983 ng/ul
RT: 18.327 min Scan# 2574
Delta R.T. -0.006 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:149 Resp: 1016517
Ion Ratio Lower Upper
149 100
150 9.1 7.4 11.0
104 6.8 4.8 7.2



#80
Fluoranthene
Concen: 26.931 ng/ul
RT: 19.398 min Scan# 2756
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:202 Resp: 1088126
Ion Ratio Lower Upper
202 100
101 10.5 8.2 12.2
100 8.3 6.4 9.6





#82

Pyrene

Concen: 27.472 ng/ul

RT: 19.751 min Scan# 2816

Delta R.T. -0.000 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:202 Resp: 1158840

Ion Ratio Lower Upper

202 100

101 12.6 9.7 14.5

100 10.7 8.0 12.0



#83

Butylbenzylphthalate

Concen: 33.836 ng/ul

RT: 20.609 min Scan# 2962

Delta R.T. -0.000 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:149 Resp: 491633


Ion Ratio Lower Upper

149 100

91 80.1 61.8 92.8

206 20.9 17.8 26.6





#84

3,3'-Dichlorobenzidine

Concen: 23.927 ng/ul

RT: 21.398 min Scan# 3095

Delta R.T. -0.000 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:252 Resp: 269210

Ion Ratio Lower Upper

252 100

254 62.4 51.4 77.0

126 10.3 7.8 11.6



#85

Benzo(a)anthracene

Concen: 33.396 ng/ul

RT: 21.462 min Scan# 3107

Delta R.T. -0.000 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:228 Resp: 1203955


Ion Ratio Lower Upper

228 100

229 19.8 15.6 23.4

226 26.9 21.8 32.6









#86
Bis(2-ethylhexyl)phthalate
Concen: 36.405 ng/ul
RT: 21.362 min Scan# 3090
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:149 Resp: 723330
Ion Ratio Lower Upper
149 100
167 27.6 21.9 32.9
279 4.4 3.7 5.5



#87
Chrysene
Concen: 34.142 ng/ul
RT: 21.521 min Scan# 3117
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:228 Resp: 1162215
Ion Ratio Lower Upper
228 100
226 29.5 23.8 35.8
229 19.8 16.1 24.1





#89

Di-n-octyl phthalate

Concen: 35.396 ng/ul

RT: 22.315 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07





Instrument :

BNA_P

ClientSampleId :

SLCS055

Tgt Ion:149 Resp: 1312302



#90

Benzo(b)fluoranthene

Concen: 33.564 ng/ul

RT: 23.209 min Scan# 3404

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:252 Resp: 1295078


Ion Ratio Lower Upper

252 100

253 22.4 18.0 27.0

125 8.5 6.8 10.2









#91
Benzo(k)fluoranthene
Concen: 33.773 ng/ul
RT: 23.262 min Scan# 3413
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07




Instrument :
BNA_P
ClientSampleId :
SLCS055


Tgt Ion:252 Resp: 1265642
Ion Ratio Lower Upper
252 100
253 21.7 17.3 25.9
125 8.9 6.9 10.3



#93
Benzo(a)pyrene
Concen: 34.568 ng/ul
RT: 23.868 min Scan# 3516
Delta R.T. -0.000 min
Lab File: BP014682.D
Acq: 12 Apr 2023 20:07

Tgt Ion:252 Resp: 1148087
Ion Ratio Lower Upper
252 100
253 21.9 16.9 25.3
125 10.2 6.8 10.2





#94

Indeno(1,2,3-cd)pyrene

Concen: 34.000 ng/ul

RT: 26.597 min Scan# 3980

Delta R.T. -0.006 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055





Tgt Ion:276 Resp: 1553551

Ion Ratio Lower Upper

276 100

138 20.3 14.9 22.3

277 24.5 20.3 30.5



#95

Dibenzo(a,h)anthracene

Concen: 33.659 ng/ul

RT: 26.615 min Scan# 3983

Delta R.T. -0.000 min

Lab File: BP014682.D

Acq: 12 Apr 2023 20:07




Tgt Ion:278 Resp: 1287670


Ion Ratio Lower Upper

278 100

139 13.3 10.4 15.6

279 23.3 18.9 28.3





#96

Benzo(g,h,i)perylene

Concen: 33.755 ng/ul

RT: 27.409 min Scan# 4118

Delta R.T. -0.000 min

Lab File: BP014682.D


Acq: 12 Apr 2023 20:07

Instrument :

BNA_P

ClientSampleId :

SLCS055



Tgt Ion:276 Resp: 1257218

Ion Ratio Lower Upper

276 100

138 18.0 13.0 19.4

277 24.1 18.4 27.6

