

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041321\
 Data File : BP005279.D
 Acq On : 13 Apr 2021 10:26
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 14 02:38:03 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 18:44:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	136	0.00
2	1,4-Dioxane	40.000	37.676	5.8	127	0.00
3	Pyridine	40.000	38.761	3.1	122	0.00
4	n-Nitrosodimethylamine	40.000	36.942	7.6	123	0.00
5 S	2-Fluorophenol	80.000	86.387	-8.0	148	0.00
6	Aniline	40.000	38.617	3.5	133	0.00
7 S	Phenol-d6	80.000	79.272	0.9	137	0.00
8	2-Chlorophenol	40.000	40.412	-1.0	138	0.00
9	Benzaldehyde	40.000	37.608	6.0	132	0.00
10 C	Phenol	40.000	39.067	2.3	135	0.00
11	bis(2-Chloroethyl)ether	40.000	39.881	0.3	136	0.00
12	1,3-Dichlorobenzene	40.000	40.349	-0.9	141	0.00
13 C	1,4-Dichlorobenzene	40.000	40.822	-2.1	140	0.00
14	1,2-Dichlorobenzene	40.000	39.203	2.0	135	0.00
15	Benzyl Alcohol	40.000	33.994	15.0	116	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.348	4.1	135	0.00
17	2-Methylphenol	40.000	35.665	10.8	123	0.00
18	Hexachloroethane	40.000	39.626	0.9	138	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	35.185	12.0	121	0.00
20	3+4-Methylphenols	40.000	35.598	11.0	123	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	124	0.00
22	Acetophenone	40.000	41.758	-4.4	129	0.00
23 S	Nitrobenzene-d5	80.000	84.662	-5.8	127	0.00
24	Nitrobenzene	40.000	42.049	-5.1	128	0.00
25	Isophorone	40.000	37.804	5.5	115	0.00
26 C	2-Nitrophenol	40.000	38.162	4.6	111	0.00
27	2,4-Dimethylphenol	40.000	39.308	1.7	119	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.719	3.2	117	0.00
29 C	2,4-Dichlorophenol	40.000	38.973	2.6	118	0.00
30	1,2,4-Trichlorobenzene	40.000	41.347	-3.4	127	0.00
31	Naphthalene	40.000	40.534	-1.3	125	0.00
32	Benzoic acid	40.000	32.073	19.8	95	0.00
33	4-Chloroaniline	40.000	37.850	5.4	114	0.00
34 C	Hexachlorobutadiene	40.000	44.239	-10.6	137	0.00
35	Caprolactam	40.000	32.182	19.5	91	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.933	7.7	111	0.00
37	2-Methylnaphthalene	40.000	39.124	2.2	119	0.00
38	1-Methylnaphthalene	40.000	38.723	3.2	119	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	113	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	45.181	-13.0	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.265	-0.7	106	0.00
42 S	2,4,6-Tribromophenol	80.000	79.234	1.0	106	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.880	-4.7	108	0.00
44	2,4,5-Trichlorophenol	40.000	40.436	-1.1	105	0.00
45 S	2-Fluorobiphenyl	80.000	87.643	-9.6	118	0.00
46	1,1'-Biphenyl	40.000	43.319	-8.3	116	0.00
47	2-Chloronaphthalene	40.000	43.364	-8.4	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041321\
 Data File : BP005279.D
 Acq On : 13 Apr 2021 10:26
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 14 02:38:03 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP040921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 09 18:44:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	35.559	11.1	99	0.00
49	Acenaphthylene	40.000	41.914	-4.8	112	0.00
50	Dimethylphthalate	40.000	39.311	1.7	105	0.00
51	2,6-Dinitrotoluene	40.000	40.324	-0.8	103	0.00
52 C	Acenaphthene	40.000	41.485	-3.7	111	0.00
53	3-Nitroaniline	40.000	35.165	12.1	95	0.00
54 P	2,4-Dinitrophenol	40.000	34.556	13.6	94	0.00
55	Dibenzofuran	40.000	42.466	-6.2	114	0.00
56 P	4-Nitrophenol	40.000	35.861	10.3	91	0.01
57	2,4-Dinitrotoluene	40.000	39.624	0.9	100	0.00
58	Fluorene	40.000	40.442	-1.1	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.962	2.6	106	0.00
60	Diethylphthalate	40.000	38.414	4.0	101	0.00
61	4-Chlorophenyl-phenylether	40.000	41.541	-3.9	110	0.00
62	4-Nitroaniline	40.000	35.128	12.2	91	0.00
63	Azobenzene	40.000	41.673	-4.2	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	36.605	8.5	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	42.089	-5.2	105	0.00
67	4-Bromophenyl-phenylether	40.000	43.270	-8.2	109	0.00
68	Hexachlorobenzene	40.000	44.465	-11.2	110	0.00
69	Atrazine	40.000	38.713	3.2	93	0.00
70 C	Pentachlorophenol	40.000	38.662	3.3	93	0.00
71	Phenanthrene	40.000	41.751	-4.4	104	0.00
72	Anthracene	40.000	41.799	-4.5	102	0.00
73	Carbazole	40.000	40.597	-1.5	98	0.00
74	Di-n-butylphthalate	40.000	38.516	3.7	93	0.00
75 C	Fluoranthene	40.000	39.743	0.6	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzidine	40.000	34.851	12.9	76	0.00
78	Pyrene	40.000	39.983	0.0	97	0.00
79 S	Terphenyl-d14	80.000	82.283	-2.9	100	0.00
80	Butylbenzylphthalate	40.000	32.288	19.3	77	0.00
81	Benzo(a)anthracene	40.000	38.540	3.7	93	0.00
82	3,3'-Dichlorobenzidine	40.000	37.207	7.0	86	0.00
83	Chrysene	40.000	39.750	0.6	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	34.273	14.3	82	0.00
85 c	Di-n-octyl phthalate	40.000	34.686	13.3	83	0.00
86 I	Perylene-d12	20.000	20.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.606	-4.0	97	0.00
88	Benzo(b)fluoranthene	40.000	40.430	-1.1	93	0.00
89	Benzo(k)fluoranthene	40.000	39.968	0.1	92	0.00
90 C	Benzo(a)pyrene	40.000	40.211	-0.5	93	0.00
91	Dibenzo(a,h)anthracene	40.000	41.309	-3.3	95	0.00
92	Benzo(g,h,i)perylene	40.000	41.413	-3.5	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041321\
Data File : BP005279.D
Acq On : 13 Apr 2021 10:26
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Apr 14 02:38:03 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP040921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 09 18:44:11 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0