

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041422\
 Data File : BP009825.D
 Acq On : 14 Apr 2022 13:34
 Operator : CG/JU
 Sample : PB144012BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS012

Quant Time: Apr 14 21:45:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 13 02:15:22 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
 Supervised By :Yogesh Patel 04/22/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.963	152	128134	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	578007	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.592	164	418271	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	963357	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.427	240	1019467	20.000	ng/u1	# 0.00
88) Perylene-d12	23.892	264	946703	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.381	96	17700m	6.165	ng/uL	0.00
4) Pyridine-d5	3.799	84	230786	29.155	ng/u1	0.00
7) Phenol-d5	7.116	99	334463	29.732	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.287	67	211313	31.553	ng/u1	0.00
11) 2-Chlorophenol-d4	7.493	132	266099	31.359	ng/u1	0.00
15) 4-Methylphenol-d8	8.669	113	285497	30.402	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	137457	32.978	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	151779	33.214	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	294998	30.849	ng/u1	0.00
31) 4-Chloroaniline-d4	10.898	131	375028	27.536	ng/u1	0.00
46) Dimethylphthalate-d6	14.004	166	1017869	30.980	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	1143348	29.373	ng/u1	0.00
54) 4-Nitrophenol-d4	14.775	143	183637	30.819	ng/u1	0.00
60) Fluorene-d10	15.586	176	867613	31.390	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.698	200	179874	31.518	ng/u1	0.00
73) Anthracene-d10	17.445	188	1399882	30.350	ng/u1	0.00
81) Pyrene-d10	19.674	212	1746820	31.373	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.733	264	1614983	32.758	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.417	88	33909	11.586	ng/uL#	21
5) Pyridine	3.817	79	239201	28.962	ng/u1#	38
6) Benzaldehyde	7.093	77	222133	32.198	ng/u1	96
8) Phenol	7.146	94	347828	30.890	ng/u1#	90
10) Bis(2-Chloroethyl)ether	7.381	93	265713	30.294	ng/u1#	76
12) 2-Chlorophenol	7.522	128	265782	31.068	ng/u1	95
13) 2-Methylphenol	8.399	108	267643	30.481	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.499	45	405205	31.253	ng/u1#	84
16) Acetophenone	8.787	105	447561	30.249	ng/u1#	79
17) N-Nitroso-di-n-propyla...	8.775	70	237422	30.951	ng/u1#	70
18) 4-Methylphenol	8.734	108	292912	30.805	ng/u1	96
19) Hexachloroethane	9.052	117	111652	30.912	ng/u1	83
22) Nitrobenzene	9.163	77	348022	32.529	ng/u1#	83
23) Isophorone	9.699	82	659736	29.925	ng/u1	96
25) 2-Nitrophenol	9.875	139	159995	32.827	ng/u1#	85
26) 2,4-Dimethylphenol	9.940	107	342733	29.632	ng/u1#	80
27) Bis(2-Chloroethoxy)met...	10.175	93	395304	30.541	ng/u1#	97
29) 2,4-Dichlorophenol	10.410	162	289946	31.276	ng/u1#	85
30) Naphthalene	10.822	128	906398	30.491	ng/u1	96
32) 4-Chloroaniline	10.922	127	368461	27.338	ng/u1	97
33) Hexachlorobutadiene	11.116	225	204682	30.655	ng/u1	95
34) Caprolactam	11.693	113	100504m	32.781	ng/u1	
35) 4-Chloro-3-methylphenol	12.045	107	354363	31.546	ng/u1#	70
36) 2-Methylnaphthalene	12.428	142	658335	30.314	ng/u1	90

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041422\
 Data File : BP009825.D
 Acq On : 14 Apr 2022 13:34
 Operator : CG/JU
 Sample : PB144012BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :

BNA_P

ClientSampleId :

SLCS012

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
 Supervised By :Yogesh Patel 04/22/2022

Quant Time: Apr 14 21:45:11 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M

Quant Title : SVOA CALIBRATION

QLast Update : Wed Apr 13 02:15:22 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.645	142	673152	30.698	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.792	216	394711	30.844	ng/ul	92
40) Hexachlorocyclopentadiene	12.781	237	229262	25.816	ng/ul	95
41) 2,4,6-Trichlorophenol	13.028	196	258807	31.668	ng/ul#	79
42) 2,4,5-Trichlorophenol	13.098	196	288807	32.701	ng/ul#	87
43) 1,1'-Biphenyl	13.434	154	936821	30.697	ng/ul	93
44) 2-Chloronaphthalene	13.469	162	726314	30.961	ng/ul	94
45) 2-Nitroaniline	13.669	65	248117	36.090	ng/ul#	81
47) Dimethylphthalate	14.051	163	974841	30.827	ng/ul#	92
48) 2,6-Dinitrotoluene	14.163	165	202548	34.822	ng/ul	96
50) Acenaphthylene	14.316	152	1185025	30.849	ng/ul	96
51) 3-Nitroaniline	14.487	138	195414	32.316	ng/ul#	82
52) Acenaphthene	14.657	153	766265	30.467	ng/ul	97
53) 2,4-Dinitrophenol	14.692	184	111755	31.785	ng/ul#	81
55) 4-Nitrophenol	14.792	109	172194	31.393	ng/ul#	52
56) Dibenzofuran	14.992	168	1140447	30.809	ng/ul	98
57) 2,4-Dinitrotoluene	14.945	165	301196	35.264	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.216	232	261057	32.419	ng/ul#	86
59) Diethylphthalate	15.416	149	1015651	31.034	ng/ul	94
61) Fluorene	15.639	166	952749	31.248	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.634	204	502380	30.988	ng/ul	96
63) 4-Nitroaniline	15.651	138	217211m	35.863	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.710	198	179170	31.585	ng/ul#	89
67) N-Nitrosodiphenylamine	15.845	169	834901	30.275	ng/ul	94
68) 4-Bromophenyl-phenylether	16.533	248	314662	30.175	ng/ul	94
69) Hexachlorobenzene	16.651	284	368964	30.771	ng/ul#	90
70) Atrazine	16.804	200	336315	29.546	ng/ul#	90
71) Pentachlorophenol	16.992	266	240523	29.634	ng/ul#	85
72) Phenanthrene	17.386	178	1567407	30.481	ng/ul	98
74) Anthracene	17.480	178	1591862	30.533	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.398	216	424254	30.460	ng/uL#	85
76) Pentachlorobenzene	14.916	250	410745	30.255	ng/uL	91
77) Carbazole	17.745	167	1440855	30.582	ng/ul#	95
78) Di-n-butylphthalate	18.298	149	1788615	30.872	ng/ul#	93
80) Fluoranthene	19.345	202	1999749	31.203	ng/ul#	96
82) Pyrene	19.698	202	2037553	31.358	ng/ul#	94
83) Butylbenzylphthalate	20.569	149	846036	31.739	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.339	252	651266	29.979	ng/ul#	96
85) Benzo(a)anthracene	21.410	228	2050096	31.721	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.333	149	1253973	31.275	ng/ul#	81
87) Chrysene	21.463	228	1980995	31.769	ng/ul	99
89) Di-n-octyl phthalate	22.286	149	2130620	32.122	ng/ul	100
90) Benzo(b)fluoranthene	23.139	252	2021966	32.146	ng/ul	99
91) Benzo(k)fluoranthene	23.186	252	1929980	33.626	ng/ul#	97
93) Benzo(a)pyrene	23.786	252	1896766	32.532	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.474	276	1928421	31.630	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.498	278	1675812	31.699	ng/ul#	97
96) Benzo(g,h,i)perylene	27.262	276	1582176	31.473	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041422\
 Data File : BP009825.D
 Acq On : 14 Apr 2022 13:34
 Operator : CG/JU
 Sample : PB144012BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS012

Quant Time: Apr 14 21:45:11 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP040722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 13 02:15:22 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
 Supervised By :Yogesh Patel 04/22/2022





#2
1,4-Dioxane
Concen: 11.586 ng/uL
RT: 3.417 min Scan# 50
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion: 88 Resp: 33909
Ion Ratio Lower Upper
88 100
43 38.1 6.8 10.2
58 81.2 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#5
Pyridine
Concen: 28.962 ng/uL
RT: 3.817 min Scan# 124
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion: 79 Resp: 239201
Ion Ratio Lower Upper
79 100
52 93.1 35.5 53.3#
51 44.7 20.2 30.4#





#6
Benzaldehyde
Concen: 32.198 ng/ul
RT: 7.093 min Scan# 681
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion: 77 Resp: 222133
Ion Ratio Lower Upper
77 100
105 94.7 77.0 115.4
106 89.4 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#8
Phenol
Concen: 30.890 ng/ul
RT: 7.146 min Scan# 690
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion: 94 Resp: 347828
Ion Ratio Lower Upper
94 100
65 33.0 19.0 28.6#
66 46.2 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 30.294 ng/ul
RT: 7.381 min Scan# 730
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion: 93 Resp: 265711
Ion Ratio Lower Upper
93 100
63 84.4 45.0 67.4
95 32.9 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#12
2-Chlorophenol
Concen: 31.068 ng/ul
RT: 7.522 min Scan# 754
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34



Tgt Ion:128 Resp: 265782
Ion Ratio Lower Upper
128 100
64 54.3 45.3 67.9
130 31.3 29.0 43.6





#13

2-Methylphenol

Concen: 30.481 ng/ul

RT: 8.399 min Scan# 90

Delta R.T. -0.006 min

Lab File: BP009825.D

Acq: 14 Apr 2022 13:34

Instrument :

BNA_P

ClientSampleId :

SLCS012

Tgt Ion:108 Resp: 26764

Ion Ratio Lower Upper

108 100





107 88.9 70.8 106.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022

Supervised By :Yogesh Patel 04/22/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 31.253 ng/ul

RT: 8.499 min Scan# 920

Delta R.T. -0.006 min

Lab File: BP009825.D

Acq: 14 Apr 2022 13:34




Tgt Ion: 45 Resp: 405205


Ion Ratio Lower Upper

45 100

77 13.3 17.1 25.7#


79 10.7 13.2 19.8#





#16
Acetophenone
Concen: 30.249 ng/ul
RT: 8.787 min Scan# 969
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion: 105 Resp: 44756
Ion Ratio Lower Upper
105 100
77 81.8 53.4 80.2
51 30.9 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#17
N-Nitroso-di-n-propylamine
Concen: 30.951 ng/ul
RT: 8.775 min Scan# 967
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34



Tgt Ion: 70 Resp: 237422
Ion Ratio Lower Upper
70 100
42 66.6 31.8 47.8#
101 10.5 15.6 23.4#
130 20.7 13.7 20.5#









#18
4-Methylphenol
Concen: 30.805 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Instrument :
BNA_P
ClientSampleId :
SLCS012

Tgt Ion:108 Resp: 292912
Ion Ratio Lower Upper
108 100
107 109.1 83.6 125.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#19
Hexachloroethane
Concen: 30.912 ng/ul
RT: 9.052 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:117 Resp: 111652
Ion Ratio Lower Upper
117 100
201 106.9 106.5 159.7
199 67.2 59.0 88.6





#22
Nitrobenzene
Concen: 32.529 ng/u1
RT: 9.163 min Scan# 1033
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion: 77 Resp: 348022
Ion Ratio Lower Upper
77 100
123 40.9 42.3 63.5
65 14.5 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#23
Isophorone
Concen: 29.925 ng/u1
RT: 9.699 min Scan# 1124
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34



Tgt Ion: 82 Resp: 659736
Ion Ratio Lower Upper
82 100
95 8.4 7.8 11.8
138 17.8 12.9 19.3





#25

2-Nitrophenol

Concen: 32.827 ng/ul

RT: 9.875 min Scan# 1154

Delta R.T. -0.006 min

Lab File: BP009825.D

Acq: 14 Apr 2022 13:34

Instrument :

BNA_P

ClientSampleId :

SLCS012

Tgt Ion:139 Resp: 15999

Ion Ratio Lower Upper

139 100

65 59.0 49.4 74.2





109 36.5 44.7 67.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022

Supervised By :Yogesh Patel 04/22/2022



#26

2,4-Dimethylphenol

Concen: 29.632 ng/ul

RT: 9.940 min Scan# 1165

Delta R.T. 0.000 min

Lab File: BP009825.D

Acq: 14 Apr 2022 13:34




Tgt Ion:107 Resp: 342733


Ion Ratio Lower Upper

107 100

121 47.3 34.1 51.1

122 79.7 46.3 69.5









#27
Bis(2-Chloroethoxy)methane
Concen: 30.541 ng/ul
RT: 10.175 min Scan# 1205
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Instrument :
BNA_P
ClientSampleId :
SLCS012

Tgt Ion: 93 Resp: 395304
Ion Ratio Lower Upper
93 100
95 32.0 26.2 39.4
123 11.5 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#29
2,4-Dichlorophenol
Concen: 31.276 ng/ul
RT: 10.410 min Scan# 1245
Delta R.T. -0.012 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:162 Resp: 289946
Ion Ratio Lower Upper
162 100
164 64.1 37.0 55.6#
98 38.3 31.0 46.6





#30
Naphthalene
Concen: 30.491 ng/ul
RT: 10.822 min Scan# 1315
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:128 Resp: 906398
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 12.9 12.2 18.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#32
4-Chloroaniline
Concen: 27.338 ng/ul
RT: 10.922 min Scan# 1332
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:127 Resp: 368461
Ion Ratio Lower Upper
127 100
129 32.2 24.6 37.0





#33
Hexachlorobutadiene
Concen: 30.655 ng/ul
RT: 11.116 min Scan# 1365
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:225 Resp: 20468
Ion Ratio Lower Upper
225 100
223 61.8 53.4 80.0
227 64.1 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#34
Caprolactam
Concen: 32.781 ng/ul m
RT: 11.693 min Scan# 1463
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34



Tgt Ion:113 Resp: 100504
Ion Ratio Lower Upper
113 100
55 189.3 58.2 87.2#
56 146.6 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 31.546 ng/ul
RT: 12.045 min Scan# 1522
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Instrument :
BNA_P
ClientSampleId :
SLCS012

Tgt Ion:107 Resp: 35436
Ion Ratio Lower Upper
107 100
144 24.3 15.8 23.8
142 74.8 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#36
2-Methylnaphthalene
Concen: 30.314 ng/ul
RT: 12.428 min Scan# 1588
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:142 Resp: 658335
Ion Ratio Lower Upper
142 100
141 89.3 64.3 96.5





#37
1-Methylnaphthalene
Concen: 30.698 ng/ul
RT: 12.645 min Scan# 1625
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:142 Resp: 67315
Ion Ratio Lower Upper
142 100
141 92.3 69.8 104.8
116 4.2 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.844 ng/ul
RT: 12.792 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:216 Resp: 394711
Ion Ratio Lower Upper
216 100
214 79.4 58.2 87.2
179 20.9 14.2 21.2
108 16.2 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 25.816 ng/ul
RT: 12.781 min Scan# 1648
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion: 237 Resp: 22926
Ion Ratio Lower Upper
237 100
235 62.2 53.2 79.8
272 12.3 10.8 16.2

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 04/15/2022
Supervised By : Yogesh Patel 04/22/2022




#41
2,4,6-Trichlorophenol
Concen: 31.668 ng/ul
RT: 13.028 min Scan# 1690
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion: 196 Resp: 258807
Ion Ratio Lower Upper
196 100
198 100.5 64.1 96.1#
200 31.5 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 32.701 ng/ul
RT: 13.098 min Scan# 1702
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:196 Resp: 28880
Ion Ratio Lower Upper
196 100
198 96.8 68.6 103.0
200 31.2 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#43
1,1'-Biphenyl
Concen: 30.697 ng/ul
RT: 13.434 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:154 Resp: 936821
Ion Ratio Lower Upper
154 100
153 40.6 28.6 43.0
76 11.4 8.5 12.7





#44
2-Chloronaphthalene
Concen: 30.961 ng/ul
RT: 13.469 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:162 Resp: 726314
Ion Ratio Lower Upper
162 100
164 33.1 27.5 41.3
127 40.9 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#45
2-Nitroaniline
Concen: 36.090 ng/ul
RT: 13.669 min Scan# 1799
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion: 65 Resp: 248117
Ion Ratio Lower Upper
65 100
92 66.6 77.9 116.9#
138 95.4 81.5 122.3





#47
Dimethylphthalate
Concen: 30.827 ng/ul
RT: 14.051 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:163 Resp: 97484
Ion Ratio Lower Upper
163 100
194 4.2 4.1 6.1
164 10.0 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#48
2,6-Dinitrotoluene
Concen: 34.822 ng/ul
RT: 14.163 min Scan# 1883
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:165 Resp: 202548
Ion Ratio Lower Upper
165 100
89 61.3 51.7 77.5
121 21.1 18.3 27.5





#50
Acenaphthylene
Concen: 30.849 ng/ul
RT: 14.316 min Scan# 1909
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion:152 Resp: 118502
Ion Ratio Lower Upper
152 100
151 19.8 18.3 27.5
153 13.3 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#51
3-Nitroaniline
Concen: 32.316 ng/ul
RT: 14.487 min Scan# 1938
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Tgt Ion:138 Resp: 195414
Ion Ratio Lower Upper
138 100
108 12.0 15.8 23.8#
92 123.3 117.1 175.7









#52
Acenaphthene
Concen: 30.467 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Instrument :
BNA_P
ClientSampleId :
SLCS012

Tgt Ion:153 Resp: 766268
Ion Ratio Lower Upper
153 100
152 47.1 34.3 51.5
154 88.3 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#53
2,4-Dinitrophenol
Concen: 31.785 ng/ul
RT: 14.692 min Scan# 1973
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:184 Resp: 111755
Ion Ratio Lower Upper
184 100
63 73.3 34.6 52.0#
154 70.6 57.7 86.5





#55
4-Nitrophenol
Concen: 31.393 ng/u1
RT: 14.792 min Scan# 1989
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion:109 Resp: 172194
Ion Ratio Lower Upper
109 100
139 113.0 61.4 92.2
65 129.4 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#56
Dibenzofuran
Concen: 30.809 ng/u1
RT: 14.992 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:168 Resp: 1140447
Ion Ratio Lower Upper
168 100
139 39.0 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 35.264 ng/ul
RT: 14.945 min Scan# 2016
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:165 Resp: 301190
Ion Ratio Lower Upper
165 100
63 49.7 29.5 44.3#
182 2.9 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 32.419 ng/ul
RT: 15.216 min Scan# 2062
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34



Tgt Ion:232 Resp: 261057
Ion Ratio Lower Upper
232 100
131 43.8 25.4 38.2#
130 2.4 4.2 6.2#
166 30.8 28.1 42.1





#59
Diethylphthalate
Concen: 31.034 ng/ul
RT: 15.416 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Instrument :

BNA_P

ClientSampleId :





SLCS012

Tgt Ion:149 Resp: 101565
Ion Ratio Lower Upper
149 100
177 22.6 22.0 33.0
150 12.8 10.2 15.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#61
Fluorene
Concen: 31.248 ng/ul
RT: 15.639 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Tgt Ion:166 Resp: 952749
Ion Ratio Lower Upper
166 100
165 96.0 70.3 105.5
167 13.5 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 30.988 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. -0.012 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Instrument :
BNA_P
ClientSampleId :
SLCS012

Tgt Ion:204 Resp: 502380
Ion Ratio Lower Upper
204 100
206 32.7 27.6 41.4
141 62.1 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#63
4-Nitroaniline
Concen: 35.863 ng/ul m
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:138 Resp: 217211
Ion Ratio Lower Upper
138 100
92 54.0 41.6 62.4
108 92.5 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 31.585 ng/ul
RT: 15.710 min Scan# 2146
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion:198 Resp: 179170
Ion Ratio Lower Upper
198 100
182 4.5 6.0 9.0
77 27.9 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#67
N-Nitrosodiphenylamine
Concen: 30.275 ng/ul
RT: 15.845 min Scan# 2169
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:169 Resp: 834901
Ion Ratio Lower Upper
169 100
168 65.2 48.3 72.5
167 34.5 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 30.175 ng/ul
RT: 16.533 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:248 Resp: 314662
Ion Ratio Lower Upper
248 100
250 98.1 79.4 119.0
141 73.4 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#69
Hexachlorobenzene
Concen: 30.771 ng/ul
RT: 16.651 min Scan# 2306
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:284 Resp: 368964
Ion Ratio Lower Upper
284 100
142 33.5 19.2 28.8#
249 32.1 27.4 41.2





#70
Atrazine
Concen: 29.546 ng/ul
RT: 16.804 min Scan# 2332
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:200 Resp: 33631
Ion Ratio Lower Upper
200 100
173 26.3 19.2 28.8
215 48.7 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#71
Pentachlorophenol
Concen: 29.634 ng/ul
RT: 16.992 min Scan# 2364
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:266 Resp: 240523
Ion Ratio Lower Upper
266 100
264 64.3 57.3 85.9
268 63.2 65.7 98.5





#72
Phenanthrene
Concen: 30.481 ng/u1
RT: 17.386 min Scan# 2431
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:178 Resp: 156740
Ion Ratio Lower Upper
178 100
179 15.5 11.7 17.5
176 18.8 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#74
Anthracene
Concen: 30.533 ng/u1
RT: 17.480 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:178 Resp: 1591862
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 17.6 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.460 ng/uL
RT: 13.398 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:216 Resp: 424254
Ion Ratio Lower Upper
216 100
214 77.8 50.4 75.6
218 49.4 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#76
Pentachlorobenzene
Concen: 30.255 ng/uL
RT: 14.916 min Scan# 2011
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:250 Resp: 410745
Ion Ratio Lower Upper
250 100
248 62.0 43.5 65.3
252 63.2 46.5 69.7





#77
 Carbazole
 Concen: 30.582 ng/ul
 RT: 17.745 min Scan# 2492
 Delta R.T. 0.000 min
 Lab File: BP009825.D
 Acq: 14 Apr 2022 13:34

Instrument :
 BNA_P
 ClientSampleId :
 SLCS012






Tgt Ion:167 Resp: 144085


Ion	Ratio	Lower	Upper
167	100		
166	21.6	15.8	23.8
139	12.9	8.3	12.5

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
 Supervised By :Yogesh Patel 04/22/2022






#78
 Di-n-butylphthalate
 Concen: 30.872 ng/ul
 RT: 18.298 min Scan# 2586
 Delta R.T. -0.006 min
 Lab File: BP009825.D
 Acq: 14 Apr 2022 13:34



Tgt Ion:149 Resp: 1788615


Ion	Ratio	Lower	Upper
149	100		
150	9.3	9.6	14.4#
104	5.5	6.2	9.4#





#80
Fluoranthene
Concen: 31.203 ng/u1
RT: 19.345 min Scan# 2764
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:202 Resp: 1999749
Ion Ratio Lower Upper
202 100
101 9.4 9.6 14.4
100 7.5 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#82
Pyrene
Concen: 31.358 ng/u1
RT: 19.698 min Scan# 2824
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:202 Resp: 2037553
Ion Ratio Lower Upper
202 100
101 11.3 10.6 15.8
100 9.7 9.8 14.6#





#83
Butylbenzylphthalate
Concen: 31.739 ng/ul
RT: 20.569 min Scan# 2973
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion:149 Resp: 846030
Ion Ratio Lower Upper
149 100
91 73.8 71.1 106.7
206 21.6 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#84
3,3'-Dichlorobenzidine
Concen: 29.979 ng/ul
RT: 21.339 min Scan# 3103
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:252 Resp: 651266
Ion Ratio Lower Upper
252 100
254 64.6 49.8 74.6
126 9.0 4.3 6.5





#85
Benzo(a)anthracene
Concen: 31.721 ng/ul
RT: 21.410 min Scan# 3115
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:228 Resp: 2050090
Ion Ratio Lower Upper
228 100
229 19.8 18.2 27.2
226 26.5 22.8 34.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 31.275 ng/ul
RT: 21.333 min Scan# 3102
Delta R.T. -0.012 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:149 Resp: 1253973
Ion Ratio Lower Upper
149 100
167 27.7 32.8 49.2#
279 4.6 5.6 8.4#





#87
Chrysene
Concen: 31.769 ng/ul
RT: 21.463 min Scan# 3124
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion:228 Resp: 198099
Ion Ratio Lower Upper
228 100
226 29.0 22.7 34.1
229 19.6 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#89
Di-n-octyl phthalate
Concen: 32.122 ng/ul
RT: 22.286 min Scan# 3264
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:149 Resp: 2130620





#90
Benzo(b)fluoranthene
Concen: 32.146 ng/ul
RT: 23.139 min Scan# 3409
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012




Tgt Ion:252 Resp: 2021960
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 8.0 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022




#91
Benzo(k)fluoranthene
Concen: 33.626 ng/ul
RT: 23.186 min Scan# 3417
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Tgt Ion:252 Resp: 1929980
Ion Ratio Lower Upper
252 100
253 21.5 17.9 26.9
125 8.4 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 32.532 ng/ul
RT: 23.786 min Scan# 3519
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion:252 Resp: 1896760
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 9.5 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#94
Indeno(1,2,3-cd)pyrene
Concen: 31.630 ng/ul
RT: 26.474 min Scan# 3976
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34


Tgt Ion:276 Resp: 1928421
Ion Ratio Lower Upper
276 100
138 19.7 12.3 18.5#
277 25.3 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 31.699 ng/ul
RT: 26.498 min Scan# 3979
Delta R.T. 0.000 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34




Instrument :
BNA_P
ClientSampleId :
SLCS012



Tgt Ion:278 Resp: 1675812
Ion Ratio Lower Upper
278 100
139 11.7 6.6 10.0
279 23.6 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/15/2022
Supervised By :Yogesh Patel 04/22/2022



#96
Benzo(g,h,i)perylene
Concen: 31.473 ng/ul
RT: 27.262 min Scan# 4110
Delta R.T. -0.006 min
Lab File: BP009825.D
Acq: 14 Apr 2022 13:34

Tgt Ion:276 Resp: 1582176
Ion Ratio Lower Upper
276 100
138 17.2 11.6 17.4
277 23.3 15.9 23.9

