


Isophorone

Response Ratio




$$\text{Response} = 9.565\text{e-}001 * \text{Amt} - 1.243\text{e-}001$$

Coef of Det (r^2) = 0.999349 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021


Hexachlorocyclopentadiene



Response = 2.906e-001 * Amt - 1.132e-001
Coef of Det (r^2) = 0.993845 Curve Fit: Linear
Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M
Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

2-Nitroaniline

Response Ratio



$$\text{Response} = 4.303\text{e-}001 * \text{Amt} - 8.705\text{e-}002$$


Coef of Det (r²) = 0.997245 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

3-Nitroaniline

Response Ratio



$$\text{Response} = 3.231\text{e-}001 * \text{Amt} - 5.021\text{e-}002$$


Coef of Det (r^2) = 0.996010 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

4-Nitrophenol

Response Ratio



$$\text{Response} = 2.774\text{e-}001 * \text{Amt} - 8.308\text{e-}002$$


Coef of Det (r^2) = 0.994627 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

2,4-Dinitrotoluene

Response Ratio



$$\text{Response} = 5.610\text{e-}001 * \text{Amt} - 9.584\text{e-}002$$


Coef of Det (r^2) = 0.995798 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

Diethylphthalate

Response Ratio



$$\text{Response} = 1.645\text{e}+000 * \text{Amt} - 3.402\text{e}-001$$


Coef of Det (r^2) = 0.996139 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

4-Nitroaniline

Response Ratio



$$\text{Response} = 3.369\text{e-}001 * \text{Amt} - 6.618\text{e-}002$$


Coef of Det (r^2) = 0.997006 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

Azobenzene

Response Ratio



$$\text{Response} = 2.263\text{e}+000 * \text{Amt} - 2.161\text{e}-001$$


Coef of Det (r^2) = 0.997141 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021

4-Bromophenyl-phenylether

Response Ratio



$$\text{Response} = 2.697\text{e-}001 * \text{Amt} - 4.653\text{e-}002$$

Coef of Det (r^2) = 0.997672 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA P\Methods\8270E-BP041521.M

Calibration Table Last Updated: Fri Apr 16 11:02:58 2021