

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041521\
 Data File : BP005307.D
 Acq On : 15 Apr 2021 20:08
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP041521

Quant Time: Apr 16 06:09:22 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 06:09:06 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	104	0.00
2	1,4-Dioxane	40.000	35.433	11.4	93	0.00
3	Pyridine	40.000	37.191	7.0	99	0.00
4	n-Nitrosodimethylamine	40.000	36.988	7.5	98	0.00
5 S	2-Fluorophenol	80.000	78.321	2.1	99	0.00
6	Aniline	40.000	42.380	-6.0	98	0.00
7 S	Phenol-d6	80.000	81.122	-1.4	99	0.00
8	2-Chlorophenol	40.000	41.826	-4.6	103	0.00
9	Benzaldehyde	40.000	40.604	-1.5	100	0.00
10 C	Phenol	40.000	40.522	-1.3	99	0.00
11	bis(2-Chloroethyl)ether	40.000	41.700	-4.3	102	0.00
12	1,3-Dichlorobenzene	40.000	40.362	-0.9	102	0.00
13 C	1,4-Dichlorobenzene	40.000	40.916	-2.3	105	0.00
14	1,2-Dichlorobenzene	40.000	40.288	-0.7	103	0.00
15	Benzyl Alcohol	40.000	40.732	-1.8	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	43.713	-9.3	107	0.00
17	2-Methylphenol	40.000	42.996	-7.5	103	0.00
18	Hexachloroethane	40.000	38.774	3.1	101	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.268	-5.7	99	0.00
20	3+4-Methylphenols	40.000	43.628	-9.1	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	39.437	1.4	103	0.00
23 S	Nitrobenzene-d5	80.000	77.796	2.8	101	0.00
24	Nitrobenzene	40.000	38.917	2.7	103	0.00
25	Isophorone	40.000	39.058	2.4	102	0.00
26 C	2-Nitrophenol	40.000	40.624	-1.6	104	0.00
27	2,4-Dimethylphenol	40.000	41.223	-3.1	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.706	-4.3	100	0.00
29 C	2,4-Dichlorophenol	40.000	40.629	-1.6	100	0.00
30	1,2,4-Trichlorobenzene	40.000	40.249	-0.6	106	0.00
31	Naphthalene	40.000	40.393	-1.0	103	0.00
32	Benzoic acid	40.000	37.123	7.2	100	0.01
33	4-Chloroaniline	40.000	42.133	-5.3	102	0.00
34 C	Hexachlorobutadiene	40.000	38.795	3.0	105	0.00
35	Caprolactam	40.000	41.789	-4.5	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	44.274	-10.7	104	0.00
37	2-Methylnaphthalene	40.000	42.470	-6.2	102	0.00
38	1-Methylnaphthalene	40.000	42.718	-6.8	103	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.643	-6.6	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.557	-1.4	109	0.00
42 S	2,4,6-Tribromophenol	80.000	89.767	-12.2	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.445	-8.6	102	0.00
44	2,4,5-Trichlorophenol	40.000	44.190	-10.5	103	0.00
45 S	2-Fluorobiphenyl	80.000	88.126	-10.2	105	0.00
46	1,1'-Biphenyl	40.000	44.013	-10.0	104	0.00
47	2-Chloronaphthalene	40.000	43.676	-9.2	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041521\
 Data File : BP005307.D
 Acq On : 15 Apr 2021 20:08
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP041521

Quant Time: Apr 16 06:09:22 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 16 06:09:06 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.451	3.9	98	0.00
49	Acenaphthylene	40.000	43.941	-9.9	102	0.00
50	Dimethylphthalate	40.000	44.844	-12.1	102	0.00
51	2,6-Dinitrotoluene	40.000	43.101	-7.8	101	0.00
52 C	Acenaphthene	40.000	42.578	-6.4	101	0.00
53	3-Nitroaniline	40.000	40.151	-0.4	101	0.00
54 P	2,4-Dinitrophenol	40.000	41.512	-3.8	102	0.00
55	Dibenzofuran	40.000	42.425	-6.1	103	0.00
56 P	4-Nitrophenol	40.000	39.570	1.1	102	0.00
57	2,4-Dinitrotoluene	40.000	40.075	-0.2	103	0.00
58	Fluorene	40.000	43.551	-8.9	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.277	-10.7	103	0.00
60	Diethylphthalate	40.000	40.268	-0.7	103	0.00
61	4-Chlorophenyl-phenylether	40.000	44.290	-10.7	104	0.00
62	4-Nitroaniline	40.000	39.284	1.8	101	0.00
63	Azobenzene	40.000	40.535	-1.3	100	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.347	-10.9	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	43.718	-9.3	103	0.00
67	4-Bromophenyl-phenylether	40.000	40.710	-1.8	102	0.00
68	Hexachlorobenzene	40.000	42.174	-5.4	102	0.00
69	Atrazine	40.000	44.793	-12.0	104	0.00
70 C	Pentachlorophenol	40.000	43.531	-8.8	105	0.00
71	Phenanthrene	40.000	41.946	-4.9	103	0.00
72	Anthracene	40.000	43.396	-8.5	101	0.00
73	Carbazole	40.000	43.848	-9.6	102	0.00
74	Di-n-butylphthalate	40.000	45.266	-13.2	100	0.00
75 C	Fluoranthene	40.000	43.315	-8.3	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	39.886	0.3	89	0.00
78	Pyrene	40.000	42.703	-6.8	100	0.00
79 S	Terphenyl-d14	80.000	83.471	-4.3	101	0.00
80	Butylbenzylphthalate	40.000	40.274	-0.7	98	0.00
81	Benzo(a)anthracene	40.000	40.371	-0.9	102	0.00
82	3,3'-Dichlorobenzidine	40.000	41.640	-4.1	98	0.00
83	Chrysene	40.000	39.644	0.9	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.326	-3.3	102	0.00
85 c	Di-n-octyl phthalate	40.000	40.469	-1.2	100	0.00
86 I	Perylene-d12	20.000	20.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.413	-1.0	100	0.00
88	Benzo(b)fluoranthene	40.000	42.439	-6.1	103	0.00
89	Benzo(k)fluoranthene	40.000	40.298	-0.7	97	0.00
90 C	Benzo(a)pyrene	40.000	40.944	-2.4	100	0.00
91	Dibenzo(a,h)anthracene	40.000	40.443	-1.1	100	0.00
92	Benzo(g,h,i)perylene	40.000	39.890	0.3	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP041521\
Data File : BP005307.D
Acq On : 15 Apr 2021 20:08
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP041521

Quant Time: Apr 16 06:09:22 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 16 06:09:06 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0