

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005346.D
 Acq On : 20 Apr 2021 09:42
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Quant Time: Apr 20 12:13:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 12:00:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.646	152	22887	20.00	ng	0.00
21) Naphthalene-d8	10.434	136	34747	20.00	ng	0.00
39) Acenaphthene-d10	14.310	164	20389	20.00	ng	0.00
64) Phenanthrene-d10	17.075	188	48634	20.00	ng	# 0.00
76) Chrysene-d12	21.169	240	77714	20.00	ng	0.00
86) Perylene-d12	23.427	264	86471	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	11781	7.71	ng	0.00
7) Phenol-d6	6.834	99	11254	6.18	ng	0.00
23) Nitrobenzene-d5	8.810	82	11036	8.76	ng	0.00
42) 2,4,6-Tribromophenol	15.816	330	3606	10.06	ng	0.00
45) 2-Fluorobiphenyl	12.922	172	13503	8.76	ng	0.00
79) Terphenyl-d14	19.645	244	30263	7.55	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.187	88	4522	4.45	ng	# 88
3) Pyridine	3.628	79	8817	4.30	ng	# 86
4) n-Nitrosodimethylamine	3.511	42	2657	3.53	ng	# 1
6) Aniline	6.987	93	5329	3.34	ng	93
8) 2-Chlorophenol	7.216	128	5052	3.53	ng	98
9) Benzaldehyde	6.799	77	3363	4.14	ng	89
10) Phenol	6.863	94	5849	3.22	ng	90
11) bis(2-Chloroethyl)ether	7.087	93	4357	3.28	ng	87
12) 1,3-Dichlorobenzene	7.534	146	8519	4.62	ng	90
13) 1,4-Dichlorobenzene	7.681	146	8137	4.60	ng	93
14) 1,2-Dichlorobenzene	7.999	146	7808	4.68	ng	94
15) Benzyl Alcohol	7.910	79	2565	3.25	ng	# 76
16) 2,2'-oxybis(1-Chloropr...	8.175	45	2601	3.22	ng	89
17) 2-Methylphenol	8.110	107	3117	3.54	ng	# 91
18) Hexachloroethane	8.710	117	3636	5.46	ng	96
19) n-Nitroso-di-n-propyla...	8.463	70	2275	2.83	ng	# 93
20) 3+4-Methylphenols	8.440	107	3363	3.15	ng	97
22) Acetophenone	8.475	105	6124	3.75	ng	# 97
24) Nitrobenzene	8.852	77	5387	4.36	ng	# 95
25) Isophorone	9.381	82	4219	3.21	ng	# 91
26) 2-Nitrophenol	9.563	139	1633	4.87	ng	# 75
27) 2,4-Dimethylphenol	9.634	122	1862	3.79	ng	90
28) bis(2-Chloroethoxy)met...	9.863	93	2297	3.43	ng	# 93
29) 2,4-Dichlorophenol	10.104	162	2632	4.46	ng	# 67
30) 1,2,4-Trichlorobenzene	10.299	180	3706	4.56	ng	95
31) Naphthalene	10.487	128	8133	3.98	ng	# 93
33) 4-Chloroaniline	10.616	127	2292	3.46	ng	# 92
34) Hexachlorobutadiene	10.763	225	3950	5.65	ng	90
35) Caprolactam	11.410	113	301	2.09	ng	96
36) 4-Chloro-3-methylphenol	11.751	107	1780	3.04	ng	# 82
37) 2-Methylnaphthalene	12.110	142	4731	3.77	ng	# 88
38) 1-Methylnaphthalene	12.110	142	4731	3.89	ng	92
40) 1,2,4,5-Tetrachloroben...	12.481	216	3766	4.99	ng	# 95
41) Hexachlorocyclopentadiene	12.451	237	503	2.15	ng	# 67
43) 2,4,6-Trichlorophenol	12.822	196	2221	5.56	ng	84

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005346.D
 Acq On : 20 Apr 2021 09:42
 Operator : CG/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Quant Time: Apr 20 12:13:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 12:00:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.822	196	2221	5.03	ng	95
46) 1,1'-Biphenyl	13.134	154	6584	4.54	ng	90
47) 2-Chloronaphthalene	13.175	162	5551	4.70	ng	98
48) 2-Nitroaniline	13.398	65	988	3.32	ng #	60
49) Acenaphthylene	14.028	152	7856	4.39	ng #	91
50) Dimethylphthalate	13.775	163	5741	4.33	ng	99
51) 2,6-Dinitrotoluene	13.898	165	1532	4.55	ng #	77
52) Acenaphthene	14.369	154	5021	4.21	ng	91
53) 3-Nitroaniline	14.240	138	1122	4.07	ng #	61
54) 2,4-Dinitrophenol	14.481	184	630	3.35	ng #	72
55) Dibenzofuran	14.710	168	8793	4.08	ng #	97
56) 4-Nitrophenol	14.604	139	626	2.94	ng #	48
57) 2,4-Dinitrotoluene	14.698	165	1959	3.82	ng #	78
58) Fluorene	15.363	166	6652	3.77	ng #	91
59) 2,3,4,6-Tetrachlorophenol	14.951	232	2379	5.00	ng #	91
60) Diethylphthalate	15.145	149	5282	3.88	ng	98
61) 4-Chlorophenyl-phenyle...	15.363	204	4015	4.02	ng #	85
62) 4-Nitroaniline	15.410	138	1116	3.94	ng #	47
63) Azobenzene	15.657	77	5576	2.85	ng	89
65) 4,6-Dinitro-2-methylph...	15.469	198	1467	4.62	ng	72
66) n-Nitrosodiphenylamine	15.581	169	5739	4.22	ng	89
67) 4-Bromophenyl-phenylether	16.263	248	2586	4.33	ng #	85
68) Hexachlorobenzene	16.369	284	4131	5.44	ng #	95
69) Atrazine	16.545	200	2298	4.71	ng #	70
70) Pentachlorophenol	16.739	266	1697	4.60	ng #	85
71) Phenanthrene	17.110	178	12787	4.37	ng	99
72) Anthracene	17.204	178	10637	3.85	ng	96
73) Carbazole	17.486	167	9853	3.94	ng #	94
74) Di-n-butylphthalate	18.045	149	8772	3.94	ng #	95
75) Fluoranthene	19.098	202	14601	4.10	ng	91
77) Benzidine	19.292	184	6207	6.07	ng	98
78) Pyrene	19.451	202	14909	3.51	ng #	91
80) Butylbenzylphthalate	20.327	149	4276	4.31	ng #	73
81) Benzo(a)anthracene	21.157	228	19988	4.09	ng	99
82) 3,3'-Dichlorobenzidine	21.098	252	6571	5.43	ng	99
83) Chrysene	21.210	228	20979	4.31	ng	97
84) Bis(2-ethylhexyl)phtha...	21.080	149	6429	4.13	ng #	94
85) Di-n-octyl phthalate	21.957	149	11901	4.12	ng #	99
87) Indeno(1,2,3-cd)pyrene	25.733	276	32910	4.86	ng	97
88) Benzo(b)fluoranthene	22.792	252	22786	3.86	ng	96
89) Benzo(k)fluoranthene	22.792	252	22786	4.02	ng	98
90) Benzo(a)pyrene	23.327	252	21728	4.10	ng	98
91) Dibenzo(a,h)anthracene	25.745	278	27717	4.69	ng	97
92) Benzo(g,h,i)perylene	26.439	276	28056	5.05	ng #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 20 12:13:55 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 20 12:00:16 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.646 min Scan# 775
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC005


Tgt Ion:	152	Resp:	22887
Ion	Ratio	Lower	Upper
152	100		
150	151.9	119.8	179.8
115	55.4	49.0	73.4



1,4-Dioxane
 Concen: 4.45 ng
 RT: 3.187 min Scan# 17
 Delta R.T. 0.006 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	88	Resp:	4522
Ion	Ratio	Lower	Upper
88	100		
58	59.2	42.5	63.7
43	15.4	19.8	29.8#









Pyridine
Concen: 4.30 ng
RT: 3.628 min Scan# 92
Delta R.T. 0.041 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion: 79 Resp: 8817
Ion Ratio Lower Upper
79 100
52 53.0 40.3 60.5
51 42.2 20.0 30.0#



n-Nitrosodimethylamine
Concen: 3.53 ng
RT: 3.511 min Scan# 72
Delta R.T. 0.012 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion: 42 Resp: 2657
Ion Ratio Lower Upper
42 100
74 57.5 122.6 183.8#
44 1025.5 146.9 220.3#









2-Fluorophenol
 Concen: 7.71 ng
 RT: 5.246 min Scan# 367
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:	112	Resp:	11781
Ion	Ratio	Lower	Upper
112	100		
64	39.0	33.6	50.4
63	21.3	18.6	27.8



Aniline
 Concen: 3.34 ng
 RT: 6.987 min Scan# 663
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	93	Resp:	5329
Ion	Ratio	Lower	Upper
93	100		
66	39.3	27.0	40.4
65	16.7	14.4	21.6









Phenol-d6
Concen: 6.18 ng
RT: 6.834 min Scan# 637
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	99	Resp:	11254
Ion	Ratio	Lower	Upper
99	100		
42	25.4	12.9	19.3#
71	38.9	26.7	40.1



2-Chlorophenol
Concen: 3.53 ng
RT: 7.216 min Scan# 702
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	128	Resp:	5052
Ion	Ratio	Lower	Upper
128	100		
130	34.8	14.1	54.1
64	39.0	17.2	57.2









Benzaldehyde
 Concen: 4.14 ng
 RT: 6.799 min Scan# 631
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC005

Tgt Ion:	77	Resp:	3363
Ion	Ratio	Lower	Upper
77	100		
105	120.0	89.4	129.4
106	88.3	79.9	119.9




Phenol
 Concen: 3.22 ng
 RT: 6.863 min Scan# 642
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	94	Resp:	5849
Ion	Ratio	Lower	Upper
94	100		
65	24.8	8.0	48.0
66	29.4	17.3	57.3



Abundance Scan 679 (7.081 min): BP005349.D\data.ms (-673) #11





bis(2-Chloroethyl)ether
Concen: 3.28 ng
RT: 7.087 min Scan# 680
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion	Ratio	Lower	Upper
93	100		
63	58.8	49.6	89.6
95	27.7	14.4	54.4


Abundance Scan 680 (7.087 min): BP005346.D\data.ms



Abundance Scan 680 (7.087 min): BP005346.D\data.ms (-628) (-)





Abundance Scan 756 (7.534 min): BP005349.D\data.ms (-746) #12




1,3-Dichlorobenzene
Concen: 4.62 ng
RT: 7.534 min Scan# 756
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion	Ratio	Lower	Upper
146	100		
148	57.0	51.4	77.0
75	24.3	23.9	35.9


Abundance Scan 756 (7.534 min): BP005346.D\data.ms



Abundance Scan 756 (7.534 min): BP005346.D\data.ms (-705) (-)






Abundance Scan 781 (7.681 min): BP005349.D\data.ms (-773) #13




1,4-Dichlorobenzene
Concen: 4.60 ng
RT: 7.681 min Scan# 781
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	146	Resp:	8137
Ion Ratio	Lower	Upper	
146	100		
148	58.6	52.2	78.4
111	40.5	34.2	51.4







Abundance Scan 834 (7.993 min): BP005349.D\data.ms (-822) #14



1,2-Dichlorobenzene
Concen: 4.68 ng
RT: 7.999 min Scan# 835
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	146	Resp:	7808
Ion Ratio	Lower	Upper	
146	100		
148	58.1	51.5	77.3
111	40.2	34.0	51.0









Benzyl Alcohol
Concen: 3.25 ng
RT: 7.910 min Scan# 820
Delta R.T. 0.012 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	79	Resp:	2565
Ion Ratio	Lower	Upper	
79	100		
108	61.4	64.0	96.0#
77	65.5	72.7	109.1#



2,2'-oxybis(1-Chloropropane)
Concen: 3.22 ng
RT: 8.175 min Scan# 865
Delta R.T. -0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	45	Resp:	2601
Ion Ratio	Lower	Upper	
45	100		
77	30.8	7.0	47.0
79	13.9	1.0	41.0









2-Methylphenol
Concen: 3.54 ng
RT: 8.110 min Scan# 854
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	107	Resp:	3117
Ion Ratio	Lower	Upper	
107	100		
108	104.1	85.8	128.8
77	40.8	41.4	62.0#
79	37.6	38.6	57.8#



Hexachloroethane
Concen: 5.46 ng
RT: 8.710 min Scan# 956
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	117	Resp:	3636
Ion Ratio	Lower	Upper	
117	100		
119	95.2	80.1	120.1
201	118.9	98.8	148.2









n-Nitroso-di-n-propylamine
Concen: 2.83 ng
RT: 8.463 min Scan# 914
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	70	Resp:	2275
Ion	Ratio	Lower	Upper
70	100		
42	44.1	39.7	59.5
101	7.2	7.4	11.0#
130	24.8	17.2	25.8



3+4-Methylphenols
Concen: 3.15 ng
RT: 8.440 min Scan# 910
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Tgt Ion:	107	Resp:	3363
Ion	Ratio	Lower	Upper
107	100		
108	87.0	69.5	109.5
77	44.3	21.5	61.5
79	39.2	16.8	56.8





Abundance Scan 1249 (10.434 min): BP005349.D\data.ms (-1198) (-)

Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1249
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005



Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.4	8.9	13.3
54	5.2	5.0	7.4
68	5.4	3.7	5.5





Abundance Scan 1249 (10.434 min): BP005346.D\data.ms (-1198) (-)




Abundance Scan 916 (8.475 min): BP005349.D\data.ms (-902) (-)

Acetophenone
Concen: 3.75 ng
RT: 8.475 min Scan# 916
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion	Ratio	Lower	Upper
105	100		
71	1.6	0.9	1.3#
51	25.1	18.2	27.2
120	21.6	17.0	25.6




Abundance Scan 916 (8.475 min): BP005346.D\data.ms (-865) (-)






Abundance Scan 973 (8.811 min): BP005349.D\data.ms (-964) #23

Nitrobenzene-d5
Concen: 8.76 ng
RT: 8.810 min Scan# 973
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005







Tgt Ion:	82	Resp:	11036
Ion	Ratio	Lower	Upper
82	100		
128	40.2	30.5	45.7
54	47.0	33.9	50.9




Abundance Scan 980 (8.852 min): BP005349.D\data.ms (-967) #24

Nitrobenzene
Concen: 4.36 ng
RT: 8.852 min Scan# 980
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	77	Resp:	5387
Ion	Ratio	Lower	Upper
77	100		
123	40.8	31.7	47.5
65	20.0	11.5	17.3#



Abundance Scan 1069 (9.375 min): BP005349.D\data.ms (-10625-)





Isophorone
Concen: 3.21 ng
RT: 9.381 min Scan# 1070
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Tgt Ion:	82	Resp:	4219
Ion	Ratio	Lower	Upper
82	100		
95	14.6	7.5	11.3#
138	21.3	14.6	22.0

Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Abundance Scan 1070 (9.381 min): BP005346.D\data.ms



Abundance Scan 1070 (9.381 min): BP005346.D\data.ms (-1018-)





Abundance Scan 1100 (9.558 min): BP005349.D\data.ms (-10826-)




2-Nitrophenol
Concen: 4.87 ng
RT: 9.563 min Scan# 1101
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	139	Resp:	1633
Ion	Ratio	Lower	Upper
139	100		
109	36.4	40.7	61.1#
65	38.5	47.8	71.6#


Abundance Scan 1101 (9.563 min): BP005346.D\data.ms



Abundance Scan 1101 (9.563 min): BP005346.D\data.ms (-1049-)



Abundance Scan 1111 (9.622 min): BP005349.D\data.ms (-11037-)





2,4-Dimethylphenol
Concen: 3.79 ng
RT: 9.634 min Scan# 1113
Delta R.T. 0.012 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	122	Resp:	1862
Ion	Ratio	Lower	Upper
122	100		
107	101.8	93.8	140.8
121	56.2	45.6	68.4


Abundance Scan 1113 (9.634 min): BP005346.D\data.ms



Abundance Scan 1113 (9.634 min): BP005346.D\data.ms (-1060-)





Abundance Scan 1151 (9.858 min): BP005349.D\data.ms (-11428-)




bis(2-Chloroethoxy)methane
Concen: 3.43 ng
RT: 9.863 min Scan# 1152
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	93	Resp:	2297
Ion	Ratio	Lower	Upper
93	100		
95	29.2	24.9	37.3
123	5.2	9.5	14.3#


Abundance Scan 1152 (9.863 min): BP005346.D\data.ms



Abundance Scan 1152 (9.863 min): BP005346.D\data.ms (-1100-)



Abundance Scan 1191 (10.093 min): BP005349.D\data.ms (-1129) (-)





2,4-Dichlorophenol
Concen: 4.46 ng
RT: 10.104 min Scan# 1193
Delta R.T. 0.012 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	162	Resp:	2632
Ion Ratio	Lower	Upper	
162	100		
164	33.1	44.1	84.1#
98	22.0	12.6	52.6


Abundance Scan 1193 (10.104 min): BP005346.D\data.ms



Abundance Scan 1193 (10.104 min): BP005346.D\data.ms (-1140) (-)





Abundance Scan 1226 (10.299 min): BP005349.D\data.ms (-1130) (-)




1,2,4-Trichlorobenzene
Concen: 4.56 ng
RT: 10.299 min Scan# 1226
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Tgt Ion:	180	Resp:	3706
Ion Ratio	Lower	Upper	
180	100		
182	101.5	77.2	115.8
145	25.4	21.7	32.5

Abundance Scan 1226 (10.299 min): BP005346.D\data.ms



Abundance Scan 1226 (10.299 min): BP005346.D\data.ms (-1175) (-)









Scan 1258 (10.487 min): BP005349.D\data.ms (-128) (-)
 Naphthalene
 Concen: 3.98 ng
 RT: 10.487 min Scan# 1258
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.0	12.0
127	18.4	11.1	16.7#



Scan 1279 (10.610 min): BP005349.D\data.ms (-127) (-)
 4-Chloroaniline
 Concen: 3.46 ng
 RT: 10.616 min Scan# 1280
 Delta R.T. 0.006 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion	Ratio	Lower	Upper
127	100		
129	29.3	25.5	38.3
65	31.9	30.0	45.0
92	31.0	20.2	30.2#









Hexachlorobutadiene
Concen: 5.65 ng
RT: 10.763 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	225	Resp:	3950
Ion Ratio	Lower	Upper	
225	100		
223	73.5	51.1	76.7
227	69.1	50.2	75.2



Caprolactam
Concen: 2.09 ng
RT: 11.410 min Scan# 1415
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Tgt Ion:	113	Resp:	301
Ion Ratio	Lower	Upper	
113	100		
55	87.1	72.0	112.0
56	88.5	65.1	105.1






Abundance Scan 1473 (11.752 min): BP005349.D\data.ms (-1485) (-)

4-Chloro-3-methylphenol
Concen: 3.04 ng
RT: 11.751 min Scan# 1473
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampled :
SSTDIC005




Tgt Ion:	107	Resp:	1780
Ion Ratio	Lower	Upper	
107	100		
144	39.0	20.1	30.1#
142	65.4	62.2	93.4






Abundance Scan 1534 (12.110 min): BP005349.D\data.ms (-1535) (-)

2-Methylnaphthalene
Concen: 3.77 ng
RT: 12.110 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42



Tgt Ion:	142	Resp:	4731
Ion Ratio	Lower	Upper	
142	100		
141	88.4	74.7	112.1
115	37.8	43.8	65.8#



Abundance Scan 1571 (12.328 min): BP005349.D\data.ms (-1534) (-)

1-Methylnaphthalene

Concen: 3.89 ng

RT: 12.110 min Scan# 1534

Delta R.T. -0.218 min

Lab File: BP005346.D


Acq: 20 Apr 2021 09:42

Instrument :

BNA_P

ClientSampleId :

SSTDIC005






Tgt Ion:142 Resp: 4731

Ion Ratio Lower Upper

142 100

141 88.4 77.0 115.6

116 5.9 5.0 7.6



Abundance Scan 1907 (14.304 min): BP005349.D\data.ms (-1908) (-)

Acenaphthene-d10

Concen: 20.00 ng

RT: 14.310 min Scan# 1908

Delta R.T. 0.006 min

Lab File: BP005346.D

Acq: 20 Apr 2021 09:42




Tgt Ion:164 Resp: 20389


Ion Ratio Lower Upper

164 100

162 100.9 73.5 110.3

160 55.1 40.6 61.0









Scan 1597 (12.481 min): BP005349.D\data.ms (-1546) (-)
 1,2,4,5-Tetrachlorobenzene
 Concen: 4.99 ng
 RT: 12.481 min Scan# 1597
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion	Ratio	Lower	Upper
216	100		
214	78.8	60.6	91.0
179	25.0	15.3	22.9#
108	16.1	13.8	20.6



Scan 1592 (12.452 min): BP005349.D\data.ms (-1546) (-)
 Hexachlorocyclopentadiene
 Concen: 2.15 ng
 RT: 12.451 min Scan# 1592
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion	Ratio	Lower	Upper
237	100		
235	89.9	40.9	80.9#
272	9.8	0.0	34.6









Scan 2163 (15.810 min): BP005349.D\data.ms (-2154) (-)
 2,4,6-Tribromophenol
 Concen: 10.06 ng
 RT: 15.816 min Scan# 2164
 Delta R.T. 0.006 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC005

Tgt Ion:	330	Resp:	3606
Ion Ratio	Lower	Upper	
330	100		
332	101.2	76.3	114.5
141	24.2	22.8	34.2



Scan 1640 (12.734 min): BP005349.D\data.ms (-1623) (-)
 2,4,6-Trichlorophenol
 Concen: 5.56 ng
 RT: 12.822 min Scan# 1655
 Delta R.T. 0.088 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42


Tgt Ion:	196	Resp:	2221
Ion Ratio	Lower	Upper	
196	100		
198	81.8	81.4	122.2
200	32.2	28.2	42.2





Abundance Scan 1654 (12.816 min): BP005349.D\data.ms (-1647) (-)

2,4,5-Trichlorophenol
Concen: 5.03 ng
RT: 12.822 min Scan# 1655
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampled :
SSTDIC005



Tgt Ion:	196	Resp:	2221
Ion	Ratio	Lower	Upper
196	100		
198	81.8	71.0	106.4
97	46.7	38.5	57.7
132	21.4	15.6	23.4




Abundance Scan 1655 (12.822 min): BP005346.D\data.ms (-1603) (-)





Abundance Scan 1672 (12.922 min): BP005349.D\data.ms (-1655) (-)


2-Fluorobiphenyl
Concen: 8.76 ng
RT: 12.922 min Scan# 1672
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Tgt Ion:	172	Resp:	13503
Ion	Ratio	Lower	Upper
172	100		
171	35.2	28.1	42.1
170	25.3	19.6	29.4



Abundance Scan 1672 (12.922 min): BP005346.D\data.ms (-1621) (-)






Abundance Scan 1708 (13.134 min): BP005349.D\data.ms (-1657) (-)




1,1'-Biphenyl
Concen: 4.54 ng
RT: 13.134 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	154	Resp:	6584
Ion Ratio	Lower	Upper	
154	100		
153	33.0	18.9	58.9
76	7.9	0.0	32.1







Abundance Scan 1715 (13.175 min): BP005349.D\data.ms (-1664) (-)



2-Chloronaphthalene
Concen: 4.70 ng
RT: 13.175 min Scan# 1715
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	162	Resp:	5551
Ion Ratio	Lower	Upper	
162	100		
127	40.2	31.4	47.0
164	31.5	26.6	39.8









2-Nitroaniline
 Concen: 3.32 ng
 RT: 13.398 min Scan# 1753
 Delta R.T. 0.006 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Tgt Ion:	65	Resp:	988
Ion	Ratio	Lower	Upper
65	100		
92	79.2	48.8	73.2#
138	134.7	69.5	104.3#



Acenaphthylene
 Concen: 4.39 ng
 RT: 14.028 min Scan# 1860
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42


Tgt Ion:	152	Resp:	7856
Ion	Ratio	Lower	Upper
152	100		
151	15.7	15.8	23.6#
153	16.9	10.5	15.7#





Abundance Scan 1817 (13.775 min): BP005349.D\data.ms (-1850) (-)

Dimethylphthalate
Concen: 4.33 ng
RT: 13.775 min Scan# 1817
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005



Tgt Ion:	163	Resp:	5741
Ion Ratio	Lower	Upper	
163	100		
194	4.5	3.9	5.9
164	10.6	8.1	12.1





Abundance Scan 1817 (13.775 min): BP005346.D\data.ms (-1766) (-)




Abundance Scan 1838 (13.898 min): BP005349.D\data.ms (-1850) (-)

2,6-Dinitrotoluene
Concen: 4.55 ng
RT: 13.898 min Scan# 1838
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Tgt Ion:	165	Resp:	1532
Ion Ratio	Lower	Upper	
165	100		
63	70.5	67.4	101.0
89	63.0	73.8	110.6#



Abundance Scan 1838 (13.898 min): BP005346.D\data.ms (-1787) (-)






Abundance Scan 1919 (14.375 min): BP005349.D\data.ms (-152) (-)




Acenaphthene
Concen: 4.21 ng
RT: 14.369 min Scan# 1918
Delta R.T. -0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	154	Resp:	5021
Ion Ratio	Lower	Upper	
154	100		
153	101.6	90.9	136.3
152	56.4	48.5	72.7






Abundance Scan 1895 (14.234 min): BP005349.D\data.ms (-153) (-)




3-Nitroaniline
Concen: 4.07 ng
RT: 14.240 min Scan# 1896
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	138	Resp:	1122
Ion Ratio	Lower	Upper	
138	100		
108	8.6	10.2	15.4#
92	85.9	108.1	162.1#



Abundance Scan 1932 (14.451 min): BP005349.D\data.ms (-1932) (-)





2,4-Dinitrophenol
Concen: 3.35 ng
RT: 14.481 min Scan# 1937
Delta R.T. 0.029 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	184	Resp:	630
Ion Ratio	Lower	Upper	
184	100		
63	46.4	51.2	76.8#
154	43.0	55.8	83.8#


Abundance Scan 1937 (14.481 min): BP005346.D\data.ms



Abundance Scan 1937 (14.481 min): BP005346.D\data.ms (-1881) (-)





Abundance Scan 1976 (14.710 min): BP005349.D\data.ms (-1955) (-)




Dibenzofuran
Concen: 4.08 ng
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	168	Resp:	8793
Ion Ratio	Lower	Upper	
168	100		
139	38.1	31.4	47.0
169	11.2	11.4	17.0#


Abundance Scan 1976 (14.710 min): BP005346.D\data.ms



Abundance Scan 1976 (14.710 min): BP005346.D\data.ms (-1925) (-)



Abundance Scan 1951 (14.563 min): BP005349.D\data.ms (-1951) (-)





4-Nitrophenol
Concen: 2.94 ng
RT: 14.604 min Scan# 1958
Delta R.T. 0.041 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	139	Resp:	626
Ion	Ratio	Lower	Upper
139	100		
109	82.2	122.8	162.8#
65	65.8	110.0	150.0#


Abundance Scan 1958 (14.604 min): BP005346.D\data.ms



Abundance Scan 1958 (14.604 min): BP005346.D\data.ms (-1900) (-)





Abundance Scan 1973 (14.693 min): BP005349.D\data.ms (-1959) (-)




2,4-Dinitrotoluene
Concen: 3.82 ng
RT: 14.698 min Scan# 1974
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	165	Resp:	1959
Ion	Ratio	Lower	Upper
165	100		
63	51.9	56.7	85.1#
89	84.0	85.8	128.6#
182	5.3	3.0	4.6#

Abundance Scan 1974 (14.698 min): BP005346.D\data.ms




Abundance Scan 1974 (14.698 min): BP005346.D\data.ms (-1922) (-)






Abundance Scan 2087 (15.363 min): BP005349.D\data.ms (-2058) (-)

Fluorene
Concen: 3.77 ng
RT: 15.363 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005




Tgt Ion:	166	Resp:	6652
Ion Ratio	Lower	Upper	
166	100		
165	107.3	78.8	118.2
167	9.0	10.9	16.3#






Abundance Scan 2016 (14.946 min): BP005349.D\data.ms (-2059) (-)

2,3,4,6-Tetrachlorophenol
Concen: 5.00 ng
RT: 14.951 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Tgt Ion:	232	Resp:	2379
Ion Ratio	Lower	Upper	
232	100		
131	27.2	26.5	39.7
130	1.4	2.1	3.1#
166	20.0	19.4	29.0





Abundance Scan 2050 (15.146 min): BP005349.D\data.ms (-2069) (-)

Diethylphthalate
Concen: 3.88 ng
RT: 15.145 min Scan# 2050
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005



Tgt Ion:	149	Resp:	5282
Ion Ratio	Lower	Upper	
149	100		
177	23.7	19.2	28.8
150	10.1	9.9	14.9




Abundance Scan 2050 (15.145 min): BP005346.D\data.ms (-1999) (-)





Abundance Scan 2087 (15.363 min): BP005349.D\data.ms (-2069) (-)


4-Chlorophenyl-phenylether
Concen: 4.02 ng
RT: 15.363 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42



Tgt Ion:	204	Resp:	4015
Ion Ratio	Lower	Upper	
204	100		
206	27.3	26.0	39.0
141	41.7	44.2	66.2#




Abundance Scan 2087 (15.363 min): BP005346.D\data.ms (-2036) (-)





Abundance Scan 2094 (15.404 min): BP005349.D\data.ms (-2043) (-)

4-Nitroaniline
Concen: 3.94 ng
RT: 15.410 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005



Tgt Ion:	138	Resp:	1116
Ion	Ratio	Lower	Upper
138	100		
92	31.1	50.9	90.9#
108	98.6	154.7	194.7#





Abundance Scan 2095 (15.410 min): BP005346.D\data.ms (-2043) (-)




Abundance Scan 2137 (15.657 min): BP005349.D\data.ms (-2128) (-)

Azobenzene
Concen: 2.85 ng
RT: 15.657 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Tgt Ion:	77	Resp:	5576
Ion	Ratio	Lower	Upper
77	100		
182	34.6	5.5	45.5
105	20.6	0.0	36.5
51	27.5	10.3	50.3



Abundance Scan 2137 (15.657 min): BP005346.D\data.ms (-2086) (-)






Abundance Scan 2377 (17.069 min): BP005349.D\data.ms (-2367) (-)




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.075 min Scan# 2378
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	188	Resp:	48634
Ion Ratio	Lower	Upper	
188	100		
94	5.3	5.0	7.6
80	5.5	5.9	8.9#





Abundance Scan 2104 (15.463 min): BP005349.D\data.ms (-2095) (-)





4,6-Dinitro-2-methylphenol
Concen: 4.62 ng
RT: 15.469 min Scan# 2105
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	198	Resp:	1467
Ion Ratio	Lower	Upper	
198	100		
51	18.3	10.3	50.3
105	14.5	13.6	53.6



Abundance Scan 2105 (15.469 min): BP005346.D\data.ms (-2053) (-)









#169 (-)
 n-Nitrosodiphenylamine
 Concen: 4.22 ng
 RT: 15.581 min Scan# 2124
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC005


Tgt Ion:	169	Resp:	5739
Ion Ratio	Lower	Upper	
169	100		
168	80.3	55.2	82.8
167	40.7	36.6	54.8



#248 (-)
 4-Bromophenyl-phenylether
 Concen: 4.33 ng
 RT: 16.263 min Scan# 2240
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	248	Resp:	2586
Ion Ratio	Lower	Upper	
248	100		
250	120.8	77.5	116.3#
141	61.6	50.1	75.1









Hexachlorobenzene
Concen: 5.44 ng
RT: 16.369 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	284	Resp:	4131
Ion Ratio	Lower	Upper	
284	100		
142	20.2	20.8	31.2#
249	29.6	23.7	35.5




Atrazine
Concen: 4.71 ng
RT: 16.545 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	200	Resp:	2298
Ion Ratio	Lower	Upper	
200	100		
173	30.4	0.7	40.7
215	72.2	29.0	69.0#





Abundance Scan 2319 (16.728 min): BP005349.D\data.ms (-2319) (-)




Pentachlorophenol
Concen: 4.60 ng
RT: 16.739 min Scan# 2321
Delta R.T. 0.012 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	266	Resp:	1697
Ion	Ratio	Lower	Upper
266	100		
268	60.3	49.4	74.0
264	45.8	53.7	80.5#



Abundance Scan 2321 (16.739 min): BP005346.D\data.ms (-2268) (-)





Abundance Scan 2384 (17.110 min): BP005349.D\data.ms (-2371) (-)





Phenanthrene
Concen: 4.37 ng
RT: 17.110 min Scan# 2384
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	178	Resp:	12787
Ion	Ratio	Lower	Upper
178	100		
176	18.9	15.4	23.2
179	15.6	12.3	18.5



Abundance Scan 2384 (17.110 min): BP005346.D\data.ms (-2333) (-)









Scan 2400 (17.204 min): BP005349.D\data.ms (-2393) (-)
 Anthracene
 Concen: 3.85 ng
 RT: 17.204 min Scan# 2400
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Tgt Ion:	178	Resp:	10637
Ion Ratio	Lower	Upper	
178	100		
176	17.5	15.2	22.8
179	13.7	12.6	19.0




Scan 2448 (17.487 min): BP005349.D\data.ms (-2459) (-)
 Carbazole
 Concen: 3.94 ng
 RT: 17.486 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	167	Resp:	9853
Ion Ratio	Lower	Upper	
167	100		
166	19.3	17.4	26.0
139	9.9	10.4	15.6#






Abundance Scan 2542 (18.039 min): BP005349.D\data.ms (-2554) (-)




Di-n-butylphthalate
Concen: 3.94 ng
RT: 18.045 min Scan# 2543
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	149	Resp:	8772
Ion Ratio	Lower	Upper	
149	100		
150	8.2	7.6	11.4
104	13.7	8.9	13.3#

Instrument :
BNA_P
ClientSampleId :
SSTDIC005







Abundance Scan 2721 (19.092 min): BP005349.D\data.ms (-2774) (-)



Fluoranthene
Concen: 4.10 ng
RT: 19.098 min Scan# 2722
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	202	Resp:	14601
Ion Ratio	Lower	Upper	
202	100		
101	4.4	0.0	26.1
203	12.8	0.0	37.2









Chrysene-d12
 Concen: 20.00 ng
 RT: 21.169 min Scan# 3074
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Tgt Ion:	240	Resp:	77714
Ion	Ratio	Lower	Upper
240	100		
120	4.5	4.2	6.2
236	26.7	20.2	30.2




Benzidine
 Concen: 6.07 ng
 RT: 19.292 min Scan# 2755
 Delta R.T. 0.006 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	184	Resp:	6207
Ion	Ratio	Lower	Upper
184	100		
185	14.9	11.1	16.7
183	11.8	9.8	14.8






Abundance Scan 2782 (19.451 min): BP005349.D\data.ms (-2773) (-)




Pyrene
Concen: 3.51 ng
RT: 19.451 min Scan# 2782
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005

Tgt Ion:	202	Resp:	14909
Ion Ratio	Lower	Upper	
202	100		
200	22.1	16.9	25.3
203	24.5	13.8	20.6#







Abundance Scan 2815 (19.645 min): BP005349.D\data.ms (-2809) (-)



Terphenyl-d14
Concen: 7.55 ng
RT: 19.645 min Scan# 2815
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	244	Resp:	30263
Ion Ratio	Lower	Upper	
244	100		
212	6.4	5.8	8.6
122	5.6	5.2	7.8









Scan 2931 (20.328 min): BP005349.D\data.ms (-2880) (-)
 Butylbenzylphthalate
 Concen: 4.31 ng
 RT: 20.327 min Scan# 2931
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Tgt Ion:	149	Resp:	4276
Ion	Ratio	Lower	Upper
149	100		
91	46.7	60.2	90.4#
206	36.5	23.9	35.9#




Scan 3072 (21.157 min): BP005349.D\data.ms (-3065) (-)
 Benzo(a)anthracene
 Concen: 4.09 ng
 RT: 21.157 min Scan# 3072
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	228	Resp:	19988
Ion	Ratio	Lower	Upper
228	100		
226	26.3	21.2	31.8
229	19.9	15.6	23.4





Abundance Scan 3061 (21.092 min): BP005349.D\data.ms (-3061) (-)




3,3'-Dichlorobenzidine
Concen: 5.43 ng
RT: 21.098 min Scan# 3062
Delta R.T. 0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	252	Resp:	6571
Ion Ratio	Lower	Upper	
252	100		
254	63.3	49.9	74.9
126	7.8	5.5	8.3



Abundance Scan 3062 (21.098 min): BP005346.D\data.ms (-3010) (-)





Abundance Scan 3081 (21.210 min): BP005349.D\data.ms (-3081) (-)





Chrysene
Concen: 4.31 ng
RT: 21.210 min Scan# 3081
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	228	Resp:	20979
Ion Ratio	Lower	Upper	
228	100		
226	28.9	24.8	37.2
229	19.0	15.4	23.0



Abundance Scan 3081 (21.210 min): BP005346.D\data.ms (-3030) (-)









Scan 3059 (21.081 min): BP005349.D\data.ms (-3059) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 4.13 ng
 RT: 21.080 min Scan# 3059
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005


Tgt Ion:	149	Resp:	6429
Ion	Ratio	Lower	Upper
149	100		
167	27.1	23.8	35.8
279	10.7	5.7	8.5#



Scan 3208 (21.957 min): BP005349.D\data.ms (-3208) (-)
 Di-n-octyl phthalate
 Concen: 4.12 ng
 RT: 21.957 min Scan# 3208
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Tgt Ion:	149	Resp:	11901
Ion	Ratio	Lower	Upper
149	100		
167	2.5	1.5	2.3#
43	6.2	5.1	7.7









Perylene-d12
Concen: 20.00 ng
RT: 23.427 min Scan# 3458
Delta R.T. -0.000 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Instrument :
BNA_P
ClientSampleId :
SSTDIC005


Tgt Ion:	264	Resp:	86471
Ion Ratio	Lower	Upper	
264	100		
260	22.8	20.6	30.8
265	21.7	19.0	28.6



Indeno(1,2,3-cd)pyrene
Concen: 4.86 ng
RT: 25.733 min Scan# 3850
Delta R.T. -0.012 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42

Tgt Ion:	276	Resp:	32910
Ion Ratio	Lower	Upper	
276	100		
138	9.5	8.4	12.6
277	24.4	20.6	31.0









Benzo(b)fluoranthene
 Concen: 3.86 ng
 RT: 22.792 min Scan# 3350
 Delta R.T. 0.047 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Tgt Ion:	252	Resp:	22786
Ion Ratio	Lower	Upper	
252	100		
253	20.4	18.2	27.2
125	4.4	3.4	5.0



Benzo(k)fluoranthene
 Concen: 4.02 ng
 RT: 22.792 min Scan# 3350
 Delta R.T. -0.000 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42


Tgt Ion:	252	Resp:	22786
Ion Ratio	Lower	Upper	
252	100		
253	20.4	17.2	25.8
125	4.4	3.8	5.6





Abundance Scan 3442 (23.333 min): BP005349.D\data.ms (-3490) (-)

Benzo(a)pyrene
Concen: 4.10 ng
RT: 23.327 min Scan# 3441
Delta R.T. -0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42


Instrument :
BNA_P
ClientSampleId :
SSTDIC005



Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.1	25.7
125	5.3	3.8	5.6




Abundance Scan 3441 (23.327 min): BP005346.D\data.ms (-3391) (-)





Abundance Scan 3853 (25.751 min): BP005349.D\data.ms (-3890) (-)


Dibenzo(a,h)anthracene
Concen: 4.69 ng
RT: 25.745 min Scan# 3852
Delta R.T. -0.006 min
Lab File: BP005346.D
Acq: 20 Apr 2021 09:42




Tgt Ion	Ratio	Lower	Upper
278	100		
139	7.2	5.0	7.6
279	25.0	18.6	28.0



Abundance Scan 3852 (25.745 min): BP005346.D\data.ms (-3802) (-)





Abundance Scan 3971 (26.445 min): BP005349.D\data.ms (-3990) (-)



Benzo(g,h,i)perylene
 Concen: 5.05 ng
 RT: 26.439 min Scan# 3970
 Delta R.T. -0.006 min
 Lab File: BP005346.D
 Acq: 20 Apr 2021 09:42

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC005

Tgt Ion:	276	Resp:	28056
Ion Ratio		Lower	Upper
276	100		
277	26.0	19.8	29.8
138	7.6	7.7	11.5#



Abundance Scan 3970 (26.439 min): BP005346.D\data.ms (-3920) (-)

