

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005348.D
 Acq On : 20 Apr 2021 10:50
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Quant Time: Apr 20 12:04:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 12:00:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.646	152	23651	20.00	ng	0.00
21) Naphthalene-d8	10.434	136	34266	20.00	ng	# 0.00
39) Acenaphthene-d10	14.310	164	27829	20.00	ng	0.00
64) Phenanthrene-d10	17.069	188	76199	20.00	ng	0.00
76) Chrysene-d12	21.174	240	126062	20.00	ng	# 0.00
86) Perylene-d12	23.427	264	130577	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	63355	40.15	ng	0.00
7) Phenol-d6	6.834	99	73938	39.26	ng	0.00
23) Nitrobenzene-d5	8.804	82	56881	45.79	ng	0.00
42) 2,4,6-Tribromophenol	15.810	330	27416	56.03	ng	0.00
45) 2-Fluorobiphenyl	12.922	172	82388	39.15	ng	0.00
79) Terphenyl-d14	19.645	244	267830	41.20	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.187	88	21793	20.77	ng	96
3) Pyridine	3.593	79	47630	22.47	ng	96
4) n-Nitrosodimethylamine	3.499	42	14382	18.48	ng	# 33
6) Aniline	6.987	93	29901	18.12	ng	95
8) 2-Chlorophenol	7.216	128	29867	20.21	ng	97
9) Benzaldehyde	6.799	77	19599	23.36	ng	95
10) Phenol	6.857	94	36415	19.41	ng	95
11) bis(2-Chloroethyl)ether	7.081	93	26881	19.57	ng	97
12) 1,3-Dichlorobenzene	7.534	146	39435	20.69	ng	99
13) 1,4-Dichlorobenzene	7.681	146	38427	21.01	ng	97
14) 1,2-Dichlorobenzene	7.993	146	35907	20.82	ng	97
15) Benzyl Alcohol	7.899	79	14415	17.66	ng	95
16) 2,2'-oxybis(1-Chloropr...	8.175	45	14056	16.83	ng	98
17) 2-Methylphenol	8.104	107	16732	18.38	ng	95
18) Hexachloroethane	8.710	117	13995	20.32	ng	92
19) n-Nitroso-di-n-propyla...	8.452	70	14232	17.14	ng	96
20) 3+4-Methylphenols	8.434	107	19048	17.28	ng	97
22) Acetophenone	8.475	105	37833	23.51	ng	98
24) Nitrobenzene	8.852	77	27084	22.24	ng	95
25) Isophorone	9.375	82	24502	18.90	ng	96
26) 2-Nitrophenol	9.557	139	7462	22.56	ng	92
27) 2,4-Dimethylphenol	9.628	122	10373	21.40	ng	98
28) bis(2-Chloroethoxy)met...	9.857	93	12369	18.70	ng	# 89
29) 2,4-Dichlorophenol	10.093	162	12360	21.24	ng	95
30) 1,2,4-Trichlorobenzene	10.299	180	18063	22.52	ng	96
31) Naphthalene	10.481	128	40506	20.08	ng	98
32) Benzoic acid	9.763	122	7103	21.49	ng	96
33) 4-Chloroaniline	10.610	127	12452	19.09	ng	97
34) Hexachlorobutadiene	10.757	225	17233	24.98	ng	94
35) Caprolactam	11.393	113	2531	17.84	ng	85
36) 4-Chloro-3-methylphenol	11.745	107	10181	17.61	ng	96
37) 2-Methylnaphthalene	12.110	142	23520	19.02	ng	99
38) 1-Methylnaphthalene	12.328	142	23146	19.27	ng	98
40) 1,2,4,5-Tetrachloroben...	12.481	216	20842	20.25	ng	96
41) Hexachlorocyclopentadiene	12.451	237	6572	20.60	ng	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005348.D
 Acq On : 20 Apr 2021 10:50
 Operator : CG/JU
 Sample : SSTDICC020
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Quant Time: Apr 20 12:04:41 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 12:00:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.734	196	10483	19.22	ng	96
44) 2,4,5-Trichlorophenol	12.816	196	12153	20.15	ng	# 94
46) 1,1'-Biphenyl	13.134	154	36522	18.44	ng	99
47) 2-Chloronaphthalene	13.175	162	31273	19.39	ng	95
48) 2-Nitroaniline	13.392	65	6554	16.14	ng	# 83
49) Acenaphthylene	14.028	152	47726	19.54	ng	99
50) Dimethylphthalate	13.775	163	33774	18.68	ng	# 95
51) 2,6-Dinitrotoluene	13.898	165	9126	19.88	ng	83
52) Acenaphthene	14.369	154	31481	19.34	ng	96
53) 3-Nitroaniline	14.234	138	6467	17.18	ng	# 95
54) 2,4-Dinitrophenol	14.451	184	5503	21.44	ng	89
55) Dibenzofuran	14.710	168	63989	21.74	ng	99
56) 4-Nitrophenol	14.569	139	5357	18.46	ng	# 87
57) 2,4-Dinitrotoluene	14.698	165	14119	20.15	ng	94
58) Fluorene	15.363	166	50482	20.97	ng	96
59) 2,3,4,6-Tetrachlorophenol	14.945	232	15746	24.23	ng	97
60) Diethylphthalate	15.145	149	35880	19.30	ng	99
61) 4-Chlorophenyl-phenyle...	15.363	204	31715	23.28	ng	96
62) 4-Nitroaniline	15.404	138	7028	18.17	ng	# 81
63) Azobenzene	15.657	77	55612	20.84	ng	97
65) 4,6-Dinitro-2-methylph...	15.463	198	10540	21.20	ng	100
66) n-Nitrosodiphenylamine	15.581	169	40192	18.86	ng	99
67) 4-Bromophenyl-phenylether	16.263	248	20831	22.26	ng	95
68) Hexachlorobenzene	16.369	284	27946	23.48	ng	96
69) Atrazine	16.545	200	15966	20.90	ng	92
70) Pentachlorophenol	16.728	266	13157	22.77	ng	97
71) Phenanthrene	17.110	178	96595	21.08	ng	98
72) Anthracene	17.204	178	91033	21.03	ng	99
73) Carbazole	17.486	167	79579	20.31	ng	100
74) Di-n-butylphthalate	18.039	149	61857	17.72	ng	97
75) Fluoranthene	19.098	202	118960	21.31	ng	98
77) Benzidine	19.286	184	29421	17.74	ng	98
78) Pyrene	19.451	202	130366	18.90	ng	99
80) Butylbenzylphthalate	20.327	149	27563	17.12	ng	91
81) Benzo(a)anthracene	21.157	228	160319	20.22	ng	99
82) 3,3'-Dichlorobenzidine	21.092	252	38502	19.60	ng	94
83) Chrysene	21.210	228	161327	20.44	ng	98
84) Bis(2-ethylhexyl)phtha...	21.080	149	39415	15.61	ng	# 91
85) Di-n-octyl phthalate	21.957	149	77214	16.49	ng	100
87) Indeno(1,2,3-cd)pyrene	25.739	276	220298	21.55	ng	100
88) Benzo(b)fluoranthene	22.745	252	185413	20.78	ng	99
89) Benzo(k)fluoranthene	22.745	252	185413	21.67	ng	99
90) Benzo(a)pyrene	23.327	252	166368	20.77	ng	99
91) Dibenzo(a,h)anthracene	25.745	278	193224	21.64	ng	100
92) Benzo(g,h,i)perylene	26.439	276	179641	21.43	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 20 12:04:41 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 20 12:00:16 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.646 min Scan# 775
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC020


Tgt Ion:	152	Resp:	23651
Ion	Ratio	Lower	Upper
152	100		
150	153.1	119.8	179.8
115	55.1	49.0	73.4



1,4-Dioxane
 Concen: 20.77 ng
 RT: 3.187 min Scan# 17
 Delta R.T. 0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	88	Resp:	21793
Ion	Ratio	Lower	Upper
88	100		
58	54.3	42.5	63.7
43	29.4	19.8	29.8









Pyridine
Concen: 22.47 ng
RT: 3.593 min Scan# 86
Delta R.T. 0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Instrument :
BNA_P
ClientSampleId :
SSTDICC020


Tgt Ion:	79	Resp:	47630
Ion	Ratio	Lower	Upper
79	100		
52	52.2	40.3	60.5
51	27.7	20.0	30.0



n-Nitrosodimethylamine
Concen: 18.48 ng
RT: 3.499 min Scan# 70
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	42	Resp:	14382
Ion	Ratio	Lower	Upper
42	100		
74	129.5	122.6	183.8
44	337.6	146.9	220.3#









2-Fluorophenol
 Concen: 40.15 ng
 RT: 5.246 min Scan# 367
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	112	Resp:	63355
Ion	Ratio	Lower	Upper
112	100		
64	40.6	33.6	50.4
63	21.1	18.6	27.8



Aniline
 Concen: 18.12 ng
 RT: 6.987 min Scan# 663
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	93	Resp:	29901
Ion	Ratio	Lower	Upper
93	100		
66	30.2	27.0	40.4
65	17.1	14.4	21.6









Phenol-d6
 Concen: 39.26 ng
 RT: 6.834 min Scan# 637
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	99	Resp:	73938
Ion	Ratio	Lower	Upper
99	100		
42	18.8	12.9	19.3
71	33.7	26.7	40.1



2-Chlorophenol
 Concen: 20.21 ng
 RT: 7.216 min Scan# 702
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	128	Resp:	29867
Ion	Ratio	Lower	Upper
128	100		
130	30.6	14.1	54.1
64	37.2	17.2	57.2









Benzaldehyde
 Concen: 23.36 ng
 RT: 6.799 min Scan# 631
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	77	Resp:	19599
Ion	Ratio	Lower	Upper
77	100		
105	101.6	89.4	129.4
106	96.5	79.9	119.9



Phenol
 Concen: 19.41 ng
 RT: 6.857 min Scan# 641
 Delta R.T. -0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	94	Resp:	36415
Ion	Ratio	Lower	Upper
94	100		
65	30.6	8.0	48.0
66	40.1	17.3	57.3









bis(2-Chloroethyl)ether
Concen: 19.57 ng
RT: 7.081 min Scan# 679
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	93	Resp:	26881
Ion	Ratio	Lower	Upper
93	100		
63	68.0	49.6	89.6
95	30.8	14.4	54.4




1,3-Dichlorobenzene
Concen: 20.69 ng
RT: 7.534 min Scan# 756
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	146	Resp:	39435
Ion	Ratio	Lower	Upper
146	100		
148	63.5	51.4	77.0
75	30.0	23.9	35.9






Abundance Scan 781 (7.681 min): BP005349.D\data.ms (-773) #13




1,4-Dichlorobenzene
Concen: 21.01 ng
RT: 7.681 min Scan# 781
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	146	Resp:	38427
Ion Ratio	Lower	Upper	
146	100		
148	63.3	52.2	78.4
111	40.3	34.2	51.4







Abundance Scan 834 (7.993 min): BP005349.D\data.ms (-822) #14



1,2-Dichlorobenzene
Concen: 20.82 ng
RT: 7.993 min Scan# 834
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	146	Resp:	35907
Ion Ratio	Lower	Upper	
146	100		
148	66.0	51.5	77.3
111	44.6	34.0	51.0









Benzyl Alcohol
 Concen: 17.66 ng
 RT: 7.899 min Scan# 818
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	79	Resp:	14415
Ion	Ratio	Lower	Upper
79	100		
108	82.8	64.0	96.0
77	96.5	72.7	109.1



2,2'-oxybis(1-Chloropropane)
 Concen: 16.83 ng
 RT: 8.175 min Scan# 865
 Delta R.T. -0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	45	Resp:	14056
Ion	Ratio	Lower	Upper
45	100		
77	28.4	7.0	47.0
79	20.7	1.0	41.0









2-Methylphenol
 Concen: 18.38 ng
 RT: 8.104 min Scan# 853
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	107	Resp:	16732
Ion	Ratio	Lower	Upper
107	100		
108	101.2	85.8	128.8
77	53.4	41.4	62.0
79	45.1	38.6	57.8



Hexachloroethane
 Concen: 20.32 ng
 RT: 8.710 min Scan# 956
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	117	Resp:	13995
Ion	Ratio	Lower	Upper
117	100		
119	90.8	80.1	120.1
201	130.6	98.8	148.2









n-Nitroso-di-n-propylamine
 Concen: 17.14 ng
 RT: 8.452 min Scan# 912
 Delta R.T. -0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion:	70	Resp:	14232
Ion	Ratio	Lower	Upper
70	100		
42	45.2	39.7	59.5
101	8.7	7.4	11.0
130	22.0	17.2	25.8



3+4-Methylphenols
 Concen: 17.28 ng
 RT: 8.434 min Scan# 909
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50


Tgt Ion:	107	Resp:	19048
Ion	Ratio	Lower	Upper
107	100		
108	91.6	69.5	109.5
77	40.8	21.5	61.5
79	33.2	16.8	56.8





Abundance Scan 1249 (10.434 min): BP005349.D\data.ms (-1198) (-)

Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1249
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion	Ratio	Lower	Upper
136	100		
137	10.3	8.9	13.3
54	4.5	5.0	7.4#
68	6.2	3.7	5.5#





Abundance Scan 1249 (10.434 min): BP005348.D\data.ms (-1198) (-)




Abundance Scan 916 (8.475 min): BP005349.D\data.ms (-902) (-)

Acetophenone
Concen: 23.51 ng
RT: 8.475 min Scan# 916
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion	Ratio	Lower	Upper
105	100		
71	1.0	0.9	1.3
51	22.7	18.2	27.2
120	23.1	17.0	25.6




Abundance Scan 916 (8.475 min): BP005348.D\data.ms (-865) (-)






Abundance Scan 973 (8.811 min): BP005349.D\data.ms (-964) #23

Nitrobenzene-d5
Concen: 45.79 ng
RT: 8.804 min Scan# 972
Delta R.T. -0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020







Tgt Ion:	82	Resp:	56881
Ion Ratio	Lower	Upper	
82	100		
128	38.5	30.5	45.7
54	43.7	33.9	50.9




Abundance Scan 980 (8.852 min): BP005349.D\data.ms (-967) #24

Nitrobenzene
Concen: 22.24 ng
RT: 8.852 min Scan# 980
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	77	Resp:	27084
Ion Ratio	Lower	Upper	
77	100		
123	42.7	31.7	47.5
65	15.7	11.5	17.3






Abundance Scan 1069 (9.375 min): BP005349.D\data.ms (-1025-)




Isophorone
Concen: 18.90 ng
RT: 9.375 min Scan# 1069
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	82	Resp:	24502
Ion	Ratio	Lower	Upper
82	100		
95	9.9	7.5	11.3
138	20.7	14.6	22.0





Abundance Scan 1100 (9.558 min): BP005349.D\data.ms (-1026-)




2-Nitrophenol
Concen: 22.56 ng
RT: 9.557 min Scan# 1100
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	139	Resp:	7462
Ion	Ratio	Lower	Upper
139	100		
109	51.6	40.7	61.1
65	49.5	47.8	71.6



Abundance Scan 1100 (9.557 min): BP005348.D\data.ms (-1049) (-)





Abundance Scan 1111 (9.622 min): BP005349.D\data.ms (-11037-)




2,4-Dimethylphenol
Concen: 21.40 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	122	Resp:	10373
Ion	Ratio	Lower	Upper
122	100		
107	116.5	93.8	140.8
121	61.5	45.6	68.4



Abundance Scan 1112 (9.628 min): BP005348.D\data.ms (-1060-)





Abundance Scan 1151 (9.858 min): BP005349.D\data.ms (-11428-)




bis(2-Chloroethoxy)methane
Concen: 18.70 ng
RT: 9.857 min Scan# 1151
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	93	Resp:	12369
Ion	Ratio	Lower	Upper
93	100		
95	38.8	24.9	37.3#
123	12.8	9.5	14.3




Abundance Scan 1151 (9.857 min): BP005348.D\data.ms (-1100-)






Abundance Scan 1191 (10.093 min): BP005349.D\data.ms (-1129) (-)

2,4-Dichlorophenol
Concen: 21.24 ng
RT: 10.093 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion	Ratio	Lower	Upper
162	100		
164	63.3	44.1	84.1
98	25.2	12.6	52.6







Abundance Scan 1226 (10.299 min): BP005349.D\data.ms (-1130) (-)

1,2,4-Trichlorobenzene
Concen: 22.52 ng
RT: 10.299 min Scan# 1226
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50



Tgt Ion	Ratio	Lower	Upper
180	100		
182	92.1	77.2	115.8
145	25.1	21.7	32.5









Scan 1258 (10.487 min): BP005349.D\data.ms (-128) (-)
 Naphthalene
 Concen: 20.08 ng
 RT: 10.481 min Scan# 1257
 Delta R.T. -0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC020

Tgt Ion:	128	Resp:	40506
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.0	12.0
127	14.2	11.1	16.7



Scan 1140 (9.793 min): BP005349.D\data.ms (-122) (-)
 Benzoic acid
 Concen: 21.49 ng
 RT: 9.763 min Scan# 1135
 Delta R.T. -0.030 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50


Tgt Ion:	122	Resp:	7103
Ion	Ratio	Lower	Upper
122	100		
105	124.7	105.1	145.1
77	97.6	69.6	109.6






Abundance Scan 1279 (10.610 min): BP005349.D\data.ms (-1288) (-)

4-Chloroaniline
Concen: 19.09 ng
RT: 10.610 min Scan# 1279
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	127	Resp:	12452
Ion	Ratio	Lower	Upper
127	100		
129	33.3	25.5	38.3
65	37.5	30.0	45.0
92	29.9	20.2	30.2







Abundance Scan 1305 (10.763 min): BP005349.D\data.ms (-1284) (-)

Hexachlorobutadiene
Concen: 24.98 ng
RT: 10.757 min Scan# 1304
Delta R.T. -0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50



Tgt Ion:	225	Resp:	17233
Ion	Ratio	Lower	Upper
225	100		
223	58.7	51.1	76.7
227	66.5	50.2	75.2









Scan 1414 (11.405 min): BP005349.D\data.ms (-1435) (-)
 Caprolactam
 Concen: 17.84 ng
 RT: 11.393 min Scan# 1412
 Delta R.T. -0.012 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion:	113	Resp:	2531
Ion	Ratio	Lower	Upper
113	100		
55	74.6	72.0	112.0
56	74.5	65.1	105.1



Scan 1473 (11.752 min): BP005349.D\data.ms (-1436) (-)
 4-Chloro-3-methylphenol
 Concen: 17.61 ng
 RT: 11.745 min Scan# 1472
 Delta R.T. -0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50


Tgt Ion:	107	Resp:	10181
Ion	Ratio	Lower	Upper
107	100		
144	26.4	20.1	30.1
142	74.3	62.2	93.4





Abundance Scan 1534 (12.110 min): BP005349.D\data.ms (-1483) (-)

2-Methylnaphthalene
Concen: 19.02 ng
RT: 12.110 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	142	Resp:	23520
Ion	Ratio	Lower	Upper
142	100		
141	93.1	74.7	112.1
115	53.4	43.8	65.8





Abundance Scan 1534 (12.110 min): BP005348.D\data.ms (-1483) (-)




Abundance Scan 1571 (12.328 min): BP005349.D\data.ms (-1520) (-)

1-Methylnaphthalene
Concen: 19.27 ng
RT: 12.328 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	142	Resp:	23146
Ion	Ratio	Lower	Upper
142	100		
141	94.8	77.0	115.6
116	5.5	5.0	7.6



Abundance Scan 1571 (12.328 min): BP005348.D\data.ms (-1520) (-)






Abundance Scan 1907 (14.304 min): BP005349.D\data.ms (-1856) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.310 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	164	Resp:	27829
Ion Ratio	Lower	Upper	
164	100		
162	100.5	73.5	110.3
160	58.8	40.6	61.0






Abundance Scan 1597 (12.481 min): BP005349.D\data.ms (-1546) (-)



1,2,4,5-Tetrachlorobenzene
Concen: 20.25 ng
RT: 12.481 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	216	Resp:	20842
Ion Ratio	Lower	Upper	
216	100		
214	80.1	60.6	91.0
179	19.7	15.3	22.9
108	17.3	13.8	20.6





Abundance Scan 1592 (12.452 min): BP005349.D\data.ms (-1541) (-)

Hexachlorocyclopentadiene
Concen: 20.60 ng
RT: 12.451 min Scan# 1592
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	237	Resp:	6572
Ion Ratio	Lower	Upper	
237	100		
235	60.7	40.9	80.9
272	17.0	0.0	34.6




Abundance Scan 1592 (12.451 min): BP005348.D\data.ms (-1541) (-)





Abundance Scan 2163 (15.810 min): BP005349.D\data.ms (-2112) (-)


2,4,6-Tribromophenol
Concen: 56.03 ng
RT: 15.810 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Tgt Ion:	330	Resp:	27416
Ion Ratio	Lower	Upper	
330	100		
332	97.1	76.3	114.5
141	26.8	22.8	34.2



Abundance Scan 2163 (15.810 min): BP005348.D\data.ms (-2112) (-)





Abundance Scan 1640 (12.734 min): BP005349.D\data.ms (-1640) (-)




2,4,6-Trichlorophenol
Concen: 19.22 ng
RT: 12.734 min Scan# 1640
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020


Tgt Ion:	196	Resp:	10483
Ion	Ratio	Lower	Upper
196	100		
198	98.3	81.4	122.2
200	31.9	28.2	42.2



Abundance Scan 1640 (12.734 min): BP005348.D\data.ms (-1589) (-)





Abundance Scan 1654 (12.816 min): BP005349.D\data.ms (-1644) (-)





2,4,5-Trichlorophenol
Concen: 20.15 ng
RT: 12.816 min Scan# 1654
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	196	Resp:	12153
Ion	Ratio	Lower	Upper
196	100		
198	83.8	71.0	106.4
97	43.6	38.5	57.7
132	14.5	15.6	23.4#



Abundance Scan 1654 (12.816 min): BP005348.D\data.ms (-1603) (-)









2-Fluorobiphenyl
Concen: 39.15 ng
RT: 12.922 min Scan# 1672
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	172	Resp:	82388
Ion	Ratio	Lower	Upper
172	100		
171	33.1	28.1	42.1
170	24.0	19.6	29.4



1,1'-Biphenyl
Concen: 18.44 ng
RT: 13.134 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	154	Resp:	36522
Ion	Ratio	Lower	Upper
154	100		
153	38.6	18.9	58.9
76	10.9	0.0	32.1





Abundance Scan 1715 (13.175 min): BP005349.D\data.ms (-1747) (-)

2-Chloronaphthalene
Concen: 19.39 ng
RT: 13.175 min Scan# 1715
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	162	Resp:	31273
Ion Ratio	Lower	Upper	
162	100		
127	36.3	31.4	47.0
164	30.5	26.6	39.8





Abundance Scan 1715 (13.175 min): BP005348.D\data.ms (-1664) (-)




Abundance Scan 1752 (13.392 min): BP005349.D\data.ms (-1748) (-)


2-Nitroaniline
Concen: 16.14 ng
RT: 13.392 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	65	Resp:	6554
Ion Ratio	Lower	Upper	
65	100		
92	69.7	48.8	73.2
138	107.0	69.5	104.3#



Abundance Scan 1752 (13.392 min): BP005348.D\data.ms (-1701) (-)









Scan 1860 (14.028 min): BP005349.D\data.ms (-1809) (-)
 Acenaphthylene
 Concen: 19.54 ng
 RT: 14.028 min Scan# 1860
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	152	Resp:	47726
Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.8	23.6
153	12.8	10.5	15.7



Scan 1817 (13.775 min): BP005349.D\data.ms (-1766) (-)
 Dimethylphthalate
 Concen: 18.68 ng
 RT: 13.775 min Scan# 1817
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	163	Resp:	33774
Ion	Ratio	Lower	Upper
163	100		
194	6.8	3.9	5.9#
164	8.7	8.1	12.1









Scan 1838 (13.899 min): BP005349.D\data.ms (-1850) (-)
 2,6-Dinitrotoluene
 Concen: 19.88 ng
 RT: 13.898 min Scan# 1838
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	165	Resp:	9126
Ion	Ratio	Lower	Upper
165	100		
63	69.0	67.4	101.0
89	76.5	73.8	110.6



Scan 1919 (14.375 min): BP005349.D\data.ms (-1952) (-)
 Acenaphthene
 Concen: 19.34 ng
 RT: 14.369 min Scan# 1918
 Delta R.T. -0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	154	Resp:	31481
Ion	Ratio	Lower	Upper
154	100		
153	110.9	90.9	136.3
152	66.4	48.5	72.7









Scan 1895 (14.234 min): BP005349.D\data.ms (-1859) (-)
 3-Nitroaniline
 Concen: 17.18 ng
 RT: 14.234 min Scan# 1895
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion:	138	Resp:	6467
Ion	Ratio	Lower	Upper
138	100		
108	15.7	10.2	15.4#
92	141.0	108.1	162.1



Scan 1932 (14.451 min): BP005349.D\data.ms (-1925) (-)
 2,4-Dinitrophenol
 Concen: 21.44 ng
 RT: 14.451 min Scan# 1932
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50


Tgt Ion:	184	Resp:	5503
Ion	Ratio	Lower	Upper
184	100		
63	70.1	51.2	76.8
154	58.1	55.8	83.8






Abundance Scan 1976 (14.710 min): BP005349.D\data.ms (-1955) (-)

Dibenzofuran
Concen: 21.74 ng
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	168	Resp:	63989
Ion	Ratio	Lower	Upper
168	100		
139	39.5	31.4	47.0
169	13.9	11.4	17.0







Abundance Scan 1951 (14.563 min): BP005349.D\data.ms (-1955) (-)

4-Nitrophenol
Concen: 18.46 ng
RT: 14.569 min Scan# 1952
Delta R.T. 0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50



Tgt Ion:	139	Resp:	5357
Ion	Ratio	Lower	Upper
139	100		
109	121.8	122.8	162.8#
65	120.8	110.0	150.0









2,4-Dinitrotoluene
Concen: 20.15 ng
RT: 14.698 min Scan# 1974
Delta R.T. 0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	165	Resp:	14119
Ion	Ratio	Lower	Upper
165	100		
63	73.4	56.7	85.1
89	98.9	85.8	128.6
182	4.4	3.0	4.6



Fluorene
Concen: 20.97 ng
RT: 15.363 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	166	Resp:	50482
Ion	Ratio	Lower	Upper
166	100		
165	103.1	78.8	118.2
167	14.5	10.9	16.3






Abundance Scan 2016 (14.946 min): BP005349.D\data.ms (-2059) (-)

2,3,4,6-Tetrachlorophenol
Concen: 24.23 ng
RT: 14.945 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion	Ratio	Lower	Upper
232	100		
131	31.2	26.5	39.7
130	2.7	2.1	3.1
166	23.4	19.4	29.0






Abundance Scan 2050 (15.146 min): BP005349.D\data.ms (-2060) (-)


Diethylphthalate
Concen: 19.30 ng
RT: 15.145 min Scan# 2050
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50



Tgt Ion	Ratio	Lower	Upper
149	100		
177	24.6	19.2	28.8
150	13.0	9.9	14.9






Abundance Scan 2087 (15.363 min): BP005349.D\data.ms (-2087) (-)




4-Chlorophenyl-phenylether
Concen: 23.28 ng
RT: 15.363 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleID :
SSTDIC020

Tgt Ion:	204	Resp:	31715
Ion	Ratio	Lower	Upper
204	100		
206	32.1	26.0	39.0
141	51.2	44.2	66.2






Abundance Scan 2094 (15.404 min): BP005349.D\data.ms (-2094) (-)



4-Nitroaniline
Concen: 18.17 ng
RT: 15.404 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	138	Resp:	7028
Ion	Ratio	Lower	Upper
138	100		
92	62.8	50.9	90.9
108	142.5	154.7	194.7#





Abundance Scan 2137 (15.657 min): BP005349.D\data.ms (-2137) (-)

Azobenzene
Concen: 20.84 ng
RT: 15.657 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	77	Resp:	55612
Ion	Ratio	Lower	Upper
77	100		
182	27.4	5.5	45.5
105	15.7	0.0	36.5
51	32.6	10.3	50.3





Abundance Scan 2137 (15.657 min): BP005348.D\data.ms (-2086) (-)




Abundance Scan 2377 (17.069 min): BP005349.D\data.ms (-2367) (-)

Phenanthrene-d10
Concen: 20.00 ng
RT: 17.069 min Scan# 2377
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	188	Resp:	76199
Ion	Ratio	Lower	Upper
188	100		
94	5.8	5.0	7.6
80	6.9	5.9	8.9




Abundance Scan 2377 (17.069 min): BP005348.D\data.ms (-2326) (-)






Abundance Scan 2104 (15.463 min): BP005349.D\data.ms (-2053) (-)

4,6-Dinitro-2-methylphenol
Concen: 21.20 ng
RT: 15.463 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion:	198	Resp:	10540
Ion Ratio	Lower	Upper	
198	100		
51	30.6	10.3	50.3
105	33.4	13.6	53.6






Abundance Scan 2124 (15.581 min): BP005349.D\data.ms (-2106) (-)

n-Nitrosodiphenylamine
Concen: 18.86 ng
RT: 15.581 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Tgt Ion:	169	Resp:	40192
Ion Ratio	Lower	Upper	
169	100		
168	68.0	55.2	82.8
167	44.9	36.6	54.8






Abundance Scan 2240 (16.263 min): BP005349.D\data.ms (-2287) (-)

4-Bromophenyl-phenylether
Concen: 22.26 ng
RT: 16.263 min Scan# 2240
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampled :
SSTDIC020




Tgt Ion:	248	Resp:	20831
Ion Ratio	Lower	Upper	
248	100		
250	98.9	77.5	116.3
141	54.4	50.1	75.1






Abundance Scan 2259 (16.375 min): BP005349.D\data.ms (-2288) (-)

Hexachlorobenzene
Concen: 23.48 ng
RT: 16.369 min Scan# 2258
Delta R.T. -0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Tgt Ion:	284	Resp:	27946
Ion Ratio	Lower	Upper	
284	100		
142	28.7	20.8	31.2
249	30.9	23.7	35.5





Abundance Scan 2288 (16.545 min): BP005349.D\data.ms (-2288) (-)

Atrazine
Concen: 20.90 ng
RT: 16.545 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion	Ratio	Lower	Upper
200	100		
173	21.1	0.7	40.7
215	56.4	29.0	69.0





Abundance Scan 2288 (16.545 min): BP005348.D\data.ms (-2237) (-)




Abundance Scan 2319 (16.728 min): BP005349.D\data.ms (-2319) (-)

Pentachlorophenol
Concen: 22.77 ng
RT: 16.728 min Scan# 2319
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion	Ratio	Lower	Upper
266	100		
268	65.3	49.4	74.0
264	66.0	53.7	80.5



Abundance Scan 2319 (16.728 min): BP005348.D\data.ms (-2268) (-)






Abundance Scan 2384 (17.110 min): BP005349.D\data.ms (-2371) (-)




Phenanthrene
Concen: 21.08 ng
RT: 17.110 min Scan# 2384
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	178	Resp:	96595
Ion Ratio	Lower	Upper	
178	100		
176	18.5	15.4	23.2
179	16.1	12.3	18.5







Abundance Scan 2400 (17.204 min): BP005349.D\data.ms (-2373) (-)



Anthracene
Concen: 21.03 ng
RT: 17.204 min Scan# 2400
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	178	Resp:	91033
Ion Ratio	Lower	Upper	
178	100		
176	19.7	15.2	22.8
179	16.0	12.6	19.0









Scan 2448 (17.487 min): BP005349.D\data.ms (-2459) (-)
 Carbazole
 Concen: 20.31 ng
 RT: 17.486 min Scan# 2448
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC020

Tgt Ion:	167	Resp:	79579
Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.4	26.0
139	13.0	10.4	15.6




Scan 2542 (18.039 min): BP005349.D\data.ms (-2554) (-)
 Di-n-butylphthalate
 Concen: 17.72 ng
 RT: 18.039 min Scan# 2542
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	149	Resp:	61857
Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.6	11.4
104	12.9	8.9	13.3






Abundance Scan 2721 (19.092 min): BP005349.D\data.ms (-27175) (-)




Fluoranthene
Concen: 21.31 ng
RT: 19.098 min Scan# 2722
Delta R.T. 0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	202	Resp:	118960
Ion Ratio	Lower	Upper	
202	100		
101	5.4	0.0	26.1
203	17.8	0.0	37.2






Abundance Scan 3074 (21.169 min): BP005349.D\data.ms (-3075) (-)



Chrysene-d12
Concen: 20.00 ng
RT: 21.174 min Scan# 3075
Delta R.T. 0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	240	Resp:	126062
Ion Ratio	Lower	Upper	
240	100		
120	4.0	4.2	6.2#
236	24.0	20.2	30.2





Abundance Scan 2754 (19.286 min): BP005349.D\data.ms (-2775) (-)

Benzidine
Concen: 17.74 ng
RT: 19.286 min Scan# 2754
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Instrument :
BNA_P
ClientSampleId :
SSTDIC020



Tgt Ion:	184	Resp:	29421
Ion	Ratio	Lower	Upper
184	100		
185	15.0	11.1	16.7
183	12.5	9.8	14.8




Abundance Scan 2754 (19.286 min): BP005348.D\data.ms (-2703) (-)





Abundance Scan 2782 (19.451 min): BP005349.D\data.ms (-2773) (-)


Pyrene
Concen: 18.90 ng
RT: 19.451 min Scan# 2782
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50




Tgt Ion:	202	Resp:	130366
Ion	Ratio	Lower	Upper
202	100		
200	20.3	16.9	25.3
203	16.9	13.8	20.6



Abundance Scan 2782 (19.451 min): BP005348.D\data.ms (-2731) (-)









Scan 2815 (19.645 min): BP005349.D\data.ms (-2809) (-)
 Terphenyl-d14
 Concen: 41.20 ng
 RT: 19.645 min Scan# 2815
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC020


Tgt Ion:	244	Resp:	267830
Ion	Ratio	Lower	Upper
244	100		
212	7.7	5.8	8.6
122	6.0	5.2	7.8



Scan 2931 (20.327 min): BP005349.D\data.ms (-2880) (-)
 Butylbenzylphthalate
 Concen: 17.12 ng
 RT: 20.327 min Scan# 2931
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	149	Resp:	27563
Ion	Ratio	Lower	Upper
149	100		
91	66.3	60.2	90.4
206	32.1	23.9	35.9









Scan 3072 (21.157 min): BP005349.D\data.ms (-3081) (-)
 Benzo(a)anthracene
 Concen: 20.22 ng
 RT: 21.157 min Scan# 3072
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	228	Resp:	160319
Ion	Ratio	Lower	Upper
228	100		
226	26.7	21.2	31.8
229	19.9	15.6	23.4



Scan 3061 (21.092 min): BP005349.D\data.ms (-3082) (-)
 3,3'-Dichlorobenzidine
 Concen: 19.60 ng
 RT: 21.092 min Scan# 3061
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	252	Resp:	38502
Ion	Ratio	Lower	Upper
252	100		
254	67.1	49.9	74.9
126	6.2	5.5	8.3









Scan 3081 (21.210 min): BP005349.D\data.ms (-3081)
 Chrysene
 Concen: 20.44 ng
 RT: 21.210 min Scan# 3081
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020

Tgt Ion:	228	Resp:	161327
Ion Ratio	Lower	Upper	
228	100		
226	29.1	24.8	37.2
229	19.4	15.4	23.0




Scan 3059 (21.081 min): BP005349.D\data.ms (-3059)
 Bis(2-ethylhexyl)phthalate
 Concen: 15.61 ng
 RT: 21.080 min Scan# 3059
 Delta R.T. -0.000 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	149	Resp:	39415
Ion Ratio	Lower	Upper	
149	100		
167	35.3	23.8	35.8
279	8.8	5.7	8.5#






Abundance Scan 3208 (21.957 min): BP005349.D\data.ms (-3208)




Di-n-octyl phthalate
Concen: 16.49 ng
RT: 21.957 min Scan# 3208
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	149	Resp:	77214
Ion Ratio	Lower	Upper	
149	100		
167	2.1	1.5	2.3
43	6.3	5.1	7.7

Instrument :
BNA_P
ClientSampleId :
SSTDIC020






Abundance Scan 3458 (23.427 min): BP005349.D\data.ms (-3458)



Perylene-d12
Concen: 20.00 ng
RT: 23.427 min Scan# 3458
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50


Tgt Ion:	264	Resp:	130577
Ion Ratio	Lower	Upper	
264	100		
260	25.4	20.6	30.8
265	23.2	19.0	28.6






Abundance Scan 3852 (25.745 min): BP005349.D\data.ms (-3852) (-)

Indeno(1,2,3-cd)pyrene
Concen: 21.55 ng
RT: 25.739 min Scan# 3851
Delta R.T. -0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020




Tgt Ion	Ratio	Lower	Upper
276	100		
138	10.5	8.4	12.6
277	25.7	20.6	31.0







Abundance Scan 3342 (22.745 min): BP005349.D\data.ms (-3342) (-)

Benzo(b)fluoranthene
Concen: 20.78 ng
RT: 22.745 min Scan# 3342
Delta R.T. -0.000 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50



Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	18.2	27.2
125	4.2	3.4	5.0









Scan 3350 (22.792 min): BP005349.D\data.ms (-3349) (-)
 Benzo(k)fluoranthene
 Concen: 21.67 ng
 RT: 22.745 min Scan# 3342
 Delta R.T. -0.047 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC020


Tgt Ion:	252	Resp:	185413
Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.2	25.8
125	4.2	3.8	5.6



Scan 3442 (23.333 min): BP005349.D\data.ms (-3440) (-)
 Benzo(a)pyrene
 Concen: 20.77 ng
 RT: 23.327 min Scan# 3441
 Delta R.T. -0.006 min
 Lab File: BP005348.D
 Acq: 20 Apr 2021 10:50

Tgt Ion:	252	Resp:	166368
Ion	Ratio	Lower	Upper
252	100		
253	20.9	17.1	25.7
125	4.4	3.8	5.6









Dibenzo(a,h)anthracene
Concen: 21.64 ng
RT: 25.745 min Scan# 3852
Delta R.T. -0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Instrument :
BNA_P
ClientSampleId :
SSTDIC020

Tgt Ion:	278	Resp:	193224
Ion	Ratio	Lower	Upper
278	100		
139	6.8	5.0	7.6
279	23.3	18.6	28.0



Benzo(g,h,i)perylene
Concen: 21.43 ng
RT: 26.439 min Scan# 3970
Delta R.T. -0.006 min
Lab File: BP005348.D
Acq: 20 Apr 2021 10:50

Tgt Ion:	276	Resp:	179641
Ion	Ratio	Lower	Upper
276	100		
277	24.0	19.8	29.8
138	8.7	7.7	11.5

