

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005350.D
 Acq On : 20 Apr 2021 11:58
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Quant Time: Apr 20 12:51:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 12:00:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.646	152	32727	20.00	ng	0.00
21) Naphthalene-d8	10.434	136	53698	20.00	ng	0.00
39) Acenaphthene-d10	14.310	164	42241	20.00	ng	0.00
64) Phenanthrene-d10	17.069	188	116066	20.00	ng	0.00
76) Chrysene-d12	21.175	240	176749	20.00	ng	0.00
86) Perylene-d12	23.433	264	173081	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	209575	95.98	ng	0.00
7) Phenol-d6	6.834	99	252856	97.03	ng	0.00
23) Nitrobenzene-d5	8.810	82	195501	100.42	ng	0.00
42) 2,4,6-Tribromophenol	15.810	330	99192	133.55	ng	0.00
45) 2-Fluorobiphenyl	12.922	172	318468	99.71	ng	0.00
79) Terphenyl-d14	19.651	244	952723	104.53	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.181	88	55043	37.92	ng	# 93
3) Pyridine	3.581	79	119662	40.80	ng	# 99
4) n-Nitrosodimethylamine	3.493	42	41309	38.36	ng	# 80
6) Aniline	6.987	93	111580	48.87	ng	99
8) 2-Chlorophenol	7.216	128	104243	50.99	ng	99
9) Benzaldehyde	6.799	77	47230	40.68	ng	99
10) Phenol	6.864	94	123677	47.64	ng	98
11) bis(2-Chloroethyl)ether	7.081	93	87988	46.29	ng	97
12) 1,3-Dichlorobenzene	7.534	146	131169	49.74	ng	98
13) 1,4-Dichlorobenzene	7.681	146	125055	49.40	ng	97
14) 1,2-Dichlorobenzene	7.993	146	118177	49.52	ng	100
15) Benzyl Alcohol	7.899	79	53404	47.29	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.175	45	54570	47.23	ng	98
17) 2-Methylphenol	8.105	107	61604	48.91	ng	98
18) Hexachloroethane	8.710	117	41154	43.18	ng	96
19) n-Nitroso-di-n-propyla...	8.458	70	53626	46.68	ng	96
20) 3+4-Methylphenols	8.434	107	72325	47.41	ng	98
22) Acetophenone	8.475	105	135415	53.70	ng	# 99
24) Nitrobenzene	8.852	77	95598	50.10	ng	99
25) Isophorone	9.375	82	102780	50.58	ng	99
26) 2-Nitrophenol	9.558	139	27413	52.88	ng	92
27) 2,4-Dimethylphenol	9.628	122	40317	53.09	ng	87
28) bis(2-Chloroethoxy)met...	9.857	93	52804	50.95	ng	94
29) 2,4-Dichlorophenol	10.093	162	49395	54.16	ng	96
30) 1,2,4-Trichlorobenzene	10.299	180	67268	53.51	ng	100
31) Naphthalene	10.487	128	162103	51.29	ng	98
32) Benzoic acid	9.805	122	26121	50.44	ng	90
33) 4-Chloroaniline	10.605	127	54260	53.08	ng	96
34) Hexachlorobutadiene	10.763	225	59920	55.42	ng	99
35) Caprolactam	11.410	113	9269	41.69	ng	97
36) 4-Chloro-3-methylphenol	11.752	107	44807	49.44	ng	97
37) 2-Methylnaphthalene	12.110	142	103757	53.54	ng	94
38) 1-Methylnaphthalene	12.328	142	99057	52.64	ng	98
40) 1,2,4,5-Tetrachloroben...	12.481	216	80737	51.67	ng	97
41) Hexachlorocyclopentadiene	12.451	237	30979	63.97	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005350.D
 Acq On : 20 Apr 2021 11:58
 Operator : CG/JU
 Sample : SSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Quant Time: Apr 20 12:51:44 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 12:00:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.734	196	41885	50.60	ng	98
44) 2,4,5-Trichlorophenol	12.810	196	46617	50.92	ng	91
46) 1,1'-Biphenyl	13.134	154	144908	48.21	ng	97
47) 2-Chloronaphthalene	13.175	162	121026	49.44	ng	100
48) 2-Nitroaniline	13.399	65	31464	51.04	ng	96
49) Acenaphthylene	14.028	152	184036	49.65	ng	99
50) Dimethylphthalate	13.775	163	139396	50.80	ng	100
51) 2,6-Dinitrotoluene	13.898	165	34997	50.22	ng	95
52) Acenaphthene	14.375	154	123101	49.82	ng	96
53) 3-Nitroaniline	14.234	138	28392	49.69	ng	93
54) 2,4-Dinitrophenol	14.445	184	21748	55.82	ng	93
55) Dibenzofuran	14.710	168	227603	50.95	ng	98
56) 4-Nitrophenol	14.563	139	22858	51.90	ng	94
57) 2,4-Dinitrotoluene	14.693	165	55602	52.28	ng	96
58) Fluorene	15.363	166	194969	53.37	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.945	232	56883	57.67	ng	99
60) Diethylphthalate	15.145	149	148595	52.65	ng	100
61) 4-Chlorophenyl-phenyle...	15.363	204	115420	55.81	ng	98
62) 4-Nitroaniline	15.404	138	31315	53.34	ng	90
63) Azobenzene	15.657	77	207779	51.29	ng	99
65) 4,6-Dinitro-2-methylph...	15.463	198	37883	50.02	ng	92
66) n-Nitrosodiphenylamine	15.581	169	152727	47.06	ng	99
67) 4-Bromophenyl-phenylether	16.257	248	80152	56.24	ng	99
68) Hexachlorobenzene	16.375	284	98928	54.56	ng	99
69) Atrazine	16.545	200	59356	51.01	ng	97
70) Pentachlorophenol	16.728	266	50528	57.42	ng	92
71) Phenanthrene	17.116	178	338834	48.55	ng	99
72) Anthracene	17.204	178	327413	49.66	ng	98
73) Carbazole	17.487	167	294208	49.29	ng	99
74) Di-n-butylphthalate	18.039	149	255652	48.07	ng	98
75) Fluoranthene	19.092	202	448183	52.72	ng	99
77) Benzidine	19.286	184	93094	40.03	ng	98
78) Pyrene	19.451	202	477211	49.35	ng	100
80) Butylbenzylphthalate	20.328	149	98537	43.64	ng	98
81) Benzo(a)anthracene	21.157	228	546157	49.12	ng	99
82) 3,3'-Dichlorobenzidine	21.092	252	134866	48.96	ng	97
83) Chrysene	21.210	228	528693	47.78	ng	98
84) Bis(2-ethylhexyl)phtha...	21.086	149	155301	43.86	ng	# 98
85) Di-n-octyl phthalate	21.963	149	279781	42.62	ng	99
87) Indeno(1,2,3-cd)pyrene	25.751	276	677385	50.00	ng	100
88) Benzo(b)fluoranthene	22.751	252	583811	49.36	ng	98
89) Benzo(k)fluoranthene	22.792	252	577083	50.89	ng	99
90) Benzo(a)pyrene	23.333	252	528598	49.78	ng	99
91) Dibenzo(a,h)anthracene	25.757	278	593754	50.17	ng	98
92) Benzo(g,h,i)perylene	26.451	276	543843	48.94	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 20 12:51:44 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 20 12:00:16 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.646 min Scan# 775
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC050


Tgt Ion:	152	Resp:	32727
Ion Ratio	Lower	Upper	
152	100		
150	157.0	119.8	179.8
115	61.1	49.0	73.4



1,4-Dioxane
 Concen: 37.92 ng
 RT: 3.181 min Scan# 16
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	88	Resp:	55043
Ion Ratio	Lower	Upper	
88	100		
58	56.6	42.5	63.7
43	30.0	19.8	29.8#









Pyridine
Concen: 40.80 ng
RT: 3.581 min Scan# 84
Delta R.T. -0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	79	Resp:	119662
Ion	Ratio	Lower	Upper
79	100		
52	51.0	40.3	60.5
51	25.0	20.0	30.0



n-Nitrosodimethylamine
Concen: 38.36 ng
RT: 3.493 min Scan# 69
Delta R.T. -0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	42	Resp:	41309
Ion	Ratio	Lower	Upper
42	100		
74	161.4	122.6	183.8
44	139.4	146.9	220.3#









2-Fluorophenol
 Concen: 95.98 ng
 RT: 5.246 min Scan# 367
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	112	Resp:	209575
Ion	Ratio	Lower	Upper
112	100		
64	41.7	33.6	50.4
63	22.6	18.6	27.8



Aniline
 Concen: 48.87 ng
 RT: 6.987 min Scan# 663
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	93	Resp:	111580
Ion	Ratio	Lower	Upper
93	100		
66	34.5	27.0	40.4
65	18.6	14.4	21.6









Phenol-d6
Concen: 97.03 ng
RT: 6.834 min Scan# 637
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion	Ratio	Lower	Upper
99	100		
42	15.6	12.9	19.3
71	33.7	26.7	40.1



2-Chlorophenol
Concen: 50.99 ng
RT: 7.216 min Scan# 702
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion	Ratio	Lower	Upper
128	100		
130	33.0	14.1	54.1
64	36.6	17.2	57.2









Benzaldehyde
 Concen: 40.68 ng
 RT: 6.799 min Scan# 631
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	77	Resp:	47230
Ion	Ratio	Lower	Upper
77	100		
105	111.2	89.4	129.4
106	99.2	79.9	119.9



Phenol
 Concen: 47.64 ng
 RT: 6.864 min Scan# 642
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	94	Resp:	123677
Ion	Ratio	Lower	Upper
94	100		
65	26.9	8.0	48.0
66	36.6	17.3	57.3









bis(2-Chloroethyl)ether
 Concen: 46.29 ng
 RT: 7.081 min Scan# 679
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC050

Tgt Ion:	93	Resp:	87988
Ion	Ratio	Lower	Upper
93	100		
63	68.0	49.6	89.6
95	32.3	14.4	54.4




1,3-Dichlorobenzene
 Concen: 49.74 ng
 RT: 7.534 min Scan# 756
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	146	Resp:	131169
Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.4	77.0
75	27.8	23.9	35.9






Abundance Scan 781 (7.681 min): BP005349.D\data.ms (-773) #13




1,4-Dichlorobenzene
Concen: 49.40 ng
RT: 7.681 min Scan# 781
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	146	Resp:	125055
Ion	Ratio	Lower	Upper
146	100		
148	64.0	52.2	78.4
111	40.1	34.2	51.4







Abundance Scan 834 (7.993 min): BP005349.D\data.ms (-822) #14



1,2-Dichlorobenzene
Concen: 49.52 ng
RT: 7.993 min Scan# 834
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	146	Resp:	118177
Ion	Ratio	Lower	Upper
146	100		
148	64.6	51.5	77.3
111	42.6	34.0	51.0









Benzyl Alcohol
Concen: 47.29 ng
RT: 7.899 min Scan# 818
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	79	Resp:	53404
Ion	Ratio	Lower	Upper
79	100		
108	78.0	64.0	96.0
77	90.8	72.7	109.1



2,2'-oxybis(1-Chloropropane)
Concen: 47.23 ng
RT: 8.175 min Scan# 865
Delta R.T. -0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	45	Resp:	54570
Ion	Ratio	Lower	Upper
45	100		
77	26.8	7.0	47.0
79	19.6	1.0	41.0









2-Methylphenol
 Concen: 48.91 ng
 RT: 8.105 min Scan# 853
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	107	Resp:	61604
Ion	Ratio	Lower	Upper
107	100		
108	105.0	85.8	128.8
77	53.0	41.4	62.0
79	47.3	38.6	57.8



Hexachloroethane
 Concen: 43.18 ng
 RT: 8.710 min Scan# 956
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58


Tgt Ion:	117	Resp:	41154
Ion	Ratio	Lower	Upper
117	100		
119	95.7	80.1	120.1
201	128.4	98.8	148.2






Abundance Scan 913 (8.458 min): BP005349.D\data.ms (-905) #19

n-Nitroso-di-n-propylamine
Concen: 46.68 ng
RT: 8.458 min Scan# 913
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion:	70	Resp:	53626
Ion	Ratio	Lower	Upper
70	100		
42	45.3	39.7	59.5
101	9.4	7.4	11.0
130	21.4	17.2	25.8







Abundance Scan 909 (8.434 min): BP005349.D\data.ms (-897) #20

3+4-Methylphenols
Concen: 47.41 ng
RT: 8.434 min Scan# 909
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58



Tgt Ion:	107	Resp:	72325
Ion	Ratio	Lower	Upper
107	100		
108	89.0	69.5	109.5
77	38.7	21.5	61.5
79	34.5	16.8	56.8









Naphthalene-d8
 Concen: 20.00 ng
 RT: 10.434 min Scan# 1249
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	136	Resp:	53698
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	5.5	5.0	7.4
68	5.1	3.7	5.5



Acetophenone
 Concen: 53.70 ng
 RT: 8.475 min Scan# 916
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58


Tgt Ion:	105	Resp:	135415
Ion	Ratio	Lower	Upper
105	100		
71	1.7	0.9	1.3#
51	22.4	18.2	27.2
120	21.9	17.0	25.6





Abundance Scan 973 (8.811 min): BP005349.D\data.ms (-964) #23

Nitrobenzene-d5
Concen: 100.42 ng
RT: 8.810 min Scan# 973
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58


Instrument :
BNA_P
ClientSampleId :
SSTDICC050



Tgt Ion:	82	Resp:	195501
Ion	Ratio	Lower	Upper
82	100		
128	38.6	30.5	45.7
54	42.6	33.9	50.9






Abundance Scan 973 (8.811 min): BP005350.D\data.ms (-922) (-)




Abundance Scan 980 (8.852 min): BP005349.D\data.ms (-967) #24

Nitrobenzene
Concen: 50.10 ng
RT: 8.852 min Scan# 980
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58


Tgt Ion:	77	Resp:	95598
Ion	Ratio	Lower	Upper
77	100		
123	39.9	31.7	47.5
65	13.7	11.5	17.3



Abundance Scan 980 (8.852 min): BP005350.D\data.ms (-929) (-)






Abundance Scan 1069 (9.375 min): BP005349.D\data.ms (-1025-)




Isophorone
Concen: 50.58 ng
RT: 9.375 min Scan# 1069
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	82	Resp:	102780
Ion	Ratio	Lower	Upper
82	100		
95	10.2	7.5	11.3
138	17.8	14.6	22.0






Abundance Scan 1100 (9.558 min): BP005349.D\data.ms (-1026-)




2-Nitrophenol
Concen: 52.88 ng
RT: 9.558 min Scan# 1100
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	139	Resp:	27413
Ion	Ratio	Lower	Upper
139	100		
109	47.7	40.7	61.1
65	52.1	47.8	71.6



Abundance Scan 1111 (9.622 min): BP005349.D\data.ms (-11037-)





2,4-Dimethylphenol
Concen: 53.09 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58


Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:	122	Resp:	40317
Ion	Ratio	Lower	Upper
122	100		
107	137.1	93.8	140.8
121	58.0	45.6	68.4


Abundance Scan 1112 (9.628 min): BP005350.D\data.ms



Abundance Scan 1112 (9.628 min): BP005350.D\data.ms (-1060) (-)





Abundance Scan 1151 (9.858 min): BP005349.D\data.ms (-11428-)




bis(2-Chloroethoxy)methane
Concen: 50.95 ng
RT: 9.857 min Scan# 1151
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	93	Resp:	52804
Ion	Ratio	Lower	Upper
93	100		
95	34.9	24.9	37.3
123	10.7	9.5	14.3

Abundance Scan 1151 (9.857 min): BP005350.D\data.ms




Abundance Scan 1151 (9.857 min): BP005350.D\data.ms (-1100) (-)





Abundance Scan 1191 (10.093 min): BP005349.D\data.ms (-1129) (-)

2,4-Dichlorophenol
Concen: 54.16 ng
RT: 10.093 min Scan# 1191
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58


Instrument :
BNA_P
ClientSampleId :
SSTDIC050



Tgt Ion:	162	Resp:	49395
Ion	Ratio	Lower	Upper
162	100		
164	66.1	44.1	84.1
98	36.4	12.6	52.6




Abundance Scan 1191 (10.093 min): BP005350.D\data.ms (-1140) (-)





Abundance Scan 1226 (10.299 min): BP005349.D\data.ms (-1130) (-)


1,2,4-Trichlorobenzene
Concen: 53.51 ng
RT: 10.299 min Scan# 1226
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Tgt Ion:	180	Resp:	67268
Ion	Ratio	Lower	Upper
180	100		
182	96.5	77.2	115.8
145	27.0	21.7	32.5



Abundance Scan 1226 (10.299 min): BP005350.D\data.ms (-1175) (-)









Naphthalene
Concen: 51.29 ng
RT: 10.487 min Scan# 1258
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	128	Resp:	162103
Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.0	12.0
127	13.5	11.1	16.7



Benzoic acid
Concen: 50.44 ng
RT: 9.805 min Scan# 1142
Delta R.T. 0.012 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58


Tgt Ion:	122	Resp:	26121
Ion	Ratio	Lower	Upper
122	100		
105	129.2	105.1	145.1
77	106.7	69.6	109.6






Abundance Scan 1279 (10.610 min): BP005349.D\data.ms (-1283) (-)

4-Chloroaniline
Concen: 53.08 ng
RT: 10.605 min Scan# 1278
Delta R.T. -0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion:	127	Resp:	54260
Ion	Ratio	Lower	Upper
127	100		
129	29.7	25.5	38.3
65	40.0	30.0	45.0
92	22.9	20.2	30.2







Abundance Scan 1305 (10.763 min): BP005349.D\data.ms (-1284) (-)

Hexachlorobutadiene
Concen: 55.42 ng
RT: 10.763 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58



Tgt Ion:	225	Resp:	59920
Ion	Ratio	Lower	Upper
225	100		
223	63.2	51.1	76.7
227	63.1	50.2	75.2









Scan 1414 (11.405 min): BP005349.D\data.ms (-1435) (-)
 Caprolactam
 Concen: 41.69 ng
 RT: 11.410 min Scan# 1415
 Delta R.T. 0.006 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	113	Resp:	9269
Ion	Ratio	Lower	Upper
113	100		
55	92.4	72.0	112.0
56	89.4	65.1	105.1



Scan 1473 (11.752 min): BP005349.D\data.ms (-1436) (-)
 4-Chloro-3-methylphenol
 Concen: 49.44 ng
 RT: 11.752 min Scan# 1473
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58


Tgt Ion:	107	Resp:	44807
Ion	Ratio	Lower	Upper
107	100		
144	22.7	20.1	30.1
142	75.7	62.2	93.4






Abundance Scan 1534 (12.110 min): BP005349.D\data.ms (-1483) (-)

2-Methylnaphthalene
Concen: 53.54 ng
RT: 12.110 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion:	142	Resp:	103757
Ion	Ratio	Lower	Upper
142	100		
141	88.2	74.7	112.1
115	49.5	43.8	65.8






Abundance Scan 1571 (12.328 min): BP005349.D\data.ms (-1520) (-)


1-Methylnaphthalene
Concen: 52.64 ng
RT: 12.328 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58



Tgt Ion:	142	Resp:	99057
Ion	Ratio	Lower	Upper
142	100		
141	93.9	77.0	115.6
116	5.7	5.0	7.6






Abundance Scan 1907 (14.304 min): BP005349.D\data.ms (-1856) (-)




Acenaphthene-d10
Concen: 20.00 ng
RT: 14.310 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	164	Resp:	42241
Ion Ratio	Lower	Upper	
164	100		
162	98.8	73.5	110.3
160	49.6	40.6	61.0







Abundance Scan 1597 (12.481 min): BP005349.D\data.ms (-1546) (-)



1,2,4,5-Tetrachlorobenzene
Concen: 51.67 ng
RT: 12.481 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	216	Resp:	80737
Ion Ratio	Lower	Upper	
216	100		
214	78.4	60.6	91.0
179	18.9	15.3	22.9
108	18.3	13.8	20.6









Hexachlorocyclopentadiene
Concen: 63.97 ng
RT: 12.451 min Scan# 1592
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	237	Resp:	30979
Ion	Ratio	Lower	Upper
237	100		
235	63.2	40.9	80.9
272	14.9	0.0	34.6



2,4,6-Tribromophenol
Concen: 133.55 ng
RT: 15.810 min Scan# 2163
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	330	Resp:	99192
Ion	Ratio	Lower	Upper
330	100		
332	96.9	76.3	114.5
141	27.7	22.8	34.2









Scan 1640 (12.734 min): BP005349.D\data.ms (-1638) (-)
 2,4,6-Trichlorophenol
 Concen: 50.60 ng
 RT: 12.734 min Scan# 1640
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	196	Resp:	41885
Ion	Ratio	Lower	Upper
196	100		
198	102.6	81.4	122.2
200	32.3	28.2	42.2




Scan 1654 (12.816 min): BP005349.D\data.ms (-1647) (-)
 2,4,5-Trichlorophenol
 Concen: 50.92 ng
 RT: 12.810 min Scan# 1653
 Delta R.T. -0.006 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	196	Resp:	46617
Ion	Ratio	Lower	Upper
196	100		
198	98.0	71.0	106.4
97	54.5	38.5	57.7
132	21.2	15.6	23.4






Abundance Scan 1672 (12.922 min): BP005349.D\data.ms (-1655) (-)




2-Fluorobiphenyl
Concen: 99.71 ng
RT: 12.922 min Scan# 1672
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	172	Resp:	318468
Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.1	42.1
170	23.5	19.6	29.4







Abundance Scan 1708 (13.134 min): BP005349.D\data.ms (-1655) (-)



1,1'-Biphenyl
Concen: 48.21 ng
RT: 13.134 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	154	Resp:	144908
Ion	Ratio	Lower	Upper
154	100		
153	41.2	18.9	58.9
76	13.0	0.0	32.1









Scan 1715 (13.175 min): BP005349.D\data.ms (-1747) (-)
 2-Chloronaphthalene
 Concen: 49.44 ng
 RT: 13.175 min Scan# 1715
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	162	Resp:	121026
Ion	Ratio	Lower	Upper
162	100		
127	39.3	31.4	47.0
164	33.5	26.6	39.8



Scan 1752 (13.393 min): BP005349.D\data.ms (-1748) (-)
 2-Nitroaniline
 Concen: 51.04 ng
 RT: 13.399 min Scan# 1753
 Delta R.T. 0.006 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	65	Resp:	31464
Ion	Ratio	Lower	Upper
65	100		
92	63.5	48.8	73.2
138	83.5	69.5	104.3









Acenaphthylene
 Concen: 49.65 ng
 RT: 14.028 min Scan# 1860
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	152	Resp:	184036
Ion	Ratio	Lower	Upper
152	100		
151	19.2	15.8	23.6
153	13.5	10.5	15.7



Dimethylphthalate
 Concen: 50.80 ng
 RT: 13.775 min Scan# 1817
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	163	Resp:	139396
Ion	Ratio	Lower	Upper
163	100		
194	5.0	3.9	5.9
164	10.2	8.1	12.1









2,6-Dinitrotoluene
Concen: 50.22 ng
RT: 13.898 min Scan# 1838
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	165	Resp:	34997
Ion	Ratio	Lower	Upper
165	100		
63	88.0	67.4	101.0
89	87.4	73.8	110.6



Acenaphthene
Concen: 49.82 ng
RT: 14.375 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	154	Resp:	123101
Ion	Ratio	Lower	Upper
154	100		
153	109.6	90.9	136.3
152	56.3	48.5	72.7









Scan 1895 (14.234 min): BP005349.D\data.ms (-1853) (-)
 3-Nitroaniline
 Concen: 49.69 ng
 RT: 14.234 min Scan# 1895
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC050


Tgt Ion:	138	Resp:	28392
Ion	Ratio	Lower	Upper
138	100		
108	14.6	10.2	15.4
92	143.8	108.1	162.1



Scan 1932 (14.451 min): BP005349.D\data.ms (-1925) (-)
 2,4-Dinitrophenol
 Concen: 55.82 ng
 RT: 14.445 min Scan# 1931
 Delta R.T. -0.006 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	184	Resp:	21748
Ion	Ratio	Lower	Upper
184	100		
63	72.8	51.2	76.8
154	67.2	55.8	83.8









Dibenzofuran
Concen: 50.95 ng
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	168	Resp:	227603
Ion	Ratio	Lower	Upper
168	100		
139	40.8	31.4	47.0
169	13.5	11.4	17.0



4-Nitrophenol
Concen: 51.90 ng
RT: 14.563 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	139	Resp:	22858
Ion	Ratio	Lower	Upper
139	100		
109	138.0	122.8	162.8
65	120.6	110.0	150.0









2,4-Dinitrotoluene
Concen: 52.28 ng
RT: 14.693 min Scan# 1973
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	165	Resp:	55602
Ion	Ratio	Lower	Upper
165	100		
63	68.0	56.7	85.1
89	102.6	85.8	128.6
182	3.3	3.0	4.6



Fluorene
Concen: 53.37 ng
RT: 15.363 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58


Tgt Ion:	166	Resp:	194969
Ion	Ratio	Lower	Upper
166	100		
165	99.7	78.8	118.2
167	13.6	10.9	16.3





Abundance Scan 2016 (14.946 min): BP005349.D\data.ms (-2059) (-)

2,3,4,6-Tetrachlorophenol
Concen: 57.67 ng
RT: 14.945 min Scan# 2016
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58


Instrument :
BNA_P
ClientSampleId :
SSTDIC050



Tgt Ion	Ratio	Lower	Upper
232	100		
131	33.9	26.5	39.7
130	2.3	2.1	3.1
166	24.7	19.4	29.0





Abundance Scan 2016 (14.945 min): BP005350.D\data.ms (-1965) (-)




Abundance Scan 2050 (15.146 min): BP005349.D\data.ms (-2060) (-)

Diethylphthalate
Concen: 52.65 ng
RT: 15.145 min Scan# 2050
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion	Ratio	Lower	Upper
149	100		
177	24.2	19.2	28.8
150	12.6	9.9	14.9




Abundance Scan 2050 (15.145 min): BP005350.D\data.ms (-1999) (-)






Abundance Scan 2087 (15.363 min): BP005349.D\data.ms (-2087) (-)

4-Chlorophenyl-phenylether
Concen: 55.81 ng
RT: 15.363 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050




Tgt Ion	Ratio	Lower	Upper
204	100		
206	34.5	26.0	39.0
141	55.5	44.2	66.2






Abundance Scan 2094 (15.404 min): BP005349.D\data.ms (-2094) (-)


4-Nitroaniline
Concen: 53.34 ng
RT: 15.404 min Scan# 2094
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58



Tgt Ion	Ratio	Lower	Upper
138	100		
92	65.2	50.9	90.9
108	158.4	154.7	194.7






Abundance Scan 2137 (15.657 min): BP005349.D\data.ms (-2137) (-)




Azobenzene
Concen: 51.29 ng
RT: 15.657 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	77	Resp:	207779
Ion	Ratio	Lower	Upper
77	100		
182	25.8	5.5	45.5
105	16.4	0.0	36.5
51	31.0	10.3	50.3






Abundance Scan 2377 (17.069 min): BP005349.D\data.ms (-2367) (-)




Phenanthrene-d10
Concen: 20.00 ng
RT: 17.069 min Scan# 2377
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	188	Resp:	116066
Ion	Ratio	Lower	Upper
188	100		
94	6.3	5.0	7.6
80	7.2	5.9	8.9






Abundance Scan 2104 (15.463 min): BP005349.D\data.ms (-2053) (-)




4,6-Dinitro-2-methylphenol
Concen: 50.02 ng
RT: 15.463 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	198	Resp:	37883
Ion	Ratio	Lower	Upper
198	100		
51	36.1	10.3	50.3
105	36.4	13.6	53.6






Abundance Scan 2124 (15.581 min): BP005349.D\data.ms (-2166) (-)




n-Nitrosodiphenylamine
Concen: 47.06 ng
RT: 15.581 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	169	Resp:	152727
Ion	Ratio	Lower	Upper
169	100		
168	68.5	55.2	82.8
167	44.3	36.6	54.8






Abundance Scan 2240 (16.263 min): BP005349.D\data.ms (-2240) (-)




4-Bromophenyl-phenylether
Concen: 56.24 ng
RT: 16.257 min Scan# 2239
Delta R.T. -0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:	248	Resp:	80152
Ion Ratio	Lower	Upper	
248	100		
250	96.6	77.5	116.3
141	64.0	50.1	75.1






Abundance Scan 2259 (16.375 min): BP005349.D\data.ms (-2259) (-)




Hexachlorobenzene
Concen: 54.56 ng
RT: 16.375 min Scan# 2259
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	284	Resp:	98928
Ion Ratio	Lower	Upper	
284	100		
142	26.2	20.8	31.2
249	30.9	23.7	35.5






Abundance Scan 2288 (16.545 min): BP005349.D\data.ms (-2288) (-)




Atrazine
Concen: 51.01 ng
RT: 16.545 min Scan# 2288
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	200	Resp:	59356
Ion Ratio	Lower	Upper	
200	100		
173	23.0	0.7	40.7
215	50.8	29.0	69.0






Abundance Scan 2319 (16.728 min): BP005349.D\data.ms (-2319) (-)




Pentachlorophenol
Concen: 57.42 ng
RT: 16.728 min Scan# 2319
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	266	Resp:	50528
Ion Ratio	Lower	Upper	
266	100		
268	59.3	49.4	74.0
264	57.3	53.7	80.5






Abundance Scan 2384 (17.110 min): BP005349.D\data.ms (-2371) (-)




Phenanthrene
Concen: 48.55 ng
RT: 17.116 min Scan# 2385
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	178	Resp:	338834
Ion	Ratio	Lower	Upper
178	100		
176	18.3	15.4	23.2
179	15.4	12.3	18.5







Abundance Scan 2400 (17.204 min): BP005349.D\data.ms (-2373) (-)



Anthracene
Concen: 49.66 ng
RT: 17.204 min Scan# 2400
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	178	Resp:	327413
Ion	Ratio	Lower	Upper
178	100		
176	17.9	15.2	22.8
179	14.9	12.6	19.0









Carbazole
Concen: 49.29 ng
RT: 17.487 min Scan# 2448
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58




Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:	167	Resp:	294208
Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.4	26.0
139	12.6	10.4	15.6




Di-n-butylphthalate
Concen: 48.07 ng
RT: 18.039 min Scan# 2542
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	149	Resp:	255652
Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.6	11.4
104	10.2	8.9	13.3






Abundance Scan 2721 (19.092 min): BP005349.D\data.ms (-2717) (-)




Fluoranthene
Concen: 52.72 ng
RT: 19.092 min Scan# 2721
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	202	Resp:	448183
Ion Ratio	Lower	Upper	
202	100		
101	6.0	0.0	26.1
203	17.8	0.0	37.2






Abundance Scan 3074 (21.169 min): BP005349.D\data.ms (-3075) (-)




Chrysene-d12
Concen: 20.00 ng
RT: 21.175 min Scan# 3075
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	240	Resp:	176749
Ion Ratio	Lower	Upper	
240	100		
120	4.4	4.2	6.2
236	25.7	20.2	30.2






Abundance Scan 2754 (19.286 min): BP005349.D\data.ms (-2775) (-)




Benzidine
Concen: 40.03 ng
RT: 19.286 min Scan# 2754
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDICC050

Tgt Ion:	184	Resp:	93094
Ion Ratio	Lower	Upper	
184	100		
185	14.9	11.1	16.7
183	13.0	9.8	14.8






Abundance Scan 2782 (19.451 min): BP005349.D\data.ms (-2773) (-)




Pyrene
Concen: 49.35 ng
RT: 19.451 min Scan# 2782
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	202	Resp:	477211
Ion Ratio	Lower	Upper	
202	100		
200	21.1	16.9	25.3
203	17.6	13.8	20.6





Abundance Scan 2815 (19.645 min): BP005349.D\data.ms (-2809) (-)




Terphenyl-d14
Concen: 104.53 ng
RT: 19.651 min Scan# 2816
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050


Tgt Ion:	244	Resp:	952723
Ion	Ratio	Lower	Upper
244	100		
212	7.3	5.8	8.6
122	5.4	5.2	7.8



Abundance Scan 2816 (19.651 min): BP005350.D\data.ms (-2764) (-)





Abundance Scan 2931 (20.328 min): BP005349.D\data.ms (-2880) (-)





Butylbenzylphthalate
Concen: 43.64 ng
RT: 20.328 min Scan# 2931
Delta R.T. -0.000 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	149	Resp:	98537
Ion	Ratio	Lower	Upper
149	100		
91	73.6	60.2	90.4
206	29.6	23.9	35.9



Abundance Scan 2931 (20.328 min): BP005350.D\data.ms (-2880) (-)









Scan 3072 (21.157 min): BP005349.D\data.ms (-3065) (-)
 Benzo(a)anthracene
 Concen: 49.12 ng
 RT: 21.157 min Scan# 3072
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	228	Resp:	546157
Ion	Ratio	Lower	Upper
228	100		
226	27.0	21.2	31.8
229	19.8	15.6	23.4



Scan 3061 (21.092 min): BP005349.D\data.ms (-3062) (-)
 3,3'-Dichlorobenzidine
 Concen: 48.96 ng
 RT: 21.092 min Scan# 3061
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	252	Resp:	134866
Ion	Ratio	Lower	Upper
252	100		
254	65.1	49.9	74.9
126	7.3	5.5	8.3









Chrysene
 Concen: 47.78 ng
 RT: 21.210 min Scan# 3081
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050

Tgt Ion:	228	Resp:	528693
Ion	Ratio	Lower	Upper
228	100		
226	29.5	24.8	37.2
229	20.0	15.4	23.0




Bis(2-ethylhexyl)phthalate
 Concen: 43.86 ng
 RT: 21.086 min Scan# 3060
 Delta R.T. 0.006 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	149	Resp:	155301
Ion	Ratio	Lower	Upper
149	100		
167	30.8	23.8	35.8
279	8.6	5.7	8.5#






Abundance Scan 3208 (21.957 min): BP005349.D\data.ms (-3208)




Di-n-octyl phthalate
Concen: 42.62 ng
RT: 21.963 min Scan# 3209
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	149	Resp:	279781
Ion Ratio	Lower	Upper	
149	100		
167	1.8	1.5	2.3
43	6.8	5.1	7.7

Instrument :
BNA_P
ClientSampleId :
SSTDIC050






Abundance Scan 3458 (23.427 min): BP005349.D\data.ms (-3458)




Perylene-d12
Concen: 20.00 ng
RT: 23.433 min Scan# 3459
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	264	Resp:	173081
Ion Ratio	Lower	Upper	
264	100		
260	23.9	20.6	30.8
265	21.3	19.0	28.6






Abundance Scan 3852 (25.745 min): BP005349.D\data.ms (-3852) (-)




Indeno(1,2,3-cd)pyrene
Concen: 50.00 ng
RT: 25.751 min Scan# 3853
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	276	Resp:	677385
Ion Ratio	Lower	Upper	
276	100		
138	10.4	8.4	12.6
277	25.8	20.6	31.0







Abundance Scan 3342 (22.745 min): BP005349.D\data.ms (-3338) (-)



Benzo(b)fluoranthene
Concen: 49.36 ng
RT: 22.751 min Scan# 3343
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	252	Resp:	583811
Ion Ratio	Lower	Upper	
252	100		
253	21.7	18.2	27.2
125	4.1	3.4	5.0









Scan 3350 (22.792 min): BP005349.D\data.ms (-3349) (-)
 Benzo(k)fluoranthene
 Concen: 50.89 ng
 RT: 22.792 min Scan# 3350
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC050


Tgt Ion:	252	Resp:	577083
Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.2	25.8
125	4.3	3.8	5.6



Scan 3442 (23.333 min): BP005349.D\data.ms (-3440) (-)
 Benzo(a)pyrene
 Concen: 49.78 ng
 RT: 23.333 min Scan# 3442
 Delta R.T. -0.000 min
 Lab File: BP005350.D
 Acq: 20 Apr 2021 11:58

Tgt Ion:	252	Resp:	528598
Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.1	25.7
125	4.9	3.8	5.6









Dibenzo(a,h)anthracene
Concen: 50.17 ng
RT: 25.757 min Scan# 3854
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Instrument :
BNA_P
ClientSampleId :
SSTDIC050

Tgt Ion:	278	Resp:	593754
Ion Ratio	Lower	Upper	
278	100		
139	7.1	5.0	7.6
279	24.0	18.6	28.0



Benzo(g,h,i)perylene
Concen: 48.94 ng
RT: 26.451 min Scan# 3972
Delta R.T. 0.006 min
Lab File: BP005350.D
Acq: 20 Apr 2021 11:58

Tgt Ion:	276	Resp:	543843
Ion Ratio	Lower	Upper	
276	100		
277	24.2	19.8	29.8
138	9.5	7.7	11.5

