

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005352.D
 Acq On : 20 Apr 2021 13:06
 Operator : CG/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Quant Time: Apr 20 13:34:52 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 13:24:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.646	152	35143	20.00	ng	0.00
21) Naphthalene-d8	10.434	136	59745	20.00	ng	0.00
39) Acenaphthene-d10	14.310	164	43545	20.00	ng	0.00
64) Phenanthrene-d10	17.075	188	120645	20.00	ng	0.00
76) Chrysene-d12	21.174	240	185795	20.00	ng	0.00
86) Perylene-d12	23.427	264	171908	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.246	112	382658	163.20	ng	0.00
7) Phenol-d6	6.840	99	466460	166.70	ng	0.00
23) Nitrobenzene-d5	8.816	82	350964	162.02	ng	0.00
42) 2,4,6-Tribromophenol	15.816	330	186996	244.23	ng	0.00
45) 2-Fluorobiphenyl	12.922	172	631741	191.87	ng	0.00
79) Terphenyl-d14	19.651	244	1784845	186.29	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.181	88	91657	58.80	ng	98
3) Pyridine	3.581	79	195747	62.16	ng	99
4) n-Nitrosodimethylamine	3.493	42	69791	60.35	ng	# 61
6) Aniline	6.987	93	206278	84.13	ng	97
8) 2-Chlorophenol	7.216	128	193633	88.20	ng	97
9) Benzaldehyde	6.799	77	68687	55.10	ng	97
10) Phenol	6.869	94	224888	80.66	ng	98
11) bis(2-Chloroethyl)ether	7.087	93	160308	78.54	ng	97
12) 1,3-Dichlorobenzene	7.534	146	233435	82.44	ng	97
13) 1,4-Dichlorobenzene	7.681	146	227059	83.53	ng	97
14) 1,2-Dichlorobenzene	7.999	146	212242	82.82	ng	99
15) Benzyl Alcohol	7.899	79	104479	86.17	ng	92
16) 2,2'-oxybis(1-Chloropr...	8.187	45	101469	81.77	ng	97
17) 2-Methylphenol	8.105	107	117343	86.76	ng	98
18) Hexachloroethane	8.710	117	71004	69.38	ng	97
19) n-Nitroso-di-n-propyla...	8.463	70	98137	79.55	ng	97
20) 3+4-Methylphenols	8.434	107	141697	86.50	ng	97
22) Acetophenone	8.475	105	247505	88.22	ng	# 99
24) Nitrobenzene	8.857	77	167929	79.10	ng	98
25) Isophorone	9.381	82	206687	91.42	ng	98
26) 2-Nitrophenol	9.557	139	53120	92.10	ng	96
27) 2,4-Dimethylphenol	9.628	122	78445	92.83	ng	94
28) bis(2-Chloroethoxy)met...	9.863	93	107861	93.54	ng	98
29) 2,4-Dichlorophenol	10.099	162	101712	100.23	ng	97
30) 1,2,4-Trichlorobenzene	10.299	180	131843	94.27	ng	99
31) Naphthalene	10.487	128	312581	88.89	ng	99
32) Benzoic acid	9.834	122	52938	91.87	ng	# 89
33) 4-Chloroaniline	10.610	127	107810	94.79	ng	98
34) Hexachlorobutadiene	10.763	225	106605	88.62	ng	96
35) Caprolactam	11.434	113	22468	90.82	ng	84
36) 4-Chloro-3-methylphenol	11.751	107	95275	94.49	ng	93
37) 2-Methylnaphthalene	12.110	142	206768	95.89	ng	96
38) 1-Methylnaphthalene	12.328	142	195399	93.33	ng	97
40) 1,2,4,5-Tetrachloroben...	12.481	216	156259	97.01	ng	99
41) Hexachlorocyclopentadiene	12.457	237	67137	134.48	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042021\
 Data File : BP005352.D
 Acq On : 20 Apr 2021 13:06
 Operator : CG/JU
 Sample : SSTDICC080
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Quant Time: Apr 20 13:34:52 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 20 13:24:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.734	196	85565	100.28	ng	93
44) 2,4,5-Trichlorophenol	12.810	196	93758	99.34	ng	92
46) 1,1'-Biphenyl	13.134	154	281033	90.70	ng	97
47) 2-Chloronaphthalene	13.175	162	232948	92.31	ng	98
48) 2-Nitroaniline	13.398	65	67719	106.56	ng	98
49) Acenaphthylene	14.028	152	357539	93.57	ng	99
50) Dimethylphthalate	13.781	163	278648	98.51	ng	99
51) 2,6-Dinitrotoluene	13.904	165	68081	94.77	ng	88
52) Acenaphthene	14.375	154	227638	89.37	ng	95
53) 3-Nitroaniline	14.240	138	57168	97.05	ng	99
54) 2,4-Dinitrophenol	14.451	184	41640	103.68	ng	# 88
55) Dibenzofuran	14.710	168	425355	92.37	ng	99
56) 4-Nitrophenol	14.563	139	46017	101.36	ng	94
57) 2,4-Dinitrotoluene	14.698	165	107073	97.66	ng	93
58) Fluorene	15.369	166	367273	97.52	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.951	232	107656	105.88	ng	97
60) Diethylphthalate	15.151	149	293294	100.81	ng	98
61) 4-Chlorophenyl-phenyle...	15.363	204	218145	102.32	ng	98
62) 4-Nitroaniline	15.410	138	61530	101.66	ng	94
63) Azobenzene	15.657	77	379223	90.80	ng	99
65) 4,6-Dinitro-2-methylph...	15.469	198	72911	92.61	ng	95
66) n-Nitrosodiphenylamine	15.586	169	291507	86.42	ng	95
67) 4-Bromophenyl-phenylether	16.263	248	149977	101.23	ng	99
68) Hexachlorobenzene	16.375	284	181159	96.11	ng	98
69) Atrazine	16.551	200	113008	93.44	ng	96
70) Pentachlorophenol	16.728	266	99312	108.57	ng	94
71) Phenanthrene	17.116	178	613910	84.63	ng	100
72) Anthracene	17.210	178	608386	88.77	ng	98
73) Carbazole	17.486	167	549917	88.64	ng	100
74) Di-n-butylphthalate	18.045	149	507214	91.75	ng	# 96
75) Fluoranthene	19.098	202	847677	95.93	ng	100
77) Benzidine	19.286	184	178649	73.08	ng	99
78) Pyrene	19.451	202	891457	87.69	ng	100
80) Butylbenzylphthalate	20.327	149	192754	81.21	ng	96
81) Benzo(a)anthracene	21.163	228	998086	85.39	ng	99
82) 3,3'-Dichlorobenzidine	21.098	252	262777	90.75	ng	97
83) Chrysene	21.216	228	985902	84.76	ng	98
84) Bis(2-ethylhexyl)phtha...	21.086	149	310665	83.46	ng	98
85) Di-n-octyl phthalate	21.963	149	550463	79.77	ng	99
87) Indeno(1,2,3-cd)pyrene	25.757	276	1203159	89.41	ng	100
88) Benzo(b)fluoranthene	22.751	252	1068002	90.90	ng	98
89) Benzo(k)fluoranthene	22.798	252	1017663	90.35	ng	100
90) Benzo(a)pyrene	23.339	252	940972	89.23	ng	100
91) Dibenzo(a,h)anthracene	25.762	278	1063184	90.44	ng	99
92) Benzo(g,h,i)perylene	26.462	276	953137	86.35	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 20 13:34:52 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 20 13:24:27 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 7.646 min Scan# 775
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC080


Tgt Ion:	152	Resp:	35143
Ion	Ratio	Lower	Upper
152	100		
150	156.7	119.8	179.8
115	60.0	49.0	73.4



1,4-Dioxane
 Concen: 58.80 ng
 RT: 3.181 min Scan# 16
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	88	Resp:	91657
Ion	Ratio	Lower	Upper
88	100		
58	54.5	42.5	63.7
43	23.6	19.8	29.8









Pyridine
 Concen: 62.16 ng
 RT: 3.581 min Scan# 84
 Delta R.T. -0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Tgt Ion:	79	Resp:	195747
Ion	Ratio	Lower	Upper
79	100		
52	50.3	40.3	60.5
51	25.9	20.0	30.0



n-Nitrosodimethylamine
 Concen: 60.35 ng
 RT: 3.493 min Scan# 69
 Delta R.T. -0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	42	Resp:	69791
Ion	Ratio	Lower	Upper
42	100		
74	155.9	122.6	183.8
44	84.4	146.9	220.3#









Scan 367 (5.246 min): BP005349.D\data.ms (-359) #5
 2-Fluorophenol
 Concen: 163.20 ng
 RT: 5.246 min Scan# 367
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Tgt Ion:	112	Resp:	382658
Ion	Ratio	Lower	Upper
112	100		
64	41.5	33.6	50.4
63	23.2	18.6	27.8



Scan 663 (6.987 min): BP005349.D\data.ms (-656) #6
 Aniline
 Concen: 84.13 ng
 RT: 6.987 min Scan# 663
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	93	Resp:	206278
Ion	Ratio	Lower	Upper
93	100		
66	36.2	27.0	40.4
65	18.4	14.4	21.6









Phenol-d6
 Concen: 166.70 ng
 RT: 6.840 min Scan# 638
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC080


Tgt Ion:	99	Resp:	466460
Ion	Ratio	Lower	Upper
99	100		
42	14.9	12.9	19.3
71	33.9	26.7	40.1



2-Chlorophenol
 Concen: 88.20 ng
 RT: 7.216 min Scan# 702
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	128	Resp:	193633
Ion	Ratio	Lower	Upper
128	100		
130	32.5	14.1	54.1
64	35.2	17.2	57.2









Benzaldehyde
 Concen: 55.10 ng
 RT: 6.799 min Scan# 631
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Tgt Ion:	77	Resp:	68687
Ion	Ratio	Lower	Upper
77	100		
105	111.7	89.4	129.4
106	104.0	79.9	119.9



Phenol
 Concen: 80.66 ng
 RT: 6.869 min Scan# 643
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	94	Resp:	224888
Ion	Ratio	Lower	Upper
94	100		
65	26.9	8.0	48.0
66	36.2	17.3	57.3









bis(2-Chloroethyl)ether
Concen: 78.54 ng
RT: 7.087 min Scan# 680
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Instrument :
BNA_P
ClientSampleId :
SSTDIC080


Tgt Ion:	93	Resp:	160308
Ion	Ratio	Lower	Upper
93	100		
63	66.8	49.6	89.6
95	32.6	14.4	54.4



1,3-Dichlorobenzene
Concen: 82.44 ng
RT: 7.534 min Scan# 756
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	146	Resp:	233435
Ion	Ratio	Lower	Upper
146	100		
148	66.1	51.4	77.0
75	28.3	23.9	35.9









Scan 781 (7.681 min): BP005349.D\data.ms (-773) #13
 1,4-Dichlorobenzene
 Concen: 83.53 ng
 RT: 7.681 min Scan# 781
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC080


Tgt Ion:	146	Resp:	227059
Ion	Ratio	Lower	Upper
146	100		
148	66.9	52.2	78.4
111	39.3	34.2	51.4



Scan 834 (7.993 min): BP005349.D\data.ms (-822) #14
 1,2-Dichlorobenzene
 Concen: 82.82 ng
 RT: 7.999 min Scan# 835
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	146	Resp:	212242
Ion	Ratio	Lower	Upper
146	100		
148	64.3	51.5	77.3
111	41.2	34.0	51.0









Benzyl Alcohol
 Concen: 86.17 ng
 RT: 7.899 min Scan# 818
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Tgt Ion:	79	Resp:	104479
Ion	Ratio	Lower	Upper
79	100		
108	76.1	64.0	96.0
77	80.8	72.7	109.1



2,2'-oxybis(1-Chloropropane)
 Concen: 81.77 ng
 RT: 8.187 min Scan# 867
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	45	Resp:	101469
Ion	Ratio	Lower	Upper
45	100		
77	25.5	7.0	47.0
79	19.2	1.0	41.0









2-Methylphenol
 Concen: 86.76 ng
 RT: 8.105 min Scan# 853
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Tgt Ion:	107	Resp:	117343
Ion	Ratio	Lower	Upper
107	100		
108	110.1	85.8	128.8
77	49.4	41.4	62.0
79	48.0	38.6	57.8



Hexachloroethane
 Concen: 69.38 ng
 RT: 8.710 min Scan# 956
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06


Tgt Ion:	117	Resp:	71004
Ion	Ratio	Lower	Upper
117	100		
119	95.1	80.1	120.1
201	125.8	98.8	148.2






Abundance Scan 913 (8.458 min): BP005349.D\data.ms (-905) #19

n-Nitroso-di-n-propylamine
Concen: 79.55 ng
RT: 8.463 min Scan# 914
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:	70	Resp:	98137
Ion	Ratio	Lower	Upper
70	100		
42	48.0	39.7	59.5
101	10.9	7.4	11.0
130	23.0	17.2	25.8






Abundance Scan 909 (8.434 min): BP005349.D\data.ms (-897) #20

3+4-Methylphenols
Concen: 86.50 ng
RT: 8.434 min Scan# 909
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	107	Resp:	141697
Ion	Ratio	Lower	Upper
107	100		
108	88.9	69.5	109.5
77	38.8	21.5	61.5
79	33.0	16.8	56.8






Abundance Scan 1249 (10.434 min): BP005349.D\data.ms (-1198) (-)

Naphthalene-d8
Concen: 20.00 ng
RT: 10.434 min Scan# 1249
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:	136	Resp:	59745
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.9	13.3
54	6.4	5.0	7.4
68	4.9	3.7	5.5






Abundance Scan 916 (8.475 min): BP005349.D\data.ms (-902) (-)

Acetophenone
Concen: 88.22 ng
RT: 8.475 min Scan# 916
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	105	Resp:	247505
Ion	Ratio	Lower	Upper
105	100		
71	3.1	0.9	1.3#
51	22.8	18.2	27.2
120	21.8	17.0	25.6





Abundance Scan 973 (8.811 min): BP005349.D\data.ms (-964) #23

Nitrobenzene-d5
Concen: 162.02 ng
RT: 8.816 min Scan# 974
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Instrument :
BNA_P
ClientSampleId :
SSTDIC080



Tgt Ion:	82	Resp:	350964
Ion	Ratio	Lower	Upper
82	100		
128	39.5	30.5	45.7
54	42.8	33.9	50.9




Abundance Scan 974 (8.816 min): BP005352.D\data.ms (-922) (-)





Abundance Scan 980 (8.852 min): BP005349.D\data.ms (-967) #24


Nitrobenzene
Concen: 79.10 ng
RT: 8.857 min Scan# 981
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	77	Resp:	167929
Ion	Ratio	Lower	Upper
77	100		
123	41.3	31.7	47.5
65	13.9	11.5	17.3



Abundance Scan 981 (8.857 min): BP005352.D\data.ms (-929) (-)






Abundance Scan 1069 (9.375 min): BP005349.D\data.ms (-1025-)




Isophorone
Concen: 91.42 ng
RT: 9.381 min Scan# 1070
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080

Tgt Ion:	82	Resp:	206687
Ion	Ratio	Lower	Upper
82	100		
95	9.6	7.5	11.3
138	17.2	14.6	22.0






Abundance Scan 1100 (9.558 min): BP005349.D\data.ms (-1049-)




2-Nitrophenol
Concen: 92.10 ng
RT: 9.557 min Scan# 1100
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	139	Resp:	53120
Ion	Ratio	Lower	Upper
139	100		
109	47.3	40.7	61.1
65	57.9	47.8	71.6






Abundance Scan 1111 (9.622 min): BP005349.D\data.ms (-1037-)




2,4-Dimethylphenol
Concen: 92.83 ng
RT: 9.628 min Scan# 1112
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080

Tgt Ion:	122	Resp:	78445
Ion	Ratio	Lower	Upper
122	100		
107	126.8	93.8	140.8
121	58.1	45.6	68.4






Abundance Scan 1151 (9.858 min): BP005349.D\data.ms (-11428-)



bis(2-Chloroethoxy)methane
Concen: 93.54 ng
RT: 9.863 min Scan# 1152
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Tgt Ion:	93	Resp:	107861
Ion	Ratio	Lower	Upper
93	100		
95	32.6	24.9	37.3
123	11.4	9.5	14.3






Abundance Scan 1191 (10.093 min): BP005349.D\data.ms (-1129) (-)

2,4-Dichlorophenol
Concen: 100.23 ng
RT: 10.099 min Scan# 1192
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleID :
SSTDIC080




Tgt Ion:	162	Resp:	101712
Ion Ratio	Lower	Upper	
162	100		
164	62.0	44.1	84.1
98	34.1	12.6	52.6







Abundance Scan 1226 (10.299 min): BP005349.D\data.ms (-1130) (-)

1,2,4-Trichlorobenzene
Concen: 94.27 ng
RT: 10.299 min Scan# 1226
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06



Tgt Ion:	180	Resp:	131843
Ion Ratio	Lower	Upper	
180	100		
182	95.4	77.2	115.8
145	26.3	21.7	32.5









Naphthalene
 Concen: 88.89 ng
 RT: 10.487 min Scan# 1258
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Tgt Ion:	128	Resp:	312581
Ion	Ratio	Lower	Upper
128	100		
129	10.4	8.0	12.0
127	13.6	11.1	16.7



Benzoic acid
 Concen: 91.87 ng
 RT: 9.834 min Scan# 1147
 Delta R.T. 0.041 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	122	Resp:	52938
Ion	Ratio	Lower	Upper
122	100		
105	122.9	105.1	145.1
77	111.6	69.6	109.6#



Abundance Scan 1279 (10.610 min): BP005349.D\data.ms (-1283) (-)

4-Chloroaniline

Concen: 94.79 ng

RT: 10.610 min Scan# 1279

Delta R.T. -0.000 min

Lab File: BP005352.D


Acq: 20 Apr 2021 13:06

Instrument :

BNA_P

ClientSampleId :

SSTDIC080



Tgt Ion:127 Resp: 107810




Ion Ratio Lower Upper

127 100

129 32.2 25.5 38.3

65 39.0 30.0 45.0

92 23.7 20.2 30.2



Abundance Scan 1305 (10.763 min): BP005349.D\data.ms (-1284) (-)

Hexachlorobutadiene


Concen: 88.62 ng

RT: 10.763 min Scan# 1305

Delta R.T. -0.000 min

Lab File: BP005352.D

Acq: 20 Apr 2021 13:06






Tgt Ion:225 Resp: 106605

Ion Ratio Lower Upper

225 100

223 62.0 51.1 76.7


227 67.4 50.2 75.2





Abundance Scan 1414 (11.405 min): BP005349.D\data.ms (-1435)

Caprolactam
Concen: 90.82 ng
RT: 11.434 min Scan# 1419
Delta R.T. 0.029 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Instrument :
BNA_P
ClientSampleId :
SSTDIC080



Tgt Ion:	113	Resp:	22468
Ion	Ratio	Lower	Upper
113	100		
55	103.9	72.0	112.0
56	103.5	65.1	105.1




Abundance Scan 1419 (11.434 min): BP005352.D\data.ms (-1363) (-)





Abundance Scan 1473 (11.752 min): BP005349.D\data.ms (-1435)


4-Chloro-3-methylphenol
Concen: 94.49 ng
RT: 11.751 min Scan# 1473
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06



Tgt Ion:	107	Resp:	95275
Ion	Ratio	Lower	Upper
107	100		
144	23.3	20.1	30.1
142	70.6	62.2	93.4




Abundance Scan 1473 (11.751 min): BP005352.D\data.ms (-1422) (-)





Abundance Scan 1534 (12.110 min): BP005349.D\data.ms (-1483) (-)

2-Methylnaphthalene
Concen: 95.89 ng
RT: 12.110 min Scan# 1534
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:	142	Resp:	206768
Ion	Ratio	Lower	Upper
142	100		
141	90.8	74.7	112.1
115	50.8	43.8	65.8




Abundance Scan 1534 (12.110 min): BP005352.D\data.ms (-1483) (-)





Abundance Scan 1571 (12.328 min): BP005349.D\data.ms (-1520) (-)


1-Methylnaphthalene
Concen: 93.33 ng
RT: 12.328 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06



Tgt Ion:	142	Resp:	195399
Ion	Ratio	Lower	Upper
142	100		
141	93.6	77.0	115.6
116	5.5	5.0	7.6




Abundance Scan 1571 (12.328 min): BP005352.D\data.ms (-1520) (-)






Abundance Scan 1907 (14.304 min): BP005349.D\data.ms (-1858) (-)

Acenaphthene-d10
Concen: 20.00 ng
RT: 14.310 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080




Tgt Ion:	164	Resp:	43545
Ion Ratio	Lower	Upper	
164	100		
162	97.7	73.5	110.3
160	52.3	40.6	61.0






Abundance Scan 1597 (12.481 min): BP005349.D\data.ms (-1546) (-)


1,2,4,5-Tetrachlorobenzene
Concen: 97.01 ng
RT: 12.481 min Scan# 1597
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06



Tgt Ion:	216	Resp:	156259
Ion Ratio	Lower	Upper	
216	100		
214	77.0	60.6	91.0
179	19.0	15.3	22.9
108	18.0	13.8	20.6





Abundance Scan 1592 (12.452 min): BP005349.D\data.ms (-1541) (-)




Hexachlorocyclopentadiene
Concen: 134.48 ng
RT: 12.457 min Scan# 1593
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080


Tgt Ion:	237	Resp:	67137
Ion Ratio	Lower	Upper	
237	100		
235	62.4	40.9	80.9
272	11.3	0.0	34.6



Abundance Scan 1593 (12.457 min): BP005352.D\data.ms (-1541) (-)





Abundance Scan 2163 (15.810 min): BP005349.D\data.ms (-2112) (-)




2,4,6-Tribromophenol
Concen: 244.23 ng
RT: 15.816 min Scan# 2164
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	330	Resp:	186996
Ion Ratio	Lower	Upper	
330	100		
332	96.9	76.3	114.5
141	27.8	22.8	34.2




Abundance Scan 2164 (15.816 min): BP005352.D\data.ms (-2112) (-)





Abundance Scan 1640 (12.734 min): BP005349.D\data.ms (-1640) (-)

2,4,6-Trichlorophenol
Concen: 100.28 ng
RT: 12.734 min Scan# 1640
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:	196	Resp:	85565
Ion	Ratio	Lower	Upper
196	100		
198	95.1	81.4	122.2
200	30.7	28.2	42.2




Abundance Scan 1640 (12.734 min): BP005352.D\data.ms (-1589) (-)





Abundance Scan 1654 (12.816 min): BP005349.D\data.ms (-1644) (-)


2,4,5-Trichlorophenol
Concen: 99.34 ng
RT: 12.810 min Scan# 1653
Delta R.T. -0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	196	Resp:	93758
Ion	Ratio	Lower	Upper
196	100		
198	96.3	71.0	106.4
97	54.4	38.5	57.7
132	19.2	15.6	23.4



Abundance Scan 1653 (12.810 min): BP005352.D\data.ms (-1603) (-)






Abundance Scan 1672 (12.922 min): BP005349.D\data.ms (-1655) (-)




2-Fluorobiphenyl
Concen: 191.87 ng
RT: 12.922 min Scan# 1672
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080

Tgt Ion:	172	Resp:	631741
Ion	Ratio	Lower	Upper
172	100		
171	36.0	28.1	42.1
170	23.2	19.6	29.4







Abundance Scan 1708 (13.134 min): BP005349.D\data.ms (-1655) (-)



1,1'-Biphenyl
Concen: 90.70 ng
RT: 13.134 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	154	Resp:	281033
Ion	Ratio	Lower	Upper
154	100		
153	41.1	18.9	58.9
76	11.7	0.0	32.1









2-Chloronaphthalene
Concen: 92.31 ng
RT: 13.175 min Scan# 1715
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Instrument :
BNA_P
ClientSampleId :
SSTDIC080

Tgt Ion:	162	Resp:	232948
Ion	Ratio	Lower	Upper
162	100		
127	40.5	31.4	47.0
164	32.5	26.6	39.8



2-Nitroaniline
Concen: 106.56 ng
RT: 13.398 min Scan# 1753
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Tgt Ion:	65	Resp:	67719
Ion	Ratio	Lower	Upper
65	100		
92	64.6	48.8	73.2
138	87.4	69.5	104.3






Abundance Scan 1860 (14.028 min): BP005349.D\data.ms (-1809) (-)

Acenaphthylene
Concen: 93.57 ng
RT: 14.028 min Scan# 1860
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:	152	Resp:	357539
Ion	Ratio	Lower	Upper
152	100		
151	19.4	15.8	23.6
153	12.9	10.5	15.7






Abundance Scan 1817 (13.775 min): BP005349.D\data.ms (-1809) (-)

Dimethylphthalate
Concen: 98.51 ng
RT: 13.781 min Scan# 1818
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	163	Resp:	278648
Ion	Ratio	Lower	Upper
163	100		
194	4.9	3.9	5.9
164	10.6	8.1	12.1






Abundance Scan 1838 (13.899 min): BP005349.D\data.ms (-1850) (-)

2,6-Dinitrotoluene
Concen: 94.77 ng
RT: 13.904 min Scan# 1839
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC080




Tgt Ion:	165	Resp:	68081
Ion	Ratio	Lower	Upper
165	100		
63	73.8	67.4	101.0
89	79.9	73.8	110.6







Abundance Scan 1919 (14.375 min): BP005349.D\data.ms (-1852) (-)

Acenaphthene
Concen: 89.37 ng
RT: 14.375 min Scan# 1919
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06



Tgt Ion:	154	Resp:	227638
Ion	Ratio	Lower	Upper
154	100		
153	109.1	90.9	136.3
152	55.4	48.5	72.7









Scan 1895 (14.234 min): BP005349.D\data.ms (-1859) (-)
 3-Nitroaniline
 Concen: 97.05 ng
 RT: 14.240 min Scan# 1896
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleID :
 SSTDICC080

Tgt Ion:	138	Resp:	57168
Ion	Ratio	Lower	Upper
138	100		
108	14.9	10.2	15.4
92	136.3	108.1	162.1



Scan 1932 (14.451 min): BP005349.D\data.ms (-1954) (-)
 2,4-Dinitrophenol
 Concen: 103.68 ng
 RT: 14.451 min Scan# 1932
 Delta R.T. -0.000 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06


Tgt Ion:	184	Resp:	41640
Ion	Ratio	Lower	Upper
184	100		
63	78.3	51.2	76.8#
154	75.1	55.8	83.8





Abundance Scan 1976 (14.710 min): BP005349.D\data.ms (-1955) (-)

Dibenzofuran
Concen: 92.37 ng
RT: 14.710 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:	168	Resp:	425355
Ion	Ratio	Lower	Upper
168	100		
139	39.4	31.4	47.0
169	13.5	11.4	17.0




Abundance Scan 1976 (14.710 min): BP005352.D\data.ms (-1925) (-)





Abundance Scan 1951 (14.563 min): BP005349.D\data.ms (-1955) (-)


4-Nitrophenol
Concen: 101.36 ng
RT: 14.563 min Scan# 1951
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	139	Resp:	46017
Ion	Ratio	Lower	Upper
139	100		
109	147.4	122.8	162.8
65	139.2	110.0	150.0



Abundance Scan 1951 (14.563 min): BP005352.D\data.ms (-1900) (-)









2,4-Dinitrotoluene
Concen: 97.66 ng
RT: 14.698 min Scan# 1974
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Instrument :
BNA_P
ClientSampleId :
SSTDIC080

Tgt Ion:	165	Resp:	107073
Ion	Ratio	Lower	Upper
165	100		
63	68.1	56.7	85.1
89	97.9	85.8	128.6
182	3.7	3.0	4.6



Fluorene
Concen: 97.52 ng
RT: 15.369 min Scan# 2088
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Tgt Ion:	166	Resp:	367273
Ion	Ratio	Lower	Upper
166	100		
165	97.1	78.8	118.2
167	13.2	10.9	16.3






Abundance Scan 2016 (14.946 min): BP005349.D\data.ms (-2059) (-)

2,3,4,6-Tetrachlorophenol
Concen: 105.88 ng
RT: 14.951 min Scan# 2017
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080




Tgt Ion	Ratio	Lower	Upper
232	100		
131	35.4	26.5	39.7
130	2.5	2.1	3.1
166	25.2	19.4	29.0






Abundance Scan 2050 (15.146 min): BP005349.D\data.ms (-2060) (-)


Diethylphthalate
Concen: 100.81 ng
RT: 15.151 min Scan# 2051
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06



Tgt Ion	Ratio	Lower	Upper
149	100		
177	25.0	19.2	28.8
150	13.0	9.9	14.9





Abundance Scan 2087 (15.363 min): BP005349.D\data.ms (-2087) (-)




4-Chlorophenyl-phenylether
Concen: 102.32 ng
RT: 15.363 min Scan# 2087
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080


Tgt Ion:	204	Resp:	218145
Ion	Ratio	Lower	Upper
204	100		
206	32.0	26.0	39.0
141	53.1	44.2	66.2



Abundance Scan 2087 (15.363 min): BP005352.D\data.ms (-2036) (-)





Abundance Scan 2094 (15.404 min): BP005349.D\data.ms (-2094) (-)




4-Nitroaniline
Concen: 101.66 ng
RT: 15.410 min Scan# 2095
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Tgt Ion:	138	Resp:	61530
Ion	Ratio	Lower	Upper
138	100		
92	65.8	50.9	90.9
108	166.0	154.7	194.7



Abundance Scan 2095 (15.410 min): BP005352.D\data.ms (-2043) (-)





Abundance Scan 2137 (15.657 min): BP005349.D\data.ms (-2137) (-)




Azobenzene
Concen: 90.80 ng
RT: 15.657 min Scan# 2137
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080


Tgt Ion:	77	Resp:	379223
Ion	Ratio	Lower	Upper
77	100		
182	24.6	5.5	45.5
105	17.1	0.0	36.5
51	30.8	10.3	50.3



Abundance Scan 2137 (15.657 min): BP005352.D\data.ms (-2086) (-)





Abundance Scan 2377 (17.069 min): BP005349.D\data.ms (-2367) (-)





Phenanthrene-d10
Concen: 20.00 ng
RT: 17.075 min Scan# 2378
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	188	Resp:	120645
Ion	Ratio	Lower	Upper
188	100		
94	6.2	5.0	7.6
80	7.0	5.9	8.9



Abundance Scan 2378 (17.075 min): BP005352.D\data.ms (-2326) (-)









Scan 2104 (15.463 min): BP005349.D\data.ms (-2053) (-)
 4,6-Dinitro-2-methylphenol
 Concen: 92.61 ng
 RT: 15.469 min Scan# 2105
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Tgt Ion:	198	Resp:	72911
Ion	Ratio	Lower	Upper
198	100		
51	31.9	10.3	50.3
105	37.1	13.6	53.6



Scan 2124 (15.581 min): BP005349.D\data.ms (-2106) (-)
 n-Nitrosodiphenylamine
 Concen: 86.42 ng
 RT: 15.586 min Scan# 2125
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06


Tgt Ion:	169	Resp:	291507
Ion	Ratio	Lower	Upper
169	100		
168	67.2	55.2	82.8
167	40.3	36.6	54.8





Abundance Scan 2240 (16.263 min): BP005349.D\data.ms (-2189) (-)

4-Bromophenyl-phenylether
Concen: 101.23 ng
RT: 16.263 min Scan# 2240
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:	248	Resp:	149977
Ion	Ratio	Lower	Upper
248	100		
250	97.8	77.5	116.3
141	61.9	50.1	75.1




Abundance Scan 2240 (16.263 min): BP005352.D\data.ms (-2189) (-)





Abundance Scan 2259 (16.375 min): BP005349.D\data.ms (-2208) (-)


Hexachlorobenzene
Concen: 96.11 ng
RT: 16.375 min Scan# 2259
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	284	Resp:	181159
Ion	Ratio	Lower	Upper
284	100		
142	27.2	20.8	31.2
249	30.5	23.7	35.5



Abundance Scan 2259 (16.375 min): BP005352.D\data.ms (-2208) (-)






Abundance Scan 2288 (16.545 min): BP005349.D\data.ms (-2288) (-)




Atrazine
Concen: 93.44 ng
RT: 16.551 min Scan# 2289
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080

Tgt Ion:	200	Resp:	113008
Ion Ratio	Lower	Upper	
200	100		
173	23.7	0.7	40.7
215	51.2	29.0	69.0






Abundance Scan 2319 (16.728 min): BP005349.D\data.ms (-2319) (-)




Pentachlorophenol
Concen: 108.57 ng
RT: 16.728 min Scan# 2319
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	266	Resp:	99312
Ion Ratio	Lower	Upper	
266	100		
268	62.4	49.4	74.0
264	58.7	53.7	80.5



Abundance Scan 2384 (17.110 min): BP005349.D\data.ms (-2371)





Phenanthrene
Concen: 84.63 ng
RT: 17.116 min Scan# 2385
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Tgt Ion:	178	Resp:	613910
Ion	Ratio	Lower	Upper
178	100		
176	19.2	15.4	23.2
179	15.6	12.3	18.5

Instrument :
BNA_P
ClientSampleId :
SSTDIC080


Abundance Scan 2385 (17.116 min): BP005352.D\data.ms



Abundance Scan 2385 (17.116 min): BP005352.D\data.ms (-2333)





Abundance Scan 2400 (17.204 min): BP005349.D\data.ms (-2373)




Anthracene
Concen: 88.77 ng
RT: 17.210 min Scan# 2401
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	178	Resp:	608386
Ion	Ratio	Lower	Upper
178	100		
176	18.5	15.2	22.8
179	14.8	12.6	19.0


Abundance Scan 2401 (17.210 min): BP005352.D\data.ms



Abundance Scan 2401 (17.210 min): BP005352.D\data.ms (-2349)






Abundance Scan 2448 (17.487 min): BP005349.D\data.ms (-2459) (-)




Carbazole
Concen: 88.64 ng
RT: 17.486 min Scan# 2448
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080

Tgt Ion:	167	Resp:	549917
Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.4	26.0
139	12.9	10.4	15.6






Abundance Scan 2542 (18.039 min): BP005349.D\data.ms (-2554) (-)




Di-n-butylphthalate
Concen: 91.75 ng
RT: 18.045 min Scan# 2543
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	149	Resp:	507214
Ion	Ratio	Lower	Upper
149	100		
150	9.2	7.6	11.4
104	8.9	8.9	13.3#






Abundance Scan 2721 (19.092 min): BP005349.D\data.ms (-2715) (-)




Fluoranthene
Concen: 95.93 ng
RT: 19.098 min Scan# 2722
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:	202	Resp:	847677
Ion Ratio	Lower	Upper	
202	100		
101	6.5	0.0	26.1
203	17.1	0.0	37.2






Abundance Scan 3074 (21.169 min): BP005349.D\data.ms (-3075) (-)



Chrysene-d12
Concen: 20.00 ng
RT: 21.174 min Scan# 3075
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Tgt Ion:	240	Resp:	185795
Ion Ratio	Lower	Upper	
240	100		
120	5.5	4.2	6.2
236	25.4	20.2	30.2





Abundance Scan 2754 (19.286 min): BP005349.D\data.ms (-2775) (-)

Benzidine
Concen: 73.08 ng
RT: 19.286 min Scan# 2754
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06


Instrument :
BNA_P
ClientSampleId :
SSTDICC080



Tgt Ion:	184	Resp:	178649
Ion	Ratio	Lower	Upper
184	100		
185	14.1	11.1	16.7
183	12.6	9.8	14.8




Abundance Scan 2754 (19.286 min): BP005352.D\data.ms (-2703) (-)





Abundance Scan 2782 (19.451 min): BP005349.D\data.ms (-2773) (-)


Pyrene
Concen: 87.69 ng
RT: 19.451 min Scan# 2782
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06




Tgt Ion:	202	Resp:	891457
Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.9	25.3
203	17.6	13.8	20.6



Abundance Scan 2782 (19.451 min): BP005352.D\data.ms (-2731) (-)






Abundance Scan 2815 (19.645 min): BP005349.D\data.ms (-2809) (-)




Terphenyl-d14
Concen: 186.29 ng
RT: 19.651 min Scan# 2816
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDICC080

Tgt Ion:	244	Resp:	1784845
Ion	Ratio	Lower	Upper
244	100		
212	7.1	5.8	8.6
122	6.7	5.2	7.8







Abundance Scan 2931 (20.328 min): BP005349.D\data.ms (-2880) (-)



Butylbenzylphthalate
Concen: 81.21 ng
RT: 20.327 min Scan# 2931
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	149	Resp:	192754
Ion	Ratio	Lower	Upper
149	100		
91	78.0	60.2	90.4
206	26.9	23.9	35.9









Scan 3072 (21.157 min): BP005349.D\data.ms (-3085) (-)
 Benzo(a)anthracene
 Concen: 85.39 ng
 RT: 21.163 min Scan# 3073
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Tgt Ion:	228	Resp:	998086
Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.2	31.8
229	19.0	15.6	23.4



Scan 3061 (21.092 min): BP005349.D\data.ms (-3082) (-)
 3,3'-Dichlorobenzidine
 Concen: 90.75 ng
 RT: 21.098 min Scan# 3062
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	252	Resp:	262777
Ion	Ratio	Lower	Upper
252	100		
254	64.9	49.9	74.9
126	6.9	5.5	8.3









Scan 3081 (21.210 min): BP005349.D\data.ms (-3081) (-)
 Chrysene
 Concen: 84.76 ng
 RT: 21.216 min Scan# 3082
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Tgt Ion:	228	Resp:	985902
Ion	Ratio	Lower	Upper
228	100		
226	29.5	24.8	37.2
229	19.3	15.4	23.0




Scan 3059 (21.081 min): BP005349.D\data.ms (-3059) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 83.46 ng
 RT: 21.086 min Scan# 3060
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	149	Resp:	310665
Ion	Ratio	Lower	Upper
149	100		
167	30.6	23.8	35.8
279	8.2	5.7	8.5






Abundance Scan 3208 (21.957 min): BP005349.D\data.ms (-3208)




Di-n-octyl phthalate
Concen: 79.77 ng
RT: 21.963 min Scan# 3209
Delta R.T. 0.006 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	149	Resp:	550463
Ion	Ratio	Lower	Upper
149	100		
167	1.8	1.5	2.3
43	6.8	5.1	7.7

Instrument :
BNA_P
ClientSampleId :
SSTDICC080







Abundance Scan 3458 (23.427 min): BP005349.D\data.ms (-3458)



Perylene-d12
Concen: 20.00 ng
RT: 23.427 min Scan# 3458
Delta R.T. -0.000 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Tgt Ion:	264	Resp:	171908
Ion	Ratio	Lower	Upper
264	100		
260	25.2	20.6	30.8
265	22.4	19.0	28.6









Indeno(1,2,3-cd)pyrene
 Concen: 89.41 ng
 RT: 25.757 min Scan# 3854
 Delta R.T. 0.012 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080


Tgt Ion:	276	Resp:	1203159
Ion	Ratio	Lower	Upper
276	100		
138	10.9	8.4	12.6
277	25.7	20.6	31.0



Benzo(b)fluoranthene
 Concen: 90.90 ng
 RT: 22.751 min Scan# 3343
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06

Tgt Ion:	252	Resp:	1068002
Ion	Ratio	Lower	Upper
252	100		
253	21.8	18.2	27.2
125	4.7	3.4	5.0









Benzo(k)fluoranthene
 Concen: 90.35 ng
 RT: 22.798 min Scan# 3351
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC080

Tgt Ion:	252	Resp:	1017663
Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.2	25.8
125	4.3	3.8	5.6



Benzo(a)pyrene
 Concen: 89.23 ng
 RT: 23.339 min Scan# 3443
 Delta R.T. 0.006 min
 Lab File: BP005352.D
 Acq: 20 Apr 2021 13:06


Tgt Ion:	252	Resp:	940972
Ion	Ratio	Lower	Upper
252	100		
253	21.2	17.1	25.7
125	4.6	3.8	5.6






Abundance Scan 3853 (25.751 min): BP005349.D\data.ms (-3853) (-)

Dibenzo(a,h)anthracene
Concen: 90.44 ng
RT: 25.762 min Scan# 3855
Delta R.T. 0.012 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06

Instrument :
BNA_P
ClientSampleId :
SSTDIC080




Tgt Ion:	278	Resp:	1063184
Ion	Ratio	Lower	Upper
278	100		
139	7.3	5.0	7.6
279	23.2	18.6	28.0



Abundance Scan 3971 (26.445 min): BP005349.D\data.ms (-3971) (-)

Benzo(g,h,i)perylene
Concen: 86.35 ng
RT: 26.462 min Scan# 3974
Delta R.T. 0.017 min
Lab File: BP005352.D
Acq: 20 Apr 2021 13:06



Tgt Ion:	276	Resp:	953137
Ion	Ratio	Lower	Upper
276	100		
277	23.2	19.8	29.8
138	8.8	7.7	11.5

