

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042024\
 Data File : BP020059.D
 Acq On : 21 Apr 2024 02:45
 Operator : MA/JU
 Sample : P2104-02
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AA3

Quant Time: Apr 22 00:29:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 01:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	96774	20.000	ng/ul	0.00
20) Naphthalene-d8	10.416	136	377243	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.316	164	251480	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.157	188	553192	20.000	ng/ul	0.00
79) Chrysene-d12	21.651	240	546059	20.000	ng/ul	0.00
88) Perylene-d12	25.045	264	613984	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	7038	2.980	ng/uL	0.00
4) Pyridine-d5	3.652	84	12827	1.845	ng/ul	0.02
7) Phenol-d5	6.875	99	249	0.030	ng/ul	0.05
9) Bis-(2-Chloroethyl)eth...	6.993	67	95322	16.169	ng/ul	0.00
11) 2-Chlorophenol-d4	7.163	132	9	0.002	ng/ul	-0.01
15) 4-Methylphenol-d8	8.322	113	6	0.001	ng/ul	-0.02
21) Nitrobenzene-d5	8.805	128	46538	17.150	ng/ul	0.00
24) 2-Nitrophenol-d4	9.510	143	46218	15.757	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.057	165	10	0.002	ng/ul	0.02
31) 4-Chloroaniline-d4	10.575	131	5	0.001	ng/ul	0.00
46) Dimethylphthalate-d6	13.734	166	342480	19.927	ng/ul	0.00
49) Acenaphthylene-d8	14.004	160	2434	0.119	ng/ul	0.00
54) 4-Nitrophenol-d4	14.563	143	8	0.003	ng/ul	0.00
60) Fluorene-d10	15.345	176	37425	2.446	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.492	200	42052	13.012	ng/ul	0.00
73) Anthracene-d10	17.263	188	5893	0.238	ng/ul	0.00
81) Pyrene-d10	19.663	212	2836	0.097	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.815	264	516	0.017	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042024\
Data File : BP020059.D
Acq On : 21 Apr 2024 02:45
Operator : MA/JU
Sample : P2104-02
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AA3

Quant Time: Apr 22 00:29:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 20 01:32:50 2024
Response via : Initial Calibration

