

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042024\
 Data File : BP020063.D
 Acq On : 21 Apr 2024 05:36
 Operator : MA/JU
 Sample : P2104-06MS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AA7MS

Quant Time: Apr 22 00:30:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 01:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	163620	20.000	ng/u1	0.00
20) Naphthalene-d8	10.410	136	652949	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	372440	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.163	188	690849	20.000	ng/u1	0.00
79) Chrysene-d12	21.669	240	587634	20.000	ng/u1	0.01
88) Perylene-d12	25.068	264	662211	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	16417	4.111	ng/uL	0.00
4) Pyridine-d5	3.640	84	44286	3.768	ng/u1	0.00
7) Phenol-d5	6.846	99	1690	0.119	ng/u1	0.02
9) Bis-(2-Chloroethyl)eth...	6.993	67	196876	19.752	ng/u1	0.00
11) 2-Chlorophenol-d4	7.181	132	4079	0.419	ng/u1	0.00
15) 4-Methylphenol-d8	8.358	113	11372	1.021	ng/u1	0.01
21) Nitrobenzene-d5	8.799	128	98989	21.076	ng/u1	0.00
24) 2-Nitrophenol-d4	9.510	143	108304	21.332	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.116	165	298	0.030	ng/u1	0.08
31) 4-Chloroaniline-d4	10.581	131	6747	0.502	ng/u1	0.01
46) Dimethylphthalate-d6	13.734	166	594727	23.365	ng/u1	0.00
49) Acenaphthylene-d8	14.010	160	4469	0.147	ng/u1	0.01
54) 4-Nitrophenol-d4	14.575	143	33452	8.634	ng/u1	0.02
60) Fluorene-d10	15.345	176	53884	2.378	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.492	200	60971	15.107	ng/u1	0.00
73) Anthracene-d10	17.163	188	690849	22.321	ng/u1	-0.10
81) Pyrene-d10	19.663	212	4335	0.137	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.827	264	2101	0.065	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	31093	6.927	ng/uL	95
5) Pyridine	3.658	79	53504	4.396	ng/u1	96
6) Benzaldehyde	6.805	77	231390	27.143	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.081	93	292733	23.857	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.158	45	425829	23.744	ng/u1	98
16) Acetophenone	8.463	105	454261	23.806	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.446	70	263718	23.660	ng/u1	93
18) 4-Methylphenol	8.422	108	12668	1.081	ng/u1	95
22) Nitrobenzene	8.840	77	358502	23.422	ng/u1	99
23) Isophorone	9.357	82	693742	24.420	ng/u1	98
25) 2-Nitrophenol	9.546	139	133831	24.748	ng/u1	94
26) 2,4-Dimethylphenol	9.610	107	26881	2.078	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.846	93	394398	24.759	ng/u1	99
32) 4-Chloroaniline	10.604	127	55840	4.186	ng/u1	99
34) Caprolactam	11.387	113	75245	24.481	ng/u1	94
43) 1,1'-Biphenyl	13.128	154	165201	6.187	ng/u1	99
45) 2-Nitroaniline	13.398	65	117315	16.928	ng/u1	97
47) Dimethylphthalate	13.781	163	699950	26.780	ng/u1	100
48) 2,6-Dinitrotoluene	13.910	165	138316	28.044	ng/u1	95
51) 3-Nitroaniline	14.251	138	94951	20.340	ng/u1	98
53) 2,4-Dinitrophenol	14.475	184	44388	14.967	ng/u1	92
55) 4-Nitrophenol	14.587	109	44845	10.159	ng/u1	94
57) 2,4-Dinitrotoluene	14.728	165	189664	27.165	ng/u1#	69
59) Diethylphthalate	15.187	149	673288	25.653	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042024\
 Data File : BP020063.D
 Acq On : 21 Apr 2024 05:36
 Operator : MA/JU
 Sample : P2104-06MS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AA7MS

Quant Time: Apr 22 00:30:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 01:32:50 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
61) Fluorene	15.398	166	54960	2.137	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.398	204	116741	8.356	ng/ul	98
63) 4-Nitroaniline	15.445	138	112823	27.221	ng/ul	89
66) 4,6-Dinitro-2-methylph...	15.504	198	79600	18.976	ng/ul	97
67) N-Nitrosodiphenylamine	15.628	169	249103	13.035	ng/ul	99
68) 4-Bromophenyl-phenylether	16.328	248	52322	6.906	ng/ul	97
70) Atrazine	16.628	200	203347	28.977	ng/ul	97
78) Di-n-butylphthalate	18.198	149	1075786	28.443	ng/ul	99
83) Butylbenzylphthalate	20.680	149	459323	31.187	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.580	252	87842	6.760	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.598	149	669385	29.482	ng/ul	98
89) Di-n-octyl phthalate	22.910	149	1169531	28.575	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042024\
Data File : BP020063.D
Acq On : 21 Apr 2024 05:36
Operator : MA/JU
Sample : P2104-06MS
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AA7MS

Quant Time: Apr 22 00:30:36 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 20 01:32:50 2024
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 6.927 ng/uL
 RT: 3.264 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BP020063.D
 Acq: 21 Apr 2024 05:36




Instrument :
 BNA_P
 ClientSampleId :
 C0AA7MS


Tgt Ion:	88	Resp:	31093
Ion	Ratio	Lower	Upper
88	100		
43	33.7	26.6	40.0
58	71.2	61.4	92.2



#5
 Pyridine
 Concen: 4.396 ng/uL
 RT: 3.658 min Scan# 97
 Delta R.T. 0.006 min
 Lab File: BP020063.D
 Acq: 21 Apr 2024 05:36

Tgt Ion:	79	Resp:	53504
Ion	Ratio	Lower	Upper
79	100		
52	79.8	61.0	91.4
51	42.0	31.5	47.3









#6
Benzaldehyde
Concen: 27.143 ng/ul
RT: 6.805 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion: 77 Resp: 231390
Ion Ratio Lower Upper
77 100
105 85.9 68.7 103.1
106 80.8 63.7 95.5



#10
Bis(2-Chloroethyl)ether
Concen: 23.857 ng/ul
RT: 7.081 min Scan# 679
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion: 93 Resp: 292733
Ion Ratio Lower Upper
93 100
63 82.4 68.6 102.8
95 33.4 26.6 39.8









#14
2,2'-oxybis(1-Chloropropane)
Concen: 23.744 ng/ul
RT: 8.158 min Scan# 862
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion: 45 Resp: 425829
Ion Ratio Lower Upper
45 100
77 15.1 12.6 18.8
79 12.0 8.8 13.2



#16
Acetophenone
Concen: 23.806 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion: 105 Resp: 454261
Ion Ratio Lower Upper
105 100
77 87.3 70.3 105.5
51 30.7 25.1 37.7









#17
N-Nitroso-di-n-propylamine
Concen: 23.660 ng/ul
RT: 8.446 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Instrument :
BNA_P
ClientSampleId :
C0AA7MS

Tgt Ion: 70 Resp: 263718




Ion	Ratio	Lower	Upper
70	100		
42	53.4	47.8	71.6
101	8.9	6.5	9.7
130	17.0	12.5	18.7




#18
4-Methylphenol
Concen: 1.081 ng/ul
RT: 8.422 min Scan# 907
Delta R.T. 0.018 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion: 108 Resp: 12668

Ion	Ratio	Lower	Upper
108	100		
107	109.1	91.5	137.3









#22
Nitrobenzene
Concen: 23.422 ng/ul
RT: 8.840 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion: 77 Resp: 358502
Ion Ratio Lower Upper
77 100
123 36.4 28.5 42.7
65 14.0 11.7 17.5



#23
Isophorone
Concen: 24.420 ng/ul
RT: 9.357 min Scan# 1066
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion: 82 Resp: 693742
Ion Ratio Lower Upper
82 100
95 7.8 6.0 9.0
138 14.5 10.9 16.3









#25
2-Nitrophenol
Concen: 24.748 ng/ul
RT: 9.546 min Scan# 1109
Delta R.T. 0.006 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion:139 Resp: 133831
Ion Ratio Lower Upper
139 100
65 66.8 58.6 87.8
109 40.6 34.8 52.2



#26
2,4-Dimethylphenol
Concen: 2.078 ng/ul
RT: 9.610 min Scan# 1109
Delta R.T. 0.012 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion:107 Resp: 26881
Ion Ratio Lower Upper
107 100
121 40.8 34.0 51.0
122 71.9 58.9 88.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 24.759 ng/ul
 RT: 9.846 min Scan# 1149
 Delta R.T. 0.006 min
 Lab File: BP020063.D
 Acq: 21 Apr 2024 05:36




Instrument :
 BNA_P
 ClientSampleId :
 C0AA7MS


Tgt Ion: 93 Resp: 394398
 Ion Ratio Lower Upper
 93 100
 95 32.1 25.4 38.0
 123 9.6 7.5 11.3



#32
 4-Chloroaniline
 Concen: 4.186 ng/ul
 RT: 10.604 min Scan# 1278
 Delta R.T. 0.012 min
 Lab File: BP020063.D
 Acq: 21 Apr 2024 05:36

Tgt Ion: 127 Resp: 55840
 Ion Ratio Lower Upper
 127 100
 129 31.9 25.9 38.9





#34

Caprolactam

Concen: 24.481 ng/ul

RT: 11.387 min Scan# 1415

Delta R.T. -0.012 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36

Instrument :

BNA_P

ClientSampleId :

C0AA7MS





Tgt Ion:113 Resp: 75245

Ion Ratio Lower Upper

113 100

55 188.1 140.5 210.7

56 146.3 113.8 170.6



#43

1,1'-Biphenyl

Concen: 6.187 ng/ul

RT: 13.128 min Scan# 1707

Delta R.T. 0.000 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36




Tgt Ion:154 Resp: 165201


Ion Ratio Lower Upper

154 100

153 40.1 32.2 48.4

76 15.2 11.3 16.9









#45
2-Nitroaniline
Concen: 16.928 ng/ul
RT: 13.398 min Scan# 1753
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion: 65 Resp: 117315
Ion Ratio Lower Upper
65 100
92 63.8 49.4 74.0
138 81.0 62.2 93.4



#47
Dimethylphthalate
Concen: 26.780 ng/ul
RT: 13.781 min Scan# 1818
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion:163 Resp: 699950
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.1 8.1 12.1





#48

2,6-Dinitrotoluene

Concen: 28.044 ng/ul

RT: 13.910 min Scan# 1841

Delta R.T. 0.000 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36

Instrument :

BNA_P

ClientSampleId :

C0AA7MS





Tgt Ion:165 Resp: 138316

Ion Ratio Lower Upper

165 100

89 70.2 61.0 91.4

121 23.2 18.7 28.1



#51

3-Nitroaniline

Concen: 20.340 ng/ul

RT: 14.251 min Scan# 1898

Delta R.T. 0.000 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36




Tgt Ion:138 Resp: 94951


Ion Ratio Lower Upper

138 100

108 12.4 10.4 15.6

92 133.1 108.4 162.6









#53
2,4-Dinitrophenol
Concen: 14.967 ng/ul
RT: 14.475 min Scan# 1936
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion:184 Resp: 44388
Ion Ratio Lower Upper
184 100
63 87.3 65.0 97.6
154 76.3 55.4 83.2



#55
4-Nitrophenol
Concen: 10.159 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. 0.012 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion:109 Resp: 44845
Ion Ratio Lower Upper
109 100
139 95.4 77.9 116.9
65 136.4 100.9 151.3









#57
2,4-Dinitrotoluene
Concen: 27.165 ng/ul
RT: 14.728 min Scan# 1979
Delta R.T. 0.006 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion:165 Resp: 189664
Ion Ratio Lower Upper
165 100
63 58.2 70.2 105.4#
182 2.6 2.0 3.0



#59
Diethylphthalate
Concen: 25.653 ng/ul
RT: 15.187 min Scan# 2057
Delta R.T. 0.006 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion:149 Resp: 673288
Ion Ratio Lower Upper
149 100
177 22.5 16.9 25.3
150 12.9 9.9 14.9









#61
Fluorene
Concen: 2.137 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion:166 Resp: 54960
Ion Ratio Lower Upper
166 100
165 97.9 78.2 117.4
167 14.6 10.9 16.3



#62
4-Chlorophenyl-phenylether
Concen: 8.356 ng/ul
RT: 15.398 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion:204 Resp: 116741
Ion Ratio Lower Upper
204 100
206 33.1 25.1 37.7
141 62.9 49.3 73.9









#63
4-Nitroaniline
Concen: 27.221 ng/ul
RT: 15.445 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36




Instrument :
BNA_P
ClientSampleId :
C0AA7MS


Tgt Ion:138 Resp: 112823
Ion Ratio Lower Upper
138 100
92 61.4 55.2 82.8
108 106.3 95.4 143.0



#66
4,6-Dinitro-2-methylphenol
Concen: 18.976 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BP020063.D
Acq: 21 Apr 2024 05:36

Tgt Ion:198 Resp: 79600
Ion Ratio Lower Upper
198 100
182 4.0 3.0 4.6
77 30.4 25.6 38.4





#67

N-Nitrosodiphenylamine

Concen: 13.035 ng/ul

RT: 15.628 min Scan# 2133

Delta R.T. 0.006 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36

Instrument :

BNA_P

ClientSampleId :

C0AA7MS





Tgt Ion:169 Resp: 249103

Ion Ratio Lower Upper

169 100

168 67.2 55.0 82.6

167 36.8 29.5 44.3



#68

4-Bromophenyl-phenylether

Concen: 6.906 ng/ul

RT: 16.328 min Scan# 2251

Delta R.T. 0.006 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36




Tgt Ion:248 Resp: 52322


Ion Ratio Lower Upper

248 100

250 97.8 79.2 118.8

141 77.7 65.1 97.7





#70

Atrazine

Concen: 28.977 ng/ul

RT: 16.628 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36

Instrument :

BNA_P

ClientSampleId :

C0AA7MS





Tgt Ion:200 Resp: 203347

Ion Ratio Lower Upper

200 100

173 24.5 19.3 28.9

215 48.4 36.6 54.8



#78

Di-n-butylphthalate

Concen: 28.443 ng/ul

RT: 18.198 min Scan# 2569

Delta R.T. 0.006 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36




Tgt Ion:149 Resp: 1075786


Ion Ratio Lower Upper

149 100

150 9.3 7.4 11.0

104 6.3 5.3 7.9





#83

Butylbenzylphthalate

Concen: 31.187 ng/ul

RT: 20.680 min Scan# 2991

Delta R.T. 0.006 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36

Instrument :

BNA_P

ClientSampleId :

C0AA7MS





Tgt Ion:149 Resp: 459323

Ion Ratio Lower Upper

149 100

91 80.8 69.6 104.4

206 23.1 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 6.760 ng/ul

RT: 21.580 min Scan# 3144

Delta R.T. 0.012 min

Lab File: BP020063.D

Acq: 21 Apr 2024 05:36




Tgt Ion:252 Resp: 87842


Ion Ratio Lower Upper

252 100

254 66.1 51.5 77.3

126 10.4 7.9 11.9









#86
 Bis(2-ethylhexyl)phthalate
 Concen: 29.482 ng/ul
 RT: 21.598 min Scan# 3146
 Delta R.T. 0.006 min
 Lab File: BP020063.D
 Acq: 21 Apr 2024 05:36

Instrument :
 BNA_P
 ClientSampleId :
 C0AA7MS

Tgt Ion:149 Resp: 669385
 Ion Ratio Lower Upper
 149 100
 167 27.6 21.3 31.9
 279 4.7 3.2 4.8



#89
 Di-n-octyl phthalate
 Concen: 28.575 ng/ul
 RT: 22.910 min Scan# 3370
 Delta R.T. 0.012 min
 Lab File: BP020063.D
 Acq: 21 Apr 2024 05:36

Tgt Ion:149 Resp: 1169531

