

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042024\
 Data File : BP020066.D
 Acq On : 21 Apr 2024 07:44
 Operator : MA/JU
 Sample : P2104-09
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AA9

Quant Time: Apr 22 00:31:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 20 01:32:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.646	152	139154	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	575584	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	352227	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.163	188	639305	20.000	ng/u1	0.00
79) Chrysene-d12	21.669	240	567630	20.000	ng/u1	0.01
88) Perylene-d12	25.074	264	630742	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.234	96	7309	2.152	ng/uL	0.00
4) Pyridine-d5	3.658	84	10458	1.046	ng/u1	0.02
7) Phenol-d5	6.875	99	933	0.077	ng/u1	0.05
9) Bis-(2-Chloroethyl)eth...	6.999	67	120032	14.160	ng/u1	0.00
11) 2-Chlorophenol-d4	7.187	132	2563	0.310	ng/u1	0.01
15) 4-Methylphenol-d8	8.364	113	4205	0.444	ng/u1	0.02
21) Nitrobenzene-d5	8.805	128	62478	15.090	ng/u1	0.00
24) 2-Nitrophenol-d4	9.516	143	67620	15.109	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.152	165	181	0.020	ng/u1	0.11
31) 4-Chloroaniline-d4	10.557	131	14	0.001	ng/u1	-0.01
46) Dimethylphthalate-d6	13.728	166	426206	17.706	ng/u1	0.00
49) Acenaphthylene-d8	14.004	160	3398	0.118	ng/u1	0.00
54) 4-Nitrophenol-d4	14.563	143	40	0.011	ng/u1	0.00
60) Fluorene-d10	15.351	176	43367	2.024	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.498	200	17048	4.565	ng/u1	0.01
73) Anthracene-d10	17.163	188	639305	22.321	ng/u1	-0.10
81) Pyrene-d10	19.681	212	3419	0.112	ng/u1	0.02
92) Benzo(a)pyrene-d12	24.833	264	3390	0.110	ng/u1	0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042024\
Data File : BP020066.D
Acq On : 21 Apr 2024 07:44
Operator : MA/JU
Sample : P2104-09
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AA9

Quant Time: Apr 22 00:31:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 20 01:32:50 2024
Response via : Initial Calibration

