


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042121\
 Data File : BP005355.D
 Acq On : 21 Apr 2021 12:19
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 21 17:40:15 2021






AutoFind: Scans 2454, 2455, 2456; Background Corrected with Scan 2447

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	32.2	17489	PASS
68	69	0.00	2	1.8	380	PASS
69	198	0.00	100	38.3	20801	PASS
70	69	0.00	2	0.8	166	PASS
127	198	10	80	37.8	20539	PASS
197	198	0.00	2	0.2	107	PASS
198	198	100	100	100.0	54328	PASS
199	198	5	9	7.0	3799	PASS
275	198	10	60	33.4	18163	PASS
365	198	1	100	5.7	3076	PASS
441	198	0.01	100	14.5	7882	PASS
442	442	100	100	100.0	52660	PASS
443	442	15	24	20.3	10693	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042121\
Data File : BP005355.D
Acq On : 21 Apr 2021 12:19
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DFTPP

Quant Time: Apr 21 13:51:59 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 21 13:46:46 2021
Response via : Initial Calibration



TIC: BP005355.D\data.ms

(77) Benzidine

19.298min (+ 0.012) 1604013.96 ng




response 344928

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	13.25
183.00	12.30	11.56
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042121\
 Data File : BP005355.D
 Acq On : 21 Apr 2021 12:19
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Apr 21 13:51:59 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 13:46:46 2021
 Response via : Initial Calibration



TIC: BP005355.D\data.ms

(70) Pentachlorophenol (C)

16.739min (+ 0.012) 1023054.17 ng

response 72354

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.70	61.80
264.00	67.10	60.48
0.00	0.00	0.00

DDT Breakdown

Instrument :
BNA_P
ClientSampleId :
DFTPP

Date	Instrument Name	DFTPP Data File
21/4/2021	BNA_P	BP005355.D
Compound Name	Response	Retention Time
DDT	226758	20.49
DDD	1441	20.06
DDE	604	19.56
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
2045	228803	0.89