

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042224\
 Data File : BP020074.D
 Acq On : 22 Apr 2024 13:01
 Operator : MA/JU
 Sample : PB160401BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK401

Quant Time: Apr 23 01:34:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 23 01:17:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	154431	20.000	ng/ul	0.00
20) Naphthalene-d8	10.416	136	593826	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.322	164	330674	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.180	188	696071	20.000	ng/ul	0.02
79) Chrysene-d12	21.674	240	648961	20.000	ng/ul	0.02
88) Perylene-d12	25.080	264	723061	20.000	ng/ul	0.03
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	24816	6.584	ng/uL	0.00
4) Pyridine-d5	3.628	84	367034	33.085	ng/ul	0.00
7) Phenol-d5	6.822	99	502610	37.343	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	340655	36.211	ng/ul	0.00
11) 2-Chlorophenol-d4	7.175	132	360343	39.212	ng/ul	0.00
15) 4-Methylphenol-d8	8.346	113	376670	35.827	ng/ul	0.00
21) Nitrobenzene-d5	8.799	128	178369	41.758	ng/ul	0.00
24) 2-Nitrophenol-d4	9.516	143	201845	43.715	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.046	165	332185	36.265	ng/ul	0.00
31) 4-Chloroaniline-d4	10.569	131	444632	36.362	ng/ul	0.00
46) Dimethylphthalate-d6	13.739	166	905246	40.057	ng/ul	0.01
49) Acenaphthylene-d8	14.010	160	1017457	37.774	ng/ul	0.01
54) 4-Nitrophenol-d4	14.575	143	138506	40.262	ng/ul	0.02
60) Fluorene-d10	15.351	176	751742	37.369	ng/ul	0.01
65) 4,6-Dinitro-2-methylph...	15.504	200	84590	20.802	ng/ul	0.02
73) Anthracene-d10	17.275	188	1195916	38.350	ng/ul	0.01
81) Pyrene-d10	19.674	212	1476209	42.327	ng/ul	0.01
92) Benzo(a)pyrene-d12	24.845	264	1399039	39.609	ng/ul	0.04

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042224\
 Data File : BP020074.D
 Acq On : 22 Apr 2024 13:01
 Operator : MA/JU
 Sample : PB160401BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK401

Quant Time: Apr 23 01:34:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP041924.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 23 01:17:52 2024
 Response via : Initial Calibration

