

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042321\
 Data File : BP005396.D
 Acq On : 23 Apr 2021 16:32
 Operator : CG/JU
 Sample : M2150-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 TWP-40

Manual Integrations
 APPROVED

mohammad
 4/26/2021 11:37:13 AM

Quant Time: Apr 23 17:08:01 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 21 17:40:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.634	152	20029	20.00	ng	-0.01
21) Naphthalene-d8	10.422	136	68814	20.00	ng	#-0.01
39) Acenaphthene-d10	14.298	164	46790	20.00	ng	0.00
64) Phenanthrene-d10	17.069	188	113196	20.00	ng	0.00
76) Chrysene-d12	21.175	240	163788	20.00	ng	# 0.00
86) Perylene-d12	23.433	264	182875	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.234	112	86474	75.00	ng	-0.01
7) Phenol-d6	6.822	99	63679	40.08	ng	-0.01
23) Nitrobenzene-d5	8.799	82	235387	148.78	ng	0.00
42) 2,4,6-Tribromophenol	15.810	330	124614	139.23	ng	0.00
45) 2-Fluorobiphenyl	12.910	172	383013	105.96	ng	0.00
79) Terphenyl-d14	19.651	244	777146	87.93	ng	0.00
Target Compounds						
10) Phenol	6.852	94	20305	12.81	ng	# 56
50) Dimethylphthalate	13.763	163	13782	4.06	ng	99
74) Di-n-butylphthalate	18.039	149	2337m	4.32	ng	
84) Bis(2-ethylhexyl)phtha...	21.086	149	1864	3.45	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042321\
Data File : BP005396.D
Acq On : 23 Apr 2021 16:32
Operator : CG/JU
Sample : M2150-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
TWP-40

Manual Integrations
APPROVED

mohammad
4/26/2021 11:37:13 AM

Quant Time: Apr 23 17:08:01 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 21 17:40:15 2021
Response via : Initial Calibration

