

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042324\
 Data File : BP020101.D
 Acq On : 23 Apr 2024 12:14
 Operator : MA/JU
 Sample : SST04097
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040619

Quant Time: Apr 23 23:57:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 23 14:20:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.640	152	146794	20.000	ng/u1	0.00
20) Naphthalene-d8	10.416	136	539926	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.322	164	284780	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.163	188	598405	20.000	ng/u1	0.00
79) Chrysene-d12	21.663	240	587934	20.000	ng/u1	0.00
88) Perylene-d12	25.057	264	638226	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.228	96	52122	16.132	ng/uL	0.00
4) Pyridine-d5	3.628	84	397658	41.040	ng/u1	0.00
7) Phenol-d5	6.828	99	508166	41.278	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.993	67	332401	41.536	ng/u1	0.00
11) 2-Chlorophenol-d4	7.181	132	362957	40.777	ng/u1	0.00
15) 4-Methylphenol-d8	8.352	113	382351	40.435	ng/u1	0.01
21) Nitrobenzene-d5	8.805	128	177517	44.162	ng/u1	0.01
24) 2-Nitrophenol-d4	9.516	143	203334	43.787	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.046	165	349656	41.211	ng/u1	0.00
31) 4-Chloroaniline-d4	10.563	131	424691	39.800	ng/u1	0.00
46) Dimethylphthalate-d6	13.734	166	793763	39.889	ng/u1	0.01
49) Acenaphthylene-d8	14.004	160	942942	40.374	ng/u1	0.00
54) 4-Nitrophenol-d4	14.575	143	126121	42.356	ng/u1	0.01
60) Fluorene-d10	15.351	176	672222	39.142	ng/u1	0.02
65) 4,6-Dinitro-2-methylph...	15.498	200	111212	38.739	ng/u1	0.00
73) Anthracene-d10	17.275	188	1081548	40.810	ng/u1	0.02
81) Pyrene-d10	19.675	212	1342782	40.135	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.815	264	1260032	40.611	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.264	88	56965	15.277	ng/uL	98
5) Pyridine	3.646	79	400005	40.332	ng/u1	97
6) Benzaldehyde	6.811	77	317995	45.832	ng/u1	95
8) Phenol	6.858	94	530613	41.268	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.087	93	425643	41.471	ng/u1	99
12) 2-Chlorophenol	7.210	128	383769	41.604	ng/u1	98
13) 2-Methylphenol	8.081	108	378872	40.591	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.163	45	587732	41.176	ng/u1	97
16) Acetophenone	8.469	105	625740	40.297	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.452	70	358835	40.728	ng/u1	99
18) 4-Methylphenol	8.416	108	400007	40.308	ng/u1	98
19) Hexachloroethane	8.693	117	180031	39.859	ng/u1	98
22) Nitrobenzene	8.846	77	526585	43.013	ng/u1	97
23) Isophorone	9.363	82	926380	41.904	ng/u1	99
25) 2-Nitrophenol	9.546	139	206423	42.913	ng/u1	94
26) 2,4-Dimethylphenol	9.604	107	434528	41.446	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.846	93	538301	42.586	ng/u1	100
29) 2,4-Dichlorophenol	10.075	162	344862	41.756	ng/u1	96
30) Naphthalene	10.463	128	1123002	40.063	ng/u1	99
32) 4-Chloroaniline	10.587	127	424956	40.614	ng/u1	97
33) Hexachlorobutadiene	10.728	225	298118	41.880	ng/u1	100
34) Caprolactam	11.404	113	67382	31.754	ng/u1	95
35) 4-Chloro-3-methylphenol	11.728	107	354627	37.626	ng/u1	99
36) 2-Methylnaphthalene	12.093	142	689338	37.172	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042324\
 Data File : BP020101.D
 Acq On : 23 Apr 2024 12:14
 Operator : MA/JU
 Sample : SST04097
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040619

Quant Time: Apr 23 23:57:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 23 14:20:59 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.316	142	677391	36.699	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.463	216	423998	42.556	ng/ul	97
40) Hexachlorocyclopentadiene	12.422	237	144020	40.030	ng/ul	97
41) 2,4,6-Trichlorophenol	12.722	196	248912	41.668	ng/ul	96
42) 2,4,5-Trichlorophenol	12.798	196	260358	42.020	ng/ul	95
43) 1,1'-Biphenyl	13.122	154	861181	41.795	ng/ul	98
44) 2-Chloronaphthalene	13.169	162	693597	41.560	ng/ul	99
45) 2-Nitroaniline	13.398	65	226883	42.046	ng/ul	95
47) Dimethylphthalate	13.781	163	820313	40.517	ng/ul	100
48) 2,6-Dinitrotoluene	13.910	165	164827	40.725	ng/ul	96
50) Acenaphthylene	14.034	152	1041130	39.857	ng/ul	99
51) 3-Nitroaniline	14.251	138	167726	42.000	ng/ul	94
52) Acenaphthene	14.387	153	680315	39.561	ng/ul	99
53) 2,4-Dinitrophenol	14.475	184	67044	38.977	ng/ul	92
55) 4-Nitrophenol	14.587	109	133765	43.113	ng/ul	96
56) Dibenzofuran	14.734	168	924905	38.862	ng/ul	99
57) 2,4-Dinitrotoluene	14.728	165	232164	40.447	ng/ul	87
58) 2,3,4,6-Tetrachlorophenol	14.975	232	217061	39.264	ng/ul	97
59) Diethylphthalate	15.187	149	762048	38.266	ng/ul	99
61) Fluorene	15.404	166	753608	38.521	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.404	204	409653	38.533	ng/ul	98
63) 4-Nitroaniline	15.451	138	158580	43.051	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.516	198	114907	38.231	ng/ul#	91
67) N-Nitrosodiphenylamine	15.634	169	621455	39.404	ng/ul	99
68) 4-Bromophenyl-phenylether	16.334	248	259630	39.805	ng/ul	98
69) Hexachlorobenzene	16.434	284	299779	40.590	ng/ul	99
70) Atrazine	16.639	200	242348	39.645	ng/ul	100
71) Pentachlorophenol	16.804	266	184157	39.920	ng/ul	98
72) Phenanthrene	17.210	178	1241356	40.457	ng/ul	99
74) Anthracene	17.310	178	1243762	39.905	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.081	216	404189	41.769	ng/ul	99
76) Pentachlorobenzene	14.651	250	355709	39.701	ng/ul	99
77) Carbazole	17.604	167	1109438	40.425	ng/ul	98
78) Di-n-butylphthalate	18.204	149	1344009	40.498	ng/ul	99
80) Fluoranthene	19.327	202	1551938	39.843	ng/ul	97
82) Pyrene	19.704	202	1590055	39.176	ng/ul	99
83) Butylbenzylphthalate	20.669	149	666849	40.846	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.569	252	456313	41.311	ng/ul	98
85) Benzo(a)anthracene	21.639	228	1629561	39.973	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.592	149	955912	39.825	ng/ul	99
87) Chrysene	21.716	228	1505604	39.924	ng/ul	100
89) Di-n-octyl phthalate	22.904	149	1610076	40.246	ng/ul	100
90) Benzo(b)fluoranthene	23.974	252	1575876	40.799	ng/ul	99
91) Benzo(k)fluoranthene	24.051	252	1536876	40.625	ng/ul	100
93) Benzo(a)pyrene	24.892	252	1473570	40.813	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.915	276	1728304	39.998	ng/ul	98
95) Dibenzo(a,h)anthracene	28.986	278	802322	22.613	ng/ul	96
96) Benzo(g,h,i)perylene	30.074	276	474266	13.987	ng/ul	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042324\
 Data File : BP020101.D
 Acq On : 23 Apr 2024 12:14
 Operator : MA/JU
 Sample : SST04097
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040619

Quant Time: Apr 23 23:57:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Apr 23 14:20:59 2024
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 15.277 ng/uL
 RT: 3.264 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BP020101.D
 Acq: 23 Apr 2024 12:14




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040619


Tgt Ion: 88 Resp: 56965
 Ion Ratio Lower Upper
 88 100
 43 31.1 24.0 36.0
 58 76.1 59.1 88.7



#5
 Pyridine
 Concen: 40.332 ng/uL
 RT: 3.646 min Scan# 95
 Delta R.T. -0.006 min
 Lab File: BP020101.D
 Acq: 23 Apr 2024 12:14

Tgt Ion: 79 Resp: 400005
 Ion Ratio Lower Upper
 79 100
 52 75.7 57.9 86.9
 51 38.2 29.6 44.4









#6
Benzaldehyde
Concen: 45.832 ng/ul
RT: 6.811 min Scan# 611
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion: 77 Resp: 317995
Ion Ratio Lower Upper
77 100
105 89.8 67.9 101.9
106 83.9 63.9 95.9



#8
Phenol
Concen: 41.268 ng/ul
RT: 6.858 min Scan# 641
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion: 94 Resp: 530613
Ion Ratio Lower Upper
94 100
65 32.2 25.5 38.3
66 45.0 36.9 55.3









#10
Bis(2-Chloroethyl)ether
Concen: 41.471 ng/ul
RT: 7.087 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion: 93 Resp: 425643
Ion Ratio Lower Upper
93 100
63 80.2 64.7 97.1
95 32.9 26.2 39.4



#12
2-Chlorophenol
Concen: 41.604 ng/ul
RT: 7.210 min Scan# 701
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion: 128 Resp: 383769
Ion Ratio Lower Upper
128 100
64 57.5 47.3 70.9
130 33.0 26.1 39.1









#13
2-Methylphenol
Concen: 40.591 ng/ul
RT: 8.081 min Scan# 849
Delta R.T. 0.000 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:108 Resp: 378872
Ion Ratio Lower Upper
108 100
107 90.6 72.8 109.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 41.176 ng/ul
RT: 8.163 min Scan# 863
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion: 45 Resp: 587732
Ion Ratio Lower Upper
45 100
77 16.3 13.4 20.2
79 11.3 10.7 16.1









#16
Acetophenone
Concen: 40.297 ng/ul
RT: 8.469 min Scan# 914
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:105 Resp: 625740
Ion Ratio Lower Upper
105 100
77 83.7 69.8 104.8
51 28.7 25.0 37.4



#17
N-Nitroso-di-n-propylamine
Concen: 40.728 ng/ul
RT: 8.452 min Scan# 912
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion: 70 Resp: 358835
Ion Ratio Lower Upper
70 100
42 55.8 45.3 67.9
101 9.0 7.5 11.3
130 17.2 14.0 21.0









#18
4-Methylphenol
Concen: 40.308 ng/ul
RT: 8.416 min Scan# 904
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:108 Resp: 400007
Ion Ratio Lower Upper
108 100
107 111.0 90.5 135.7



#19
Hexachloroethane
Concen: 39.859 ng/ul
RT: 8.693 min Scan# 953
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:117 Resp: 180031
Ion Ratio Lower Upper
117 100
201 109.8 85.2 127.8
199 67.6 54.3 81.5









#22
Nitrobenzene
Concen: 43.013 ng/ul
RT: 8.846 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion: 77 Resp: 526585
Ion Ratio Lower Upper
77 100
123 37.5 27.8 41.6
65 14.3 11.5 17.3



#23
Isophorone
Concen: 41.904 ng/ul
RT: 9.363 min Scan# 1067
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion: 82 Resp: 926380
Ion Ratio Lower Upper
82 100
95 7.6 6.5 9.7
138 14.5 11.4 17.2









#25
2-Nitrophenol
Concen: 42.913 ng/ul
RT: 9.546 min Scan# 1097
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:139 Resp: 206423
Ion Ratio Lower Upper
139 100
65 63.2 55.3 82.9
109 41.0 34.7 52.1



#26
2,4-Dimethylphenol
Concen: 41.446 ng/ul
RT: 9.604 min Scan# 1108
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:107 Resp: 434528
Ion Ratio Lower Upper
107 100
121 45.9 35.0 52.4
122 75.0 61.0 91.4





#27

Bis(2-Chloroethoxy)methane

Concen: 42.586 ng/ul

RT: 9.846 min Scan# 1148

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion: 93 Resp: 538301

Ion Ratio Lower Upper

93 100

95 32.9 26.2 39.4

123 10.1 8.2 12.4



#29

2,4-Dichlorophenol

Concen: 41.756 ng/ul

RT: 10.075 min Scan# 1188

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:162 Resp: 344862


Ion Ratio Lower Upper

162 100

164 67.0 50.5 75.7

98 38.8 31.8 47.6









#30
Naphthalene
Concen: 40.063 ng/ul
RT: 10.463 min Scan# 1253
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:128 Resp: 1123002
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.6
127 13.2 10.3 15.5



#32
4-Chloroaniline
Concen: 40.614 ng/ul
RT: 10.587 min Scan# 1275
Delta R.T. 0.000 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:127 Resp: 424956
Ion Ratio Lower Upper
127 100
129 32.8 25.0 37.4









#33
Hexachlorobutadiene
Concen: 41.880 ng/ul
RT: 10.728 min Scan# 1298
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:225 Resp: 298118
Ion Ratio Lower Upper
225 100
223 63.5 50.8 76.2
227 63.8 50.7 76.1



#34
Caprolactam
Concen: 31.754 ng/ul
RT: 11.404 min Scan# 1414
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:113 Resp: 67382
Ion Ratio Lower Upper
113 100
55 184.2 140.2 210.4
56 140.7 116.2 174.4









#35
4-Chloro-3-methylphenol
Concen: 37.626 ng/ul
RT: 11.728 min Scan# 1469
Delta R.T. 0.000 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:107 Resp: 354627
Ion Ratio Lower Upper
107 100
144 22.8 17.9 26.9
142 71.5 56.2 84.4



#36
2-Methylnaphthalene
Concen: 37.172 ng/ul
RT: 12.093 min Scan# 1531
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:142 Resp: 689338
Ion Ratio Lower Upper
142 100
141 90.8 71.4 107.2





#37

1-Methylnaphthalene

Concen: 36.699 ng/ul

RT: 12.316 min Scan# 1568

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:142 Resp: 677391

Ion Ratio Lower Upper

142 100

141 93.5 73.2 109.8

116 4.5 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 42.556 ng/ul

RT: 12.463 min Scan# 1594

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Tgt Ion:216 Resp: 423998




Ion Ratio Lower Upper


216 100

214 77.5 64.5 96.7

179 20.9 17.8 26.6

108 16.8 14.3 21.5





#40

Hexachlorocyclopentadiene

Concen: 40.030 ng/ul

RT: 12.422 min Scan# 1587

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:237 Resp: 144020

Ion Ratio Lower Upper

237 100

235 61.6 47.6 71.4

272 12.9 9.3 13.9



#41

2,4,6-Trichlorophenol

Concen: 41.668 ng/ul

RT: 12.722 min Scan# 1638

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:196 Resp: 248912


Ion Ratio Lower Upper

196 100

198 97.8 74.3 111.5

200 31.2 24.3 36.5





#42

2,4,5-Trichlorophenol

Concen: 42.020 ng/ul

RT: 12.798 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:196 Resp: 260358

Ion Ratio Lower Upper

196 100

198 95.5 81.1 121.7

200 30.3 26.0 39.0



#43

1,1'-Biphenyl

Concen: 41.795 ng/ul

RT: 13.122 min Scan# 1706

Delta R.T. -0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:154 Resp: 861181


Ion Ratio Lower Upper

154 100

153 40.5 31.9 47.9

76 15.2 10.9 16.3





#44

2-Chloronaphthalene

Concen: 41.560 ng/ul

RT: 13.169 min Scan# 1713

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:162 Resp: 693597

Ion Ratio Lower Upper

162 100

164 33.1 26.6 39.8

127 39.3 32.1 48.1



#45

2-Nitroaniline

Concen: 42.046 ng/ul

RT: 13.398 min Scan# 1753

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion: 65 Resp: 226883


Ion Ratio Lower Upper

65 100

92 64.7 51.3 76.9

138 88.5 64.8 97.2









#47
Dimethylphthalate
Concen: 40.517 ng/ul
RT: 13.781 min Scan# 1817
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:163 Resp: 820313
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 40.725 ng/ul
RT: 13.910 min Scan# 1840
Delta R.T. 0.000 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:165 Resp: 164827
Ion Ratio Lower Upper
165 100
89 74.2 56.1 84.1
121 21.9 18.1 27.1





#50

Acenaphthylene

Concen: 39.857 ng/ul

RT: 14.034 min Scan# 1861

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:152 Resp: 1041130

Ion Ratio Lower Upper

152 100

151 20.0 15.7 23.5

153 13.3 10.3 15.5



#51

3-Nitroaniline

Concen: 42.000 ng/ul

RT: 14.251 min Scan# 1898

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:138 Resp: 167726


Ion Ratio Lower Upper

138 100

108 12.2 9.8 14.6

92 139.3 105.1 157.7





#52

Acenaphthene

Concen: 39.561 ng/ul

RT: 14.387 min Scan# 1919

Delta R.T. 0.012 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNP_P

ClientSampleId :

SSTD040619





Tgt Ion:153 Resp: 680315

Ion Ratio Lower Upper

153 100

152 47.9 36.9 55.3

154 89.4 71.6 107.4



#53

2,4-Dinitrophenol

Concen: 38.977 ng/ul

RT: 14.475 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:184 Resp: 67044


Ion Ratio Lower Upper

184 100

63 76.6 53.3 79.9

154 67.0 51.8 77.6









#55
4-Nitrophenol
Concen: 43.113 ng/ul
RT: 14.587 min Scan# 1955
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:	109	Resp:	133765
Ion	Ratio	Lower	Upper
109	100		
139	99.7	81.4	122.2
65	123.0	104.3	156.5



#56
Dibenzofuran
Concen: 38.862 ng/ul
RT: 14.734 min Scan# 1980
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:	168	Resp:	924905
Ion	Ratio	Lower	Upper
168	100		
139	39.2	30.6	46.0





#57

2,4-Dinitrotoluene

Concen: 40.447 ng/ul

RT: 14.728 min Scan# 1978

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:165 Resp: 232164

Ion Ratio Lower Upper

165 100

63 67.3 63.2 94.8

182 2.7 2.2 3.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 39.264 ng/ul

RT: 14.975 min Scan# 2021

Delta R.T. 0.018 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Tgt Ion:232 Resp: 217061




Ion Ratio Lower Upper


232 100

131 43.6 33.2 49.8

130 2.3 2.2 3.2

166 30.3 25.2 37.8









#59
Diethylphthalate
Concen: 38.266 ng/ul
RT: 15.187 min Scan# 2055
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:149 Resp: 762048
Ion Ratio Lower Upper
149 100
177 23.0 18.3 27.5
150 12.9 10.0 15.0



#61
Fluorene
Concen: 38.521 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:166 Resp: 753608
Ion Ratio Lower Upper
166 100
165 98.7 78.0 117.0
167 13.5 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 38.533 ng/ul
RT: 15.404 min Scan# 2094
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:204 Resp: 409653
Ion Ratio Lower Upper
204 100
206 33.5 27.4 41.2
141 59.8 48.8 73.2



#63
4-Nitroaniline
Concen: 43.051 ng/ul
RT: 15.451 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:138 Resp: 158580
Ion Ratio Lower Upper
138 100
92 65.5 49.8 74.8
108 112.3 85.4 128.0









#66
4,6-Dinitro-2-methylphenol
Concen: 38.231 ng/ul
RT: 15.516 min Scan# 2113
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:198 Resp: 114907
Ion Ratio Lower Upper
198 100
182 2.9 3.0 4.6#
77 26.0 24.9 37.3



#67
N-Nitrosodiphenylamine
Concen: 39.404 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:169 Resp: 621455
Ion Ratio Lower Upper
169 100
168 67.2 53.0 79.4
167 36.3 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 39.805 ng/ul
RT: 16.334 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:248 Resp: 259630
Ion Ratio Lower Upper
248 100
250 97.1 77.8 116.6
141 73.7 61.4 92.0



#69
Hexachlorobenzene
Concen: 40.590 ng/ul
RT: 16.434 min Scan# 2269
Delta R.T. 0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:284 Resp: 299779
Ion Ratio Lower Upper
284 100
142 33.4 26.2 39.2
249 33.2 27.0 40.4





#70

Atrazine

Concen: 39.645 ng/ul

RT: 16.639 min Scan# 21

Delta R.T. 0.018 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:200 Resp: 242348

Ion Ratio Lower Upper

200 100

173 23.5 19.1 28.7

215 47.5 38.1 57.1



#71

Pentachlorophenol

Concen: 39.920 ng/ul

RT: 16.804 min Scan# 2332

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:266 Resp: 184157


Ion Ratio Lower Upper

266 100

264 63.0 52.8 79.2

268 65.4 51.6 77.4





#72

Phenanthrene

Concen: 40.457 ng/ul

RT: 17.210 min Scan# 2418

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:178 Resp: 1241356

Ion Ratio Lower Upper

178 100

179 15.4 12.6 18.8

176 18.9 15.0 22.4



#74

Anthracene

Concen: 39.905 ng/ul

RT: 17.310 min Scan# 2418

Delta R.T. 0.012 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:178 Resp: 1243762


Ion Ratio Lower Upper

178 100

179 15.7 12.6 18.8

176 18.5 14.6 21.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 41.769 ng/uL

RT: 13.081 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:216 Resp: 404189

Ion Ratio Lower Upper

216 100

214 79.3 63.0 94.6

218 48.9 38.8 58.2



#76

Pentachlorobenzene

Concen: 39.701 ng/uL

RT: 14.651 min Scan# 1966

Delta R.T. 0.018 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:250 Resp: 355709


Ion Ratio Lower Upper

250 100

248 62.2 49.5 74.3

252 64.0 50.5 75.7





#77

Carbazole

Concen: 40.425 ng/ul

RT: 17.604 min Scan# 2465

Delta R.T. 0.018 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:167 Resp: 1109438

Ion Ratio Lower Upper

167 100

166 21.4 18.0 27.0

139 13.2 10.6 16.0



#78

Di-n-butylphthalate

Concen: 40.498 ng/ul

RT: 18.204 min Scan# 2570

Delta R.T. 0.018 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:149 Resp: 1344009


Ion Ratio Lower Upper

149 100

150 9.1 7.6 11.4

104 6.1 4.9 7.3





#80

Fluoranthene

Concen: 39.843 ng/ul

RT: 19.327 min Scan# 21

Delta R.T. 0.012 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:202 Resp: 1551938

Ion Ratio Lower Upper

202 100

101 8.7 7.7 11.5

100 6.5 5.9 8.9



#82

Pyrene

Concen: 39.176 ng/ul

RT: 19.704 min Scan# 2825

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:202 Resp: 1590055


Ion Ratio Lower Upper

202 100

101 10.8 9.1 13.7

100 8.8 7.3 10.9





#83

Butylbenzylphthalate

Concen: 40.846 ng/ul

RT: 20.669 min Scan# 2990

Delta R.T. -0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:149 Resp: 666849

Ion Ratio Lower Upper

149 100

91 78.7 63.3 94.9

206 21.3 18.4 27.6



#84

3,3'-Dichlorobenzidine

Concen: 41.311 ng/ul

RT: 21.569 min Scan# 3142

Delta R.T. -0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:252 Resp: 456313


Ion Ratio Lower Upper

252 100

254 64.9 50.6 75.8

126 7.9 7.8 11.6





#85

Benzo(a)anthracene

Concen: 39.973 ng/ul

RT: 21.639 min Scan# 3154

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:228 Resp: 1629561

Ion Ratio Lower Upper

228 100

229 20.0 15.8 23.6

226 27.0 21.8 32.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 39.825 ng/ul

RT: 21.592 min Scan# 3146

Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:149 Resp: 955912


Ion Ratio Lower Upper

149 100

167 27.6 21.8 32.8

279 5.0 3.8 5.8





#87

Chrysene

Concen: 39.924 ng/ul

RT: 21.716 min Scan# 3166

Delta R.T. 0.006 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:228 Resp: 1505604

Ion Ratio Lower Upper

228 100

226 29.8 24.2 36.2

229 19.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 40.246 ng/ul




RT: 22.904 min Scan# 3369


Delta R.T. 0.000 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Tgt Ion:149 Resp: 1610076









#90
Benzo(b)fluoranthene
Concen: 40.799 ng/ul
RT: 23.974 min Scan# 3551
Delta R.T. -0.006 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14




Instrument :
BNA_P
ClientSampleId :
SSTD040619


Tgt Ion:252 Resp: 1575876
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 7.3 5.7 8.5



#91
Benzo(k)fluoranthene
Concen: 40.625 ng/ul
RT: 24.051 min Scan# 3564
Delta R.T. 0.000 min
Lab File: BP020101.D
Acq: 23 Apr 2024 12:14

Tgt Ion:252 Resp: 1536876
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 7.7 6.2 9.4





#93

Benzo(a)pyrene

Concen: 40.813 ng/ul

RT: 24.892 min Scan# 31

Delta R.T. -0.012 min

Lab File: BP020101.D


Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619






Tgt Ion:252 Resp: 1473570

Ion Ratio Lower Upper

252 100

253 21.6 16.8 25.2

125 8.3 6.5 9.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 39.998 ng/ul

RT: 28.915 min Scan# 4391

Delta R.T. 0.012 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14




Tgt Ion:276 Resp: 1728304


Ion Ratio Lower Upper

276 100

138 14.8 11.7 17.5

277 24.1 18.4 27.6





#95

Dibenzo(a,h)anthracene

Concen: 22.613 ng/ul

RT: 28.986 min Scan# 4406

Delta R.T. -0.018 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Instrument :

BNA_P

ClientSampleId :

SSTD040619





Tgt Ion:278 Resp: 802322

Ion Ratio Lower Upper

278 100

139 12.0 8.8 13.2

279 23.2 20.3 30.5



#96

Benzo(g,h,i)perylene

Concen: 13.987 ng/ul

RT: 30.074 min Scan# 4588

Delta R.T. -0.041 min

Lab File: BP020101.D

Acq: 23 Apr 2024 12:14

Tgt Ion:276 Resp: 474266

Ion Ratio Lower Upper

276 100

138 15.2 14.2 21.4

277 25.1 18.7 28.1

