

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042425\
 Data File : BP024408.D
 Acq On : 24 Apr 2025 12:31
 Operator : RC/JU
 Sample : PB167711BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB167711BL

Quant Time: Apr 24 12:56:17 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 18 12:04:48 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.722	152	288156	20.000	ng	0.00
21) Naphthalene-d8	10.486	136	1116360	20.000	ng	-0.01
39) Acenaphthene-d10	14.339	164	654539	20.000	ng	-0.01
64) Phenanthrene-d10	17.133	188	1181165	20.000	ng	-0.01
76) Chrysene-d12	21.574	240	1074963	20.000	ng	-0.02
86) Perylene-d12	24.915	264	1210803	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	2338269	134.171	ng	0.00
7) Phenol-d6	6.910	99	2874772	120.469	ng	0.00
23) Nitrobenzene-d5	8.869	82	1792266	91.545	ng	0.00
42) 2,4,6-Tribromophenol	15.857	330	1257538	138.864	ng	0.00
45) 2-Fluorobiphenyl	12.957	172	3941973	92.211	ng	0.00
79) Terphenyl-d14	19.857	244	5252367	98.486	ng	-0.02


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042425\
Data File : BP024408.D
Acq On : 24 Apr 2025 12:31
Operator : RC/JU
Sample : PB167711BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB167711BL

Quant Time: Apr 24 12:56:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP041425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 18 12:04:48 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.722 min Scan# 71
Delta R.T. -0.000 min
Lab File: BP024408.D
Acq: 24 Apr 2025 12:31




Instrument :
BNA_P
ClientSampleId :
PB167711BL


Tgt Ion:152 Resp: 288156
Ion Ratio Lower Upper
152 100
150 154.7 124.9 187.3
115 59.9 47.7 71.5



#5
2-Fluorophenol
Concen: 134.171 ng
RT: 5.346 min Scan# 384
Delta R.T. -0.000 min
Lab File: BP024408.D
Acq: 24 Apr 2025 12:31

Tgt Ion:112 Resp: 2338269
Ion Ratio Lower Upper
112 100
64 53.3 41.1 61.7
63 27.7 20.6 30.8









#7
Phenol-d6
Concen: 120.469 ng
RT: 6.910 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP024408.D
Acq: 24 Apr 2025 12:31




Instrument :
BNA_P
ClientSampleId :
PB167711BL


Tgt Ion: 99 Resp: 2874772
Ion Ratio Lower Upper
99 100
42 14.8 10.6 16.0
71 33.1 23.7 35.5



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.486 min Scan# 1258
Delta R.T. -0.012 min
Lab File: BP024408.D
Acq: 24 Apr 2025 12:31

Tgt Ion:136 Resp: 1116360
Ion Ratio Lower Upper
136 100
137 10.8 9.2 13.8
54 6.8 6.0 9.0
68 5.7 5.5 8.3









#23
 Nitrobenzene-d5
 Concen: 91.545 ng
 RT: 8.869 min Scan# 913
 Delta R.T. -0.000 min
 Lab File: BP024408.D
 Acq: 24 Apr 2025 12:31




Instrument :
 BNA_P
 ClientSampleId :
 PB167711BL


Tgt Ion: 82 Resp: 1792266
 Ion Ratio Lower Upper
 82 100
 128 47.2 36.5 54.7
 54 44.0 33.4 50.2



#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 14.339 min Scan# 1913
 Delta R.T. -0.012 min
 Lab File: BP024408.D
 Acq: 24 Apr 2025 12:31

Tgt Ion:164 Resp: 654539
 Ion Ratio Lower Upper
 164 100
 162 99.3 80.6 120.8
 160 44.6 35.3 52.9





#42

2,4,6-Tribromophenol

Concen: 138.864 ng

RT: 15.857 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP024408.D

Acq: 24 Apr 2025 12:31

Instrument :

BNA_P

ClientSampleId :

PB167711BL





Tgt Ion:330 Resp: 1257538

Ion Ratio Lower Upper

330 100

332 96.7 77.1 115.7

141 29.7 24.7 37.1



#45

2-Fluorobiphenyl

Concen: 92.211 ng

RT: 12.957 min Scan# 1678

Delta R.T. -0.006 min

Lab File: BP024408.D

Acq: 24 Apr 2025 12:31




Tgt Ion:172 Resp: 3941973


Ion Ratio Lower Upper

172 100

171 35.5 28.2 42.2

170 23.4 18.6 27.8









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.133 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP024408.D
Acq: 24 Apr 2025 12:31




Instrument :
BNA_P
ClientSampleId :
PB167711BL


Tgt Ion:188 Resp: 1181165
Ion Ratio Lower Upper
188 100
94 10.4 8.6 13.0
80 11.6 9.8 14.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.574 min Scan# 3143
Delta R.T. -0.018 min
Lab File: BP024408.D
Acq: 24 Apr 2025 12:31

Tgt Ion:240 Resp: 1074963
Ion Ratio Lower Upper
240 100
120 12.3 9.8 14.8
236 25.8 20.9 31.3





#79

Terphenyl-d14

Concen: 98.486 ng

RT: 19.857 min Scan# 21

Delta R.T. -0.018 min

Lab File: BP024408.D

Acq: 24 Apr 2025 12:31

Instrument :

BNA_P

ClientSampleId :

PB167711BL





Tgt Ion:244 Resp: 5252367

Ion Ratio Lower Upper

244 100

212 7.1 5.6 8.4

122 10.5 9.4 14.0



#86

Perylene-d12

Concen: 20.000 ng

RT: 24.915 min Scan# 3711

Delta R.T. -0.018 min

Lab File: BP024408.D

Acq: 24 Apr 2025 12:31

Tgt Ion:264 Resp: 1210803

Ion Ratio Lower Upper

264 100

260 23.6 18.9 28.3

265 21.5 17.8 26.6

