

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042522\
 Data File : BP010093.D
 Acq On : 25 Apr 2022 21:02
 Operator : CG/JU
 Sample : N2475-11MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A0BM8MS

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 04/26/2022
 Supervised By :Yogesh Patel 04/28/2022

Quant Time: Apr 26 01:05:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	164120	20.000	ng/ul	0.00
20) Naphthalene-d8	10.740	136	700466	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.569	164	467274	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.322	188	969301	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.404	240	746533	20.000	ng/ul	# 0.00
88) Perylene-d12	23.851	264	559687	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	17831	4.738	ng/uL	0.00
4) Pyridine-d5	3.787	84	59335	5.604	ng/ul	0.00
7) Phenol-d5	7.093	99	96238	6.447	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.258	67	290586m	32.440	ng/ul	-0.01
11) 2-Chlorophenol-d4	7.463	132	275230	24.895	ng/ul	0.00
15) 4-Methylphenol-d8	8.646	113	190662	15.415	ng/ul	0.00
21) Nitrobenzene-d5	9.093	128	185359	33.271	ng/ul	0.00
24) 2-Nitrophenol-d4	9.816	143	201529	33.122	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.363	165	341372	28.820	ng/ul	0.00
31) 4-Chloroaniline-d4	10.869	131	421911	24.842	ng/ul	-0.01
46) Dimethylphthalate-d6	13.975	166	1338340	36.819	ng/ul	0.00
49) Acenaphthylene-d8	14.263	160	1449919	32.959	ng/ul	0.00
54) 4-Nitrophenol-d4	14.757	143	47259m	7.062	ng/ul	0.00
60) Fluorene-d10	15.563	176	1123240	35.924	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	239791	39.124	ng/ul	0.00
73) Anthracene-d10	17.422	188	1771124	37.769	ng/ul	0.00
81) Pyrene-d10	19.651	212	1962189	46.217	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.692	264	1191926	40.622	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	18622	4.761	ng/uL#	14
5) Pyridine	3.805	79	66919	6.080	ng/ul#	35
6) Benzaldehyde	7.069	77	325140	40.756	ng/ul	97
8) Phenol	7.122	94	111940	7.463	ng/ul#	92
10) Bis(2-Chloroethyl)ether	7.352	93	392034	33.924	ng/ul#	76
12) 2-Chlorophenol	7.499	128	296186	26.560	ng/ul	93
13) 2-Methylphenol	8.375	108	225990	19.645	ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.469	45	588971	34.347	ng/ul#	85
16) Acetophenone	8.757	105	633530	32.830	ng/ul#	82
17) N-Nitroso-di-n-propyla...	8.746	70	339013	33.531	ng/ul#	75
18) 4-Methylphenol	8.705	108	212268	16.759	ng/ul	90
19) Hexachloroethane	9.022	117	153816	32.825	ng/ul	85
22) Nitrobenzene	9.134	77	488324	35.776	ng/ul#	84
23) Isophorone	9.663	82	947276	35.137	ng/ul	97
25) 2-Nitrophenol	9.852	139	225161	34.941	ng/ul#	81
26) 2,4-Dimethylphenol	9.910	107	351621	25.203	ng/ul#	79
27) Bis(2-Chloroethoxy)met...	10.146	93	557489	35.280	ng/ul#	98
29) 2,4-Dichlorophenol	10.387	162	360746	31.683	ng/ul#	85
30) Naphthalene	10.793	128	1284427	35.034	ng/ul	97
32) 4-Chloroaniline	10.899	127	439928	26.286	ng/ul	99
33) Hexachlorobutadiene	11.087	225	280764	34.792	ng/ul	95
34) Caprolactam	11.675	113	17437m	4.257	ng/ul	
35) 4-Chloro-3-methylphenol	12.016	107	390390	28.382	ng/ul#	67
36) 2-Methylnaphthalene	12.398	142	923277	34.724	ng/ul	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042522\
 Data File : BP010093.D
 Acq On : 25 Apr 2022 21:02
 Operator : CG/JU
 Sample : N2475-11MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A0BM8MS


Quant Time: Apr 26 01:05:00 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Jagrut Upadhyay 04/26/2022
 Supervised By : Yogesh Patel 04/28/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	929099	34.414	ng/ul#	94
39) 1,2,4,5-Tetrachloroben...	12.769	216	536842	37.186	ng/ul	93
40) Hexachlorocyclopentadiene	12.751	237	254794	31.318	ng/ul	95
41) 2,4,6-Trichlorophenol	13.004	196	344112	36.840	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.069	196	374794	36.878	ng/ul#	87
43) 1,1'-Biphenyl	13.404	154	1289286	37.410	ng/ul	94
44) 2-Chloronaphthalene	13.445	162	985511	36.951	ng/ul	93
45) 2-Nitroaniline	13.645	65	340167	39.334	ng/ul#	81
47) Dimethylphthalate	14.022	163	1368927	38.751	ng/ul#	92
48) 2,6-Dinitrotoluene	14.140	165	294583	40.634	ng/ul	96
50) Acenaphthylene	14.287	152	1611570	36.983	ng/ul	95
51) 3-Nitroaniline	14.469	138	248072	37.774	ng/ul#	80
52) Acenaphthene	14.634	153	1055009	37.234	ng/ul	96
53) 2,4-Dinitrophenol	14.675	184	151236	37.332	ng/ul#	80
55) 4-Nitrophenol	14.775	109	46070	7.924	ng/ul#	53
56) Dibenzofuran	14.969	168	1555087	37.064	ng/ul	99
57) 2,4-Dinitrotoluene	14.928	165	424026	40.037	ng/ul#	85
58) 2,3,4,6-Tetrachlorophenol	15.192	232	347038m	38.415	ng/ul	
59) Diethylphthalate	15.387	149	1437342	40.002	ng/ul	94
61) Fluorene	15.616	166	1289382	37.429	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.610	204	684442	37.753	ng/ul	99
63) 4-Nitroaniline	15.628	138	262747m	41.117	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.692	198	247242	41.811	ng/ul#	92
67) N-Nitrosodiphenylamine	15.822	169	1142704	41.087	ng/ul	95
68) 4-Bromophenyl-phenylether	16.504	248	435263	42.099	ng/ul	95
69) Hexachlorobenzene	16.628	284	502025	41.833	ng/ul#	91
70) Atrazine	16.781	200	445663	39.415	ng/ul#	89
71) Pentachlorophenol	16.969	266	293954	38.462	ng/ul#	84
72) Phenanthrene	17.363	178	2125019	40.335	ng/ul	99
74) Anthracene	17.457	178	2153916	40.590	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.369	216	569601	40.252	ng/ul#	84
76) Pentachlorobenzene	14.892	250	559160	40.295	ng/ul	91
77) Carbazole	17.722	167	1846660	39.326	ng/ul#	96
78) Di-n-butylphthalate	18.269	149	2459180	42.568	ng/ul#	93
80) Fluoranthene	19.327	202	2439317	50.939	ng/ul#	95
82) Pyrene	19.680	202	2438130	49.418	ng/ul#	93
83) Butylbenzylphthalate	20.545	149	979227	48.016	ng/ul#	80
84) 3,3'-Dichlorobenzidine	21.316	252	486292	36.202	ng/ul#	95
85) Benzo(a)anthracene	21.386	228	1997866	42.169	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.310	149	1369227	45.300	ng/ul#	82
87) Chrysene	21.439	228	1928586	41.342	ng/ul	99
89) Di-n-octyl phthalate	22.251	149	2049612	47.107	ng/ul	100
90) Benzo(b)fluoranthene	23.104	252	1616012	42.949	ng/ul	100
91) Benzo(k)fluoranthene	23.157	252	1614961	45.311	ng/ul#	98
93) Benzo(a)pyrene	23.745	252	1500904	43.136	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.415	276	1559234	45.048	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.433	278	1319342	44.546	ng/ul#	96
96) Benzo(g,h,i)perylene	27.204	276	1349567	46.208	ng/ul	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 4.761 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. -0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion: 88 Resp: 1862
Ion Ratio Lower Upper
88 100
43 35.4 6.8 10.2
58 87.0 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#5
Pyridine
Concen: 6.080 ng/uL
RT: 3.805 min Scan# 122
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02



Tgt Ion: 79 Resp: 66919
Ion Ratio Lower Upper
79 100
52 93.9 35.5 53.3#
51 48.2 20.2 30.4#









#6
Benzaldehyde
Concen: 40.756 ng/ul
RT: 7.069 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Instrument :
BNA_P
ClientSampleId :
A0BM8MS

Tgt Ion: 77 Resp: 325140
Ion Ratio Lower Upper
77 100
105 94.3 77.0 115.4
106 92.0 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#8
Phenol
Concen: 7.463 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Tgt Ion: 94 Resp: 111940
Ion Ratio Lower Upper
94 100
65 32.0 19.0 28.6#
66 41.3 34.6 51.8





#10

Bis(2-Chloroethyl)ether

Concen: 33.924 ng/ul

RT: 7.352 min Scan# 70

Delta R.T. -0.006 min

Lab File: BP010093.D

Acq: 25 Apr 2022 21:02

Instrument :

BNA_P

ClientSampleId :

A0BM8MS





Tgt Ion: 93 Resp: 392034

Ion	Ratio	Lower	Upper
93	100		
63	83.1	45.0	67.4
95	31.2	26.3	39.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#12

2-Chlorophenol

Concen: 26.560 ng/ul

RT: 7.499 min Scan# 750




Delta R.T. -0.006 min


Lab File: BP010093.D

Acq: 25 Apr 2022 21:02

Tgt Ion:128 Resp: 296186


Ion	Ratio	Lower	Upper
128	100		
64	50.7	45.3	67.9
130	33.0	29.0	43.6





#13
2-Methylphenol
Concen: 19.645 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:108 Resp: 225990
Ion Ratio Lower Upper
108 100
107 92.5 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 34.347 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02



Tgt Ion: 45 Resp: 588971
Ion Ratio Lower Upper
45 100
77 13.6 17.1 25.7#
79 10.7 13.2 19.8#





#16
Acetophenone
Concen: 32.830 ng/ul
RT: 8.757 min Scan# 964
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Instrument :

BNA_P

ClientSampleId :





A0BM8MS

Tgt Ion: 105 Resp: 633530
Ion Ratio Lower Upper
105 100
77 79.2 53.4 80.2
51 30.1 13.8 20.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#17
N-Nitroso-di-n-propylamine
Concen: 33.531 ng/ul
RT: 8.746 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion: 70 Resp: 339013
Ion Ratio Lower Upper
70 100
42 61.3 31.8 47.8#
101 10.2 15.6 23.4#
130 21.0 13.7 20.5#





#18
4-Methylphenol
Concen: 16.759 ng/ul
RT: 8.705 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:108 Resp: 21226
Ion Ratio Lower Upper
108 100
107 114.3 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#19
Hexachloroethane
Concen: 32.825 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:117 Resp: 153816
Ion Ratio Lower Upper
117 100
201 109.6 106.5 159.7
199 67.9 59.0 88.6





#22
Nitrobenzene
Concen: 35.776 ng/ul
RT: 9.134 min Scan# 1028
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion: 77 Resp: 488324
Ion Ratio Lower Upper
77 100
123 42.3 42.3 63.5
65 14.5 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#23
Isophorone
Concen: 35.137 ng/ul
RT: 9.663 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion: 82 Resp: 947276
Ion Ratio Lower Upper
82 100
95 8.1 7.8 11.8
138 17.2 12.9 19.3





#25
2-Nitrophenol
Concen: 34.941 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:139 Resp: 22516
Ion Ratio Lower Upper
139 100
65 56.0 49.4 74.2
109 33.6 44.7 67.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#26
2,4-Dimethylphenol
Concen: 25.203 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:107 Resp: 351621
Ion Ratio Lower Upper
107 100
121 48.0 34.1 51.1
122 80.9 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 35.280 ng/ul
RT: 10.146 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion: 93 Resp: 557489
Ion Ratio Lower Upper
93 100
95 32.9 26.2 39.4
123 11.4 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#29
2,4-Dichlorophenol
Concen: 31.683 ng/ul
RT: 10.387 min Scan# 1241
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:162 Resp: 360746
Ion Ratio Lower Upper
162 100
164 64.1 37.0 55.6#
98 39.0 31.0 46.6





#30
Naphthalene
Concen: 35.034 ng/ul
RT: 10.793 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:128 Resp: 128442
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.3 12.2 18.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#32
4-Chloroaniline
Concen: 26.286 ng/ul
RT: 10.899 min Scan# 1328
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:127 Resp: 439928
Ion Ratio Lower Upper
127 100
129 31.2 24.6 37.0





#33
Hexachlorobutadiene
Concen: 34.792 ng/ul
RT: 11.087 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:225 Resp: 280764
Ion Ratio Lower Upper
225 100
223 61.9 53.4 80.0
227 63.5 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#34
Caprolactam
Concen: 4.257 ng/ul m
RT: 11.675 min Scan# 1460
Delta R.T. 0.012 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02



Tgt Ion:113 Resp: 17437
Ion Ratio Lower Upper
113 100
55 168.9 58.2 87.2#
56 148.9 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 28.382 ng/ul
RT: 12.016 min Scan# 1519
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Instrument :
BNA_P
ClientSampleId :
A0BM8MS

Tgt Ion:107 Resp: 390390
Ion Ratio Lower Upper
107 100
144 24.4 15.8 23.8
142 77.4 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#36
2-Methylnaphthalene
Concen: 34.724 ng/ul
RT: 12.398 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:142 Resp: 923277
Ion Ratio Lower Upper
142 100
141 87.9 64.3 96.5





#37
1-Methylnaphthalene
Concen: 34.414 ng/ul
RT: 12.616 min Scan# 1620
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:142 Resp: 929099
Ion Ratio Lower Upper
142 100
141 92.8 69.8 104.8
116 4.1 5.6 8.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 37.186 ng/ul
RT: 12.769 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:216 Resp: 536842
Ion Ratio Lower Upper
216 100
214 79.0 58.2 87.2
179 20.8 14.2 21.2
108 16.3 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 31.318 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:237 Resp: 254794
Ion Ratio Lower Upper
237 100
235 62.3 53.2 79.8
272 13.1 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#41
2,4,6-Trichlorophenol
Concen: 36.840 ng/ul
RT: 13.004 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Tgt Ion:196 Resp: 344112
Ion Ratio Lower Upper
196 100
198 95.6 64.1 96.1
200 31.5 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 36.878 ng/ul
RT: 13.069 min Scan# 1698
Delta R.T. -0.012 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Instrument :
BNA_P
ClientSampleId :
A0BM8MS

Tgt Ion:196 Resp: 374794
Ion Ratio Lower Upper
196 100
198 95.5 68.6 103.0
200 30.3 32.8 49.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#43
1,1'-Biphenyl
Concen: 37.410 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:154 Resp: 1289286
Ion Ratio Lower Upper
154 100
153 39.9 28.6 43.0
76 12.0 8.5 12.7





#44
2-Chloronaphthalene
Concen: 36.951 ng/ul
RT: 13.445 min Scan# 1761
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:162 Resp: 98551
Ion Ratio Lower Upper
162 100
164 32.5 27.5 41.3
127 40.1 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#45
2-Nitroaniline
Concen: 39.334 ng/ul
RT: 13.645 min Scan# 1795
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion: 65 Resp: 340167
Ion Ratio Lower Upper
65 100
92 64.3 77.9 116.9#
138 95.8 81.5 122.3





#47
Dimethylphthalate
Concen: 38.751 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:163 Resp: 136892
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 10.3 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#48
2,6-Dinitrotoluene
Concen: 40.634 ng/ul
RT: 14.140 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:165 Resp: 294583
Ion Ratio Lower Upper
165 100
89 60.9 51.7 77.5
121 21.6 18.3 27.5





#50
Acenaphthylene
Concen: 36.983 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:152 Resp: 1611570
Ion Ratio Lower Upper
152 100
151 19.3 18.3 27.5
153 13.2 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#51
3-Nitroaniline
Concen: 37.774 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:138 Resp: 248072
Ion Ratio Lower Upper
138 100
108 12.2 15.8 23.8#
92 121.5 117.1 175.7





#52
Acenaphthene
Concen: 37.234 ng/ul
RT: 14.634 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:153 Resp: 1055009
Ion Ratio Lower Upper
153 100
152 46.4 34.3 51.5
154 90.4 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#53
2,4-Dinitrophenol
Concen: 37.332 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:184 Resp: 151236
Ion Ratio Lower Upper
184 100
63 69.0 34.6 52.0#
154 65.9 57.7 86.5





#55
4-Nitrophenol
Concen: 7.924 ng/ul
RT: 14.775 min Scan# 1987
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:109 Resp: 46070
Ion Ratio Lower Upper
109 100
139 114.3 61.4 92.2
65 127.0 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#56
Dibenzofuran
Concen: 37.064 ng/ul
RT: 14.969 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Tgt Ion:168 Resp: 1555087
Ion Ratio Lower Upper
168 100
139 38.0 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 40.037 ng/ul
RT: 14.928 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Instrument :
BNA_P
ClientSampleId :
A0BM8MS

Tgt Ion:165 Resp: 424020
Ion Ratio Lower Upper
165 100
63 46.8 29.5 44.3#
182 3.4 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 38.415 ng/ul m
RT: 15.192 min Scan# 2058
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:232 Resp: 347038
Ion Ratio Lower Upper
232 100
131 43.9 25.4 38.2#
130 2.1 4.2 6.2#
166 29.6 28.1 42.1





#59
Diethylphthalate
Concen: 40.002 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:149 Resp: 143734
Ion Ratio Lower Upper
149 100
177 22.7 22.0 33.0
150 13.0 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#61
Fluorene
Concen: 37.429 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:166 Resp: 1289382
Ion Ratio Lower Upper
166 100
165 98.5 70.3 105.5
167 13.4 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 37.753 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:204 Resp: 68444
Ion Ratio Lower Upper
204 100
206 33.6 27.6 41.4
141 58.8 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#63
4-Nitroaniline
Concen: 41.117 ng/ul m
RT: 15.628 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:138 Resp: 262747
Ion Ratio Lower Upper
138 100
92 54.0 41.6 62.4
108 85.4 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 41.811 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:198 Resp: 247241
Ion Ratio Lower Upper
198 100
182 3.7 6.0 9.0
77 25.2 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#67
N-Nitrosodiphenylamine
Concen: 41.087 ng/ul
RT: 15.822 min Scan# 2165
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:169 Resp: 1142704
Ion Ratio Lower Upper
169 100
168 65.6 48.3 72.5
167 35.2 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 42.099 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:248 Resp: 43526
Ion Ratio Lower Upper
248 100
250 98.0 79.4 119.0
141 75.3 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#69
Hexachlorobenzene
Concen: 41.833 ng/ul
RT: 16.628 min Scan# 2302
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:284 Resp: 502025
Ion Ratio Lower Upper
284 100
142 33.6 19.2 28.8#
249 33.0 27.4 41.2





#70
Atrazine
Concen: 39.415 ng/ul
RT: 16.781 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:200 Resp: 44566
Ion Ratio Lower Upper
200 100
173 25.2 19.2 28.8
215 49.9 32.2 48.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#71
Pentachlorophenol
Concen: 38.462 ng/ul
RT: 16.969 min Scan# 2360
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:266 Resp: 293954
Ion Ratio Lower Upper
266 100
264 62.2 57.3 85.9
268 64.1 65.7 98.5





#72
Phenanthrene
Concen: 40.335 ng/u1
RT: 17.363 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:178 Resp: 2125019
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 18.0 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#74
Anthracene
Concen: 40.590 ng/u1
RT: 17.457 min Scan# 2443
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Tgt Ion:178 Resp: 2153916
Ion Ratio Lower Upper
178 100
179 15.1 12.3 18.5
176 17.9 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 40.252 ng/uL
RT: 13.369 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Instrument :
BNA_P
ClientSampleId :
A0BM8MS

Tgt Ion:216 Resp: 56960
Ion Ratio Lower Upper
216 100
214 80.2 50.4 75.6
218 49.0 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#76
Pentachlorobenzene
Concen: 40.295 ng/uL
RT: 14.892 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:250 Resp: 559160
Ion Ratio Lower Upper
250 100
248 62.5 43.5 65.3
252 63.2 46.5 69.7





#77
 Carbazole
 Concen: 39.326 ng/ul
 RT: 17.722 min Scan# 2488
 Delta R.T. -0.000 min
 Lab File: BP010093.D
 Acq: 25 Apr 2022 21:02




Instrument :
 BNA_P
 ClientSampleId :
 A0BM8MS




Tgt Ion:167 Resp: 1846660
 Ion Ratio Lower Upper
 167 100
 166 21.0 15.8 23.8
 139 13.2 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
 Supervised By :Yogesh Patel 04/28/2022




#78
 Di-n-butylphthalate
 Concen: 42.568 ng/ul
 RT: 18.269 min Scan# 2581
 Delta R.T. -0.006 min
 Lab File: BP010093.D
 Acq: 25 Apr 2022 21:02




Tgt Ion:149 Resp: 2459180
 Ion Ratio Lower Upper
 149 100
 150 9.4 9.6 14.4#
 104 5.3 6.2 9.4#





#80
Fluoranthene
Concen: 50.939 ng/ul
RT: 19.327 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:202 Resp: 243931
Ion Ratio Lower Upper
202 100
101 9.3 9.6 14.4
100 7.3 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#82
Pyrene
Concen: 49.418 ng/ul
RT: 19.680 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:202 Resp: 2438130
Ion Ratio Lower Upper
202 100
101 11.0 10.6 15.8
100 9.2 9.8 14.6





#83
Butylbenzylphthalate
Concen: 48.016 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:149 Resp: 979227
Ion Ratio Lower Upper
149 100
91 72.0 71.1 106.7
206 22.2 28.8 43.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#84
3,3'-Dichlorobenzidine
Concen: 36.202 ng/ul
RT: 21.316 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:252 Resp: 486292
Ion Ratio Lower Upper
252 100
254 65.2 49.8 74.6
126 9.8 4.3 6.5





#85
Benzo(a)anthracene
Concen: 42.169 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:228 Resp: 1997860
Ion Ratio Lower Upper
228 100
229 19.7 18.2 27.2
226 26.1 22.8 34.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 45.300 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:149 Resp: 1369227
Ion Ratio Lower Upper
149 100
167 28.6 32.8 49.2#
279 5.1 5.6 8.4#





#87
Chrysene
Concen: 41.342 ng/ul
RT: 21.439 min Scan# 3120
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS



Tgt Ion:228 Resp: 1928580
Ion Ratio Lower Upper
228 100
226 28.7 22.7 34.1
229 19.7 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#89
Di-n-octyl phthalate
Concen: 47.107 ng/ul
RT: 22.251 min Scan# 3258
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Tgt Ion:149 Resp: 2049612









#90
Benzo(b)fluoranthene
Concen: 42.949 ng/ul
RT: 23.104 min Scan# 3403
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02

Instrument :
BNA_P
ClientSampleId :
A0BM8MS

Tgt Ion:252 Resp: 161601
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.0
125 8.9 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#91
Benzo(k)fluoranthene
Concen: 45.311 ng/ul
RT: 23.157 min Scan# 3412
Delta R.T. 0.000 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02


Tgt Ion:252 Resp: 1614961
Ion Ratio Lower Upper
252 100
253 21.9 17.9 26.9
125 8.4 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 43.136 ng/ul
RT: 23.745 min Scan# 3512
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:252 Resp: 1500904
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 9.4 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 45.048 ng/ul
RT: 26.415 min Scan# 3966
Delta R.T. -0.012 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Tgt Ion:276 Resp: 1559234
Ion Ratio Lower Upper
276 100
138 20.5 12.3 18.5#
277 25.0 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 44.546 ng/ul
RT: 26.433 min Scan# 3968
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02




Instrument :
BNA_P
ClientSampleId :
A0BM8MS




Tgt Ion:278 Resp: 131934
Ion Ratio Lower Upper
278 100
139 13.3 6.6 10.0
279 23.2 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/26/2022
Supervised By :Yogesh Patel 04/28/2022



#96
Benzo(g,h,i)perylene
Concen: 46.208 ng/ul
RT: 27.204 min Scan# 4100
Delta R.T. -0.006 min
Lab File: BP010093.D
Acq: 25 Apr 2022 21:02



Tgt Ion:276 Resp: 1349567
Ion Ratio Lower Upper
276 100
138 17.2 11.6 17.4
277 23.3 15.9 23.9

