

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042522\
 Data File : BP010107.D
 Acq On : 26 Apr 2022 10:10
 Operator : CG/JU
 Sample : PB144362BS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS362

Quant Time: Apr 27 00:44:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/27/2022
 Supervised By :Yogesh Patel 04/28/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	166163	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	710016	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.569	164	487081	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.322	188	1026425	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.404	240	920262	20.000	ng/u1	# 0.00
88) Perylene-d12	23.857	264	715772	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	22883m	6.005	ng/uL	0.00
4) Pyridine-d5	3.781	84	308426	28.772	ng/u1	0.00
7) Phenol-d5	7.093	99	445707	29.490	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.258	67	289060	31.873	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.463	132	360015	32.163	ng/u1	0.00
15) 4-Methylphenol-d8	8.646	113	383702	30.641	ng/u1	0.00
21) Nitrobenzene-d5	9.093	128	186860	33.090	ng/u1	0.00
24) 2-Nitrophenol-d4	9.816	143	209628	33.990	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.357	165	381372	31.764	ng/u1	0.00
31) 4-Chloroaniline-d4	10.869	131	502417	29.184	ng/u1	-0.01
46) Dimethylphthalate-d6	13.975	166	1233148	32.546	ng/u1	0.00
49) Acenaphthylene-d8	14.257	160	1426694	31.112	ng/u1	0.00
54) 4-Nitrophenol-d4	14.763	143	204521	29.321	ng/u1	0.00
60) Fluorene-d10	15.557	176	1047676	32.144	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.681	200	212292	32.709	ng/u1	0.00
73) Anthracene-d10	17.422	188	1591393	32.047	ng/u1	0.00
81) Pyrene-d10	19.651	212	1866963	35.673	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.704	264	1296476	34.550	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	44215	11.166	ng/uL#	16
5) Pyridine	3.799	79	316039	28.362	ng/u1#	43
6) Benzaldehyde	7.069	77	283500	35.100	ng/u1	99
8) Phenol	7.122	94	463799	30.540	ng/u1#	93
10) Bis(2-Chloroethyl)ether	7.352	93	368926	31.532	ng/u1#	78
12) 2-Chlorophenol	7.499	128	361491	32.017	ng/u1	94
13) 2-Methylphenol	8.375	108	355601	30.532	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.463	45	551702	31.778	ng/u1#	84
16) Acetophenone	8.758	105	585634	29.975	ng/u1#	81
17) N-Nitroso-di-n-propyla...	8.746	70	311688	30.450	ng/u1#	74
18) 4-Methylphenol	8.705	108	387663	30.230	ng/u1	94
19) Hexachloroethane	9.016	117	154757	32.620	ng/u1#	78
22) Nitrobenzene	9.134	77	454309	32.836	ng/u1#	85
23) Isophorone	9.663	82	871147	31.878	ng/u1	97
25) 2-Nitrophenol	9.852	139	219126	33.547	ng/u1#	85
26) 2,4-Dimethylphenol	9.910	107	423823	29.970	ng/u1#	79
27) Bis(2-Chloroethoxy)met...	10.146	93	522702	32.634	ng/u1#	98
29) 2,4-Dichlorophenol	10.387	162	379907	32.917	ng/u1#	83
30) Naphthalene	10.787	128	1199390	32.275	ng/u1	96
32) 4-Chloroaniline	10.899	127	495436	29.205	ng/u1	98
33) Hexachlorobutadiene	11.081	225	269616	32.961	ng/u1	95
34) Caprolactam	11.669	113	123595m	29.771	ng/u1	
35) 4-Chloro-3-methylphenol	12.022	107	438076	31.421	ng/u1#	69
36) 2-Methylnaphthalene	12.398	142	853190	31.656	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042522\
 Data File : BP010107.D
 Acq On : 26 Apr 2022 10:10
 Operator : CG/JU
 Sample : PB144362BS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS362

Quant Time: Apr 27 00:44:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/27/2022
 Supervised By :Yogesh Patel 04/28/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	878123m	32.088	ng/ul	
39) 1,2,4,5-Tetrachloroben...	12.763	216	498879	33.151	ng/ul	94
40) Hexachlorocyclopentadiene	12.751	237	276491	32.603	ng/ul	96
41) 2,4,6-Trichlorophenol	13.004	196	324193	33.296	ng/ul#	83
42) 2,4,5-Trichlorophenol	13.075	196	352782	33.301	ng/ul#	87
43) 1,1'-Biphenyl	13.404	154	1191830	33.176	ng/ul	93
44) 2-Chloronaphthalene	13.445	162	918089	33.023	ng/ul	93
45) 2-Nitroaniline	13.645	65	299845	33.261	ng/ul#	83
47) Dimethylphthalate	14.022	163	1190813	32.338	ng/ul#	92
48) 2,6-Dinitrotoluene	14.140	165	255605	33.824	ng/ul	94
50) Acenaphthylene	14.287	152	1477662	32.531	ng/ul	96
51) 3-Nitroaniline	14.469	138	234690	34.283	ng/ul#	79
52) Acenaphthene	14.634	153	949140	32.135	ng/ul	97
53) 2,4-Dinitrophenol	14.675	184	122433	28.993	ng/ul#	82
55) 4-Nitrophenol	14.775	109	173931	28.701	ng/ul#	45
56) Dibenzofuran	14.963	168	1408390	32.202	ng/ul	98
57) 2,4-Dinitrotoluene	14.928	165	362107	32.800	ng/ul#	86
58) 2,3,4,6-Tetrachlorophenol	15.192	232	304166	32.300	ng/ul#	88
59) Diethylphthalate	15.387	149	1228392	32.797	ng/ul	94
61) Fluorene	15.616	166	1143909	31.856	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.610	204	600056	31.752	ng/ul	99
63) 4-Nitroaniline	15.634	138	237892	35.713	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.692	198	205176	32.766	ng/ul#	91
67) N-Nitrosodiphenylamine	15.822	169	988966	33.580	ng/ul	96
68) 4-Bromophenyl-phenylether	16.504	248	372506	34.024	ng/ul	95
69) Hexachlorobenzene	16.628	284	425275	33.465	ng/ul#	89
70) Atrazine	16.781	200	387741	32.384	ng/ul#	90
71) Pentachlorophenol	16.969	266	254679	31.469	ng/ul#	84
72) Phenanthrene	17.363	178	1822140	32.661	ng/ul	99
74) Anthracene	17.457	178	1813358	32.271	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.369	216	530317	35.391	ng/uL#	86
76) Pentachlorobenzene	14.892	250	497703	33.870	ng/uL	92
77) Carbazole	17.722	167	1570098	31.576	ng/ul#	95
78) Di-n-butylphthalate	18.275	149	2092279	34.201	ng/ul#	92
80) Fluoranthene	19.328	202	2136895	36.200	ng/ul#	96
82) Pyrene	19.680	202	2173120	35.732	ng/ul#	94
83) Butylbenzylphthalate	20.545	149	931101	37.037	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.316	252	628774	37.973	ng/ul#	98
85) Benzo(a)anthracene	21.386	228	1965165	33.649	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.310	149	1356558	36.408	ng/ul#	81
87) Chrysene	21.445	228	1934306	33.637	ng/ul	100
89) Di-n-octyl phthalate	22.251	149	2223638	39.962	ng/ul	100
90) Benzo(b)fluoranthene	23.110	252	1723117	35.809	ng/ul	99
91) Benzo(k)fluoranthene	23.157	252	1628660	35.731	ng/ul#	98
93) Benzo(a)pyrene	23.751	252	1528346	34.346	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.421	276	1447366	32.697	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.439	278	1223956	32.314	ng/ul#	95
96) Benzo(g,h,i)perylene	27.209	276	1236028	33.092	ng/ul#	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042522\
 Data File : BP010107.D
 Acq On : 26 Apr 2022 10:10
 Operator : CG/JU
 Sample : PB144362BS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS362

Quant Time: Apr 27 00:44:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Apr 21 14:49:38 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 04/27/2022
 Supervised By :Yogesh Patel 04/28/2022









#2
1,4-Dioxane
Concen: 11.166 ng/uL
RT: 3.399 min Scan# 51
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :
BNA_P
ClientSampleId :
SLCS362

Tgt Ion: 88 Resp: 44215
Ion Ratio Lower Upper
88 100
43 36.2 6.8 10.2
58 85.7 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#5
Pyridine
Concen: 28.362 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion: 79 Resp: 316039
Ion Ratio Lower Upper
79 100
52 89.4 35.5 53.3#
51 43.0 20.2 30.4#





#6
Benzaldehyde
Concen: 35.100 ng/ul
RT: 7.069 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion: 77 Resp: 283500
Ion Ratio Lower Upper
77 100
105 96.8 77.0 115.4
106 94.1 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#8
Phenol
Concen: 30.540 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion: 94 Resp: 463799
Ion Ratio Lower Upper
94 100
65 32.3 19.0 28.6
66 44.5 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.532 ng/ul
RT: 7.352 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion: 93 Resp: 368920
Ion Ratio Lower Upper
93 100
63 81.8 45.0 67.4
95 33.1 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#12
2-Chlorophenol
Concen: 32.017 ng/ul
RT: 7.499 min Scan# 750
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:128 Resp: 361491
Ion Ratio Lower Upper
128 100
64 52.1 45.3 67.9
130 32.5 29.0 43.6





#13
2-Methylphenol
Concen: 30.532 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:108 Resp: 35560
Ion Ratio Lower Upper
108 100
107 91.1 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.778 ng/ul
RT: 8.463 min Scan# 914
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10



Tgt Ion: 45 Resp: 551702
Ion Ratio Lower Upper
45 100
77 13.5 17.1 25.7#
79 10.3 13.2 19.8#









#16
Acetophenone
Concen: 29.975 ng/ul
RT: 8.758 min Scan# 964
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :
BNA_P
ClientSampleId :
SLCS362

Tgt Ion:105 Resp: 585634
Ion Ratio Lower Upper
105 100
77 79.8 53.4 80.2
51 30.6 13.8 20.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#17
N-Nitroso-di-n-propylamine
Concen: 30.450 ng/ul
RT: 8.746 min Scan# 962
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion: 70 Resp: 311688
Ion Ratio Lower Upper
70 100
42 61.5 31.8 47.8#
101 10.1 15.6 23.4#
130 21.4 13.7 20.5#





#18
4-Methylphenol
Concen: 30.230 ng/ul
RT: 8.705 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:108 Resp: 38766
Ion Ratio Lower Upper
108 100
107 111.1 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#19
Hexachloroethane
Concen: 32.620 ng/ul
RT: 9.016 min Scan# 1008
Delta R.T. -0.012 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:117 Resp: 154757
Ion Ratio Lower Upper
117 100
201 102.5 106.5 159.7#
199 62.2 59.0 88.6





#22
Nitrobenzene
Concen: 32.836 ng/ul
RT: 9.134 min Scan# 1028
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion: 77 Resp: 454309
Ion Ratio Lower Upper
77 100
123 43.3 42.3 63.5
65 14.9 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#23
Isophorone
Concen: 31.878 ng/ul
RT: 9.663 min Scan# 1118
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10



Tgt Ion: 82 Resp: 871147
Ion Ratio Lower Upper
82 100
95 8.0 7.8 11.8
138 17.1 12.9 19.3









#25
2-Nitrophenol
Concen: 33.547 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :
BNA_P
ClientSampleId :
SLCS362

Tgt Ion:139 Resp: 219120
Ion Ratio Lower Upper
139 100
65 58.9 49.4 74.2
109 35.6 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#26
2,4-Dimethylphenol
Concen: 29.970 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:107 Resp: 423823
Ion Ratio Lower Upper
107 100
121 48.8 34.1 51.1
122 79.6 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 32.634 ng/ul
RT: 10.146 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion: 93 Resp: 522702
Ion Ratio Lower Upper
93 100
95 33.1 26.2 39.4
123 11.4 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#29
2,4-Dichlorophenol
Concen: 32.917 ng/ul
RT: 10.387 min Scan# 1241
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:162 Resp: 379907
Ion Ratio Lower Upper
162 100
164 64.8 37.0 55.6
98 36.6 31.0 46.6





#30
Naphthalene
Concen: 32.275 ng/ul
RT: 10.787 min Scan# 1199390
Delta R.T. -0.012 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:128 Resp: 1199390
Ion Ratio Lower Upper
128 100
129 11.1 8.6 13.0
127 13.0 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#32
4-Chloroaniline
Concen: 29.205 ng/ul
RT: 10.899 min Scan# 1328
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:127 Resp: 495436
Ion Ratio Lower Upper
127 100
129 32.1 24.6 37.0





#33
Hexachlorobutadiene
Concen: 32.961 ng/ul
RT: 11.081 min Scan# 1359
Delta R.T. -0.012 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:225 Resp: 269610
Ion Ratio Lower Upper
225 100
223 62.2 53.4 80.0
227 63.6 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#34
Caprolactam
Concen: 29.771 ng/ul m
RT: 11.669 min Scan# 1459
Delta R.T. 0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10



Tgt Ion:113 Resp: 123595
Ion Ratio Lower Upper
113 100
55 182.4 58.2 87.2#
56 145.1 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 31.421 ng/ul
RT: 12.022 min Scan# 1519
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :
BNA_P
ClientSampleId :
SLCS362

Tgt Ion:107 Resp: 438070
Ion Ratio Lower Upper
107 100
144 24.6 15.8 23.8
142 75.6 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#36
2-Methylnaphthalene
Concen: 31.656 ng/ul
RT: 12.398 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:142 Resp: 853190
Ion Ratio Lower Upper
142 100
141 88.1 64.3 96.5





#37
1-Methylnaphthalene
Concen: 32.088 ng/ul m
RT: 12.616 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:142 Resp: 87812
Ion Ratio Lower Upper
142 100
141 91.2 69.8 104.8
116 4.0 5.6 8.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 33.151 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. -0.012 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:216 Resp: 498879
Ion Ratio Lower Upper
216 100
214 77.6 58.2 87.2
179 20.7 14.2 21.2
108 16.4 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 32.603 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:237 Resp: 27649
Ion Ratio Lower Upper
237 100
235 62.9 53.2 79.8
272 13.1 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#41
2,4,6-Trichlorophenol
Concen: 33.296 ng/ul
RT: 13.004 min Scan# 1686
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:196 Resp: 324193
Ion Ratio Lower Upper
196 100
198 96.4 64.1 96.1#
200 32.0 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 33.301 ng/ul
RT: 13.075 min Scan# 1698
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:196 Resp: 35278
Ion Ratio Lower Upper
196 100
198 96.1 68.6 103.0
200 30.0 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#43
1,1'-Biphenyl
Concen: 33.176 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10



Tgt Ion:154 Resp: 1191830
Ion Ratio Lower Upper
154 100
153 40.2 28.6 43.0
76 11.9 8.5 12.7









#44
2-Chloronaphthalene
Concen: 33.023 ng/ul
RT: 13.445 min Scan# 1760
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :
BNA_P
ClientSampleId :
SLCS362

Tgt Ion:162 Resp: 918089
Ion Ratio Lower Upper
162 100
164 32.5 27.5 41.3
127 39.7 37.1 55.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#45
2-Nitroaniline
Concen: 33.261 ng/ul
RT: 13.645 min Scan# 1795
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion: 65 Resp: 299845
Ion Ratio Lower Upper
65 100
92 67.1 77.9 116.9#
138 98.6 81.5 122.3





#47
Dimethylphthalate
Concen: 32.338 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:163 Resp: 119081
Ion Ratio Lower Upper
163 100
194 4.4 4.1 6.1
164 10.0 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#48
2,6-Dinitrotoluene
Concen: 33.824 ng/ul
RT: 14.140 min Scan# 1879
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:165 Resp: 255605
Ion Ratio Lower Upper
165 100
89 58.9 51.7 77.5
121 21.4 18.3 27.5





#50
Acenaphthylene
Concen: 32.531 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:152 Resp: 1477662
Ion Ratio Lower Upper
152 100
151 19.8 18.3 27.5
153 13.3 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#51
3-Nitroaniline
Concen: 34.283 ng/ul
RT: 14.469 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:138 Resp: 234690
Ion Ratio Lower Upper
138 100
108 11.8 15.8 23.8#
92 119.2 117.1 175.7





#52
Acenaphthene
Concen: 32.135 ng/ul
RT: 14.634 min Scan# 1963
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:153 Resp: 949140
Ion Ratio Lower Upper
153 100
152 46.0 34.3 51.5
154 88.7 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#53
2,4-Dinitrophenol
Concen: 28.993 ng/ul
RT: 14.675 min Scan# 1970
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:184 Resp: 122433
Ion Ratio Lower Upper
184 100
63 70.9 34.6 52.0#
154 69.1 57.7 86.5





#55
4-Nitrophenol
Concen: 28.701 ng/ul
RT: 14.775 min Scan# 1986
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:109 Resp: 17393
Ion Ratio Lower Upper
109 100
139 119.3 61.4 92.2
65 135.9 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#56
Dibenzofuran
Concen: 32.202 ng/ul
RT: 14.963 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Tgt Ion:168 Resp: 1408390
Ion Ratio Lower Upper
168 100
139 38.6 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 32.800 ng/ul
RT: 14.928 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :
BNA_P
ClientSampleId :
SLCS362

Tgt Ion:165 Resp: 362107
Ion Ratio Lower Upper
165 100
63 45.9 29.5 44.3#
182 3.1 2.8 4.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 32.300 ng/ul
RT: 15.192 min Scan# 2058
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Tgt Ion:232 Resp: 304166
Ion Ratio Lower Upper
232 100
131 42.6 25.4 38.2#
130 2.2 4.2 6.2#
166 31.3 28.1 42.1





#59
Diethylphthalate
Concen: 32.797 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :

BNA_P

ClientSampleId :





SLCS362

Tgt Ion:149 Resp: 1228392
Ion Ratio Lower Upper
149 100
177 22.7 22.0 33.0
150 12.5 10.2 15.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#61
Fluorene
Concen: 31.856 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:166 Resp: 1143909
Ion Ratio Lower Upper
166 100
165 95.9 70.3 105.5
167 13.2 9.5 14.3





#62
4-Chlorophenyl-phenylether
Concen: 31.752 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion: 204 Resp: 600050
Ion Ratio Lower Upper
204 100
206 33.5 27.6 41.4
141 58.9 46.6 69.8

Manual Integrations
APPROVED



Reviewed By : Jagrut Upadhyay 04/27/2022
Supervised By : Yogesh Patel 04/28/2022




#63
4-Nitroaniline
Concen: 35.713 ng/ul
RT: 15.634 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion: 138 Resp: 237892
Ion Ratio Lower Upper
138 100
92 53.4 41.6 62.4
108 85.8 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 32.766 ng/ul
RT: 15.692 min Scan# 2143
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:198 Resp: 205170
Ion Ratio Lower Upper
198 100
182 3.9 6.0 9.0
77 26.0 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#67
N-Nitrosodiphenylamine
Concen: 33.580 ng/ul
RT: 15.822 min Scan# 2165
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Tgt Ion:169 Resp: 988966
Ion Ratio Lower Upper
169 100
168 64.3 48.3 72.5
167 35.4 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 34.024 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :

BNA_P

ClientSampleId :





SLCS362

Tgt Ion:248 Resp: 372500
Ion Ratio Lower Upper
248 100
250 97.2 79.4 119.0
141 75.5 67.2 100.8

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#69
Hexachlorobenzene
Concen: 33.465 ng/ul
RT: 16.628 min Scan# 2302
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:284 Resp: 425275
Ion Ratio Lower Upper
284 100
142 33.4 19.2 28.8#
249 31.6 27.4 41.2





#70
Atrazine
Concen: 32.384 ng/ul
RT: 16.781 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:200 Resp: 38774
Ion Ratio Lower Upper
200 100
173 25.1 19.2 28.8
215 49.4 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#71
Pentachlorophenol
Concen: 31.469 ng/ul
RT: 16.969 min Scan# 2360
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:266 Resp: 254679
Ion Ratio Lower Upper
266 100
264 62.7 57.3 85.9
268 62.9 65.7 98.5





#72
Phenanthrene
Concen: 32.661 ng/u1
RT: 17.363 min Scan# 2427
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:178 Resp: 1822140
Ion Ratio Lower Upper
178 100
179 15.2 11.7 17.5
176 18.7 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#74
Anthracene
Concen: 32.271 ng/u1
RT: 17.457 min Scan# 2443
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:178 Resp: 1813358
Ion Ratio Lower Upper
178 100
179 15.2 12.3 18.5
176 17.9 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 35.391 ng/uL
RT: 13.369 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:216 Resp: 530317
Ion Ratio Lower Upper
216 100
214 78.2 50.4 75.6
218 47.6 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#76
Pentachlorobenzene
Concen: 33.870 ng/uL
RT: 14.892 min Scan# 2007
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:250 Resp: 497703
Ion Ratio Lower Upper
250 100
248 60.7 43.5 65.3
252 63.4 46.5 69.7





#77
Carbazole
Concen: 31.576 ng/ul
RT: 17.722 min Scan# 2487
Delta R.T. -0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:167 Resp: 1570098
Ion Ratio Lower Upper
167 100
166 21.7 15.8 23.8
139 12.7 8.3 12.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#78
Di-n-butylphthalate
Concen: 34.201 ng/ul
RT: 18.275 min Scan# 2582
Delta R.T. -0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:149 Resp: 2092279
Ion Ratio Lower Upper
149 100
150 9.1 9.6 14.4#
104 5.1 6.2 9.4#





#80
Fluoranthene
Concen: 36.200 ng/u1
RT: 19.328 min Scan# 21
Delta R.T. 0.001 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:202 Resp: 213689
Ion Ratio Lower Upper
202 100
101 9.5 9.6 14.4
100 7.6 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#82
Pyrene
Concen: 35.732 ng/u1
RT: 19.680 min Scan# 2821
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10



Tgt Ion:202 Resp: 2173120
Ion Ratio Lower Upper
202 100
101 11.0 10.6 15.8
100 9.7 9.8 14.6





#83
Butylbenzylphthalate
Concen: 37.037 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Instrument :

BNA_P

ClientSampleId :





SLCS362

Tgt Ion:149 Resp: 93110
Ion Ratio Lower Upper
149 100
91 72.3 71.1 106.7
206 22.6 28.8 43.2

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#84
3,3'-Dichlorobenzidine
Concen: 37.973 ng/ul
RT: 21.316 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:252 Resp: 628774
Ion Ratio Lower Upper
252 100
254 62.9 49.8 74.6
126 9.6 4.3 6.5





#85
Benzo(a)anthracene
Concen: 33.649 ng/ul
RT: 21.386 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:228 Resp: 1965168
Ion Ratio Lower Upper
228 100
229 19.9 18.2 27.2
226 26.4 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 36.408 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:149 Resp: 1356558
Ion Ratio Lower Upper
149 100
167 28.0 32.8 49.2#
279 4.8 5.6 8.4#





#87
Chrysene
Concen: 33.637 ng/ul
RT: 21.445 min Scan# 3121
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:228 Resp: 1934300
Ion Ratio Lower Upper
228 100
226 28.3 22.7 34.1
229 19.3 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#89
Di-n-octyl phthalate
Concen: 39.962 ng/ul
RT: 22.251 min Scan# 3258
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10


Tgt Ion:149 Resp: 2223638





#90
Benzo(b)fluoranthene
Concen: 35.809 ng/ul
RT: 23.110 min Scan# 3404
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:252 Resp: 172311
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 8.2 7.3 10.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#91
Benzo(k)fluoranthene
Concen: 35.731 ng/ul
RT: 23.157 min Scan# 3412
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:252 Resp: 1628660
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 8.4 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 34.346 ng/ul
RT: 23.751 min Scan# 3513
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362




Tgt Ion:252 Resp: 1528340
Ion Ratio Lower Upper
252 100
253 21.6 17.5 26.3
125 9.8 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 32.697 ng/ul
RT: 26.421 min Scan# 3967
Delta R.T. -0.006 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Tgt Ion:276 Resp: 1447366
Ion Ratio Lower Upper
276 100
138 20.4 12.3 18.5
277 25.7 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 32.314 ng/ul
RT: 26.439 min Scan# 3968
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10




Instrument :
BNA_P
ClientSampleId :
SLCS362



Tgt Ion:278 Resp: 1223950
Ion Ratio Lower Upper
278 100
139 13.2 6.6 10.0#
279 24.2 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 04/27/2022
Supervised By :Yogesh Patel 04/28/2022



#96
Benzo(g,h,i)perylene
Concen: 33.092 ng/ul
RT: 27.209 min Scan# 4101
Delta R.T. 0.000 min
Lab File: BP010107.D
Acq: 26 Apr 2022 10:10

Tgt Ion:276 Resp: 1236028
Ion Ratio Lower Upper
276 100
138 18.2 11.6 17.4#
277 24.3 15.9 23.9#

