

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042723\
 Data File : BP014866.D
 Acq On : 27 Apr 2023 10:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020688

Quant Time: Apr 27 23:22:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 26 02:53:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.199	152	268552	20.000	ng/u1	0.00
20) Naphthalene-d8	11.034	136	1175620	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.834	164	748357	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.586	188	1653419	20.000	ng/u1	0.00
79) Chrysene-d12	21.657	240	1243791	20.000	ng/u1	0.00
88) Perylene-d12	24.257	264	1139430	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.499	96	42410	6.209	ng/uL	0.00
4) Pyridine-d5	3.940	84	322372	16.273	ng/u1	0.00
7) Phenol-d5	7.340	99	412088	16.880	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.516	67	242911	16.557	ng/u1	0.00
11) 2-Chlorophenol-d4	7.722	132	321767	17.849	ng/u1	0.00
15) 4-Methylphenol-d8	8.904	113	330668	16.863	ng/u1	0.00
21) Nitrobenzene-d5	9.375	128	159440	18.118	ng/u1	0.00
24) 2-Nitrophenol-d4	10.104	143	169271	18.714	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.646	165	342051	18.193	ng/u1	0.00
31) 4-Chloroaniline-d4	11.163	131	495517	18.367	ng/u1	0.00
46) Dimethylphthalate-d6	14.234	166	990147	17.051	ng/u1	0.00
49) Acenaphthylene-d8	14.528	160	1207529	18.139	ng/u1	0.00
54) 4-Nitrophenol-d4	15.004	143	185109	17.498	ng/u1	0.00
60) Fluorene-d10	15.822	176	892798	18.059	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.933	200	124819	13.605	ng/u1	0.00
73) Anthracene-d10	17.680	188	1394028	17.955	ng/u1	0.00
81) Pyrene-d10	19.892	212	1541582	20.386	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.086	264	1057743	17.906	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.540	88	46104	6.255	ng/uL	98
5) Pyridine	3.958	79	324445	15.974	ng/u1	96
6) Benzaldehyde	7.328	77	145801	12.437	ng/u1	98
8) Phenol	7.369	94	430405	16.952	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.610	93	337662	16.591	ng/u1	98
12) 2-Chlorophenol	7.757	128	328879	17.492	ng/u1	98
13) 2-Methylphenol	8.640	108	330021	17.233	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.734	45	424124	16.721	ng/u1	99
16) Acetophenone	9.034	105	530607	17.298	ng/u1	98
17) N-Nitroso-di-n-propyla...	9.016	70	261515	16.659	ng/u1	97
18) 4-Methylphenol	8.969	108	356046	16.608	ng/u1	98
19) Hexachloroethane	9.299	117	133521	17.604	ng/u1	96
22) Nitrobenzene	9.416	77	389241	17.387	ng/u1	97
23) Isophorone	9.946	82	765226	16.547	ng/u1	99
25) 2-Nitrophenol	10.134	139	182925	18.116	ng/u1	97
26) 2,4-Dimethylphenol	10.187	107	392908	17.536	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.428	93	473926	16.279	ng/u1	98
29) 2,4-Dichlorophenol	10.669	162	336072	17.939	ng/u1	97
30) Naphthalene	11.087	128	1141292	17.551	ng/u1	99
32) 4-Chloroaniline	11.187	127	496629	18.271	ng/u1	99
33) Hexachlorobutadiene	11.369	225	215726	18.105	ng/u1	99
34) Caprolactam	11.951	113	108341	17.540	ng/u1	94
35) 4-Chloro-3-methylphenol	12.293	107	359128	17.312	ng/u1	98
36) 2-Methylnaphthalene	12.681	142	770368	17.448	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042723\
 Data File : BP014866.D
 Acq On : 27 Apr 2023 10:10
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020688


Quant Time: Apr 27 23:22:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Apr 26 02:53:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.892	142	771028	17.517	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	13.040	216	427709	18.243	ng/ul	99
40) Hexachlorocyclopentadiene	13.016	237	189842	13.974	ng/ul	98
41) 2,4,6-Trichlorophenol	13.269	196	278784	18.734	ng/ul	98
42) 2,4,5-Trichlorophenol	13.339	196	304859	18.538	ng/ul	97
43) 1,1'-Biphenyl	13.669	154	1045253	17.768	ng/ul	99
44) 2-Chloronaphthalene	13.716	162	830106	18.025	ng/ul	99
45) 2-Nitroaniline	13.910	65	212218	18.400	ng/ul	97
47) Dimethylphthalate	14.281	163	984087	16.909	ng/ul	100
48) 2,6-Dinitrotoluene	14.398	165	187265	17.959	ng/ul	98
50) Acenaphthylene	14.557	152	1297665	18.049	ng/ul	100
51) 3-Nitroaniline	14.728	138	188307	16.464	ng/ul	100
52) Acenaphthene	14.898	153	885115	17.910	ng/ul	100
53) 2,4-Dinitrophenol	14.934	184	66202	10.777	ng/ul	96
55) 4-Nitrophenol	15.022	109	139690	17.648	ng/ul	97
56) Dibenzofuran	15.228	168	1245654	17.898	ng/ul	100
57) 2,4-Dinitrotoluene	15.186	165	274790	17.885	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	15.451	232	259023	18.678	ng/ul	99
59) Diethylphthalate	15.639	149	978608	17.364	ng/ul	99
61) Fluorene	15.881	166	1011114	18.417	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.869	204	499149	17.934	ng/ul	97
63) 4-Nitroaniline	15.886	138	181157	17.559	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.945	198	126041	13.668	ng/ul	98
67) N-Nitrosodiphenylamine	16.081	169	850704	17.494	ng/ul	98
68) 4-Bromophenyl-phenylether	16.763	248	313954	17.548	ng/ul	98
69) Hexachlorobenzene	16.886	284	372029	17.931	ng/ul	99
70) Atrazine	17.028	200	295916	17.806	ng/ul	99
71) Pentachlorophenol	17.228	266	211398	17.369	ng/ul	99
72) Phenanthrene	17.628	178	1602666	17.700	ng/ul	100
74) Anthracene	17.716	178	1629158	17.846	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.639	216	441165	17.923	ng/ul	98
76) Pentachlorobenzene	15.151	250	432439	17.795	ng/ul	99
77) Carbazole	17.975	167	1458587	18.772	ng/ul	99
78) Di-n-butylphthalate	18.510	149	1704273	18.481	ng/ul	100
80) Fluoranthene	19.569	202	1831577	20.552	ng/ul	100
82) Pyrene	19.922	202	1877150	20.366	ng/ul	98
83) Butylbenzylphthalate	20.774	149	716175	23.158	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.557	252	426357	16.427	ng/ul	100
85) Benzo(a)anthracene	21.639	228	1558754	18.015	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.539	149	1048379	22.836	ng/ul	98
87) Chrysene	21.692	228	1440392	17.560	ng/ul	99
89) Di-n-octyl phthalate	22.527	149	1614711	21.316	ng/ul	100
90) Benzo(b)fluoranthene	23.451	252	1366404	17.510	ng/ul	99
91) Benzo(k)fluoranthene	23.504	252	1278915	16.899	ng/ul	100
93) Benzo(a)pyrene	24.139	252	1151597	17.424	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.015	276	1485493	19.931	ng/ul	99
95) Dibenzo(a,h)anthracene	27.039	278	1226286	19.729	ng/ul	99
96) Benzo(g,h,i)perylene	27.868	276	1198265	20.293	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Apr 27 23:22:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042523.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Apr 26 02:53:24 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.255 ng/uL
RT: 3.540 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion: 88 Resp: 46104
Ion Ratio Lower Upper
88 100
43 28.8 21.3 31.9
58 70.3 55.4 83.0



#5
Pyridine
Concen: 15.974 ng/uL
RT: 3.958 min Scan# 97
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion: 79 Resp: 324445
Ion Ratio Lower Upper
79 100
52 72.0 54.7 82.1
51 32.9 25.5 38.3









#6
Benzaldehyde
Concen: 12.437 ng/ul
RT: 7.328 min Scan# 61
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion: 77 Resp: 145801
Ion Ratio Lower Upper
77 100
105 99.6 79.9 119.9
106 96.9 75.2 112.8



#8
Phenol
Concen: 16.952 ng/ul
RT: 7.369 min Scan# 677
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion: 94 Resp: 430405
Ion Ratio Lower Upper
94 100
65 25.9 21.0 31.6
66 32.9 27.0 40.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 16.591 ng/ul
 RT: 7.610 min Scan# 718
 Delta R.T. -0.000 min
 Lab File: BP014866.D
 Acq: 27 Apr 2023 10:10

Instrument :
 BNA_P
 ClientSampleId :
 SST020688

Tgt Ion: 93 Resp: 337662




Ion	Ratio	Lower	Upper
93	100		
63	72.6	56.3	84.5
95	32.3	25.9	38.9




#12
 2-Chlorophenol
 Concen: 17.492 ng/ul
 RT: 7.757 min Scan# 743
 Delta R.T. -0.000 min
 Lab File: BP014866.D
 Acq: 27 Apr 2023 10:10

Tgt Ion:128 Resp: 328879

Ion	Ratio	Lower	Upper
128	100		
64	45.6	35.6	53.4
130	33.0	25.0	37.6





#13

2-Methylphenol

Concen: 17.233 ng/ul

RT: 8.640 min Scan# 893

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :





SSTD020688

Tgt Ion:108 Resp: 330021

Ion Ratio Lower Upper

108 100

107 88.9 71.0 106.4



#14

2,2'-oxybis(1-Chloropropane)

Concen: 16.721 ng/ul

RT: 8.734 min Scan# 909

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion: 45 Resp: 424124


Ion Ratio Lower Upper

45 100

77 15.8 12.5 18.7

79 12.6 9.4 14.2









#16
Acetophenone
Concen: 17.298 ng/ul
RT: 9.034 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:105 Resp: 530607
Ion Ratio Lower Upper
105 100
77 77.2 59.8 89.8
51 24.9 19.6 29.4



#17
N-Nitroso-di-n-propylamine
Concen: 16.659 ng/ul
RT: 9.016 min Scan# 957
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion: 70 Resp: 261515
Ion Ratio Lower Upper
70 100
42 51.5 43.5 65.3
101 9.4 7.9 11.9
130 20.2 16.4 24.6









#18
4-Methylphenol
Concen: 16.608 ng/ul
RT: 8.969 min Scan# 949
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:108 Resp: 356046
Ion Ratio Lower Upper
108 100
107 113.3 88.6 132.8



#19
Hexachloroethane
Concen: 17.604 ng/ul
RT: 9.299 min Scan# 1005
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:117 Resp: 133521
Ion Ratio Lower Upper
117 100
201 106.4 80.3 120.5
199 62.9 49.0 73.6









#22
Nitrobenzene
Concen: 17.387 ng/ul
RT: 9.416 min Scan# 1025
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion: 77 Resp: 389241
Ion Ratio Lower Upper
77 100
123 47.1 35.8 53.6
65 12.7 10.2 15.2



#23
Isophorone
Concen: 16.547 ng/ul
RT: 9.946 min Scan# 1115
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion: 82 Resp: 765226
Ion Ratio Lower Upper
82 100
95 6.8 5.4 8.2
138 17.1 14.1 21.1









#25
2-Nitrophenol
Concen: 18.116 ng/ul
RT: 10.134 min Scan# 1147
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:139 Resp: 182925
Ion Ratio Lower Upper
139 100
65 47.7 40.7 61.1
109 28.9 23.4 35.0



#26
2,4-Dimethylphenol
Concen: 17.536 ng/ul
RT: 10.187 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:107 Resp: 392908
Ion Ratio Lower Upper
107 100
121 49.8 41.3 61.9
122 85.3 70.4 105.6





#27

Bis(2-Chloroethoxy)methane

Concen: 16.279 ng/ul

RT: 10.428 min Scan# 1198

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion: 93 Resp: 473926

Ion Ratio Lower Upper

93 100

95 33.2 25.2 37.8

123 10.4 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 17.939 ng/ul

RT: 10.669 min Scan# 1238

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:162 Resp: 336072


Ion Ratio Lower Upper

162 100

164 65.5 50.6 75.8

98 36.7 28.5 42.7









#30
Naphthalene
Concen: 17.551 ng/ul
RT: 11.087 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:128 Resp: 1141292
Ion Ratio Lower Upper
128 100
129 10.9 9.1 13.7
127 12.6 10.6 16.0



#32
4-Chloroaniline
Concen: 18.271 ng/ul
RT: 11.187 min Scan# 1326
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:127 Resp: 496629
Ion Ratio Lower Upper
127 100
129 32.6 25.8 38.8









#33
Hexachlorobutadiene
Concen: 18.105 ng/ul
RT: 11.369 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:225 Resp: 215726
Ion Ratio Lower Upper
225 100
223 63.2 49.9 74.9
227 64.3 51.5 77.3



#34
Caprolactam
Concen: 17.540 ng/ul
RT: 11.951 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:113 Resp: 108341
Ion Ratio Lower Upper
113 100
55 151.9 129.0 193.6
56 117.9 99.4 149.2









#35
4-Chloro-3-methylphenol
Concen: 17.312 ng/ul
RT: 12.293 min Scan# 1514
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:107 Resp: 359128
Ion Ratio Lower Upper
107 100
144 25.7 21.4 32.0
142 80.1 65.0 97.6



#36
2-Methylnaphthalene
Concen: 17.448 ng/ul
RT: 12.681 min Scan# 1580
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:142 Resp: 770368
Ion Ratio Lower Upper
142 100
141 87.2 70.6 106.0





#37

1-Methylnaphthalene

Concen: 17.517 ng/ul

RT: 12.892 min Scan# 1617

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:142 Resp: 771028

Ion Ratio Lower Upper

142 100

141 91.3 73.7 110.5

116 3.6 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.243 ng/ul

RT: 13.040 min Scan# 1641

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Tgt Ion:216 Resp: 427709




Ion Ratio Lower Upper


216 100

214 79.8 63.3 94.9

179 20.4 16.7 25.1

108 17.4 14.5 21.7





#40

Hexachlorocyclopentadiene

Concen: 13.974 ng/ul

RT: 13.016 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:237 Resp: 189842

Ion Ratio Lower Upper

237 100

235 63.8 50.0 75.0

272 13.1 10.1 15.1



#41

2,4,6-Trichlorophenol

Concen: 18.734 ng/ul

RT: 13.269 min Scan# 1680

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:196 Resp: 278784


Ion Ratio Lower Upper

196 100

198 94.1 76.7 115.1

200 30.3 25.3 37.9









#42
2,4,5-Trichlorophenol
Concen: 18.538 ng/ul
RT: 13.339 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:196 Resp: 304859
Ion Ratio Lower Upper
196 100
198 96.9 75.3 112.9
200 31.1 24.2 36.4



#43
1,1'-Biphenyl
Concen: 17.768 ng/ul
RT: 13.669 min Scan# 1748
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:154 Resp: 1045253
Ion Ratio Lower Upper
154 100
153 40.4 31.9 47.9
76 13.3 11.2 16.8





#44

2-Chloronaphthalene

Concen: 18.025 ng/ul

RT: 13.716 min Scan# 1756

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:162 Resp: 830106

Ion Ratio Lower Upper

162 100

164 32.9 26.2 39.2

127 38.1 29.7 44.5



#45

2-Nitroaniline

Concen: 18.400 ng/ul

RT: 13.910 min Scan# 1789

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion: 65 Resp: 212218


Ion Ratio Lower Upper

65 100

92 70.7 58.5 87.7

138 107.1 87.6 131.4









#47
Dimethylphthalate
Concen: 16.909 ng/ul
RT: 14.281 min Scan# 1852
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:163 Resp: 984087
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 17.959 ng/ul
RT: 14.398 min Scan# 1872
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:165 Resp: 187265
Ion Ratio Lower Upper
165 100
89 61.3 47.6 71.4
121 20.9 17.5 26.3









#50
Acenaphthylene
Concen: 18.049 ng/ul
RT: 14.557 min Scan# 1899
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:152 Resp: 1297665
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.6
153 12.8 10.4 15.6



#51
3-Nitroaniline
Concen: 16.464 ng/ul
RT: 14.728 min Scan# 1928
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:138 Resp: 188307
Ion Ratio Lower Upper
138 100
108 10.9 8.4 12.6
92 116.7 93.2 139.8





#52

Acenaphthene

Concen: 17.910 ng/ul

RT: 14.898 min Scan# 1957

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:153 Resp: 885115

Ion Ratio Lower Upper

153 100

152 47.3 38.2 57.4

154 88.6 71.0 106.4



#53

2,4-Dinitrophenol

Concen: 10.777 ng/ul

RT: 14.934 min Scan# 1963

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:184 Resp: 66202


Ion Ratio Lower Upper

184 100

63 61.7 53.6 80.4

154 60.1 49.1 73.7









#55
4-Nitrophenol
Concen: 17.648 ng/ul
RT: 15.022 min Scan# 1979
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:109 Resp: 139690
Ion Ratio Lower Upper
109 100
139 136.4 112.8 169.2
65 125.9 103.6 155.4



#56
Dibenzofuran
Concen: 17.898 ng/ul
RT: 15.228 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:168 Resp: 1245654
Ion Ratio Lower Upper
168 100
139 35.8 28.7 43.1





#57

2,4-Dinitrotoluene

Concen: 17.885 ng/ul

RT: 15.186 min Scan# 2006

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:165 Resp: 274790

Ion Ratio Lower Upper

165 100

63 38.7 33.8 50.6

182 2.7 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.678 ng/ul

RT: 15.451 min Scan# 2051

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Tgt Ion:232 Resp: 259023




Ion Ratio Lower Upper


232 100

131 42.9 35.0 52.4

130 2.2 1.8 2.6

166 30.4 25.0 37.6









#59
Diethylphthalate
Concen: 17.364 ng/ul
RT: 15.639 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:149 Resp: 978608
Ion Ratio Lower Upper
149 100
177 22.7 17.9 26.9
150 11.9 9.8 14.6



#61
Fluorene
Concen: 18.417 ng/ul
RT: 15.881 min Scan# 2124
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:166 Resp: 1011114
Ion Ratio Lower Upper
166 100
165 98.0 79.1 118.7
167 13.3 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 17.934 ng/ul
RT: 15.869 min Scan# 2122
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:204 Resp: 499149
Ion Ratio Lower Upper
204 100
206 32.9 26.3 39.5
141 53.6 45.9 68.9



#63
4-Nitroaniline
Concen: 17.559 ng/ul
RT: 15.886 min Scan# 2125
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:138 Resp: 181157
Ion Ratio Lower Upper
138 100
92 51.3 39.0 58.6
108 74.5 59.2 88.8









#66
4,6-Dinitro-2-methylphenol
Concen: 13.668 ng/ul
RT: 15.945 min Scan# 2136
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:198 Resp: 126041
Ion Ratio Lower Upper
198 100
182 4.5 3.5 5.3
77 24.8 18.7 28.1



#67
N-Nitrosodiphenylamine
Concen: 17.494 ng/ul
RT: 16.081 min Scan# 2158
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:169 Resp: 850704
Ion Ratio Lower Upper
169 100
168 65.5 53.6 80.4
167 36.2 29.4 44.0









#68
4-Bromophenyl-phenylether
Concen: 17.548 ng/ul
RT: 16.763 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:248 Resp: 313954
Ion Ratio Lower Upper
248 100
250 98.0 79.1 118.7
141 69.3 53.8 80.6



#69
Hexachlorobenzene
Concen: 17.931 ng/ul
RT: 16.886 min Scan# 2295
Delta R.T. 0.006 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:284 Resp: 372029
Ion Ratio Lower Upper
284 100
142 32.1 26.5 39.7
249 30.3 24.4 36.6





#70

Atrazine

Concen: 17.806 ng/ul

RT: 17.028 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:200 Resp: 295916

Ion Ratio Lower Upper

200 100

173 26.4 21.5 32.3

215 47.7 38.4 57.6



#71

Pentachlorophenol

Concen: 17.369 ng/ul

RT: 17.228 min Scan# 2353

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:266 Resp: 211398


Ion Ratio Lower Upper

266 100

264 60.8 49.0 73.4

268 64.8 51.4 77.2





#72

Phenanthrene

Concen: 17.700 ng/ul

RT: 17.628 min Scan# 2422

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:178 Resp: 1602666

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 19.5 15.7 23.5



#74

Anthracene

Concen: 17.846 ng/ul

RT: 17.716 min Scan# 2436

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:178 Resp: 1629158


Ion Ratio Lower Upper

178 100

179 15.5 12.1 18.1

176 18.5 14.9 22.3





#75

1,2,3,4-Tetrachlorobenzene

Concen: 17.923 ng/uL

RT: 13.639 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:216 Resp: 441165

Ion Ratio Lower Upper

216 100

214 77.5 62.9 94.3

218 47.0 39.5 59.3



#76

Pentachlorobenzene

Concen: 17.795 ng/uL

RT: 15.151 min Scan# 2000

Delta R.T. 0.006 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:250 Resp: 432439


Ion Ratio Lower Upper

250 100

248 63.0 50.4 75.6

252 64.0 50.3 75.5





#77

Carbazole

Concen: 18.772 ng/ul

RT: 17.975 min Scan# 2481

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:167 Resp: 1458587

Ion Ratio Lower Upper

167 100

166 21.9 17.2 25.8

139 12.9 9.8 14.8



#78

Di-n-butylphthalate

Concen: 18.481 ng/ul

RT: 18.510 min Scan# 2571

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:149 Resp: 1704273


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 4.8 3.9 5.9









#80
Fluoranthene
Concen: 20.552 ng/ul
RT: 19.569 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10




Instrument :
BNA_P
ClientSampleId :
SSTD020688


Tgt Ion:202 Resp: 1831577
Ion Ratio Lower Upper
202 100
101 11.3 9.1 13.7
100 8.7 7.0 10.4



#82
Pyrene
Concen: 20.366 ng/ul
RT: 19.922 min Scan# 2811
Delta R.T. -0.000 min
Lab File: BP014866.D
Acq: 27 Apr 2023 10:10

Tgt Ion:202 Resp: 1877150
Ion Ratio Lower Upper
202 100
101 12.9 11.1 16.7
100 10.7 9.0 13.6





#83

Butylbenzylphthalate

Concen: 23.158 ng/ul

RT: 20.774 min Scan# 2957

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:149 Resp: 716175

Ion Ratio Lower Upper

149 100

91 70.1 58.5 87.7

206 21.2 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 16.427 ng/ul

RT: 21.557 min Scan# 3089

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:252 Resp: 426357


Ion Ratio Lower Upper

252 100

254 64.4 51.4 77.0

126 11.1 9.2 13.8





#85

Benzo(a)anthracene

Concen: 18.015 ng/ul

RT: 21.639 min Scan# 3104

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:228 Resp: 1558754

Ion Ratio Lower Upper

228 100

229 19.9 15.7 23.5

226 27.5 22.0 33.0



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.836 ng/ul

RT: 21.539 min Scan# 3086

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:149 Resp: 1048379


Ion Ratio Lower Upper

149 100

167 28.1 21.8 32.6

279 4.5 3.4 5.0





#87

Chrysene

Concen: 17.560 ng/ul

RT: 21.692 min Scan# 3113

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:228 Resp: 1440392

Ion Ratio Lower Upper

228 100

226 30.2 24.5 36.7

229 19.5 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 21.316 ng/ul




RT: 22.527 min Scan# 3254


Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Tgt Ion:149 Resp: 1614711





#90

Benzo(b)fluoranthene

Concen: 17.510 ng/ul

RT: 23.451 min Scan# 3413

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:252 Resp: 1366404

Ion Ratio Lower Upper

252 100

253 22.1 17.2 25.8

125 9.4 7.8 11.8



#91

Benzo(k)fluoranthene

Concen: 16.899 ng/ul

RT: 23.504 min Scan# 3420

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:252 Resp: 1278915


Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 9.5 7.9 11.9





#93

Benzo(a)pyrene

Concen: 17.424 ng/ul

RT: 24.139 min Scan# 3

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10

Instrument :

BNA_P

ClientSampleId :

SSTD020688





Tgt Ion:252 Resp: 1151597

Ion Ratio Lower Upper

252 100

253 22.1 17.5 26.3

125 10.1 8.6 12.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.931 ng/ul

RT: 27.015 min Scan# 4017

Delta R.T. -0.000 min

Lab File: BP014866.D

Acq: 27 Apr 2023 10:10




Tgt Ion:276 Resp: 1485493


Ion Ratio Lower Upper

276 100

138 21.5 17.7 26.5

277 24.4 19.6 29.4









#95
 Dibenzo(a,h)anthracene
 Concen: 19.729 ng/ul
 RT: 27.039 min Scan# 4023
 Delta R.T. 0.006 min
 Lab File: BP014866.D
 Acq: 27 Apr 2023 10:10

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020688

Tgt Ion:278 Resp: 1226286
 Ion Ratio Lower Upper
 278 100
 139 14.5 12.0 18.0
 279 22.8 18.9 28.3



#96
 Benzo(g,h,i)perylene
 Concen: 20.293 ng/ul
 RT: 27.868 min Scan# 4162
 Delta R.T. 0.006 min
 Lab File: BP014866.D
 Acq: 27 Apr 2023 10:10

Tgt Ion:276 Resp: 1198265
 Ion Ratio Lower Upper
 276 100
 138 19.2 16.2 24.2
 277 24.3 18.9 28.3

