

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042823\
 Data File : BP014875.D
 Acq On : 28 Apr 2023 14:29
 Operator : CG/JU
 Sample : SSTD04062
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040649

Quant Time: Apr 28 23:57:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 28 17:04:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.175	152	216811	20.000	ng/ul	0.00
20) Naphthalene-d8	11.016	136	828882	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.822	164	477175	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.581	188	989867	20.000	ng/ul	0.00
79) Chrysene-d12	21.663	240	905465	20.000	ng/ul	0.00
88) Perylene-d12	24.268	264	1015350	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.470	96	94220	15.661	ng/uL	0.00
4) Pyridine-d5	3.905	84	586791	40.577	ng/ul	0.00
7) Phenol-d5	7.316	99	683018	40.251	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.493	67	422974	39.742	ng/ul	0.00
11) 2-Chlorophenol-d4	7.699	132	544815	42.404	ng/ul	0.00
15) 4-Methylphenol-d8	8.887	113	515662	39.880	ng/ul	0.00
21) Nitrobenzene-d5	9.352	128	258549	43.443	ng/ul	0.00
24) 2-Nitrophenol-d4	10.081	143	258962	47.401	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.622	165	509259	43.106	ng/ul	0.00
31) 4-Chloroaniline-d4	11.146	131	698496	41.174	ng/ul	0.00
46) Dimethylphthalate-d6	14.222	166	1420171	41.529	ng/ul	0.00
49) Acenaphthylene-d8	14.516	160	1686619	41.289	ng/ul	0.00
54) 4-Nitrophenol-d4	14.998	143	222696	43.602	ng/ul	0.00
60) Fluorene-d10	15.816	176	1215300	40.769	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.922	200	176418	42.293	ng/ul	0.00
73) Anthracene-d10	17.675	188	1834959	40.590	ng/ul	0.00
81) Pyrene-d10	19.892	212	2066374	41.089	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.098	264	2099478	41.864	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.505	88	100574	15.701	ng/uL	96
5) Pyridine	3.928	79	609231	40.171	ng/ul	98
6) Benzaldehyde	7.299	77	221867	46.616	ng/ul	98
8) Phenol	7.346	94	709031	40.288	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.587	93	608685	40.362	ng/ul	99
12) 2-Chlorophenol	7.734	128	573954	41.631	ng/ul	98
13) 2-Methylphenol	8.616	108	541372	39.709	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.710	45	756424	39.869	ng/ul	99
16) Acetophenone	9.010	105	859001	40.900	ng/ul	100
17) N-Nitroso-di-n-propyla...	8.993	70	417967	41.920	ng/ul	98
18) 4-Methylphenol	8.952	108	570934	39.682	ng/ul	98
19) Hexachloroethane	9.275	117	259330	39.928	ng/ul	95
22) Nitrobenzene	9.399	77	646756	42.838	ng/ul	98
23) Isophorone	9.928	82	1174872	43.199	ng/ul	99
25) 2-Nitrophenol	10.116	139	282950	46.728	ng/ul	97
26) 2,4-Dimethylphenol	10.169	107	620218	42.356	ng/ul	99
27) Bis(2-Chloroethoxy)met...	10.410	93	786802	41.951	ng/ul	99
29) 2,4-Dichlorophenol	10.651	162	518879	43.932	ng/ul	99
30) Naphthalene	11.063	128	1839672	40.364	ng/ul	100
32) 4-Chloroaniline	11.169	127	725565	40.992	ng/ul	98
33) Hexachlorobutadiene	11.351	225	363952	40.432	ng/ul	99
34) Caprolactam	11.934	113	58574	19.734	ng/ul	98
35) 4-Chloro-3-methylphenol	12.275	107	531280	44.191	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP042823\
 Data File : BP014875.D
 Acq On : 28 Apr 2023 14:29
 Operator : CG/JU
 Sample : SST04062
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST040649

Quant Time: Apr 28 23:57:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 28 17:04:54 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.663	142	1177072	40.928	ng/ul	98
37) 1-Methylnaphthalene	12.881	142	1176443	40.743	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.022	216	644072	40.392	ng/ul	99
40) Hexachlorocyclopentadiene	13.004	237	411233	41.082	ng/ul	99
41) 2,4,6-Trichlorophenol	13.257	196	369871	45.652	ng/ul	95
42) 2,4,5-Trichlorophenol	13.328	196	444456	47.170	ng/ul	100
43) 1,1'-Biphenyl	13.657	154	1554626	40.605	ng/ul	99
44) 2-Chloronaphthalene	13.704	162	1221295	40.502	ng/ul	99
45) 2-Nitroaniline	13.904	65	305531	46.381	ng/ul	95
47) Dimethylphthalate	14.269	163	1471724	41.631	ng/ul	99
48) 2,6-Dinitrotoluene	14.392	165	291027	45.699	ng/ul	99
50) Acenaphthylene	14.545	152	1866350	40.872	ng/ul	100
51) 3-Nitroaniline	14.722	138	220332	42.442	ng/ul	97
52) Acenaphthene	14.887	153	1268265	40.295	ng/ul	99
53) 2,4-Dinitrophenol	14.928	184	88175	37.538	ng/ul	89
55) 4-Nitrophenol	15.016	109	165164	42.922	ng/ul	95
56) Dibenzofuran	15.222	168	1752642	40.400	ng/ul	100
57) 2,4-Dinitrotoluene	15.175	165	401325	45.717	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.439	232	351756	47.180	ng/ul	98
59) Diethylphthalate	15.628	149	1445939	41.941	ng/ul	100
61) Fluorene	15.869	166	1397834	40.303	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.857	204	699608	39.913	ng/ul	98
63) 4-Nitroaniline	15.881	138	176406	43.475	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.939	198	188919	43.557	ng/ul	97
67) N-Nitrosodiphenylamine	16.069	169	1191765	40.808	ng/ul	99
68) 4-Bromophenyl-phenylether	16.757	248	434273	40.322	ng/ul	99
69) Hexachlorobenzene	16.881	284	501882	39.534	ng/ul	97
70) Atrazine	17.022	200	433927	41.042	ng/ul	99
71) Pentachlorophenol	17.222	266	227831	40.503	ng/ul	97
72) Phenanthrene	17.622	178	2197189	40.361	ng/ul	99
74) Anthracene	17.710	178	2219171	40.692	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.628	216	643522	39.470	ng/uL	99
76) Pentachlorobenzene	15.139	250	615579	39.332	ng/uL	99
77) Carbazole	17.975	167	1905221	40.957	ng/ul	100
78) Di-n-butylphthalate	18.510	149	2459216	42.615	ng/ul	100
80) Fluoranthene	19.569	202	2526254	41.517	ng/ul	99
82) Pyrene	19.922	202	2611372	40.380	ng/ul	99
83) Butylbenzylphthalate	20.780	149	1046980	43.019	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.569	252	831472	42.309	ng/ul	97
85) Benzo(a)anthracene	21.645	228	2502423	41.203	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.545	149	1599355	43.351	ng/ul	99
87) Chrysene	21.704	228	2415024	39.954	ng/ul	99
89) Di-n-octyl phthalate	22.539	149	2770918	42.079	ng/ul	100
90) Benzo(b)fluoranthene	23.463	252	2687081	42.530	ng/ul	100
91) Benzo(k)fluoranthene	23.515	252	2601233	39.684	ng/ul	99
93) Benzo(a)pyrene	24.151	252	2390601	41.604	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.039	276	3208755	42.051	ng/ul	98
95) Dibenzo(a,h)anthracene	27.062	278	2639007	41.817	ng/ul	99
96) Benzo(g,h,i)perylene	27.898	276	2625264	41.685	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed





Instrument :
BNA_P
ClientSampleId :
SSTD040649



1,4-Dioxane
 Concen: 15.701 ng/uL
 RT: 3.505 min Scan# 2
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	88	Resp:	100574
Ion	Ratio	Lower	Upper
88	100		
43	25.1	22.2	33.4
58	65.9	55.1	82.7



Pyridine
 Concen: 40.171 ng/uL
 RT: 3.928 min Scan# 92
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	79	Resp:	609231
Ion	Ratio	Lower	Upper
79	100		
52	65.0	53.2	79.8
51	31.9	26.5	39.7









Benzaldehyde
 Concen: 46.616 ng/ul
 RT: 7.299 min Scan# 666
 Delta R.T. -0.006 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	77	Resp:	221867
Ion	Ratio	Lower	Upper
77	100		
105	99.8	79.4	119.0
106	93.0	77.6	116.4



Phenol
 Concen: 40.288 ng/ul
 RT: 7.346 min Scan# 673
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	94	Resp:	709031
Ion	Ratio	Lower	Upper
94	100		
65	25.6	21.1	31.7
66	33.8	26.3	39.5









Bis(2-Chloroethyl)ether
 Concen: 40.362 ng/ul
 RT: 7.587 min Scan#
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	93	Resp:	608685
Ion	Ratio	Lower	Upper
93	100		
63	71.0	57.9	86.9
95	32.4	26.4	39.6



2-Chlorophenol
 Concen: 41.631 ng/ul
 RT: 7.734 min Scan# 739
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	128	Resp:	573954
Ion	Ratio	Lower	Upper
128	100		
64	46.2	35.5	53.3
130	32.9	26.8	40.2





2-Methylphenol

Concen: 39.709 ng/u1

RT: 8.616 min Scan# 889

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Instrument :

BNA_P

ClientSampleId :





SSTD040649

Tgt Ion:108 Resp: 541372

Ion Ratio Lower Upper

108 100

107 87.2 71.3 106.9



2,2'-oxybis(1-Chloropropane)

Concen: 39.869 ng/u1

RT: 8.710 min Scan# 905

Delta R.T. -0.006 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29




Tgt Ion: 45 Resp: 756424


Ion Ratio Lower Upper

45 100

77 15.4 12.7 19.1

79 12.8 10.3 15.5





Acetophenone

Concen: 40.900 ng/ul

RT: 9.010 min Scan# 956

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD040649





Tgt Ion:105 Resp: 859001

Ion Ratio Lower Upper

105 100

77 74.8 59.8 89.6

51 23.9 19.4 29.2



N-Nitroso-di-n-propylamine

Concen: 41.920 ng/ul

RT: 8.993 min Scan# 953

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Tgt Ion: 70 Resp: 417967




Ion Ratio Lower Upper


70 100

42 52.8 40.9 61.3

101 10.0 8.2 12.4

130 21.2 16.6 24.8





4-Methylphenol

Concen: 39.682 ng/ul

RT: 8.952 min Scan# 945

Delta R.T. 0.006 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Instrument :

BNP_P

ClientSampleId :





SSTD040649

Tgt Ion:108 Resp: 570934

Ion Ratio Lower Upper

108 100

107 109.9 89.2 133.8



Hexachloroethane

Concen: 39.928 ng/ul

RT: 9.275 min Scan# 1001

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29




Tgt Ion:117 Resp: 259330

Ion Ratio Lower Upper


117 100

201 103.7 78.9 118.3

199 64.7 49.4 74.0



Abundance Scan 1021 (9.393 min): BP014874.D\data.ms (-1021-)



Nitrobenzene

Concen: 42.838 ng/ul

RT: 9.399 min Scan#

Delta R.T. 0.006 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD040649

Tgt Ion: 77 Resp: 646756



Ion Ratio Lower Upper

77 100


123 45.9 35.7 53.5

65 12.7 10.0 15.0


Abundance Scan 1022 (9.399 min): BP014875.D\data.ms



Abundance Scan 1022 (9.399 min): BP014875.D\data.ms (-987) (-)



Abundance Scan 1111 (9.922 min): BP014874.D\data.ms (-1111-)



Isophorone

Concen: 43.199 ng/ul

RT: 9.928 min Scan# 1112

Delta R.T. 0.006 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Tgt Ion: 82 Resp: 1174872



Ion Ratio Lower Upper

82 100


95 6.8 5.4 8.2


138 17.0 13.3 19.9

Abundance Scan 1112 (9.928 min): BP014875.D\data.ms



Abundance Scan 1112 (9.928 min): BP014875.D\data.ms (-1077) (-)









2-Nitrophenol
 Concen: 46.728 ng/ul
 RT: 10.116 min Scan# 1144
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	139	Resp:	282950
Ion	Ratio	Lower	Upper
139	100		
65	49.0	37.0	55.4
109	28.7	22.0	33.0



2,4-Dimethylphenol
 Concen: 42.356 ng/ul
 RT: 10.169 min Scan# 1153
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	107	Resp:	620218
Ion	Ratio	Lower	Upper
107	100		
121	50.4	40.4	60.6
122	88.0	69.3	103.9









Bis(2-Chloroethoxy)methane
Concen: 41.951 ng/ul
RT: 10.410 min Scan# 1194
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	93	Resp:	786802
Ion	Ratio	Lower	Upper
93	100		
95	32.0	26.1	39.1
123	11.3	8.8	13.2



2,4-Dichlorophenol
Concen: 43.932 ng/ul
RT: 10.651 min Scan# 1235
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	162	Resp:	518879
Ion	Ratio	Lower	Upper
162	100		
164	66.2	52.7	79.1
98	34.9	29.4	44.2









Naphthalene
Concen: 40.364 ng/ul
RT: 11.063 min Scan# 1305
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	128	Resp:	1839672
Ion	Ratio	Lower	Upper
128	100		
129	11.0	8.9	13.3
127	13.2	10.4	15.6



4-Chloroaniline
Concen: 40.992 ng/ul
RT: 11.169 min Scan# 1323
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	127	Resp:	725565
Ion	Ratio	Lower	Upper
127	100		
129	31.6	25.9	38.9









Hexachlorobutadiene
 Concen: 40.432 ng/ul
 RT: 11.351 min Scan# 1348
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040649


Tgt Ion:	225	Resp:	363952
Ion	Ratio	Lower	Upper
225	100		
223	62.4	49.7	74.5
227	64.2	51.8	77.6



Caprolactam
 Concen: 19.734 ng/ul
 RT: 11.934 min Scan# 1453
 Delta R.T. -0.012 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	113	Resp:	58574
Ion	Ratio	Lower	Upper
113	100		
55	163.6	133.7	200.5
56	121.8	98.7	148.1









Scan 1511 (12.275 min): BP014874.D\data.ms (-1535) (-)
 4-Chloro-3-methylphenol
 Concen: 44.191 ng/ul
 RT: 12.275 min Scan# 1511
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	107	Resp:	531280
Ion	Ratio	Lower	Upper
107	100		
144	26.0	19.8	29.8
142	79.3	62.4	93.6



Scan 1577 (12.663 min): BP014874.D\data.ms (-1536) (-)
 2-Methylnaphthalene
 Concen: 40.928 ng/ul
 RT: 12.663 min Scan# 1577
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	142	Resp:	1177072
Ion	Ratio	Lower	Upper
142	100		
141	89.5	70.5	105.7









Scan 1614 (12.881 min): BP014874.D\data.ms (-1637) (-)
 1-Methylnaphthalene
 Concen: 40.743 ng/u1
 RT: 12.881 min Scan# 1614
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040649


Tgt Ion:	142	Resp:	1176443
Ion	Ratio	Lower	Upper
142	100		
141	93.1	73.3	109.9
116	3.6	3.0	4.4



Scan 1638 (13.022 min): BP014874.D\data.ms (-1639) (-)
 1,2,4,5-Tetrachlorobenzene
 Concen: 40.392 ng/u1
 RT: 13.022 min Scan# 1638
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	216	Resp:	644072
Ion	Ratio	Lower	Upper
216	100		
214	77.8	63.0	94.6
179	20.9	16.2	24.4
108	17.2	13.6	20.4









Hexachlorocyclopentadiene
Concen: 41.082 ng/ul
RT: 13.004 min Scan# 1635
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	237	Resp:	411233
Ion	Ratio	Lower	Upper
237	100		
235	62.2	50.6	75.8
272	13.0	10.7	16.1



2,4,6-Trichlorophenol
Concen: 45.652 ng/ul
RT: 13.257 min Scan# 1678
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	196	Resp:	369871
Ion	Ratio	Lower	Upper
196	100		
198	94.1	79.1	118.7
200	30.8	26.5	39.7









2,4,5-Trichlorophenol
Concen: 47.170 ng/ul
RT: 13.328 min Scan# 1690
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649

Tgt Ion:	196	Resp:	444456
Ion	Ratio	Lower	Upper
196	100		
198	95.6	76.6	115.0
200	30.5	24.0	36.0




1,1'-Biphenyl
Concen: 40.605 ng/ul
RT: 13.657 min Scan# 1746
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	154	Resp:	1554626
Ion	Ratio	Lower	Upper
154	100		
153	40.1	32.5	48.7
76	13.5	10.8	16.2





Abundance Scan 1754 (13.704 min): BP014874.D\data.ms (-1746) (-)




2-Chloronaphthalene
Concen: 40.502 ng/ul
RT: 13.704 min Scan# 1754
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	162	Resp:	1221295
Ion	Ratio	Lower	Upper
162	100		
164	32.3	25.6	38.4
127	37.6	30.3	45.5



Abundance Scan 1754 (13.704 min): BP014875.D\data.ms (-1720) (-)





Abundance Scan 1788 (13.904 min): BP014874.D\data.ms (-1705) (-)





2-Nitroaniline
Concen: 46.381 ng/ul
RT: 13.904 min Scan# 1788
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	65	Resp:	305531
Ion	Ratio	Lower	Upper
65	100		
92	72.8	58.6	87.8
138	114.4	84.8	127.2



Abundance Scan 1788 (13.904 min): BP014875.D\data.ms (-1754) (-)





Dimethylphthalate

Concen: 41.631 ng/ul

RT: 14.269 min Scan# 1850

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD040649





Tgt Ion:163 Resp: 1471724

Ion Ratio Lower Upper

163 100

194 3.9 3.2 4.8

164 10.1 8.3 12.5



2,6-Dinitrotoluene

Concen: 45.699 ng/ul

RT: 14.392 min Scan# 1871

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29




Tgt Ion:165 Resp: 291027


Ion Ratio Lower Upper

165 100

89 55.6 45.2 67.8

121 20.5 16.3 24.5









Acenaphthylene
 Concen: 40.872 ng/ul
 RT: 14.545 min Scan# 1897
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	152	Resp:	1866350
Ion	Ratio	Lower	Upper
152	100		
151	20.0	16.0	24.0
153	13.1	10.6	15.8



3-Nitroaniline
 Concen: 42.442 ng/ul
 RT: 14.722 min Scan# 1927
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	138	Resp:	220332
Ion	Ratio	Lower	Upper
138	100		
108	11.0	9.0	13.4
92	121.7	94.2	141.4









Acenaphthene
Concen: 40.295 ng/u1
RT: 14.887 min Scan# 1955
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	153	Resp:	1268265
Ion	Ratio	Lower	Upper
153	100		
152	47.7	38.5	57.7
154	88.5	69.4	104.2



2,4-Dinitrophenol
Concen: 37.538 ng/u1
RT: 14.928 min Scan# 1962
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	184	Resp:	88175
Ion	Ratio	Lower	Upper
184	100		
63	61.1	57.6	86.4
154	61.8	54.6	82.0









4-Nitrophenol
 Concen: 42.922 ng/ul
 RT: 15.016 min Scan# 1976
 Delta R.T. 0.006 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	109	Resp:	165164
Ion	Ratio	Lower	Upper
109	100		
139	141.8	107.8	161.8
65	130.3	108.6	162.8



Dibenzofuran
 Concen: 40.400 ng/ul
 RT: 15.222 min Scan# 2012
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	168	Resp:	1752642
Ion	Ratio	Lower	Upper
168	100		
139	34.9	27.8	41.8









2,4-Dinitrotoluene
 Concen: 45.717 ng/u1
 RT: 15.175 min Scan# 2004
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649

Tgt Ion:	165	Resp:	401325
Ion Ratio	Lower	Upper	
165	100		
63	40.4	33.9	50.9
182	2.6	2.1	3.1




2,3,4,6-Tetrachlorophenol
 Concen: 47.180 ng/u1
 RT: 15.439 min Scan# 2049
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	232	Resp:	351756
Ion Ratio	Lower	Upper	
232	100		
131	44.1	34.4	51.6
130	2.5	1.8	2.6
166	31.2	25.9	38.9






Abundance Scan 2081 (15.628 min): BP014874.D\data.ms (-2075) (-)




Diethylphthalate
Concen: 41.941 ng/u1
RT: 15.628 min Scan# 2081
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Instrument :
BNA_P
ClientSampleId :
SSTD040649

Tgt Ion:	149	Resp:	1445939
Ion	Ratio	Lower	Upper
149	100		
177	22.5	18.0	27.0
150	12.6	9.8	14.8






Abundance Scan 2122 (15.869 min): BP014874.D\data.ms (-2165) (-)



Fluorene
Concen: 40.303 ng/u1
RT: 15.869 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	166	Resp:	1397834
Ion	Ratio	Lower	Upper
166	100		
165	98.4	79.8	119.6
167	13.7	11.0	16.6



Abundance Scan 2120 (15.857 min): BP014874.D\data.ms (-2160) (-)

4-Chlorophenyl-phenylether

Concen: 39.913 ng/ul

RT: 15.857 min Scan# 2120

Delta R.T. -0.000 min

Lab File: BP014875.D


Acq: 28 Apr 2023 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD040649






Tgt Ion:204 Resp: 699608

Ion Ratio Lower Upper

204 100

206 33.4 26.1 39.1

141 57.2 44.6 67.0



Abundance Scan 2124 (15.881 min): BP014874.D\data.ms (-2163) (-)

4-Nitroaniline

Concen: 43.475 ng/ul

RT: 15.881 min Scan# 2124

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29




Tgt Ion:138 Resp: 176406


Ion Ratio Lower Upper

138 100

92 49.3 44.2 66.2

108 69.5 60.1 90.1









4,6-Dinitro-2-methylphenol
 Concen: 43.557 ng/ul
 RT: 15.939 min Scan# 2134
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649

Tgt Ion:	198	Resp:	188919
Ion	Ratio	Lower	Upper
198	100		
182	4.3	4.2	6.2
77	23.9	20.4	30.6



N-Nitrosodiphenylamine
 Concen: 40.808 ng/ul
 RT: 16.069 min Scan# 2156
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	169	Resp:	1191765
Ion	Ratio	Lower	Upper
169	100		
168	66.9	53.1	79.7
167	36.8	29.9	44.9



Abundance Scan 2273 (16.757 min): BP014874.D\data.ms (-2268) (-)

4-Bromophenyl-phenylether

Concen: 40.322 ng/ul

RT: 16.757 min Scan# 2273

Delta R.T. -0.000 min

Lab File: BP014875.D


Acq: 28 Apr 2023 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD040649






Tgt Ion:248 Resp: 434273

Ion Ratio Lower Upper

248 100

250 98.0 77.3 115.9

141 69.5 55.0 82.6



Abundance Scan 2293 (16.875 min): BP014874.D\data.ms (-2269) (-)

Hexachlorobenzene

Concen: 39.534 ng/ul

RT: 16.881 min Scan# 2294

Delta R.T. 0.006 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29




Tgt Ion:284 Resp: 501882


Ion Ratio Lower Upper

284 100

142 31.8 27.6 41.4

249 30.4 24.4 36.6









Atrazine
Concen: 41.042 ng/ul
RT: 17.022 min Scan# 2318
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	200	Resp:	433927
Ion Ratio	Lower	Upper	
200	100		
173	26.3	21.0	31.4
215	47.0	38.0	57.0



Pentachlorophenol
Concen: 40.503 ng/ul
RT: 17.222 min Scan# 2352
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	266	Resp:	227831
Ion Ratio	Lower	Upper	
266	100		
264	63.4	47.8	71.6
268	64.8	51.1	76.7









Scan 2420 (17.622 min): BP014874.D\data.ms (-2473) (-)
 Phenanthrene
 Concen: 40.361 ng/u1
 RT: 17.622 min Scan# 2420
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	178	Resp:	2197189
Ion	Ratio	Lower	Upper
178	100		
179	15.5	12.3	18.5
176	19.4	15.2	22.8



Scan 2435 (17.710 min): BP014874.D\data.ms (-2474) (-)
 Anthracene
 Concen: 40.692 ng/u1
 RT: 17.710 min Scan# 2435
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	178	Resp:	2219171
Ion	Ratio	Lower	Upper
178	100		
179	15.8	12.3	18.5
176	18.9	15.2	22.8









Scan 1741 (13.628 min): BP014874.D\data.ms (-1735) (-)
 1,2,3,4-Tetrachlorobenzene
 Concen: 39.470 ng/uL
 RT: 13.628 min Scan# 1741
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29




Instrument :
 BNA_P
 ClientSampleId :
 SST040649


Tgt Ion:	216	Resp:	643522
Ion	Ratio	Lower	Upper
216	100		
214	78.1	62.1	93.1
218	48.6	38.6	58.0



Scan 1998 (15.140 min): BP014874.D\data.ms (-1976) (-)
 Pentachlorobenzene
 Concen: 39.332 ng/uL
 RT: 15.139 min Scan# 1998
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	250	Resp:	615579
Ion	Ratio	Lower	Upper
250	100		
248	63.1	50.5	75.7
252	63.5	52.3	78.5









Carbazole
Concen: 40.957 ng/ul
RT: 17.975 min Scan# 2480
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649

Tgt Ion:	167	Resp:	1905221
Ion	Ratio	Lower	Upper
167	100		
166	21.6	17.2	25.8
139	12.3	10.2	15.2




Di-n-butylphthalate
Concen: 42.615 ng/ul
RT: 18.510 min Scan# 2571
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	149	Resp:	2459216
Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.4	11.0
104	4.8	4.0	6.0



Abundance Scan 2751 (19.569 min): BP014874.D\data.ms (-2748) (-)



Fluoranthene

Concen: 41.517 ng/ul

RT: 19.569 min Scan# 2751

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29

Instrument :

BNA_P

ClientSampleId :

SSTD040649



Tgt Ion:202 Resp: 2526254

Ion Ratio Lower Upper


202 100

101 12.2 9.4 14.2


100 9.0 7.3 10.9



Abundance Scan 2751 (19.569 min): BP014875.D\data.ms (-2717) (-)



Abundance Scan 2811 (19.922 min): BP014874.D\data.ms (-2802) (-)



Pyrene

Concen: 40.380 ng/ul

RT: 19.922 min Scan# 2811

Delta R.T. -0.000 min

Lab File: BP014875.D

Acq: 28 Apr 2023 14:29



Tgt Ion:202 Resp: 2611372

Ion Ratio Lower Upper


202 100


101 13.8 10.9 16.3

100 11.7 8.6 12.8



Abundance Scan 2811 (19.922 min): BP014875.D\data.ms (-2777) (-)









Butylbenzylphthalate
Concen: 43.019 ng/ul
RT: 20.780 min Scan# 2957
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	149	Resp:	1046980
Ion	Ratio	Lower	Upper
149	100		
91	71.4	58.3	87.5
206	21.3	17.4	26.0



3,3'-Dichlorobenzidine
Concen: 42.309 ng/ul
RT: 21.569 min Scan# 3091
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	252	Resp:	831472
Ion	Ratio	Lower	Upper
252	100		
254	63.5	52.6	78.8
126	10.4	8.9	13.3









Scan 3104 (21.645 min): BP014874.D\data.ms (-3085) (-)
 Benzo(a)anthracene
 Concen: 41.203 ng/u1
 RT: 21.645 min Scan# 3104
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Instrument :
 BNA_P
 ClientSampleId :
 SST040649

Tgt Ion:228 Resp: 2502423




Ion	Ratio	Lower	Upper
228	100		
229	19.7	15.8	23.6
226	27.5	21.7	32.5




Scan 3087 (21.545 min): BP014874.D\data.ms (-3086) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 43.351 ng/u1
 RT: 21.545 min Scan# 3087
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:149 Resp: 1599355

Ion	Ratio	Lower	Upper
149	100		
167	27.4	22.2	33.2
279	4.1	3.4	5.0









Chrysene
Concen: 39.954 ng/u1
RT: 21.704 min Scan# 3114
Delta R.T. 0.006 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:228 Resp: 2415024
Ion Ratio Lower Upper
228 100
226 30.7 24.0 36.0
229 19.8 15.6 23.4



Di-n-octyl phthalate
Concen: 42.079 ng/u1
RT: 22.539 min Scan# 3256
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:149 Resp: 2770918









Scan 3413 (23.463 min): BP014874.D\data.ms (-3490) (-)
 Benzo(b)fluoranthene
 Concen: 42.530 ng/u1
 RT: 23.463 min Scan# 3413
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040649

Tgt Ion:252 Resp: 2687081




Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.5	26.3
125	10.1	8.1	12.1




Scan 3422 (23.516 min): BP014874.D\data.ms (-3491) (-)
 Benzo(k)fluoranthene
 Concen: 39.684 ng/u1
 RT: 23.515 min Scan# 3422
 Delta R.T. -0.000 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:252 Resp: 2601233

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.5	26.3
125	10.3	7.8	11.6









Benzo(a)pyrene
Concen: 41.604 ng/u1
RT: 24.151 min Scan# 3530
Delta R.T. -0.000 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29




Instrument :
BNA_P
ClientSampleId :
SSTD040649


Tgt Ion:	252	Resp:	2390601
Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.6	26.4
125	10.7	8.4	12.6



Indeno(1,2,3-cd)pyrene
Concen: 42.051 ng/u1
RT: 27.039 min Scan# 4021
Delta R.T. 0.006 min
Lab File: BP014875.D
Acq: 28 Apr 2023 14:29

Tgt Ion:	276	Resp:	3208755
Ion	Ratio	Lower	Upper
276	100		
138	23.1	17.4	26.0
277	25.2	19.4	29.2









Scan 4024 (27.057 min): BP014874.D\data.ms (-4095) (-)
 Dibenzo(a,h)anthracene
 Concen: 41.817 ng/ul
 RT: 27.062 min Scan# 4166
 Delta R.T. 0.006 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040649

Tgt Ion:	278	Resp:	2639007
Ion	Ratio	Lower	Upper
278	100		
139	15.6	12.1	18.1
279	23.4	18.7	28.1



Scan 4166 (27.892 min): BP014874.D\data.ms (-4148) (-)
 Benzo(g,h,i)perylene
 Concen: 41.685 ng/ul
 RT: 27.898 min Scan# 4167
 Delta R.T. 0.006 min
 Lab File: BP014875.D
 Acq: 28 Apr 2023 14:29

Tgt Ion:	276	Resp:	2625264
Ion	Ratio	Lower	Upper
276	100		
138	20.4	15.8	23.6
277	24.0	19.0	28.6

