

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP042921\  
 Data File : BP005408.D  
 Acq On : 29 Apr 2021 13:54  
 Operator : CG/JU  
 Sample : SSTDICCC040  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICCC040

Manual Integrations  
 APPROVED

mohammad  
 5/3/2021 10:48:04 AM

Quant Time: Apr 29 14:42:33 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP042921.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Apr 29 14:39:11 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.628  | 152  | 25637    | 20.00 | ng    | # 0.00   |
| 21) Naphthalene-d8            | 10.416 | 136  | 87361    | 20.00 | ng    | 0.00     |
| 39) Acenaphthene-d10          | 14.293 | 164  | 66764    | 20.00 | ng    | 0.00     |
| 64) Phenanthrene-d10          | 17.057 | 188  | 164152   | 20.00 | ng    | # 0.00   |
| 76) Chrysene-d12              | 21.163 | 240  | 224964   | 20.00 | ng    | 0.00     |
| 86) Perylene-d12              | 23.416 | 264  | 241498   | 20.00 | ng    | 0.00     |
|                               |        |      |          |       |       |          |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 5) 2-Fluorophenol             | 5.228  | 112  | 122356   | 82.90 | ng    | 0.00     |
| 7) Phenol-d6                  | 6.817  | 99   | 166693   | 81.98 | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 8.793  | 82   | 200630   | 99.89 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 15.798 | 330  | 110444   | 86.48 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 12.904 | 172  | 445824   | 86.44 | ng    | 0.00     |
| 79) Terphenyl-d14             | 19.639 | 244  | 1000594  | 82.42 | ng    | 0.00     |
|                               |        |      |          |       |       |          |
| Target Compounds              |        |      |          |       |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.164  | 88   | 32775    | 41.67 | ng    | # 91     |
| 3) Pyridine                   | 3.564  | 79   | 77273    | 45.27 | ng    | # 83     |
| 4) n-Nitrosodimethylamine     | 3.481  | 42   | 32195    | 36.59 | ng    | # 31     |
| 6) Aniline                    | 6.964  | 93   | 91555    | 41.49 | ng    | # 82     |
| 8) 2-Chlorophenol             | 7.199  | 128  | 61643    | 40.79 | ng    | 87       |
| 9) Benzaldehyde               | 6.781  | 77   | 42201    | 35.49 | ng    | # 70     |
| 10) Phenol                    | 6.840  | 94   | 82734    | 40.79 | ng    | # 58     |
| 11) bis(2-Chloroethyl)ether   | 7.064  | 93   | 58713    | 40.15 | ng    | 96       |
| 12) 1,3-Dichlorobenzene       | 7.516  | 146  | 74138    | 40.58 | ng    | # 86     |
| 13) 1,4-Dichlorobenzene       | 7.664  | 146  | 74493    | 39.98 | ng    | 98       |
| 14) 1,2-Dichlorobenzene       | 7.975  | 146  | 72608    | 40.74 | ng    | 96       |
| 15) Benzyl Alcohol            | 7.881  | 79   | 77645    | 47.88 | ng    | # 74     |
| 16) 2,2'-oxybis(1-Chloropr... | 8.164  | 45   | 24370    | 19.86 | ng    | # 51     |
| 17) 2-Methylphenol            | 8.081  | 107  | 52376    | 41.06 | ng    | # 81     |
| 18) Hexachloroethane          | 8.693  | 117  | 32318    | 45.36 | ng    | 94       |
| 19) n-Nitroso-di-n-propyla... | 8.440  | 70   | 65014    | 49.23 | ng    | # 82     |
| 20) 3+4-Methylphenols         | 8.411  | 107  | 72706    | 42.46 | ng    | 86       |
| 22) Acetophenone              | 8.458  | 105  | 106312   | 42.40 | ng    | # 99     |
| 24) Nitrobenzene              | 8.834  | 77   | 102753   | 47.32 | ng    | # 88     |
| 25) Isophorone                | 9.358  | 82   | 162473   | 46.55 | ng    | # 93     |
| 26) 2-Nitrophenol             | 9.540  | 139  | 35298    | 50.50 | ng    | 87       |
| 27) 2,4-Dimethylphenol        | 9.605  | 122  | 51434    | 43.63 | ng    | # 84     |
| 28) bis(2-Chloroethoxy)met... | 9.840  | 93   | 70660    | 42.10 | ng    | 98       |
| 29) 2,4-Dichlorophenol        | 10.075 | 162  | 69776    | 44.73 | ng    | 97       |
| 30) 1,2,4-Trichlorobenzene    | 10.281 | 180  | 86752    | 42.33 | ng    | 97       |
| 31) Naphthalene               | 10.463 | 128  | 187501   | 40.67 | ng    | 99       |
| 32) Benzoic acid              | 9.781  | 122  | 39946    | 52.06 | ng    | # 74     |
| 33) 4-Chloroaniline           | 10.587 | 127  | 76734    | 43.88 | ng    | # 94     |
| 34) Hexachlorobutadiene       | 10.746 | 225  | 76152    | 41.88 | ng    | 96       |
| 35) Caprolactam               | 11.393 | 113  | 19194    | 44.12 | ng    | 92       |
| 36) 4-Chloro-3-methylphenol   | 11.734 | 107  | 74174    | 44.08 | ng    | 90       |
| 37) 2-Methylnaphthalene       | 12.087 | 142  | 144944   | 42.94 | ng    | 96       |
| 38) 1-Methylnaphthalene       | 12.310 | 142  | 138622   | 42.99 | ng    | 91       |
| 40) 1,2,4,5-Tetrachloroben... | 12.463 | 216  | 116099   | 39.94 | ng    | 98       |
| 41) Hexachlorocyclopentadiene | 12.434 | 237  | 42883    | 45.53 | ng    | 94       |
| 43) 2,4,6-Trichlorophenol     | 12.716 | 196  | 71134    | 45.18 | ng    | 94       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP042921\  
 Data File : BP005408.D  
 Acq On : 29 Apr 2021 13:54  
 Operator : CG/JU  
 Sample : SSTDICCC040  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICCC040

Manual Integrations  
 APPROVED


mohammad  
 5/3/2021 10:48:04 AM

Quant Time: Apr 29 14:42:33 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP042921.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Apr 29 14:39:11 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| 44) 2,4,5-Trichlorophenol     | 12.799 | 196  | 75926    | 44.17 | ng    | 92       |
| 46) 1,1'-Biphenyl             | 13.116 | 154  | 201536   | 42.41 | ng    | 97       |
| 47) 2-Chloronaphthalene       | 13.157 | 162  | 165076   | 42.19 | ng    | 97       |
| 48) 2-Nitroaniline            | 13.381 | 65   | 55855    | 52.00 | ng    | 94       |
| 49) Acenaphthylene            | 14.010 | 152  | 240687   | 42.51 | ng    | 99       |
| 50) Dimethylphthalate         | 13.763 | 163  | 212506   | 43.85 | ng    | 100      |
| 51) 2,6-Dinitrotoluene        | 13.887 | 165  | 47380    | 44.83 | ng    | 93       |
| 52) Acenaphthene              | 14.357 | 154  | 148964   | 41.70 | ng    | 97       |
| 53) 3-Nitroaniline            | 14.222 | 138  | 39130    | 44.78 | ng    | 99       |
| 54) 2,4-Dinitrophenol         | 14.440 | 184  | 32017    | 53.16 | ng    | 91       |
| 55) Dibenzofuran              | 14.698 | 168  | 268428   | 41.95 | ng    | 95       |
| 56) 4-Nitrophenol             | 14.551 | 139  | 29507    | 42.42 | ng    | 83       |
| 57) 2,4-Dinitrotoluene        | 14.687 | 165  | 69606    | 47.08 | ng    | 94       |
| 58) Fluorene                  | 15.351 | 166  | 226152   | 44.05 | ng    | 98       |
| 59) 2,3,4,6-Tetrachlorophenol | 14.934 | 232  | 74035    | 41.39 | ng    | 92       |
| 60) Diethylphthalate          | 15.134 | 149  | 211730   | 46.29 | ng    | 97       |
| 61) 4-Chlorophenyl-phenyle... | 15.351 | 204  | 140449   | 42.64 | ng    | 98       |
| 62) 4-Nitroaniline            | 15.393 | 138  | 37396    | 45.61 | ng    | 92       |
| 63) Azobenzene                | 15.645 | 77   | 261618   | 49.08 | ng    | 96       |
| 65) 4,6-Dinitro-2-methylph... | 15.457 | 198  | 50000    | 46.33 | ng    | 92       |
| 66) n-Nitrosodiphenylamine    | 15.569 | 169  | 188431   | 42.01 | ng    | # 94     |
| 67) 4-Bromophenyl-phenylether | 16.245 | 248  | 93489    | 41.68 | ng    | 92       |
| 68) Hexachlorobenzene         | 16.357 | 284  | 106976   | 40.67 | ng    | # 90     |
| 69) Atrazine                  | 16.534 | 200  | 85756    | 45.80 | ng    | 99       |
| 70) Pentachlorophenol         | 16.716 | 266  | 59244    | 41.89 | ng    | 97       |
| 71) Phenanthrene              | 17.098 | 178  | 358176   | 40.94 | ng    | 99       |
| 72) Anthracene                | 17.192 | 178  | 359365   | 42.18 | ng    | 99       |
| 73) Carbazole                 | 17.475 | 167  | 319817   | 41.71 | ng    | 99       |
| 74) Di-n-butylphthalate       | 18.034 | 149  | 358404   | 49.50 | ng    | # 96     |
| 75) Fluoranthene              | 19.086 | 202  | 493635   | 41.10 | ng    | 98       |
| 77) Benzidine                 | 19.281 | 184  | 167473   | 47.11 | ng    | 100      |
| 78) Pyrene                    | 19.439 | 202  | 509013   | 40.35 | ng    | 99       |
| 80) Butylbenzylphthalate      | 20.322 | 149  | 172608   | 60.24 | ng    | 90       |
| 81) Benzo(a)anthracene        | 21.151 | 228  | 575340   | 39.75 | ng    | 99       |
| 82) 3,3'-Dichlorobenzidine    | 21.086 | 252  | 201250   | 44.73 | ng    | 92       |
| 83) Chrysene                  | 21.204 | 228  | 567846   | 39.38 | ng    | 99       |
| 84) Bis(2-ethylhexyl)phtha... | 21.075 | 149  | 273763   | 62.19 | ng    | 98       |
| 85) Di-n-octyl phthalate      | 21.951 | 149  | 461381   | 53.97 | ng    | 99       |
| 87) Indeno(1,2,3-cd)pyrene    | 25.727 | 276  | 798177   | 43.26 | ng    | 98       |
| 88) Benzo(b)fluoranthene      | 22.739 | 252  | 628590   | 39.53 | ng    | 99       |
| 89) Benzo(k)fluoranthene      | 22.780 | 252  | 623453m  | 39.94 | ng    |          |
| 90) Benzo(a)pyrene            | 23.322 | 252  | 583438   | 40.24 | ng    | 99       |
| 91) Dibenzo(a,h)anthracene    | 25.733 | 278  | 693779   | 43.42 | ng    | 100      |
| 92) Benzo(g,h,i)perylene      | 26.427 | 276  | 645522   | 43.13 | ng    | # 97     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed









#1  
1,4-Dichlorobenzene-d4  
Concen: 20.00 ng  
RT: 7.628 min Scan# 772  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion: 152 Resp: 25637  
Ion Ratio Lower Upper  
152 100  
150 170.3 119.8 179.8  
115 74.2 49.0 73.4#

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#2  
1,4-Dioxane  
Concen: 41.67 ng  
RT: 3.164 min Scan# 13  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion: 88 Resp: 32775  
Ion Ratio Lower Upper  
88 100  
58 46.6 42.5 63.7  
43 19.8 19.8 29.8#





#3  
Pyridine  
Concen: 45.27 ng  
RT: 3.564 min Scan# 81  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 79 Resp: 77273  
Ion Ratio Lower Upper  
79 100  
52 34.1 40.3 60.5#  
51 26.7 20.0 30.0




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#4  
n-Nitrosodimethylamine  
Concen: 36.59 ng  
RT: 3.481 min Scan# 67  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion: 42 Resp: 32195  
Ion Ratio Lower Upper  
42 100  
74 150.8 122.6 183.8  
44 2.9 146.9 220.3#





#5  
2-Fluorophenol  
Concen: 82.90 ng  
RT: 5.228 min Scan# 364  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:112 Resp: 122356  
Ion Ratio Lower Upper  
112 100  
64 64.9 33.6 50.4#  
63 44.0 18.6 27.8#




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#6  
Aniline  
Concen: 41.49 ng  
RT: 6.964 min Scan# 659  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion: 93 Resp: 91555  
Ion Ratio Lower Upper  
93 100  
66 46.0 27.0 40.4#  
65 23.2 14.4 21.6#





#7  
Phenol-d6  
Concen: 81.98 ng  
RT: 6.817 min Scan# 634  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 99 Resp: 166693  
Ion Ratio Lower Upper  
99 100  
42 17.3 12.9 19.3  
71 63.9 26.7 40.1#




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#8  
2-Chlorophenol  
Concen: 40.79 ng  
RT: 7.199 min Scan# 699  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion: 128 Resp: 61643  
Ion Ratio Lower Upper  
128 100  
130 34.5 14.1 54.1  
64 51.1 17.2 57.2





#9  
Benzaldehyde  
Concen: 35.49 ng  
RT: 6.781 min Scan# 628  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 77 Resp: 42201  
Ion Ratio Lower Upper  
77 100  
105 75.1 89.4 129.4#  
106 72.8 79.9 119.9#

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#10  
Phenol  
Concen: 40.79 ng  
RT: 6.840 min Scan# 638  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54



Tgt Ion: 94 Resp: 82734  
Ion Ratio Lower Upper  
94 100  
65 40.4 8.0 48.0  
66 70.3 17.3 57.3#









#11  
bis(2-Chloroethyl)ether  
Concen: 40.15 ng  
RT: 7.064 min Scan# 676  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 93    | Resp: | 58713 |
| Ion      | Ratio | Lower | Upper |
| 93       | 100   |       |       |
| 63       | 67.7  | 49.6  | 89.6  |
| 95       | 29.3  | 14.4  | 54.4  |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#12  
1,3-Dichlorobenzene  
Concen: 40.58 ng  
RT: 7.516 min Scan# 753  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 146   | Resp: | 74138 |
| Ion      | Ratio | Lower | Upper |
| 146      | 100   |       |       |
| 148      | 62.9  | 51.4  | 77.0  |
| 75       | 52.4  | 23.9  | 35.9# |









#13  
1,4-Dichlorobenzene  
Concen: 39.98 ng  
RT: 7.664 min Scan# 778  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 146   | Resp: | 74493 |
| Ion      | Ratio | Lower | Upper |
| 146      | 100   |       |       |
| 148      | 66.1  | 52.2  | 78.4  |
| 111      | 45.1  | 34.2  | 51.4  |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#14  
1,2-Dichlorobenzene  
Concen: 40.74 ng  
RT: 7.975 min Scan# 831  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 146   | Resp: | 72608 |
| Ion      | Ratio | Lower | Upper |
| 146      | 100   |       |       |
| 148      | 65.2  | 51.5  | 77.3  |
| 111      | 47.6  | 34.0  | 51.0  |





#15  
Benzyl Alcohol  
Concen: 47.88 ng  
RT: 7.881 min Scan# 815  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion: 79 Resp: 77645  
Ion Ratio Lower Upper  
79 100  
108 52.8 64.0 96.0#  
77 70.7 72.7 109.1#

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 19.86 ng  
RT: 8.164 min Scan# 863  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Tgt Ion: 45 Resp: 24370  
Ion Ratio Lower Upper  
45 100  
77 56.5 7.0 47.0#  
79 38.7 1.0 41.0





#17  
2-Methylphenol  
Concen: 41.06 ng  
RT: 8.081 min Scan# 849  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 107   | Resp: | 52376 |
| Ion      | Ratio | Lower | Upper |
| 107      | 100   |       |       |
| 108      | 106.4 | 85.8  | 128.8 |
| 77       | 77.0  | 41.4  | 62.0# |
| 79       | 76.1  | 38.6  | 57.8# |




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#18  
Hexachloroethane  
Concen: 45.36 ng  
RT: 8.693 min Scan# 953  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 117   | Resp: | 32318 |
| Ion      | Ratio | Lower | Upper |
| 117      | 100   |       |       |
| 119      | 94.5  | 80.1  | 120.1 |
| 201      | 131.1 | 98.8  | 148.2 |





#19  
n-Nitroso-di-n-propylamine  
Concen: 49.23 ng  
RT: 8.440 min Scan# 910  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 70    | Resp: | 65014 |
| Ion      | Ratio | Lower | Upper |
| 70       | 100   |       |       |
| 42       | 33.7  | 39.7  | 59.5# |
| 101      | 8.8   | 7.4   | 11.0  |
| 130      | 15.6  | 17.2  | 25.8# |

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#20  
3+4-Methylphenols  
Concen: 42.46 ng  
RT: 8.411 min Scan# 905  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 107   | Resp: | 72706 |
| Ion      | Ratio | Lower | Upper |
| 107      | 100   |       |       |
| 108      | 82.0  | 69.5  | 109.5 |
| 77       | 56.5  | 21.5  | 61.5  |
| 79       | 47.0  | 16.8  | 56.8  |





#21  
Naphthalene-d8  
Concen: 20.00 ng  
RT: 10.416 min Scan# 1246  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:136 Resp: 87361  
Ion Ratio Lower Upper  
136 100  
137 10.1 8.9 13.3  
54 6.0 5.0 7.4  
68 4.0 3.7 5.5

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#22  
Acetophenone  
Concen: 42.40 ng  
RT: 8.458 min Scan# 913  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54



Tgt Ion:105 Resp: 106312  
Ion Ratio Lower Upper  
105 100  
71 1.6 0.9 1.3#  
51 22.6 18.2 27.2  
120 20.9 17.0 25.6





#23

Nitrobenzene-d5

Concen: 99.89 ng

RT: 8.793 min Scan# 970

Delta R.T. 0.000 min

Lab File: BP005408.D

Acq: 29 Apr 2021 13:54

Tgt Ion: 82 Resp: 200630

Ion Ratio Lower Upper

|     |      |      |       |
|-----|------|------|-------|
| 82  | 100  |      |       |
| 128 | 30.0 | 30.5 | 45.7# |
| 54  | 36.4 | 33.9 | 50.9  |

Instrument :

BNA\_P

ClientSampleId :





SSTDICCC040

Manual Integrations

APPROVED

mohammad

5/3/2021 10:48:04 AM



#24

Nitrobenzene

Concen: 47.32 ng

RT: 8.834 min Scan# 977

Delta R.T. 0.000 min




Lab File: BP005408.D


Acq: 29 Apr 2021 13:54

Tgt Ion: 77 Resp: 102753

Ion Ratio Lower Upper


|     |      |      |       |
|-----|------|------|-------|
| 77  | 100  |      |       |
| 123 | 29.8 | 31.7 | 47.5# |
| 65  | 14.5 | 11.5 | 17.3  |





#25  
Isophorone  
Concen: 46.55 ng  
RT: 9.358 min Scan# 1066  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 82 Resp: 162473  
Ion Ratio Lower Upper  
82 100  
95 10.0 7.5 11.3  
138 13.8 14.6 22.0#




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#26  
2-Nitrophenol  
Concen: 50.50 ng  
RT: 9.540 min Scan# 1097  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion: 139 Resp: 35298  
Ion Ratio Lower Upper  
139 100  
109 59.4 40.7 61.1  
65 69.8 47.8 71.6









#27  
2,4-Dimethylphenol  
Concen: 43.63 ng  
RT: 9.605 min Scan# 1108  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

| Tgt Ion | Ratio | Lower | Upper  |
|---------|-------|-------|--------|
| 122     | 100   |       |        |
| 107     | 142.1 | 93.8  | 140.8# |
| 121     | 58.7  | 45.6  | 68.4   |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#28  
bis(2-Chloroethoxy)methane  
Concen: 42.10 ng  
RT: 9.840 min Scan# 1148  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 93      | 100   |       |       |
| 95      | 30.8  | 24.9  | 37.3  |
| 123     | 9.8   | 9.5   | 14.3  |









#29  
2,4-Dichlorophenol  
Concen: 44.73 ng  
RT: 10.075 min Scan# 1188  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:162 Resp: 69776  
Ion Ratio Lower Upper  
162 100  
164 64.9 44.1 84.1  
98 36.9 12.6 52.6

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#30  
1,2,4-Trichlorobenzene  
Concen: 42.33 ng  
RT: 10.281 min Scan# 1223  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:180 Resp: 86752  
Ion Ratio Lower Upper  
180 100  
182 99.7 77.2 115.8  
145 28.7 21.7 32.5





#31  
Naphthalene  
Concen: 40.67 ng  
RT: 10.463 min Scan# 1254  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:128 Resp: 187501  
Ion Ratio Lower Upper  
128 100  
129 11.3 8.0 12.0  
127 13.9 11.1 16.7




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#32  
Benzoic acid  
Concen: 52.06 ng  
RT: 9.781 min Scan# 1138  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:122 Resp: 39946  
Ion Ratio Lower Upper  
122 100  
105 132.5 105.1 145.1  
77 139.6 69.6 109.6#





#33  
4-Chloroaniline  
Concen: 43.88 ng  
RT: 10.587 min Scan# 1275  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 127   | Resp: | 76734 |
| Ion Ratio | Lower | Upper |       |
| 127       | 100   |       |       |
| 129       | 30.5  | 25.5  | 38.3  |
| 65        | 45.4  | 30.0  | 45.0# |
| 92        | 24.6  | 20.2  | 30.2  |




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM




#34  
Hexachlorobutadiene  
Concen: 41.88 ng  
RT: 10.746 min Scan# 1302  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 225   | Resp: | 76152 |
| Ion Ratio | Lower | Upper |       |
| 225       | 100   |       |       |
| 223       | 60.0  | 51.1  | 76.7  |
| 227       | 60.1  | 50.2  | 75.2  |









#35  
Caprolactam  
Concen: 44.12 ng  
RT: 11.393 min Scan# 1412  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54





Tgt Ion:113 Resp: 19194  
Ion Ratio Lower Upper  
113 100  
55 103.4 72.0 112.0  
56 87.8 65.1 105.1



#36  
4-Chloro-3-methylphenol  
Concen: 44.08 ng  
RT: 11.734 min Scan# 1470  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Tgt Ion:107 Resp: 74174  
Ion Ratio Lower Upper  
107 100  
144 21.0 20.1 30.1  
142 69.0 62.2 93.4



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Manual Integrations  
APPROVED

mohammad  
5/3/2021 10:48:04 AM







#37  
2-Methylnaphthalene  
Concen: 42.94 ng  
RT: 12.087 min Scan# 1530  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:142 Resp: 144944  
Ion Ratio Lower Upper  
142 100  
141 89.4 74.7 112.1  
115 56.8 43.8 65.8

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#38  
1-Methylnaphthalene  
Concen: 42.99 ng  
RT: 12.310 min Scan# 1568  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:142 Resp: 138622  
Ion Ratio Lower Upper  
142 100  
141 87.5 77.0 115.6  
116 6.8 5.0 7.6





#39  
Acenaphthene-d10  
Concen: 20.00 ng  
RT: 14.293 min Scan# 1905  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:164 Resp: 66764  
Ion Ratio Lower Upper  
164 100  
162 104.8 73.5 110.3  
160 46.5 40.6 61.0

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 39.94 ng  
RT: 12.463 min Scan# 1594  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Tgt Ion:216 Resp: 116099  
Ion Ratio Lower Upper  
216 100  
214 77.4 60.6 91.0  
179 19.6 15.3 22.9  
108 18.3 13.8 20.6





#41  
Hexachlorocyclopentadiene  
Concen: 45.53 ng  
RT: 12.434 min Scan# 1589  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion: 237 Resp: 42883  
Ion Ratio Lower Upper  
237 100  
235 66.3 40.9 80.9  
272 14.4 0.0 34.6




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#42  
2,4,6-Tribromophenol  
Concen: 86.48 ng  
RT: 15.798 min Scan# 2161  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion: 330 Resp: 110444  
Ion Ratio Lower Upper  
330 100  
332 98.2 76.3 114.5  
141 37.1 22.8 34.2#









#43  
2,4,6-Trichlorophenol  
Concen: 45.18 ng  
RT: 12.716 min Scan# 1637  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 196     | 100   |       |       |
| 198     | 96.8  | 81.4  | 122.2 |
| 200     | 31.0  | 28.2  | 42.2  |

Instrument :  
BNA\_P  
ClientSampled :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#44  
2,4,5-Trichlorophenol  
Concen: 44.17 ng  
RT: 12.799 min Scan# 1651  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 196     | 100   |       |       |
| 198     | 97.3  | 71.0  | 106.4 |
| 97      | 53.0  | 38.5  | 57.7  |
| 132     | 19.2  | 15.6  | 23.4  |





#45  
2-Fluorobiphenyl  
Concen: 86.44 ng  
RT: 12.904 min Scan# 1669  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:172 Resp: 445824  
Ion Ratio Lower Upper  
172 100  
171 35.5 28.1 42.1  
170 23.9 19.6 29.4

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#46  
1,1'-Biphenyl  
Concen: 42.41 ng  
RT: 13.116 min Scan# 1705  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Tgt Ion:154 Resp: 201536  
Ion Ratio Lower Upper  
154 100  
153 41.2 18.9 58.9  
76 11.8 0.0 32.1








#47  
2-Chloronaphthalene  
Concen: 42.19 ng  
RT: 13.157 min Scan# 1712  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040





Tgt Ion:162 Resp: 165076  
Ion Ratio Lower Upper  
162 100  
127 41.8 31.4 47.0  
164 34.2 26.6 39.8




#48  
2-Nitroaniline  
Concen: 52.00 ng  
RT: 13.381 min Scan# 1750  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Tgt Ion: 65 Resp: 55855  
Ion Ratio Lower Upper  
65 100  
92 64.1 48.8 73.2  
138 80.0 69.5 104.3








#49  
Acenaphthylene  
Concen: 42.51 ng  
RT: 14.010 min Scan# 1857  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040





Tgt Ion:152 Resp: 240687  
Ion Ratio Lower Upper  
152 100  
151 20.2 15.8 23.6  
153 12.9 10.5 15.7




#50  
Dimethylphthalate  
Concen: 43.85 ng  
RT: 13.763 min Scan# 1815  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54



Tgt Ion:163 Resp: 212506  
Ion Ratio Lower Upper  
163 100  
194 5.0 3.9 5.9  
164 10.2 8.1 12.1









#51  
2,6-Dinitrotoluene  
Concen: 44.83 ng  
RT: 13.887 min Scan# 1836  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:165 Resp: 47380  
Ion Ratio Lower Upper  
165 100  
63 77.7 67.4 101.0  
89 84.8 73.8 110.6




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#52  
Acenaphthene  
Concen: 41.70 ng  
RT: 14.357 min Scan# 1916  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:154 Resp: 148964  
Ion Ratio Lower Upper  
154 100  
153 113.4 90.9 136.3  
152 54.9 48.5 72.7









#53  
3-Nitroaniline  
Concen: 44.78 ng  
RT: 14.222 min Scan# 1893  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:138 Resp: 39130  
Ion Ratio Lower Upper  
138 100  
108 11.4 10.2 15.4  
92 136.3 108.1 162.1

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#54  
2,4-Dinitrophenol  
Concen: 53.16 ng  
RT: 14.440 min Scan# 1930  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:184 Resp: 32017  
Ion Ratio Lower Upper  
184 100  
63 70.5 51.2 76.8  
154 78.0 55.8 83.8





#55  
Dibenzofuran  
Concen: 41.95 ng  
RT: 14.698 min Scan# 1974  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:168 Resp: 268428  
Ion Ratio Lower Upper  
168 100  
139 43.2 31.4 47.0  
169 13.3 11.4 17.0




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#56  
4-Nitrophenol  
Concen: 42.42 ng  
RT: 14.551 min Scan# 1949  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:139 Resp: 29507  
Ion Ratio Lower Upper  
139 100  
109 162.7 122.8 162.8  
65 149.9 110.0 150.0









#57  
2,4-Dinitrotoluene  
Concen: 47.08 ng  
RT: 14.687 min Scan# 1972  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 165   | Resp: | 69606 |
| Ion Ratio | Lower | Upper |       |
| 165       | 100   |       |       |
| 63        | 79.7  | 56.7  | 85.1  |
| 89        | 103.6 | 85.8  | 128.6 |
| 182       | 4.2   | 3.0   | 4.6   |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#58  
Fluorene  
Concen: 44.05 ng  
RT: 15.351 min Scan# 2085  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 166   | Resp: | 226152 |
| Ion Ratio | Lower | Upper |        |
| 166       | 100   |       |        |
| 165       | 96.7  | 78.8  | 118.2  |
| 167       | 13.5  | 10.9  | 16.3   |





#59  
2,3,4,6-Tetrachlorophenol  
Concen: 41.39 ng  
RT: 14.934 min Scan# 2014  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampled :  
SSTDICCC040




|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 232   | Resp: | 74035 |
| Ion Ratio | Lower | Upper |       |
| 232       | 100   |       |       |
| 131       | 39.6  | 26.5  | 39.7  |
| 130       | 2.9   | 2.1   | 3.1   |
| 166       | 26.4  | 19.4  | 29.0  |

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#60  
Diethylphthalate  
Concen: 46.29 ng  
RT: 15.134 min Scan# 2048  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54



|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 149   | Resp: | 211730 |
| Ion Ratio | Lower | Upper |        |
| 149       | 100   |       |        |
| 177       | 25.8  | 19.2  | 28.8   |
| 150       | 13.0  | 9.9   | 14.9   |









#61  
4-Chlorophenyl-phenylether  
Concen: 42.64 ng  
RT: 15.351 min Scan# 2085  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 204   | Resp: | 140449 |
| Ion      | Ratio | Lower | Upper  |
| 204      | 100   |       |        |
| 206      | 33.1  | 26.0  | 39.0   |
| 141      | 53.3  | 44.2  | 66.2   |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#62  
4-Nitroaniline  
Concen: 45.61 ng  
RT: 15.393 min Scan# 2092  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 138   | Resp: | 37396 |
| Ion      | Ratio | Lower | Upper |
| 138      | 100   |       |       |
| 92       | 66.7  | 50.9  | 90.9  |
| 108      | 187.0 | 154.7 | 194.7 |





#63  
Azobenzene  
Concen: 49.08 ng  
RT: 15.645 min Scan# 2135  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Instrument :  
BNA\_P  
ClientSampled :  
SSTDICCC040






Tgt Ion: 77 Resp: 261618

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 77  | 100   |       |       |
| 182 | 25.2  | 5.5   | 45.5  |
| 105 | 14.6  | 0.0   | 36.5  |
| 51  | 26.7  | 10.3  | 50.3  |

Manual Integrations  
APPROVED




mohammad  
5/3/2021 10:48:04 AM




#64  
Phenanthrene-d10  
Concen: 20.00 ng  
RT: 17.057 min Scan# 2375  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion: 188 Resp: 164152

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 188 | 100   |       |       |
| 94  | 4.7   | 5.0   | 7.6#  |
| 80  | 6.0   | 5.9   | 8.9   |









#65  
4,6-Dinitro-2-methylphenol  
Concen: 46.33 ng  
RT: 15.457 min Scan# 2103  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 198   | Resp: | 50000 |
| Ion Ratio | Lower | Upper |       |
| 198       | 100   |       |       |
| 51        | 23.2  | 10.3  | 50.3  |
| 105       | 36.1  | 13.6  | 53.6  |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#66  
n-Nitrosodiphenylamine  
Concen: 42.01 ng  
RT: 15.569 min Scan# 2122  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 169   | Resp: | 188431 |
| Ion Ratio | Lower | Upper |        |
| 169       | 100   |       |        |
| 168       | 69.9  | 55.2  | 82.8   |
| 167       | 36.3  | 36.6  | 54.8#  |





#67  
4-Bromophenyl-phenylether  
Concen: 41.68 ng  
RT: 16.245 min Scan# 2237  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:248 Resp: 93489  
Ion Ratio Lower Upper  
248 100  
250 100.8 77.5 116.3  
141 73.1 50.1 75.1




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#68  
Hexachlorobenzene  
Concen: 40.67 ng  
RT: 16.357 min Scan# 2256  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:284 Resp: 106976  
Ion Ratio Lower Upper  
284 100  
142 33.0 20.8 31.2#  
249 33.6 23.7 35.5









#69  
Atrazine  
Concen: 45.80 ng  
RT: 16.534 min Scan# 2286  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:200 Resp: 85756  
Ion Ratio Lower Upper  
200 100  
173 19.8 0.7 40.7  
215 49.1 29.0 69.0

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#70  
Pentachlorophenol  
Concen: 41.89 ng  
RT: 16.716 min Scan# 2317  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:266 Resp: 59244  
Ion Ratio Lower Upper  
266 100  
268 63.3 49.4 74.0  
264 63.5 53.7 80.5









#71  
Phenanthrene  
Concen: 40.94 ng  
RT: 17.098 min Scan# 2382  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:178 Resp: 358176  
Ion Ratio Lower Upper  
178 100  
176 19.0 15.4 23.2  
179 14.5 12.3 18.5

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#72  
Anthracene  
Concen: 42.18 ng  
RT: 17.192 min Scan# 2398  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:178 Resp: 359365  
Ion Ratio Lower Upper  
178 100  
176 18.1 15.2 22.8  
179 15.8 12.6 19.0





#73  
 Carbazole  
 Concen: 41.71 ng  
 RT: 17.475 min Scan# 2446  
 Delta R.T. 0.000 min  
 Lab File: BP005408.D  
 Acq: 29 Apr 2021 13:54




Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDICC040



Tgt Ion:167 Resp: 319817  
 Ion Ratio Lower Upper  
 167 100  
 166 21.2 17.4 26.0  
 139 13.3 10.4 15.6




Manual Integrations  
 APPROVED


mohammad  
 5/3/2021 10:48:04 AM



#74  
 Di-n-butylphthalate  
 Concen: 49.50 ng  
 RT: 18.034 min Scan# 2541  
 Delta R.T. 0.000 min  
 Lab File: BP005408.D  
 Acq: 29 Apr 2021 13:54


Tgt Ion:149 Resp: 358404  
 Ion Ratio Lower Upper  
 149 100  
 150 9.4 7.6 11.4  
 104 8.6 8.9 13.3#





#75  
Fluoranthene  
Concen: 41.10 ng  
RT: 19.086 min Scan# 2720  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:202 Resp: 493635  
Ion Ratio Lower Upper  
202 100  
101 4.8 0.0 26.1  
203 17.5 0.0 37.2




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#76  
Chrysene-d12  
Concen: 20.00 ng  
RT: 21.163 min Scan# 3073  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:240 Resp: 224964  
Ion Ratio Lower Upper  
240 100  
120 4.9 4.2 6.2  
236 25.2 20.2 30.2









#77  
Benzidine  
Concen: 47.11 ng  
RT: 19.281 min Scan# 2753  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040

Tgt Ion:184 Resp: 167473  
Ion Ratio Lower Upper  
184 100  
185 14.0 11.1 16.7  
183 12.2 9.8 14.8




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#78  
Pyrene  
Concen: 40.35 ng  
RT: 19.439 min Scan# 2780  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:202 Resp: 509013  
Ion Ratio Lower Upper  
202 100  
200 20.8 16.9 25.3  
203 16.8 13.8 20.6





#79  
Terphenyl-d14  
Concen: 82.42 ng  
RT: 19.639 min Scan# 2814  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040



Tgt Ion:244 Resp: 1000594  
Ion Ratio Lower Upper  
244 100  
212 8.2 5.8 8.6  
122 5.5 5.2 7.8




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#80  
Butylbenzylphthalate  
Concen: 60.24 ng  
RT: 20.322 min Scan# 2930  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:149 Resp: 172608  
Ion Ratio Lower Upper  
149 100  
91 87.1 60.2 90.4  
206 29.0 23.9 35.9





#81  
Benzo(a)anthracene  
Concen: 39.75 ng  
RT: 21.151 min Scan# 3071  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampled :  
SSTDICCC040



Tgt Ion:228 Resp: 575340  
Ion Ratio Lower Upper  
228 100  
226 27.1 21.2 31.8  
229 19.7 15.6 23.4




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#82  
3,3'-Dichlorobenzidine  
Concen: 44.73 ng  
RT: 21.086 min Scan# 3060  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

Tgt Ion:252 Resp: 201250  
Ion Ratio Lower Upper  
252 100  
254 69.4 49.9 74.9  
126 7.7 5.5 8.3









#83  
Chrysene  
Concen: 39.38 ng  
RT: 21.204 min Scan# 3080  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 228   | Resp: | 567846 |
| Ion Ratio | Lower | Upper |        |
| 228       | 100   |       |        |
| 226       | 31.1  | 24.8  | 37.2   |
| 229       | 19.9  | 15.4  | 23.0   |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#84  
Bis(2-ethylhexyl)phthalate  
Concen: 62.19 ng  
RT: 21.075 min Scan# 3058  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 149   | Resp: | 273763 |
| Ion Ratio | Lower | Upper |        |
| 149       | 100   |       |        |
| 167       | 29.0  | 23.8  | 35.8   |
| 279       | 6.5   | 5.7   | 8.5    |









#85  
Di-n-octyl phthalate  
Concen: 53.97 ng  
RT: 21.951 min Scan# 3207  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 149   | Resp: | 461381 |
| Ion Ratio | Lower | Upper |        |
| 149       | 100   |       |        |
| 167       | 1.6   | 1.5   | 2.3    |
| 43        | 6.8   | 5.1   | 7.7    |

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#86  
Perylene-d12  
Concen: 20.00 ng  
RT: 23.416 min Scan# 3456  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 264   | Resp: | 241498 |
| Ion Ratio | Lower | Upper |        |
| 264       | 100   |       |        |
| 260       | 24.3  | 20.6  | 30.8   |
| 265       | 21.5  | 19.0  | 28.6   |





#87  
Indeno(1,2,3-cd)pyrene  
Concen: 43.26 ng  
RT: 25.727 min Scan# 3849  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampled :  
SSTDICCC040




Tgt Ion: 276 Resp: 798177  
Ion Ratio Lower Upper  
276 100  
138 8.8 8.4 12.6  
277 25.2 20.6 31.0

Manual Integrations  
APPROVED



mohammad  
5/3/2021 10:48:04 AM




#88  
Benzo(b)fluoranthene  
Concen: 39.53 ng  
RT: 22.739 min Scan# 3341  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Tgt Ion: 252 Resp: 628590  
Ion Ratio Lower Upper  
252 100  
253 22.0 18.2 27.2  
125 3.8 3.4 5.0





#89  
Benzo(k)fluoranthene  
Concen: 39.94 ng m  
RT: 22.780 min Scan# 3348  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampled :  
SSTDICCC040



Tgt Ion:252 Resp: 623453  
Ion Ratio Lower Upper  
252 100  
253 21.4 17.2 25.8  
125 3.7 3.8 5.6#




Manual Integrations  
APPROVED


mohammad  
5/3/2021 10:48:04 AM



#90  
Benzo(a)pyrene  
Concen: 40.24 ng  
RT: 23.322 min Scan# 3440  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54


Tgt Ion:252 Resp: 583438  
Ion Ratio Lower Upper  
252 100  
253 21.6 17.1 25.7  
125 3.8 3.8 5.6





#91  
Dibenzo(a,h)anthracene  
Concen: 43.42 ng  
RT: 25.733 min Scan# 3850  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54




Instrument :  
BNA\_P  
ClientSampleId :  
SSTDICCC040




Tgt Ion:278 Resp: 693779  
Ion Ratio Lower Upper  
278 100  
139 6.1 5.0 7.6  
279 23.3 18.6 28.0

Manual Integrations  
APPROVED

mohammad  
5/3/2021 10:48:04 AM



#92  
Benzo(g,h,i)perylene  
Concen: 43.13 ng  
RT: 26.427 min Scan# 3968  
Delta R.T. 0.000 min  
Lab File: BP005408.D  
Acq: 29 Apr 2021 13:54



Tgt Ion:276 Resp: 645522  
Ion Ratio Lower Upper  
276 100  
277 24.5 19.8 29.8  
138 6.8 7.7 11.5#

