

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP043021\
 Data File : BP005432.D
 Acq On : 30 Apr 2021 14:04
 Operator : CG/JU
 Sample : M2229-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 TP-6

Quant Time: Apr 30 14:32:42 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 11:27:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	32401	20.00	ng	0.00
21) Naphthalene-d8	10.410	136	114181	20.00	ng	0.00
39) Acenaphthene-d10	14.287	164	81273	20.00	ng	0.00
64) Phenanthrene-d10	17.051	188	196762	20.00	ng	0.00
76) Chrysene-d12	21.163	240	260459	20.00	ng	# 0.00
86) Perylene-d12	23.416	264	278349	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.228	112	210014	113.40	ng	0.00
7) Phenol-d6	6.816	99	270898	104.51	ng	0.00
23) Nitrobenzene-d5	8.787	82	215181	66.79	ng	0.00
42) 2,4,6-Tribromophenol	15.798	330	177715	105.42	ng	0.00
45) 2-Fluorobiphenyl	12.899	172	432126	63.47	ng	0.00
79) Terphenyl-d14	19.639	244	1613092	111.28	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.840	94	6662	2.58	ng	# 30
50) Dimethylphthalate	13.757	163	13626	2.07	ng	# 98
71) Phenanthrene	17.098	178	50684	4.74	ng	99
74) Di-n-butylphthalate	18.028	149	25792	2.47	ng	# 94
75) Fluoranthene	19.081	202	102112	6.93	ng	97
78) Pyrene	19.433	202	103612	6.96	ng	99
80) Butylbenzylphthalate	20.316	149	49461	10.37	ng	# 85
81) Benzo(a)anthracene	21.145	228	134659	7.97	ng	97
83) Chrysene	21.198	228	131633	7.99	ng	99
84) Bis(2-ethylhexyl)phtha...	21.075	149	290026	39.11	ng	97
85) Di-n-octyl phthalate	21.945	149	50364	4.00	ng	# 88
88) Benzo(b)fluoranthene	22.733	252	87413	4.72	ng	99
89) Benzo(k)fluoranthene	22.774	252	56990	3.22	ng	94
90) Benzo(a)pyrene	23.316	252	64007	3.79	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP043021\
Data File : BP005432.D
Acq On : 30 Apr 2021 14:04
Operator : CG/JU
Sample : M2229-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
TP-6

Quant Time: Apr 30 14:32:42 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP042921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 30 11:27:53 2021
Response via : Initial Calibration

