

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050125\  
 Data File : BP024486.D  
 Acq On : 01 May 2025 10:37  
 Operator : RC/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040

Manual Integrations  
 APPROVED

Reviewed By :Rahul  
 Chavli

Quant Time: May 01 12:08:53 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP041425.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Apr 18 12:04:48 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |                                   |
|-------------------------------|--------|------|----------|--------|-------|----------|-----------------------------------|
| Internal Standards            |        |      |          |        |       |          | 05/02/2025                        |
| 1) 1,4-Dichlorobenzene-d4     | 7.711  | 152  | 147752   | 20.000 | ng    | -0.01    | Supervised By :Jagrut<br>Upadhyay |
| 21) Naphthalene-d8            | 10.481 | 136  | 596585   | 20.000 | ng    | #-0.02   |                                   |
| 39) Acenaphthene-d10          | 14.346 | 164  | 373130   | 20.000 | ng    | 0.00     |                                   |
| 64) Phenanthrene-d10          | 17.145 | 188  | 737949   | 20.000 | ng    | # 0.00   |                                   |
| 76) Chrysene-d12              | 21.592 | 240  | 820131   | 20.000 | ng    | 0.00     |                                   |
| 86) Perylene-d12              | 24.939 | 264  | 979134   | 20.000 | ng    | 0.00     | 05/02/2025                        |
| System Monitoring Compounds   |        |      |          |        |       |          |                                   |
| 5) 2-Fluorophenol             | 5.334  | 112  | 673943   | 75.419 | ng    | -0.01    |                                   |
| 7) Phenol-d6                  | 6.899  | 99   | 882966   | 72.162 | ng    | -0.01    |                                   |
| 23) Nitrobenzene-d5           | 8.863  | 82   | 811892   | 77.600 | ng    | 0.00     |                                   |
| 42) 2,4,6-Tribromophenol      | 15.863 | 330  | 491410   | 95.189 | ng    | 0.00     |                                   |
| 45) 2-Fluorobiphenyl          | 12.951 | 172  | 1973498  | 80.981 | ng    | -0.01    |                                   |
| 79) Terphenyl-d14             | 19.869 | 244  | 3273021  | 80.441 | ng    | 0.00     |                                   |
| Target Compounds              |        |      |          |        |       |          | Qvalue                            |
| 2) 1,4-Dioxane                | 3.281  | 88   | 129194   | 34.181 | ng    |          | 96                                |
| 3) Pyridine                   | 3.676  | 79   | 331009   | 33.165 | ng    |          | 93                                |
| 4) n-Nitrosodimethylamine     | 3.581  | 42   | 136126   | 41.096 | ng    | #        | 74                                |
| 6) Aniline                    | 7.052  | 93   | 392841   | 35.929 | ng    |          | 98                                |
| 8) 2-Chlorophenol             | 7.287  | 128  | 375237   | 37.611 | ng    |          | 98                                |
| 9) Benzaldehyde               | 6.864  | 77   | 245883   | 40.624 | ng    |          | 96                                |
| 10) Phenol                    | 6.922  | 94   | 435521   | 35.482 | ng    |          | 90                                |
| 11) bis(2-Chloroethyl)ether   | 7.146  | 93   | 334846   | 33.859 | ng    |          | 100                               |
| 12) 1,3-Dichlorobenzene       | 7.605  | 146  | 404068   | 37.620 | ng    |          | 100                               |
| 13) 1,4-Dichlorobenzene       | 7.746  | 146  | 410261   | 37.646 | ng    |          | 99                                |
| 14) 1,2-Dichlorobenzene       | 8.058  | 146  | 392890   | 37.336 | ng    |          | 98                                |
| 15) Benzyl Alcohol            | 7.952  | 79   | 312145   | 39.447 | ng    |          | 97                                |
| 16) 2,2'-oxybis(1-Chloropr... | 8.234  | 45   | 361149   | 33.785 | ng    |          | 99                                |
| 17) 2-Methylphenol            | 8.158  | 107  | 295423   | 37.205 | ng    |          | 97                                |
| 18) Hexachloroethane          | 8.775  | 117  | 151714   | 38.522 | ng    |          | 96                                |
| 19) n-Nitroso-di-n-propyla... | 8.511  | 70   | 265410   | 35.062 | ng    |          | 94                                |
| 20) 3+4-Methylphenols         | 8.481  | 107  | 414549   | 37.626 | ng    |          | 98                                |
| 22) Acetophenone              | 8.522  | 105  | 566686   | 38.378 | ng    | #        | 97                                |
| 24) Nitrobenzene              | 8.899  | 77   | 390213   | 37.701 | ng    |          | 99                                |
| 25) Isophorone                | 9.422  | 82   | 713636   | 37.683 | ng    |          | 97                                |
| 26) 2-Nitrophenol             | 9.605  | 139  | 220031   | 43.441 | ng    |          | 96                                |
| 27) 2,4-Dimethylphenol        | 9.669  | 122  | 250979   | 39.463 | ng    |          | 98                                |
| 28) bis(2-Chloroethoxy)met... | 9.893  | 93   | 450730   | 35.232 | ng    |          | 100                               |
| 29) 2,4-Dichlorophenol        | 10.146 | 162  | 355334   | 40.977 | ng    |          | 98                                |
| 30) 1,2,4-Trichlorobenzene    | 10.346 | 180  | 375943   | 40.463 | ng    |          | 99                                |
| 31) Naphthalene               | 10.528 | 128  | 1212817  | 38.593 | ng    |          | 99                                |
| 32) Benzoic acid              | 9.828  | 122  | 317100m  | 42.790 | ng    |          |                                   |
| 33) 4-Chloroaniline           | 10.646 | 127  | 446104   | 40.310 | ng    |          | 98                                |
| 34) Hexachlorobutadiene       | 10.816 | 225  | 237029   | 43.415 | ng    |          | 99                                |
| 35) Caprolactam               | 11.440 | 113  | 127906   | 39.975 | ng    |          | 96                                |
| 36) 4-Chloro-3-methylphenol   | 11.781 | 107  | 411730   | 40.764 | ng    |          | 99                                |
| 37) 2-Methylnaphthalene       | 12.146 | 142  | 834151   | 38.274 | ng    |          | 97                                |
| 38) 1-Methylnaphthalene       | 12.369 | 142  | 821873   | 38.548 | ng    |          | 99                                |
| 40) 1,2,4,5-Tetrachloroben... | 12.516 | 216  | 426228   | 41.538 | ng    |          | 99                                |
| 41) Hexachlorocyclopentadiene | 12.493 | 237  | 148794   | 41.000 | ng    |          | 99                                |
| 43) 2,4,6-Trichlorophenol     | 12.763 | 196  | 291945   | 42.794 | ng    |          | 98                                |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050125\  
 Data File : BP024486.D  
 Acq On : 01 May 2025 10:37  
 Operator : RC/JU  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SSTDCCC040

Manual Integrations  
 APPROVED


Reviewed By :Rahul  
 Chavli

Quant Time: May 01 12:08:53 2025  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP041425.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Apr 18 12:04:48 2025  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |                                |
|-------------------------------|--------|------|----------|--------|-------|----------|--------------------------------|
| 44) 2,4,5-Trichlorophenol     | 12.846 | 196  | 318969   | 42.222 | ng    | 98       | 05/02/2025                     |
| 46) 1,1'-Biphenyl             | 13.163 | 154  | 1082855  | 39.481 | ng    | 97       | Supervised By :Jagrut Upadhyay |
| 47) 2-Chloronaphthalene       | 13.204 | 162  | 812334   | 39.629 | ng    | 98       |                                |
| 48) 2-Nitroaniline            | 13.416 | 65   | 244988   | 42.607 | ng    | 99       |                                |
| 49) Acenaphthylene            | 14.063 | 152  | 1249284  | 38.709 | ng    | 99       |                                |
| 50) Dimethylphthalate         | 13.799 | 163  | 1077582  | 39.712 | ng    | 100      |                                |
| 51) 2,6-Dinitrotoluene        | 13.916 | 165  | 234942   | 40.774 | ng    | 95       | 05/02/2025                     |
| 52) Acenaphthene              | 14.404 | 154  | 776175   | 37.935 | ng    | 99       |                                |
| 53) 3-Nitroaniline            | 14.257 | 138  | 245201   | 40.489 | ng    | 96       |                                |
| 54) 2,4-Dinitrophenol         | 14.469 | 184  | 145747   | 42.941 | ng    | 95       |                                |
| 55) Dibenzofuran              | 14.746 | 168  | 1272161  | 38.039 | ng    | 96       |                                |
| 56) 4-Nitrophenol             | 14.593 | 139  | 208593   | 39.882 | ng    | 92       |                                |
| 57) 2,4-Dinitrotoluene        | 14.716 | 165  | 334668   | 42.578 | ng    | 96       |                                |
| 58) Fluorene                  | 15.404 | 166  | 1008200  | 38.942 | ng    | 100      |                                |
| 59) 2,3,4,6-Tetrachlorophenol | 14.987 | 232  | 288836   | 41.453 | ng    | 99       |                                |
| 60) Diethylphthalate          | 15.181 | 149  | 1119859  | 40.048 | ng    | 99       |                                |
| 61) 4-Chlorophenyl-phenyle... | 15.404 | 204  | 506457   | 40.284 | ng    | 99       |                                |
| 62) 4-Nitroaniline            | 15.440 | 138  | 265592   | 41.428 | ng    | 90       |                                |
| 63) Azobenzene                | 15.698 | 77   | 958644   | 37.300 | ng    | 96       |                                |
| 65) 4,6-Dinitro-2-methylph... | 15.498 | 198  | 196422   | 42.219 | ng    | 96       |                                |
| 66) n-Nitrosodiphenylamine    | 15.622 | 169  | 862970   | 39.179 | ng    | 99       |                                |
| 67) 4-Bromophenyl-phenylether | 16.316 | 248  | 335283   | 41.543 | ng    | 100      |                                |
| 68) Hexachlorobenzene         | 16.440 | 284  | 420624   | 43.859 | ng    | 99       |                                |
| 69) Atrazine                  | 16.598 | 200  | 223377   | 39.819 | ng    | 99       |                                |
| 70) Pentachlorophenol         | 16.792 | 266  | 293539   | 43.874 | ng    | 99       |                                |
| 71) Phenanthrene              | 17.187 | 178  | 1550769  | 38.522 | ng    | 100      |                                |
| 72) Anthracene                | 17.281 | 178  | 1549476  | 39.935 | ng    | 99       |                                |
| 73) Carbazole                 | 17.563 | 167  | 1482653  | 39.109 | ng    | 99       |                                |
| 74) Di-n-butylphthalate       | 18.151 | 149  | 1947580  | 41.235 | ng    | 100      |                                |
| 75) Fluoranthene              | 19.275 | 202  | 1916893  | 40.036 | ng    | 96       |                                |
| 77) Benzidine                 | 19.469 | 184  | 631788   | 60.773 | ng    | 99       |                                |
| 78) Pyrene                    | 19.651 | 202  | 1964755  | 37.697 | ng    | 99       |                                |
| 80) Butylbenzylphthalate      | 20.604 | 149  | 923015   | 41.346 | ng    | 94       |                                |
| 81) Benzo(a)anthracene        | 21.575 | 228  | 1961943  | 38.487 | ng    | 100      |                                |
| 82) 3,3'-Dichlorobenzidine    | 21.498 | 252  | 777836   | 43.474 | ng    | 99       |                                |
| 83) Chrysene                  | 21.645 | 228  | 1881857  | 38.533 | ng    | 100      |                                |
| 84) Bis(2-ethylhexyl)phtha... | 21.510 | 149  | 1332713  | 40.723 | ng    | 99       |                                |
| 85) Di-n-octyl phthalate      | 22.786 | 149  | 2299618  | 43.223 | ng    | 100      |                                |
| 87) Indeno(1,2,3-cd)pyrene    | 28.751 | 276  | 2727671m | 40.127 | ng    |          |                                |
| 88) Benzo(b)fluoranthene      | 23.880 | 252  | 2206536  | 37.596 | ng    | #        | 97                             |
| 89) Benzo(k)fluoranthene      | 23.957 | 252  | 2146431  | 37.862 | ng    | #        | 97                             |
| 90) Benzo(a)pyrene            | 24.792 | 252  | 1971970  | 38.952 | ng    | #        | 97                             |
| 91) Dibenzo(a,h)anthracene    | 28.821 | 278  | 2241033  | 39.719 | ng    |          | 96                             |
| 92) Benzo(g,h,i)perylene      | 29.933 | 276  | 2266414  | 39.464 | ng    | #        | 94                             |

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#1  
1,4-Dichlorobenzene-d4  
Concen: 20.000 ng  
RT: 7.711 min Scan# 71  
Delta R.T. -0.011 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion:152 Resp: 147752  
Ion Ratio Lower Upper  
152 100  
150 154.7 124.9 187.3  
115 60.5 47.7 71.5


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#2  
1,4-Dioxane  
Concen: 34.181 ng  
RT: 3.281 min Scan# 33  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37




Tgt Ion: 88 Resp: 129194  
Ion Ratio Lower Upper  
88 100  
58 66.6 51.4 77.2  
43 25.5 17.1 25.7





#3  
Pyridine  
Concen: 33.165 ng  
RT: 3.676 min Scan# 103  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion: 79 Resp: 331009  
Ion Ratio Lower Upper  
79 100  
52 61.6 45.4 68.2  
51 32.1 22.6 33.8


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#4  
n-Nitrosodimethylamine  
Concen: 41.096 ng  
RT: 3.581 min Scan# 84  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion: 42 Resp: 136126  
Ion Ratio Lower Upper  
42 100  
74 137.5 139.5 209.3#  
44 6.5 5.8 8.8





#5  
2-Fluorophenol  
Concen: 75.419 ng  
RT: 5.334 min Scan# 385  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:112 Resp: 67394  
Ion Ratio Lower Upper  
112 100  
64 51.6 41.1 61.7  
63 27.7 20.6 30.8



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#6  
Aniline  
Concen: 35.929 ng  
RT: 7.052 min Scan# 674  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion: 93 Resp: 392841  
Ion Ratio Lower Upper  
93 100  
66 37.8 29.1 43.7  
65 18.5 14.7 22.1





#7  
Phenol-d6  
Concen: 72.162 ng  
RT: 6.899 min Scan# 648  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion: 99 Resp: 882960  
Ion Ratio Lower Upper  
99 100  
42 15.7 10.6 16.0  
71 34.5 23.7 35.5



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#8  
2-Chlorophenol  
Concen: 37.611 ng  
RT: 7.287 min Scan# 714  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37


Tgt Ion:128 Resp: 375237  
Ion Ratio Lower Upper  
128 100  
130 32.9 12.9 52.9  
64 40.3 22.7 62.7





#9  
Benzaldehyde  
Concen: 40.624 ng  
RT: 6.864 min Scan# 64  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion: 77 Resp: 24588  
Ion Ratio Lower Upper  
77 100  
105 103.6 79.2 119.2  
106 100.2 76.2 116.2


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#10  
Phenol  
Concen: 35.482 ng  
RT: 6.922 min Scan# 652  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion: 94 Resp: 435521  
Ion Ratio Lower Upper  
94 100  
65 27.7 4.5 44.5  
66 39.1 12.4 52.4






#11  
bis(2-Chloroethyl)ether  
Concen: 33.859 ng  
RT: 7.146 min Scan# 693  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion: 93 Resp: 334840  
Ion Ratio Lower Upper  
93 100  
63 68.6 48.6 88.6  
95 32.6 12.3 52.3


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#12  
1,3-Dichlorobenzene  
Concen: 37.620 ng  
RT: 7.605 min Scan# 768  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion:146 Resp: 404068  
Ion Ratio Lower Upper  
146 100  
148 64.2 51.4 77.0  
75 29.4 22.9 34.3





#13

1,4-Dichlorobenzene

Concen: 37.646 ng

RT: 7.746 min Scan# 795

Delta R.T. -0.012 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040

Tgt Ion:146 Resp: 41026

Ion Ratio Lower Upper

146 100

148 64.5 51.1 76.7



111 42.0 32.6 48.8

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli





05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#14

1,2-Dichlorobenzene

Concen: 37.336 ng

RT: 8.058 min Scan# 845

Delta R.T. -0.017 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37




Tgt Ion:146 Resp: 392890


Ion Ratio Lower Upper

146 100

148 62.5 51.0 76.4


111 43.5 33.8 50.8





#15  
Benzyl Alcohol  
Concen: 39.447 ng  
RT: 7.952 min Scan# 81  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion: 79 Resp: 31214  
Ion Ratio Lower Upper  
79 100  
108 78.1 65.1 97.7  
77 61.6 50.7 76.1


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#16  
2,2'-oxybis(1-Chloropropane)  
Concen: 33.785 ng  
RT: 8.234 min Scan# 875  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion: 45 Resp: 361149  
Ion Ratio Lower Upper  
45 100  
77 17.4 0.0 37.7  
79 12.9 0.0 33.5






#17  
2-Methylphenol  
Concen: 37.205 ng  
RT: 8.158 min Scan# 80  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :


SSTDCCC040



Tgt Ion:107 Resp: 29542  
Ion Ratio Lower Upper  
107 100  
108 109.3 89.8 134.6  
77 44.8 34.2 51.4  
79 45.1 33.8 50.8


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#18  
Hexachloroethane  
Concen: 38.522 ng  
RT: 8.775 min Scan# 967  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:117 Resp: 151714  
Ion Ratio Lower Upper  
117 100  
119 97.6 77.8 116.6  
201 101.5 75.9 113.9






#19  
n-Nitroso-di-n-propylamine  
Concen: 35.062 ng  
RT: 8.511 min Scan# 917  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion: 70 Resp: 265410  
Ion Ratio Lower Upper  
70 100  
42 51.0 37.3 55.9  
101 11.3 8.2 12.4  
130 26.0 18.6 28.0


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#20  
3+4-Methylphenols  
Concen: 37.626 ng  
RT: 8.481 min Scan# 917  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion:107 Resp: 414549  
Ion Ratio Lower Upper  
107 100  
108 88.4 69.8 109.8  
77 32.7 11.5 51.5  
79 28.3 7.8 47.8





#21  
Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.481 min Scan# 1257  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:136 Resp: 596585  
Ion Ratio Lower Upper  
136 100  
137 10.8 9.2 13.8  
54 6.4 6.0 9.0  
68 5.3 5.5 8.3#



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#22  
Acetophenone  
Concen: 38.378 ng  
RT: 8.522 min Scan# 924  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:105 Resp: 566686  
Ion Ratio Lower Upper  
105 100  
71 3.0 1.7 2.5#  
51 23.2 16.9 25.3  
120 22.9 18.6 28.0





#23  
Nitrobenzene-d5  
Concen: 77.600 ng  
RT: 8.863 min Scan# 982  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 811892  
Ion Ratio Lower Upper  
82 100  
128 48.1 36.5 54.7  
54 44.7 33.4 50.2



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#24  
Nitrobenzene  
Concen: 37.701 ng  
RT: 8.899 min Scan# 988  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion: 77 Resp: 390213  
Ion Ratio Lower Upper  
77 100  
123 49.4 39.2 58.8  
65 13.6 9.8 14.6





#25  
Isophorone  
Concen: 37.683 ng  
RT: 9.422 min Scan# 1080  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 713630  
Ion Ratio Lower Upper  
82 100  
95 7.9 5.4 8.2  
138 19.5 14.4 21.6



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#26  
2-Nitrophenol  
Concen: 43.441 ng  
RT: 9.605 min Scan# 1108  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:139 Resp: 220031  
Ion Ratio Lower Upper  
139 100  
109 31.5 21.8 32.6  
65 44.5 36.2 54.2






#27  
2,4-Dimethylphenol  
Concen: 39.463 ng  
RT: 9.669 min Scan# 1119  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion:122 Resp: 250979  
Ion Ratio Lower Upper  
122 100  
107 119.8 93.7 140.5  
121 60.4 48.2 72.2


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#28  
bis(2-Chloroethoxy)methane  
Concen: 35.232 ng  
RT: 9.893 min Scan# 1157  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion: 93 Resp: 450730  
Ion Ratio Lower Upper  
93 100  
95 32.5 26.0 39.0  
123 11.9 9.1 13.7







#29  
2,4-Dichlorophenol  
Concen: 40.977 ng  
RT: 10.146 min Scan# 1200  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040

Tgt Ion:162 Resp: 355334  
Ion Ratio Lower Upper  
162 100  
164 64.1 43.3 83.3  
98 34.7 16.6 56.6



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#30  
1,2,4-Trichlorobenzene  
Concen: 40.463 ng  
RT: 10.346 min Scan# 1234  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37


Tgt Ion:180 Resp: 375943  
Ion Ratio Lower Upper  
180 100  
182 96.2 77.3 115.9  
145 32.1 25.1 37.7





#31  
Naphthalene  
Concen: 38.593 ng  
RT: 10.528 min Scan# 1265  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:128 Resp: 121281  
Ion Ratio Lower Upper  
128 100  
129 11.1 8.6 13.0  
127 13.2 10.3 15.5


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#32  
Benzoic acid  
Concen: 42.790 ng m  
RT: 9.828 min Scan# 1146  
Delta R.T. 0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:122 Resp: 317100  
Ion Ratio Lower Upper  
122 100  
105 121.5 104.8 144.8  
77 86.6 69.2 109.2






#33  
4-Chloroaniline  
Concen: 40.310 ng  
RT: 10.646 min Scan# 1285  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion:127 Resp: 446104  
Ion Ratio Lower Upper  
127 100  
129 31.7 26.1 39.1  
65 28.3 22.1 33.1  
92 19.5 15.0 22.4


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#34  
Hexachlorobutadiene  
Concen: 43.415 ng  
RT: 10.816 min Scan# 1314  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37


Tgt Ion:225 Resp: 237029  
Ion Ratio Lower Upper  
225 100  
223 62.6 49.5 74.3  
227 65.0 50.9 76.3





#35  
Caprolactam  
Concen: 39.975 ng  
RT: 11.440 min Scan# 1422  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:113 Resp: 127900  
Ion Ratio Lower Upper  
113 100  
55 129.5 115.9 155.9  
56 103.3 81.1 121.1


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#36  
4-Chloro-3-methylphenol  
Concen: 40.764 ng  
RT: 11.781 min Scan# 1478  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37




Tgt Ion:107 Resp: 411730  
Ion Ratio Lower Upper  
107 100  
144 26.2 20.5 30.7  
142 80.1 63.4 95.2





#37  
2-Methylnaphthalene  
Concen: 38.274 ng  
RT: 12.146 min Scan# 1543  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:142 Resp: 83415  
Ion Ratio Lower Upper  
142 100  
141 89.3 70.3 105.5  
115 36.5 26.9 40.3


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#38  
1-Methylnaphthalene  
Concen: 38.548 ng  
RT: 12.369 min Scan# 1578  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion:142 Resp: 821873  
Ion Ratio Lower Upper  
142 100  
141 92.7 73.6 110.4  
116 3.9 3.0 4.6






#39  
Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.346 min Scan# 1917  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNP\_P

ClientSampleId :



SSTDCCC040



Tgt Ion:164 Resp: 373130  
Ion Ratio Lower Upper  
164 100  
162 100.2 80.6 120.8  
160 44.1 35.3 52.9


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#40  
1,2,4,5-Tetrachlorobenzene  
Concen: 41.538 ng  
RT: 12.516 min Scan# 1603  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37




Tgt Ion:216 Resp: 426228  
Ion Ratio Lower Upper  
216 100  
214 78.4 62.6 93.8  
179 21.9 17.2 25.8  
108 18.2 15.0 22.6





#41  
Hexachlorocyclopentadiene  
Concen: 41.000 ng  
RT: 12.493 min Scan# 11  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion: 237 Resp: 148794  
Ion Ratio Lower Upper  
237 100  
235 62.8 43.9 83.9  
272 13.8 0.0 33.6


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#42  
2,4,6-Tribromophenol  
Concen: 95.189 ng  
RT: 15.863 min Scan# 2172  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion: 330 Resp: 491410  
Ion Ratio Lower Upper  
330 100  
332 96.8 77.1 115.7  
141 28.3 24.7 37.1





#43

2,4,6-Trichlorophenol

Concen: 42.794 ng

RT: 12.763 min Scan# 1645

Delta R.T. -0.012 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040

Tgt Ion:196 Resp: 29194

Ion Ratio Lower Upper

196 100



198 95.1 74.3 111.5

200 30.6 24.0 36.0


Manual Integrations

APPROVED


Reviewed By :Rahul  
Chavli



05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025



#44

2,4,5-Trichlorophenol

Concen: 42.222 ng

RT: 12.846 min Scan# 1659

Delta R.T. -0.012 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37

Tgt Ion:196 Resp: 318969




Ion Ratio Lower Upper


196 100

198 95.2 77.3 115.9

97 48.9 40.4 60.6


132 28.1 23.4 35.0





#45  
2-Fluorobiphenyl  
Concen: 80.981 ng  
RT: 12.951 min Scan# 1677  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:172 Resp: 1973498  
Ion Ratio Lower Upper  
172 100  
171 35.3 28.2 42.2  
170 23.5 18.6 27.8


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#46  
1,1'-Biphenyl  
Concen: 39.481 ng  
RT: 13.163 min Scan# 1713  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion:154 Resp: 1082855  
Ion Ratio Lower Upper  
154 100  
153 41.2 20.3 60.3  
76 11.0 0.0 33.5






#47  
2-Chloronaphthalene  
Concen: 39.629 ng  
RT: 13.204 min Scan# 1720  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion:162 Resp: 812334  
Ion Ratio Lower Upper  
162 100  
127 41.9 31.9 47.9  
164 32.6 26.2 39.4


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#48  
2-Nitroaniline  
Concen: 42.607 ng  
RT: 13.416 min Scan# 1756  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion: 65 Resp: 244988  
Ion Ratio Lower Upper  
65 100  
92 74.9 62.3 93.5  
138 121.5 97.2 145.8






#49  
Acenaphthylene  
Concen: 38.709 ng  
RT: 14.063 min Scan# 1866  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :


SSTDCCC040



Tgt Ion:152 Resp: 1249284  
Ion Ratio Lower Upper  
152 100  
151 20.3 15.9 23.9  
153 12.9 10.6 15.8


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#50  
Dimethylphthalate  
Concen: 39.712 ng  
RT: 13.799 min Scan# 1821  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37




Tgt Ion:163 Resp: 1077582  
Ion Ratio Lower Upper  
163 100  
194 4.2 3.5 5.3  
164 10.1 8.0 12.0





#51  
2,6-Dinitrotoluene  
Concen: 40.774 ng  
RT: 13.916 min Scan# 1841  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:165 Resp: 23494  
Ion Ratio Lower Upper  
165 100  
63 52.2 38.0 57.0  
89 54.8 41.7 62.5


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#52  
Acenaphthene  
Concen: 37.935 ng  
RT: 14.404 min Scan# 1924  
Delta R.T. -0.017 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Tgt Ion:154 Resp: 776175  
Ion Ratio Lower Upper  
154 100  
153 113.9 89.8 134.6  
152 53.6 42.6 63.8





#53  
3-Nitroaniline  
Concen: 40.489 ng  
RT: 14.257 min Scan# 1899  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:138 Resp: 24520  
Ion Ratio Lower Upper  
138 100  
108 10.3 8.3 12.5  
92 108.0 89.8 134.8



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#54  
2,4-Dinitrophenol  
Concen: 42.941 ng  
RT: 14.469 min Scan# 1935  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:184 Resp: 145747  
Ion Ratio Lower Upper  
184 100  
63 55.2 42.6 64.0  
154 65.0 47.2 70.8






#55  
Dibenzenofuran  
Concen: 38.039 ng  
RT: 14.746 min Scan# 1982  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :


SSTDCCC040



Tgt Ion:168 Resp: 127216  
Ion Ratio Lower Upper  
168 100  
139 40.9 30.3 45.5  
169 13.4 10.6 16.0


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#56  
4-Nitrophenol  
Concen: 39.882 ng  
RT: 14.593 min Scan# 1956  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:139 Resp: 208593  
Ion Ratio Lower Upper  
139 100  
109 76.5 46.1 86.1  
65 89.3 65.0 105.0





#57  
2,4-Dinitrotoluene  
Concen: 42.578 ng  
RT: 14.716 min Scan# 1980  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:165 Resp: 334668  
Ion Ratio Lower Upper  
165 100  
63 38.7 30.4 45.6  
89 63.9 54.6 82.0  
182 3.1 2.2 3.2



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#58  
Fluorene  
Concen: 38.942 ng  
RT: 15.404 min Scan# 2094  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:166 Resp: 1008200  
Ion Ratio Lower Upper  
166 100  
165 97.4 78.1 117.1  
167 13.5 10.9 16.3





#59  
2,3,4,6-Tetrachlorophenol  
Concen: 41.453 ng  
RT: 14.987 min Scan# 2023  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNL\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:232 Resp: 288830  
Ion Ratio Lower Upper  
232 100  
131 46.9 36.8 55.2  
130 2.6 1.9 2.9  
166 31.4 25.0 37.6



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#60  
Diethylphthalate  
Concen: 40.048 ng  
RT: 15.181 min Scan# 2056  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:149 Resp: 1119859  
Ion Ratio Lower Upper  
149 100  
177 22.2 17.4 26.2  
150 12.7 10.1 15.1





#61  
4-Chlorophenyl-phenylether  
Concen: 40.284 ng  
RT: 15.404 min Scan# 2098  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:204 Resp: 50645  
Ion Ratio Lower Upper  
204 100  
206 33.5 26.5 39.7  
141 62.6 49.4 74.0



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#62  
4-Nitroaniline  
Concen: 41.428 ng  
RT: 15.440 min Scan# 2100  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37


Tgt Ion:138 Resp: 265592  
Ion Ratio Lower Upper  
138 100  
92 47.8 27.8 67.8  
108 82.3 48.9 88.9





#63  
Azobenzene  
Concen: 37.300 ng  
RT: 15.698 min Scan# 2149  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



| Tgt Ion:  | 77   | Resp: | 95864 |
|-----------|------|-------|-------|
| Ion Ratio | 100  | Lower | Upper |
| 182       | 27.1 | 7.3   | 47.3  |
| 105       | 22.4 | 2.3   | 42.3  |
| 51        | 28.5 | 3.4   | 43.4  |


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli




05/02/2025  
Supervised By :Jagrut  
Upadhyay



05/02/2025




#64  
Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.145 min Scan# 2390  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37




| Tgt Ion:  | 188 | Resp: | 737949 |
|-----------|-----|-------|--------|
| Ion Ratio | 100 | Lower | Upper  |
| 94        | 8.6 | 8.6   | 13.0   |
| 80        | 9.5 | 9.8   | 14.6   |





#65  
4,6-Dinitro-2-methylphenol  
Concen: 42.219 ng  
RT: 15.498 min Scan# 2110  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37



Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:198 Resp: 19642  
Ion Ratio Lower Upper  
198 100  
51 37.8 18.4 58.4  
105 42.5 26.5 66.5


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#66  
n-Nitrosodiphenylamine  
Concen: 39.179 ng  
RT: 15.622 min Scan# 2131  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:169 Resp: 862970  
Ion Ratio Lower Upper  
169 100  
168 67.2 53.0 79.4  
167 36.2 28.9 43.3






#67  
4-Bromophenyl-phenylether  
Concen: 41.543 ng  
RT: 16.316 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion:248 Resp: 33528  
Ion Ratio Lower Upper  
248 100  
250 97.2 78.2 117.4  
141 72.4 57.7 86.5


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#68  
Hexachlorobenzene  
Concen: 43.859 ng  
RT: 16.440 min Scan# 2270  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:284 Resp: 420624  
Ion Ratio Lower Upper  
284 100  
142 29.3 24.6 37.0  
249 29.6 23.8 35.6





#69

Atrazine

Concen: 39.819 ng

RT: 16.598 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:200 Resp: 22337  
 Ion Ratio Lower Upper  
 200 100  
 173 27.4 7.5 47.5  
 215 48.3 28.8 68.8



Manual Integrations  
 APPROVED

Reviewed By :Rahul  
 Chavli



05/02/2025  
 Supervised By :Jagrut  
 Upadhyay

05/02/2025



#70

Pentachlorophenol

Concen: 43.874 ng




RT: 16.792 min Scan# 2330


Delta R.T. -0.006 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37

Tgt Ion:266 Resp: 293539  
 Ion Ratio Lower Upper  
 266 100  
 268 64.2 51.0 76.6  
 264 63.0 49.8 74.6





#71

Phenanthrene

Concen: 38.522 ng

RT: 17.187 min Scan# 2401

Delta R.T. -0.006 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :

SSTDCCC040

Tgt Ion:178 Resp: 1550769

Ion Ratio Lower Upper

178 100

176 19.2 15.4 23.0



179 15.2 12.5 18.7

Manual Integrations

APPROVED

Reviewed By :Rahul

Chavli





05/02/2025

Supervised By :Jagrut

Upadhyay

05/02/2025



#72

Anthracene

Concen: 39.935 ng

RT: 17.281 min Scan# 2413

Delta R.T. -0.006 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37




Tgt Ion:178 Resp: 1549476


Ion Ratio Lower Upper

178 100

176 18.8 14.7 22.1

179 15.3 12.4 18.6






#73  
 Carbazole  
 Concen: 39.109 ng  
 RT: 17.563 min Scan# 2461  
 Delta R.T. -0.006 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :


SSTDCCC040



Tgt Ion:167 Resp: 148265  
 Ion Ratio Lower Upper  
 167 100  
 166 21.8 17.1 25.7  
 139 13.9 10.6 16.0


Manual Integrations  
 APPROVED

Reviewed By :Rahul  
 Chavli




05/02/2025  
 Supervised By :Jagrut  
 Upadhyay



05/02/2025




#74  
 Di-n-butylphthalate  
 Concen: 41.235 ng  
 RT: 18.151 min Scan# 2561  
 Delta R.T. -0.006 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37



Tgt Ion:149 Resp: 1947580  
 Ion Ratio Lower Upper  
 149 100  
 150 9.4 7.4 11.2  
 104 5.0 3.9 5.9






#75  
Fluoranthene  
Concen: 40.036 ng  
RT: 19.275 min Scan# 21  
Delta R.T. -0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040



Tgt Ion:202 Resp: 191689  
Ion Ratio Lower Upper  
202 100  
101 9.7 0.0 33.2  
203 17.4 0.0 37.7


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#76  
Chrysene-d12  
Concen: 20.000 ng  
RT: 21.592 min Scan# 3146  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37


Tgt Ion:240 Resp: 820131  
Ion Ratio Lower Upper  
240 100  
120 10.3 9.8 14.8  
236 25.4 20.9 31.3





#77  
Benzidine  
Concen: 60.773 ng  
RT: 19.469 min Scan# 21  
Delta R.T. -0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:184 Resp: 631788  
Ion Ratio Lower Upper  
184 100  
185 14.4 11.6 17.4  
183 12.2 9.2 13.8


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#78  
Pyrene  
Concen: 37.697 ng  
RT: 19.651 min Scan# 2816  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:202 Resp: 1964755  
Ion Ratio Lower Upper  
202 100  
200 21.1 17.2 25.8  
203 17.5 14.4 21.6






#79  
 Terphenyl-d14  
 Concen: 80.441 ng  
 RT: 19.869 min Scan# 2853  
 Delta R.T. -0.006 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :


SSTDCCC040



Tgt Ion:244 Resp: 327302  
 Ion Ratio Lower Upper  
 244 100  
 212 7.0 5.6 8.4  
 122 9.5 9.4 14.0


Manual Integrations  
 APPROVED

Reviewed By :Rahul  
 Chavli




05/02/2025  
 Supervised By :Jagrut  
 Upadhyay



05/02/2025




#80  
 Butylbenzylphthalate  
 Concen: 41.346 ng  
 RT: 20.604 min Scan# 2978  
 Delta R.T. 0.000 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37



Tgt Ion:149 Resp: 923015  
 Ion Ratio Lower Upper  
 149 100  
 91 64.9 56.4 84.6  
 206 21.9 16.3 24.5





#81  
Benzo(a)anthracene  
Concen: 38.487 ng  
RT: 21.575 min Scan# 3143  
Delta R.T. 0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:228 Resp: 196194  
Ion Ratio Lower Upper  
228 100  
226 27.2 21.8 32.8  
229 19.8 15.9 23.9



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#82  
3,3'-Dichlorobenzidine  
Concen: 43.474 ng  
RT: 21.498 min Scan# 3130  
Delta R.T. 0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:252 Resp: 777836  
Ion Ratio Lower Upper  
252 100  
254 64.0 51.7 77.5  
126 10.1 8.8 13.2





#83

Chrysene

Concen: 38.533 ng

RT: 21.645 min Scan# 31

Delta R.T. 0.006 min

Lab File: BP024486.D

Acq: 01 May 2025 10:37

Instrument :

BNA\_P

ClientSampleId :



SSTDCCC040

Tgt Ion:228 Resp: 188185

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 228 | 100   |       |       |
| 226 | 29.9  | 24.2  | 36.2  |
| 229 | 19.5  | 15.6  | 23.4  |



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli



05/02/2025  
Supervised By :Jagrut  
Upadhyay

05/02/2025



#84

Bis(2-ethylhexyl)phthalate

Concen: 40.723 ng

RT: 21.510 min Scan# 3132




Delta R.T. 0.000 min


Lab File: BP024486.D

Acq: 01 May 2025 10:37

Tgt Ion:149 Resp: 1332713


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 149 | 100   |       |       |
| 167 | 28.2  | 22.4  | 33.6  |
| 279 | 5.6   | 4.0   | 6.0   |





#85  
Di-n-octyl phthalate  
Concen: 43.223 ng  
RT: 22.786 min Scan# 311  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37


Instrument :  
BNA\_P  
ClientSampleId :  
SSTDCCC040



Tgt Ion:149 Resp: 2299618  
Ion Ratio Lower Upper  
149 100  
167 1.6 1.2 1.8  
43 7.5 6.0 9.0


Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#86  
Perylene-d12  
Concen: 20.000 ng  
RT: 24.939 min Scan# 3715  
Delta R.T. 0.006 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:264 Resp: 979134  
Ion Ratio Lower Upper  
264 100  
260 23.4 18.9 28.3  
265 22.1 17.8 26.6





#87  
 Indeno(1,2,3-cd)pyrene  
 Concen: 40.127 ng m  
 RT: 28.751 min Scan# 41  
 Delta R.T. 0.012 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37



Instrument : BNA\_P  
 ClientSampleId : SSTDCCC040

Tgt Ion: 276 Resp: 272767

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 276 | 100   |       |       |
| 138 | 14.4  | 14.5  | 21.7  |
| 277 | 20.8  | 20.2  | 30.2  |



Manual Integrations  
**APPROVED**

Reviewed By :Rahul Chavli



05/02/2025  
 Supervised By :Jagrut Upadhyay




05/02/2025




#88  
 Benzo(b)fluoranthene  
 Concen: 37.596 ng  
 RT: 23.880 min Scan# 3535  
 Delta R.T. -0.006 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37

Tgt Ion: 252 Resp: 2206536

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 252 | 100   |       |       |
| 253 | 21.9  | 17.7  | 26.5  |
| 125 | 9.0   | 9.4   | 14.0  |





#89  
Benzo(k)fluoranthene  
Concen: 37.862 ng  
RT: 23.957 min Scan# 3548  
Delta R.T. 0.000 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Instrument :

BNA\_P



ClientSampleId :

SSTDCCC040

Tgt Ion:252 Resp: 214643  
Ion Ratio Lower Upper  
252 100  
253 22.0 17.7 26.5  
125 8.6 9.1 13.7



Manual Integrations  
APPROVED

Reviewed By :Rahul  
Chavli






05/02/2025  
Supervised By :Jagrut  
Upadhyay


05/02/2025



#90  
Benzo(a)pyrene  
Concen: 38.952 ng  
RT: 24.792 min Scan# 3690  
Delta R.T. 0.012 min  
Lab File: BP024486.D  
Acq: 01 May 2025 10:37

Tgt Ion:252 Resp: 1971970  
Ion Ratio Lower Upper  
252 100  
253 22.2 17.6 26.4  
125 10.0 10.1 15.1








#91  
 Dibenzo(a,h)anthracene  
 Concen: 39.719 ng  
 RT: 28.821 min Scan# 41  
 Delta R.T. 0.018 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37

Instrument : BNA\_P  
 ClientSampleId : SSTDCCC040

Tgt Ion: 278 Resp: 224103


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 278 | 100   |       |       |
| 139 | 13.2  | 12.6  | 19.0  |
| 279 | 24.3  | 18.5  | 27.7  |

Manual Integrations  
**APPROVED**  
 Reviewed By :Rahul Chavli



05/02/2025  
 Supervised By :Jagrut Upadhyay

05/02/2025



#92  
 Benzo(g,h,i)perylene  
 Concen: 39.464 ng  
 RT: 29.933 min Scan# 4564  
 Delta R.T. 0.036 min  
 Lab File: BP024486.D  
 Acq: 01 May 2025 10:37

Tgt Ion: 276 Resp: 2266414

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 276 | 100   |       |       |
| 277 | 24.0  | 18.9  | 28.3  |
| 138 | 17.7  | 18.7  | 28.1  |

