

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010201.D
 Acq On : 03 May 2022 09:46
 Operator : CG/JU
 Sample : PB144528BL
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK528

Quant Time: May 04 01:07:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	138186	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	562772	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	356708	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.304	188	736747	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.386	240	660969	20.000	ng/u1	# 0.00
88) Perylene-d12	23.827	264	617725	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	18791	5.930	ng/uL	0.00
4) Pyridine-d5	3.775	84	246413	27.641	ng/u1	0.00
7) Phenol-d5	7.081	99	303940	24.181	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.246	67	218956	29.031	ng/u1	0.00
11) 2-Chlorophenol-d4	7.452	132	256629	27.569	ng/u1	0.00
15) 4-Methylphenol-d8	8.628	113	252352	24.232	ng/u1	0.00
21) Nitrobenzene-d5	9.081	128	130033	29.051	ng/u1	0.00
24) 2-Nitrophenol-d4	9.805	143	139786	28.596	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.346	165	239911	25.210	ng/u1	0.00
31) 4-Chloroaniline-d4	10.857	131	335877	24.615	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	802382	28.917	ng/u1	0.00
49) Acenaphthylene-d8	14.245	160	910132	27.101	ng/u1	0.00
54) 4-Nitrophenol-d4	14.751	143	98890	19.359	ng/u1	0.00
60) Fluorene-d10	15.545	176	697333	29.215	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	95977	20.602	ng/u1	0.00
73) Anthracene-d10	17.404	188	1067708	29.956	ng/u1	0.00
81) Pyrene-d10	19.633	212	1219690	32.448	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.669	264	1025463	31.665	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
Data File : BP010201.D
Acq On : 03 May 2022 09:46
Operator : CG/JU
Sample : PB144528BL
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK528

Quant Time: May 04 01:07:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 03 00:58:32 2022
Response via : Initial Calibration

