




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(3) 1,4-Dioxane-d8 (S)**

**3.417min (+ 0.059) 0.11 ng/uL**

**response 327**

| Ion   | Exp%   | Act%   |
|-------|--------|--------|
| 96.00 | 100.00 | 100.00 |
| 64.00 | 55.40  | 63.53  |
| 34.00 | 0.00   | 0.00   |
| 0.00  | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(3) 1,4-Dioxane-d8 (S)**

**3.358min ( 0.000) 8.22 ng/uL m**

**response 25524**

| Ion   | Exp%   | Act%   |
|-------|--------|--------|
| 96.00 | 100.00 | 100.00 |
| 64.00 | 55.40  | 77.07# |
| 34.00 | 0.00   | 0.00   |
| 0.00  | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(29) 2,4-Dichlorophenol (C)**

**10.481min (+ 0.106) 0.04 ng/ul**

**response 342**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 162.00 | 100.00 | 100.00 |
| 164.00 | 46.30  | 50.80  |
| 98.00  | 38.80  | 33.56  |
| 0.00   | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
Data File : BP010166.D  
Acq On : 02 May 2022 11:14  
Operator : CG/JU  
Sample : SSTDCCC020  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_P  
LabSampleId :  
SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Tue May 03 00:58:32 2022  
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

(29) 2,4-Dichlorophenol (C)

10.375min ( 0.000) 20.54 ng/ul m

response 183474

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 162.00 | 100.00 | 100.00 |
| 164.00 | 46.30  | 65.28# |
| 98.00  | 38.80  | 39.39  |
| 0.00   | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(31) 4-Chloroaniline-d4 (S)**

**10.963min (+ 0.100) 0.04 ng/ul**

**response 585**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 131.00 | 100.00 | 100.00 |
| 133.00 | 29.20  | 30.28  |
| 69.00  | 26.10  | 29.13  |
| 0.00   | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(31) 4-Chloroaniline-d4 (S)**

**10.863min ( 0.000) 18.94 ng/ul m**

**response 252455**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 131.00 | 100.00 | 100.00 |
| 133.00 | 29.20  | 32.93  |
| 69.00  | 26.10  | 19.60# |
| 0.00   | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
Data File : BP010166.D  
Acq On : 02 May 2022 11:14  
Operator : CG/JU  
Sample : SSTDCCC020  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_P  
LabSampleId :  
SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Tue May 03 00:58:32 2022  
Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(63) 4-Nitroaniline**

**15.622min ( 0.000) 12.58 ng/ul**

**response 59752**

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 138.00 | 100.00 | 100.00 |
| 92.00  | 52.00  | 53.65  |
| 108.00 | 122.00 | 83.05# |
| 0.00   | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
Data File : BP010166.D  
Acq On : 02 May 2022 11:14  
Operator : CG/JU  
Sample : SSTDCCC020  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_P  
LabSampleId :  
SSTDCCC020

## Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 05/03/2022  
Supervised By :Yogesh Patel 05/05/2022

Quant Time: May 03 00:58:41 2022  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Tue May 03 00:58:32 2022  
Response via : Initial Calibration



TIC: BP010166.D\data.ms

**(63) 4-Nitroaniline**

15.622min ( 0.000) 12.88 ng/ul m

response 61196

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 138.00 | 100.00 | 100.00 |
| 92.00  | 52.00  | 53.65  |
| 108.00 | 122.00 | 83.05# |
| 0.00   | 0.00   | 0.00   |






Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(86) Bis(2-ethylhexyl)phthalate**

21.257min (-0.041) 0.02 ng/ul

response 647

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 149.00 | 100.00 | 100.00 |
| 167.00 | 41.00  | 33.08  |
| 279.00 | 7.00   | 6.37   |
| 0.00   | 0.00   | 0.00   |




Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010166.D\data.ms

**(86) Bis(2-ethylhexyl)phthalate**

21.298min ( 0.000) 23.90 ng/ul m

response 645118

| Ion    | Exp%   | Act%   |
|--------|--------|--------|
| 149.00 | 100.00 | 100.00 |
| 167.00 | 41.00  | 28.01# |
| 279.00 | 7.00   | 4.46#  |
| 0.00   | 0.00   | 0.00   |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| -----                         |        |      |          |        |        |          |
| Internal Standards            |        |      |          |        |        |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.922  | 152  | 135364   | 20.000 | ng/ul  | 0.00     |
| 20) Naphthalene-d8            | 10.728 | 136  | 549652   | 20.000 | ng/ul  | 0.00     |
| 38) Acenaphthene-d10          | 14.557 | 164  | 347303   | 20.000 | ng/ul  | 0.00     |
| 64) Phenanthrene-d10          | 17.310 | 188  | 711813   | 20.000 | ng/ul  | # 0.00   |
| 79) Chrysene-d12              | 21.392 | 240  | 666755   | 20.000 | ng/ul  | # 0.00   |
| 88) Perylene-d12              | 23.839 | 264  | 645002   | 20.000 | ng/ul  | 0.00     |
| System Monitoring Compounds   |        |      |          |        |        |          |
| 3) 1,4-Dioxane-d8             | 3.358  | 96   | 25524m   | 8.222  | ng/uL  | 0.00     |
| 4) Pyridine-d5                | 3.775  | 84   | 165957   | 19.004 | ng/ul  | 0.00     |
| 7) Phenol-d5                  | 7.087  | 99   | 237697   | 19.305 | ng/ul  | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.252  | 67   | 149729   | 20.266 | ng/ul  | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.458  | 132  | 191825   | 21.037 | ng/ul  | 0.00     |
| 15) 4-Methylphenol-d8         | 8.634  | 113  | 187868   | 18.416 | ng/ul  | 0.00     |
| 21) Nitrobenzene-d5           | 9.081  | 128  | 93668    | 21.426 | ng/ul  | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.805  | 143  | 104471   | 21.881 | ng/ul  | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.352 | 165  | 187159   | 20.136 | ng/ul  | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.863 | 131  | 252455m  | 18.943 | ng/ul  | 0.00     |
| 46) Dimethylphthalate-d6      | 13.963 | 166  | 547584   | 20.269 | ng/ul  | 0.00     |
| 49) Acenaphthylene-d8         | 14.251 | 160  | 683691   | 20.910 | ng/ul  | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.751 | 143  | 87296    | 17.552 | ng/ul  | 0.00     |
| 60) Fluorene-d10              | 15.551 | 176  | 466367   | 20.068 | ng/ul  | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.669 | 200  | 86251    | 19.163 | ng/ul  | 0.00     |
| 73) Anthracene-d10            | 17.410 | 188  | 711034   | 20.647 | ng/ul  | 0.00     |
| 81) Pyrene-d10                | 19.639 | 212  | 826969   | 21.809 | ng/ul  | 0.00     |
| 92) Benzo(a)pyrene-d12        | 23.680 | 264  | 718265   | 21.241 | ng/ul  | 0.00     |
| Target Compounds              |        |      |          |        |        |          |
|                               |        |      |          |        | Qvalue |          |
| 2) 1,4-Dioxane                | 3.393  | 88   | 26077    | 8.084  | ng/ul# | 14       |
| 5) Pyridine                   | 3.793  | 79   | 174439   | 19.217 | ng/ul# | 42       |
| 6) Benzaldehyde               | 7.058  | 77   | 144361   | 21.940 | ng/ul  | 97       |
| 8) Phenol                     | 7.116  | 94   | 238291   | 19.261 | ng/ul# | 92       |
| 10) Bis(2-Chloroethyl)ether   | 7.340  | 93   | 195784   | 20.541 | ng/ul# | 78       |
| 12) 2-Chlorophenol            | 7.487  | 128  | 189529   | 20.606 | ng/ul  | 96       |
| 13) 2-Methylphenol            | 8.369  | 108  | 182498   | 19.234 | ng/ul  | 98       |
| 14) 2,2'-oxybis(1-Chloropr... | 8.458  | 45   | 290857   | 20.565 | ng/ul# | 85       |
| 16) Acetophenone              | 8.746  | 105  | 297822   | 18.712 | ng/ul# | 80       |
| 17) N-Nitroso-di-n-propyla... | 8.734  | 70   | 159495   | 19.127 | ng/ul# | 75       |
| 18) 4-Methylphenol            | 8.699  | 108  | 192230   | 18.401 | ng/ul  | 98       |
| 19) Hexachloroethane          | 9.010  | 117  | 84727    | 21.922 | ng/ul  | 83       |
| 22) Nitrobenzene              | 9.122  | 77   | 223907   | 20.905 | ng/ul# | 84       |
| 23) Isophorone                | 9.657  | 82   | 425316   | 20.105 | ng/ul# | 96       |
| 25) 2-Nitrophenol             | 9.840  | 139  | 106296   | 21.021 | ng/ul# | 86       |
| 26) 2,4-Dimethylphenol        | 9.905  | 107  | 223753   | 20.438 | ng/ul# | 77       |
| 27) Bis(2-Chloroethoxy)met... | 10.134 | 93   | 264199   | 21.307 | ng/ul# | 98       |
| 29) 2,4-Dichlorophenol        | 10.375 | 162  | 183474m  | 20.535 | ng/ul  |          |
| 30) Naphthalene               | 10.781 | 128  | 608313   | 21.145 | ng/ul  | 97       |
| 32) 4-Chloroaniline           | 10.887 | 127  | 249533   | 19.001 | ng/ul  | 96       |
| 33) Hexachlorobutadiene       | 11.075 | 225  | 138597   | 21.887 | ng/ul  | 97       |
| 34) Caprolactam               | 11.657 | 113  | 53845    | 16.754 | ng/ul# | 12       |
| 35) 4-Chloro-3-methylphenol   | 12.010 | 107  | 199890   | 18.520 | ng/ul# | 71       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP050222\  
 Data File : BP010166.D  
 Acq On : 02 May 2022 11:14  
 Operator : CG/JU  
 Sample : SSTDCCC020  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 LabSampleId :  
 SSTDCCC020

Manual IntegrationsAPPROVED

Quant Time: May 03 00:58:41 2022  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP042122.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Tue May 03 00:58:32 2022  
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022  
 Supervised By :Yogesh Patel 05/05/2022

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| 36) 2-Methylnaphthalene       | 12.387 | 142  | 415671   | 19.923 | ng/ul  | 93       |
| 37) 1-Methylnaphthalene       | 12.604 | 142  | 416372   | 19.654 | ng/ul# | 97       |
| 39) 1,2,4,5-Tetrachloroben... | 12.757 | 216  | 242990   | 22.646 | ng/ul  | 94       |
| 40) Hexachlorocyclopentadiene | 12.740 | 237  | 137037   | 22.662 | ng/ul  | 94       |
| 41) 2,4,6-Trichlorophenol     | 12.993 | 196  | 149793   | 21.576 | ng/ul# | 85       |
| 42) 2,4,5-Trichlorophenol     | 13.063 | 196  | 156745   | 20.751 | ng/ul# | 85       |
| 43) 1,1'-Biphenyl             | 13.393 | 154  | 557396   | 21.760 | ng/ul# | 93       |
| 44) 2-Chloronaphthalene       | 13.434 | 162  | 434989   | 21.944 | ng/ul  | 92       |
| 45) 2-Nitroaniline            | 13.634 | 65   | 128955   | 20.062 | ng/ul# | 79       |
| 47) Dimethylphthalate         | 14.010 | 163  | 528307   | 20.121 | ng/ul# | 92       |
| 48) 2,6-Dinitrotoluene        | 14.128 | 165  | 107833   | 20.012 | ng/ul  | 97       |
| 50) Acenaphthylene            | 14.275 | 152  | 683483   | 21.103 | ng/ul  | 95       |
| 51) 3-Nitroaniline            | 14.457 | 138  | 64922    | 13.301 | ng/ul# | 77       |
| 52) Acenaphthene              | 14.622 | 153  | 438851   | 20.838 | ng/ul  | 97       |
| 53) 2,4-Dinitrophenol         | 14.663 | 184  | 50201    | 16.673 | ng/ul# | 81       |
| 55) 4-Nitrophenol             | 14.769 | 109  | 70943    | 16.418 | ng/ul# | 47       |
| 56) Dibenzofuran              | 14.957 | 168  | 636949   | 20.425 | ng/ul  | 98       |
| 57) 2,4-Dinitrotoluene        | 14.916 | 165  | 150473   | 19.116 | ng/ul# | 81       |
| 58) 2,3,4,6-Tetrachlorophenol | 15.181 | 232  | 132544   | 19.740 | ng/ul# | 85       |
| 59) Diethylphthalate          | 15.375 | 149  | 562547   | 21.064 | ng/ul  | 94       |
| 61) Fluorene                  | 15.604 | 166  | 512937   | 20.034 | ng/ul  | 91       |
| 62) 4-Chlorophenyl-phenyle... | 15.598 | 204  | 272740   | 20.241 | ng/ul  | 99       |
| 63) 4-Nitroaniline            | 15.622 | 138  | 61196m   | 12.884 | ng/ul  |          |
| 66) 4,6-Dinitro-2-methylph... | 15.681 | 198  | 81948    | 18.871 | ng/ul# | 90       |
| 67) N-Nitrosodiphenylamine    | 15.810 | 169  | 444142   | 21.746 | ng/ul  | 94       |
| 68) 4-Bromophenyl-phenylether | 16.498 | 248  | 167407   | 22.049 | ng/ul  | 92       |
| 69) Hexachlorobenzene         | 16.616 | 284  | 184501   | 20.935 | ng/ul# | 88       |
| 70) Atrazine                  | 16.769 | 200  | 171076   | 20.603 | ng/ul# | 87       |
| 71) Pentachlorophenol         | 16.963 | 266  | 101613   | 18.105 | ng/ul# | 82       |
| 72) Phenanthrene              | 17.351 | 178  | 798128   | 20.629 | ng/ul  | 99       |
| 74) Anthracene                | 17.445 | 178  | 804509   | 20.645 | ng/ul  | 99       |
| 75) 1,2,3,4-Tetrachloroben... | 13.357 | 216  | 249367   | 23.997 | ng/ul# | 89       |
| 76) Pentachlorobenzene        | 14.881 | 250  | 225461   | 22.125 | ng/uL  | 87       |
| 77) Carbazole                 | 17.710 | 167  | 690514   | 20.024 | ng/ul  | 96       |
| 78) Di-n-butylphthalate       | 18.263 | 149  | 999148   | 23.551 | ng/ul# | 92       |
| 80) Fluoranthene              | 19.316 | 202  | 941529   | 22.014 | ng/ul  | 97       |
| 82) Pyrene                    | 19.669 | 202  | 960485   | 21.797 | ng/ul# | 94       |
| 83) Butylbenzylphthalate      | 20.533 | 149  | 412560   | 22.650 | ng/ul# | 81       |
| 84) 3,3'-Dichlorobenzidine    | 21.310 | 252  | 232869   | 19.410 | ng/ul# | 95       |
| 85) Benzo(a)anthracene        | 21.374 | 228  | 903037   | 21.341 | ng/ul  | 95       |
| 86) Bis(2-ethylhexyl)phtha... | 21.298 | 149  | 645118m  | 23.897 | ng/ul  |          |
| 87) Chrysene                  | 21.427 | 228  | 889646   | 21.353 | ng/ul  | 99       |
| 89) Di-n-octyl phthalate      | 22.239 | 149  | 1020800  | 20.358 | ng/ul  | 100      |
| 90) Benzo(b)fluoranthene      | 23.092 | 252  | 896522   | 20.675 | ng/ul  | 99       |
| 91) Benzo(k)fluoranthene      | 23.139 | 252  | 868557   | 21.146 | ng/ul# | 97       |
| 93) Benzo(a)pyrene            | 23.727 | 252  | 854812   | 21.318 | ng/ul# | 98       |
| 94) Indeno(1,2,3-cd)pyrene    | 26.398 | 276  | 956406   | 23.977 | ng/ul# | 94       |
| 95) Dibenzo(a,h)anthracene    | 26.410 | 278  | 812775   | 23.813 | ng/ul# | 95       |
| 96) Benzo(g,h,i)perylene      | 27.174 | 276  | 818379   | 24.314 | ng/ul# | 91       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

**Instrument :**  
BNA\_P  
**LabSampleId :**  
SSTDCCC020

**Manual IntegrationsAPPROVED**

Reviewed By :Jagrut Upadhyay 05/03/2022  
Supervised By :Yogesh Patel 05/05/2022

