




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010186.D
 Acq On : 02 May 2022 23:09
 Operator : CG/JU
 Sample : N2615-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH0D5

Manual IntegrationsAPPROVED

Quant Time: May 03 01:39:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010186.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.405min (+ 0.047) 0.07 ng/uL

response 249

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	55.40	61.33
34.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010186.D
 Acq On : 02 May 2022 23:09
 Operator : CG/JU
 Sample : N2615-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH0D5

Manual IntegrationsAPPROVED

Quant Time: May 03 01:39:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Reviewed By :Jagrut Upadhyay 05/03/2022
 Supervised By :Yogesh Patel 05/05/2022



TIC: BP010186.D\data.ms

(3) 1,4-Dioxane-d8 (S)

3.358min (-0.000) 4.81 ng/uL m

response 17990

Ion	Exp%	Act%
96.00	100.00	100.00
64.00	55.40	88.09#
34.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010186.D
 Acq On : 02 May 2022 23:09
 Operator : CG/JU
 Sample : N2615-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BH0D5

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 05/03/2022
 Supervised By :Yogesh Patel 05/05/2022

Quant Time: May 03 01:39:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.922	152	162979	20.000	ng/ul	0.00
20) Naphthalene-d8	10.728	136	685488	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.551	164	417524	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.304	188	767994	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.386	240	652708	20.000	ng/ul	0.00
88) Perylene-d12	23.827	264	624821	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	17990m	4.813	ng/uL	0.00
4) Pyridine-d5	3.787	84	44976	4.278	ng/ul	0.01
7) Phenol-d5	7.087	99	87901	5.929	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.246	67	280860	31.574	ng/ul	0.00
11) 2-Chlorophenol-d4	7.452	132	255992	23.317	ng/ul	0.00
15) 4-Methylphenol-d8	8.628	113	171815	13.989	ng/ul	0.00
21) Nitrobenzene-d5	9.081	128	172406	31.623	ng/ul	0.00
24) 2-Nitrophenol-d4	9.804	143	179134	30.085	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.346	165	276088	23.818	ng/ul	0.00
31) 4-Chloroaniline-d4	10.863	131	328037	19.737	ng/ul	0.00
46) Dimethylphthalate-d6	13.963	166	1016758	31.305	ng/ul	0.00
49) Acenaphthylene-d8	14.245	160	1168459	29.726	ng/ul	0.00
54) 4-Nitrophenol-d4	14.769	143	23305	3.898	ng/ul	0.02
60) Fluorene-d10	15.545	176	858030	30.712	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	139699	28.768	ng/ul	0.00
73) Anthracene-d10	17.404	188	1423280	38.307	ng/ul	0.00
81) Pyrene-d10	19.633	212	1569685	42.287	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.674	264	1327034	40.512	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
Data File : BP010186.D
Acq On : 02 May 2022 23:09
Operator : CG/JU
Sample : N2615-01
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BH0D5

Manual IntegrationsAPPROVED

Reviewed By :Jagrut Upadhyay 05/03/2022
Supervised By :Yogesh Patel 05/05/2022

Quant Time: May 03 01:39:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 03 00:58:32 2022
Response via : Initial Calibration

