

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010212.D
 Acq On : 03 May 2022 17:50
 Operator : CG/JU
 Sample : PB144543BS
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS543

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022
 Supervised By :Yogesh Patel 05/05/2022

Quant Time: May 04 09:54:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	175566	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	776107	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	508369	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.304	188	976658	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.386	240	710952	20.000	ng/u1	0.00
88) Perylene-d12	23.827	264	669071	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	24410	6.063	ng/uL	0.00
4) Pyridine-d5	3.770	84	330107	29.145	ng/u1	0.00
7) Phenol-d5	7.087	99	462441	28.958	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.246	67	307193	32.058	ng/u1	0.00
11) 2-Chlorophenol-d4	7.452	132	374964	31.705	ng/u1	0.00
15) 4-Methylphenol-d8	8.634	113	396508	29.968	ng/u1	0.00
21) Nitrobenzene-d5	9.081	128	194888	31.572	ng/u1	0.00
24) 2-Nitrophenol-d4	9.805	143	217591	32.277	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.346	165	392725	29.924	ng/u1	0.00
31) 4-Chloroaniline-d4	10.857	131	470233	24.989	ng/u1	0.00
46) Dimethylphthalate-d6	13.963	166	1230498	31.116	ng/u1	0.00
49) Acenaphthylene-d8	14.245	160	1421633	29.704	ng/u1	0.00
54) 4-Nitrophenol-d4	14.751	143	186322m	25.593	ng/u1	0.00
60) Fluorene-d10	15.545	176	1029868	30.275	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	188046	30.450	ng/u1	0.00
73) Anthracene-d10	17.404	188	1478763	31.297	ng/u1	0.00
81) Pyrene-d10	19.633	212	1470269	36.364	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.669	264	1152013	32.843	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	46658	11.152	ng/uL#	25
5) Pyridine	3.793	79	341965	29.045	ng/u1#	49
6) Benzaldehyde	7.058	77	314866	36.895	ng/u1	95
8) Phenol	7.111	94	487497	30.382	ng/u1#	94
10) Bis(2-Chloroethyl)ether	7.340	93	387113	31.314	ng/u1#	78
12) 2-Chlorophenol	7.481	128	377562	31.650	ng/u1	95
13) 2-Methylphenol	8.363	108	371697	30.204	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.452	45	602603	32.851	ng/u1#	83
16) Acetophenone	8.746	105	611891	29.642	ng/u1#	80
17) N-Nitroso-di-n-propyla...	8.734	70	329959	30.508	ng/u1#	72
18) 4-Methylphenol	8.693	108	402524	29.708	ng/u1	90
19) Hexachloroethane	9.005	117	158911	31.702	ng/u1	82
22) Nitrobenzene	9.122	77	473800	31.328	ng/u1#	84
23) Isophorone	9.652	82	917272	30.708	ng/u1#	97
25) 2-Nitrophenol	9.834	139	227877	31.916	ng/u1#	82
26) 2,4-Dimethylphenol	9.899	107	468753	30.324	ng/u1#	78
27) Bis(2-Chloroethoxy)met...	10.134	93	553012	31.586	ng/u1#	96
29) 2,4-Dichlorophenol	10.369	162	388976	30.833	ng/u1#	84
30) Naphthalene	10.775	128	1259316	31.002	ng/u1	97
32) 4-Chloroaniline	10.881	127	450077	24.272	ng/u1	97
33) Hexachlorobutadiene	11.069	225	275557	30.819	ng/u1	96
34) Caprolactam	11.657	113	127892m	28.183	ng/u1	
35) 4-Chloro-3-methylphenol	12.010	107	445859	29.256	ng/u1#	70
36) 2-Methylnaphthalene	12.381	142	883299	29.983	ng/u1	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010212.D
 Acq On : 03 May 2022 17:50
 Operator : CG/JU
 Sample : PB144543BS
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS543

Manual Integrations
 APPROVED




Reviewed By :Jagrut Upadhyay 05/04/2022
 Supervised By :Yogesh Patel 05/05/2022

Quant Time: May 04 09:54:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.599	142	904262	30.230	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.751	216	512028	32.600	ng/ul	94
40) Hexachlorocyclopentadiene	12.734	237	274040	30.960	ng/ul	93
41) 2,4,6-Trichlorophenol	12.987	196	331235	32.595	ng/ul	86
42) 2,4,5-Trichlorophenol	13.063	196	360559m	32.610	ng/ul	
43) 1,1'-Biphenyl	13.387	154	1209842	32.267	ng/ul#	92
44) 2-Chloronaphthalene	13.428	162	938364	32.339	ng/ul	93
45) 2-Nitroaniline	13.628	65	296540	31.517	ng/ul#	81
47) Dimethylphthalate	14.010	163	1190645	30.980	ng/ul#	92
48) 2,6-Dinitrotoluene	14.128	165	246711	31.280	ng/ul	92
50) Acenaphthylene	14.275	152	1496504	31.566	ng/ul	95
51) 3-Nitroaniline	14.451	138	208158	29.134	ng/ul#	81
52) Acenaphthene	14.616	153	963116	31.243	ng/ul	97
53) 2,4-Dinitrophenol	14.663	184	113422	25.734	ng/ul#	81
55) 4-Nitrophenol	14.763	109	152745	24.149	ng/ul#	34
56) Dibenzofuran	14.951	168	1392525	30.506	ng/ul	100
57) 2,4-Dinitrotoluene	14.910	165	343123	29.779	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.181	232	295602	30.076	ng/ul#	86
59) Diethylphthalate	15.375	149	1201373	30.732	ng/ul	93
61) Fluorene	15.604	166	1118422	29.842	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.592	204	597339	30.285	ng/ul	97
63) 4-Nitroaniline	15.616	138	222640m	32.024	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.681	198	180748	30.336	ng/ul#	93
67) N-Nitrosodiphenylamine	15.804	169	974673	34.781	ng/ul	95
68) 4-Bromophenyl-phenylether	16.492	248	358397	34.403	ng/ul	94
69) Hexachlorobenzene	16.610	284	394524	32.627	ng/ul#	89
70) Atrazine	16.763	200	368213	32.320	ng/ul#	89
71) Pentachlorophenol	16.957	266	221294	28.737	ng/ul#	82
72) Phenanthrene	17.345	178	1673873	31.533	ng/ul	99
74) Anthracene	17.439	178	1685848	31.530	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.357	216	523334	36.704	ng/uL#	86
76) Pentachlorobenzene	14.875	250	488916	34.968	ng/uL	91
77) Carbazole	17.704	167	1371687	28.991	ng/ul#	96
78) Di-n-butylphthalate	18.257	149	1914757	32.894	ng/ul#	92
80) Fluoranthene	19.310	202	1725094	37.827	ng/ul	97
82) Pyrene	19.663	202	1735592	36.939	ng/ul#	95
83) Butylbenzylphthalate	20.528	149	706877	36.396	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.298	252	443466	34.667	ng/ul#	98
85) Benzo(a)anthracene	21.369	228	1481139	32.827	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.292	149	1010284	35.098	ng/ul#	81
87) Chrysene	21.422	228	1452305	32.690	ng/ul	99
89) Di-n-octyl phthalate	22.227	149	1574085	30.263	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	1413036	31.415	ng/ul	99
91) Benzo(k)fluoranthene	23.133	252	1408957	33.069	ng/ul#	97
93) Benzo(a)pyrene	23.716	252	1386249	33.327	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.374	276	1539286	37.201	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.392	278	1320777	37.304	ng/ul#	94
96) Benzo(g,h,i)perylene	27.157	276	1324597	37.939	ng/ul#	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Abundance Scan 52 (3.393 min): BP010200.D\data.ms (-47)







#2
1,4-Dioxane
Concen: 11.152 ng/uL
RT: 3.393 min Scan# 51
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Tgt Ion: 88 Resp: 46658
Ion Ratio Lower Upper
88 100
43 33.7 6.8 10.2
58 80.0 28.4 42.6

Instrument :
BNA_P
ClientSampleId :
SLCS543


Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#5
Pyridine
Concen: 29.045 ng/uL
RT: 3.793 min Scan# 120
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion: 79 Resp: 341965
Ion Ratio Lower Upper
79 100
52 84.6 35.5 53.3#
51 41.5 20.2 30.4#





#6
Benzaldehyde
Concen: 36.895 ng/ul
RT: 7.058 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion: 77 Resp: 314860
Ion Ratio Lower Upper
77 100
105 91.9 77.0 115.4
106 90.8 77.0 115.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#8
Phenol
Concen: 30.382 ng/ul
RT: 7.111 min Scan# 684
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion: 94 Resp: 487497
Ion Ratio Lower Upper
94 100
65 31.6 19.0 28.6
66 43.5 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.314 ng/ul
RT: 7.340 min Scan# 711
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion: 93 Resp: 387111
Ion Ratio Lower Upper
93 100
63 82.0 45.0 67.4
95 32.6 26.3 39.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#12
2-Chlorophenol
Concen: 31.650 ng/ul
RT: 7.481 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50



Tgt Ion:128 Resp: 377562
Ion Ratio Lower Upper
128 100
64 54.1 45.3 67.9
130 31.6 29.0 43.6









#13
2-Methylphenol
Concen: 30.204 ng/ul
RT: 8.363 min Scan# 897
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:108 Resp: 37169
Ion Ratio Lower Upper
108 100
107 89.2 70.8 106.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.851 ng/ul
RT: 8.452 min Scan# 912
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion: 45 Resp: 602603
Ion Ratio Lower Upper
45 100
77 13.4 17.1 25.7#
79 9.6 13.2 19.8#





#16
Acetophenone
Concen: 29.642 ng/ul
RT: 8.746 min Scan# 962
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:105 Resp: 61189
Ion Ratio Lower Upper
105 100
77 80.2 53.4 80.2
51 31.8 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#17
N-Nitroso-di-n-propylamine
Concen: 30.508 ng/ul
RT: 8.734 min Scan# 960
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion: 70 Resp: 329959
Ion Ratio Lower Upper
70 100
42 64.0 31.8 47.8#
101 10.3 15.6 23.4#
130 20.8 13.7 20.5#





#18
4-Methylphenol
Concen: 29.708 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:108 Resp: 402524
Ion Ratio Lower Upper
108 100
107 115.1 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#19
Hexachloroethane
Concen: 31.702 ng/ul
RT: 9.005 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:117 Resp: 158911
Ion Ratio Lower Upper
117 100
201 106.7 106.5 159.7
199 65.6 59.0 88.6





#22
Nitrobenzene
Concen: 31.328 ng/ul
RT: 9.122 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion: 77 Resp: 473800
Ion Ratio Lower Upper
77 100
123 42.0 42.3 63.5
65 13.9 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#23
Isophorone
Concen: 30.708 ng/ul
RT: 9.652 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50



Tgt Ion: 82 Resp: 917272
Ion Ratio Lower Upper
82 100
95 7.6 7.8 11.8
138 16.7 12.9 19.3





#25

2-Nitrophenol

Concen: 31.916 ng/ul

RT: 9.834 min Scan# 1147

Delta R.T. -0.006 min

Lab File: BP010212.D

Acq: 03 May 2022 17:50

Instrument :

BNA_P

ClientSampleId :

SLCS543

Tgt Ion:139 Resp: 22787

Ion Ratio Lower Upper

139 100

65 56.8 49.4 74.2





109 34.3 44.7 67.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022

Supervised By :Yogesh Patel 05/05/2022



#26

2,4-Dimethylphenol

Concen: 30.324 ng/ul

RT: 9.899 min Scan# 1158

Delta R.T. -0.006 min

Lab File: BP010212.D

Acq: 03 May 2022 17:50




Tgt Ion:107 Resp: 468753


Ion Ratio Lower Upper

107 100

121 48.8 34.1 51.1


122 81.4 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 31.586 ng/ul
RT: 10.134 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543



Tgt Ion: 93 Resp: 553012
Ion Ratio Lower Upper
93 100
95 31.6 26.2 39.4
123 11.2 6.2 9.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#29
2,4-Dichlorophenol
Concen: 30.833 ng/ul
RT: 10.369 min Scan# 1238
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:162 Resp: 388976
Ion Ratio Lower Upper
162 100
164 65.2 37.0 55.6#
98 38.8 31.0 46.6





#30
Naphthalene
Concen: 31.002 ng/ul
RT: 10.775 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543



Tgt Ion:128 Resp: 1259310
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.2 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#32
4-Chloroaniline
Concen: 24.272 ng/ul
RT: 10.881 min Scan# 1325
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:127 Resp: 450077
Ion Ratio Lower Upper
127 100
129 32.7 24.6 37.0





#33
Hexachlorobutadiene
Concen: 30.819 ng/ul
RT: 11.069 min Scan# 1357
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:225 Resp: 27555
Ion Ratio Lower Upper
225 100
223 63.1 53.4 80.0
227 64.0 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#34
Caprolactam
Concen: 28.183 ng/ul m
RT: 11.657 min Scan# 1457
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:113 Resp: 127892
Ion Ratio Lower Upper
113 100
55 175.6 58.2 87.2#
56 139.1 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 29.256 ng/ul
RT: 12.010 min Scan# 1517
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543



Tgt Ion:107 Resp: 445859
Ion Ratio Lower Upper
107 100
144 24.0 15.8 23.8
142 75.1 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#36
2-Methylnaphthalene
Concen: 29.983 ng/ul
RT: 12.381 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Tgt Ion:142 Resp: 883299
Ion Ratio Lower Upper
142 100
141 87.9 64.3 96.5









#37
1-Methylnaphthalene
Concen: 30.230 ng/ul
RT: 12.599 min Scan# 1617
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:142 Resp: 904262
Ion Ratio Lower Upper
142 100
141 91.6 69.8 104.8
116 4.1 5.6 8.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.600 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:216 Resp: 512028
Ion Ratio Lower Upper
216 100
214 77.3 58.2 87.2
179 20.4 14.2 21.2
108 16.5 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 30.960 ng/ul
RT: 12.734 min Scan# 1640
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:237 Resp: 274040
Ion Ratio Lower Upper
237 100
235 60.7 53.2 79.8
272 12.4 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#41
2,4,6-Trichlorophenol
Concen: 32.595 ng/ul
RT: 12.987 min Scan# 1683
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:196 Resp: 331235
Ion Ratio Lower Upper
196 100
198 94.1 64.1 96.1
200 29.6 19.9 29.9





#42
2,4,5-Trichlorophenol
Concen: 32.610 ng/ul m
RT: 13.063 min Scan# 1
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:196 Resp: 360559
Ion Ratio Lower Upper
196 100
198 98.3 68.6 103.0
200 32.3 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#43
1,1'-Biphenyl
Concen: 32.267 ng/ul
RT: 13.387 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:154 Resp: 1209842
Ion Ratio Lower Upper
154 100
153 40.7 28.6 43.0
76 12.8 8.5 12.7





#44
2-Chloronaphthalene
Concen: 32.339 ng/ul
RT: 13.428 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion: 162 Resp: 938364
Ion Ratio Lower Upper
162 100
164 32.7 27.5 41.3
127 40.1 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#45
2-Nitroaniline
Concen: 31.517 ng/ul
RT: 13.628 min Scan# 1792
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion: 65 Resp: 296540
Ion Ratio Lower Upper
65 100
92 66.1 77.9 116.9#
138 95.9 81.5 122.3





#47
Dimethylphthalate
Concen: 30.980 ng/ul
RT: 14.010 min Scan# 1857
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:163 Resp: 119064
Ion Ratio Lower Upper
163 100
194 4.4 4.1 6.1
164 9.8 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#48
2,6-Dinitrotoluene
Concen: 31.280 ng/ul
RT: 14.128 min Scan# 1877
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:165 Resp: 246711
Ion Ratio Lower Upper
165 100
89 57.7 51.7 77.5
121 20.2 18.3 27.5





#50
Acenaphthylene
Concen: 31.566 ng/ul
RT: 14.275 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:152 Resp: 1496504
Ion Ratio Lower Upper
152 100
151 19.7 18.3 27.5
153 12.9 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#51
3-Nitroaniline
Concen: 29.134 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50



Tgt Ion:138 Resp: 208158
Ion Ratio Lower Upper
138 100
108 11.4 15.8 23.8#
92 122.8 117.1 175.7









#52
Acenaphthene
Concen: 31.243 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:153 Resp: 963110
Ion Ratio Lower Upper
153 100
152 46.4 34.3 51.5
154 89.1 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#53
2,4-Dinitrophenol
Concen: 25.734 ng/ul
RT: 14.663 min Scan# 1968
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:184 Resp: 113422
Ion Ratio Lower Upper
184 100
63 72.9 34.6 52.0#
154 69.9 57.7 86.5





#55
4-Nitrophenol
Concen: 24.149 ng/ul
RT: 14.763 min Scan# 1985
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543



Tgt Ion:109 Resp: 152745
Ion Ratio Lower Upper
109 100
139 128.8 61.4 92.2
65 145.7 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#56
Dibenzofuran
Concen: 30.506 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:168 Resp: 1392525
Ion Ratio Lower Upper
168 100
139 37.5 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 29.779 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:165 Resp: 34312
Ion Ratio Lower Upper
165 100
63 49.9 29.5 44.3#
182 3.1 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 30.076 ng/ul
RT: 15.181 min Scan# 2056
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:232 Resp: 295602
Ion Ratio Lower Upper
232 100
131 41.6 25.4 38.2#
130 2.3 4.2 6.2#
166 28.4 28.1 42.1





#59
Diethylphthalate
Concen: 30.732 ng/ul
RT: 15.375 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:149 Resp: 120137
Ion Ratio Lower Upper
149 100
177 22.6 22.0 33.0
150 12.5 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#61
Fluorene
Concen: 29.842 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50



Tgt Ion:166 Resp: 1118422
Ion Ratio Lower Upper
166 100
165 96.5 70.3 105.5
167 13.3 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 30.285 ng/ul
RT: 15.592 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:204 Resp: 597339
Ion Ratio Lower Upper
204 100
206 33.2 27.6 41.4
141 60.5 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#63
4-Nitroaniline
Concen: 32.024 ng/ul m
RT: 15.616 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:138 Resp: 222640
Ion Ratio Lower Upper
138 100
92 51.6 41.6 62.4
108 83.7 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 30.336 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:198 Resp: 180748
Ion Ratio Lower Upper
198 100
182 4.0 6.0 9.0
77 25.0 17.7 26.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#67
N-Nitrosodiphenylamine
Concen: 34.781 ng/ul
RT: 15.804 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50



Tgt Ion:169 Resp: 974673
Ion Ratio Lower Upper
169 100
168 65.6 48.3 72.5
167 35.5 29.6 44.4






#68
4-Bromophenyl-phenylether
Concen: 34.403 ng/ul
RT: 16.492 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :

BNA_P

ClientSampleId :

SLCS543






Tgt Ion:248 Resp: 35839
Ion Ratio Lower Upper
248 100
250 98.3 79.4 119.0
141 72.2 67.2 100.8


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#69
Hexachlorobenzene
Concen: 32.627 ng/ul
RT: 16.610 min Scan# 2299
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:284 Resp: 394524
Ion Ratio Lower Upper
284 100
142 35.4 19.2 28.8#
249 33.1 27.4 41.2





#70
Atrazine
Concen: 32.320 ng/ul
RT: 16.763 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:200 Resp: 36821
Ion Ratio Lower Upper
200 100
173 26.0 19.2 28.8
215 49.4 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#71
Pentachlorophenol
Concen: 28.737 ng/ul
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50



Tgt Ion:266 Resp: 221294
Ion Ratio Lower Upper
266 100
264 60.5 57.3 85.9
268 62.4 65.7 98.5









#72
Phenanthrene
Concen: 31.533 ng/u1
RT: 17.345 min Scan# 2425
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:178 Resp: 167387
Ion Ratio Lower Upper
178 100
179 15.1 11.7 17.5
176 18.7 14.5 21.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#74
Anthracene
Concen: 31.530 ng/u1
RT: 17.439 min Scan# 2440
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Tgt Ion:178 Resp: 1685848
Ion Ratio Lower Upper
178 100
179 15.0 12.3 18.5
176 17.8 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 36.704 ng/uL
RT: 13.357 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:216 Resp: 523334
Ion Ratio Lower Upper
216 100
214 77.4 50.4 75.6
218 48.2 35.2 52.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#76
Pentachlorobenzene
Concen: 34.968 ng/uL
RT: 14.875 min Scan# 2004
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:250 Resp: 488916
Ion Ratio Lower Upper
250 100
248 63.2 43.5 65.3
252 62.7 46.5 69.7





#77
 Carbazole
 Concen: 28.991 ng/ul
 RT: 17.704 min Scan# 2486
 Delta R.T. -0.006 min
 Lab File: BP010212.D
 Acq: 03 May 2022 17:50




Instrument :
 BNA_P
 ClientSampleId :
 SLCS543




Tgt Ion:167 Resp: 137168
 Ion Ratio Lower Upper
 167 100
 166 20.8 15.8 23.8
 139 12.6 8.3 12.5

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
 Supervised By :Yogesh Patel 05/05/2022




#78
 Di-n-butylphthalate
 Concen: 32.894 ng/ul
 RT: 18.257 min Scan# 2579
 Delta R.T. -0.006 min
 Lab File: BP010212.D
 Acq: 03 May 2022 17:50



Tgt Ion:149 Resp: 1914757
 Ion Ratio Lower Upper
 149 100
 150 9.0 9.6 14.4#
 104 5.3 6.2 9.4#









#80
Fluoranthene
Concen: 37.827 ng/ul
RT: 19.310 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:202 Resp: 1725094
Ion Ratio Lower Upper
202 100
101 10.2 9.6 14.4
100 8.1 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#82
Pyrene
Concen: 36.939 ng/ul
RT: 19.663 min Scan# 2818
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Tgt Ion:202 Resp: 1735592
Ion Ratio Lower Upper
202 100
101 11.7 10.6 15.8
100 9.7 9.8 14.6#









#83
Butylbenzylphthalate
Concen: 36.396 ng/ul
RT: 20.528 min Scan# 2965
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:149 Resp: 70687
Ion Ratio Lower Upper
149 100
91 73.8 71.1 106.7
206 21.5 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#84
3,3'-Dichlorobenzidine
Concen: 34.667 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. -0.012 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:252 Resp: 443466
Ion Ratio Lower Upper
252 100
254 63.0 49.8 74.6
126 10.2 4.3 6.5





#85
Benzo(a)anthracene
Concen: 32.827 ng/ul
RT: 21.369 min Scan# 3109
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543



Tgt Ion:228 Resp: 1481139
Ion Ratio Lower Upper
228 100
229 19.9 18.2 27.2
226 26.2 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 35.098 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:149 Resp: 1010284
Ion Ratio Lower Upper
149 100
167 28.0 32.8 49.2#
279 4.8 5.6 8.4#





#87
Chrysene
Concen: 32.690 ng/u1
RT: 21.422 min Scan# 3118
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543



Tgt Ion:228 Resp: 1452309
Ion Ratio Lower Upper
228 100
226 28.8 22.7 34.1
229 19.3 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#89
Di-n-octyl phthalate
Concen: 30.263 ng/u1
RT: 22.227 min Scan# 3254
Delta R.T. -0.012 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Tgt Ion:149 Resp: 1574085









#90
Benzo(b)fluoranthene
Concen: 31.415 ng/ul
RT: 23.080 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50

Instrument :
BNA_P
ClientSampleId :
SLCS543

Tgt Ion:252 Resp: 141303
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 8.7 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#91
Benzo(k)fluoranthene
Concen: 33.069 ng/ul
RT: 23.133 min Scan# 3408
Delta R.T. -0.006 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50


Tgt Ion:252 Resp: 1408957
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 8.4 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 33.327 ng/ul
RT: 23.716 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:252 Resp: 1386249
Ion Ratio Lower Upper
252 100
253 21.3 17.5 26.3
125 9.9 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 37.201 ng/ul
RT: 26.374 min Scan# 3959
Delta R.T. -0.023 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Tgt Ion:276 Resp: 1539286
Ion Ratio Lower Upper
276 100
138 21.1 12.3 18.5
277 24.3 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 37.304 ng/ul
RT: 26.392 min Scan# 3962
Delta R.T. -0.018 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50




Instrument :
BNA_P
ClientSampleId :
SLCS543




Tgt Ion:278 Resp: 1320777
Ion Ratio Lower Upper
278 100
139 13.9 6.6 10.0#
279 24.0 18.4 27.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#96
Benzo(g,h,i)perylene
Concen: 37.939 ng/ul
RT: 27.157 min Scan# 4092
Delta R.T. -0.018 min
Lab File: BP010212.D
Acq: 03 May 2022 17:50



Tgt Ion:276 Resp: 1324597
Ion Ratio Lower Upper
276 100
138 18.4 11.6 17.4#
277 24.3 15.9 23.9#

