

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010221.D
 Acq On : 03 May 2022 23:27
 Operator : CG/JU
 Sample : PB144558BS
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS558

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022
 Supervised By :Yogesh Patel 05/05/2022

Quant Time: May 04 01:26:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.916	152	167668	20.000	ng/u1	0.00
20) Naphthalene-d8	10.722	136	723288	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.551	164	452436	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.304	188	834953	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.386	240	683216	20.000	ng/u1	# 0.00
88) Perylene-d12	23.821	264	669175	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	20734m	5.392	ng/uL	0.00
4) Pyridine-d5	3.769	84	297230	27.479	ng/u1	0.00
7) Phenol-d5	7.081	99	409149	26.828	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.246	67	278665	30.451	ng/u1	0.00
11) 2-Chlorophenol-d4	7.452	132	330440	29.256	ng/u1	0.00
15) 4-Methylphenol-d8	8.628	113	343939	27.219	ng/u1	0.00
21) Nitrobenzene-d5	9.081	128	170020	29.555	ng/u1	0.00
24) 2-Nitrophenol-d4	9.804	143	186979	29.761	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.346	165	336358	27.501	ng/u1	0.00
31) 4-Chloroaniline-d4	10.857	131	401689	22.905	ng/u1	0.00
46) Dimethylphthalate-d6	13.957	166	1013408	28.795	ng/u1	0.00
49) Acenaphthylene-d8	14.245	160	1191744	27.979	ng/u1	0.00
54) 4-Nitrophenol-d4	14.745	143	143625m	22.167	ng/u1	0.00
60) Fluorene-d10	15.545	176	844575	27.897	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	146452	27.740	ng/u1	0.00
73) Anthracene-d10	17.404	188	1177738	29.156	ng/u1	0.00
81) Pyrene-d10	19.633	212	1229920	31.654	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.668	264	1066992	30.414	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	43924	10.993	ng/uL#	13
5) Pyridine	3.793	79	310565	27.621	ng/u1#	44
6) Benzaldehyde	7.052	77	280762	34.449	ng/u1	96
8) Phenol	7.110	94	436285	28.471	ng/u1#	93
10) Bis(2-Chloroethyl)ether	7.340	93	355645	30.124	ng/u1#	74
12) 2-Chlorophenol	7.481	128	342955	30.103	ng/u1	95
13) 2-Methylphenol	8.363	108	330497	28.122	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.446	45	553299	31.584	ng/u1#	84
16) Acetophenone	8.740	105	540359	27.409	ng/u1#	79
17) N-Nitroso-di-n-propyla...	8.734	70	295612	28.620	ng/u1#	73
18) 4-Methylphenol	8.693	108	356993	27.589	ng/u1	94
19) Hexachloroethane	9.005	117	146603	30.624	ng/u1	83
22) Nitrobenzene	9.122	77	427790	30.352	ng/u1#	86
23) Isophorone	9.652	82	798181	28.672	ng/u1#	97
25) 2-Nitrophenol	9.834	139	201365	30.262	ng/u1#	84
26) 2,4-Dimethylphenol	9.899	107	419799	29.140	ng/u1#	78
27) Bis(2-Chloroethoxy)met...	10.134	93	489704	30.012	ng/u1#	97
29) 2,4-Dichlorophenol	10.369	162	342917	29.167	ng/u1#	84
30) Naphthalene	10.769	128	1130200	29.855	ng/u1	96
32) 4-Chloroaniline	10.881	127	398248	23.045	ng/u1	97
33) Hexachlorobutadiene	11.063	225	249746	29.972	ng/u1	97
34) Caprolactam	11.657	113	100443	23.750	ng/u1#	1
35) 4-Chloro-3-methylphenol	12.004	107	381805	26.882	ng/u1#	70
36) 2-Methylnaphthalene	12.381	142	777940	28.335	ng/u1	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050222\
 Data File : BP010221.D
 Acq On : 03 May 2022 23:27
 Operator : CG/JU
 Sample : PB144558BS
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS558


Quant Time: May 04 01:26:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP042122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 03 00:58:32 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022
 Supervised By :Yogesh Patel 05/05/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.598	142	796302	28.565	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.751	216	446562	31.947	ng/ul	93
40) Hexachlorocyclopentadiene	12.734	237	244959	31.096	ng/ul	96
41) 2,4,6-Trichlorophenol	12.987	196	277067	30.635	ng/ul#	85
42) 2,4,5-Trichlorophenol	13.057	196	297027	30.185	ng/ul#	88
43) 1,1'-Biphenyl	13.387	154	1059059	31.737	ng/ul#	93
44) 2-Chloronaphthalene	13.428	162	809339	31.341	ng/ul	94
45) 2-Nitroaniline	13.628	65	248500	29.676	ng/ul#	80
47) Dimethylphthalate	14.004	163	1004957	29.381	ng/ul#	92
48) 2,6-Dinitrotoluene	14.122	165	206360	29.398	ng/ul	96
50) Acenaphthylene	14.275	152	1274781	30.214	ng/ul	95
51) 3-Nitroaniline	14.451	138	167195	26.294	ng/ul#	86
52) Acenaphthene	14.616	153	824486	30.052	ng/ul	97
53) 2,4-Dinitrophenol	14.657	184	92570	23.600	ng/ul#	82
55) 4-Nitrophenol	14.757	109	127988	22.737	ng/ul#	45
56) Dibenzofuran	14.951	168	1178599	29.012	ng/ul	99
57) 2,4-Dinitrotoluene	14.910	165	275130	26.830	ng/ul#	82
58) 2,3,4,6-Tetrachlorophenol	15.175	232	238930	27.316	ng/ul#	86
59) Diethylphthalate	15.369	149	1001347	28.782	ng/ul	94
61) Fluorene	15.598	166	943259	28.280	ng/ul	93
62) 4-Chlorophenyl-phenyle...	15.592	204	508204	28.951	ng/ul	97
63) 4-Nitroaniline	15.616	138	178541m	28.856	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.675	198	142450	27.966	ng/ul#	86
67) N-Nitrosodiphenylamine	15.804	169	801689	33.464	ng/ul	94
68) 4-Bromophenyl-phenylether	16.486	248	295364	33.164	ng/ul	96
69) Hexachlorobenzene	16.610	284	326494	31.584	ng/ul#	89
70) Atrazine	16.763	200	294030	30.189	ng/ul#	89
71) Pentachlorophenol	16.957	266	180424	27.406	ng/ul#	84
72) Phenanthrene	17.345	178	1369390	30.175	ng/ul	99
74) Anthracene	17.439	178	1379353	30.176	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.351	216	451781	37.063	ng/uL	89
76) Pentachlorobenzene	14.875	250	414871	34.708	ng/uL	92
77) Carbazole	17.704	167	1117481	27.627	ng/ul#	96
78) Di-n-butylphthalate	18.257	149	1495550	30.053	ng/ul#	92
80) Fluoranthene	19.310	202	1468481	33.507	ng/ul	97
82) Pyrene	19.663	202	1471851	32.598	ng/ul#	94
83) Butylbenzylphthalate	20.527	149	580066	31.079	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.298	252	389529	31.686	ng/ul#	93
85) Benzo(a)anthracene	21.369	228	1335998	30.813	ng/ul	94
86) Bis(2-ethylhexyl)phtha...	21.292	149	835139	30.191	ng/ul#	82
87) Chrysene	21.422	228	1343554	31.470	ng/ul	100
89) Di-n-octyl phthalate	22.227	149	1380779	26.542	ng/ul	100
90) Benzo(b)fluoranthene	23.080	252	1354816	30.116	ng/ul	100
91) Benzo(k)fluoranthene	23.127	252	1314842	30.855	ng/ul#	96
93) Benzo(a)pyrene	23.721	252	1314931	31.608	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.374	276	1485355	35.892	ng/ul#	94
95) Dibenzo(a,h)anthracene	26.392	278	1263622	35.684	ng/ul#	94
96) Benzo(g,h,i)perylene	27.156	276	1270281	36.377	ng/ul#	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.993 ng/uL
RT: 3.393 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion: 88 Resp: 43924
Ion Ratio Lower Upper
88 100
43 37.0 6.8 10.2
58 87.2 28.4 42.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#5
Pyridine
Concen: 27.621 ng/uL
RT: 3.793 min Scan# 120
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Tgt Ion: 79 Resp: 310565
Ion Ratio Lower Upper
79 100
52 87.9 35.5 53.3#
51 43.3 20.2 30.4#









#6
Benzaldehyde
Concen: 34.449 ng/ul
RT: 7.052 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Instrument :
BNA_P
ClientSampleId :
SLCS558

Tgt Ion: 77 Resp: 280762
Ion Ratio Lower Upper
77 100
105 93.2 77.0 115.4
106 90.4 77.0 115.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#8
Phenol
Concen: 28.471 ng/ul
RT: 7.110 min Scan# 684
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion: 94 Resp: 436285
Ion Ratio Lower Upper
94 100
65 32.4 19.0 28.6#
66 43.5 34.6 51.8





#10
Bis(2-Chloroethyl)ether
Concen: 30.124 ng/ul
RT: 7.340 min Scan# 723
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion: 93 Resp: 35564
Ion Ratio Lower Upper
93 100
63 84.6 45.0 67.4
95 31.3 26.3 39.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#12
2-Chlorophenol
Concen: 30.103 ng/ul
RT: 7.481 min Scan# 747
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:128 Resp: 342955
Ion Ratio Lower Upper
128 100
64 53.7 45.3 67.9
130 31.9 29.0 43.6





#13
2-Methylphenol
Concen: 28.122 ng/ul
RT: 8.363 min Scan# 897
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion: 108 Resp: 330497
Ion Ratio Lower Upper
108 100
107 90.9 70.8 106.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.584 ng/ul
RT: 8.446 min Scan# 911
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion: 45 Resp: 553299
Ion Ratio Lower Upper
45 100
77 13.8 17.1 25.7#
79 10.1 13.2 19.8#





#16
Acetophenone
Concen: 27.409 ng/ul
RT: 8.740 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion: 105 Resp: 540359
Ion Ratio Lower Upper
105 100
77 80.7 53.4 80.2
51 31.6 13.8 20.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#17
N-Nitroso-di-n-propylamine
Concen: 28.620 ng/ul
RT: 8.734 min Scan# 960
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion: 70 Resp: 295612
Ion Ratio Lower Upper
70 100
42 64.1 31.8 47.8#
101 10.4 15.6 23.4#
130 20.6 13.7 20.5#





#18
4-Methylphenol
Concen: 27.589 ng/ul
RT: 8.693 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:108 Resp: 35699
Ion Ratio Lower Upper
108 100
107 111.0 83.6 125.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#19
Hexachloroethane
Concen: 30.624 ng/ul
RT: 9.005 min Scan# 1006
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:117 Resp: 146603
Ion Ratio Lower Upper
117 100
201 107.0 106.5 159.7
199 66.3 59.0 88.6





#22
Nitrobenzene
Concen: 30.352 ng/ul
RT: 9.122 min Scan# 1026
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion: 77 Resp: 427790
Ion Ratio Lower Upper
77 100
123 43.8 42.3 63.5
65 14.5 5.8 8.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#23
Isophorone
Concen: 28.672 ng/ul
RT: 9.652 min Scan# 1116
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27



Tgt Ion: 82 Resp: 798181
Ion Ratio Lower Upper
82 100
95 7.4 7.8 11.8#
138 16.7 12.9 19.3









#25
2-Nitrophenol
Concen: 30.262 ng/ul
RT: 9.834 min Scan# 1147
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Instrument :
BNA_P
ClientSampleId :
SLCS558

Tgt Ion:139 Resp: 20136
Ion Ratio Lower Upper
139 100
65 58.8 49.4 74.2
109 35.0 44.7 67.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#26
2,4-Dimethylphenol
Concen: 29.140 ng/ul
RT: 9.899 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:107 Resp: 419799
Ion Ratio Lower Upper
107 100
121 48.8 34.1 51.1
122 81.1 46.3 69.5





#27
Bis(2-Chloroethoxy)methane
Concen: 30.012 ng/ul
RT: 10.134 min Scan# 1197
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion: 93 Resp: 489704
Ion Ratio Lower Upper
93 100
95 32.0 26.2 39.4
123 11.1 6.2 9.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#29
2,4-Dichlorophenol
Concen: 29.167 ng/ul
RT: 10.369 min Scan# 1238
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27



Tgt Ion:162 Resp: 342917
Ion Ratio Lower Upper
162 100
164 65.0 37.0 55.6#
98 38.2 31.0 46.6









#30
Naphthalene
Concen: 29.855 ng/ul
RT: 10.769 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Instrument :
BNA_P
ClientSampleId :
SLCS558

Tgt Ion:128 Resp: 1130200
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 12.5 12.2 18.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#32
4-Chloroaniline
Concen: 23.045 ng/ul
RT: 10.881 min Scan# 1325
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:127 Resp: 398248
Ion Ratio Lower Upper
127 100
129 32.4 24.6 37.0





#33
Hexachlorobutadiene
Concen: 29.972 ng/ul
RT: 11.063 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:225 Resp: 249740
Ion Ratio Lower Upper
225 100
223 64.8 53.4 80.0
227 64.0 53.8 80.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#34
Caprolactam
Concen: 23.750 ng/ul
RT: 11.657 min Scan# 1457
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:113 Resp: 100443
Ion Ratio Lower Upper
113 100
55 192.5 58.2 87.2#
56 145.5 68.6 103.0#





#35
4-Chloro-3-methylphenol
Concen: 26.882 ng/ul
RT: 12.004 min Scan# 1516
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:107 Resp: 38180
Ion Ratio Lower Upper
107 100
144 24.0 15.8 23.8
142 75.0 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#36
2-Methylnaphthalene
Concen: 28.335 ng/ul
RT: 12.381 min Scan# 1580
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:142 Resp: 777940
Ion Ratio Lower Upper
142 100
141 87.8 64.3 96.5





#37
1-Methylnaphthalene
Concen: 28.565 ng/ul
RT: 12.598 min Scan# 1617
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:142 Resp: 796302
Ion Ratio Lower Upper
142 100
141 90.9 69.8 104.8
116 4.2 5.6 8.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.947 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:216 Resp: 446562
Ion Ratio Lower Upper
216 100
214 78.9 58.2 87.2
179 21.1 14.2 21.2
108 15.9 15.3 22.9





#40
Hexachlorocyclopentadiene
Concen: 31.096 ng/ul
RT: 12.734 min Scan# 1640
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:237 Resp: 244958
Ion Ratio Lower Upper
237 100
235 63.0 53.2 79.8
272 12.3 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#41
2,4,6-Trichlorophenol
Concen: 30.635 ng/ul
RT: 12.987 min Scan# 1683
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:196 Resp: 277067
Ion Ratio Lower Upper
196 100
198 94.2 64.1 96.1
200 30.8 19.9 29.9#





#42
2,4,5-Trichlorophenol
Concen: 30.185 ng/ul
RT: 13.057 min Scan# 1695
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:196 Resp: 29702
Ion Ratio Lower Upper
196 100
198 94.9 68.6 103.0
200 31.2 32.8 49.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#43
1,1'-Biphenyl
Concen: 31.737 ng/ul
RT: 13.387 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:154 Resp: 1059059
Ion Ratio Lower Upper
154 100
153 40.1 28.6 43.0
76 12.7 8.5 12.7





#44
2-Chloronaphthalene
Concen: 31.341 ng/ul
RT: 13.428 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:162 Resp: 809339
Ion Ratio Lower Upper
162 100
164 32.8 27.5 41.3
127 40.2 37.1 55.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#45
2-Nitroaniline
Concen: 29.676 ng/ul
RT: 13.628 min Scan# 1792
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion: 65 Resp: 248500
Ion Ratio Lower Upper
65 100
92 64.4 77.9 116.9#
138 93.8 81.5 122.3





#47
Dimethylphthalate
Concen: 29.381 ng/ul
RT: 14.004 min Scan# 1856
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:163 Resp: 100495
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 10.1 11.3 16.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#48
2,6-Dinitrotoluene
Concen: 29.398 ng/ul
RT: 14.122 min Scan# 1876
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:165 Resp: 206360
Ion Ratio Lower Upper
165 100
89 61.4 51.7 77.5
121 21.9 18.3 27.5





#50
Acenaphthylene
Concen: 30.214 ng/ul
RT: 14.275 min Scan# 1901
Delta R.T. -0.000 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:152 Resp: 127478
Ion Ratio Lower Upper
152 100
151 19.8 18.3 27.5
153 12.7 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#51
3-Nitroaniline
Concen: 26.294 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27



Tgt Ion:138 Resp: 167195
Ion Ratio Lower Upper
138 100
108 12.4 15.8 23.8#
92 128.8 117.1 175.7









#52
Acenaphthene
Concen: 30.052 ng/ul
RT: 14.616 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Instrument :
BNA_P
ClientSampleId :
SLCS558

Tgt Ion:153 Resp: 824480
Ion Ratio Lower Upper
153 100
152 45.8 34.3 51.5
154 89.7 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#53
2,4-Dinitrophenol
Concen: 23.600 ng/ul
RT: 14.657 min Scan# 1967
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:184 Resp: 92570
Ion Ratio Lower Upper
184 100
63 74.8 34.6 52.0#
154 72.1 57.7 86.5





#55
4-Nitrophenol
Concen: 22.737 ng/ul
RT: 14.757 min Scan# 1985
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:109 Resp: 127988
Ion Ratio Lower Upper
109 100
139 115.9 61.4 92.2
65 139.5 65.3 97.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#56
Dibenzofuran
Concen: 29.012 ng/ul
RT: 14.951 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:168 Resp: 1178599
Ion Ratio Lower Upper
168 100
139 37.3 30.1 45.1





#57
2,4-Dinitrotoluene
Concen: 26.830 ng/ul
RT: 14.910 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:165 Resp: 275130
Ion Ratio Lower Upper
165 100
63 48.5 29.5 44.3#
182 2.8 2.8 4.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 27.316 ng/ul
RT: 15.175 min Scan# 2055
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:232 Resp: 238930
Ion Ratio Lower Upper
232 100
131 44.3 25.4 38.2#
130 2.5 4.2 6.2#
166 31.4 28.1 42.1





#59
Diethylphthalate
Concen: 28.782 ng/ul
RT: 15.369 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:149 Resp: 100134
Ion Ratio Lower Upper
149 100
177 22.7 22.0 33.0
150 12.9 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#61
Fluorene
Concen: 28.280 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Tgt Ion:166 Resp: 943259
Ion Ratio Lower Upper
166 100
165 95.4 70.3 105.5
167 13.3 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 28.951 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Instrument :
BNA_P
ClientSampleId :
SLCS558

Tgt Ion:204 Resp: 508204
Ion Ratio Lower Upper
204 100
206 32.4 27.6 41.4
141 60.1 46.6 69.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#63
4-Nitroaniline
Concen: 28.856 ng/ul m
RT: 15.616 min Scan# 2130
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:138 Resp: 178541
Ion Ratio Lower Upper
138 100
92 50.1 41.6 62.4
108 83.0 97.6 146.4#





#66
4,6-Dinitro-2-methylphenol
Concen: 27.966 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:198 Resp: 142450
Ion Ratio Lower Upper
198 100
182 4.6 6.0 9.0
77 29.4 17.7 26.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#67
N-Nitrosodiphenylamine
Concen: 33.464 ng/ul
RT: 15.804 min Scan# 2162
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:169 Resp: 801689
Ion Ratio Lower Upper
169 100
168 65.9 48.3 72.5
167 35.1 29.6 44.4





#68
4-Bromophenyl-phenylether
Concen: 33.164 ng/ul
RT: 16.486 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:248 Resp: 295364
Ion Ratio Lower Upper
248 100
250 97.4 79.4 119.0
141 77.1 67.2 100.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#69
Hexachlorobenzene
Concen: 31.584 ng/ul
RT: 16.610 min Scan# 2299
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:284 Resp: 326494
Ion Ratio Lower Upper
284 100
142 33.7 19.2 28.8#
249 31.1 27.4 41.2





#70
Atrazine
Concen: 30.189 ng/ul
RT: 16.763 min Scan# 2325
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:200 Resp: 294030
Ion Ratio Lower Upper
200 100
173 25.8 19.2 28.8
215 49.4 32.2 48.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#71
Pentachlorophenol
Concen: 27.406 ng/ul
RT: 16.957 min Scan# 2358
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:266 Resp: 180424
Ion Ratio Lower Upper
266 100
264 63.6 57.3 85.9
268 63.2 65.7 98.5





#72
Phenanthrene
Concen: 30.175 ng/u1
RT: 17.345 min Scan# 2424
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:178 Resp: 1369390
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 18.3 14.5 21.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#74
Anthracene
Concen: 30.176 ng/u1
RT: 17.439 min Scan# 2440
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Tgt Ion:178 Resp: 1379353
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 17.8 13.4 20.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.063 ng/uL
RT: 13.351 min Scan# 1745
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:216 Resp: 45178
Ion Ratio Lower Upper
216 100
214 75.2 50.4 75.6
218 47.1 35.2 52.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#76
Pentachlorobenzene
Concen: 34.708 ng/uL
RT: 14.875 min Scan# 2004
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27



Tgt Ion:250 Resp: 414871
Ion Ratio Lower Upper
250 100
248 60.5 43.5 65.3
252 64.0 46.5 69.7





#77
 Carbazole
 Concen: 27.627 ng/ul
 RT: 17.704 min Scan# 2485
 Delta R.T. -0.006 min
 Lab File: BP010221.D
 Acq: 03 May 2022 23:27

Instrument :

BNA_P

ClientSampleId :





SLCS558

Tgt Ion:167 Resp: 111748
 Ion Ratio Lower Upper
 167 100
 166 20.8 15.8 23.8
 139 12.8 8.3 12.5

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
 Supervised By :Yogesh Patel 05/05/2022



#78
 Di-n-butylphthalate
 Concen: 30.053 ng/ul
 RT: 18.257 min Scan# 2579
 Delta R.T. -0.006 min
 Lab File: BP010221.D
 Acq: 03 May 2022 23:27


Tgt Ion:149 Resp: 1495550
 Ion Ratio Lower Upper
 149 100
 150 8.9 9.6 14.4#
 104 5.2 6.2 9.4#





#80
Fluoranthene
Concen: 33.507 ng/u1
RT: 19.310 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:202 Resp: 146848
Ion Ratio Lower Upper
202 100
101 10.0 9.6 14.4
100 7.7 6.3 9.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#82
Pyrene
Concen: 32.598 ng/u1
RT: 19.663 min Scan# 2818
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Tgt Ion:202 Resp: 1471851
Ion Ratio Lower Upper
202 100
101 11.2 10.6 15.8
100 9.6 9.8 14.6#









#83
Butylbenzylphthalate
Concen: 31.079 ng/ul
RT: 20.527 min Scan# 2965
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Instrument :
BNA_P
ClientSampleId :
SLCS558

Tgt Ion:149 Resp: 580060
Ion Ratio Lower Upper
149 100
91 74.5 71.1 106.7
206 21.7 28.8 43.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#84
3,3'-Dichlorobenzidine
Concen: 31.686 ng/ul
RT: 21.298 min Scan# 3096
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:252 Resp: 389529
Ion Ratio Lower Upper
252 100
254 66.6 49.8 74.6
126 10.8 4.3 6.5





#85
Benzo(a)anthracene
Concen: 30.813 ng/ul
RT: 21.369 min Scan# 3108
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:228 Resp: 1335998
Ion Ratio Lower Upper
228 100
229 19.4 18.2 27.2
226 25.6 22.8 34.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 30.191 ng/ul
RT: 21.292 min Scan# 3095
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:149 Resp: 835139
Ion Ratio Lower Upper
149 100
167 28.8 32.8 49.2#
279 4.8 5.6 8.4#





#87
Chrysene
Concen: 31.470 ng/u1
RT: 21.422 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:228 Resp: 1343554
Ion Ratio Lower Upper
228 100
226 28.3 22.7 34.1
229 19.2 15.2 22.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#89
Di-n-octyl phthalate
Concen: 26.542 ng/u1
RT: 22.227 min Scan# 3254
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:149 Resp: 1380779





#90
Benzo(b)fluoranthene
Concen: 30.116 ng/u1
RT: 23.080 min Scan# 3399
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558



Tgt Ion:252 Resp: 1354810
Ion Ratio Lower Upper
252 100
253 21.5 17.4 26.0
125 9.3 7.3 10.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#91
Benzo(k)fluoranthene
Concen: 30.855 ng/u1
RT: 23.127 min Scan# 3407
Delta R.T. -0.012 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27


Tgt Ion:252 Resp: 1314842
Ion Ratio Lower Upper
252 100
253 21.2 17.9 26.9
125 9.3 5.1 7.7#





#93
Benzo(a)pyrene
Concen: 31.608 ng/ul
RT: 23.721 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27




Instrument :
BNA_P
ClientSampleId :
SLCS558




Tgt Ion:252 Resp: 131493
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 9.0 6.4 9.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 35.892 ng/ul
RT: 26.374 min Scan# 3959
Delta R.T. -0.024 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27



Tgt Ion:276 Resp: 1485355
Ion Ratio Lower Upper
276 100
138 20.8 12.3 18.5#
277 24.6 19.3 28.9





#95
Dibenzo(a,h)anthracene
Concen: 35.684 ng/ul
RT: 26.392 min Scan# 3961
Delta R.T. -0.018 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Instrument :

BNA_P

ClientSampleId :





SLCS558

Tgt Ion:278 Resp: 126362
Ion Ratio Lower Upper
278 100
139 14.3 6.6 10.0#
279 24.3 18.4 27.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 05/04/2022
Supervised By :Yogesh Patel 05/05/2022



#96
Benzo(g,h,i)perylene
Concen: 36.377 ng/ul
RT: 27.156 min Scan# 4092
Delta R.T. -0.018 min
Lab File: BP010221.D
Acq: 03 May 2022 23:27

Tgt Ion:276 Resp: 1270281
Ion Ratio Lower Upper
276 100
138 19.0 11.6 17.4#
277 24.0 15.9 23.9#

