

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050224\
 Data File : BP020170.D
 Acq On : 03 May 2024 09:55
 Operator : MA/JU
 Sample : PB160684BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS684

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay

Quant Time: May 03 11:55:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 03 06:07:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.622	152	85071	20.000	ng/u1	0.00
20) Naphthalene-d8	10.399	136	381391	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.293	164	287944	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.134	188	673264	20.000	ng/u1	0.00
79) Chrysene-d12	21.622	240	736789	20.000	ng/u1	0.00
88) Perylene-d12	24.980	264	816433	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.217	96	13379	6.642	ng/uL	0.00
4) Pyridine-d5	3.617	84	190526	32.552	ng/u1	0.00
7) Phenol-d5	6.811	99	264815	36.247	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.981	67	153875	34.672	ng/u1	0.00
11) 2-Chlorophenol-d4	7.158	132	205198	39.289	ng/u1	0.00
15) 4-Methylphenol-d8	8.334	113	225185	38.124	ng/u1	0.00
21) Nitrobenzene-d5	8.781	128	104617	36.631	ng/u1	0.00
24) 2-Nitrophenol-d4	9.493	143	124536	38.078	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.028	165	241270	40.426	ng/u1	0.00
31) 4-Chloroaniline-d4	10.552	131	309681	37.074	ng/u1	0.00
46) Dimethylphthalate-d6	13.710	166	801805	38.131	ng/u1	0.00
49) Acenaphthylene-d8	13.981	160	849043	36.552	ng/u1	0.00
54) 4-Nitrophenol-d4	14.540	143	130838	40.485	ng/u1	0.00
60) Fluorene-d10	15.316	176	678653	38.695	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.469	200	166881	39.061	ng/u1	0.00
73) Anthracene-d10	17.239	188	1107932	38.040	ng/u1	0.00
81) Pyrene-d10	19.639	212	1396548	32.426	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.763	264	1488273	37.701	ng/u1	0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.252	88	29313	12.824	ng/uL	94
5) Pyridine	3.634	79	189125	31.947	ng/u1	96
6) Benzaldehyde	6.793	77	175301	47.737	ng/u1	98
8) Phenol	6.834	94	266997	35.408	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.069	93	199448	33.802	ng/u1	99
12) 2-Chlorophenol	7.193	128	204706	37.541	ng/u1	94
13) 2-Methylphenol	8.063	108	201532	35.830	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.146	45	236134	32.081	ng/u1	97
16) Acetophenone	8.452	105	353984	36.109	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.434	70	186892	35.131	ng/u1	96
18) 4-Methylphenol	8.393	108	226511	37.051	ng/u1	96
19) Hexachloroethane	8.669	117	84715	33.973	ng/u1	87
22) Nitrobenzene	8.822	77	268298	32.725	ng/u1	99
23) Isophorone	9.346	82	529767	33.604	ng/u1	100
25) 2-Nitrophenol	9.528	139	129675	38.278	ng/u1	95
26) 2,4-Dimethylphenol	9.587	107	257068	34.782	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.828	93	293078	33.533	ng/u1	99
29) 2,4-Dichlorophenol	10.057	162	232381	40.030	ng/u1	97
30) Naphthalene	10.446	128	681086	34.646	ng/u1	99
32) 4-Chloroaniline	10.575	127	281317	34.441	ng/u1	99
33) Hexachlorobutadiene	10.710	225	170791	37.878	ng/u1	97
34) Caprolactam	11.399	113	81681	40.297	ng/u1	86
35) 4-Chloro-3-methylphenol	11.710	107	280703	38.836	ng/u1	97
36) 2-Methylnaphthalene	12.069	142	498804	36.468	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050224\
 Data File : BP020170.D
 Acq On : 03 May 2024 09:55
 Operator : MA/JU
 Sample : PB160684BS
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS684


Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay

Quant Time: May 03 11:55:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050124.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 03 06:07:25 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
37) 1-Methylnaphthalene	12.293	142	508379	36.935	ng/ul	98	05/03/2024
39) 1,2,4,5-Tetrachloroben...	12.446	216	341995	38.083	ng/ul	99	Supervised By :mohammad ahmed
40) Hexachlorocyclopentadiene	12.404	237	151106	30.939	ng/ul	96	
41) 2,4,6-Trichlorophenol	12.698	196	221183	39.275	ng/ul	99	
42) 2,4,5-Trichlorophenol	12.781	196	238446	39.348	ng/ul	95	
43) 1,1'-Biphenyl	13.104	154	698529	34.299	ng/ul	99	
44) 2-Chloronaphthalene	13.146	162	558775	34.576	ng/ul	98	05/06/2024
45) 2-Nitroaniline	13.381	65	192162	35.848	ng/ul	94	
47) Dimethylphthalate	13.757	163	785157	36.967	ng/ul	99	
48) 2,6-Dinitrotoluene	13.887	165	161860	38.334	ng/ul	97	
50) Acenaphthylene	14.010	152	899513	34.429	ng/ul	99	
51) 3-Nitroaniline	14.228	138	158776	40.318	ng/ul	94	
52) Acenaphthene	14.363	153	599678	34.155	ng/ul	100	
53) 2,4-Dinitrophenol	14.451	184	108370	41.814	ng/ul	90	
55) 4-Nitrophenol	14.551	109	130301	38.726	ng/ul	95	
56) Dibenzofuran	14.704	168	892104	36.944	ng/ul	96	
57) 2,4-Dinitrotoluene	14.704	165	246976	40.684	ng/ul	93	
58) 2,3,4,6-Tetrachlorophenol	14.940	232	234592	43.110	ng/ul	92	
59) Diethylphthalate	15.157	149	791415	37.328	ng/ul	99	
61) Fluorene	15.375	166	737285	36.747	ng/ul	98	
62) 4-Chlorophenyl-phenyle...	15.375	204	407623	38.295	ng/ul	91	
63) 4-Nitroaniline	15.428	138	153990	46.921	ng/ul	95	
66) 4,6-Dinitro-2-methylph...	15.487	198	171672	38.535	ng/ul#	89	
67) N-Nitrosodiphenylamine	15.604	169	648707	34.605	ng/ul	100	
68) 4-Bromophenyl-phenylether	16.304	248	277364	38.033	ng/ul	98	
69) Hexachlorobenzene	16.398	284	337950	40.011	ng/ul	95	
70) Atrazine	16.610	200	276000	41.522	ng/ul	99	
71) Pentachlorophenol	16.769	266	212386	41.600	ng/ul	97	
72) Phenanthrene	17.175	178	1223626	35.622	ng/ul	99	
74) Anthracene	17.275	178	1220349	34.929	ng/ul	99	
75) 1,2,3,4-Tetrachloroben...	13.063	216	340645	34.043	ng/uL	99	
76) Pentachlorobenzene	14.616	250	366956	36.647	ng/uL	99	
77) Carbazole	17.569	167	1111427	36.927	ng/ul	99	
78) Di-n-butylphthalate	18.169	149	1379869	37.823	ng/ul	99	
80) Fluoranthene	19.292	202	1559714	30.772	ng/ul	98	
82) Pyrene	19.669	202	1675212	31.765	ng/ul	97	
83) Butylbenzylphthalate	20.639	149	653859	32.661	ng/ul	94	
84) 3,3'-Dichlorobenzidine	21.533	252	611738	40.796	ng/ul	99	
85) Benzo(a)anthracene	21.598	228	1789784	35.815	ng/ul	100	
86) Bis(2-ethylhexyl)phtha...	21.557	149	967463	33.499	ng/ul#	96	
87) Chrysene	21.674	228	1693457	36.562	ng/ul	98	
89) Di-n-octyl phthalate	22.857	149	1653739	31.628	ng/ul	100	
90) Benzo(b)fluoranthene	23.933	252	1777029	36.590	ng/ul	98	
91) Benzo(k)fluoranthene	23.998	252	1770168	35.765	ng/ul	99	
93) Benzo(a)pyrene	24.833	252	1498324	32.424	ng/ul	99	
94) Indeno(1,2,3-cd)pyrene	28.804	276	1906409m	33.653	ng/ul		
95) Dibenzo(a,h)anthracene	28.904	278	1563190m	33.362	ng/ul		
96) Benzo(g,h,i)perylene	29.968	276	1440501m	31.529	ng/ul		

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 12.824 ng/uL
RT: 3.252 min Scan# 28
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion: 88 Resp: 2931
Ion Ratio Lower Upper
88 100
43 32.2 23.9 35.9
58 67.1 58.2 87.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#5
Pyridine
Concen: 31.947 ng/uL
RT: 3.634 min Scan# 93
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion: 79 Resp: 189125
Ion Ratio Lower Upper
79 100
52 77.3 60.7 91.1
51 36.6 33.1 49.7





#6
Benzaldehyde
Concen: 47.737 ng/ul
RT: 6.793 min Scan# 61
Delta R.T. -0.011 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion: 77 Resp: 17530
Ion Ratio Lower Upper
77 100
105 94.6 75.4 113.2
106 89.3 68.2 102.2



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#8
Phenol
Concen: 35.408 ng/ul
RT: 6.834 min Scan# 637
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion: 94 Resp: 266997
Ion Ratio Lower Upper
94 100
65 31.6 26.1 39.1
66 46.1 38.6 57.8






#10
Bis(2-Chloroethyl)ether
Concen: 33.802 ng/ul
RT: 7.069 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P


ClientSampleId :
SLCS684



Tgt Ion: 93 Resp: 199448
Ion Ratio Lower Upper
93 100
63 74.8 59.5 89.3
95 30.4 25.3 37.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#12
2-Chlorophenol
Concen: 37.541 ng/ul
RT: 7.193 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:128 Resp: 204706
Ion Ratio Lower Upper
128 100
64 48.4 43.8 65.6
130 32.3 26.5 39.7







#13
2-Methylphenol
Concen: 35.830 ng/ul
RT: 8.063 min Scan# 847
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:108 Resp: 20153
Ion Ratio Lower Upper
108 100
107 92.1 73.0 109.6



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#14
2,2'-oxybis(1-Chloropropane)
Concen: 32.081 ng/ul
RT: 8.146 min Scan# 860
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion: 45 Resp: 236134
Ion Ratio Lower Upper
45 100
77 19.2 13.8 20.8
79 14.2 10.9 16.3





#16
Acetophenone
Concen: 36.109 ng/ul
RT: 8.452 min Scan# 912
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:105 Resp: 353984
Ion Ratio Lower Upper
105 100
77 80.8 66.4 99.6
51 28.1 23.9 35.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/03/2024
Supervised By :mohammad
ahmed



05/06/2024




#17
N-Nitroso-di-n-propylamine
Concen: 35.131 ng/ul
RT: 8.434 min Scan# 909
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Tgt Ion: 70 Resp: 186892
Ion Ratio Lower Upper
70 100
42 54.1 46.2 69.4
101 9.9 7.8 11.6
130 19.9 15.4 23.0





#18
4-Methylphenol
Concen: 37.051 ng/ul
RT: 8.393 min Scan# 904
Delta R.T. -0.011 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion:108 Resp: 22651
Ion Ratio Lower Upper
108 100
107 114.5 88.2 132.2



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#19
Hexachloroethane
Concen: 33.973 ng/ul
RT: 8.669 min Scan# 949
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:117 Resp: 84715
Ion Ratio Lower Upper
117 100
201 119.0 84.0 126.0
199 76.0 53.4 80.2





#22
Nitrobenzene
Concen: 32.725 ng/ul
RT: 8.822 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion: 77 Resp: 268298
Ion Ratio Lower Upper
77 100
123 40.1 31.4 47.2
65 14.4 11.4 17.2


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/03/2024
Supervised By :mohammad
ahmed



05/06/2024




#23
Isophorone
Concen: 33.604 ng/ul
RT: 9.346 min Scan# 1064
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Tgt Ion: 82 Resp: 529767
Ion Ratio Lower Upper
82 100
95 8.2 6.6 9.8
138 16.1 12.9 19.3





#25
2-Nitrophenol
Concen: 38.278 ng/ul
RT: 9.528 min Scan# 1095
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion:139 Resp: 129675
Ion Ratio Lower Upper
139 100
65 61.5 53.8 80.6
109 43.3 35.3 52.9



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#26
2,4-Dimethylphenol
Concen: 34.782 ng/ul
RT: 9.587 min Scan# 1105
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:107 Resp: 257068
Ion Ratio Lower Upper
107 100
121 45.9 36.0 54.0
122 76.5 59.4 89.2





#27
Bis(2-Chloroethoxy)methane
Concen: 33.533 ng/ul
RT: 9.828 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion: 93 Resp: 293078
Ion Ratio Lower Upper
93 100
95 32.5 25.6 38.4
123 10.8 8.3 12.5



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#29
2,4-Dichlorophenol
Concen: 40.030 ng/ul
RT: 10.057 min Scan# 1185
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:162 Resp: 232381
Ion Ratio Lower Upper
162 100
164 66.1 53.1 79.7
98 35.1 31.4 47.2







#30
Naphthalene
Concen: 34.646 ng/ul
RT: 10.446 min Scan# 1251
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:128 Resp: 681086
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.5 10.6 16.0



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#32
4-Chloroaniline
Concen: 34.441 ng/ul
RT: 10.575 min Scan# 1273
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:127 Resp: 281317
Ion Ratio Lower Upper
127 100
129 30.7 24.8 37.2





#33
Hexachlorobutadiene
Concen: 37.878 ng/ul
RT: 10.710 min Scan# 1296
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:225 Resp: 17079
Ion Ratio Lower Upper
225 100
223 62.3 53.2 79.8
227 64.1 51.4 77.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#34
Caprolactam
Concen: 40.297 ng/ul
RT: 11.399 min Scan# 1413
Delta R.T. 0.012 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:113 Resp: 81681
Ion Ratio Lower Upper
113 100
55 144.6 131.0 196.6
56 115.7 105.4 158.2





#35
4-Chloro-3-methylphenol
Concen: 38.836 ng/ul
RT: 11.710 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion:107 Resp: 28070
Ion Ratio Lower Upper
107 100
144 24.2 18.3 27.5
142 73.0 56.6 84.8



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#36
2-Methylnaphthalene
Concen: 36.468 ng/ul
RT: 12.069 min Scan# 1527
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:142 Resp: 498804
Ion Ratio Lower Upper
142 100
141 91.5 74.6 111.8





#37
1-Methylnaphthalene
Concen: 36.935 ng/ul
RT: 12.293 min Scan# 1566
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNP_P



ClientSampleId :

SLCS684

Tgt Ion:142 Resp: 508379
Ion Ratio Lower Upper
142 100
141 92.4 75.5 113.3
116 4.8 3.9 5.9



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#39
1,2,4,5-Tetrachlorobenzene
Concen: 38.083 ng/ul
RT: 12.446 min Scan# 1591
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:216 Resp: 341995
Ion Ratio Lower Upper
216 100
214 76.6 61.3 91.9
179 20.1 16.9 25.3
108 14.8 13.2 19.8





#40
Hexachlorocyclopentadiene
Concen: 30.939 ng/ul
RT: 12.404 min Scan# 1584
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:237 Resp: 151100
Ion Ratio Lower Upper
237 100
235 62.2 52.4 78.6
272 13.0 10.0 15.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/03/2024
Supervised By :mohammad
ahmed



05/06/2024




#41
2,4,6-Trichlorophenol
Concen: 39.275 ng/ul
RT: 12.698 min Scan# 1634
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Tgt Ion:196 Resp: 221183
Ion Ratio Lower Upper
196 100
198 94.4 75.8 113.8
200 31.7 24.8 37.2






#42
2,4,5-Trichlorophenol
Concen: 39.348 ng/ul
RT: 12.781 min Scan# 1649
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P

ClientSampleId :


SLCS684



Tgt Ion:196 Resp: 238440
Ion Ratio Lower Upper
196 100
198 97.0 73.5 110.3
200 32.2 24.4 36.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#43
1,1'-Biphenyl
Concen: 34.299 ng/ul
RT: 13.104 min Scan# 1703
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:154 Resp: 698529
Ion Ratio Lower Upper
154 100
153 40.5 33.1 49.7
76 11.5 9.7 14.5






#44
2-Chloronaphthalene
Concen: 34.576 ng/ul
RT: 13.146 min Scan# 1711
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P

ClientSampleId :



SLCS684



Tgt Ion:162 Resp: 55877
Ion Ratio Lower Upper
162 100
164 33.3 26.1 39.1
127 40.3 33.6 50.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/03/2024
Supervised By :mohammad
ahmed



05/06/2024




#45
2-Nitroaniline
Concen: 35.848 ng/ul
RT: 13.381 min Scan# 1750
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Tgt Ion: 65 Resp: 192162
Ion Ratio Lower Upper
65 100
92 67.1 55.6 83.4
138 99.4 73.1 109.7





#47
Dimethylphthalate
Concen: 36.967 ng/ul
RT: 13.757 min Scan# 1814
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion:163 Resp: 78515
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.0 8.2 12.4



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#48
2,6-Dinitrotoluene
Concen: 38.334 ng/ul
RT: 13.887 min Scan# 1836
Delta R.T. -0.011 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:165 Resp: 161860
Ion Ratio Lower Upper
165 100
89 62.3 51.0 76.6
121 20.6 18.5 27.7





#50
Acenaphthylene
Concen: 34.429 ng/ul
RT: 14.010 min Scan# 1857
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion:152 Resp: 89951
Ion Ratio Lower Upper
152 100
151 20.3 16.5 24.7
153 13.0 10.3 15.5



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#51
3-Nitroaniline
Concen: 40.318 ng/ul
RT: 14.228 min Scan# 1894
Delta R.T. -0.011 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:138 Resp: 158776
Ion Ratio Lower Upper
138 100
108 12.6 10.8 16.2
92 121.6 103.3 154.9







#52
Acenaphthene
Concen: 34.155 ng/ul
RT: 14.363 min Scan# 1916
Delta R.T. 0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:153 Resp: 599678
Ion Ratio Lower Upper
153 100
152 46.5 37.4 56.2
154 87.7 70.3 105.5



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#53
2,4-Dinitrophenol
Concen: 41.814 ng/ul
RT: 14.451 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:184 Resp: 108370
Ion Ratio Lower Upper
184 100
63 70.0 65.6 98.4
154 71.7 53.8 80.8





#55
4-Nitrophenol
Concen: 38.726 ng/ul
RT: 14.551 min Scan# 1950
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:109 Resp: 13030
Ion Ratio Lower Upper
109 100
139 96.3 78.4 117.6
65 117.7 101.8 152.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#56
Dibenzofuran
Concen: 36.944 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:168 Resp: 892104
Ion Ratio Lower Upper
168 100
139 38.3 32.7 49.1







#57
2,4-Dinitrotoluene
Concen: 40.684 ng/ul
RT: 14.704 min Scan# 1975
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:165 Resp: 246970
Ion Ratio Lower Upper
165 100
63 68.3 59.7 89.5
182 2.8 2.4 3.6



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#58
2,3,4,6-Tetrachlorophenol
Concen: 43.110 ng/ul
RT: 14.940 min Scan# 2015
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:232 Resp: 234592
Ion Ratio Lower Upper
232 100
131 39.5 36.7 55.1
130 2.0 1.9 2.9
166 27.3 24.4 36.6





#59
Diethylphthalate
Concen: 37.328 ng/ul
RT: 15.157 min Scan# 2052
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion:149 Resp: 791411
Ion Ratio Lower Upper
149 100
177 22.9 17.9 26.9
150 12.8 10.2 15.2



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#61
Fluorene
Concen: 36.747 ng/ul
RT: 15.375 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:166 Resp: 737285
Ion Ratio Lower Upper
166 100
165 97.5 76.2 114.2
167 13.4 10.3 15.5





#62
4-Chlorophenyl-phenylether
Concen: 38.295 ng/ul
RT: 15.375 min Scan# 2089
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P



ClientSampleId :

SLCS684

Tgt Ion:204 Resp: 40762
Ion Ratio Lower Upper
204 100
206 33.1 25.4 38.2
141 56.6 53.0 79.6



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#63
4-Nitroaniline
Concen: 46.921 ng/ul
RT: 15.428 min Scan# 2098
Delta R.T. -0.011 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:138 Resp: 153990
Ion Ratio Lower Upper
138 100
92 56.0 47.6 71.4
108 107.4 90.5 135.7







#66
4,6-Dinitro-2-methylphenol
Concen: 38.535 ng/ul
RT: 15.487 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:198 Resp: 171672
Ion Ratio Lower Upper
198 100
182 3.6 3.3 4.9
77 23.5 24.4 36.6



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#67
N-Nitrosodiphenylamine
Concen: 34.605 ng/ul
RT: 15.604 min Scan# 2128
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:169 Resp: 648707
Ion Ratio Lower Upper
169 100
168 67.3 54.2 81.2
167 36.2 29.0 43.6





#68
4-Bromophenyl-phenylether
Concen: 38.033 ng/ul
RT: 16.304 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:248 Resp: 277364
Ion Ratio Lower Upper
248 100
250 96.3 76.2 114.4
141 69.0 57.2 85.8



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#69
Hexachlorobenzene
Concen: 40.011 ng/ul
RT: 16.398 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:284 Resp: 337950
Ion Ratio Lower Upper
284 100
142 29.6 27.8 41.8
249 31.6 25.4 38.0







#70
Atrazine
Concen: 41.522 ng/ul
RT: 16.610 min Scan# 21
Delta R.T. 0.018 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:200 Resp: 276000
Ion Ratio Lower Upper
200 100
173 23.6 20.3 30.5
215 47.3 37.8 56.8



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#71
Pentachlorophenol
Concen: 41.600 ng/ul
RT: 16.769 min Scan# 2326
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:266 Resp: 212386
Ion Ratio Lower Upper
266 100
264 64.5 49.8 74.6
268 64.9 50.4 75.6







#72
Phenanthrene
Concen: 35.622 ng/u1
RT: 17.175 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:178 Resp: 1223620
Ion Ratio Lower Upper
178 100
179 15.3 13.0 19.6
176 19.2 15.2 22.8



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#74
Anthracene
Concen: 34.929 ng/u1
RT: 17.275 min Scan# 2412
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:178 Resp: 1220349
Ion Ratio Lower Upper
178 100
179 15.9 12.0 18.0
176 18.5 14.8 22.2







#75
1,2,3,4-Tetrachlorobenzene
Concen: 34.043 ng/uL
RT: 13.063 min Scan# 1696
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :
BNA_P
ClientSampleId :
SLCS684

Tgt Ion:216 Resp: 340649
Ion Ratio Lower Upper
216 100
214 76.8 60.1 90.1
218 47.9 38.7 58.1



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#76
Pentachlorobenzene
Concen: 36.647 ng/uL
RT: 14.616 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:250 Resp: 366956
Ion Ratio Lower Upper
250 100
248 63.3 51.0 76.6
252 63.4 52.1 78.1





#77
 Carbazole
 Concen: 36.927 ng/ul
 RT: 17.569 min Scan# 2463
 Delta R.T. -0.006 min
 Lab File: BP020170.D
 Acq: 03 May 2024 09:55


Instrument :
 BNA_P
 ClientSampleId :
 SLCS684



Tgt Ion:167 Resp: 111142
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.4 26.0
 139 12.8 11.0 16.4


Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay






05/03/2024
 Supervised By :mohammad
 ahmed


05/06/2024



#78
 Di-n-butylphthalate
 Concen: 37.823 ng/ul
 RT: 18.169 min Scan# 2564
 Delta R.T. 0.006 min
 Lab File: BP020170.D
 Acq: 03 May 2024 09:55


Tgt Ion:149 Resp: 1379869
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 5.9 5.0 7.6





#80
Fluoranthene
Concen: 30.772 ng/u1
RT: 19.292 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:202 Resp: 1559714
Ion Ratio Lower Upper
202 100
101 7.7 7.0 10.4
100 6.3 5.5 8.3


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




05/03/2024
Supervised By :mohammad
ahmed



05/06/2024




#82
Pyrene
Concen: 31.765 ng/u1
RT: 19.669 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Tgt Ion:202 Resp: 1675212
Ion Ratio Lower Upper
202 100
101 8.8 8.1 12.1
100 7.6 6.8 10.2






#83
Butylbenzylphthalate
Concen: 32.661 ng/ul
RT: 20.639 min Scan# 2984
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Instrument :

BNA_P

ClientSampleId :

SLCS684





Tgt Ion:149 Resp: 653859
Ion Ratio Lower Upper
149 100
91 73.4 62.7 94.1
206 24.1 17.3 25.9

Manual Integrations


APPROVED

Reviewed By :Jagrut
Upadhyay




05/03/2024
Supervised By :mohammad
ahmed



05/06/2024




#84
3,3'-Dichlorobenzidine
Concen: 40.796 ng/ul
RT: 21.533 min Scan# 3136
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55




Tgt Ion:252 Resp: 611738
Ion Ratio Lower Upper
252 100
254 63.8 50.2 75.4
126 8.2 7.3 10.9





#85
Benzo(a)anthracene
Concen: 35.815 ng/ul
RT: 21.598 min Scan# 3148
Delta R.T. -0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:228 Resp: 1789784
Ion Ratio Lower Upper
228 100
229 19.4 15.8 23.8
226 26.9 21.5 32.3


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#86
Bis(2-ethylhexyl)phthalate
Concen: 33.499 ng/ul
RT: 21.557 min Scan# 3140
Delta R.T. 0.000 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:149 Resp: 967463
Ion Ratio Lower Upper
149 100
167 28.8 21.4 32.0
279 5.7 3.4 5.2#





#87
Chrysene
Concen: 36.562 ng/u1
RT: 21.674 min Scan# 3159
Delta R.T. 0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:228 Resp: 169345
Ion Ratio Lower Upper
228 100
226 29.3 24.2 36.4
229 19.3 15.9 23.9



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#89
Di-n-octyl phthalate
Concen: 31.628 ng/u1
RT: 22.857 min Scan# 3361
Delta R.T. 0.012 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:149 Resp: 1653739





#90
Benzo(b)fluoranthene
Concen: 36.590 ng/u1
RT: 23.933 min Scan# 3544
Delta R.T. 0.024 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55



Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:252 Resp: 1777029
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 6.3 6.2 9.2


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#91
Benzo(k)fluoranthene
Concen: 35.765 ng/u1
RT: 23.998 min Scan# 3555
Delta R.T. 0.012 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:252 Resp: 1770168
Ion Ratio Lower Upper
252 100
253 21.8 17.3 25.9
125 6.2 5.6 8.4





#93
Benzo(a)pyrene
Concen: 32.424 ng/ul
RT: 24.833 min Scan# 30
Delta R.T. 0.018 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:252 Resp: 1498324
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 7.2 6.6 10.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






05/03/2024
Supervised By :mohammad
ahmed


05/06/2024



#94
Indeno(1,2,3-cd)pyrene
Concen: 33.653 ng/ul m
RT: 28.804 min Scan# 4372
Delta R.T. 0.036 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Tgt Ion:276 Resp: 1906409
Ion Ratio Lower Upper
276 100
138 12.2 12.3 18.5#
277 23.8 19.0 28.4





#95
Dibenzo(a,h)anthracene
Concen: 33.362 ng/ul m
RT: 28.904 min Scan# 4389
Delta R.T. 0.047 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55


Instrument :
BNA_P
ClientSampleId :
SLCS684



Tgt Ion:278 Resp: 1563190
Ion Ratio Lower Upper
278 100
139 9.9 10.2 15.2
279 24.3 19.3 28.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay



05/03/2024
Supervised By :mohammad
ahmed

05/06/2024



#96
Benzo(g,h,i)perylene
Concen: 31.529 ng/ul m
RT: 29.968 min Scan# 4570
Delta R.T. 0.006 min
Lab File: BP020170.D
Acq: 03 May 2024 09:55

Tgt Ion:276 Resp: 1440501
Ion Ratio Lower Upper
276 100
138 13.5 11.6 17.4
277 23.3 19.4 29.0

