

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050323\
 Data File : BP014926.D
 Acq On : 03 May 2023 11:19
 Operator : CG/JU
 Sample : SSTD00565
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD005604

Quant Time: May 03 22:40:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 14:42:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.163	152	237521	20.000	ng/u1	0.00
20) Naphthalene-d8	10.999	136	1010829	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.810	164	590965	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.569	188	1108018	20.000	ng/u1	0.00
79) Chrysene-d12	21.651	240	678716	20.000	ng/u1	0.00
88) Perylene-d12	24.251	264	632443	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.458	96	12001	1.879	ng/uL	0.00
4) Pyridine-d5	3.905	84	68636	3.931	ng/u1	0.00
7) Phenol-d5	7.305	99	72610	3.526	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.481	67	55661	4.373	ng/u1	0.00
11) 2-Chlorophenol-d4	7.687	132	60372	3.987	ng/u1	0.00
15) 4-Methylphenol-d8	8.869	113	56502	3.560	ng/u1	0.00
21) Nitrobenzene-d5	9.340	128	25866	3.598	ng/u1	0.00
24) 2-Nitrophenol-d4	10.069	143	23170	3.139	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.610	165	54525	3.762	ng/u1	0.00
31) 4-Chloroaniline-d4	11.134	131	71146	3.393	ng/u1	0.00
46) Dimethylphthalate-d6	14.210	166	181096	4.330	ng/u1	0.00
49) Acenaphthylene-d8	14.504	160	207783	4.142	ng/u1	0.00
54) 4-Nitrophenol-d4	14.987	143	12436	1.709	ng/u1	0.00
60) Fluorene-d10	15.804	176	157356	4.373	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.910	200	13963	2.239	ng/u1	0.00
73) Anthracene-d10	17.663	188	206751	4.153	ng/u1	0.00
81) Pyrene-d10	19.886	212	207056	4.523	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.074	264	122623	3.979	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.499	88	13239	1.913	ng/uL#	96
5) Pyridine	3.928	79	77210	4.248	ng/u1	97
6) Benzaldehyde	7.287	77	17596	3.125	ng/u1	90
8) Phenol	7.334	94	82270	3.805	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.575	93	76944	4.282	ng/u1	96
12) 2-Chlorophenol	7.722	128	67289	4.102	ng/u1	98
13) 2-Methylphenol	8.605	108	57875	3.590	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.699	45	106313	4.664	ng/u1	99
16) Acetophenone	8.999	105	108362	4.176	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.975	70	55241	4.177	ng/u1	98
18) 4-Methylphenol	8.934	108	64124	3.598	ng/u1	99
19) Hexachloroethane	9.263	117	32143	4.517	ng/u1	96
22) Nitrobenzene	9.381	77	74917	3.966	ng/u1	99
23) Isophorone	9.910	82	153471	4.111	ng/u1	97
25) 2-Nitrophenol	10.105	139	28321	3.336	ng/u1	98
26) 2,4-Dimethylphenol	10.152	107	75094	4.076	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.393	93	103388	4.332	ng/u1	98
29) 2,4-Dichlorophenol	10.640	162	58758	3.919	ng/u1	96
30) Naphthalene	11.052	128	251369	4.524	ng/u1	99
32) 4-Chloroaniline	11.157	127	76755	3.508	ng/u1	98
33) Hexachlorobutadiene	11.334	225	44865	4.638	ng/u1	95
34) Caprolactam	11.916	113	13077	2.827	ng/u1	94
35) 4-Chloro-3-methylphenol	12.263	107	55374	3.593	ng/u1	98
36) 2-Methylnaphthalene	12.651	142	146186	4.254	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050323\
 Data File : BP014926.D
 Acq On : 03 May 2023 11:19
 Operator : CG/JU
 Sample : SST00565
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005604

Quant Time: May 03 22:40:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 14:42:28 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.863	142	158936	4.487	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	13.010	216	82290	4.464	ng/ul	97
40) Hexachlorocyclopentadiene	12.993	237	45412	3.653	ng/ul	99
41) 2,4,6-Trichlorophenol	13.246	196	43172	3.801	ng/ul	98
42) 2,4,5-Trichlorophenol	13.316	196	46139	3.769	ng/ul	99
43) 1,1'-Biphenyl	13.646	154	212020	4.488	ng/ul	99
44) 2-Chloronaphthalene	13.693	162	163263	4.436	ng/ul	99
45) 2-Nitroaniline	13.887	65	24710	2.914	ng/ul#	85
47) Dimethylphthalate	14.257	163	194075	4.413	ng/ul	99
48) 2,6-Dinitrotoluene	14.375	165	23892	3.074	ng/ul#	85
50) Acenaphthylene	14.534	152	246393	4.333	ng/ul	99
51) 3-Nitroaniline	14.710	138	11910	1.694	ng/ul#	82
52) Acenaphthene	14.875	153	178966	4.548	ng/ul	99
53) 2,4-Dinitrophenol	14.910	184	7285	1.540	ng/ul	96
55) 4-Nitrophenol	14.998	109	11517	2.079	ng/ul	91
56) Dibenzofuran	15.210	168	237052	4.464	ng/ul	99
57) 2,4-Dinitrotoluene	15.163	165	32972	3.029	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.428	232	37193	3.733	ng/ul	97
59) Diethylphthalate	15.616	149	172258	4.145	ng/ul	98
61) Fluorene	15.857	166	190864	4.557	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.845	204	93860	4.579	ng/ul	99
63) 4-Nitroaniline	15.869	138	8349	1.450	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.928	198	15533	2.348	ng/ul#	91
67) N-Nitrosodiphenylamine	16.063	169	142047	4.224	ng/ul	100
68) 4-Bromophenyl-phenylether	16.745	248	48249	4.186	ng/ul	97
69) Hexachlorobenzene	16.863	284	62852	4.629	ng/ul	99
70) Atrazine	17.010	200	39322	3.475	ng/ul	99
71) Pentachlorophenol	17.210	266	23733	2.994	ng/ul	94
72) Phenanthrene	17.610	178	279611	4.568	ng/ul	99
74) Anthracene	17.698	178	263671	4.326	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.610	216	85603	4.667	ng/ul	97
76) Pentachlorobenzene	15.128	250	81781	4.687	ng/ul	99
77) Carbazole	17.963	167	178911	3.490	ng/ul	99
78) Di-n-butylphthalate	18.498	149	203644	3.553	ng/ul#	98
80) Fluoranthene	19.563	202	245045	4.363	ng/ul	97
82) Pyrene	19.910	202	278154	4.674	ng/ul	97
83) Butylbenzylphthalate	20.769	149	59878	3.351	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.557	252	33796	2.977	ng/ul#	96
85) Benzo(a)anthracene	21.633	228	193975	4.282	ng/ul	97
86) Bis(2-ethylhexyl)phtha...	21.533	149	68300	3.038	ng/ul	100
87) Chrysene	21.686	228	211373	4.634	ng/ul	99
89) Di-n-octyl phthalate	22.527	149	102637	2.779	ng/ul	100
90) Benzo(b)fluoranthene	23.445	252	170574	4.173	ng/ul	98
91) Benzo(k)fluoranthene	23.498	252	167313	4.022	ng/ul	99
93) Benzo(a)pyrene	24.133	252	144706	4.037	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.004	276	172332	3.956	ng/ul	95
95) Dibenzo(a,h)anthracene	27.027	278	132039	3.719	ng/ul	98
96) Benzo(g,h,i)perylene	27.851	276	161285	4.334	ng/ul	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050323\
 Data File : BP014926.D
 Acq On : 03 May 2023 11:19
 Operator : CG/JU
 Sample : SST00565
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005604

Quant Time: May 03 22:40:10 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 14:42:28 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 1.913 ng/uL
RT: 3.499 min Scan# 19
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion: 88 Resp: 13239
Ion Ratio Lower Upper
88 100
43 32.6 20.6 31.0#
58 67.6 54.4 81.6



#5
Pyridine
Concen: 4.248 ng/uL
RT: 3.928 min Scan# 92
Delta R.T. 0.012 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion: 79 Resp: 77210
Ion Ratio Lower Upper
79 100
52 70.7 54.7 82.1
51 34.6 25.5 38.3









#6
Benzaldehyde
Concen: 3.125 ng/ul
RT: 7.287 min Scan# 604
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion: 77 Resp: 17596
Ion Ratio Lower Upper
77 100
105 94.2 80.7 121.1
106 83.1 77.6 116.4



#8
Phenol
Concen: 3.805 ng/ul
RT: 7.334 min Scan# 671
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion: 94 Resp: 82270
Ion Ratio Lower Upper
94 100
65 25.7 21.4 32.0
66 31.9 25.8 38.8









#10
Bis(2-Chloroethyl)ether
Concen: 4.282 ng/ul
RT: 7.575 min Scan# 712
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion: 93 Resp: 76944
Ion Ratio Lower Upper
93 100
63 76.8 58.1 87.1
95 32.3 26.0 39.0



#12
2-Chlorophenol
Concen: 4.102 ng/ul
RT: 7.722 min Scan# 737
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion: 128 Resp: 67289
Ion Ratio Lower Upper
128 100
64 45.1 36.6 55.0
130 30.7 25.8 38.8





#13

2-Methylphenol

Concen: 3.590 ng/ul

RT: 8.605 min Scan# 887

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :





SSTD005604

Tgt Ion:108 Resp: 57875

Ion Ratio Lower Upper

108 100

107 92.7 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 4.664 ng/ul

RT: 8.699 min Scan# 903

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion: 45 Resp: 106313


Ion Ratio Lower Upper

45 100

77 15.6 12.3 18.5

79 12.3 9.5 14.3









#16
Acetophenone
Concen: 4.176 ng/ul
RT: 8.999 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:105 Resp: 108362
Ion Ratio Lower Upper
105 100
77 77.6 61.2 91.8
51 23.6 19.4 29.2



#17
N-Nitroso-di-n-propylamine
Concen: 4.177 ng/ul
RT: 8.975 min Scan# 950
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion: 70 Resp: 55241
Ion Ratio Lower Upper
70 100
42 53.4 41.9 62.9
101 10.1 7.7 11.5
130 18.9 16.2 24.2









#18
4-Methylphenol
Concen: 3.598 ng/ul
RT: 8.934 min Scan# 944
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:108 Resp: 64124
Ion Ratio Lower Upper
108 100
107 107.5 87.2 130.8



#19
Hexachloroethane
Concen: 4.517 ng/ul
RT: 9.263 min Scan# 999
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:117 Resp: 32143
Ion Ratio Lower Upper
117 100
201 96.8 78.9 118.3
199 66.7 48.3 72.5









#22
Nitrobenzene
Concen: 3.966 ng/ul
RT: 9.381 min Scan# 1109
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion: 77 Resp: 74917
Ion Ratio Lower Upper
77 100
123 44.3 34.8 52.2
65 13.0 10.2 15.2



#23
Isophorone
Concen: 4.111 ng/ul
RT: 9.910 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion: 82 Resp: 153471
Ion Ratio Lower Upper
82 100
95 6.9 5.3 7.9
138 15.9 13.9 20.9









#25
2-Nitrophenol
Concen: 3.336 ng/ul
RT: 10.105 min Scan# 1142
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:139 Resp: 28321
Ion Ratio Lower Upper
139 100
65 52.3 40.5 60.7
109 28.6 22.4 33.6



#26
2,4-Dimethylphenol
Concen: 4.076 ng/ul
RT: 10.152 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:107 Resp: 75094
Ion Ratio Lower Upper
107 100
121 49.6 40.2 60.2
122 87.9 69.6 104.4









#27
Bis(2-Chloroethoxy)methane
Concen: 4.332 ng/ul
RT: 10.393 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion: 93 Resp: 103388
Ion Ratio Lower Upper
93 100
95 32.2 26.1 39.1
123 9.4 9.0 13.4



#29
2,4-Dichlorophenol
Concen: 3.919 ng/ul
RT: 10.640 min Scan# 1233
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion: 162 Resp: 58758
Ion Ratio Lower Upper
162 100
164 67.6 51.5 77.3
98 37.8 28.8 43.2









#30
Naphthalene
Concen: 4.524 ng/ul
RT: 11.052 min Scan# 1303
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:128 Resp: 251369
Ion Ratio Lower Upper
128 100
129 11.2 8.7 13.1
127 13.3 10.4 15.6



#32
4-Chloroaniline
Concen: 3.508 ng/ul
RT: 11.157 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:127 Resp: 76755
Ion Ratio Lower Upper
127 100
129 33.3 25.5 38.3









#33
Hexachlorobutadiene
Concen: 4.638 ng/ul
RT: 11.334 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:225 Resp: 44865
Ion Ratio Lower Upper
225 100
223 67.4 50.6 76.0
227 68.1 51.4 77.0



#34
Caprolactam
Concen: 2.827 ng/ul
RT: 11.916 min Scan# 1450
Delta R.T. -0.012 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:113 Resp: 13077
Ion Ratio Lower Upper
113 100
55 149.1 128.5 192.7
56 114.7 94.3 141.5









#35
4-Chloro-3-methylphenol
Concen: 3.593 ng/ul
RT: 12.263 min Scan# 1509
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:107 Resp: 55374
Ion Ratio Lower Upper
107 100
144 24.8 20.6 30.8
142 81.3 63.2 94.8



#36
2-Methylnaphthalene
Concen: 4.254 ng/ul
RT: 12.651 min Scan# 1575
Delta R.T. 0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:142 Resp: 146186
Ion Ratio Lower Upper
142 100
141 89.4 70.5 105.7





#37

1-Methylnaphthalene

Concen: 4.487 ng/ul

RT: 12.863 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:142 Resp: 158936

Ion Ratio Lower Upper

142 100

141 92.5 74.4 111.6

116 3.9 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.464 ng/ul

RT: 13.010 min Scan# 1636

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Tgt Ion:216 Resp: 82290




Ion Ratio Lower Upper


216 100

214 75.8 63.0 94.4

179 21.2 16.9 25.3

108 20.7 14.9 22.3





#40

Hexachlorocyclopentadiene

Concen: 3.653 ng/ul

RT: 12.993 min Scan# 1633

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:237 Resp: 45412

Ion Ratio Lower Upper

237 100

235 61.9 49.9 74.9

272 13.6 10.0 15.0



#41

2,4,6-Trichlorophenol

Concen: 3.801 ng/ul

RT: 13.246 min Scan# 1676

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:196 Resp: 43172


Ion Ratio Lower Upper

196 100

198 93.4 76.8 115.2

200 30.0 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 3.769 ng/ul
RT: 13.316 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:196 Resp: 46139
Ion Ratio Lower Upper
196 100
198 92.5 74.2 111.4
200 29.4 24.6 37.0



#43
1,1'-Biphenyl
Concen: 4.488 ng/ul
RT: 13.646 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:154 Resp: 212020
Ion Ratio Lower Upper
154 100
153 39.2 31.9 47.9
76 13.7 11.3 16.9





#44

2-Chloronaphthalene

Concen: 4.436 ng/ul

RT: 13.693 min Scan# 1752

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:162 Resp: 163263

Ion Ratio Lower Upper

162 100

164 33.4 26.3 39.5

127 38.3 30.4 45.6



#45

2-Nitroaniline

Concen: 2.914 ng/ul

RT: 13.887 min Scan# 1785

Delta R.T. -0.006 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion: 65 Resp: 24710


Ion Ratio Lower Upper

65 100

92 70.9 58.3 87.5

138 84.5 87.9 131.9#









#47
Dimethylphthalate
Concen: 4.413 ng/ul
RT: 14.257 min Scan# 1848
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:163 Resp: 194075
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 9.3 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 3.074 ng/ul
RT: 14.375 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:165 Resp: 23892
Ion Ratio Lower Upper
165 100
89 71.5 48.1 72.1
121 28.4 17.3 25.9#









#50
Acenaphthylene
Concen: 4.333 ng/ul
RT: 14.534 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:152 Resp: 246393
Ion Ratio Lower Upper
152 100
151 20.0 16.1 24.1
153 13.4 10.4 15.6



#51
3-Nitroaniline
Concen: 1.694 ng/ul
RT: 14.710 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:138 Resp: 11910
Ion Ratio Lower Upper
138 100
108 13.6 8.5 12.7#
92 140.7 95.8 143.6





#52

Acenaphthene

Concen: 4.548 ng/ul

RT: 14.875 min Scan# 1953

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:153 Resp: 178966

Ion Ratio Lower Upper

153 100

152 49.5 38.8 58.2

154 87.5 69.6 104.4



#53

2,4-Dinitrophenol

Concen: 1.540 ng/ul

RT: 14.910 min Scan# 1959

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:184 Resp: 7285


Ion Ratio Lower Upper

184 100

63 74.5 56.3 84.5

154 63.5 48.9 73.3









#55
4-Nitrophenol
Concen: 2.079 ng/ul
RT: 14.998 min Scan# 1975
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:109 Resp: 11517
Ion Ratio Lower Upper
109 100
139 137.5 115.1 172.7
65 154.6 110.6 166.0



#56
Dibenzofuran
Concen: 4.464 ng/ul
RT: 15.210 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:168 Resp: 237052
Ion Ratio Lower Upper
168 100
139 36.3 28.6 42.8









#57
2,4-Dinitrotoluene
Concen: 3.029 ng/ul
RT: 15.163 min Scan# 2002
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:165 Resp: 32972
Ion Ratio Lower Upper
165 100
63 44.1 35.7 53.5
182 2.1 2.2 3.4#



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.733 ng/ul
RT: 15.428 min Scan# 2047
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:232 Resp: 37193
Ion Ratio Lower Upper
232 100
131 49.9 38.2 57.4
130 2.5 2.1 3.1
166 34.3 26.6 39.8









#59
Diethylphthalate
Concen: 4.145 ng/ul
RT: 15.616 min Scan# 2080
Delta R.T. -0.006 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:149 Resp: 172258
Ion Ratio Lower Upper
149 100
177 23.2 17.5 26.3
150 12.4 9.6 14.4



#61
Fluorene
Concen: 4.557 ng/ul
RT: 15.857 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:166 Resp: 190864
Ion Ratio Lower Upper
166 100
165 96.8 79.3 118.9
167 13.3 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 4.579 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:204 Resp: 93860
Ion Ratio Lower Upper
204 100
206 33.3 26.2 39.2
141 57.9 46.8 70.2



#63
4-Nitroaniline
Concen: 1.450 ng/ul
RT: 15.869 min Scan# 2122
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:138 Resp: 8349
Ion Ratio Lower Upper
138 100
92 56.8 43.7 65.5
108 68.2 57.4 86.2









#66
4,6-Dinitro-2-methylphenol
Concen: 2.348 ng/ul
RT: 15.928 min Scan# 2132
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:198 Resp: 15533
Ion Ratio Lower Upper
198 100
182 8.4 3.9 5.9#
77 31.4 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 4.224 ng/ul
RT: 16.063 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:169 Resp: 142047
Ion Ratio Lower Upper
169 100
168 66.9 53.5 80.3
167 36.7 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 4.186 ng/ul
RT: 16.745 min Scan# 2271
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19




Instrument :
BNA_P
ClientSampleId :
SSTD005604


Tgt Ion:248 Resp: 48249
Ion Ratio Lower Upper
248 100
250 95.1 79.5 119.3
141 73.9 60.2 90.4



#69
Hexachlorobenzene
Concen: 4.629 ng/ul
RT: 16.863 min Scan# 2291
Delta R.T. -0.000 min
Lab File: BP014926.D
Acq: 03 May 2023 11:19

Tgt Ion:284 Resp: 62852
Ion Ratio Lower Upper
284 100
142 37.5 30.2 45.2
249 31.1 24.5 36.7





#70

Atrazine

Concen: 3.475 ng/ul

RT: 17.010 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:200 Resp: 39322

Ion Ratio Lower Upper

200 100

173 28.0 21.8 32.6

215 48.5 39.0 58.4



#71

Pentachlorophenol

Concen: 2.994 ng/ul

RT: 17.210 min Scan# 2350

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:266 Resp: 23733


Ion Ratio Lower Upper

266 100

264 64.9 48.9 73.3

268 67.7 49.6 74.4





#72

Phenanthrene

Concen: 4.568 ng/ul

RT: 17.610 min Scan# 2418

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:178 Resp: 279611

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 19.7 15.4 23.0



#74

Anthracene

Concen: 4.326 ng/ul

RT: 17.698 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:178 Resp: 263671


Ion Ratio Lower Upper

178 100

179 15.8 12.6 18.8

176 19.2 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 4.667 ng/uL

RT: 13.610 min Scan# 1739

Delta R.T. -0.006 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:216 Resp: 85603

Ion Ratio Lower Upper

216 100

214 79.3 60.5 90.7

218 47.3 38.6 58.0



#76

Pentachlorobenzene

Concen: 4.687 ng/uL

RT: 15.128 min Scan# 1996

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:250 Resp: 81781


Ion Ratio Lower Upper

250 100

248 62.9 50.5 75.7

252 61.0 50.5 75.7





#77

Carbazole

Concen: 3.490 ng/ul

RT: 17.963 min Scan# 2478

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:167 Resp: 178911

Ion Ratio Lower Upper

167 100

166 21.8 17.1 25.7

139 13.1 10.1 15.1



#78

Di-n-butylphthalate

Concen: 3.553 ng/ul

RT: 18.498 min Scan# 2569

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:149 Resp: 203644


Ion Ratio Lower Upper

149 100

150 9.0 7.4 11.2

104 6.1 4.1 6.1#





#80

Fluoranthene

Concen: 4.363 ng/ul

RT: 19.563 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:202 Resp: 245045

Ion Ratio Lower Upper

202 100

101 10.7 10.1 15.1

100 9.3 7.7 11.5



#82

Pyrene

Concen: 4.674 ng/ul

RT: 19.910 min Scan# 2809

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:202 Resp: 278154


Ion Ratio Lower Upper

202 100

101 13.1 11.9 17.9

100 11.4 9.7 14.5





#83

Butylbenzylphthalate

Concen: 3.351 ng/ul

RT: 20.769 min Scan# 2955

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:149 Resp: 59878

Ion Ratio Lower Upper

149 100

91 80.1 60.1 90.1

206 19.7 16.0 24.0



#84

3,3'-Dichlorobenzidine

Concen: 2.977 ng/ul

RT: 21.557 min Scan# 3089

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:252 Resp: 33796


Ion Ratio Lower Upper

252 100

254 63.7 53.3 79.9

126 14.0 9.0 13.6#





#85

Benzo(a)anthracene

Concen: 4.282 ng/ul

RT: 21.633 min Scan# 3102

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:228 Resp: 193975

Ion Ratio Lower Upper

228 100

229 20.8 15.9 23.9

226 25.5 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.038 ng/ul

RT: 21.533 min Scan# 3085

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:149 Resp: 68300


Ion Ratio Lower Upper

149 100

167 27.3 22.0 33.0

279 3.5 2.9 4.3





#87

Chrysene

Concen: 4.634 ng/ul

RT: 21.686 min Scan# 3111

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:228 Resp: 211373

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 20.4 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 2.779 ng/ul




RT: 22.527 min Scan# 3254


Delta R.T. 0.006 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Tgt Ion:149 Resp: 102637





#90

Benzo(b)fluoranthene

Concen: 4.173 ng/ul

RT: 23.445 min Scan# 3410

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:252 Resp: 170574

Ion Ratio Lower Upper

252 100

253 23.1 17.4 26.2

125 10.5 8.4 12.6



#91

Benzo(k)fluoranthene

Concen: 4.022 ng/ul

RT: 23.498 min Scan# 3419

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:252 Resp: 167313


Ion Ratio Lower Upper

252 100

253 22.6 17.8 26.6

125 10.9 8.9 13.3





#93

Benzo(a)pyrene

Concen: 4.037 ng/ul

RT: 24.133 min Scan# 3527

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:252 Resp: 144706

Ion Ratio Lower Upper

252 100

253 22.4 17.5 26.3

125 11.3 9.4 14.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 3.956 ng/ul

RT: 27.004 min Scan# 4015

Delta R.T. -0.006 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19




Tgt Ion:276 Resp: 172332


Ion Ratio Lower Upper

276 100

138 20.7 19.4 29.0

277 23.0 19.4 29.2





#95

Dibenzo(a,h)anthracene

Concen: 3.719 ng/ul

RT: 27.027 min Scan# 4019

Delta R.T. -0.000 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Instrument :

BNA_P

ClientSampleId :

SSTD005604





Tgt Ion:278 Resp: 132039

Ion Ratio Lower Upper

278 100

139 15.0 12.9 19.3

279 23.3 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 4.334 ng/ul

RT: 27.851 min Scan# 4159

Delta R.T. -0.006 min

Lab File: BP014926.D

Acq: 03 May 2023 11:19

Tgt Ion:276 Resp: 161285

Ion Ratio Lower Upper

276 100

138 20.1 18.0 27.0

277 23.9 19.8 29.8

