

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050323\
 Data File : BP014935.D
 Acq On : 03 May 2023 18:04
 Operator : CG/JU
 Sample : 02561-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK

Quant Time: May 04 06:18:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 23:33:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.169	152	261261	20.000	ng/ul	0.00
20) Naphthalene-d8	11.005	136	1104554	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.810	164	757016	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.569	188	1593769	20.000	ng/ul	0.00
79) Chrysene-d12	21.651	240	1301509	20.000	ng/ul	0.00
88) Perylene-d12	24.251	264	1281770	20.000	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.464	96	243	0.035	ng/uL	0.00
4) Pyridine-d5	3.905	84	149	0.008	ng/ul	0.00
7) Phenol-d5	7.299	99	51	0.002	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.487	67	64	0.005	ng/ul	0.00
11) 2-Chlorophenol-d4	7.705	132	99	0.006	ng/ul	0.02
15) 4-Methylphenol-d8	8.887	113	25	0.001	ng/ul	0.01
21) Nitrobenzene-d5	9.340	128	32	0.004	ng/ul	0.00
24) 2-Nitrophenol-d4	10.063	143	19	0.002	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.622	165	12	0.001	ng/ul	0.01
31) 4-Chloroaniline-d4	11.146	131	34	0.001	ng/ul	0.01
46) Dimethylphthalate-d6	14.222	166	27	0.001	ng/ul	0.01
49) Acenaphthylene-d8	14.510	160	97	0.002	ng/ul	0.00
54) 4-Nitrophenol-d4	14.987	143	6	0.001	ng/ul	0.00
60) Fluorene-d10	15.810	176	359	0.008	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.916	200	15	0.002	ng/ul	0.00
73) Anthracene-d10	17.569	188	1593769	22.256	ng/ul	-0.09
81) Pyrene-d10	19.886	212	307	0.003	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.251	264	857998	13.738	ng/ul	0.17

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.517	88	335	0.044	ng/uL#	16
5) Pyridine	3.917	79	148	0.007	ng/ul#	1
6) Benzaldehyde	7.299	77	95	0.015	ng/ul#	20
8) Phenol	7.328	94	68	0.003	ng/ul#	1
10) Bis(2-Chloroethyl)ether	7.569	93	54	0.003	ng/ul#	75
12) 2-Chlorophenol	7.740	128	22	0.001	ng/ul#	45
13) 2-Methylphenol	8.628	108	15	0.001	ng/ul	89
14) 2,2'-oxybis(1-Chloropr...	8.705	45	110	0.004	ng/ul#	1
16) Acetophenone	9.011	105	74	0.003	ng/ul#	13
17) N-Nitroso-di-n-propyla...	9.011	70	96	0.007	ng/ul#	1
18) 4-Methylphenol	9.063	108	8	0.000	ng/ul	96
19) Hexachloroethane	9.263	117	47	0.006	ng/ul#	14
22) Nitrobenzene	9.393	77	60	0.003	ng/ul#	26
23) Isophorone	9.922	82	73	0.002	ng/ul#	1
25) 2-Nitrophenol	10.099	139	13	0.001	ng/ul#	1
26) 2,4-Dimethylphenol	10.163	107	34	0.002	ng/ul#	52
27) Bis(2-Chloroethoxy)met...	10.363	93	37	0.001	ng/ul	97
29) 2,4-Dichlorophenol	10.640	162	13	0.001	ng/ul#	69
30) Naphthalene	11.116	128	108	0.002	ng/ul	98
32) 4-Chloroaniline	11.181	127	11	0.000	ng/ul#	1
33) Hexachlorobutadiene	11.363	225	6	0.001	ng/ul	86
34) Caprolactam	11.940	113	17	0.003	ng/ul#	1
35) 4-Chloro-3-methylphenol	12.257	107	9	0.001	ng/ul#	1
36) 2-Methylnaphthalene	12.652	142	10	0.000	ng/ul	94

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050323\
 Data File : BP014935.D
 Acq On : 03 May 2023 18:04
 Operator : CG/JU
 Sample : 02561-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK


Quant Time: May 04 06:18:48 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 03 23:33:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.887	142	6	0.000	ng/ul#	55
39) 1,2,4,5-Tetrachloroben...	13.016	216	19	0.001	ng/ul#	1
40) Hexachlorocyclopentadiene	13.010	237	36	0.002	ng/ul#	36
41) 2,4,6-Trichlorophenol	13.246	196	17	0.001	ng/ul#	24
42) 2,4,5-Trichlorophenol	13.310	196	18	0.001	ng/ul#	21
43) 1,1'-Biphenyl	13.646	154	38	0.001	ng/ul#	4
44) 2-Chloronaphthalene	13.693	162	24	0.001	ng/ul#	1
45) 2-Nitroaniline	13.893	65	58	0.005	ng/ul#	29
47) Dimethylphthalate	14.257	163	62	0.001	ng/ul#	90
48) 2,6-Dinitrotoluene	14.246	165	39	0.004	ng/ul	93
50) Acenaphthylene	14.540	152	44	0.001	ng/ul#	90
51) 3-Nitroaniline	14.716	138	17	0.002	ng/ul#	1
52) Acenaphthene	14.875	153	46	0.001	ng/ul#	73
53) 2,4-Dinitrophenol	14.904	184	20	0.003	ng/ul	94
55) 4-Nitrophenol	15.010	109	19	0.003	ng/ul#	1
56) Dibenzofuran	15.269	168	37	0.001	ng/ul	94
57) 2,4-Dinitrotoluene	15.169	165	62	0.004	ng/ul#	15
58) 2,3,4,6-Tetrachlorophenol	15.428	232	5	0.000	ng/ul#	1
59) Diethylphthalate	15.628	149	126	0.002	ng/ul#	1
61) Fluorene	15.845	166	23	0.000	ng/ul#	1
62) 4-Chlorophenyl-phenyle...	15.857	204	21	0.001	ng/ul#	66
63) 4-Nitroaniline	15.875	138	9	0.001	ng/ul#	54
66) 4,6-Dinitro-2-methylph...	15.934	198	4	0.000	ng/ul#	1
67) N-Nitrosodiphenylamine	16.057	169	24	0.000	ng/ul	98
68) 4-Bromophenyl-phenylether	16.845	248	14	0.001	ng/ul	89
69) Hexachlorobenzene	16.845	284	75	0.004	ng/ul#	72
70) Atrazine	16.963	200	26	0.002	ng/ul	92
71) Pentachlorophenol	17.263	266	30	0.003	ng/ul	97
72) Phenanthrene	17.610	178	267	0.003	ng/ul#	80
74) Anthracene	17.698	178	129	0.001	ng/ul#	1
75) 1,2,3,4-Tetrachloroben...	13.604	216	29	0.001	ng/ul#	66
76) Pentachlorobenzene	15.151	250	19	0.001	ng/ul	91
77) Carbazole	17.963	167	62	0.001	ng/ul#	59
78) Di-n-butylphthalate	18.498	149	306	0.004	ng/ul#	38
80) Fluoranthene	19.569	202	322	0.003	ng/ul#	81
82) Pyrene	19.922	202	578	0.005	ng/ul#	85
83) Butylbenzylphthalate	20.780	149	140	0.004	ng/ul#	61
84) 3,3'-Dichlorobenzidine	21.545	252	37	0.002	ng/ul#	29
85) Benzo(a)anthracene	21.651	228	3553	0.041	ng/ul#	64
86) Bis(2-ethylhexyl)phtha...	21.539	149	691	0.016	ng/ul#	79
87) Chrysene	21.680	228	626	0.007	ng/ul#	59
89) Di-n-octyl phthalate	22.527	149	283	0.004	ng/ul	100
90) Benzo(b)fluoranthene	23.445	252	453	0.005	ng/ul#	1
91) Benzo(k)fluoranthene	23.516	252	359	0.004	ng/ul#	10
93) Benzo(a)pyrene	24.133	252	156	0.002	ng/ul#	1
94) Indeno(1,2,3-cd)pyrene	27.068	276	105	0.001	ng/ul	95
95) Dibenzo(a,h)anthracene	27.010	278	624	0.009	ng/ul#	16
96) Benzo(g,h,i)perylene	27.839	276	60	0.001	ng/ul#	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 04 06:18:48 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 03 23:33:17 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 0.044 ng/uL
RT: 3.517 min Scan# 21
Delta R.T. 0.018 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion: 88 Resp: 335
Ion Ratio Lower Upper
88 100
43 78.8 20.6 31.0#
58 6.1 54.4 81.6#



#5
Pyridine
Concen: 0.007 ng/uL
RT: 3.917 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion: 79 Resp: 148
Ion Ratio Lower Upper
79 100
52 48.6 54.7 82.1#
51 429.3 25.5 38.3#









#6
Benzaldehyde
Concen: 0.015 ng/ul
RT: 7.299 min Scan# 664
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion: 77 Resp: 95
Ion Ratio Lower Upper
77 100
105 19.4 80.7 121.1#
106 19.4 77.6 116.4#



#8
Phenol
Concen: 0.003 ng/ul
RT: 7.328 min Scan# 670
Delta R.T. -0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion: 94 Resp: 68
Ion Ratio Lower Upper
94 100
65 129.4 21.4 32.0#
66 76.5 25.8 38.8#









#10
Bis(2-Chloroethyl)ether
Concen: 0.003 ng/ul
RT: 7.569 min Scan# 711
Delta R.T. -0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion: 93	Resp:	54
Ion	Ratio	Lower Upper
93	100	
63	45.9	58.1 87.1#
95	27.5	26.0 39.0



#12
2-Chlorophenol
Concen: 0.001 ng/ul
RT: 7.740 min Scan# 740
Delta R.T. 0.018 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion: 128	Resp:	22
Ion	Ratio	Lower Upper
128	100	
64	0.0	36.6 55.0#
130	13.5	25.8 38.8#





#13

2-Methylphenol

Concen: 0.001 ng/ul

RT: 8.628 min Scan# 891

Delta R.T. 0.023 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04

Instrument :

BNA_P

ClientSampleId :





SVOC-GPC2-BLANK

Tgt Ion:108 Resp: 15

Ion Ratio Lower Upper

108 100

107 100.0 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.004 ng/ul

RT: 8.705 min Scan# 904

Delta R.T. 0.006 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04




Tgt Ion: 45 Resp: 110


Ion Ratio Lower Upper

45 100

77 126.3 12.3 18.5#

79 114.9 9.5 14.3#









#16
Acetophenone
Concen: 0.003 ng/ul
RT: 9.011 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:105 Resp: 74
Ion Ratio Lower Upper
105 100
77 97.0 61.2 91.8#
51 165.2 19.4 29.2#



#17
N-Nitroso-di-n-propylamine
Concen: 0.007 ng/ul
RT: 9.011 min Scan# 956
Delta R.T. 0.029 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion: 70 Resp: 96
Ion Ratio Lower Upper
70 100
42 292.2 41.9 62.9#
101 15.6 7.7 11.5#
130 31.2 16.2 24.2#





#18

4-Methylphenol

Concen: 0.000 ng/ul

RT: 9.063 min Scan# 9

Delta R.T. 0.123 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04

Instrument :

BNA_P

ClientSampleId :





SVOC-GPC2-BLANK

Tgt Ion:108 Resp: 8

Ion Ratio Lower Upper

108 100

107 104.5 87.2 130.8



#19

Hexachloroethane

Concen: 0.006 ng/ul

RT: 9.263 min Scan# 999

Delta R.T. -0.000 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04




Tgt Ion:117 Resp: 47


Ion Ratio Lower Upper

117 100

201 0.0 78.9 118.3#

199 10.9 48.3 72.5#









#22
Nitrobenzene
Concen: 0.003 ng/ul
RT: 9.393 min Scan# 1101
Delta R.T. 0.012 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion: 77 Resp: 60
Ion Ratio Lower Upper
77 100
123 10.5 34.8 52.2#
65 72.8 10.2 15.2#



#23
Isophorone
Concen: 0.002 ng/ul
RT: 9.922 min Scan# 1111
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion: 82 Resp: 73
Ion Ratio Lower Upper
82 100
95 143.0 5.3 7.9#
138 24.4 13.9 20.9#









#25
2-Nitrophenol
Concen: 0.001 ng/ul
RT: 10.099 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:139 Resp: 13
Ion Ratio Lower Upper
139 100
65 340.0 40.5 60.7#
109 280.0 22.4 33.6#



#26
2,4-Dimethylphenol
Concen: 0.002 ng/ul
RT: 10.163 min Scan# 1152
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:107 Resp: 34
Ion Ratio Lower Upper
107 100
121 10.8 40.2 60.2#
122 47.0 69.6 104.4#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 0.001 ng/ul
 RT: 10.363 min Scan# 1192
 Delta R.T. -0.035 min
 Lab File: BP014935.D
 Acq: 03 May 2023 18:04




Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK


Tgt Ion: 93	Resp:	37
Ion Ratio	Lower	Upper
93	100	
95	31.7	26.1 39.1
123	13.4	9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 0.001 ng/ul
 RT: 10.640 min Scan# 1233
 Delta R.T. -0.000 min
 Lab File: BP014935.D
 Acq: 03 May 2023 18:04

Tgt Ion: 162	Resp:	13
Ion Ratio	Lower	Upper
162	100	
164	42.2	51.5 77.3#
98	57.8	28.8 43.2#









#30
Naphthalene
Concen: 0.002 ng/ul
RT: 11.116 min Scan# 11
Delta R.T. 0.065 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:128 Resp: 108
Ion Ratio Lower Upper
128 100
129 9.2 8.7 13.1
127 13.3 10.4 15.6



#32
4-Chloroaniline
Concen: 0.000 ng/ul
RT: 11.181 min Scan# 1325
Delta R.T. 0.023 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:127 Resp: 11
Ion Ratio Lower Upper
127 100
129 100.0 25.5 38.3#









#33
Hexachlorobutadiene
Concen: 0.001 ng/ul
RT: 11.363 min Scan# 11
Delta R.T. 0.029 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:	225	Resp:	6
Ion	Ratio	Lower	Upper
225	100		
223	52.9	50.6	76.0
227	52.9	51.4	77.0



#34
Caprolactam
Concen: 0.003 ng/ul
RT: 11.940 min Scan# 1454
Delta R.T. 0.012 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:	113	Resp:	17
Ion	Ratio	Lower	Upper
113	100		
55	585.0	128.5	192.7#
56	385.0	94.3	141.5#









#35
4-Chloro-3-methylphenol
Concen: 0.001 ng/ul
RT: 12.257 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:107 Resp: 9
Ion Ratio Lower Upper
107 100
144 34.3 20.6 30.8#
142 211.4 63.2 94.8#



#36
2-Methylnaphthalene
Concen: 0.000 ng/ul
RT: 12.652 min Scan# 1575
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:142 Resp: 10
Ion Ratio Lower Upper
142 100
141 93.3 70.5 105.7









#37
1-Methylnaphthalene
Concen: 0.000 ng/ul
RT: 12.887 min Scan# 1612
Delta R.T. 0.018 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:142 Resp: 6
Ion Ratio Lower Upper
142 100
141 68.2 74.4 111.6#
116 177.3 2.9 4.3#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 0.001 ng/ul
RT: 13.016 min Scan# 1637
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:216 Resp: 19
Ion Ratio Lower Upper
216 100
214 38.9 63.0 94.4#
179 161.1 16.9 25.3#
108 483.3 14.9 22.3#









#40
Hexachlorocyclopentadiene
Concen: 0.002 ng/ul
RT: 13.010 min Scan# 1636
Delta R.T. 0.018 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:237 Resp: 36
Ion Ratio Lower Upper
237 100
235 90.9 49.9 74.9#
272 90.9 10.0 15.0#



#41
2,4,6-Trichlorophenol
Concen: 0.001 ng/ul
RT: 13.246 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:196 Resp: 17
Ion Ratio Lower Upper
196 100
198 0.0 76.8 115.2#
200 34.1 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 0.001 ng/ul
RT: 13.310 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:196 Resp: 18
Ion Ratio Lower Upper
196 100
198 0.0 74.2 111.4#
200 44.8 24.6 37.0#



#43
1,1'-Biphenyl
Concen: 0.001 ng/ul
RT: 13.646 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:154 Resp: 38
Ion Ratio Lower Upper
154 100
153 90.0 31.9 47.9#
76 70.0 11.3 16.9#









#44
2-Chloronaphthalene
Concen: 0.001 ng/ul
RT: 13.693 min Scan# 1752
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:162 Resp: 24
Ion Ratio Lower Upper
162 100
164 104.0 26.3 39.5#
127 232.0 30.4 45.6#



#45
2-Nitroaniline
Concen: 0.005 ng/ul
RT: 13.893 min Scan# 1786
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion: 65 Resp: 58
Ion Ratio Lower Upper
65 100
92 145.3 58.3 87.5#
138 45.3 87.9 131.9#









#47
Dimethylphthalate
Concen: 0.001 ng/ul
RT: 14.257 min Scan# 1848
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:163 Resp: 62
Ion Ratio Lower Upper
163 100
194 14.1 3.1 4.7#
164 10.9 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 0.004 ng/ul
RT: 14.246 min Scan# 1846
Delta R.T. -0.135 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:165 Resp: 39
Ion Ratio Lower Upper
165 100
89 66.0 48.1 72.1
121 24.5 17.3 25.9









#50
Acenaphthylene
Concen: 0.001 ng/ul
RT: 14.540 min Scan# 1896
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:152 Resp: 44
Ion Ratio Lower Upper
152 100
151 26.9 16.1 24.1#
153 11.5 10.4 15.6



#51
3-Nitroaniline
Concen: 0.002 ng/ul
RT: 14.716 min Scan# 1926
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:138 Resp: 17
Ion Ratio Lower Upper
138 100
108 47.8 8.5 12.7#
92 447.8 95.8 143.6#





#52

Acenaphthene

Concen: 0.001 ng/ul

RT: 14.875 min Scan# 1953

Delta R.T. -0.000 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04

Instrument :

BNA_P

ClientSampleId :

SVOC-GPC2-BLANK





Tgt Ion:153 Resp: 46

Ion Ratio Lower Upper

153 100

152 12.9 38.8 58.2#

154 75.8 69.6 104.4



#53

2,4-Dinitrophenol

Concen: 0.003 ng/ul

RT: 14.904 min Scan# 1958

Delta R.T. -0.006 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04




Tgt Ion:184 Resp: 20


Ion Ratio Lower Upper

184 100

63 71.0 56.3 84.5

154 71.0 48.9 73.3









#55
4-Nitrophenol
Concen: 0.003 ng/ul
RT: 15.010 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:109 Resp: 19
Ion Ratio Lower Upper
109 100
139 23.3 115.1 172.7#
65 380.0 110.6 166.0#



#56
Dibenzofuran
Concen: 0.001 ng/ul
RT: 15.269 min Scan# 2020
Delta R.T. 0.059 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:168 Resp: 37
Ion Ratio Lower Upper
168 100
139 32.1 28.6 42.8









#57
2,4-Dinitrotoluene
Concen: 0.004 ng/ul
RT: 15.169 min Scan# 2002
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:165 Resp: 62
Ion Ratio Lower Upper
165 100
63 101.2 35.7 53.5#
182 18.1 2.2 3.4#



#58
2,3,4,6-Tetrachlorophenol
Concen: 0.000 ng/ul
RT: 15.428 min Scan# 2047
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:232 Resp: 5
Ion Ratio Lower Upper
232 100
131 900.0 38.2 57.4#
130 325.0 2.1 3.1#
166 262.5 26.6 39.8#









#59
Diethylphthalate
Concen: 0.002 ng/ul
RT: 15.628 min Scan# 2080
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:149 Resp: 126
Ion Ratio Lower Upper
149 100
177 157.8 17.5 26.3#
150 15.7 9.6 14.4#



#61
Fluorene
Concen: 0.000 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. -0.012 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:166 Resp: 23
Ion Ratio Lower Upper
166 100
165 624.0 79.3 118.9#
167 80.0 10.7 16.1#









#62
4-Chlorophenyl-phenylether
Concen: 0.001 ng/ul
RT: 15.857 min Scan# 2118
Delta R.T. 0.012 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion: 204 Resp: 21
Ion Ratio Lower Upper
204 100
206 40.0 26.2 39.2#
141 24.0 46.8 70.2#



#63
4-Nitroaniline
Concen: 0.001 ng/ul
RT: 15.875 min Scan# 2123
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion: 138 Resp: 9
Ion Ratio Lower Upper
138 100
92 121.1 43.7 65.5#
108 63.2 57.4 86.2









#66
4,6-Dinitro-2-methylphenol
Concen: 0.000 ng/ul
RT: 15.934 min Scan# 2132
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:198 Resp: 4
Ion Ratio Lower Upper
198 100
182 120.0 3.9 5.9#
77 1960.0 21.4 32.0#



#67
N-Nitrosodiphenylamine
Concen: 0.000 ng/ul
RT: 16.057 min Scan# 2154
Delta R.T. -0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:169 Resp: 24
Ion Ratio Lower Upper
169 100
168 66.7 53.5 80.3
167 40.0 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 0.001 ng/ul
RT: 16.845 min Scan# 2271
Delta R.T. 0.100 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:248	Resp:	14
Ion Ratio	Lower	Upper
248	100	
250	90.0	79.5 119.3
141	65.0	60.2 90.4



#69
Hexachlorobenzene
Concen: 0.004 ng/ul
RT: 16.845 min Scan# 2288
Delta R.T. -0.018 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:284	Resp:	75
Ion Ratio	Lower	Upper
284	100	
142	34.4	30.2 45.2
249	0.0	24.5 36.7#









#70
Atrazine
Concen: 0.002 ng/ul
RT: 16.963 min Scan# 21
Delta R.T. -0.047 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:200 Resp: 26
Ion Ratio Lower Upper
200 100
173 30.0 21.8 32.6
215 55.0 39.0 58.4



#71
Pentachlorophenol
Concen: 0.003 ng/ul
RT: 17.263 min Scan# 2359
Delta R.T. 0.053 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:266 Resp: 30
Ion Ratio Lower Upper
266 100
264 57.6 48.9 73.3
268 63.6 49.6 74.4









#72
Phenanthrene
Concen: 0.003 ng/ul
RT: 17.610 min Scan# 2418
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:178	Resp:	267
Ion Ratio	Lower	Upper
178	100	
179	18.3	12.2 18.4
176	5.6	15.4 23.0#



#74
Anthracene
Concen: 0.001 ng/ul
RT: 17.698 min Scan# 2433
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:178	Resp:	129
Ion Ratio	Lower	Upper
178	100	
179	94.4	12.6 18.8#
176	53.2	15.0 22.4#





#75

1,2,3,4-Tetrachlorobenzene

Concen: 0.001 ng/uL

RT: 13.604 min Scan# 1

Delta R.T. -0.012 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04

Instrument :

BNA_P

ClientSampleId :

SVOC-GPC2-BLANK





Tgt Ion:216 Resp: 29

Ion Ratio Lower Upper

216 100

214 57.7 60.5 90.7#

218 84.6 38.6 58.0#



#76

Pentachlorobenzene

Concen: 0.001 ng/uL

RT: 15.151 min Scan# 2000

Delta R.T. 0.023 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04




Tgt Ion:250 Resp: 19


Ion Ratio Lower Upper

250 100

248 66.7 50.5 75.7

252 52.4 50.5 75.7









#77
 Carbazole
 Concen: 0.001 ng/ul
 RT: 17.963 min Scan# 2478
 Delta R.T. -0.000 min
 Lab File: BP014935.D
 Acq: 03 May 2023 18:04




Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK


Tgt Ion:167 Resp: 62
 Ion Ratio Lower Upper
 167 100
 166 0.0 17.1 25.7#
 139 26.0 10.1 15.1#



#78
 Di-n-butylphthalate
 Concen: 0.004 ng/ul
 RT: 18.498 min Scan# 2569
 Delta R.T. -0.000 min
 Lab File: BP014935.D
 Acq: 03 May 2023 18:04

Tgt Ion:149 Resp: 306
 Ion Ratio Lower Upper
 149 100
 150 3.3 7.4 11.2#
 104 53.0 4.1 6.1#









#80
Fluoranthene
Concen: 0.003 ng/ul
RT: 19.569 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:202 Resp: 322
Ion Ratio Lower Upper
202 100
101 3.7 10.1 15.1#
100 15.1 7.7 11.5#



#82
Pyrene
Concen: 0.005 ng/ul
RT: 19.922 min Scan# 2811
Delta R.T. 0.012 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:202 Resp: 578
Ion Ratio Lower Upper
202 100
101 19.9 11.9 17.9#
100 19.2 9.7 14.5#





#83

Butylbenzylphthalate

Concen: 0.004 ng/ul

RT: 20.780 min Scan# 2955

Delta R.T. 0.012 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04

Instrument :

BNA_P

ClientSampleId :

SVOC-GPC2-BLANK





Tgt Ion:149 Resp: 140

Ion Ratio Lower Upper

149 100

91 40.0 60.1 90.1#

206 33.8 16.0 24.0#



#84

3,3'-Dichlorobenzidine

Concen: 0.002 ng/ul

RT: 21.545 min Scan# 3087

Delta R.T. -0.012 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04




Tgt Ion:252 Resp: 37


Ion Ratio Lower Upper

252 100

254 5.2 53.3 79.9#

126 24.5 9.0 13.6#





#85

Benzo(a)anthracene

Concen: 0.041 ng/ul

RT: 21.651 min Scan# 3105

Delta R.T. 0.018 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04

Instrument :

BNA_P

ClientSampleId :

SVOC-GPC2-BLANK





Tgt Ion:228 Resp: 3553

Ion Ratio Lower Upper

228 100

229 38.4 15.9 23.9#

226 10.2 22.1 33.1#



#86

Bis(2-ethylhexyl)phthalate

Concen: 0.016 ng/ul

RT: 21.539 min Scan# 3086

Delta R.T. 0.006 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04




Tgt Ion:149 Resp: 691


Ion Ratio Lower Upper

149 100

167 15.3 22.0 33.0#

279 2.9 2.9 4.3









#87
Chrysene
Concen: 0.007 ng/ul
RT: 21.680 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:228 Resp: 626
Ion Ratio Lower Upper
228 100
226 11.3 24.3 36.5#
229 42.3 15.8 23.8#



#89
Di-n-octyl phthalate
Concen: 0.004 ng/ul
RT: 22.527 min Scan# 3254
Delta R.T. 0.006 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:149 Resp: 283









#90
Benzo(b)fluoranthene
Concen: 0.005 ng/ul
RT: 23.445 min Scan# 3410
Delta R.T. -0.000 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04




Instrument :
BNA_P
ClientSampleId :
SVOC-GPC2-BLANK


Tgt Ion:252 Resp: 453
Ion Ratio Lower Upper
252 100
253 79.3 17.4 26.2#
125 124.4 8.4 12.6#



#91
Benzo(k)fluoranthene
Concen: 0.004 ng/ul
RT: 23.516 min Scan# 3422
Delta R.T. 0.018 min
Lab File: BP014935.D
Acq: 03 May 2023 18:04

Tgt Ion:252 Resp: 359
Ion Ratio Lower Upper
252 100
253 45.5 17.8 26.6#
125 76.4 8.9 13.3#





#93

Benzo(a)pyrene

Concen: 0.002 ng/ul

RT: 24.133 min Scan# 3

Delta R.T. -0.000 min

Lab File: BP014935.D

Acq: 03 May 2023 18:04

Instrument :





BNA_P

ClientSampleId :

SVOC-GPC2-BLANK

Tgt Ion:252 Resp: 156

Ion	Ratio	Lower	Upper
252	100		
253	92.8	17.5	26.3#
125	158.0	9.4	14.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 0.001 ng/ul

RT: 27.068 min Scan# 4026




Delta R.T. 0.059 min


Lab File: BP014935.D

Acq: 03 May 2023 18:04

Tgt Ion:276 Resp: 105

Ion	Ratio	Lower	Upper
276	100		
138	25.2	19.4	29.0
277	20.1	19.4	29.2









#95
 Dibenzo(a,h)anthracene
 Concen: 0.009 ng/ul
 RT: 27.010 min Scan# 4157
 Delta R.T. -0.018 min
 Lab File: BP014935.D
 Acq: 03 May 2023 18:04

Instrument :
 BNA_P
 ClientSampleId :
 SVOC-GPC2-BLANK

Tgt Ion:278 Resp: 624
 Ion Ratio Lower Upper
 278 100
 139 71.3 12.9 19.3#
 279 49.6 19.0 28.6#



#96
 Benzo(g,h,i)perylene
 Concen: 0.001 ng/ul
 RT: 27.839 min Scan# 4157
 Delta R.T. -0.018 min
 Lab File: BP014935.D
 Acq: 03 May 2023 18:04

Tgt Ion:276 Resp: 60
 Ion Ratio Lower Upper
 276 100
 138 12.0 18.0 27.0#
 277 9.0 19.8 29.8#

