

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050523\
 Data File : BP014977.D
 Acq On : 05 May 2023 11:28
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698

Quant Time: May 05 12:04:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 04 11:40:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.163	152	289003	20.000	ng/u1	0.00
20) Naphthalene-d8	10.998	136	1283433	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.810	164	829570	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.569	188	1812387	20.000	ng/u1	0.00
79) Chrysene-d12	21.645	240	1556726	20.000	ng/u1	0.00
88) Perylene-d12	24.251	264	1372349	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.458	96	52116	6.705	ng/uL	0.00
4) Pyridine-d5	3.899	84	358256	16.863	ng/u1	0.00
7) Phenol-d5	7.310	99	445820	17.793	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.481	67	265630	17.152	ng/u1	0.00
11) 2-Chlorophenol-d4	7.687	132	337583	18.323	ng/u1	0.00
15) 4-Methylphenol-d8	8.875	113	355679	18.421	ng/u1	0.00
21) Nitrobenzene-d5	9.340	128	175366	19.212	ng/u1	0.00
24) 2-Nitrophenol-d4	10.075	143	195384	20.845	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.616	165	350820	19.062	ng/u1	0.00
31) 4-Chloroaniline-d4	11.134	131	521388	19.586	ng/u1	0.00
46) Dimethylphthalate-d6	14.210	166	1074668	18.304	ng/u1	0.00
49) Acenaphthylene-d8	14.504	160	1234527	17.532	ng/u1	0.00
54) 4-Nitrophenol-d4	14.998	143	205173	20.087	ng/u1	0.01
60) Fluorene-d10	15.804	176	919866	18.211	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.922	200	188329	18.459	ng/u1	0.01
73) Anthracene-d10	17.663	188	1436920	17.645	ng/u1	0.00
81) Pyrene-d10	19.880	212	1614068	15.373	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.074	264	1188729	17.778	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.499	88	56974	6.766	ng/uL	96
5) Pyridine	3.917	79	379232	17.147	ng/u1	97
6) Benzaldehyde	7.293	77	151225	22.075	ng/u1	98
8) Phenol	7.340	94	464357	17.652	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.575	93	378914	17.330	ng/u1	98
12) 2-Chlorophenol	7.722	128	356944	17.884	ng/u1	99
13) 2-Methylphenol	8.610	108	356753	18.188	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.699	45	471680	17.007	ng/u1	99
16) Acetophenone	8.999	105	590254	18.696	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.981	70	303127	18.839	ng/u1	99
18) 4-Methylphenol	8.940	108	392484	18.100	ng/u1	100
19) Hexachloroethane	9.263	117	150735	17.407	ng/u1	99
22) Nitrobenzene	9.387	77	431404	17.987	ng/u1	97
23) Isophorone	9.916	82	883338	18.636	ng/u1	100
25) 2-Nitrophenol	10.104	139	211020	19.576	ng/u1	100
26) 2,4-Dimethylphenol	10.157	107	424318	18.138	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.399	93	534531	17.640	ng/u1	100
29) 2,4-Dichlorophenol	10.640	162	354024	18.598	ng/u1	97
30) Naphthalene	11.051	128	1202373	17.043	ng/u1	100
32) 4-Chloroaniline	11.157	127	541509	19.494	ng/u1	100
33) Hexachlorobutadiene	11.334	225	217392	17.699	ng/u1	99
34) Caprolactam	11.928	113	131793	22.473	ng/u1	95
35) 4-Chloro-3-methylphenol	12.269	107	396762	20.275	ng/u1	99
36) 2-Methylnaphthalene	12.651	142	802264	18.389	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050523\
 Data File : BP014977.D
 Acq On : 05 May 2023 11:28
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698

Quant Time: May 05 12:04:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 04 11:40:49 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.863	142	817927	18.188	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	13.010	216	428964	16.577	ng/ul	99
40) Hexachlorocyclopentadiene	12.987	237	367811	21.080	ng/ul	100
41) 2,4,6-Trichlorophenol	13.245	196	296591	18.601	ng/ul	100
42) 2,4,5-Trichlorophenol	13.322	196	318154	18.513	ng/ul	97
43) 1,1'-Biphenyl	13.645	154	1102156	16.618	ng/ul	100
44) 2-Chloronaphthalene	13.692	162	864637	16.736	ng/ul	99
45) 2-Nitroaniline	13.892	65	254681	21.170	ng/ul	98
47) Dimethylphthalate	14.257	163	1121848	18.172	ng/ul	100
48) 2,6-Dinitrotoluene	14.381	165	229176	21.007	ng/ul	100
50) Acenaphthylene	14.534	152	1379829	17.284	ng/ul	99
51) 3-Nitroaniline	14.710	138	186655	18.909	ng/ul	96
52) Acenaphthene	14.875	153	945855	17.122	ng/ul	98
53) 2,4-Dinitrophenol	14.922	184	131674	19.831	ng/ul	94
55) 4-Nitrophenol	15.010	109	155086	19.947	ng/ul	99
56) Dibenzofuran	15.204	168	1304492	17.500	ng/ul	98
57) 2,4-Dinitrotoluene	15.169	165	332530	21.762	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.434	232	283773	20.290	ng/ul	92
59) Diethylphthalate	15.616	149	1124488	19.277	ng/ul	99
61) Fluorene	15.857	166	1070475	18.205	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.845	204	519712	18.061	ng/ul	98
63) 4-Nitroaniline	15.875	138	169610	20.586	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.934	198	197359	18.237	ng/ul	99
67) N-Nitrosodiphenylamine	16.057	169	925173	16.819	ng/ul	99
68) 4-Bromophenyl-phenylether	16.745	248	325247	17.251	ng/ul	98
69) Hexachlorobenzene	16.869	284	384991	17.335	ng/ul	95
70) Atrazine	17.010	200	348453	18.825	ng/ul	100
71) Pentachlorophenol	17.210	266	238386	18.388	ng/ul	98
72) Phenanthrene	17.610	178	1706114	17.040	ng/ul	100
74) Anthracene	17.698	178	1737204	17.427	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.616	216	442839	14.760	ng/uL	98
76) Pentachlorobenzene	15.128	250	449032	15.732	ng/uL	99
77) Carbazole	17.963	167	1570339	18.729	ng/ul	100
78) Di-n-butylphthalate	18.492	149	1975961	21.075	ng/ul	99
80) Fluoranthene	19.557	202	1993989	15.477	ng/ul	99
82) Pyrene	19.910	202	2061897	15.106	ng/ul	98
83) Butylbenzylphthalate	20.763	149	895725	21.857	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.551	252	588233	22.451	ng/ul	98
85) Benzo(a)anthracene	21.633	228	1870015	17.997	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.527	149	1294883	25.112	ng/ul#	99
87) Chrysene	21.686	228	1752238	16.748	ng/ul	100
89) Di-n-octyl phthalate	22.516	149	2116588	26.415	ng/ul	100
90) Benzo(b)fluoranthene	23.445	252	1623818	18.308	ng/ul	99
91) Benzo(k)fluoranthene	23.498	252	1625026	18.003	ng/ul	98
93) Benzo(a)pyrene	24.127	252	1332822	17.135	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.004	276	1423828	15.006	ng/ul	98
95) Dibenzo(a,h)anthracene	27.027	278	1180154	15.176	ng/ul	98
96) Benzo(g,h,i)perylene	27.856	276	1178127	14.589	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050523\
Data File : BP014977.D
Acq On : 05 May 2023 11:28
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTD020698

Quant Time: May 05 12:04:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 04 11:40:49 2023
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 6.766 ng/uL
 RT: 3.499 min Scan# 19
 Delta R.T. -0.000 min
 Lab File: BP014977.D
 Acq: 05 May 2023 11:28




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698


Tgt Ion: 88 Resp: 56974
 Ion Ratio Lower Upper
 88 100
 43 24.5 20.6 31.0
 58 64.2 54.4 81.6



#5
 Pyridine
 Concen: 17.147 ng/uL
 RT: 3.917 min Scan# 90
 Delta R.T. -0.000 min
 Lab File: BP014977.D
 Acq: 05 May 2023 11:28

Tgt Ion: 79 Resp: 379232
 Ion Ratio Lower Upper
 79 100
 52 70.9 54.7 82.1
 51 32.9 25.5 38.3









#6
Benzaldehyde
Concen: 22.075 ng/ul
RT: 7.293 min Scan# 60
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion: 77 Resp: 151225
Ion Ratio Lower Upper
77 100
105 98.5 80.7 121.1
106 94.7 77.6 116.4



#8
Phenol
Concen: 17.652 ng/ul
RT: 7.340 min Scan# 672
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion: 94 Resp: 464357
Ion Ratio Lower Upper
94 100
65 26.7 21.4 32.0
66 33.1 25.8 38.8









#10
Bis(2-Chloroethyl)ether
Concen: 17.330 ng/ul
RT: 7.575 min Scan# 712
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion: 93 Resp: 378914
Ion Ratio Lower Upper
93 100
63 70.9 58.1 87.1
95 31.9 26.0 39.0



#12
2-Chlorophenol
Concen: 17.884 ng/ul
RT: 7.722 min Scan# 737
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:128 Resp: 356944
Ion Ratio Lower Upper
128 100
64 44.7 36.6 55.0
130 32.1 25.8 38.8





#13

2-Methylphenol

Concen: 18.188 ng/ul

RT: 8.610 min Scan# 8

Delta R.T. 0.006 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :





SSTD020698

Tgt Ion:108 Resp: 356753

Ion Ratio Lower Upper

108 100

107 86.9 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 17.007 ng/ul

RT: 8.699 min Scan# 903

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28




Tgt Ion: 45 Resp: 471680


Ion Ratio Lower Upper

45 100

77 15.7 12.3 18.5

79 12.3 9.5 14.3









#16
Acetophenone
Concen: 18.696 ng/ul
RT: 8.999 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:105 Resp: 590254
Ion Ratio Lower Upper
105 100
77 75.8 61.2 91.8
51 24.0 19.4 29.2



#17
N-Nitroso-di-n-propylamine
Concen: 18.839 ng/ul
RT: 8.981 min Scan# 951
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion: 70 Resp: 303127
Ion Ratio Lower Upper
70 100
42 51.8 41.9 62.9
101 9.7 7.7 11.5
130 21.2 16.2 24.2









#18
4-Methylphenol
Concen: 18.100 ng/ul
RT: 8.940 min Scan# 944
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:108 Resp: 392484
Ion Ratio Lower Upper
108 100
107 108.9 87.2 130.8



#19
Hexachloroethane
Concen: 17.407 ng/ul
RT: 9.263 min Scan# 999
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:117 Resp: 150735
Ion Ratio Lower Upper
117 100
201 97.4 78.9 118.3
199 60.9 48.3 72.5









#22
Nitrobenzene
Concen: 17.987 ng/ul
RT: 9.387 min Scan# 110
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion: 77 Resp: 431404
Ion Ratio Lower Upper
77 100
123 46.2 34.8 52.2
65 13.0 10.2 15.2



#23
Isophorone
Concen: 18.636 ng/ul
RT: 9.916 min Scan# 1110
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion: 82 Resp: 883338
Ion Ratio Lower Upper
82 100
95 7.0 5.3 7.9
138 17.3 13.9 20.9









#25
2-Nitrophenol
Concen: 19.576 ng/ul
RT: 10.104 min Scan# 1142
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:139 Resp: 211020
Ion Ratio Lower Upper
139 100
65 50.4 40.5 60.7
109 28.4 22.4 33.6



#26
2,4-Dimethylphenol
Concen: 18.138 ng/ul
RT: 10.157 min Scan# 1151
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:107 Resp: 424318
Ion Ratio Lower Upper
107 100
121 51.0 40.2 60.2
122 86.7 69.6 104.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.640 ng/ul
 RT: 10.399 min Scan# 1192
 Delta R.T. 0.006 min
 Lab File: BP014977.D
 Acq: 05 May 2023 11:28




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698


Tgt Ion: 93 Resp: 534531
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.1 39.1
 123 11.0 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 18.598 ng/ul
 RT: 10.640 min Scan# 1233
 Delta R.T. -0.000 min
 Lab File: BP014977.D
 Acq: 05 May 2023 11:28

Tgt Ion:162 Resp: 354024
 Ion Ratio Lower Upper
 162 100
 164 66.6 51.5 77.3
 98 37.6 28.8 43.2





#30

Naphthalene

Concen: 17.043 ng/ul

RT: 11.051 min Scan# 1303

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:128 Resp: 1202373

Ion Ratio Lower Upper

128 100

129 10.9 8.7 13.1

127 13.3 10.4 15.6



#32

4-Chloroaniline

Concen: 19.494 ng/ul

RT: 11.157 min Scan# 1321

Delta R.T. -0.000 min

Lab File: BP014977.D




Acq: 05 May 2023 11:28


Tgt Ion:127 Resp: 541509

Ion Ratio Lower Upper

127 100

129 31.9 25.5 38.3









#33
Hexachlorobutadiene
Concen: 17.699 ng/ul
RT: 11.334 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:225 Resp: 217392
Ion Ratio Lower Upper
225 100
223 62.6 50.6 76.0
227 62.7 51.4 77.0



#34
Caprolactam
Concen: 22.473 ng/ul
RT: 11.928 min Scan# 1452
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:113 Resp: 131793
Ion Ratio Lower Upper
113 100
55 150.7 128.5 192.7
56 115.5 94.3 141.5









#35
4-Chloro-3-methylphenol
Concen: 20.275 ng/ul
RT: 12.269 min Scan# 1510
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:107 Resp: 396762
Ion Ratio Lower Upper
107 100
144 26.9 20.6 30.8
142 79.5 63.2 94.8



#36
2-Methylnaphthalene
Concen: 18.389 ng/ul
RT: 12.651 min Scan# 1575
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:142 Resp: 802264
Ion Ratio Lower Upper
142 100
141 89.2 70.5 105.7





#37

1-Methylnaphthalene

Concen: 18.188 ng/ul

RT: 12.863 min Scan# 1611

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:142 Resp: 817927

Ion Ratio Lower Upper

142 100

141 90.0 74.4 111.6

116 3.9 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 16.577 ng/ul

RT: 13.010 min Scan# 1636

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Tgt Ion:216 Resp: 428964




Ion Ratio Lower Upper


216 100

214 77.8 63.0 94.4

179 20.7 16.9 25.3

108 17.7 14.9 22.3









#40
Hexachlorocyclopentadiene
Concen: 21.080 ng/ul
RT: 12.987 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:237 Resp: 367811
Ion Ratio Lower Upper
237 100
235 62.6 49.9 74.9
272 12.6 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 18.601 ng/ul
RT: 13.245 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:196 Resp: 296591
Ion Ratio Lower Upper
196 100
198 96.1 76.8 115.2
200 31.5 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 18.513 ng/ul
RT: 13.322 min Scan# 1
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:196 Resp: 318154
Ion Ratio Lower Upper
196 100
198 96.6 74.2 111.4
200 31.4 24.6 37.0



#43
1,1'-Biphenyl
Concen: 16.618 ng/ul
RT: 13.645 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:154 Resp: 1102156
Ion Ratio Lower Upper
154 100
153 39.9 31.9 47.9
76 13.9 11.3 16.9









#44
2-Chloronaphthalene
Concen: 16.736 ng/ul
RT: 13.692 min Scan# 1752
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:162 Resp: 864637
Ion Ratio Lower Upper
162 100
164 32.4 26.3 39.5
127 37.6 30.4 45.6



#45
2-Nitroaniline
Concen: 21.170 ng/ul
RT: 13.892 min Scan# 1786
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion: 65 Resp: 254681
Ion Ratio Lower Upper
65 100
92 72.4 58.3 87.5
138 112.6 87.9 131.9









#47
Dimethylphthalate
Concen: 18.172 ng/ul
RT: 14.257 min Scan# 1848
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:163 Resp: 1121848
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.0 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 21.007 ng/ul
RT: 14.381 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:165 Resp: 229176
Ion Ratio Lower Upper
165 100
89 59.8 48.1 72.1
121 21.6 17.3 25.9









#50
Acenaphthylene
Concen: 17.284 ng/ul
RT: 14.534 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:152 Resp: 1379829
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.1 10.4 15.6



#51
3-Nitroaniline
Concen: 18.909 ng/ul
RT: 14.710 min Scan# 1925
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:138 Resp: 186655
Ion Ratio Lower Upper
138 100
108 10.3 8.5 12.7
92 114.5 95.8 143.6





#52

Acenaphthene

Concen: 17.122 ng/ul

RT: 14.875 min Scan# 1953

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:153 Resp: 945855

Ion Ratio Lower Upper

153 100

152 47.3 38.8 58.2

154 88.4 69.6 104.4



#53

2,4-Dinitrophenol

Concen: 19.831 ng/ul

RT: 14.922 min Scan# 1961

Delta R.T. 0.012 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28




Tgt Ion:184 Resp: 131674


Ion Ratio Lower Upper

184 100

63 62.6 56.3 84.5

154 59.1 48.9 73.3









#55
4-Nitrophenol
Concen: 19.947 ng/ul
RT: 15.010 min Scan# 1975
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:109 Resp: 155086
Ion Ratio Lower Upper
109 100
139 143.8 115.1 172.7
65 134.9 110.6 166.0



#56
Dibenzofuran
Concen: 17.500 ng/ul
RT: 15.204 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:168 Resp: 1304492
Ion Ratio Lower Upper
168 100
139 36.6 28.6 42.8









#57
2,4-Dinitrotoluene
Concen: 21.762 ng/ul
RT: 15.169 min Scan# 2002
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:165 Resp: 332530
Ion Ratio Lower Upper
165 100
63 39.8 35.7 53.5
182 2.6 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.290 ng/ul
RT: 15.434 min Scan# 2048
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:232 Resp: 283773
Ion Ratio Lower Upper
232 100
131 41.1 38.2 57.4
130 2.1 2.1 3.1
166 29.4 26.6 39.8









#59
Diethylphthalate
Concen: 19.277 ng/ul
RT: 15.616 min Scan# 2079
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:149 Resp: 1124488
Ion Ratio Lower Upper
149 100
177 22.3 17.5 26.3
150 12.3 9.6 14.4



#61
Fluorene
Concen: 18.205 ng/ul
RT: 15.857 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:166 Resp: 1070475
Ion Ratio Lower Upper
166 100
165 97.3 79.3 118.9
167 13.4 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 18.061 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:204 Resp: 519712
Ion Ratio Lower Upper
204 100
206 32.6 26.2 39.2
141 56.4 46.8 70.2



#63
4-Nitroaniline
Concen: 20.586 ng/ul
RT: 15.875 min Scan# 2123
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:138 Resp: 169610
Ion Ratio Lower Upper
138 100
92 52.4 43.7 65.5
108 76.7 57.4 86.2









#66
4,6-Dinitro-2-methylphenol
Concen: 18.237 ng/ul
RT: 15.934 min Scan# 2133
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:198 Resp: 197359
Ion Ratio Lower Upper
198 100
182 4.5 3.9 5.9
77 26.2 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 16.819 ng/ul
RT: 16.057 min Scan# 2154
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:169 Resp: 925173
Ion Ratio Lower Upper
169 100
168 67.5 53.5 80.3
167 36.1 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 17.251 ng/ul
RT: 16.745 min Scan# 211
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:248 Resp: 325247
Ion Ratio Lower Upper
248 100
250 99.7 79.5 119.3
141 70.9 60.2 90.4



#69
Hexachlorobenzene
Concen: 17.335 ng/ul
RT: 16.869 min Scan# 2292
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:284 Resp: 384991
Ion Ratio Lower Upper
284 100
142 32.7 30.2 45.2
249 30.0 24.5 36.7





#70

Atrazine

Concen: 18.825 ng/ul

RT: 17.010 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:200 Resp: 348453

Ion Ratio Lower Upper

200 100

173 27.2 21.8 32.6

215 48.8 39.0 58.4



#71

Pentachlorophenol

Concen: 18.388 ng/ul

RT: 17.210 min Scan# 2350

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28




Tgt Ion:266 Resp: 238386


Ion Ratio Lower Upper

266 100

264 63.3 48.9 73.3

268 63.0 49.6 74.4





#72

Phenanthrene

Concen: 17.040 ng/ul

RT: 17.610 min Scan# 2418

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:178 Resp: 1706114

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.4 15.4 23.0



#74

Anthracene

Concen: 17.427 ng/ul

RT: 17.698 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28




Tgt Ion:178 Resp: 1737204


Ion Ratio Lower Upper

178 100

179 15.6 12.6 18.8

176 18.7 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 14.760 ng/uL
RT: 13.616 min Scan# 1739
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:216 Resp: 442839
Ion Ratio Lower Upper
216 100
214 78.9 60.5 90.7
218 48.0 38.6 58.0



#76
Pentachlorobenzene
Concen: 15.732 ng/uL
RT: 15.128 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:250 Resp: 449032
Ion Ratio Lower Upper
250 100
248 64.3 50.5 75.7
252 63.6 50.5 75.7





#77

Carbazole

Concen: 18.729 ng/ul

RT: 17.963 min Scan# 2478

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:167 Resp: 1570339

Ion Ratio Lower Upper

167 100

166 21.7 17.1 25.7

139 12.6 10.1 15.1



#78

Di-n-butylphthalate

Concen: 21.075 ng/ul

RT: 18.492 min Scan# 2568

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28




Tgt Ion:149 Resp: 1975961


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 5.0 4.1 6.1









#80
Fluoranthene
Concen: 15.477 ng/ul
RT: 19.557 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:202 Resp: 1993989
Ion Ratio Lower Upper
202 100
101 12.0 10.1 15.1
100 9.2 7.7 11.5



#82
Pyrene
Concen: 15.106 ng/ul
RT: 19.910 min Scan# 2809
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:202 Resp: 2061897
Ion Ratio Lower Upper
202 100
101 14.0 11.9 17.9
100 11.5 9.7 14.5





#83

Butylbenzylphthalate

Concen: 21.857 ng/ul

RT: 20.763 min Scan# 2954

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:149 Resp: 895725

Ion Ratio Lower Upper

149 100

91 74.1 60.1 90.1

206 20.4 16.0 24.0



#84

3,3'-Dichlorobenzidine

Concen: 22.451 ng/ul

RT: 21.551 min Scan# 3088

Delta R.T. -0.000 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28




Tgt Ion:252 Resp: 588233


Ion Ratio Lower Upper

252 100

254 64.9 53.3 79.9

126 11.1 9.0 13.6









#85
Benzo(a)anthracene
Concen: 17.997 ng/ul
RT: 21.633 min Scan# 3101
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:228 Resp: 1870015
Ion Ratio Lower Upper
228 100
229 19.4 15.9 23.9
226 27.4 22.1 33.1



#86
Bis(2-ethylhexyl)phthalate
Concen: 25.112 ng/ul
RT: 21.527 min Scan# 3084
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:149 Resp: 1294883
Ion Ratio Lower Upper
149 100
167 27.0 22.0 33.0
279 4.3 2.9 4.3#









#87
Chrysene
Concen: 16.748 ng/ul
RT: 21.686 min Scan# 3111
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:228 Resp: 1752238
Ion Ratio Lower Upper
228 100
226 30.5 24.3 36.5
229 19.7 15.8 23.8



#89
Di-n-octyl phthalate
Concen: 26.415 ng/ul
RT: 22.516 min Scan# 3252
Delta R.T. -0.000 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:149 Resp: 2116588









#90
Benzo(b)fluoranthene
Concen: 18.308 ng/ul
RT: 23.445 min Scan# 3409
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28




Instrument :
BNA_P
ClientSampleId :
SSTD020698


Tgt Ion:252 Resp: 1623818
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 10.1 8.4 12.6



#91
Benzo(k)fluoranthene
Concen: 18.003 ng/ul
RT: 23.498 min Scan# 3419
Delta R.T. 0.006 min
Lab File: BP014977.D
Acq: 05 May 2023 11:28

Tgt Ion:252 Resp: 1625026
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.6
125 9.8 8.9 13.3









#93
 Benzo(a)pyrene
 Concen: 17.135 ng/ul
 RT: 24.127 min Scan# 3526
 Delta R.T. -0.000 min
 Lab File: BP014977.D
 Acq: 05 May 2023 11:28

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020698

Tgt Ion:252 Resp: 1332822




Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.5	26.3
125	11.2	9.4	14.2




#94
 Indeno(1,2,3-cd)pyrene
 Concen: 15.006 ng/ul
 RT: 27.004 min Scan# 4015
 Delta R.T. 0.006 min
 Lab File: BP014977.D
 Acq: 05 May 2023 11:28

Tgt Ion:276 Resp: 1423828

Ion	Ratio	Lower	Upper
276	100		
138	23.6	19.4	29.0
277	25.5	19.4	29.2





#95

Dibenzo(a,h)anthracene

Concen: 15.176 ng/ul

RT: 27.027 min Scan# 4018

Delta R.T. 0.012 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Instrument :

BNA_P

ClientSampleId :

SSTD020698





Tgt Ion:278 Resp: 1180154

Ion Ratio Lower Upper

278 100

139 14.8 12.9 19.3

279 23.0 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 14.589 ng/ul

RT: 27.856 min Scan# 4160

Delta R.T. 0.012 min

Lab File: BP014977.D

Acq: 05 May 2023 11:28

Tgt Ion:276 Resp: 1178127

Ion Ratio Lower Upper

276 100

138 21.3 18.0 27.0

277 24.7 19.8 29.8

