

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050621\
 Data File : BP005452.D
 Acq On : 06 May 2021 16:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 06 16:48:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 05 19:00:25 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	80	0.00
2	1,4-Dioxane	40.000	46.323	-15.8	92	0.00
3	Pyridine	40.000	48.458	-21.1	97	0.00
4	n-Nitrosodimethylamine	40.000	41.677	-4.2	82	0.00
5 S	2-Fluorophenol	80.000	77.225	3.5	81	0.00
6	Aniline	40.000	36.923	7.7	76	0.00
7 S	Phenol-d6	80.000	74.973	6.3	75	0.00
8	2-Chlorophenol	40.000	39.053	2.4	79	0.00
9	Benzaldehyde	40.000	39.047	2.4	77	0.00
10 C	Phenol	40.000	38.112	4.7	79	0.00
11	bis(2-Chloroethyl)ether	40.000	37.466	6.3	79	0.00
12	1,3-Dichlorobenzene	40.000	36.161	9.6	79	0.00
13 C	1,4-Dichlorobenzene	40.000	37.981	5.0	81	0.00
14	1,2-Dichlorobenzene	40.000	37.113	7.2	77	0.00
15	Benzyl Alcohol	40.000	37.183	7.0	74	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.948	12.6	70	-0.01
17	2-Methylphenol	40.000	37.243	6.9	76	0.00
18	Hexachloroethane	40.000	36.906	7.7	78	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.647	3.4	77	0.00
20	3+4-Methylphenols	40.000	37.491	6.3	75	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	76	0.00
22	Acetophenone	40.000	38.844	2.9	75	0.00
23 S	Nitrobenzene-d5	80.000	79.960	0.1	77	0.00
24	Nitrobenzene	40.000	40.927	-2.3	80	0.00
25	Isophorone	40.000	39.094	2.3	76	0.00
26 C	2-Nitrophenol	40.000	37.235	6.9	75	0.00
27	2,4-Dimethylphenol	40.000	38.347	4.1	75	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.571	1.1	77	0.00
29 C	2,4-Dichlorophenol	40.000	38.911	2.7	75	0.00
30	1,2,4-Trichlorobenzene	40.000	39.057	2.4	76	0.00
31	Naphthalene	40.000	38.934	2.7	75	0.00
32	Benzoic acid	40.000	40.562	-1.4	76	0.00
33	4-Chloroaniline	40.000	39.296	1.8	76	0.00
34 C	Hexachlorobutadiene	40.000	41.264	-3.2	83	0.00
35	Caprolactam	40.000	39.449	1.4	73	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.666	0.8	76	0.00
37	2-Methylnaphthalene	40.000	40.468	-1.2	79	0.00
38	1-Methylnaphthalene	40.000	39.972	0.1	77	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	79	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.373	6.6	81	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.938	5.2	81	0.00
42 S	2,4,6-Tribromophenol	80.000	75.732	5.3	80	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.883	5.3	78	0.00
44	2,4,5-Trichlorophenol	40.000	37.780	5.5	79	0.00
45 S	2-Fluorobiphenyl	80.000	74.426	7.0	78	0.00
46	1,1'-Biphenyl	40.000	37.400	6.5	78	0.00
47	2-Chloronaphthalene	40.000	37.138	7.2	78	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050621\
 Data File : BP005452.D
 Acq On : 06 May 2021 16:16
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 06 16:48:21 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 05 19:00:25 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.370	4.1	76	0.00
49	Acenaphthylene	40.000	38.032	4.9	78	0.00
50	Dimethylphthalate	40.000	38.151	4.6	80	0.00
51	2,6-Dinitrotoluene	40.000	38.046	4.9	79	0.00
52 C	Acenaphthene	40.000	37.472	6.3	77	0.00
53	3-Nitroaniline	40.000	37.280	6.8	73	0.00
54 P	2,4-Dinitrophenol	40.000	41.309	-3.3	77	0.00
55	Dibenzofuran	40.000	37.944	5.1	79	0.00
56 P	4-Nitrophenol	40.000	40.989	-2.5	79	0.00
57	2,4-Dinitrotoluene	40.000	38.081	4.8	81	0.00
58	Fluorene	40.000	38.325	4.2	78	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.657	0.9	82	0.00
60	Diethylphthalate	40.000	38.296	4.3	79	0.00
61	4-Chlorophenyl-phenylether	40.000	39.729	0.7	84	0.00
62	4-Nitroaniline	40.000	40.975	-2.4	82	0.00
63	Azobenzene	40.000	40.188	-0.5	81	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	84	0.01
65	4,6-Dinitro-2-methylphenol	40.000	39.395	1.5	80	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.243	6.9	78	0.00
67	4-Bromophenyl-phenylether	40.000	37.202	7.0	83	0.00
68	Hexachlorobenzene	40.000	37.251	6.9	84	0.00
69	Atrazine	40.000	37.220	7.0	82	0.01
70 C	Pentachlorophenol	40.000	37.420	6.4	81	0.00
71	Phenanthrene	40.000	37.892	5.3	81	0.01
72	Anthracene	40.000	37.597	6.0	81	0.00
73	Carbazole	40.000	38.165	4.6	81	0.00
74	Di-n-butylphthalate	40.000	37.273	6.8	80	0.00
75 C	Fluoranthene	40.000	37.855	5.4	82	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	86	0.01
77	Benzydine	40.000	39.348	1.6	79	0.00
78	Pyrene	40.000	36.893	7.8	84	0.01
79 S	Terphenyl-d14	80.000	74.862	6.4	83	0.01
80	Butylbenzylphthalate	40.000	37.100	7.2	82	0.01
81	Benzo(a)anthracene	40.000	38.425	3.9	87	0.00
82	3,3'-Dichlorobenzidine	40.000	39.277	1.8	89	0.01
83	Chrysene	40.000	38.583	3.5	86	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.033	7.4	80	0.01
85 c	Di-n-octyl phthalate	40.000	37.888	5.3	83	0.01
86 I	Perylene-d12	20.000	20.000	0.0	92	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	37.783	5.5	90	0.02
88	Benzo(b)fluoranthene	40.000	38.604	3.5	88	0.02
89	Benzo(k)fluoranthene	40.000	37.444	6.4	92	0.02
90 C	Benzo(a)pyrene	40.000	38.416	4.0	91	0.02
91	Dibenzo(a,h)anthracene	40.000	37.628	5.9	89	0.02
92	Benzo(g,h,i)perylene	40.000	38.216	4.5	90	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050621\
Data File : BP005452.D
Acq On : 06 May 2021 16:16
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 06 16:48:21 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 05 19:00:25 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0