

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050823\
 Data File : BPCCC003.D
 Acq On : 08 May 2023 20:41
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020701

Quant Time: May 08 22:53:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 05 12:12:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.163	152	386493	20.000	ng/u1	0.00
20) Naphthalene-d8	10.999	136	1752244	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.810	164	1092103	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.563	188	2267374	20.000	ng/u1	0.00
79) Chrysene-d12	21.645	240	1738598	20.000	ng/u1	0.00
88) Perylene-d12	24.245	264	1494012	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.458	96	61636	5.930	ng/uL	0.00
4) Pyridine-d5	3.899	84	441418	15.536	ng/u1	0.00
7) Phenol-d5	7.305	99	573005	17.100	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.475	67	336274	16.236	ng/u1	0.00
11) 2-Chlorophenol-d4	7.687	132	429834	17.446	ng/u1	0.00
15) 4-Methylphenol-d8	8.875	113	455565	17.642	ng/u1	0.00
21) Nitrobenzene-d5	9.340	128	220891	17.725	ng/u1	0.00
24) 2-Nitrophenol-d4	10.069	143	245797	19.208	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.610	165	451616	17.974	ng/u1	0.00
31) 4-Chloroaniline-d4	11.134	131	646325	17.784	ng/u1	0.00
46) Dimethylphthalate-d6	14.210	166	1289172	16.679	ng/u1	0.00
49) Acenaphthylene-d8	14.504	160	1536549	16.576	ng/u1	0.00
54) 4-Nitrophenol-d4	14.992	143	241099	17.930	ng/u1	0.00
60) Fluorene-d10	15.798	176	1121475	16.865	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.916	200	208913	16.368	ng/u1	0.00
73) Anthracene-d10	17.663	188	1701981	16.706	ng/u1	0.00
81) Pyrene-d10	19.880	212	1831735	15.621	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.074	264	1211716	16.646	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.493	88	64996	5.771	ng/uL	95
5) Pyridine	3.917	79	452665	15.304	ng/u1	98
6) Benzaldehyde	7.287	77	186152	20.319	ng/u1	97
8) Phenol	7.334	94	593655	16.875	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.575	93	476152	16.284	ng/u1	98
12) 2-Chlorophenol	7.716	128	454299	17.021	ng/u1	100
13) 2-Methylphenol	8.605	108	456848	17.416	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.693	45	602377	16.241	ng/u1	99
16) Acetophenone	8.999	105	737174	17.460	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.981	70	370807	17.232	ng/u1	98
18) 4-Methylphenol	8.940	108	505389	17.428	ng/u1	100
19) Hexachloroethane	9.257	117	185517	16.020	ng/u1	98
22) Nitrobenzene	9.381	77	545897	16.671	ng/u1	97
23) Isophorone	9.910	82	1087234	16.800	ng/u1	100
25) 2-Nitrophenol	10.099	139	266792	18.128	ng/u1	98
26) 2,4-Dimethylphenol	10.151	107	539509	16.892	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.393	93	672127	16.246	ng/u1	100
29) 2,4-Dichlorophenol	10.640	162	454567	17.490	ng/u1	100
30) Naphthalene	11.051	128	1539337	15.982	ng/u1	100
32) 4-Chloroaniline	11.157	127	673896	17.769	ng/u1	100
33) Hexachlorobutadiene	11.328	225	267605	15.958	ng/u1	100
34) Caprolactam	11.928	113	102709	12.828	ng/u1	99
35) 4-Chloro-3-methylphenol	12.263	107	452750	16.946	ng/u1	99
36) 2-Methylnaphthalene	12.645	142	923714	15.508	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050823\
 Data File : BPCCC003.D
 Acq On : 08 May 2023 20:41
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020701


Quant Time: May 08 22:53:33 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 05 12:12:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.863	142	1030856	16.790	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	13.010	216	532205	15.623	ng/ul	99
40) Hexachlorocyclopentadiene	12.987	237	419973	18.283	ng/ul	99
41) 2,4,6-Trichlorophenol	13.245	196	363765	17.330	ng/ul	100
42) 2,4,5-Trichlorophenol	13.316	196	393424	17.390	ng/ul	96
43) 1,1'-Biphenyl	13.640	154	1383551	15.846	ng/ul	100
44) 2-Chloronaphthalene	13.687	162	1090360	16.032	ng/ul	99
45) 2-Nitroaniline	13.887	65	301003	19.005	ng/ul	100
47) Dimethylphthalate	14.257	163	1350803	16.620	ng/ul	100
48) 2,6-Dinitrotoluene	14.381	165	268155	18.671	ng/ul	97
50) Acenaphthylene	14.534	152	1716914	16.337	ng/ul	99
51) 3-Nitroaniline	14.710	138	228690	17.598	ng/ul	95
52) Acenaphthene	14.875	153	1162917	15.990	ng/ul	99
53) 2,4-Dinitrophenol	14.916	184	145426	16.637	ng/ul	97
55) 4-Nitrophenol	15.010	109	184641	18.040	ng/ul	98
56) Dibenzofuran	15.204	168	1616518	16.473	ng/ul	100
57) 2,4-Dinitrotoluene	15.169	165	384407	19.109	ng/ul	93
58) 2,3,4,6-Tetrachlorophenol	15.428	232	333167	18.095	ng/ul	94
59) Diethylphthalate	15.616	149	1330710	17.328	ng/ul	99
61) Fluorene	15.857	166	1294303	16.720	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.845	204	628422	16.589	ng/ul	97
63) 4-Nitroaniline	15.869	138	191614	17.666	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.934	198	222205	16.412	ng/ul	97
67) N-Nitrosodiphenylamine	16.057	169	1100720	15.995	ng/ul	100
68) 4-Bromophenyl-phenylether	16.739	248	384330	16.294	ng/ul	99
69) Hexachlorobenzene	16.863	284	442162	15.914	ng/ul	98
70) Atrazine	17.010	200	395673	17.086	ng/ul	99
71) Pentachlorophenol	17.210	266	268930	16.581	ng/ul	97
72) Phenanthrene	17.604	178	2033447	16.233	ng/ul	99
74) Anthracene	17.698	178	2069509	16.594	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.610	216	547429	14.585	ng/ul	97
76) Pentachlorobenzene	15.122	250	533441	14.939	ng/ul	99
77) Carbazole	17.963	167	1841400	17.555	ng/ul	99
78) Di-n-butylphthalate	18.492	149	2296800	19.582	ng/ul	99
80) Fluoranthene	19.557	202	2286184	15.889	ng/ul	97
82) Pyrene	19.910	202	2333064	15.304	ng/ul	96
83) Butylbenzylphthalate	20.763	149	997832	21.801	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.545	252	593431	20.280	ng/ul	97
85) Benzo(a)anthracene	21.627	228	1989795	17.147	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.521	149	1452245	25.217	ng/ul	99
87) Chrysene	21.680	228	1870807	16.011	ng/ul	100
89) Di-n-octyl phthalate	22.510	149	2383615	27.325	ng/ul	100
90) Benzo(b)fluoranthene	23.439	252	1612147	16.697	ng/ul	100
91) Benzo(k)fluoranthene	23.492	252	1621795	16.504	ng/ul	99
93) Benzo(a)pyrene	24.127	252	1379699	16.293	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.998	276	1695607	16.416	ng/ul	99
95) Dibenzo(a,h)anthracene	27.021	278	1400699	16.545	ng/ul	100
96) Benzo(g,h,i)perylene	27.851	276	1396047	15.880	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 08 22:53:33 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 05 12:12:36 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 5.771 ng/uL
RT: 3.493 min Scan# 18
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:	88	Resp:	64996
Ion	Ratio	Lower	Upper
88	100		
43	28.7	20.6	31.0
58	71.7	54.4	81.6



#5
Pyridine
Concen: 15.304 ng/uL
RT: 3.917 min Scan# 90
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:	79	Resp:	452665
Ion	Ratio	Lower	Upper
79	100		
52	70.0	54.7	82.1
51	32.9	25.5	38.3









#6
Benzaldehyde
Concen: 20.319 ng/ul
RT: 7.287 min Scan# 60
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion: 77 Resp: 186152
Ion Ratio Lower Upper
77 100
105 98.2 80.7 121.1
106 94.3 77.6 116.4



#8
Phenol
Concen: 16.875 ng/ul
RT: 7.334 min Scan# 671
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion: 94 Resp: 593655
Ion Ratio Lower Upper
94 100
65 26.1 21.4 32.0
66 34.1 25.8 38.8









#10
Bis(2-Chloroethyl)ether
Concen: 16.284 ng/ul
RT: 7.575 min Scan# 712
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion: 93 Resp: 476152
Ion Ratio Lower Upper
93 100
63 70.6 58.1 87.1
95 32.3 26.0 39.0



#12
2-Chlorophenol
Concen: 17.021 ng/ul
RT: 7.716 min Scan# 736
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion: 128 Resp: 454299
Ion Ratio Lower Upper
128 100
64 45.6 36.6 55.0
130 32.0 25.8 38.8









#13
2-Methylphenol
Concen: 17.416 ng/ul
RT: 8.605 min Scan# 888
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:108 Resp: 456848
Ion Ratio Lower Upper
108 100
107 90.5 71.7 107.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 16.241 ng/ul
RT: 8.693 min Scan# 902
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion: 45 Resp: 602377
Ion Ratio Lower Upper
45 100
77 15.2 12.3 18.5
79 11.6 9.5 14.3









#16
Acetophenone
Concen: 17.460 ng/ul
RT: 8.999 min Scan# 91
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:105 Resp: 737174
Ion Ratio Lower Upper
105 100
77 74.9 61.2 91.8
51 24.1 19.4 29.2



#17
N-Nitroso-di-n-propylamine
Concen: 17.232 ng/ul
RT: 8.981 min Scan# 951
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion: 70 Resp: 370807
Ion Ratio Lower Upper
70 100
42 52.9 41.9 62.9
101 10.2 7.7 11.5
130 22.1 16.2 24.2





#18

4-Methylphenol

Concen: 17.428 ng/ul

RT: 8.940 min Scan# 944

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :





SSTD020701

Tgt Ion:108 Resp: 505389

Ion Ratio Lower Upper

108 100

107 108.6 87.2 130.8



#19

Hexachloroethane

Concen: 16.020 ng/ul

RT: 9.257 min Scan# 998

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:117 Resp: 185517


Ion Ratio Lower Upper

117 100

201 100.1 78.9 118.3

199 62.1 48.3 72.5









#22
Nitrobenzene
Concen: 16.671 ng/ul
RT: 9.381 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion: 77 Resp: 545897
Ion Ratio Lower Upper
77 100
123 46.1 34.8 52.2
65 12.3 10.2 15.2



#23
Isophorone
Concen: 16.800 ng/ul
RT: 9.910 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion: 82 Resp: 1087234
Ion Ratio Lower Upper
82 100
95 6.6 5.3 7.9
138 17.4 13.9 20.9









#25
2-Nitrophenol
Concen: 18.128 ng/ul
RT: 10.099 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:139 Resp: 266792
Ion Ratio Lower Upper
139 100
65 48.6 40.5 60.7
109 27.7 22.4 33.6



#26
2,4-Dimethylphenol
Concen: 16.892 ng/ul
RT: 10.151 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:107 Resp: 539509
Ion Ratio Lower Upper
107 100
121 49.9 40.2 60.2
122 87.3 69.6 104.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 16.246 ng/ul
 RT: 10.393 min Scan# 1191
 Delta R.T. -0.006 min
 Lab File: BPCCC003.D
 Acq: 08 May 2023 20:41




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020701


Tgt Ion: 93 Resp: 672127
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.1 39.1
 123 11.2 9.0 13.4



#29
 2,4-Dichlorophenol
 Concen: 17.490 ng/ul
 RT: 10.640 min Scan# 1233
 Delta R.T. 0.000 min
 Lab File: BPCCC003.D
 Acq: 08 May 2023 20:41

Tgt Ion:162 Resp: 454567
 Ion Ratio Lower Upper
 162 100
 164 64.4 51.5 77.3
 98 35.8 28.8 43.2









#30
Naphthalene
Concen: 15.982 ng/ul
RT: 11.051 min Scan# 1303
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:128 Resp: 1539337
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.1 10.4 15.6



#32
4-Chloroaniline
Concen: 17.769 ng/ul
RT: 11.157 min Scan# 1321
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:127 Resp: 673896
Ion Ratio Lower Upper
127 100
129 31.7 25.5 38.3





#33

Hexachlorobutadiene

Concen: 15.958 ng/ul

RT: 11.328 min Scan# 1

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:225 Resp: 267605

Ion Ratio Lower Upper

225 100

223 63.5 50.6 76.0

227 64.5 51.4 77.0



#34

Caprolactam

Concen: 12.828 ng/ul

RT: 11.928 min Scan# 1452

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:113 Resp: 102709


Ion Ratio Lower Upper

113 100

55 159.0 128.5 192.7

56 117.6 94.3 141.5









#35
4-Chloro-3-methylphenol
Concen: 16.946 ng/ul
RT: 12.263 min Scan# 1510
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:107 Resp: 452750
Ion Ratio Lower Upper
107 100
144 25.8 20.6 30.8
142 80.1 63.2 94.8



#36
2-Methylnaphthalene
Concen: 15.508 ng/ul
RT: 12.645 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:142 Resp: 923714
Ion Ratio Lower Upper
142 100
141 88.8 70.5 105.7





#37

1-Methylnaphthalene

Concen: 16.790 ng/ul

RT: 12.863 min Scan# 1

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:142 Resp: 1030856

Ion Ratio Lower Upper

142 100

141 91.7 74.4 111.6

116 3.7 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 15.623 ng/ul

RT: 13.010 min Scan# 1636

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Tgt Ion:216 Resp: 532205




Ion Ratio Lower Upper


216 100

214 79.9 63.0 94.4

179 20.8 16.9 25.3

108 17.5 14.9 22.3









#40
Hexachlorocyclopentadiene
Concen: 18.283 ng/ul
RT: 12.987 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:237 Resp: 419973
Ion Ratio Lower Upper
237 100
235 61.2 49.9 74.9
272 12.9 10.0 15.0



#41
2,4,6-Trichlorophenol
Concen: 17.330 ng/ul
RT: 13.245 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:196 Resp: 363765
Ion Ratio Lower Upper
196 100
198 96.2 76.8 115.2
200 31.1 24.6 36.8





#42

2,4,5-Trichlorophenol

Concen: 17.390 ng/ul

RT: 13.316 min Scan# 1

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:196 Resp: 393424

Ion Ratio Lower Upper

196 100

198 97.8 74.2 111.4

200 30.9 24.6 37.0



#43

1,1'-Biphenyl

Concen: 15.846 ng/ul

RT: 13.640 min Scan# 1743

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:154 Resp: 1383551


Ion Ratio Lower Upper

154 100

153 40.0 31.9 47.9

76 14.0 11.3 16.9





#44

2-Chloronaphthalene

Concen: 16.032 ng/ul

RT: 13.687 min Scan# 1

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:162 Resp: 1090360

Ion Ratio Lower Upper

162 100

164 33.0 26.3 39.5

127 38.7 30.4 45.6



#45

2-Nitroaniline

Concen: 19.005 ng/ul

RT: 13.887 min Scan# 1785

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion: 65 Resp: 301003


Ion Ratio Lower Upper

65 100

92 73.5 58.3 87.5

138 110.1 87.9 131.9









#47
Dimethylphthalate
Concen: 16.620 ng/ul
RT: 14.257 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:163 Resp: 1350803
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.1 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 18.671 ng/ul
RT: 14.381 min Scan# 1869
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:165 Resp: 268155
Ion Ratio Lower Upper
165 100
89 57.7 48.1 72.1
121 20.7 17.3 25.9









#50
Acenaphthylene
Concen: 16.337 ng/ul
RT: 14.534 min Scan# 1895
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:152 Resp: 1716914
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.1 10.4 15.6



#51
3-Nitroaniline
Concen: 17.598 ng/ul
RT: 14.710 min Scan# 1925
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:138 Resp: 228690
Ion Ratio Lower Upper
138 100
108 10.9 8.5 12.7
92 114.3 95.8 143.6









#52
Acenaphthene
Concen: 15.990 ng/ul
RT: 14.875 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:153 Resp: 1162917
Ion Ratio Lower Upper
153 100
152 47.5 38.8 58.2
154 86.9 69.6 104.4



#53
2,4-Dinitrophenol
Concen: 16.637 ng/ul
RT: 14.916 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:184 Resp: 145426
Ion Ratio Lower Upper
184 100
63 65.5 56.3 84.5
154 60.6 48.9 73.3









#55
4-Nitrophenol
Concen: 18.040 ng/ul
RT: 15.010 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:109 Resp: 184641
Ion Ratio Lower Upper
109 100
139 145.9 115.1 172.7
65 135.0 110.6 166.0



#56
Dibenzofuran
Concen: 16.473 ng/ul
RT: 15.204 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:168 Resp: 1616518
Ion Ratio Lower Upper
168 100
139 35.8 28.6 42.8





#57

2,4-Dinitrotoluene

Concen: 19.109 ng/ul

RT: 15.169 min Scan# 2003

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:165 Resp: 384407

Ion Ratio Lower Upper

165 100

63 39.8 35.7 53.5

182 2.6 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 18.095 ng/ul

RT: 15.428 min Scan# 2047

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Tgt Ion:232 Resp: 333167




Ion Ratio Lower Upper


232 100

131 43.3 38.2 57.4

130 2.2 2.1 3.1

166 29.5 26.6 39.8









#59
Diethylphthalate
Concen: 17.328 ng/ul
RT: 15.616 min Scan# 2079
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:149 Resp: 1330710
Ion Ratio Lower Upper
149 100
177 22.6 17.5 26.3
150 12.2 9.6 14.4



#61
Fluorene
Concen: 16.720 ng/ul
RT: 15.857 min Scan# 2120
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:166 Resp: 1294303
Ion Ratio Lower Upper
166 100
165 98.0 79.3 118.9
167 13.1 10.7 16.1









#62
4-Chlorophenyl-phenylether
Concen: 16.589 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:204 Resp: 628422
Ion Ratio Lower Upper
204 100
206 32.7 26.2 39.2
141 54.5 46.8 70.2



#63
4-Nitroaniline
Concen: 17.666 ng/ul
RT: 15.869 min Scan# 2122
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:138 Resp: 191614
Ion Ratio Lower Upper
138 100
92 53.8 43.7 65.5
108 74.6 57.4 86.2









#66
4,6-Dinitro-2-methylphenol
Concen: 16.412 ng/ul
RT: 15.934 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:198 Resp: 222205
Ion Ratio Lower Upper
198 100
182 4.2 3.9 5.9
77 24.9 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 15.995 ng/ul
RT: 16.057 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:169 Resp: 1100720
Ion Ratio Lower Upper
169 100
168 66.9 53.5 80.3
167 37.2 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 16.294 ng/ul
RT: 16.739 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:248 Resp: 384330
Ion Ratio Lower Upper
248 100
250 98.0 79.5 119.3
141 75.1 60.2 90.4



#69
Hexachlorobenzene
Concen: 15.914 ng/ul
RT: 16.863 min Scan# 2291
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:284 Resp: 442162
Ion Ratio Lower Upper
284 100
142 35.6 30.2 45.2
249 30.8 24.5 36.7





#70

Atrazine

Concen: 17.086 ng/ul

RT: 17.010 min Scan# 21

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:200 Resp: 395673

Ion Ratio Lower Upper

200 100

173 26.3 21.8 32.6

215 48.9 39.0 58.4



#71

Pentachlorophenol

Concen: 16.581 ng/ul

RT: 17.210 min Scan# 2350

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:266 Resp: 268930


Ion Ratio Lower Upper

266 100

264 63.3 48.9 73.3

268 64.5 49.6 74.4





#72

Phenanthrene

Concen: 16.233 ng/ul

RT: 17.604 min Scan# 2418

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:178 Resp: 2033447

Ion Ratio Lower Upper

178 100

179 15.9 12.2 18.4

176 19.2 15.4 23.0



#74

Anthracene

Concen: 16.594 ng/ul

RT: 17.698 min Scan# 2433

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:178 Resp: 2069509


Ion Ratio Lower Upper

178 100

179 15.2 12.6 18.8

176 18.7 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 14.585 ng/uL

RT: 13.610 min Scan# 1

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:216 Resp: 547429

Ion Ratio Lower Upper

216 100

214 79.0 60.5 90.7

218 47.8 38.6 58.0



#76

Pentachlorobenzene

Concen: 14.939 ng/uL

RT: 15.122 min Scan# 1995

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:250 Resp: 533441


Ion Ratio Lower Upper

250 100

248 62.0 50.5 75.7

252 63.5 50.5 75.7





#77

Carbazole

Concen: 17.555 ng/ul

RT: 17.963 min Scan# 2478

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:167 Resp: 1841400

Ion Ratio Lower Upper

167 100

166 22.0 17.1 25.7

139 12.8 10.1 15.1



#78

Di-n-butylphthalate

Concen: 19.582 ng/ul

RT: 18.492 min Scan# 2568

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:149 Resp: 2296800


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2

104 4.8 4.1 6.1





#80

Fluoranthene

Concen: 15.889 ng/ul

RT: 19.557 min Scan# 21

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:202 Resp: 2286184

Ion Ratio Lower Upper

202 100

101 11.2 10.1 15.1

100 8.8 7.7 11.5



#82

Pyrene

Concen: 15.304 ng/ul

RT: 19.910 min Scan# 2809

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:202 Resp: 2333064


Ion Ratio Lower Upper

202 100

101 13.1 11.9 17.9

100 11.0 9.7 14.5





#83

Butylbenzylphthalate

Concen: 21.801 ng/ul

RT: 20.763 min Scan# 2954

Delta R.T. 0.000 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:149 Resp: 997832

Ion Ratio Lower Upper

149 100

91 70.5 60.1 90.1

206 20.4 16.0 24.0



#84

3,3'-Dichlorobenzidine

Concen: 20.280 ng/ul

RT: 21.545 min Scan# 3087

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:252 Resp: 593431


Ion Ratio Lower Upper

252 100

254 63.8 53.3 79.9

126 12.1 9.0 13.6





#85

Benzo(a)anthracene

Concen: 17.147 ng/ul

RT: 21.627 min Scan# 3101

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:228 Resp: 1989795

Ion Ratio Lower Upper

228 100

229 19.7 15.9 23.9

226 27.3 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 25.217 ng/ul

RT: 21.521 min Scan# 3083

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41




Tgt Ion:149 Resp: 1452245


Ion Ratio Lower Upper

149 100

167 27.0 22.0 33.0

279 3.9 2.9 4.3





#87

Chrysene

Concen: 16.011 ng/ul

RT: 21.680 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:228 Resp: 1870807

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 19.9 15.8 23.8



#89

Di-n-octyl phthalate

Concen: 27.325 ng/ul




RT: 22.510 min Scan# 3251


Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Tgt Ion:149 Resp: 2383615









#90
Benzo(b)fluoranthene
Concen: 16.697 ng/ul
RT: 23.439 min Scan# 3410
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:252 Resp: 1612147
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.2
125 10.5 8.4 12.6



#91
Benzo(k)fluoranthene
Concen: 16.504 ng/ul
RT: 23.492 min Scan# 3418
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:252 Resp: 1621795
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 10.2 8.9 13.3









#93
Benzo(a)pyrene
Concen: 16.293 ng/ul
RT: 24.127 min Scan# 31
Delta R.T. 0.000 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41




Instrument :
BNA_P
ClientSampleId :
SSTD020701


Tgt Ion:252 Resp: 1379699
Ion Ratio Lower Upper
252 100
253 21.8 17.5 26.3
125 11.6 9.4 14.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 16.416 ng/ul
RT: 26.998 min Scan# 4014
Delta R.T. -0.006 min
Lab File: BPCCC003.D
Acq: 08 May 2023 20:41

Tgt Ion:276 Resp: 1695607
Ion Ratio Lower Upper
276 100
138 24.0 19.4 29.0
277 24.6 19.4 29.2





#95

Dibenzo(a,h)anthracene

Concen: 16.545 ng/ul

RT: 27.021 min Scan# 4019

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Instrument :

BNA_P

ClientSampleId :

SSTD020701





Tgt Ion:278 Resp: 1400699

Ion Ratio Lower Upper

278 100

139 15.8 12.9 19.3

279 23.8 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 15.880 ng/ul

RT: 27.851 min Scan# 4159

Delta R.T. -0.006 min

Lab File: BPCCC003.D

Acq: 08 May 2023 20:41

Tgt Ion:276 Resp: 1396047

Ion Ratio Lower Upper

276 100

138 23.1 18.0 27.0

277 23.5 19.8 29.8

