

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050823\
 Data File : BP015006.D
 Acq On : 08 May 2023 22:56
 Operator : CG/JU
 Sample : 02609-15MS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JREJ4MS

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/09/2023
 Supervised By : Jagrut Upadhyay 05/09/2023

Quant Time: May 08 23:59:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 05 12:12:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.163	152	362738	20.000	ng/u1	0.00
20) Naphthalene-d8	10.998	136	1380879	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.810	164	861301	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.563	188	1890970	20.000	ng/u1	0.00
79) Chrysene-d12	21.645	240	1547039	20.000	ng/u1	0.00
88) Perylene-d12	24.245	264	1441162	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.458	96	33502	3.434	ng/uL	0.00
4) Pyridine-d5	3.899	84	181336	6.800	ng/u1	0.00
7) Phenol-d5	7.305	99	213762	6.797	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.475	67	556696	28.639	ng/u1	0.00
11) 2-Chlorophenol-d4	7.687	132	569022	24.607	ng/u1	0.00
15) 4-Methylphenol-d8	8.875	113	400224	16.514	ng/u1	0.00
21) Nitrobenzene-d5	9.340	128	380416	38.735	ng/u1	0.00
24) 2-Nitrophenol-d4	10.069	143	411787	40.833	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.610	165	674561	34.067	ng/u1	0.00
31) 4-Chloroaniline-d4	11.128	131	328995	11.487	ng/u1	0.00
46) Dimethylphthalate-d6	14.210	166	2076228	34.060	ng/u1	0.00
49) Acenaphthylene-d8	14.504	160	2281467	31.207	ng/u1	0.00
54) 4-Nitrophenol-d4	14.992	143	88678	8.362	ng/u1	0.00
60) Fluorene-d10	15.798	176	1720397	32.804	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.922	200	347059	32.603	ng/u1	0.00
73) Anthracene-d10	17.663	188	2708128	31.874	ng/u1	0.00
81) Pyrene-d10	19.880	212	3086968	29.586	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.074	264	2472862	35.216	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.493	88	42511	4.022	ng/uL	96
5) Pyridine	3.916	79	313803	11.304	ng/u1	99
6) Benzaldehyde	7.287	77	557418m	64.828	ng/u1	
8) Phenol	7.334	94	245952	7.449	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.575	93	803563	29.281	ng/u1	99
12) 2-Chlorophenol	7.716	128	612807	24.463	ng/u1	100
13) 2-Methylphenol	8.604	108	483815	19.652	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.699	45	993857	28.551	ng/u1	100
16) Acetophenone	8.999	105	1251438	31.581	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.981	70	637773	31.580	ng/u1	99
18) 4-Methylphenol	8.940	108	467311	17.170	ng/u1	99
19) Hexachloroethane	9.257	117	244399	22.487	ng/u1	99
22) Nitrobenzene	9.381	77	938205	36.358	ng/u1	97
23) Isophorone	9.916	82	1904961	37.353	ng/u1	99
25) 2-Nitrophenol	10.098	139	451019	38.887	ng/u1	98
26) 2,4-Dimethylphenol	10.151	107	719770	28.596	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.393	93	1158166	35.523	ng/u1	100
29) 2,4-Dichlorophenol	10.640	162	695667	33.966	ng/u1	99
30) Naphthalene	11.046	128	2003731	26.398	ng/u1	100
32) 4-Chloroaniline	11.157	127	712679	23.845	ng/u1	99
33) Hexachlorobutadiene	11.328	225	329304	24.918	ng/u1	99
34) Caprolactam	11.945	113	41076	6.510	ng/u1	93
35) 4-Chloro-3-methylphenol	12.263	107	652667	30.998	ng/u1	99
36) 2-Methylnaphthalene	12.645	142	1422825	30.312	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050823\
 Data File : BP015006.D
 Acq On : 08 May 2023 22:56
 Operator : CG/JU
 Sample : 02609-15MS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JREJ4MS


Quant Time: May 08 23:59:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 05 12:12:36 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 05/09/2023
 Supervised By : Jagrut Upadhyay 05/09/2023


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.863	142	1471005	30.402	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	13.010	216	770922	28.695	ng/ul	99
40) Hexachlorocyclopentadiene	12.987	237	357950	19.759	ng/ul	98
41) 2,4,6-Trichlorophenol	13.245	196	560403	33.852	ng/ul	99
42) 2,4,5-Trichlorophenol	13.316	196	619650	34.729	ng/ul	96
43) 1,1'-Biphenyl	13.639	154	2009836	29.187	ng/ul	100
44) 2-Chloronaphthalene	13.687	162	1568151	29.235	ng/ul	99
45) 2-Nitroaniline	13.887	65	497874	39.860	ng/ul	99
47) Dimethylphthalate	14.257	163	2121464	33.098	ng/ul	100
48) 2,6-Dinitrotoluene	14.381	165	467962	41.315	ng/ul	96
50) Acenaphthylene	14.534	152	2498760	30.148	ng/ul	99
51) 3-Nitroaniline	14.710	138	449920	43.901	ng/ul	95
52) Acenaphthene	14.869	153	1720477	29.996	ng/ul	98
53) 2,4-Dinitrophenol	14.916	184	175369	25.439	ng/ul	93
55) 4-Nitrophenol	15.010	109	69204	8.573	ng/ul	96
56) Dibenzofuran	15.204	168	2441102	31.542	ng/ul	99
57) 2,4-Dinitrotoluene	15.169	165	667907	42.100	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	15.428	232	529833	36.488	ng/ul	96
59) Diethylphthalate	15.616	149	2169377	35.819	ng/ul	98
61) Fluorene	15.857	166	1904894	31.202	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.845	204	955124	31.970	ng/ul	98
63) 4-Nitroaniline	15.875	138	451084	52.733	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.933	198	358144	31.719	ng/ul#	95
67) N-Nitrosodiphenylamine	16.057	169	1466677	25.555	ng/ul	100
68) 4-Bromophenyl-phenylether	16.739	248	652105	33.150	ng/ul	98
69) Hexachlorobenzene	16.863	284	759461	32.776	ng/ul	96
70) Atrazine	17.010	200	553320	28.650	ng/ul	99
71) Pentachlorophenol	17.210	266	439899	32.522	ng/ul	97
72) Phenanthrene	17.610	178	3271940	31.320	ng/ul	99
74) Anthracene	17.698	178	3249850	31.246	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.610	216	810186	25.882	ng/ul	99
76) Pentachlorobenzene	15.128	250	827418	27.784	ng/ul	99
77) Carbazole	17.963	167	3047337	34.835	ng/ul	100
78) Di-n-butylphthalate	18.492	149	3862802	39.488	ng/ul	100
80) Fluoranthene	19.557	202	3851824	30.085	ng/ul	98
82) Pyrene	19.910	202	3838069	28.294	ng/ul	98
83) Butylbenzylphthalate	20.763	149	1793568	44.040	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.551	252	563391	21.637	ng/ul	98
85) Benzo(a)anthracene	21.627	228	3616667	35.025	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.527	149	2675075	52.203	ng/ul#	99
87) Chrysene	21.686	228	3417139	32.866	ng/ul	100
89) Di-n-octyl phthalate	22.515	149	4920516	58.475	ng/ul	100
90) Benzo(b)fluoranthene	23.445	252	3346725	35.932	ng/ul	99
91) Benzo(k)fluoranthene	23.498	252	3273286	34.531	ng/ul	98
93) Benzo(a)pyrene	24.133	252	2787065	34.120	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	27.009	276	3422336	34.348	ng/ul	98
95) Dibenzo(a,h)anthracene	27.027	278	2812567	34.440	ng/ul	100
96) Benzo(g,h,i)perylene	27.856	276	2811596	33.155	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 4.022 ng/uL
RT: 3.493 min Scan# 18
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion: 88 Resp: 4251
Ion Ratio Lower Upper
88 100
43 27.6 20.6 31.0
58 64.4 54.4 81.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#5
Pyridine
Concen: 11.304 ng/uL
RT: 3.916 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion: 79 Resp: 313803
Ion Ratio Lower Upper
79 100
52 68.8 54.7 82.1
51 33.4 25.5 38.3





#6
Benzaldehyde
Concen: 64.828 ng/ul m
RT: 7.287 min Scan# 663
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion: 77 Resp: 557418
Ion Ratio Lower Upper
77 100
105 99.4 80.7 121.1
106 97.7 77.6 116.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#8
Phenol
Concen: 7.449 ng/ul
RT: 7.334 min Scan# 671
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion: 94 Resp: 245952
Ion Ratio Lower Upper
94 100
65 26.1 21.4 32.0
66 32.8 25.8 38.8





#10
Bis(2-Chloroethyl)ether
Concen: 29.281 ng/ul
RT: 7.575 min Scan# 712
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion: 93 Resp: 80356
Ion Ratio Lower Upper
93 100
63 71.0 58.1 87.1
95 32.9 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#12
2-Chlorophenol
Concen: 24.463 ng/ul
RT: 7.716 min Scan# 736
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Tgt Ion:128 Resp: 612807
Ion Ratio Lower Upper
128 100
64 45.7 36.6 55.0
130 32.1 25.8 38.8









#13
2-Methylphenol
Concen: 19.652 ng/ul
RT: 8.604 min Scan# 887
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Instrument :
BNA_P
ClientSampleId :
JREJ4MS

Tgt Ion:108 Resp: 48381
Ion Ratio Lower Upper
108 100
107 89.8 71.7 107.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#14
2,2'-oxybis(1-Chloropropane)
Concen: 28.551 ng/ul
RT: 8.699 min Scan# 903
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion: 45 Resp: 993857
Ion Ratio Lower Upper
45 100
77 15.5 12.3 18.5
79 11.9 9.5 14.3





#16
Acetophenone
Concen: 31.581 ng/ul
RT: 8.999 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion: 105 Resp: 1251438
Ion Ratio Lower Upper
105 100
77 74.3 61.2 91.8
51 23.7 19.4 29.2




Manual Integrations
APPROVED


Reviewed By : Christian Giraldo 05/09/2023
Supervised By : Jagrut Upadhyay 05/09/2023



#17
N-Nitroso-di-n-propylamine
Concen: 31.580 ng/ul
RT: 8.981 min Scan# 951
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion: 70 Resp: 637773
Ion Ratio Lower Upper
70 100
42 52.1 41.9 62.9
101 10.4 7.7 11.5
130 21.6 16.2 24.2





#18
4-Methylphenol
Concen: 17.170 ng/ul
RT: 8.940 min Scan# 944
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:108 Resp: 46731
Ion Ratio Lower Upper
108 100
107 107.8 87.2 130.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#19
Hexachloroethane
Concen: 22.487 ng/ul
RT: 9.257 min Scan# 998
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:117 Resp: 244399
Ion Ratio Lower Upper
117 100
201 99.2 78.9 118.3
199 61.8 48.3 72.5





#22
Nitrobenzene
Concen: 36.358 ng/ul
RT: 9.381 min Scan# 1109
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion: 77 Resp: 938209
Ion Ratio Lower Upper
77 100
123 46.0 34.8 52.2
65 12.7 10.2 15.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#23
Isophorone
Concen: 37.353 ng/ul
RT: 9.916 min Scan# 1110
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion: 82 Resp: 1904961
Ion Ratio Lower Upper
82 100
95 6.9 5.3 7.9
138 18.1 13.9 20.9





#25
2-Nitrophenol
Concen: 38.887 ng/ul
RT: 10.098 min Scan# 1141
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:139 Resp: 451019
Ion Ratio Lower Upper
139 100
65 48.7 40.5 60.7
109 27.6 22.4 33.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#26
2,4-Dimethylphenol
Concen: 28.596 ng/ul
RT: 10.151 min Scan# 1150
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion:107 Resp: 719770
Ion Ratio Lower Upper
107 100
121 51.6 40.2 60.2
122 89.4 69.6 104.4





#27
Bis(2-Chloroethoxy)methane
Concen: 35.523 ng/ul
RT: 10.393 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion: 93 Resp: 115816
Ion Ratio Lower Upper
93 100
95 32.7 26.1 39.1
123 11.0 9.0 13.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#29
2,4-Dichlorophenol
Concen: 33.966 ng/ul
RT: 10.640 min Scan# 1233
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Tgt Ion:162 Resp: 695667
Ion Ratio Lower Upper
162 100
164 65.4 51.5 77.3
98 36.4 28.8 43.2









#30
Naphthalene
Concen: 26.398 ng/ul
RT: 11.046 min Scan# 1303
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Instrument :
BNA_P
ClientSampleId :
JREJ4MS

Tgt Ion:128 Resp: 200373
Ion Ratio Lower Upper
128 100
129 10.9 8.7 13.1
127 13.0 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#32
4-Chloroaniline
Concen: 23.845 ng/ul
RT: 11.157 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:127 Resp: 712679
Ion Ratio Lower Upper
127 100
129 32.2 25.5 38.3





#33
Hexachlorobutadiene
Concen: 24.918 ng/ul
RT: 11.328 min Scan# 1135
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:225 Resp: 329304
Ion Ratio Lower Upper
225 100
223 62.7 50.6 76.0
227 63.2 51.4 77.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#34
Caprolactam
Concen: 6.510 ng/ul
RT: 11.945 min Scan# 1455
Delta R.T. 0.018 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Tgt Ion:113 Resp: 41076
Ion Ratio Lower Upper
113 100
55 176.8 128.5 192.7
56 118.5 94.3 141.5









#35
4-Chloro-3-methylphenol
Concen: 30.998 ng/ul
RT: 12.263 min Scan# 1509
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Instrument :
BNA_P
ClientSampleId :
JREJ4MS

Tgt Ion:107 Resp: 65266
Ion Ratio Lower Upper
107 100
144 26.5 20.6 30.8
142 79.7 63.2 94.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#36
2-Methylnaphthalene
Concen: 30.312 ng/ul
RT: 12.645 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:142 Resp: 1422825
Ion Ratio Lower Upper
142 100
141 88.6 70.5 105.7





#37
1-Methylnaphthalene
Concen: 30.402 ng/ul
RT: 12.863 min Scan# 1611
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:142 Resp: 147100
Ion Ratio Lower Upper
142 100
141 91.3 74.4 111.6
116 3.7 2.9 4.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#39
1,2,4,5-Tetrachlorobenzene
Concen: 28.695 ng/ul
RT: 13.010 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:216 Resp: 770922
Ion Ratio Lower Upper
216 100
214 79.0 63.0 94.4
179 20.6 16.9 25.3
108 16.3 14.9 22.3





#40
Hexachlorocyclopentadiene
Concen: 19.759 ng/ul
RT: 12.987 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:237 Resp: 357950
Ion Ratio Lower Upper
237 100
235 63.6 49.9 74.9
272 13.0 10.0 15.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#41
2,4,6-Trichlorophenol
Concen: 33.852 ng/ul
RT: 13.245 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion:196 Resp: 560403
Ion Ratio Lower Upper
196 100
198 95.4 76.8 115.2
200 31.3 24.6 36.8





#42
2,4,5-Trichlorophenol
Concen: 34.729 ng/ul
RT: 13.316 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:196 Resp: 619650
Ion Ratio Lower Upper
196 100
198 98.3 74.2 111.4
200 31.2 24.6 37.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#43
1,1'-Biphenyl
Concen: 29.187 ng/ul
RT: 13.639 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:154 Resp: 2009836
Ion Ratio Lower Upper
154 100
153 40.0 31.9 47.9
76 13.9 11.3 16.9





#44
2-Chloronaphthalene
Concen: 29.235 ng/ul
RT: 13.687 min Scan# 1751
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:162 Resp: 156815
Ion Ratio Lower Upper
162 100
164 33.2 26.3 39.5
127 38.6 30.4 45.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#45
2-Nitroaniline
Concen: 39.860 ng/ul
RT: 13.887 min Scan# 1785
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion: 65 Resp: 497874
Ion Ratio Lower Upper
65 100
92 71.3 58.3 87.5
138 109.8 87.9 131.9





#47
Dimethylphthalate
Concen: 33.098 ng/ul
RT: 14.257 min Scan# 1848
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:163 Resp: 2121464
Ion Ratio Lower Upper
163 100
194 4.1 3.1 4.7
164 10.1 8.0 12.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#48
2,6-Dinitrotoluene
Concen: 41.315 ng/ul
RT: 14.381 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:165 Resp: 467962
Ion Ratio Lower Upper
165 100
89 56.4 48.1 72.1
121 21.7 17.3 25.9





#50
Acenaphthylene
Concen: 30.148 ng/ul
RT: 14.534 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:152 Resp: 2498760
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.2 10.4 15.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#51
3-Nitroaniline
Concen: 43.901 ng/ul
RT: 14.710 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion:138 Resp: 449920
Ion Ratio Lower Upper
138 100
108 10.4 8.5 12.7
92 114.0 95.8 143.6





#52
Acenaphthene
Concen: 29.996 ng/ul
RT: 14.869 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:153 Resp: 172047
Ion Ratio Lower Upper
153 100
152 47.7 38.8 58.2
154 89.0 69.6 104.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#53
2,4-Dinitrophenol
Concen: 25.439 ng/ul
RT: 14.916 min Scan# 1960
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion:184 Resp: 175369
Ion Ratio Lower Upper
184 100
63 60.4 56.3 84.5
154 60.3 48.9 73.3





#55
4-Nitrophenol
Concen: 8.573 ng/ul
RT: 15.010 min Scan# 1976
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:109 Resp: 69204
Ion Ratio Lower Upper
109 100
139 142.3 115.1 172.7
65 130.6 110.6 166.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#56
Dibenzofuran
Concen: 31.542 ng/ul
RT: 15.204 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:168 Resp: 2441102
Ion Ratio Lower Upper
168 100
139 36.5 28.6 42.8





#57
2,4-Dinitrotoluene
Concen: 42.100 ng/ul
RT: 15.169 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:165 Resp: 66790
Ion Ratio Lower Upper
165 100
63 38.3 35.7 53.5
182 2.8 2.2 3.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.488 ng/ul
RT: 15.428 min Scan# 2047
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Tgt Ion:232 Resp: 529833
Ion Ratio Lower Upper
232 100
131 44.6 38.2 57.4
130 2.5 2.1 3.1
166 30.9 26.6 39.8









#59
Diethylphthalate
Concen: 35.819 ng/ul
RT: 15.616 min Scan# 2079
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Instrument :
BNA_P
Client SampleId :
JREJ4MS

Tgt Ion:149 Resp: 216937
Ion Ratio Lower Upper
149 100
177 22.9 17.5 26.3
150 12.6 9.6 14.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#61
Fluorene
Concen: 31.202 ng/ul
RT: 15.857 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:166 Resp: 1904894
Ion Ratio Lower Upper
166 100
165 97.5 79.3 118.9
167 13.2 10.7 16.1





#62
4-Chlorophenyl-phenylether
Concen: 31.970 ng/ul
RT: 15.845 min Scan# 2118
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion: 204 Resp: 955124
Ion Ratio Lower Upper
204 100
206 33.3 26.2 39.2
141 56.2 46.8 70.2

Manual Integrations
APPROVED



Reviewed By : Christian Giraldo 05/09/2023
Supervised By : Jagrut Upadhyay 05/09/2023




#63
4-Nitroaniline
Concen: 52.733 ng/ul
RT: 15.875 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56



Tgt Ion: 138 Resp: 451084
Ion Ratio Lower Upper
138 100
92 49.5 43.7 65.5
108 72.9 57.4 86.2









#66
4,6-Dinitro-2-methylphenol
Concen: 31.719 ng/ul
RT: 15.933 min Scan# 2133
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Instrument :
BNA_P
ClientSampleId :
JREJ4MS

Tgt Ion:198 Resp: 358144
Ion Ratio Lower Upper
198 100
182 3.9 3.9 5.9
77 23.8 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#67
N-Nitrosodiphenylamine
Concen: 25.555 ng/ul
RT: 16.057 min Scan# 2154
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Tgt Ion:169 Resp: 1466677
Ion Ratio Lower Upper
169 100
168 66.8 53.5 80.3
167 36.0 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 33.150 ng/ul
RT: 16.739 min Scan# 2270
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Instrument :
BNA_P
ClientSampleId :
JREJ4MS

Tgt Ion:248 Resp: 65210
Ion Ratio Lower Upper
248 100
250 98.2 79.5 119.3
141 73.3 60.2 90.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#69
Hexachlorobenzene
Concen: 32.776 ng/ul
RT: 16.863 min Scan# 2291
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56

Tgt Ion:284 Resp: 759461
Ion Ratio Lower Upper
284 100
142 33.7 30.2 45.2
249 30.5 24.5 36.7





#70

Atrazine

Concen: 28.650 ng/ul

RT: 17.010 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56

Instrument :

BNA_P

ClientSampleId :

JREJ4MS





Tgt Ion:200	Resp: 553320
Ion Ratio	Lower Upper
200 100	
173 27.5	21.8 32.6
215 49.9	39.0 58.4

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023

Supervised By :Jagrut Upadhyay 05/09/2023



#71

Pentachlorophenol

Concen: 32.522 ng/ul




RT: 17.210 min Scan# 2350


Delta R.T. -0.000 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56

Tgt Ion:266	Resp: 439899
Ion Ratio	Lower Upper
266 100	
264 63.1	48.9 73.3
268 64.4	49.6 74.4





#72

Phenanthrene

Concen: 31.320 ng/ul

RT: 17.610 min Scan# 2417

Delta R.T. -0.000 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56

Instrument :

BNP_P

ClientSampleId :

JREJ4MS

Tgt Ion:178 Resp: 3271940

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4





176 19.5 15.4 23.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023

Supervised By :Jagrut Upadhyay 05/09/2023



#74

Anthracene

Concen: 31.246 ng/ul

RT: 17.698 min Scan# 2433

Delta R.T. -0.000 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56




Tgt Ion:178 Resp: 3249850


Ion Ratio Lower Upper

178 100

179 15.6 12.6 18.8


176 19.0 15.0 22.4





#75
1,2,3,4-Tetrachlorobenzene
Concen: 25.882 ng/uL
RT: 13.610 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:216 Resp: 810180
Ion Ratio Lower Upper
216 100
214 77.4 60.5 90.7
218 48.5 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#76
Pentachlorobenzene
Concen: 27.784 ng/uL
RT: 15.128 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:250 Resp: 827418
Ion Ratio Lower Upper
250 100
248 63.2 50.5 75.7
252 64.5 50.5 75.7





#77
 Carbazole
 Concen: 34.835 ng/ul
 RT: 17.963 min Scan# 2478
 Delta R.T. -0.000 min
 Lab File: BP015006.D
 Acq: 08 May 2023 22:56




Instrument :
 BNA_P
 ClientSampleId :
 JREJ4MS




Tgt Ion:167 Resp: 304733
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.1 25.7
 139 12.5 10.1 15.1

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 05/09/2023
 Supervised By :Jagrut Upadhyay 05/09/2023




#78
 Di-n-butylphthalate
 Concen: 39.488 ng/ul
 RT: 18.492 min Scan# 2568
 Delta R.T. -0.000 min
 Lab File: BP015006.D
 Acq: 08 May 2023 22:56




Tgt Ion:149 Resp: 3862802
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 4.9 4.1 6.1





#80
Fluoranthene
Concen: 30.085 ng/u1
RT: 19.557 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:202 Resp: 3851824
Ion Ratio Lower Upper
202 100
101 11.7 10.1 15.1
100 9.0 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#82
Pyrene
Concen: 28.294 ng/u1
RT: 19.910 min Scan# 2809
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion:202 Resp: 3838069
Ion Ratio Lower Upper
202 100
101 14.0 11.9 17.9
100 11.4 9.7 14.5





#83
Butylbenzylphthalate
Concen: 44.040 ng/ul
RT: 20.763 min Scan# 2954
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:149 Resp: 1793568
Ion Ratio Lower Upper
149 100
91 70.7 60.1 90.1
206 21.4 16.0 24.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#84
3,3'-Dichlorobenzidine
Concen: 21.637 ng/ul
RT: 21.551 min Scan# 3088
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Tgt Ion:252 Resp: 563391
Ion Ratio Lower Upper
252 100
254 64.7 53.3 79.9
126 11.2 9.0 13.6





#85
Benzo(a)anthracene
Concen: 35.025 ng/ul
RT: 21.627 min Scan# 3101
Delta R.T. -0.006 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:228 Resp: 361666
Ion Ratio Lower Upper
228 100
229 19.8 15.9 23.9
226 27.5 22.1 33.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#86
Bis(2-ethylhexyl)phthalate
Concen: 52.203 ng/ul
RT: 21.527 min Scan# 3084
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:149 Resp: 2675075
Ion Ratio Lower Upper
149 100
167 28.0 22.0 33.0
279 4.7 2.9 4.3





#87
Chrysene
Concen: 32.866 ng/ul
RT: 21.686 min Scan# 3111
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS



Tgt Ion:228 Resp: 3417139
Ion Ratio Lower Upper
228 100
226 30.4 24.3 36.5
229 20.1 15.8 23.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023



#89
Di-n-octyl phthalate
Concen: 58.475 ng/ul
RT: 22.515 min Scan# 3252
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56


Tgt Ion:149 Resp: 4920516





#90
Benzo(b)fluoranthene
Concen: 35.932 ng/ul
RT: 23.445 min Scan# 3409
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56




Instrument :
BNA_P
ClientSampleId :
JREJ4MS




Tgt Ion:252 Resp: 3346729
Ion Ratio Lower Upper
252 100
253 22.4 17.4 26.2
125 10.7 8.4 12.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/09/2023
Supervised By :Jagrut Upadhyay 05/09/2023




#91
Benzo(k)fluoranthene
Concen: 34.531 ng/ul
RT: 23.498 min Scan# 3419
Delta R.T. -0.000 min
Lab File: BP015006.D
Acq: 08 May 2023 22:56



Tgt Ion:252 Resp: 3273286
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.6
125 10.2 8.9 13.3





#93

Benzo(a)pyrene

Concen: 34.120 ng/ul

RT: 24.133 min Scan# 3526

Delta R.T. 0.006 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56

Instrument :

BNA_P

ClientSampleId :

JREJ4MS

Tgt Ion:252 Resp: 278706

Ion Ratio Lower Upper

252 100

253 22.0 17.5 26.3





125 11.6 9.4 14.2

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023

Supervised By :Jagrut Upadhyay 05/09/2023



#94

Indeno(1,2,3-cd)pyrene

Concen: 34.348 ng/ul

RT: 27.009 min Scan# 4016

Delta R.T. 0.006 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56




Tgt Ion:276 Resp: 3422336


Ion Ratio Lower Upper

276 100

138 25.3 19.4 29.0

277 25.3 19.4 29.2





#95

Dibenzo(a,h)anthracene

Concen: 34.440 ng/ul

RT: 27.027 min Scan# 4018

Delta R.T. -0.000 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56

Instrument :

BNA_P

ClientSampleId :

JREJ4MS

Tgt Ion:278 Resp: 281256

Ion Ratio Lower Upper

278 100

139 16.3 12.9 19.3





279 23.6 19.0 28.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/09/2023

Supervised By :Jagrut Upadhyay 05/09/2023



#96

Benzo(g,h,i)perylene

Concen: 33.155 ng/ul

RT: 27.856 min Scan# 4160

Delta R.T. -0.000 min

Lab File: BP015006.D

Acq: 08 May 2023 22:56

Tgt Ion:276 Resp: 2811596

Ion Ratio Lower Upper

276 100

138 22.5 18.0 27.0

277 23.9 19.8 29.8

