

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050923\
 Data File : BP015030.D
 Acq On : 09 May 2023 22:01
 Operator : CG/JU
 Sample : 02609-24
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 JREL3

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/10/2023
 Supervised By : Jagrut Upadhyay 05/10/2023

Quant Time: May 09 23:11:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri May 05 12:12:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.158	152	393584	20.000	ng/u1	0.00
20) Naphthalene-d8	10.993	136	1882008	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.804	164	1239024	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.557	188	2601598	20.000	ng/u1	-0.01
79) Chrysene-d12	21.633	240	1744915	20.000	ng/u1	-0.01
88) Perylene-d12	24.227	264	1554535	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.458	96	35510	3.355	ng/uL	0.00
4) Pyridine-d5	3.899	84	115195	3.981	ng/u1	0.00
7) Phenol-d5	7.305	99	39567	1.160	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.475	67	490697	23.266	ng/u1	0.00
11) 2-Chlorophenol-d4	7.681	132	55120	2.197	ng/u1	0.00
15) 4-Methylphenol-d8	8.863	113	178803	6.800	ng/u1	-0.01
21) Nitrobenzene-d5	9.334	128	339807	25.387	ng/u1	0.00
24) 2-Nitrophenol-d4	10.063	143	14390	1.047	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.605	165	28794	1.067	ng/u1	-0.01
31) 4-Chloroaniline-d4	11.128	131	332928	8.529	ng/u1	0.00
46) Dimethylphthalate-d6	14.204	166	2064058	23.538	ng/u1	0.00
49) Acenaphthylene-d8	14.498	160	2635738	25.062	ng/u1	0.00
54) 4-Nitrophenol-d4	14.987	143	206	0.014	ng/u1	-0.01
60) Fluorene-d10	15.793	176	1976048	26.192	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.904	200	828m	0.057	ng/u1	-0.02
73) Anthracene-d10	17.657	188	3096203	26.487	ng/u1	0.00
81) Pyrene-d10	19.875	212	3296651	28.012	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.057	264	2081755	27.484	ng/u1	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP050923\
Data File : BP015030.D
Acq On : 09 May 2023 22:01
Operator : CG/JU
Sample : 02609-24
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
JREL3

Quant Time: May 09 23:11:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP050323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri May 05 12:12:36 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 05/10/2023
Supervised By : Jagrut Upadhyay 05/10/2023

