

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051021\
 Data File : BP005469.D
 Acq On : 10 May 2021 15:28
 Operator : CG/JU
 Sample : M2323-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 AST2-COPR3-1

Quant Time: May 10 17:11:40 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 07 12:37:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.705	152	16155	20.00	ng	0.00
21) Naphthalene-d8	10.499	136	45462	20.00	ng	# 0.00
39) Acenaphthene-d10	14.345	164	30865	20.00	ng	#-0.01
64) Phenanthrene-d10	17.104	188	67324	20.00	ng	# 0.00
76) Chrysene-d12	21.198	240	84484	20.00	ng	# 0.00
86) Perylene-d12	23.510	264	120188	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.293	112	50892	60.46	ng	0.00
7) Phenol-d6	6.893	99	46212	44.03	ng	0.00
23) Nitrobenzene-d5	8.869	82	49185	42.58	ng	-0.01
42) 2,4,6-Tribromophenol	15.845	330	45629	52.10	ng	-0.01
45) 2-Fluorobiphenyl	12.969	172	113616	41.52	ng	-0.01
79) Terphenyl-d14	19.669	244	215360	43.29	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051021\
Data File : BP005469.D
Acq On : 10 May 2021 15:28
Operator : CG/JU
Sample : M2323-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
AST2-COPR3-1

Quant Time: May 10 17:11:40 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 07 12:37:06 2021
Response via : Initial Calibration

