

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051021\
 Data File : BP005462.D
 Acq On : 10 May 2021 09:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 10 11:02:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 07 12:37:06 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	61	0.00
2	1,4-Dioxane	0.459	0.518	-12.9	78	0.01
3	Pyridine	0.923	1.178	-27.6#	78	0.00
4	n-Nitrosodimethylamine	0.742	0.851	-14.7	70	0.00
5 S	2-Fluorophenol	1.042	1.006	3.5	62	0.00
6	Aniline	1.403	1.307	6.8	59	0.01
7 S	Phenol-d6	1.299	1.214	6.5	57	0.00
8	2-Chlorophenol	1.119	1.071	4.3	59	0.01
9	Benzaldehyde	0.750	0.732	2.4	66	0.01
10 C	Phenol	1.300	1.163	10.5	57	0.00
11	bis(2-Chloroethyl)ether	0.898	0.818	8.9	59	0.00
12	1,3-Dichlorobenzene	1.489	1.347	9.5	61	0.01
13 C	1,4-Dichlorobenzene	1.484	1.370	7.7	60	0.01
14	1,2-Dichlorobenzene	1.423	1.367	3.9	61	0.00
15	Benzyl Alcohol	1.254	1.201	4.2	59	0.01
16	2,2'-oxybis(1-Chloropropane	0.443	0.408	7.9	57	0.00
17	2-Methylphenol	0.893	0.836	6.4	58	0.00
18	Hexachloroethane	0.615	0.631	-2.6	67	0.00
19 P	n-Nitroso-di-n-propylamine	0.952	0.919	3.5	59	0.00
20	3+4-Methylphenols	1.187	1.113	6.2	57	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
22	Acetophenone	0.564	0.531	5.9	61	0.00
23 S	Nitrobenzene-d5	0.508	0.496	2.4	63	0.00
24	Nitrobenzene	0.510	0.484	5.1	62	0.01
25	Isophorone	0.787	0.739	6.1	61	0.00
26 C	2-Nitrophenol	0.213	0.188	11.7	60	0.01
27	2,4-Dimethylphenol	0.283	0.262	7.4	61	0.00
28	bis(2-Chloroethoxy)methane	0.329	0.290	11.9	57	0.00
29 C	2,4-Dichlorophenol	0.444	0.419	5.6	62	0.00
30	1,2,4-Trichlorobenzene	0.567	0.559	1.4	65	0.01
31	Naphthalene	1.075	1.007	6.3	61	0.00
32	Benzoic acid	0.218	0.200	8.3	58	0.00
33	4-Chloroaniline	0.419	0.374	10.7	58	0.00
34 C	Hexachlorobutadiene	0.473	0.503	-6.3	72	0.00
35	Caprolactam	0.102	0.086	15.7	53	0.00
36 C	4-Chloro-3-methylphenol	0.396	0.368	7.1	60	0.00
37	2-Methylnaphthalene	0.863	0.830	3.8	63	0.00
38	1-Methylnaphthalene	0.814	0.783	3.8	62	0.01
39 I	Acenaphthene-d10	1.000	1.000	0.0	69	0.01
40	1,2,4,5-Tetrachlorobenzene	0.919	0.837	8.9	70	0.00
41 P	Hexachlorocyclopentadiene	0.313	0.349	-11.5	77	0.00
42 S	2,4,6-Tribromophenol	0.568	0.545	4.0	71	0.00
43 C	2,4,6-Trichlorophenol	0.574	0.528	8.0	66	0.00
44	2,4,5-Trichlorophenol	0.620	0.556	10.3	67	0.01
45 S	2-Fluorobiphenyl	1.773	1.588	10.4	66	0.00
46	1,1'-Biphenyl	1.540	1.333	13.4	63	0.01
47	2-Chloronaphthalene	1.282	1.152	10.1	66	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051021\
 Data File : BP005462.D
 Acq On : 10 May 2021 09:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 10 11:02:54 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 07 12:37:06 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.364	0.339	6.9	65	0.00
49	Acenaphthylene	1.847	1.635	11.5	64	0.01
50	Dimethylphthalate	1.726	1.549	10.3	66	0.00
51	2,6-Dinitrotoluene	0.388	0.342	11.9	64	0.00
52 C	Acenaphthene	1.149	1.025	10.8	65	0.00
53	3-Nitroaniline	0.278	0.238	14.4	59	0.00
54 P	2,4-Dinitrophenol	0.234	0.233	0.4	66	0.00
55	Dibenzofuran	2.162	1.968	9.0	67	0.00
56 P	4-Nitrophenol	0.210	0.181	13.8	59	0.00
57	2,4-Dinitrotoluene	0.572	0.507	11.4	66	0.00
58	Fluorene	1.765	1.663	5.8	67	0.01
59	2,3,4,6-Tetrachlorophenol	0.619	0.587	5.2	69	0.00
60	Diethylphthalate	1.672	1.542	7.8	67	0.00
61	4-Chlorophenyl-phenylether	1.105	1.088	1.5	74	0.00
62	4-Nitroaniline	0.267	0.247	7.5	66	0.00
63	Azobenzene	1.724	1.753	-1.7	72	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
65	4,6-Dinitro-2-methylphenol	0.158	0.148	6.3	68	0.00
66 c	n-Nitrosodiphenylamine	0.584	0.540	7.5	69	0.00
67	4-Bromophenyl-phenylether	0.312	0.289	7.4	73	0.01
68	Hexachlorobenzene	0.396	0.367	7.3	74	0.00
69	Atrazine	0.273	0.252	7.7	72	0.00
70 C	Pentachlorophenol	0.210	0.199	5.2	73	0.00
71	Phenanthrene	1.136	1.057	7.0	70	0.00
72	Anthracene	1.128	1.042	7.6	71	0.01
73	Carbazole	0.993	0.910	8.4	69	0.00
74	Di-n-butylphthalate	1.090	1.009	7.4	71	0.00
75 C	Fluoranthene	1.563	1.511	3.3	74	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
77	Benzydine	0.310	0.352	-13.5	80	0.00
78	Pyrene	1.132	1.072	5.3	75	0.00
79 S	Terphenyl-d14	1.178	1.156	1.9	76	0.00
80	Butylbenzylphthalate	0.366	0.347	5.2	73	0.00
81	Benzo(a)anthracene	1.296	1.226	5.4	75	0.01
82	3,3'-Dichlorobenzidine	0.488	0.459	5.9	75	0.00
83	Chrysene	1.256	1.213	3.4	76	0.00
84	Bis(2-ethylhexyl)phthalate	0.559	0.530	5.2	72	0.00
85 c	Di-n-octyl phthalate	0.940	0.897	4.6	74	0.01
86 I	Perylene-d12	1.000	1.000	0.0	79	0.00
87	Indeno(1,2,3-cd)pyrene	1.683	1.540	8.5	75	0.01
88	Benzo(b)fluoranthene	1.333	1.233	7.5	73	0.01
89	Benzo(k)fluoranthene	1.303	1.238	5.0	81	0.01
90 C	Benzo(a)pyrene	1.222	1.141	6.6	76	0.01
91	Dibenzo(a,h)anthracene	1.485	1.371	7.7	75	0.01
92	Benzo(g,h,i)perylene	1.374	1.237	10.0	73	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051021\
Data File : BP005462.D
Acq On : 10 May 2021 09:51
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 10 11:02:54 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 07 12:37:06 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0