

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010259.D
 Acq On : 10 May 2022 10:59
 Operator : CG/JU
 Sample : SST00530
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST005630

Quant Time: May 10 15:18:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.910	152	142096	20.000	ng/ul	0.00
20) Naphthalene-d8	10.716	136	597366	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.545	164	431358	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.298	188	972436	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.380	240	1025338	20.000	ng/ul	0.00
88) Perylene-d12	23.816	264	1020273	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.358	96	6627	2.034	ng/uL	0.00
4) Pyridine-d5	3.781	84	41342	4.510	ng/ul	0.01
7) Phenol-d5	7.075	99	54191	4.193	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.240	67	37590	4.847	ng/ul	0.00
11) 2-Chlorophenol-d4	7.446	132	43289	4.522	ng/ul	0.00
15) 4-Methylphenol-d8	8.622	113	44410	4.147	ng/ul	0.00
21) Nitrobenzene-d5	9.069	128	21184	4.459	ng/ul	0.00
24) 2-Nitrophenol-d4	9.793	143	24266	4.677	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	43742	4.330	ng/ul	0.00
31) 4-Chloroaniline-d4	10.852	131	61123	4.220	ng/ul	0.00
46) Dimethylphthalate-d6	13.951	166	165210	4.924	ng/ul	0.00
49) Acenaphthylene-d8	14.234	160	184925	4.554	ng/ul	0.00
54) 4-Nitrophenol-d4	14.740	143	23091	3.738	ng/ul	0.00
60) Fluorene-d10	15.534	176	137361	4.759	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.657	200	22365	3.637	ng/ul	0.00
73) Anthracene-d10	17.392	188	220218	4.681	ng/ul	-0.01
81) Pyrene-d10	19.622	212	271120	4.650	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.657	264	246983	4.617	ng/ul	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.393	88	6919	2.043	ng/ul#	32
5) Pyridine	3.799	79	40076	4.206	ng/ul#	36
6) Benzaldehyde	7.052	77	38620	5.591	ng/ul	97
8) Phenol	7.105	94	55644	4.285	ng/ul#	91
10) Bis(2-Chloroethyl)ether	7.334	93	45895	4.587	ng/ul#	77
12) 2-Chlorophenol	7.475	128	44097	4.567	ng/ul	94
13) 2-Methylphenol	8.358	108	40797	4.096	ng/ul	98
14) 2,2'-oxybis(1-Chloropr...	8.440	45	75607	5.093	ng/ul#	85
16) Acetophenone	8.734	105	75214	4.502	ng/ul#	74
17) N-Nitroso-di-n-propyla...	8.716	70	39319	4.492	ng/ul#	68
18) 4-Methylphenol	8.681	108	45202	4.122	ng/ul	90
19) Hexachloroethane	8.993	117	19377	4.776	ng/ul#	81
22) Nitrobenzene	9.110	77	55892	4.801	ng/ul#	85
23) Isophorone	9.640	82	110265	4.796	ng/ul#	97
25) 2-Nitrophenol	9.828	139	23888	4.347	ng/ul#	86
26) 2,4-Dimethylphenol	9.887	107	54918	4.616	ng/ul#	77
27) Bis(2-Chloroethoxy)met...	10.122	93	65615	4.869	ng/ul#	96
29) 2,4-Dichlorophenol	10.363	162	43892	4.520	ng/ul#	84
30) Naphthalene	10.763	128	150652	4.818	ng/ul	96
32) 4-Chloroaniline	10.875	127	61556	4.313	ng/ul	99
33) Hexachlorobutadiene	11.057	225	32166	4.674	ng/ul	94
34) Caprolactam	11.651	113	15893	4.550	ng/ul#	1
35) 4-Chloro-3-methylphenol	11.993	107	52901	4.510	ng/ul#	81
36) 2-Methylnaphthalene	12.375	142	105568	4.656	ng/ul	94

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010259.D
 Acq On : 10 May 2022 10:59
 Operator : CG/JU
 Sample : SST00530
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST005630


Quant Time: May 10 15:18:35 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.593	142	109340	4.749	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.740	216	61271	4.598	ng/ul	93
40) Hexachlorocyclopentadiene	12.722	237	24778	3.299	ng/ul	98
41) 2,4,6-Trichlorophenol	12.981	196	38409	4.454	ng/ul	86
42) 2,4,5-Trichlorophenol	13.051	196	41467	4.420	ng/ul#	88
43) 1,1'-Biphenyl	13.381	154	151999	4.778	ng/ul	95
44) 2-Chloronaphthalene	13.416	162	116135	4.717	ng/ul	94
45) 2-Nitroaniline	13.622	65	35375	4.431	ng/ul#	79
47) Dimethylphthalate	13.998	163	160181	4.912	ng/ul#	91
48) 2,6-Dinitrotoluene	14.116	165	28445	4.250	ng/ul	95
50) Acenaphthylene	14.263	152	186283	4.631	ng/ul	94
51) 3-Nitroaniline	14.445	138	26482	4.368	ng/ul#	83
52) Acenaphthene	14.610	153	125247	4.788	ng/ul	98
53) 2,4-Dinitrophenol	14.657	184	12125	3.242	ng/ul#	73
55) 4-Nitrophenol	14.751	109	22333	4.161	ng/ul#	50
56) Dibenzofuran	14.940	168	184557	4.765	ng/ul	97
57) 2,4-Dinitrotoluene	14.904	165	41027	4.196	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.251	232	197	0.024	ng/ul	91
59) Diethylphthalate	15.363	149	166883	5.031	ng/ul	93
61) Fluorene	15.592	166	152703	4.802	ng/ul	92
62) 4-Chlorophenyl-phenyle...	15.587	204	80009	4.781	ng/ul	99
63) 4-Nitroaniline	15.610	138	25560	4.333	ng/ul#	75
66) 4,6-Dinitro-2-methylph...	15.669	198	22365	3.770	ng/ul	94
67) N-Nitrosodiphenylamine	15.798	169	132539	4.750	ng/ul	92
68) 4-Bromophenyl-phenylether	16.481	248	48253	4.652	ng/ul	96
69) Hexachlorobenzene	16.657	284	93	0.008	ng/ul	91
70) Atrazine	16.751	200	50790	4.477	ng/ul#	89
71) Pentachlorophenol	16.945	266	31183	4.067	ng/ul#	81
72) Phenanthrene	17.339	178	253082	4.788	ng/ul	99
74) Anthracene	17.428	178	256615	4.820	ng/ul	97
75) 1,2,3,4-Tetrachloroben...	13.345	216	65445	4.610	ng/ul#	86
76) Pentachlorobenzene	14.869	250	64731	4.650	ng/ul	97
77) Carbazole	17.698	167	216343	4.592	ng/ul#	98
78) Di-n-butylphthalate	18.251	149	292728	5.051	ng/ul#	93
80) Fluoranthene	19.304	202	307521	4.676	ng/ul#	95
82) Pyrene	19.651	202	325899	4.809	ng/ul	96
83) Butylbenzylphthalate	20.522	149	131175	4.683	ng/ul#	82
84) 3,3'-Dichlorobenzidine	21.292	252	95578	5.181	ng/ul#	96
85) Benzo(a)anthracene	21.363	228	321480	4.940	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.286	149	209902	5.056	ng/ul#	81
87) Chrysene	21.416	228	318688	4.974	ng/ul	99
89) Di-n-octyl phthalate	22.221	149	330015	4.161	ng/ul	100
90) Benzo(b)fluoranthene	23.069	252	321592	4.689	ng/ul	99
91) Benzo(k)fluoranthene	23.116	252	287773	4.429	ng/ul#	98
93) Benzo(a)pyrene	23.704	252	294502	4.643	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	26.357	276	283905	4.499	ng/ul#	93
95) Dibenzo(a,h)anthracene	26.374	278	237950	4.407	ng/ul#	95
96) Benzo(g,h,i)perylene	27.139	276	236781	4.447	ng/ul#	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 10 15:18:35 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue May 10 15:17:10 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 2.043 ng/uL
RT: 3.393 min Scan# 51
Delta R.T. 0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion: 88 Resp: 6919
Ion Ratio Lower Upper
88 100
43 44.8 6.8 10.2#
58 70.4 28.4 42.6#



#5
Pyridine
Concen: 4.206 ng/uL
RT: 3.799 min Scan# 121
Delta R.T. 0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion: 79 Resp: 40076
Ion Ratio Lower Upper
79 100
52 92.0 35.5 53.3#
51 49.2 20.2 30.4#









#6
Benzaldehyde
Concen: 5.591 ng/ul
RT: 7.052 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion: 77 Resp: 38620
Ion Ratio Lower Upper
77 100
105 93.5 77.0 115.4
106 92.9 77.0 115.4



#8
Phenol
Concen: 4.285 ng/ul
RT: 7.105 min Scan# 683
Delta R.T. 0.000 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion: 94 Resp: 55644
Ion Ratio Lower Upper
94 100
65 30.9 19.0 28.6#
66 39.4 34.6 51.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 4.587 ng/ul
 RT: 7.334 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BP010259.D
 Acq: 10 May 2022 10:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005630


Tgt Ion: 93 Resp: 45895
 Ion Ratio Lower Upper
 93 100
 63 81.5 45.0 67.4#
 95 35.1 26.3 39.5



#12
 2-Chlorophenol
 Concen: 4.567 ng/ul
 RT: 7.475 min Scan# 746
 Delta R.T. -0.006 min
 Lab File: BP010259.D
 Acq: 10 May 2022 10:59

Tgt Ion:128 Resp: 44097
 Ion Ratio Lower Upper
 128 100
 64 53.8 45.3 67.9
 130 31.3 29.0 43.6





#13

2-Methylphenol

Concen: 4.096 ng/ul

RT: 8.358 min Scan# 896

Delta R.T. 0.000 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Instrument :

BNA_P

ClientSampleId :





SSTD005630

Tgt Ion:108 Resp: 40797

Ion Ratio Lower Upper

108 100

107 90.4 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 5.093 ng/ul

RT: 8.440 min Scan# 910

Delta R.T. -0.006 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59




Tgt Ion: 45 Resp: 75607


Ion Ratio Lower Upper

45 100

77 14.2 17.1 25.7#

79 9.8 13.2 19.8#









#16
Acetophenone
Concen: 4.502 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:105 Resp: 75214
Ion Ratio Lower Upper
105 100
77 85.2 53.4 80.2#
51 33.1 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 4.492 ng/ul
RT: 8.716 min Scan# 957
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion: 70 Resp: 39319
Ion Ratio Lower Upper
70 100
42 67.4 31.8 47.8#
101 10.4 15.6 23.4#
130 23.0 13.7 20.5#









#18
4-Methylphenol
Concen: 4.122 ng/ul
RT: 8.681 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:108 Resp: 45202
Ion Ratio Lower Upper
108 100
107 114.3 83.6 125.4



#19
Hexachloroethane
Concen: 4.776 ng/ul
RT: 8.993 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:117 Resp: 19377
Ion Ratio Lower Upper
117 100
201 103.2 106.5 159.7#
199 68.0 59.0 88.6









#22
Nitrobenzene
Concen: 4.801 ng/ul
RT: 9.110 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion: 77 Resp: 55892
Ion Ratio Lower Upper
77 100
123 42.6 42.3 63.5
65 14.7 5.8 8.8#



#23
Isophorone
Concen: 4.796 ng/ul
RT: 9.640 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion: 82 Resp: 110265
Ion Ratio Lower Upper
82 100
95 7.2 7.8 11.8#
138 15.9 12.9 19.3









#25
2-Nitrophenol
Concen: 4.347 ng/ul
RT: 9.828 min Scan# 1146
Delta R.T. 0.000 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:139 Resp: 23888
Ion Ratio Lower Upper
139 100
65 57.4 49.4 74.2
109 38.7 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 4.616 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:107 Resp: 54918
Ion Ratio Lower Upper
107 100
121 46.0 34.1 51.1
122 84.1 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 4.869 ng/ul
 RT: 10.122 min Scan# 1197
 Delta R.T. -0.006 min
 Lab File: BP010259.D
 Acq: 10 May 2022 10:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005630


Tgt Ion: 93 Resp: 65615
 Ion Ratio Lower Upper
 93 100
 95 30.8 26.2 39.4
 123 10.6 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 4.520 ng/ul
 RT: 10.363 min Scan# 1237
 Delta R.T. 0.000 min
 Lab File: BP010259.D
 Acq: 10 May 2022 10:59

Tgt Ion: 162 Resp: 43892
 Ion Ratio Lower Upper
 162 100
 164 64.3 37.0 55.6#
 98 36.8 31.0 46.6









#30
Naphthalene
Concen: 4.818 ng/ul
RT: 10.763 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:128 Resp: 150652
Ion Ratio Lower Upper
128 100
129 11.6 8.6 13.0
127 12.8 12.2 18.4



#32
4-Chloroaniline
Concen: 4.313 ng/ul
RT: 10.875 min Scan# 1324
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:127 Resp: 61556
Ion Ratio Lower Upper
127 100
129 30.2 24.6 37.0









#33
Hexachlorobutadiene
Concen: 4.674 ng/ul
RT: 11.057 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:225 Resp: 32166
Ion Ratio Lower Upper
225 100
223 60.3 53.4 80.0
227 70.0 53.8 80.8



#34
Caprolactam
Concen: 4.550 ng/ul
RT: 11.651 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:113 Resp: 15893
Ion Ratio Lower Upper
113 100
55 190.2 58.2 87.2#
56 147.6 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 4.510 ng/ul
RT: 11.993 min Scan# 11
Delta R.T. -0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:107 Resp: 52901
Ion Ratio Lower Upper
107 100
144 21.7 15.8 23.8
142 65.9 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 4.656 ng/ul
RT: 12.375 min Scan# 1579
Delta R.T. 0.000 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:142 Resp: 105568
Ion Ratio Lower Upper
142 100
141 86.1 64.3 96.5









#37
1-Methylnaphthalene
Concen: 4.749 ng/ul
RT: 12.593 min Scan# 1616
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:142 Resp: 109340
Ion Ratio Lower Upper
142 100
141 91.9 69.8 104.8
116 3.4 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 4.598 ng/ul
RT: 12.740 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:216 Resp: 61271
Ion Ratio Lower Upper
216 100
214 79.2 58.2 87.2
179 20.8 14.2 21.2
108 17.4 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 3.299 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:237 Resp: 24778
Ion Ratio Lower Upper
237 100
235 65.7 53.2 79.8
272 11.6 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 4.454 ng/ul
RT: 12.981 min Scan# 1682
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:196 Resp: 38409
Ion Ratio Lower Upper
196 100
198 94.0 64.1 96.1
200 29.9 19.9 29.9









#42
2,4,5-Trichlorophenol
Concen: 4.420 ng/ul
RT: 13.051 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:196 Resp: 41467
Ion Ratio Lower Upper
196 100
198 95.4 68.6 103.0
200 31.7 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 4.778 ng/ul
RT: 13.381 min Scan# 1750
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:154 Resp: 151999
Ion Ratio Lower Upper
154 100
153 38.9 28.6 43.0
76 12.0 8.5 12.7









#44
2-Chloronaphthalene
Concen: 4.717 ng/ul
RT: 13.416 min Scan# 1756
Delta R.T. -0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument : BNA_P
ClientSampleId : SSTD005630


Tgt Ion:162 Resp: 116135
Ion Ratio Lower Upper
162 100
164 32.7 27.5 41.3
127 40.4 37.1 55.7



#45
2-Nitroaniline
Concen: 4.431 ng/ul
RT: 13.622 min Scan# 1791
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion: 65 Resp: 35375
Ion Ratio Lower Upper
65 100
92 66.8 77.9 116.9#
138 90.4 81.5 122.3









#47
Dimethylphthalate
Concen: 4.912 ng/ul
RT: 13.998 min Scan# 1855
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:163 Resp: 160181
Ion Ratio Lower Upper
163 100
194 4.5 4.1 6.1
164 9.6 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 4.250 ng/ul
RT: 14.116 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:165 Resp: 28445
Ion Ratio Lower Upper
165 100
89 60.3 51.7 77.5
121 25.6 18.3 27.5









#50
Acenaphthylene
Concen: 4.631 ng/ul
RT: 14.263 min Scan# 1900
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:152 Resp: 186283
Ion Ratio Lower Upper
152 100
151 19.1 18.3 27.5
153 13.1 11.0 16.6



#51
3-Nitroaniline
Concen: 4.368 ng/ul
RT: 14.445 min Scan# 1931
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:138 Resp: 26482
Ion Ratio Lower Upper
138 100
108 11.9 15.8 23.8#
92 125.8 117.1 175.7





#52

Acenaphthene

Concen: 4.788 ng/ul

RT: 14.610 min Scan# 1959

Delta R.T. -0.006 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Instrument :

BNA_P

ClientSampleId :

SSTD005630





Tgt Ion:153 Resp: 125247

Ion Ratio Lower Upper

153 100

152 46.1 34.3 51.5

154 86.7 70.0 105.0



#53

2,4-Dinitrophenol

Concen: 3.242 ng/ul

RT: 14.657 min Scan# 1967

Delta R.T. 0.000 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59




Tgt Ion:184 Resp: 12125


Ion Ratio Lower Upper

184 100

63 82.5 34.6 52.0#

154 67.0 57.7 86.5









#55
4-Nitrophenol
Concen: 4.161 ng/ul
RT: 14.751 min Scan# 1983
Delta R.T. -0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:109 Resp: 22333
Ion Ratio Lower Upper
109 100
139 106.5 61.4 92.2#
65 138.9 65.3 97.9#



#56
Dibenzofuran
Concen: 4.765 ng/ul
RT: 14.940 min Scan# 2015
Delta R.T. -0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:168 Resp: 184557
Ion Ratio Lower Upper
168 100
139 39.3 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 4.196 ng/ul
RT: 14.904 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:165 Resp: 41027
Ion Ratio Lower Upper
165 100
63 49.9 29.5 44.3#
182 3.2 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 0.024 ng/ul
RT: 15.251 min Scan# 2068
Delta R.T. 0.077 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:232 Resp: 197
Ion Ratio Lower Upper
232 100
131 35.3 25.4 38.2
130 4.7 4.2 6.2
166 28.2 28.1 42.1









#59
Diethylphthalate
Concen: 5.031 ng/ul
RT: 15.363 min Scan# 2088
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:149 Resp: 166883
Ion Ratio Lower Upper
149 100
177 22.5 22.0 33.0
150 13.5 10.2 15.2



#61
Fluorene
Concen: 4.802 ng/ul
RT: 15.592 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:166 Resp: 152703
Ion Ratio Lower Upper
166 100
165 95.8 70.3 105.5
167 14.2 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 4.781 ng/ul
RT: 15.587 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:204 Resp: 80009
Ion Ratio Lower Upper
204 100
206 34.6 27.6 41.4
141 59.4 46.6 69.8



#63
4-Nitroaniline
Concen: 4.333 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:138 Resp: 25560
Ion Ratio Lower Upper
138 100
92 51.5 41.6 62.4
108 81.7 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 3.770 ng/ul
RT: 15.669 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:198 Resp: 22365
Ion Ratio Lower Upper
198 100
182 7.6 6.0 9.0
77 26.1 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 4.750 ng/ul
RT: 15.798 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:169 Resp: 132539
Ion Ratio Lower Upper
169 100
168 67.3 48.3 72.5
167 33.9 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 4.652 ng/ul
RT: 16.481 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:248	Resp:	48253
Ion Ratio	Lower	Upper
248	100	
250	96.6	79.4 119.0
141	78.5	67.2 100.8



#69
Hexachlorobenzene
Concen: 0.008 ng/ul
RT: 16.657 min Scan# 2307
Delta R.T. 0.047 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:284	Resp:	93
Ion Ratio	Lower	Upper
284	100	
142	19.4	19.2 28.8
249	39.5	27.4 41.2





#70

Atrazine

Concen: 4.477 ng/ul

RT: 16.751 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Instrument :

BNA_P

ClientSampleId :

SSTD005630





Tgt Ion:200 Resp: 50790

Ion Ratio Lower Upper

200 100

173 27.0 19.2 28.8

215 49.4 32.2 48.4#



#71

Pentachlorophenol

Concen: 4.067 ng/ul

RT: 16.945 min Scan# 2356

Delta R.T. -0.012 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59




Tgt Ion:266 Resp: 31183


Ion Ratio Lower Upper

266 100

264 58.2 57.3 85.9

268 62.3 65.7 98.5#





#72

Phenanthrene

Concen: 4.788 ng/u1

RT: 17.339 min Scan# 2423

Delta R.T. -0.006 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Instrument :

BNA_P

ClientSampleId :

SSTD005630





Tgt Ion:178 Resp: 253082

Ion Ratio Lower Upper

178 100

179 15.4 11.7 17.5

176 18.3 14.5 21.7



#74

Anthracene

Concen: 4.820 ng/u1

RT: 17.428 min Scan# 2438

Delta R.T. -0.012 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59




Tgt Ion:178 Resp: 256615


Ion Ratio Lower Upper

178 100

179 14.5 12.3 18.5

176 18.2 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.610 ng/uL
RT: 13.345 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:216 Resp: 65445
Ion Ratio Lower Upper
216 100
214 76.7 50.4 75.6#
218 49.4 35.2 52.8



#76
Pentachlorobenzene
Concen: 4.650 ng/uL
RT: 14.869 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:250 Resp: 64731
Ion Ratio Lower Upper
250 100
248 58.6 43.5 65.3
252 58.7 46.5 69.7









#77
 Carbazole
 Concen: 4.592 ng/ul
 RT: 17.698 min Scan# 2484
 Delta R.T. -0.006 min
 Lab File: BP010259.D
 Acq: 10 May 2022 10:59




Instrument :
 BNA_P
 ClientSampleId :
 SSTD005630


Tgt Ion:167 Resp: 216343
 Ion Ratio Lower Upper
 167 100
 166 20.1 15.8 23.8
 139 12.7 8.3 12.5#



#78
 Di-n-butylphthalate
 Concen: 5.051 ng/ul
 RT: 18.251 min Scan# 2578
 Delta R.T. -0.006 min
 Lab File: BP010259.D
 Acq: 10 May 2022 10:59

Tgt Ion:149 Resp: 292728
 Ion Ratio Lower Upper
 149 100
 150 9.5 9.6 14.4#
 104 5.4 6.2 9.4#









#80
Fluoranthene
Concen: 4.676 ng/ul
RT: 19.304 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:202 Resp: 307521
Ion Ratio Lower Upper
202 100
101 9.3 9.6 14.4#
100 7.6 6.3 9.5



#82
Pyrene
Concen: 4.809 ng/ul
RT: 19.651 min Scan# 2816
Delta R.T. -0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:202 Resp: 325899
Ion Ratio Lower Upper
202 100
101 11.6 10.6 15.8
100 11.0 9.8 14.6





#83

Butylbenzylphthalate

Concen: 4.683 ng/ul

RT: 20.522 min Scan# 2964

Delta R.T. -0.012 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Instrument :

BNA_P

ClientSampleId :

SSTD005630





Tgt Ion:149 Resp: 131175

Ion Ratio Lower Upper

149 100

91 74.0 71.1 106.7

206 21.5 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 5.181 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59




Tgt Ion:252 Resp: 95578


Ion Ratio Lower Upper

252 100

254 64.6 49.8 74.6

126 10.8 4.3 6.5#





#85

Benzo(a)anthracene

Concen: 4.940 ng/ul

RT: 21.363 min Scan# 3108

Delta R.T. -0.006 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Instrument :

BNA_P

ClientSampleId :

SSTD005630





Tgt Ion:228 Resp: 321480

Ion Ratio Lower Upper

228 100

229 20.3 18.2 27.2

226 26.1 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 5.056 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59




Tgt Ion:149 Resp: 209902


Ion Ratio Lower Upper

149 100

167 27.7 32.8 49.2#

279 4.4 5.6 8.4#









#87
Chrysene
Concen: 4.974 ng/ul
RT: 21.416 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:228 Resp: 318688
Ion Ratio Lower Upper
228 100
226 28.7 22.7 34.1
229 20.1 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 4.161 ng/ul
RT: 22.221 min Scan# 3253
Delta R.T. -0.006 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:149 Resp: 330015









#90
Benzo(b)fluoranthene
Concen: 4.689 ng/ul
RT: 23.069 min Scan# 31
Delta R.T. -0.012 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:252 Resp: 321592
Ion Ratio Lower Upper
252 100
253 22.2 17.4 26.0
125 9.4 7.3 10.9



#91
Benzo(k)fluoranthene
Concen: 4.429 ng/ul
RT: 23.116 min Scan# 3405
Delta R.T. -0.018 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:252 Resp: 287773
Ion Ratio Lower Upper
252 100
253 21.8 17.9 26.9
125 8.6 5.1 7.7#









#93
Benzo(a)pyrene
Concen: 4.643 ng/ul
RT: 23.704 min Scan# 31
Delta R.T. -0.018 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59




Instrument :
BNA_P
ClientSampleId :
SSTD005630


Tgt Ion:252 Resp: 294502
Ion Ratio Lower Upper
252 100
253 22.0 17.5 26.3
125 11.7 6.4 9.6#



#94
Indeno(1,2,3-cd)pyrene
Concen: 4.499 ng/ul
RT: 26.357 min Scan# 3956
Delta R.T. -0.023 min
Lab File: BP010259.D
Acq: 10 May 2022 10:59

Tgt Ion:276 Resp: 283905
Ion Ratio Lower Upper
276 100
138 20.3 12.3 18.5#
277 26.2 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 4.407 ng/ul

RT: 26.374 min Scan# 3963

Delta R.T. -0.023 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Instrument :

BNA_P

ClientSampleId :

SSTD005630





Tgt Ion:278 Resp: 237950

Ion Ratio Lower Upper

278 100

139 13.9 6.6 10.0#

279 22.4 18.4 27.6



#96

Benzo(g,h,i)perylene

Concen: 4.447 ng/ul

RT: 27.139 min Scan# 4089

Delta R.T. -0.023 min

Lab File: BP010259.D

Acq: 10 May 2022 10:59

Tgt Ion:276 Resp: 236781

Ion Ratio Lower Upper

276 100

138 19.5 11.6 17.4#

277 23.5 15.9 23.9

