

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010261.D
 Acq On : 10 May 2022 12:06
 Operator : CG/JU
 Sample : SST02032
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020632

Quant Time: May 10 15:20:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.910	152	137475	20.000	ng/ul	0.00
20) Naphthalene-d8	10.710	136	619406	20.000	ng/ul	-0.01
38) Acenaphthene-d10	14.545	164	431191	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.298	188	968986	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.380	240	995143	20.000	ng/ul	# 0.00
88) Perylene-d12	23.815	264	989016	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	43713	13.866	ng/uL	0.00
4) Pyridine-d5	3.769	84	292804	33.015	ng/ul	0.00
7) Phenol-d5	7.075	99	418118	33.437	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.240	67	263465	35.113	ng/ul	0.00
11) 2-Chlorophenol-d4	7.440	132	319659	34.517	ng/ul	0.00
15) 4-Methylphenol-d8	8.622	113	343797	33.184	ng/ul	0.00
21) Nitrobenzene-d5	9.069	128	168503	34.204	ng/ul	0.00
24) 2-Nitrophenol-d4	9.793	143	192684	35.813	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.334	165	350443	33.458	ng/ul	0.00
31) 4-Chloroaniline-d4	10.845	131	511466	34.056	ng/ul	-0.01
46) Dimethylphthalate-d6	13.951	166	1132611	33.767	ng/ul	0.00
49) Acenaphthylene-d8	14.233	160	1369503	33.736	ng/ul	0.00
54) 4-Nitrophenol-d4	14.739	143	212055	34.341	ng/ul	0.00
60) Fluorene-d10	15.533	176	952532	33.013	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.657	200	210499	34.356	ng/ul	0.00
73) Anthracene-d10	17.398	188	1548664	33.036	ng/ul	0.00
81) Pyrene-d10	19.627	212	1850952	32.706	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.662	264	1767016	34.080	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.387	88	44845	13.689	ng/uL#	21
5) Pyridine	3.787	79	309471	33.568	ng/ul#	36
6) Benzaldehyde	7.046	77	181097	27.100	ng/ul	94
8) Phenol	7.099	94	422657	33.639	ng/ul#	93
10) Bis(2-Chloroethyl)ether	7.328	93	332410	34.339	ng/ul#	75
12) 2-Chlorophenol	7.475	128	319068	34.157	ng/ul	93
13) 2-Methylphenol	8.351	108	323244	33.545	ng/ul	97
14) 2,2'-oxybis(1-Chloropr...	8.446	45	539398	37.553	ng/ul#	82
16) Acetophenone	8.734	105	543017	33.594	ng/ul#	80
17) N-Nitroso-di-n-propyla...	8.722	70	294499	34.774	ng/ul#	70
18) 4-Methylphenol	8.681	108	353226	33.293	ng/ul	90
19) Hexachloroethane	8.993	117	135322	34.476	ng/ul#	80
22) Nitrobenzene	9.110	77	418123	34.641	ng/ul#	83
23) Isophorone	9.640	82	825658	34.633	ng/ul#	98
25) 2-Nitrophenol	9.828	139	196509	34.485	ng/ul#	85
26) 2,4-Dimethylphenol	9.887	107	412482	33.434	ng/ul#	77
27) Bis(2-Chloroethoxy)met...	10.122	93	481728	34.475	ng/ul#	97
29) 2,4-Dichlorophenol	10.363	162	335979	33.369	ng/ul#	85
30) Naphthalene	10.763	128	1076299	33.199	ng/ul	97
32) 4-Chloroaniline	10.869	127	502313	33.941	ng/ul	99
33) Hexachlorobutadiene	11.057	225	226002	31.671	ng/ul	95
34) Caprolactam	11.645	113	71206	19.661	ng/ul#	8
35) 4-Chloro-3-methylphenol	11.998	107	404515	33.258	ng/ul#	70
36) 2-Methylnaphthalene	12.369	142	769695	32.736	ng/ul	92

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010261.D
 Acq On : 10 May 2022 12:06
 Operator : CG/JU
 Sample : SSTD02032
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020632

Quant Time: May 10 15:20:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.592	142	776999	32.547	ng/ul#	95
39) 1,2,4,5-Tetrachloroben...	12.739	216	435901	32.721	ng/ul	92
40) Hexachlorocyclopentadiene	12.722	237	239999	31.968	ng/ul	95
41) 2,4,6-Trichlorophenol	12.981	196	295578	34.292	ng/ul#	82
42) 2,4,5-Trichlorophenol	13.051	196	320598	34.186	ng/ul#	88
43) 1,1'-Biphenyl	13.381	154	1068128	33.586	ng/ul	93
44) 2-Chloronaphthalene	13.422	162	819865	33.313	ng/ul	93
45) 2-Nitroaniline	13.622	65	291345	36.507	ng/ul#	80
47) Dimethylphthalate	13.998	163	1085650	33.304	ng/ul#	93
48) 2,6-Dinitrotoluene	14.116	165	234897	35.113	ng/ul	96
50) Acenaphthylene	14.263	152	1358554	33.786	ng/ul	95
51) 3-Nitroaniline	14.445	138	163219	26.933	ng/ul#	83
52) Acenaphthene	14.610	153	871858	33.345	ng/ul	97
53) 2,4-Dinitrophenol	14.651	184	141244	37.783	ng/ul#	77
55) 4-Nitrophenol	14.757	109	178669	33.304	ng/ul#	45
56) Dibenzofuran	14.939	168	1272973	32.879	ng/ul	99
57) 2,4-Dinitrotoluene	14.904	165	334357	34.212	ng/ul#	80
58) 2,3,4,6-Tetrachlorophenol	15.169	232	282529	33.891	ng/ul#	86
59) Diethylphthalate	15.363	149	1126216	33.966	ng/ul	93
61) Fluorene	15.592	166	1051226	33.070	ng/ul	90
62) 4-Chlorophenyl-phenyle...	15.586	204	546608	32.673	ng/ul	98
63) 4-Nitroaniline	15.610	138	147705	25.048	ng/ul#	76
66) 4,6-Dinitro-2-methylph...	15.675	198	208327	35.242	ng/ul#	93
67) N-Nitrosodiphenylamine	15.798	169	924659	33.258	ng/ul	96
68) 4-Bromophenyl-phenylether	16.480	248	337881	32.691	ng/ul	96
69) Hexachlorobenzene	16.604	284	379689	31.649	ng/ul#	91
70) Atrazine	16.757	200	369193	32.662	ng/ul#	88
71) Pentachlorophenol	16.945	266	249804	32.696	ng/ul#	86
72) Phenanthrene	17.339	178	1723285	32.720	ng/ul	99
74) Anthracene	17.433	178	1749529	32.980	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.345	216	465222	32.887	ng/uL#	88
76) Pentachlorobenzene	14.869	250	439694	31.696	ng/uL	92
77) Carbazole	17.698	167	1566314	33.367	ng/ul#	96
78) Di-n-butylphthalate	18.251	149	1984480	34.362	ng/ul#	92
80) Fluoranthene	19.304	202	2126819	33.318	ng/ul	97
82) Pyrene	19.657	202	2166745	32.946	ng/ul	96
83) Butylbenzylphthalate	20.521	149	959691	35.302	ng/ul#	81
84) 3,3'-Dichlorobenzidine	21.292	252	648562	36.221	ng/ul#	98
85) Benzo(a)anthracene	21.363	228	2127469	33.687	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.286	149	1412001	35.045	ng/ul#	81
87) Chrysene	21.415	228	2060770	33.139	ng/ul	99
89) Di-n-octyl phthalate	22.221	149	2401381	31.233	ng/ul	100
90) Benzo(b)fluoranthene	23.074	252	2205247	33.167	ng/ul	100
91) Benzo(k)fluoranthene	23.121	252	2075655	32.957	ng/ul#	98
93) Benzo(a)pyrene	23.709	252	2088570	33.969	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.368	276	2142500	35.029	ng/ul#	91
95) Dibenzo(a,h)anthracene	26.380	278	1832587	35.015	ng/ul#	94
96) Benzo(g,h,i)perylene	27.145	276	1769633	34.289	ng/ul#	91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010261.D
 Acq On : 10 May 2022 12:06
 Operator : CG/JU
 Sample : SST020632
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020632

Quant Time: May 10 15:20:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 13.689 ng/uL
RT: 3.387 min Scan# 51
Delta R.T. -0.000 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion: 88 Resp: 44845
Ion Ratio Lower Upper
88 100
43 35.2 6.8 10.2#
58 82.2 28.4 42.6#



#5
Pyridine
Concen: 33.568 ng/uL
RT: 3.787 min Scan# 119
Delta R.T. -0.000 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion: 79 Resp: 309471
Ion Ratio Lower Upper
79 100
52 96.8 35.5 53.3#
51 43.4 20.2 30.4#









#6
Benzaldehyde
Concen: 27.100 ng/ul
RT: 7.046 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion: 77 Resp: 181097
Ion Ratio Lower Upper
77 100
105 92.3 77.0 115.4
106 88.7 77.0 115.4



#8
Phenol
Concen: 33.639 ng/ul
RT: 7.099 min Scan# 682
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion: 94 Resp: 422657
Ion Ratio Lower Upper
94 100
65 32.2 19.0 28.6#
66 44.4 34.6 51.8









#10
Bis(2-Chloroethyl)ether
Concen: 34.339 ng/ul
RT: 7.328 min Scan# 71
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion: 93 Resp: 332410
Ion Ratio Lower Upper
93 100
63 85.5 45.0 67.4#
95 32.6 26.3 39.5



#12
2-Chlorophenol
Concen: 34.157 ng/ul
RT: 7.475 min Scan# 746
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:128 Resp: 319068
Ion Ratio Lower Upper
128 100
64 52.7 45.3 67.9
130 31.3 29.0 43.6





#13

2-Methylphenol

Concen: 33.545 ng/ul

RT: 8.351 min Scan# 896

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06

Instrument :

BNA_P

ClientSampleId :





SSTD020632

Tgt Ion:108 Resp: 323244

Ion Ratio Lower Upper

108 100

107 91.7 70.8 106.2



#14

2,2'-oxybis(1-Chloropropane)

Concen: 37.553 ng/ul

RT: 8.446 min Scan# 911

Delta R.T. -0.000 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06




Tgt Ion: 45 Resp: 539398


Ion Ratio Lower Upper

45 100

77 12.5 17.1 25.7#

79 9.8 13.2 19.8#









#16
Acetophenone
Concen: 33.594 ng/ul
RT: 8.734 min Scan# 90
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:105 Resp: 543017
Ion Ratio Lower Upper
105 100
77 80.2 53.4 80.2#
51 32.0 13.8 20.8#



#17
N-Nitroso-di-n-propylamine
Concen: 34.774 ng/ul
RT: 8.722 min Scan# 958
Delta R.T. -0.000 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion: 70 Resp: 294499
Ion Ratio Lower Upper
70 100
42 66.1 31.8 47.8#
101 10.0 15.6 23.4#
130 21.1 13.7 20.5#









#18
4-Methylphenol
Concen: 33.293 ng/ul
RT: 8.681 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:108 Resp: 353226
Ion Ratio Lower Upper
108 100
107 114.8 83.6 125.4



#19
Hexachloroethane
Concen: 34.476 ng/ul
RT: 8.993 min Scan# 1004
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:117 Resp: 135322
Ion Ratio Lower Upper
117 100
201 102.4 106.5 159.7#
199 64.9 59.0 88.6









#22
Nitrobenzene
Concen: 34.641 ng/ul
RT: 9.110 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion: 77 Resp: 418123
Ion Ratio Lower Upper
77 100
123 41.3 42.3 63.5#
65 14.5 5.8 8.8#



#23
Isophorone
Concen: 34.633 ng/ul
RT: 9.640 min Scan# 1114
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion: 82 Resp: 825658
Ion Ratio Lower Upper
82 100
95 7.7 7.8 11.8#
138 16.0 12.9 19.3









#25
2-Nitrophenol
Concen: 34.485 ng/ul
RT: 9.828 min Scan# 1146
Delta R.T. -0.000 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:139 Resp: 196509
Ion Ratio Lower Upper
139 100
65 60.3 49.4 74.2
109 34.6 44.7 67.1#



#26
2,4-Dimethylphenol
Concen: 33.434 ng/ul
RT: 9.887 min Scan# 1156
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:107 Resp: 412482
Ion Ratio Lower Upper
107 100
121 49.5 34.1 51.1
122 81.3 46.3 69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 34.475 ng/ul
 RT: 10.122 min Scan# 1196
 Delta R.T. -0.006 min
 Lab File: BP010261.D
 Acq: 10 May 2022 12:06




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020632


Tgt Ion: 93 Resp: 481728
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.2 39.4
 123 10.5 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 33.369 ng/ul
 RT: 10.363 min Scan# 1237
 Delta R.T. -0.000 min
 Lab File: BP010261.D
 Acq: 10 May 2022 12:06

Tgt Ion:162 Resp: 335979
 Ion Ratio Lower Upper
 162 100
 164 64.3 37.0 55.6#
 98 38.3 31.0 46.6









#30
Naphthalene
Concen: 33.199 ng/ul
RT: 10.763 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:128 Resp: 1076299
Ion Ratio Lower Upper
128 100
129 10.7 8.6 13.0
127 13.3 12.2 18.4



#32
4-Chloroaniline
Concen: 33.941 ng/ul
RT: 10.869 min Scan# 1323
Delta R.T. -0.012 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:127 Resp: 502313
Ion Ratio Lower Upper
127 100
129 31.3 24.6 37.0









#33
Hexachlorobutadiene
Concen: 31.671 ng/ul
RT: 11.057 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:225 Resp: 226002
Ion Ratio Lower Upper
225 100
223 62.0 53.4 80.0
227 64.5 53.8 80.8



#34
Caprolactam
Concen: 19.661 ng/ul
RT: 11.645 min Scan# 1455
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:113 Resp: 71206
Ion Ratio Lower Upper
113 100
55 182.4 58.2 87.2#
56 139.7 68.6 103.0#









#35
4-Chloro-3-methylphenol
Concen: 33.258 ng/ul
RT: 11.998 min Scan# 1515
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:107 Resp: 404515
Ion Ratio Lower Upper
107 100
144 24.0 15.8 23.8#
142 74.6 38.9 58.3#



#36
2-Methylnaphthalene
Concen: 32.736 ng/ul
RT: 12.369 min Scan# 1578
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:142 Resp: 769695
Ion Ratio Lower Upper
142 100
141 87.7 64.3 96.5









#37
1-Methylnaphthalene
Concen: 32.547 ng/ul
RT: 12.592 min Scan# 1616
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:142 Resp: 776999
Ion Ratio Lower Upper
142 100
141 92.0 69.8 104.8
116 4.1 5.6 8.4#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.721 ng/ul
RT: 12.739 min Scan# 1641
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:216 Resp: 435901
Ion Ratio Lower Upper
216 100
214 80.2 58.2 87.2
179 21.0 14.2 21.2
108 16.9 15.3 22.9









#40
Hexachlorocyclopentadiene
Concen: 31.968 ng/ul
RT: 12.722 min Scan# 1638
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:237 Resp: 239999
Ion Ratio Lower Upper
237 100
235 61.9 53.2 79.8
272 13.1 10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 34.292 ng/ul
RT: 12.981 min Scan# 1682
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:196 Resp: 295578
Ion Ratio Lower Upper
196 100
198 97.2 64.1 96.1#
200 31.5 19.9 29.9#









#42
2,4,5-Trichlorophenol
Concen: 34.186 ng/ul
RT: 13.051 min Scan# 1
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:196 Resp: 320598
Ion Ratio Lower Upper
196 100
198 95.3 68.6 103.0
200 30.6 32.8 49.2#



#43
1,1'-Biphenyl
Concen: 33.586 ng/ul
RT: 13.381 min Scan# 1750
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:154 Resp: 1068128
Ion Ratio Lower Upper
154 100
153 40.2 28.6 43.0
76 12.1 8.5 12.7





#44

2-Chloronaphthalene

Concen: 33.313 ng/ul

RT: 13.422 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06

Instrument :

BNA_P

ClientSampleId :

SSTD020632





Tgt Ion:162 Resp: 819865

Ion Ratio Lower Upper

162 100

164 32.9 27.5 41.3

127 39.1 37.1 55.7



#45

2-Nitroaniline

Concen: 36.507 ng/ul

RT: 13.622 min Scan# 1791

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06




Tgt Ion: 65 Resp: 291345


Ion Ratio Lower Upper

65 100

92 67.2 77.9 116.9#

138 92.7 81.5 122.3









#47
Dimethylphthalate
Concen: 33.304 ng/ul
RT: 13.998 min Scan# 1855
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:163 Resp: 1085650
Ion Ratio Lower Upper
163 100
194 4.3 4.1 6.1
164 10.3 11.3 16.9#



#48
2,6-Dinitrotoluene
Concen: 35.113 ng/ul
RT: 14.116 min Scan# 1875
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:165 Resp: 234897
Ion Ratio Lower Upper
165 100
89 61.2 51.7 77.5
121 21.2 18.3 27.5









#50
Acenaphthylene
Concen: 33.786 ng/ul
RT: 14.263 min Scan# 1900
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:152 Resp: 1358554
Ion Ratio Lower Upper
152 100
151 19.5 18.3 27.5
153 13.1 11.0 16.6



#51
3-Nitroaniline
Concen: 26.933 ng/ul
RT: 14.445 min Scan# 1931
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:138 Resp: 163219
Ion Ratio Lower Upper
138 100
108 11.5 15.8 23.8#
92 125.4 117.1 175.7









#52
Acenaphthene
Concen: 33.345 ng/ul
RT: 14.610 min Scan# 1959
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:153 Resp: 871858
Ion Ratio Lower Upper
153 100
152 46.6 34.3 51.5
154 88.5 70.0 105.0



#53
2,4-Dinitrophenol
Concen: 37.783 ng/ul
RT: 14.651 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:184 Resp: 141244
Ion Ratio Lower Upper
184 100
63 76.0 34.6 52.0#
154 66.3 57.7 86.5









#55
4-Nitrophenol
Concen: 33.304 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:109 Resp: 178669
Ion Ratio Lower Upper
109 100
139 118.9 61.4 92.2#
65 136.5 65.3 97.9#



#56
Dibenzofuran
Concen: 32.879 ng/ul
RT: 14.939 min Scan# 2015
Delta R.T. -0.012 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:168 Resp: 1272973
Ion Ratio Lower Upper
168 100
139 38.1 30.1 45.1









#57
2,4-Dinitrotoluene
Concen: 34.212 ng/ul
RT: 14.904 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:165 Resp: 334357
Ion Ratio Lower Upper
165 100
63 49.9 29.5 44.3#
182 3.2 2.8 4.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 33.891 ng/ul
RT: 15.169 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:232 Resp: 282529
Ion Ratio Lower Upper
232 100
131 44.2 25.4 38.2#
130 2.3 4.2 6.2#
166 30.8 28.1 42.1





#59

Diethylphthalate

Concen: 33.966 ng/ul

RT: 15.363 min Scan# 2088

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06

Instrument :

BNA_P

ClientSampleId :

SSTD020632





Tgt Ion:149 Resp: 1126216

Ion Ratio Lower Upper

149 100

177 22.1 22.0 33.0

150 12.8 10.2 15.2



#61

Fluorene

Concen: 33.070 ng/ul

RT: 15.592 min Scan# 2126

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06




Tgt Ion:166 Resp: 1051226


Ion Ratio Lower Upper

166 100

165 98.5 70.3 105.5

167 13.3 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 32.673 ng/ul
RT: 15.586 min Scan# 2125
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:204 Resp: 546608
Ion Ratio Lower Upper
204 100
206 32.8 27.6 41.4
141 58.7 46.6 69.8



#63
4-Nitroaniline
Concen: 25.048 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:138 Resp: 147705
Ion Ratio Lower Upper
138 100
92 54.5 41.6 62.4
108 85.0 97.6 146.4#









#66
4,6-Dinitro-2-methylphenol
Concen: 35.242 ng/ul
RT: 15.675 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:198 Resp: 208327
Ion Ratio Lower Upper
198 100
182 3.9 6.0 9.0#
77 24.9 17.7 26.5



#67
N-Nitrosodiphenylamine
Concen: 33.258 ng/ul
RT: 15.798 min Scan# 2161
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:169 Resp: 924659
Ion Ratio Lower Upper
169 100
168 64.3 48.3 72.5
167 35.0 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 32.691 ng/ul
RT: 16.480 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:248 Resp: 337881
Ion Ratio Lower Upper
248 100
250 97.3 79.4 119.0
141 77.4 67.2 100.8



#69
Hexachlorobenzene
Concen: 31.649 ng/ul
RT: 16.604 min Scan# 2298
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:284 Resp: 379689
Ion Ratio Lower Upper
284 100
142 33.8 19.2 28.8#
249 33.1 27.4 41.2









#70
Atrazine
Concen: 32.662 ng/ul
RT: 16.757 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:200 Resp: 369193
Ion Ratio Lower Upper
200 100
173 26.3 19.2 28.8
215 50.6 32.2 48.4#



#71
Pentachlorophenol
Concen: 32.696 ng/ul
RT: 16.945 min Scan# 2356
Delta R.T. -0.012 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:266 Resp: 249804
Ion Ratio Lower Upper
266 100
264 63.6 57.3 85.9
268 65.3 65.7 98.5#









#72
Phenanthrene
Concen: 32.720 ng/u1
RT: 17.339 min Scan# 2424
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:178 Resp: 1723285
Ion Ratio Lower Upper
178 100
179 15.3 11.7 17.5
176 18.5 14.5 21.7



#74
Anthracene
Concen: 32.980 ng/u1
RT: 17.433 min Scan# 2439
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:178 Resp: 1749529
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 17.7 13.4 20.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 32.887 ng/uL
RT: 13.345 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:216 Resp: 465222
Ion Ratio Lower Upper
216 100
214 76.5 50.4 75.6#
218 47.2 35.2 52.8



#76
Pentachlorobenzene
Concen: 31.696 ng/uL
RT: 14.869 min Scan# 2003
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:250 Resp: 439694
Ion Ratio Lower Upper
250 100
248 62.4 43.5 65.3
252 62.2 46.5 69.7









#77
Carbazole
Concen: 33.367 ng/ul
RT: 17.698 min Scan# 2484
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:167 Resp: 1566314
Ion Ratio Lower Upper
167 100
166 21.2 15.8 23.8
139 12.9 8.3 12.5#



#78
Di-n-butylphthalate
Concen: 34.362 ng/ul
RT: 18.251 min Scan# 2578
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:149 Resp: 1984480
Ion Ratio Lower Upper
149 100
150 8.9 9.6 14.4#
104 5.3 6.2 9.4#









#80
Fluoranthene
Concen: 33.318 ng/ul
RT: 19.304 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:202 Resp: 2126819
Ion Ratio Lower Upper
202 100
101 10.0 9.6 14.4
100 8.1 6.3 9.5



#82
Pyrene
Concen: 32.946 ng/ul
RT: 19.657 min Scan# 2817
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:202 Resp: 2166745
Ion Ratio Lower Upper
202 100
101 12.0 10.6 15.8
100 10.2 9.8 14.6





#83

Butylbenzylphthalate

Concen: 35.302 ng/ul

RT: 20.521 min Scan# 2964

Delta R.T. -0.012 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06

Instrument :

BNA_P

ClientSampleId :

SSTD020632





Tgt Ion:149 Resp: 959691

Ion Ratio Lower Upper

149 100

91 74.0 71.1 106.7

206 20.9 28.8 43.2#



#84

3,3'-Dichlorobenzidine

Concen: 36.221 ng/ul

RT: 21.292 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06




Tgt Ion:252 Resp: 648562


Ion Ratio Lower Upper

252 100

254 62.5 49.8 74.6

126 10.4 4.3 6.5#





#85

Benzo(a)anthracene

Concen: 33.687 ng/ul

RT: 21.363 min Scan# 3108

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06

Instrument :

BNA_P

ClientSampleId :

SSTD020632





Tgt Ion:228 Resp: 2127469

Ion Ratio Lower Upper

228 100

229 19.6 18.2 27.2

226 26.5 22.8 34.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 35.045 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06




Tgt Ion:149 Resp: 1412001


Ion Ratio Lower Upper

149 100

167 27.8 32.8 49.2#

279 4.9 5.6 8.4#









#87
Chrysene
Concen: 33.139 ng/ul
RT: 21.415 min Scan# 3116
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06




Instrument :
BNA_P
ClientSampleId :
SSTD020632


Tgt Ion:228 Resp: 2060770
Ion Ratio Lower Upper
228 100
226 28.9 22.7 34.1
229 19.7 15.2 22.8



#89
Di-n-octyl phthalate
Concen: 31.233 ng/ul
RT: 22.221 min Scan# 3253
Delta R.T. -0.006 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:149 Resp: 2401381





#90

Benzo(b)fluoranthene

Concen: 33.167 ng/ul

RT: 23.074 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06

Instrument :

BNA_P

ClientSampleId :

SSTD020632





Tgt Ion:252 Resp: 2205247

Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 9.1 7.3 10.9



#91

Benzo(k)fluoranthene

Concen: 32.957 ng/ul

RT: 23.121 min Scan# 3406

Delta R.T. -0.012 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06




Tgt Ion:252 Resp: 2075655


Ion Ratio Lower Upper

252 100

253 21.9 17.9 26.9

125 8.9 5.1 7.7#





#93

Benzo(a)pyrene

Concen: 33.969 ng/ul

RT: 23.709 min Scan# 31

Delta R.T. -0.012 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06

Instrument :

BNA_P

ClientSampleId :

SSTD020632





Tgt Ion:252 Resp: 2088570

Ion Ratio Lower Upper

252 100

253 22.4 17.5 26.3

125 9.6 6.4 9.6#



#94

Indeno(1,2,3-cd)pyrene

Concen: 35.029 ng/ul

RT: 26.368 min Scan# 3958

Delta R.T. -0.012 min

Lab File: BP010261.D

Acq: 10 May 2022 12:06




Tgt Ion:276 Resp: 2142500


Ion Ratio Lower Upper

276 100

138 22.6 12.3 18.5#

277 25.7 19.3 28.9









#95
Dibenzo(a,h)anthracene
Concen: 35.015 ng/ul
RT: 26.380 min Scan# 3963
Delta R.T. -0.018 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Instrument :
BNA_P
ClientSampleId :
SSTD020632

Tgt Ion:278 Resp: 1832587
Ion Ratio Lower Upper
278 100
139 14.5 6.6 10.0#
279 24.1 18.4 27.6



#96
Benzo(g,h,i)perylene
Concen: 34.289 ng/ul
RT: 27.145 min Scan# 4090
Delta R.T. -0.018 min
Lab File: BP010261.D
Acq: 10 May 2022 12:06

Tgt Ion:276 Resp: 1769633
Ion Ratio Lower Upper
276 100
138 18.7 11.6 17.4#
277 23.3 15.9 23.9

