

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010263.D
 Acq On : 10 May 2022 13:13
 Operator : CG/JU
 Sample : SSTD08034
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634

Quant Time: May 10 15:22:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.910	152	114783	20.000	ng/ul	0.00
20) Naphthalene-d8	10.716	136	507429	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.545	164	347758	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.298	188	780913	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.380	240	875179	20.000	ng/ul	# 0.00
88) Perylene-d12	23.821	264	961331	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.346	96	84076	31.941	ng/uL	0.00
4) Pyridine-d5	3.764	84	613211	82.811	ng/ul	0.00
7) Phenol-d5	7.075	99	874702	83.779	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.240	67	519130	82.865	ng/ul	0.00
11) 2-Chlorophenol-d4	7.446	132	657336	85.013	ng/ul	0.00
15) 4-Methylphenol-d8	8.628	113	705052	81.506	ng/ul	0.00
21) Nitrobenzene-d5	9.069	128	341598	84.642	ng/ul	0.00
24) 2-Nitrophenol-d4	9.799	143	393967	89.383	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.340	165	705276	82.194	ng/ul	0.00
31) 4-Chloroaniline-d4	10.851	131	996534	80.998	ng/ul	0.00
46) Dimethylphthalate-d6	13.957	166	2191538	81.013	ng/ul	0.00
49) Acenaphthylene-d8	14.240	160	2689809	82.157	ng/ul	0.00
54) 4-Nitrophenol-d4	14.751	143	443228	88.999	ng/ul	0.00
60) Fluorene-d10	15.539	176	1861518	79.996	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.663	200	455555	92.258	ng/ul	0.00
73) Anthracene-d10	17.398	188	3065628	81.145	ng/ul	0.00
81) Pyrene-d10	19.633	212	3796612	76.280	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.674	264	4183841	83.015	ng/ul	0.00
Target Compounds						
2) 1,4-Dioxane	3.381	88	85174	31.139	ng/ul#	17
5) Pyridine	3.781	79	640559	83.218	ng/ul#	35
6) Benzaldehyde	7.046	77	312367	55.985	ng/ul	94
8) Phenol	7.105	94	874821	83.391	ng/ul#	93
10) Bis(2-Chloroethyl)ether	7.334	93	662944	82.024	ng/ul#	75
12) 2-Chlorophenol	7.475	128	657212	84.266	ng/ul	96
13) 2-Methylphenol	8.357	108	663323	82.446	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.440	45	1044566	87.099	ng/ul#	82
16) Acetophenone	8.740	105	1086108	80.475	ng/ul#	81
17) N-Nitroso-di-n-propyla...	8.734	70	570078	80.622	ng/ul#	71
18) 4-Methylphenol	8.693	108	722909	81.607	ng/ul	96
19) Hexachloroethane	8.993	117	269812	82.329	ng/ul#	82
22) Nitrobenzene	9.116	77	840518	85.004	ng/ul#	85
23) Isophorone	9.652	82	1622766	83.090	ng/ul#	97
25) 2-Nitrophenol	9.828	139	398837	85.437	ng/ul#	85
26) 2,4-Dimethylphenol	9.893	107	829716	82.095	ng/ul#	79
27) Bis(2-Chloroethoxy)met...	10.122	93	932841	81.491	ng/ul#	97
29) 2,4-Dichlorophenol	10.363	162	675336	81.875	ng/ul#	84
30) Naphthalene	10.763	128	2151548	81.012	ng/ul	97
32) 4-Chloroaniline	10.875	127	986651	81.380	ng/ul	97
33) Hexachlorobutadiene	11.057	225	444794	76.087	ng/ul	96
34) Caprolactam	11.669	113	126397	42.601	ng/ul#	6
35) 4-Chloro-3-methylphenol	12.004	107	809381	81.229	ng/ul#	68
36) 2-Methylnaphthalene	12.375	142	1514840	78.646	ng/ul	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010263.D
 Acq On : 10 May 2022 13:13
 Operator : CG/JU
 Sample : SSTD08034
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634

Quant Time: May 10 15:22:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.593	142	1541150	78.801	ng/ul#	96
39) 1,2,4,5-Tetrachloroben...	12.740	216	863412	80.362	ng/ul#	92
40) Hexachlorocyclopentadiene	12.722	237	512466	84.637	ng/ul	97
41) 2,4,6-Trichlorophenol	12.981	196	584252	84.046	ng/ul#	82
42) 2,4,5-Trichlorophenol	13.051	196	647715	85.636	ng/ul#	87
43) 1,1'-Biphenyl	13.381	154	2052592	80.026	ng/ul#	93
44) 2-Chloronaphthalene	13.422	162	1608845	81.054	ng/ul	93
45) 2-Nitroaniline	13.628	65	598482	92.986	ng/ul#	80
47) Dimethylphthalate	14.004	163	2101777	79.944	ng/ul#	92
48) 2,6-Dinitrotoluene	14.122	165	473285	87.721	ng/ul	92
50) Acenaphthylene	14.269	152	2656588	81.917	ng/ul	95
51) 3-Nitroaniline	14.451	138	403964	82.652	ng/ul#	82
52) Acenaphthene	14.610	153	1703703	80.792	ng/ul	97
53) 2,4-Dinitrophenol	14.657	184	322186	106.863	ng/ul#	79
55) 4-Nitrophenol	14.769	109	371251	85.804	ng/ul#	44
56) Dibenzofuran	14.945	168	2462889	78.874	ng/ul	100
57) 2,4-Dinitrotoluene	14.910	165	684608	86.857	ng/ul#	81
58) 2,3,4,6-Tetrachlorophenol	15.169	232	561510	83.517	ng/ul#	85
59) Diethylphthalate	15.369	149	2200134	82.274	ng/ul	93
61) Fluorene	15.592	166	2045837	79.799	ng/ul	91
62) 4-Chlorophenyl-phenyle...	15.587	204	1049956	77.817	ng/ul	97
63) 4-Nitroaniline	15.616	138	392376	82.504	ng/ul#	77
66) 4,6-Dinitro-2-methylph...	15.681	198	436722	91.670	ng/ul#	92
67) N-Nitrosodiphenylamine	15.804	169	1803707	80.499	ng/ul	95
68) 4-Bromophenyl-phenylether	16.486	248	659319	79.153	ng/ul	93
69) Hexachlorobenzene	16.604	284	756201	78.214	ng/ul#	89
70) Atrazine	16.763	200	741280	81.375	ng/ul#	89
71) Pentachlorophenol	16.951	266	523298	84.988	ng/ul#	85
72) Phenanthrene	17.339	178	3430920	80.833	ng/ul	99
74) Anthracene	17.433	178	3471324	81.197	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.345	216	912589	80.048	ng/ul#	86
76) Pentachlorobenzene	14.869	250	854887	76.469	ng/ul	89
77) Carbazole	17.704	167	3179713	84.050	ng/ul#	95
78) Di-n-butylphthalate	18.251	149	3999227	85.925	ng/ul#	93
80) Fluoranthene	19.304	202	4345146	77.400	ng/ul	97
82) Pyrene	19.657	202	4387792	75.863	ng/ul	97
83) Butylbenzylphthalate	20.527	149	1988062	83.155	ng/ul#	80
84) 3,3'-Dichlorobenzidine	21.298	252	1552213	98.570	ng/ul#	96
85) Benzo(a)anthracene	21.369	228	4525315	81.477	ng/ul	95
86) Bis(2-ethylhexyl)phtha...	21.286	149	2895226	81.707	ng/ul#	81
87) Chrysene	21.422	228	4409282	80.625	ng/ul	98
89) Di-n-octyl phthalate	22.227	149	5145697	68.854	ng/ul	100
90) Benzo(b)fluoranthene	23.086	252	4999854	77.363	ng/ul	99
91) Benzo(k)fluoranthene	23.133	252	4714068	77.004	ng/ul#	97
93) Benzo(a)pyrene	23.727	252	4926347	82.430	ng/ul#	98
94) Indeno(1,2,3-cd)pyrene	26.404	276	5845888	98.329	ng/ul#	89
95) Dibenzo(a,h)anthracene	26.409	278	5014947	98.580	ng/ul#	95
96) Benzo(g,h,i)perylene	27.180	276	4930061	98.277	ng/ul#	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051022\
 Data File : BP010263.D
 Acq On : 10 May 2022 13:13
 Operator : CG/JU
 Sample : SSTD08034
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634

Quant Time: May 10 15:22:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue May 10 15:17:10 2022
 Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 31.139 ng/uL
 RT: 3.381 min Scan# 50
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion: 88 Resp: 85174
 Ion Ratio Lower Upper
 88 100
 43 34.7 6.8 10.2#
 58 85.7 28.4 42.6#



#5
 Pyridine
 Concen: 83.218 ng/uL
 RT: 3.781 min Scan# 118
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion: 79 Resp: 640559
 Ion Ratio Lower Upper
 79 100
 52 98.3 35.5 53.3#
 51 42.6 20.2 30.4#









#6
Benzaldehyde
Concen: 55.985 ng/uI
RT: 7.046 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP010263.D
Acq: 10 May 2022 13:13

Instrument :
BNA_P
ClientSampleId :
SSTD080634

Tgt Ion: 77 Resp: 312367




Ion	Ratio	Lower	Upper
77	100		
105	91.9	77.0	115.4
106	88.6	77.0	115.4




#8
Phenol
Concen: 83.391 ng/uI
RT: 7.105 min Scan# 683
Delta R.T. 0.000 min
Lab File: BP010263.D
Acq: 10 May 2022 13:13

Tgt Ion: 94 Resp: 874821


Ion	Ratio	Lower	Upper
94	100		
65	32.3	19.0	28.6
66	44.0	34.6	51.8








#10
 Bis(2-Chloroethyl)ether
 Concen: 82.024 ng/uI
 RT: 7.334 min Scan# 710
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634







Tgt Ion: 93 Resp: 662944
 Ion Ratio Lower Upper
 93 100
 63 84.4 45.0 67.4#
 95 32.2 26.3 39.5



#12
 2-Chlorophenol
 Concen: 84.266 ng/uI
 RT: 7.475 min Scan# 746
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion: 128 Resp: 657212
 Ion Ratio Lower Upper
 128 100
 64 55.0 45.3 67.9
 130 32.4 29.0 43.6









#13
 2-Methylphenol
 Concen: 82.446 ng/ul
 RT: 8.357 min Scan# 896
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion:108 Resp: 663323
 Ion Ratio Lower Upper
 108 100
 107 89.6 70.8 106.2



#14
 2,2'-oxybis(1-Chloropropane)
 Concen: 87.099 ng/ul
 RT: 8.440 min Scan# 910
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion: 45 Resp: 1044566
 Ion Ratio Lower Upper
 45 100
 77 12.3 17.1 25.7#
 79 9.7 13.2 19.8#








#16
 Acetophenone
 Concen: 80.475 ng/ul
 RT: 8.740 min Scan# 90
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634





Tgt Ion:105 Resp: 1086108
 Ion Ratio Lower Upper
 105 100
 77 79.7 53.4 80.2
 51 31.1 13.8 20.8#




#17
 N-Nitroso-di-n-propylamine
 Concen: 80.622 ng/ul
 RT: 8.734 min Scan# 960
 Delta R.T. 0.012 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13



Tgt Ion: 70 Resp: 570078
 Ion Ratio Lower Upper
 70 100
 42 65.6 31.8 47.8#
 101 9.8 15.6 23.4#
 130 20.7 13.7 20.5#









#18
 4-Methylphenol
 Concen: 81.607 ng/uI
 RT: 8.693 min Scan# 91
 Delta R.T. 0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion:108 Resp: 722909
 Ion Ratio Lower Upper
 108 100
 107 108.9 83.6 125.4



#19
 Hexachloroethane
 Concen: 82.329 ng/uI
 RT: 8.993 min Scan# 1004
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:117 Resp: 269812
 Ion Ratio Lower Upper
 117 100
 201 106.0 106.5 159.7#
 199 65.7 59.0 88.6









#22
 Nitrobenzene
 Concen: 85.004 ng/u1
 RT: 9.116 min Scan# 1025
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion: 77 Resp: 840518
 Ion Ratio Lower Upper
 77 100
 123 42.9 42.3 63.5
 65 14.4 5.8 8.8#



#23
 Isophorone
 Concen: 83.090 ng/u1
 RT: 9.652 min Scan# 1116
 Delta R.T. 0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion: 82 Resp: 1622766
 Ion Ratio Lower Upper
 82 100
 95 7.6 7.8 11.8#
 138 16.7 12.9 19.3









#25
 2-Nitrophenol
 Concen: 85.437 ng/ul
 RT: 9.828 min Scan# 1146
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 ClientSampleId : SSTD080634

Tgt Ion:139 Resp: 398837




Ion	Ratio	Lower	Upper
139	100		
65	59.9	49.4	74.2
109	34.7	44.7	67.1#




#26
 2,4-Dimethylphenol
 Concen: 82.095 ng/ul
 RT: 9.893 min Scan# 1157
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:107 Resp: 829716

Ion	Ratio	Lower	Upper
107	100		
121	47.9	34.1	51.1
122	80.5	46.3	69.5#









#27
 Bis(2-Chloroethoxy)methane
 Concen: 81.491 ng/ul
 RT: 10.122 min Scan# 1197
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion: 93 Resp: 932841
 Ion Ratio Lower Upper
 93 100
 95 31.8 26.2 39.4
 123 10.7 6.2 9.4#



#29
 2,4-Dichlorophenol
 Concen: 81.875 ng/ul
 RT: 10.363 min Scan# 1237
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:162 Resp: 675336
 Ion Ratio Lower Upper
 162 100
 164 65.5 37.0 55.6#
 98 39.1 31.0 46.6









#30
 Naphthalene
 Concen: 81.012 ng/uI
 RT: 10.763 min Scan# 111
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 ClientSampleId : SSTD080634

Tgt Ion:128 Resp: 2151548




Ion	Ratio	Lower	Upper
128	100		
129	10.9	8.6	13.0
127	13.0	12.2	18.4




#32
 4-Chloroaniline
 Concen: 81.380 ng/uI
 RT: 10.875 min Scan# 1324
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:127 Resp: 986651

Ion	Ratio	Lower	Upper
127	100		
129	32.4	24.6	37.0









#33
 Hexachlorobutadiene
 Concen: 76.087 ng/uI
 RT: 11.057 min Scan# 111
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument : BNA_P
 Client Sample Id : SSTD080634


Tgt Ion	Resp	Lower	Upper
225	100		
223	63.3	53.4	80.0
227	64.0	53.8	80.8



#34
 Caprolactam
 Concen: 42.601 ng/uI
 RT: 11.669 min Scan# 1459
 Delta R.T. 0.018 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion	Resp	Lower	Upper
113	100		
55	181.3	58.2	87.2#
56	144.2	68.6	103.0#









#35
 4-Chloro-3-methylphenol
 Concen: 81.229 ng/ul
 RT: 12.004 min Scan# 1516
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument : BNA_P
 ClientSampleId : SSTD080634


Tgt Ion	Resp	Lower	Upper
107	809381		
144	25.1	15.8	23.8#
142	76.1	38.9	58.3#



#36
 2-Methylnaphthalene
 Concen: 78.646 ng/ul
 RT: 12.375 min Scan# 1579
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion	Resp	Lower	Upper
142	1514840		
141	88.2	64.3	96.5








#37
 1-Methylnaphthalene
 Concen: 78.801 ng/ul
 RT: 12.593 min Scan# 1616
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 Client Sample Id : SSTD080634







Tgt Ion:142 Resp: 1541150
 Ion Ratio Lower Upper
 142 100
 141 91.0 69.8 104.8
 116 3.9 5.6 8.4#



#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 80.362 ng/ul
 RT: 12.740 min Scan# 1641
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:216 Resp: 863412
 Ion Ratio Lower Upper
 216 100
 214 80.3 58.2 87.2
 179 21.7 14.2 21.2#
 108 17.7 15.3 22.9









#40
 Hexachlorocyclopentadiene
 Concen: 84.637 ng/ul
 RT: 12.722 min Scan# 1638
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument : BNA_P
 Client Sample Id : SSTD080634


Tgt Ion	Resp	Lower	Upper
237	100		
235	63.7	53.2	79.8
272	13.0	10.8	16.2



#41
 2,4,6-Trichlorophenol
 Concen: 84.046 ng/ul
 RT: 12.981 min Scan# 1682
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion	Resp	Lower	Upper
196	100		
198	97.0	64.1	96.1#
200	31.7	19.9	29.9#









#42
 2,4,5-Trichlorophenol
 Concen: 85.636 ng/ul
 RT: 13.051 min Scan# 1694
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument : BNA_P
 Client Sample Id : SSTD080634


Tgt Ion	Resp	Lower	Upper
196	647715	100	
198	95.9	68.6	103.0
200	31.1	32.8	49.2#



#43
 1,1'-Biphenyl
 Concen: 80.026 ng/ul
 RT: 13.381 min Scan# 1750
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion	Resp	Lower	Upper
154	2052592	100	
153	40.1	28.6	43.0
76	12.8	8.5	12.7#









#44
 2-Chloronaphthalene
 Concen: 81.054 ng/uI
 RT: 13.422 min Scan# 1757
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 Client Sample Id : SSTD080634

Tgt Ion: 162 Resp: 1608845




Ion	Ratio	Lower	Upper
162	100		
164	32.6	27.5	41.3
127	39.7	37.1	55.7




#45
 2-Nitroaniline
 Concen: 92.986 ng/uI
 RT: 13.628 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion: 65 Resp: 598482

Ion	Ratio	Lower	Upper
65	100		
92	64.2	77.9	116.9
138	95.0	81.5	122.3









#47
 Dimethylphthalate
 Concen: 79.944 ng/ul
 RT: 14.004 min Scan# 1856
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument : BNA_P
 ClientSampleId : SSTD080634


Tgt Ion:163 Resp: 2101777
 Ion Ratio Lower Upper
 163 100
 194 4.3 4.1 6.1
 164 10.0 11.3 16.9#



#48
 2,6-Dinitrotoluene
 Concen: 87.721 ng/ul
 RT: 14.122 min Scan# 1876
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion:165 Resp: 473285
 Ion Ratio Lower Upper
 165 100
 89 58.0 51.7 77.5
 121 20.2 18.3 27.5








#50
 Acenaphthylene
 Concen: 81.917 ng/ul
 RT: 14.269 min Scan# 1901
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 Client Sample Id : SSTD080634







Tgt Ion:152 Resp: 2656588
 Ion Ratio Lower Upper
 152 100
 151 19.9 18.3 27.5
 153 13.1 11.0 16.6



#51
 3-Nitroaniline
 Concen: 82.652 ng/ul
 RT: 14.451 min Scan# 1932
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:138 Resp: 403964
 Ion Ratio Lower Upper
 138 100
 108 11.7 15.8 23.8#
 92 123.3 117.1 175.7









#52
 Acenaphthene
 Concen: 80.792 ng/uI
 RT: 14.610 min Scan# 1959
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634

Tgt Ion:153 Resp: 1703703




Ion	Ratio	Lower	Upper
153	100		
152	46.5	34.3	51.5
154	89.3	70.0	105.0




#53
 2,4-Dinitrophenol
 Concen: 106.863 ng/uI
 RT: 14.657 min Scan# 1967
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:184 Resp: 322186

Ion	Ratio	Lower	Upper
184	100		
63	71.1	34.6	52.0
154	66.0	57.7	86.5









#55
 4-Nitrophenol
 Concen: 85.804 ng/u1
 RT: 14.769 min Scan# 1985
 Delta R.T. 0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion:109 Resp: 371251
 Ion Ratio Lower Upper
 109 100
 139 121.2 61.4 92.2#
 65 134.9 65.3 97.9#



#56
 Dibenzofuran
 Concen: 78.874 ng/u1
 RT: 14.945 min Scan# 2016
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion:168 Resp: 2462889
 Ion Ratio Lower Upper
 168 100
 139 37.6 30.1 45.1








#57
 2,4-Dinitrotoluene
 Concen: 86.857 ng/ul
 RT: 14.910 min Scan# 2010
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 Client Sample Id : SSTD080634



Tgt Ion:165 Resp: 684608




Ion	Ratio	Lower	Upper
165	100		
63	49.0	29.5	44.3#
182	3.1	2.8	4.2




#58
 2,3,4,6-Tetrachlorophenol
 Concen: 83.517 ng/ul
 RT: 15.169 min Scan# 2054
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:232 Resp: 561510


Ion	Ratio	Lower	Upper
232	100		
131	46.0	25.4	38.2#
130	2.5	4.2	6.2#
166	31.9	28.1	42.1








#59
 Diethylphthalate
 Concen: 82.274 ng/u1
 RT: 15.369 min Scan# 2088
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634





Tgt Ion:149 Resp: 2200134
 Ion Ratio Lower Upper
 149 100
 177 22.4 22.0 33.0
 150 12.6 10.2 15.2




#61
 Fluorene
 Concen: 79.799 ng/u1
 RT: 15.592 min Scan# 2126
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13



Tgt Ion:166 Resp: 2045837
 Ion Ratio Lower Upper
 166 100
 165 97.0 70.3 105.5
 167 13.5 9.5 14.3









#62
4-Chlorophenyl-phenylether
Concen: 77.817 ng/ul
RT: 15.587 min Scan# 2126
Delta R.T. -0.006 min
Lab File: BP010263.D
Acq: 10 May 2022 13:13




Instrument :
BNA_P
ClientSampleId :
SSTD080634


Tgt Ion:204 Resp: 1049956
Ion Ratio Lower Upper
204 100
206 32.7 27.6 41.4
141 59.8 46.6 69.8



#63
4-Nitroaniline
Concen: 82.504 ng/ul
RT: 15.616 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP010263.D
Acq: 10 May 2022 13:13

Tgt Ion:138 Resp: 392376
Ion Ratio Lower Upper
138 100
92 54.9 41.6 62.4
108 87.1 97.6 146.4#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 91.670 ng/ul
 RT: 15.681 min Scan# 2141
 Delta R.T. 0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion:198 Resp: 436722
 Ion Ratio Lower Upper
 198 100
 182 3.3 6.0 9.0#
 77 25.1 17.7 26.5



#67
 N-Nitrosodiphenylamine
 Concen: 80.499 ng/ul
 RT: 15.804 min Scan# 2162
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion:169 Resp: 1803707
 Ion Ratio Lower Upper
 169 100
 168 64.8 48.3 72.5
 167 34.7 29.6 44.4








#68
 4-Bromophenyl-phenylether
 Concen: 79.153 ng/ul
 RT: 16.486 min Scan# 2278
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 ClientSampleId : SSTD080634



Tgt Ion: 248 Resp: 659319




Ion	Ratio	Lower	Upper
248	100		
250	98.1	79.4	119.0
141	71.2	67.2	100.8




#69
 Hexachlorobenzene
 Concen: 78.214 ng/ul
 RT: 16.604 min Scan# 2298
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion: 284 Resp: 756201

Ion	Ratio	Lower	Upper
284	100		
142	35.1	19.2	28.8#
249	32.9	27.4	41.2









#70
 Atrazine
 Concen: 81.375 ng/ul
 RT: 16.763 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634

Tgt Ion:200 Resp: 741280




Ion	Ratio	Lower	Upper
200	100		
173	26.6	19.2	28.8
215	49.7	32.2	48.4#




#71
 Pentachlorophenol
 Concen: 84.988 ng/ul
 RT: 16.951 min Scan# 2357
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:266 Resp: 523298

Ion	Ratio	Lower	Upper
266	100		
264	63.8	57.3	85.9
268	64.9	65.7	98.5#









#72
 Phenanthrene
 Concen: 80.833 ng/u1
 RT: 17.339 min Scan# 2439
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634


Tgt Ion:178 Resp: 3430920
 Ion Ratio Lower Upper
 178 100
 179 15.3 11.7 17.5
 176 18.3 14.5 21.7



#74
 Anthracene
 Concen: 81.197 ng/u1
 RT: 17.433 min Scan# 2439
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion:178 Resp: 3471324
 Ion Ratio Lower Upper
 178 100
 179 15.8 12.3 18.5
 176 18.1 13.4 20.2








#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 80.048 ng/uL
 RT: 13.345 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634







Tgt Ion:216 Resp: 912589
 Ion Ratio Lower Upper
 216 100
 214 77.8 50.4 75.6#
 218 48.0 35.2 52.8



#76
 Pentachlorobenzene
 Concen: 76.469 ng/uL
 RT: 14.869 min Scan# 2003
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:250 Resp: 854887
 Ion Ratio Lower Upper
 250 100
 248 64.0 43.5 65.3
 252 64.6 46.5 69.7









#77
 Carbazole
 Concen: 84.050 ng/uI
 RT: 17.704 min Scan# 2485
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 Client Sample Id : SSTD080634

Tgt Ion:167 Resp: 3179713




Ion	Ratio	Lower	Upper
167	100		
166	21.5	15.8	23.8
139	12.8	8.3	12.5#




#78
 Di-n-butylphthalate
 Concen: 85.925 ng/uI
 RT: 18.251 min Scan# 2578
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:149 Resp: 3999227

Ion	Ratio	Lower	Upper
149	100		
150	9.2	9.6	14.4#
104	5.4	6.2	9.4#









#80
 Fluoranthene
 Concen: 77.400 ng/u1
 RT: 19.304 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 Client Sample Id : SSTD080634

Tgt Ion:202 Resp: 4345146




Ion	Ratio	Lower	Upper
202	100		
101	10.4	9.6	14.4
100	8.1	6.3	9.5




#82
 Pyrene
 Concen: 75.863 ng/u1
 RT: 19.657 min Scan# 2817
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:202 Resp: 4387792

Ion	Ratio	Lower	Upper
202	100		
101	12.3	10.6	15.8
100	10.6	9.8	14.6





#83
 Butylbenzylphthalate
 Concen: 83.155 ng/ul
 RT: 20.527 min Scan# 2965
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13





Instrument :

BNA_P

Client Sample Id :




SSTD080634


Tgt Ion:149 Resp: 1988062
 Ion Ratio Lower Upper
 149 100
 91 71.4 71.1 106.7
 206 23.1 28.8 43.2#



#84
 3,3'-Dichlorobenzidine
 Concen: 98.570 ng/ul
 RT: 21.298 min Scan# 3096
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion:252 Resp: 1552213
 Ion Ratio Lower Upper
 252 100
 254 64.7 49.8 74.6
 126 9.6 4.3 6.5#








#85
 Benzo(a)anthracene
 Concen: 81.477 ng/ul
 RT: 21.369 min Scan# 3108
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 ClientSampleId : SSTD080634







Tgt Ion:228 Resp: 4525315
 Ion Ratio Lower Upper
 228 100
 229 19.7 18.2 27.2
 226 26.5 22.8 34.2



#86
 Bis(2-ethylhexyl)phthalate
 Concen: 81.707 ng/ul
 RT: 21.286 min Scan# 3094
 Delta R.T. -0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Tgt Ion:149 Resp: 2895226
 Ion Ratio Lower Upper
 149 100
 167 27.9 32.8 49.2#
 279 4.6 5.6 8.4#








#87
 Chrysene
 Concen: 80.625 ng/uI
 RT: 21.422 min Scan# 3117
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument : BNA_P
 Client Sample Id : SSTD080634







Tgt Ion:228 Resp: 4409282
 Ion Ratio Lower Upper
 228 100
 226 29.6 22.7 34.1
 229 19.6 15.2 22.8



#89
 Di-n-octyl phthalate
 Concen: 68.854 ng/uI
 RT: 22.227 min Scan# 3254
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:149 Resp: 5145697









#90
 Benzo(b)fluoranthene
 Concen: 77.363 ng/u1
 RT: 23.086 min Scan# 3400
 Delta R.T. 0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument :
 BNA_P
ClientSampleId :
 SSTD080634

Tgt Ion:252 Resp: 4999854




Ion	Ratio	Lower	Upper
252	100		
253	21.8	17.4	26.0
125	8.5	7.3	10.9




#91
 Benzo(k)fluoranthene
 Concen: 77.004 ng/u1
 RT: 23.133 min Scan# 3408
 Delta R.T. 0.000 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:252 Resp: 4714068

Ion	Ratio	Lower	Upper
252	100		
253	21.7	17.9	26.9
125	8.9	5.1	7.7#









#93
 Benzo(a)pyrene
 Concen: 82.430 ng/u1
 RT: 23.727 min Scan# 31
 Delta R.T. 0.006 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080634

Tgt Ion:252 Resp: 4926347




Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.5	26.3
125	9.6	6.4	9.6#




#94
 Indeno(1,2,3-cd)pyrene
 Concen: 98.329 ng/u1
 RT: 26.404 min Scan# 3964
 Delta R.T. 0.024 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13

Tgt Ion:276 Resp: 5845888


Ion	Ratio	Lower	Upper
276	100		
138	24.4	12.3	18.5#
277	26.3	19.3	28.9








#95
 Dibenzo(a,h)anthracene
 Concen: 98.580 ng/ul
 RT: 26.409 min Scan# 3965
 Delta R.T. 0.012 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13


Instrument : BNA_P
 Client Sample Id : SSTD080634



Tgt Ion: 278 Resp: 5014947
 Ion Ratio Lower Upper
 278 100
 139 13.9 6.6 10.0#
 279 23.8 18.4 27.6



#96
 Benzo(g,h,i)perylene
 Concen: 98.277 ng/ul
 RT: 27.180 min Scan# 4096
 Delta R.T. 0.018 min
 Lab File: BP010263.D
 Acq: 10 May 2022 13:13



Tgt Ion: 276 Resp: 4930061
 Ion Ratio Lower Upper
 276 100
 138 17.8 11.6 17.4#
 277 23.1 15.9 23.9

