

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP051024\  
 Data File : BP020327.D  
 Acq On : 11 May 2024 18:39  
 Operator : MA/JU  
 Sample : P2371-11  
 Misc :  
 ALS Vial : 47 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 A43S3

Quant Time: May 12 23:15:09 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP050124.MA.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Wed May 08 22:57:55 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.611  | 152  | 78106    | 20.000 | ng/ul | 0.00     |
| 20) Naphthalene-d8            | 10.387 | 136  | 328457   | 20.000 | ng/ul | 0.00     |
| 38) Acenaphthene-d10          | 14.298 | 164  | 231372   | 20.000 | ng/ul | 0.00     |
| 64) Phenanthrene-d10          | 17.134 | 188  | 529562   | 20.000 | ng/ul | 0.00     |
| 79) Chrysene-d12              | 21.633 | 240  | 543874   | 20.000 | ng/ul | 0.00     |
| 88) Perylene-d12              | 24.998 | 264  | 622312   | 20.000 | ng/ul | -0.01    |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 3) 1,4-Dioxane-d8             | 3.217  | 96   | 7442     | 4.024  | ng/uL | 0.00     |
| 4) Pyridine-d5                | 3.617  | 84   | 111362   | 20.723 | ng/ul | 0.00     |
| 7) Phenol-d5                  | 6.805  | 99   | 157976   | 23.552 | ng/ul | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 6.963  | 67   | 96640    | 23.717 | ng/ul | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.152  | 132  | 122967   | 25.644 | ng/ul | 0.00     |
| 15) 4-Methylphenol-d8         | 8.322  | 113  | 132213   | 24.380 | ng/ul | 0.00     |
| 21) Nitrobenzene-d5           | 8.775  | 128  | 62226    | 25.299 | ng/ul | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.487  | 143  | 71938    | 25.541 | ng/ul | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.028 | 165  | 131900   | 25.662 | ng/ul | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.546 | 131  | 182573   | 25.380 | ng/ul | 0.00     |
| 46) Dimethylphthalate-d6      | 13.710 | 166  | 502351   | 29.732 | ng/ul | 0.00     |
| 49) Acenaphthylene-d8         | 13.981 | 160  | 521063   | 27.917 | ng/ul | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.581 | 143  | 51430    | 19.805 | ng/ul | 0.03     |
| 60) Fluorene-d10              | 15.322 | 176  | 422260   | 29.963 | ng/ul | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.481 | 200  | 60708    | 18.065 | ng/ul | 0.00     |
| 73) Anthracene-d10            | 17.234 | 188  | 710515   | 31.014 | ng/ul | -0.01    |
| 81) Pyrene-d10                | 19.633 | 212  | 891259   | 28.034 | ng/ul | -0.01    |
| 92) Benzo(a)pyrene-d12        | 24.757 | 264  | 901623   | 29.965 | ng/ul | -0.02    |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP051024\  
Data File : BP020327.D  
Acq On : 11 May 2024 18:39  
Operator : MA/JU  
Sample : P2371-11  
Misc :  
ALS Vial : 47 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
A43S3

Quant Time: May 12 23:15:09 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP050124.MA.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Wed May 08 22:57:55 2024  
Response via : Initial Calibration

