

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051121\
 Data File : BP005482.D
 Acq On : 11 May 2021 10:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 11 11:22:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 11 11:20:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	65	0.00
2	1,4-Dioxane	0.459	0.633	-37.9#	102	0.00
3	Pyridine	0.923	1.415	-53.3#	100	0.00
4	n-Nitrosodimethylamine	0.742	0.988	-33.2#	86	0.00
5 S	2-Fluorophenol	1.042	1.129	-8.3	74	0.00
6	Aniline	1.403	1.357	3.3	65	0.00
7 S	Phenol-d6	1.299	1.253	3.5	63	0.00
8	2-Chlorophenol	1.119	1.104	1.3	65	0.00
9	Benzaldehyde	0.750	0.708	5.6	68	0.00
10 C	Phenol	1.300	1.221	6.1	63	0.00
11	bis(2-Chloroethyl)ether	0.898	0.807	10.1	62	0.00
12	1,3-Dichlorobenzene	1.489	1.406	5.6	68	0.00
13 C	1,4-Dichlorobenzene	1.484	1.416	4.6	66	0.00
14	1,2-Dichlorobenzene	1.423	1.388	2.5	66	0.00
15	Benzyl Alcohol	1.254	1.214	3.2	63	0.00
16	2,2'-oxybis(1-Chloropropane	0.443	0.393	11.3	58	0.00
17	2-Methylphenol	0.893	0.854	4.4	64	0.00
18	Hexachloroethane	0.615	0.604	1.8	68	0.00
19 P	n-Nitroso-di-n-propylamine	0.952	0.900	5.5	62	0.00
20	3+4-Methylphenols	1.187	1.102	7.2	61	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	64	0.00
22	Acetophenone	0.564	0.551	2.3	64	0.00
23 S	Nitrobenzene-d5	0.508	0.505	0.6	65	0.00
24	Nitrobenzene	0.510	0.496	2.7	64	0.00
25	Isophorone	0.787	0.740	6.0	62	0.00
26 C	2-Nitrophenol	0.213	0.195	8.5	62	0.00
27	2,4-Dimethylphenol	0.283	0.254	10.2	59	0.00
28	bis(2-Chloroethoxy)methane	0.329	0.311	5.5	62	0.00
29 C	2,4-Dichlorophenol	0.444	0.430	3.2	63	0.00
30	1,2,4-Trichlorobenzene	0.567	0.525	7.4	61	0.00
31	Naphthalene	1.075	1.000	7.0	61	0.00
32	Benzoic acid	0.218	0.207	5.0	60	0.00
33	4-Chloroaniline	0.419	0.379	9.5	59	0.00
34 C	Hexachlorobutadiene	0.473	0.482	-1.9	69	0.00
35	Caprolactam	0.102	0.088	13.7	54	0.00
36 C	4-Chloro-3-methylphenol	0.396	0.362	8.6	60	0.00
37	2-Methylnaphthalene	0.863	0.813	5.8	62	0.00
38	1-Methylnaphthalene	0.814	0.781	4.1	62	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	66	0.00
40	1,2,4,5-Tetrachlorobenzene	0.919	0.843	8.3	67	0.00
41 P	Hexachlorocyclopentadiene	0.313	0.342	-9.3	72	0.00
42 S	2,4,6-Tribromophenol	0.568	0.546	3.9	68	0.00
43 C	2,4,6-Trichlorophenol	0.574	0.527	8.2	63	0.00
44	2,4,5-Trichlorophenol	0.620	0.542	12.6	62	0.00
45 S	2-Fluorobiphenyl	1.773	1.614	9.0	64	0.00
46	1,1'-Biphenyl	1.540	1.364	11.4	62	0.00
47	2-Chloronaphthalene	1.282	1.158	9.7	64	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051121\
 Data File : BP005482.D
 Acq On : 11 May 2021 10:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 11 11:22:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 11 11:20:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.364	0.349	4.1	64	0.00
49	Acenaphthylene	1.847	1.642	11.1	62	0.00
50	Dimethylphthalate	1.726	1.520	11.9	62	0.00
51	2,6-Dinitrotoluene	0.388	0.343	11.6	62	0.00
52 C	Acenaphthene	1.149	1.015	11.7	61	0.00
53	3-Nitroaniline	0.278	0.236	15.1	56	0.00
54 P	2,4-Dinitrophenol	0.234	0.229	2.1	62	0.00
55	Dibenzofuran	2.162	1.990	8.0	64	0.00
56 P	4-Nitrophenol	0.210	0.190	9.5	59	0.00
57	2,4-Dinitrotoluene	0.572	0.512	10.5	64	0.00
58	Fluorene	1.765	1.619	8.3	63	0.00
59	2,3,4,6-Tetrachlorophenol	0.619	0.598	3.4	67	0.00
60	Diethylphthalate	1.672	1.518	9.2	63	0.00
61	4-Chlorophenyl-phenylether	1.105	1.070	3.2	69	0.00
62	4-Nitroaniline	0.267	0.242	9.4	61	0.00
63	Azobenzene	1.724	1.757	-1.9	69	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	68	0.00
65	4,6-Dinitro-2-methylphenol	0.158	0.155	1.9	65	0.00
66 c	n-Nitrosodiphenylamine	0.584	0.543	7.0	64	0.00
67	4-Bromophenyl-phenylether	0.312	0.297	4.8	69	0.00
68	Hexachlorobenzene	0.396	0.373	5.8	69	0.00
69	Atrazine	0.273	0.268	1.8	70	0.00
70 C	Pentachlorophenol	0.210	0.196	6.7	66	0.00
71	Phenanthrene	1.136	1.057	7.0	65	0.00
72	Anthracene	1.128	1.059	6.1	66	0.00
73	Carbazole	0.993	0.901	9.3	63	0.00
74	Di-n-butylphthalate	1.090	1.029	5.6	66	0.00
75 C	Fluoranthene	1.563	1.512	3.3	68	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
77	Benzydine	0.310	0.318	-2.6	67	0.00
78	Pyrene	1.132	1.064	6.0	69	0.00
79 S	Terphenyl-d14	1.178	1.125	4.5	69	0.00
80	Butylbenzylphthalate	0.366	0.352	3.8	69	0.00
81	Benzo(a)anthracene	1.296	1.205	7.0	68	0.00
82	3,3'-Dichlorobenzidine	0.488	0.461	5.5	70	0.00
83	Chrysene	1.256	1.154	8.1	67	0.00
84	Bis(2-ethylhexyl)phthalate	0.559	0.521	6.8	66	0.00
85 c	Di-n-octyl phthalate	0.940	0.869	7.6	66	0.00
86 I	Perylene-d12	1.000	1.000	0.0	72	0.00
87	Indeno(1,2,3-cd)pyrene	1.683	1.544	8.3	69	0.00
88	Benzo(b)fluoranthene	1.333	1.228	7.9	66	0.00
89	Benzo(k)fluoranthene	1.303	1.196	8.2	71	0.00
90 C	Benzo(a)pyrene	1.222	1.117	8.6	68	0.00
91	Dibenzo(a,h)anthracene	1.485	1.359	8.5	68	0.00
92	Benzo(g,h,i)perylene	1.374	1.260	8.3	68	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051121\
Data File : BP005482.D
Acq On : 11 May 2021 10:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 11 11:22:26 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 11 11:20:59 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0