

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051121\
 Data File : BP005482.D
 Acq On : 11 May 2021 10:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 11 11:22:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 11 11:20:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	65	0.00
2	1,4-Dioxane	40.000	64.963	-62.4#	102	0.00
3	Pyridine	40.000	61.352	-53.4#	100	0.00
4	n-Nitrosodimethylamine	40.000	53.283	-33.2#	86	0.00
5 S	2-Fluorophenol	80.000	86.655	-8.3	74	0.00
6	Aniline	40.000	38.697	3.3	65	0.00
7 S	Phenol-d6	80.000	77.167	3.5	63	0.00
8	2-Chlorophenol	40.000	39.439	1.4	65	0.00
9	Benzaldehyde	40.000	43.002	-7.5	68	0.00
10 C	Phenol	40.000	37.565	6.1	63	0.00
11	bis(2-Chloroethyl)ether	40.000	35.978	10.1	62	0.00
12	1,3-Dichlorobenzene	40.000	37.775	5.6	68	0.00
13 C	1,4-Dichlorobenzene	40.000	38.146	4.6	66	0.00
14	1,2-Dichlorobenzene	40.000	39.019	2.5	66	0.00
15	Benzyl Alcohol	40.000	38.726	3.2	63	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.450	11.4	58	0.00
17	2-Methylphenol	40.000	38.247	4.4	64	0.00
18	Hexachloroethane	40.000	39.280	1.8	68	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.817	5.5	62	0.00
20	3+4-Methylphenols	40.000	37.157	7.1	61	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	64	0.00
22	Acetophenone	40.000	39.070	2.3	64	0.00
23 S	Nitrobenzene-d5	80.000	79.492	0.6	65	0.00
24	Nitrobenzene	40.000	38.947	2.6	64	0.00
25	Isophorone	40.000	37.612	6.0	62	0.00
26 C	2-Nitrophenol	40.000	36.525	8.7	62	0.00
27	2,4-Dimethylphenol	40.000	35.932	10.2	59	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.711	5.7	62	0.00
29 C	2,4-Dichlorophenol	40.000	38.731	3.2	63	0.00
30	1,2,4-Trichlorobenzene	40.000	37.068	7.3	61	0.00
31	Naphthalene	40.000	37.210	7.0	61	0.00
32	Benzoic acid	40.000	37.961	5.1	60	0.00
33	4-Chloroaniline	40.000	36.163	9.6	59	0.00
34 C	Hexachlorobutadiene	40.000	40.808	-2.0	69	0.00
35	Caprolactam	40.000	34.355	14.1	54	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.629	8.4	60	0.00
37	2-Methylnaphthalene	40.000	37.683	5.8	62	0.00
38	1-Methylnaphthalene	40.000	38.382	4.0	62	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	66	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.677	8.3	67	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.837	0.4	72	0.00
42 S	2,4,6-Tribromophenol	80.000	76.992	3.8	68	0.00
43 C	2,4,6-Trichlorophenol	40.000	36.699	8.3	63	0.00
44	2,4,5-Trichlorophenol	40.000	34.993	12.5	62	0.00
45 S	2-Fluorobiphenyl	80.000	72.842	8.9	64	0.00
46	1,1'-Biphenyl	40.000	35.440	11.4	62	0.00
47	2-Chloronaphthalene	40.000	36.113	9.7	64	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051121\
 Data File : BP005482.D
 Acq On : 11 May 2021 10:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: May 11 11:22:26 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 11 11:20:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.383	4.0	64	0.00
49	Acenaphthylene	40.000	35.550	11.1	62	0.00
50	Dimethylphthalate	40.000	35.213	12.0	62	0.00
51	2,6-Dinitrotoluene	40.000	35.374	11.6	62	0.00
52 C	Acenaphthene	40.000	35.335	11.7	61	0.00
53	3-Nitroaniline	40.000	33.930	15.2	56	0.00
54 P	2,4-Dinitrophenol	40.000	39.106	2.2	62	0.00
55	Dibenzofuran	40.000	36.816	8.0	64	0.00
56 P	4-Nitrophenol	40.000	36.222	9.4	59	0.00
57	2,4-Dinitrotoluene	40.000	35.832	10.4	64	0.00
58	Fluorene	40.000	36.697	8.3	63	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.608	3.5	67	0.00
60	Diethylphthalate	40.000	36.305	9.2	63	0.00
61	4-Chlorophenyl-phenylether	40.000	38.768	3.1	69	0.00
62	4-Nitroaniline	40.000	36.249	9.4	61	0.00
63	Azobenzene	40.000	40.770	-1.9	69	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	68	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.045	2.4	65	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.243	6.9	64	0.00
67	4-Bromophenyl-phenylether	40.000	38.055	4.9	69	0.00
68	Hexachlorobenzene	40.000	37.641	5.9	69	0.00
69	Atrazine	40.000	39.262	1.8	70	0.00
70 C	Pentachlorophenol	40.000	37.267	6.8	66	0.00
71	Phenanthrene	40.000	37.191	7.0	65	0.00
72	Anthracene	40.000	37.573	6.1	66	0.00
73	Carbazole	40.000	36.303	9.2	63	0.00
74	Di-n-butylphthalate	40.000	37.741	5.6	66	0.00
75 C	Fluoranthene	40.000	38.699	3.3	68	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	70	0.00
77	Benzydine	40.000	41.084	-2.7	67	0.00
78	Pyrene	40.000	37.597	6.0	69	0.00
79 S	Terphenyl-d14	80.000	76.453	4.4	69	0.00
80	Butylbenzylphthalate	40.000	38.528	3.7	69	0.00
81	Benzo(a)anthracene	40.000	37.179	7.1	68	0.00
82	3,3'-Dichlorobenzidine	40.000	37.739	5.7	70	0.00
83	Chrysene	40.000	36.757	8.1	67	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.227	6.9	66	0.00
85 c	Di-n-octyl phthalate	40.000	37.004	7.5	66	0.00
86 I	Perylene-d12	20.000	20.000	0.0	72	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.714	8.2	69	0.00
88	Benzo(b)fluoranthene	40.000	36.856	7.9	66	0.00
89	Benzo(k)fluoranthene	40.000	36.702	8.2	71	0.00
90 C	Benzo(a)pyrene	40.000	36.585	8.5	68	0.00
91	Dibenzo(a,h)anthracene	40.000	36.609	8.5	68	0.00
92	Benzo(g,h,i)perylene	40.000	36.688	8.3	68	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051121\
Data File : BP005482.D
Acq On : 11 May 2021 10:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: May 11 11:22:26 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP050521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 11 11:20:59 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0