

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051123\
 Data File : BP015062.D
 Acq On : 11 May 2023 12:47
 Operator : CG/JU
 Sample : SST02073
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020613

Quant Time: May 11 22:12:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 11 15:04:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.140	152	320699	20.000	ng/u1	0.00
20) Naphthalene-d8	10.975	136	1461252	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.787	164	931217	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.539	188	2037838	20.000	ng/u1	0.00
79) Chrysene-d12	21.621	240	1830070	20.000	ng/u1	0.00
88) Perylene-d12	24.198	264	2159086	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	69807	8.457	ng/uL	0.00
4) Pyridine-d5	3.881	84	500861	21.408	ng/u1	0.00
7) Phenol-d5	7.293	99	629480	21.140	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.458	67	377121	21.929	ng/u1	0.00
11) 2-Chlorophenol-d4	7.669	132	474072	21.654	ng/u1	0.00
15) 4-Methylphenol-d8	8.857	113	515021	21.706	ng/u1	0.00
21) Nitrobenzene-d5	9.316	128	250278	21.797	ng/u1	0.00
24) 2-Nitrophenol-d4	10.046	143	279188	21.852	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.593	165	504941	21.902	ng/u1	0.00
31) 4-Chloroaniline-d4	11.110	131	767568	22.502	ng/u1	0.00
46) Dimethylphthalate-d6	14.187	166	1551576	22.049	ng/u1	0.00
49) Acenaphthylene-d8	14.481	160	1797956	22.340	ng/u1	0.00
54) 4-Nitrophenol-d4	14.975	143	295524	21.628	ng/u1	0.00
60) Fluorene-d10	15.775	176	1308430	22.036	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.898	200	274805	21.398	ng/u1	0.00
73) Anthracene-d10	17.639	188	2062939	22.108	ng/u1	0.00
81) Pyrene-d10	19.857	212	2321688	21.874	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.027	264	2271038	21.189	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.475	88	75825	8.549	ng/uL	100
5) Pyridine	3.899	79	521624	21.437	ng/u1	100
6) Benzaldehyde	7.269	77	165146	20.178	ng/u1	100
8) Phenol	7.322	94	652646	21.375	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.552	93	532971	21.818	ng/u1	100
12) 2-Chlorophenol	7.699	128	502852	21.715	ng/u1	100
13) 2-Methylphenol	8.587	108	511505	21.676	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.681	45	677731	22.144	ng/u1	100
16) Acetophenone	8.975	105	853038	23.036	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.957	70	436605	22.481	ng/u1	100
18) 4-Methylphenol	8.922	108	565074	21.756	ng/u1	100
19) Hexachloroethane	9.234	117	206047	21.792	ng/u1	100
22) Nitrobenzene	9.363	77	620524	21.881	ng/u1	100
23) Isophorone	9.893	82	1281547	21.985	ng/u1	100
25) 2-Nitrophenol	10.081	139	308783	22.040	ng/u1	100
26) 2,4-Dimethylphenol	10.134	107	613592	21.947	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.375	93	776161	21.884	ng/u1	100
29) 2,4-Dichlorophenol	10.616	162	509439	21.876	ng/u1	100
30) Naphthalene	11.028	128	1725364	21.718	ng/u1	100
32) 4-Chloroaniline	11.134	127	796924	22.731	ng/u1	100
33) Hexachlorobutadiene	11.304	225	301151	21.457	ng/u1	100
34) Caprolactam	11.910	113	109862	12.514	ng/u1	100
35) 4-Chloro-3-methylphenol	12.246	107	579450	22.037	ng/u1	100
36) 2-Methylnaphthalene	12.622	142	1171363	21.883	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP051123\
 Data File : BP015062.D
 Acq On : 11 May 2023 12:47
 Operator : CG/JU
 Sample : SST02073
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST020613


Quant Time: May 11 22:12:21 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 11 15:04:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.840	142	1186118	22.066	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.987	216	608573	21.851	ng/ul	100
40) Hexachlorocyclopentadiene	12.963	237	374026	20.728	ng/ul	100
41) 2,4,6-Trichlorophenol	13.222	196	423747	21.547	ng/ul	100
42) 2,4,5-Trichlorophenol	13.293	196	457706	21.465	ng/ul	100
43) 1,1'-Biphenyl	13.622	154	1593829	22.060	ng/ul	100
44) 2-Chloronaphthalene	13.663	162	1236622	21.917	ng/ul	100
45) 2-Nitroaniline	13.869	65	361230	22.375	ng/ul	100
47) Dimethylphthalate	14.234	163	1618000	22.078	ng/ul	100
48) 2,6-Dinitrotoluene	14.357	165	325493	22.419	ng/ul	100
50) Acenaphthylene	14.510	152	2000431	22.277	ng/ul	100
51) 3-Nitroaniline	14.687	138	261176	19.978	ng/ul	100
52) Acenaphthene	14.851	153	1354378	22.083	ng/ul	100
53) 2,4-Dinitrophenol	14.898	184	193554	20.078	ng/ul	100
55) 4-Nitrophenol	14.992	109	221704	22.060	ng/ul	100
56) Dibenzofuran	15.181	168	1886700	22.117	ng/ul	100
57) 2,4-Dinitrotoluene	15.145	165	471674	22.451	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.404	232	397236	22.074	ng/ul	100
59) Diethylphthalate	15.592	149	1620620	22.172	ng/ul	100
61) Fluorene	15.834	166	1528920	22.302	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.822	204	741256	22.130	ng/ul	100
63) 4-Nitroaniline	15.851	138	234888	20.806	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.910	198	288769	21.787	ng/ul	100
67) N-Nitrosodiphenylamine	16.034	169	1323896	22.094	ng/ul	100
68) 4-Bromophenyl-phenylether	16.716	248	452312	21.549	ng/ul	100
69) Hexachlorobenzene	16.839	284	536284	21.620	ng/ul	100
70) Atrazine	16.986	200	494501	22.327	ng/ul	100
71) Pentachlorophenol	17.186	266	329458	21.234	ng/ul	100
72) Phenanthrene	17.581	178	2436952	21.961	ng/ul	100
74) Anthracene	17.675	178	2495170	22.236	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.587	216	626494	21.683	ng/ul	100
76) Pentachlorobenzene	15.104	250	629659	21.704	ng/ul	100
77) Carbazole	17.939	167	2254668	22.914	ng/ul	100
78) Di-n-butylphthalate	18.469	149	2807568	22.157	ng/ul	100
80) Fluoranthene	19.533	202	2822914	21.821	ng/ul	100
82) Pyrene	19.886	202	2933992	21.962	ng/ul	100
83) Butylbenzylphthalate	20.739	149	1298664	21.927	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.522	252	887846	22.090	ng/ul	100
85) Benzo(a)anthracene	21.604	228	2741357	21.664	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.498	149	1878783	22.034	ng/ul	100
87) Chrysene	21.657	228	2625177	21.702	ng/ul	100
89) Di-n-octyl phthalate	22.480	149	3230113	21.844	ng/ul	100
90) Benzo(b)fluoranthene	23.404	252	2894907	21.150	ng/ul	100
91) Benzo(k)fluoranthene	23.451	252	2811859	21.394	ng/ul	100
93) Benzo(a)pyrene	24.086	252	2572446	21.263	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.933	276	3352487	21.401	ng/ul	100
95) Dibenzo(a,h)anthracene	26.951	278	2763269	21.489	ng/ul	100
96) Benzo(g,h,i)perylene	27.780	276	2733689	21.400	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: May 11 22:12:21 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP051123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 11 15:04:16 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.549 ng/uL
RT: 3.475 min Scan# 15
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion: 88 Resp: 75825
Ion Ratio Lower Upper
88 100
43 25.5 20.4 30.6
58 69.3 55.4 83.2



#5
Pyridine
Concen: 21.437 ng/uL
RT: 3.899 min Scan# 87
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion: 79 Resp: 521624
Ion Ratio Lower Upper
79 100
52 69.3 55.4 83.2
51 31.8 25.4 38.2









#6
Benzaldehyde
Concen: 20.178 ng/ul
RT: 7.269 min Scan# 660
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion: 77 Resp: 165146
Ion Ratio Lower Upper
77 100
105 96.1 76.9 115.3
106 93.8 75.0 112.6



#8
Phenol
Concen: 21.375 ng/ul
RT: 7.322 min Scan# 669
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion: 94 Resp: 652646
Ion Ratio Lower Upper
94 100
65 25.8 20.6 31.0
66 32.6 26.1 39.1









#10
 Bis(2-Chloroethyl)ether
 Concen: 21.818 ng/ul
 RT: 7.552 min Scan# 708
 Delta R.T. 0.000 min
 Lab File: BP015062.D
 Acq: 11 May 2023 12:47

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020613

Tgt Ion: 93 Resp: 532971




Ion	Ratio	Lower	Upper
93	100		
63	72.5	58.0	87.0
95	33.0	26.4	39.6




#12
 2-Chlorophenol
 Concen: 21.715 ng/ul
 RT: 7.699 min Scan# 733
 Delta R.T. 0.000 min
 Lab File: BP015062.D
 Acq: 11 May 2023 12:47

Tgt Ion: 128 Resp: 502852

Ion	Ratio	Lower	Upper
128	100		
64	45.2	36.2	54.2
130	32.9	26.3	39.5





#13

2-Methylphenol

Concen: 21.676 ng/ul

RT: 8.587 min Scan# 884

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :





SSTD020613

Tgt Ion:108 Resp: 511505

Ion Ratio Lower Upper

108 100

107 89.7 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.144 ng/ul

RT: 8.681 min Scan# 900

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion: 45 Resp: 677731


Ion Ratio Lower Upper

45 100

77 16.0 12.8 19.2

79 11.9 9.5 14.3









#16
Acetophenone
Concen: 23.036 ng/ul
RT: 8.975 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:105 Resp: 853038
Ion Ratio Lower Upper
105 100
77 74.3 59.4 89.2
51 24.0 19.2 28.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.481 ng/ul
RT: 8.957 min Scan# 947
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion: 70 Resp: 436605
Ion Ratio Lower Upper
70 100
42 51.3 41.0 61.6
101 9.9 7.9 11.9
130 20.1 16.1 24.1









#18
4-Methylphenol
Concen: 21.756 ng/ul
RT: 8.922 min Scan# 941
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:108 Resp: 565074
Ion Ratio Lower Upper
108 100
107 110.6 88.5 132.7



#19
Hexachloroethane
Concen: 21.792 ng/ul
RT: 9.234 min Scan# 994
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:117 Resp: 206047
Ion Ratio Lower Upper
117 100
201 94.8 75.8 113.8
199 58.1 46.5 69.7









#22
Nitrobenzene
Concen: 21.881 ng/ul
RT: 9.363 min Scan# 1016
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion: 77 Resp: 620524
Ion Ratio Lower Upper
77 100
123 45.2 36.2 54.2
65 12.6 10.1 15.1



#23
Isophorone
Concen: 21.985 ng/ul
RT: 9.893 min Scan# 1106
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion: 82 Resp: 1281547
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 17.0 13.6 20.4









#25
2-Nitrophenol
Concen: 22.040 ng/ul
RT: 10.081 min Scan# 1138
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:139 Resp: 308783
Ion Ratio Lower Upper
139 100
65 48.0 38.4 57.6
109 27.7 22.2 33.2



#26
2,4-Dimethylphenol
Concen: 21.947 ng/ul
RT: 10.134 min Scan# 1147
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:107 Resp: 613592
Ion Ratio Lower Upper
107 100
121 51.3 41.0 61.6
122 86.9 69.5 104.3









#27
 Bis(2-Chloroethoxy)methane
 Concen: 21.884 ng/ul
 RT: 10.375 min Scan# 1188
 Delta R.T. 0.000 min
 Lab File: BP015062.D
 Acq: 11 May 2023 12:47




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020613


Tgt Ion: 93 Resp: 776161
 Ion Ratio Lower Upper
 93 100
 95 33.4 26.7 40.1
 123 11.4 9.1 13.7



#29
 2,4-Dichlorophenol
 Concen: 21.876 ng/ul
 RT: 10.616 min Scan# 1229
 Delta R.T. 0.000 min
 Lab File: BP015062.D
 Acq: 11 May 2023 12:47

Tgt Ion: 162 Resp: 509439
 Ion Ratio Lower Upper
 162 100
 164 65.0 52.0 78.0
 98 36.1 28.9 43.3









#30
Naphthalene
Concen: 21.718 ng/ul
RT: 11.028 min Scan# 1299
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:128 Resp: 1725364
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 13.2 10.6 15.8



#32
4-Chloroaniline
Concen: 22.731 ng/ul
RT: 11.134 min Scan# 1317
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:127 Resp: 796924
Ion Ratio Lower Upper
127 100
129 32.3 25.8 38.8





#33

Hexachlorobutadiene

Concen: 21.457 ng/ul

RT: 11.304 min Scan# 1346

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:225 Resp: 301151

Ion Ratio Lower Upper

225 100

223 63.8 51.0 76.6

227 65.3 52.2 78.4



#34

Caprolactam

Concen: 12.514 ng/ul

RT: 11.910 min Scan# 1449

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:113 Resp: 109862


Ion Ratio Lower Upper

113 100

55 151.5 121.2 181.8

56 112.8 90.2 135.4









#35
4-Chloro-3-methylphenol
Concen: 22.037 ng/ul
RT: 12.246 min Scan# 1506
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:107 Resp: 579450
Ion Ratio Lower Upper
107 100
144 25.4 20.3 30.5
142 79.0 63.2 94.8



#36
2-Methylnaphthalene
Concen: 21.883 ng/ul
RT: 12.622 min Scan# 1570
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:142 Resp: 1171363
Ion Ratio Lower Upper
142 100
141 88.8 71.0 106.6









#37
1-Methylnaphthalene
Concen: 22.066 ng/ul
RT: 12.840 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:142 Resp: 1186118
Ion Ratio Lower Upper
142 100
141 93.3 74.6 112.0
116 3.6 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 21.851 ng/ul
RT: 12.987 min Scan# 1632
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:216 Resp: 608573
Ion Ratio Lower Upper
216 100
214 76.7 61.4 92.0
179 19.9 15.9 23.9
108 17.1 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 20.728 ng/ul
RT: 12.963 min Scan# 1628
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:237 Resp: 374026
Ion Ratio Lower Upper
237 100
235 61.7 49.4 74.0
272 13.0 10.4 15.6



#41
2,4,6-Trichlorophenol
Concen: 21.547 ng/ul
RT: 13.222 min Scan# 1672
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:196 Resp: 423747
Ion Ratio Lower Upper
196 100
198 94.4 75.5 113.3
200 29.5 23.6 35.4





#42

2,4,5-Trichlorophenol

Concen: 21.465 ng/ul

RT: 13.293 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:196 Resp: 457706

Ion Ratio Lower Upper

196 100

198 96.2 77.0 115.4

200 30.4 24.3 36.5



#43

1,1'-Biphenyl

Concen: 22.060 ng/ul

RT: 13.622 min Scan# 1740

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:154 Resp: 1593829


Ion Ratio Lower Upper

154 100

153 40.9 32.7 49.1

76 13.4 10.7 16.1





#44

2-Chloronaphthalene

Concen: 21.917 ng/ul

RT: 13.663 min Scan# 1747

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:162 Resp: 1236622

Ion Ratio Lower Upper

162 100

164 32.4 25.9 38.9

127 38.5 30.8 46.2



#45

2-Nitroaniline

Concen: 22.375 ng/ul

RT: 13.869 min Scan# 1782

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion: 65 Resp: 361230


Ion Ratio Lower Upper

65 100

92 74.8 59.8 89.8

138 115.4 92.3 138.5









#47
Dimethylphthalate
Concen: 22.078 ng/ul
RT: 14.234 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:163 Resp: 1618000
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.3 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 22.419 ng/ul
RT: 14.357 min Scan# 1865
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:165 Resp: 325493
Ion Ratio Lower Upper
165 100
89 60.0 48.0 72.0
121 21.9 17.5 26.3





#50

Acenaphthylene

Concen: 22.277 ng/ul

RT: 14.510 min Scan# 1891

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:152 Resp: 2000431

Ion Ratio Lower Upper

152 100

151 20.1 16.1 24.1

153 12.9 10.3 15.5



#51

3-Nitroaniline

Concen: 19.978 ng/ul

RT: 14.687 min Scan# 1921

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:138 Resp: 261176


Ion Ratio Lower Upper

138 100

108 11.0 8.8 13.2

92 116.7 93.4 140.0









#52
Acenaphthene
Concen: 22.083 ng/ul
RT: 14.851 min Scan# 1949
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:153 Resp: 1354378
Ion Ratio Lower Upper
153 100
152 47.5 38.0 57.0
154 88.7 71.0 106.4



#53
2,4-Dinitrophenol
Concen: 20.078 ng/ul
RT: 14.898 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:184 Resp: 193554
Ion Ratio Lower Upper
184 100
63 62.1 49.7 74.5
154 59.4 47.5 71.3









#55
4-Nitrophenol
Concen: 22.060 ng/ul
RT: 14.992 min Scan# 1973
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:109 Resp: 221704
Ion Ratio Lower Upper
109 100
139 144.1 115.3 172.9
65 134.9 107.9 161.9



#56
Dibenzofuran
Concen: 22.117 ng/ul
RT: 15.181 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:168 Resp: 1886700
Ion Ratio Lower Upper
168 100
139 36.7 29.4 44.0









#57
2,4-Dinitrotoluene
Concen: 22.451 ng/ul
RT: 15.145 min Scan# 1999
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:165 Resp: 471674
Ion Ratio Lower Upper
165 100
63 41.4 33.1 49.7
182 2.6 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 22.074 ng/ul
RT: 15.404 min Scan# 2043
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:232 Resp: 397236
Ion Ratio Lower Upper
232 100
131 46.0 36.8 55.2
130 2.4 1.9 2.9
166 30.8 24.6 37.0









#59
Diethylphthalate
Concen: 22.172 ng/ul
RT: 15.592 min Scan# 2075
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:149 Resp: 1620620
Ion Ratio Lower Upper
149 100
177 21.9 17.5 26.3
150 12.3 9.8 14.8



#61
Fluorene
Concen: 22.302 ng/ul
RT: 15.834 min Scan# 2116
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:166 Resp: 1528920
Ion Ratio Lower Upper
166 100
165 97.2 77.8 116.6
167 13.2 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 22.130 ng/ul
RT: 15.822 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:204 Resp: 741256
Ion Ratio Lower Upper
204 100
206 32.6 26.1 39.1
141 55.6 44.5 66.7



#63
4-Nitroaniline
Concen: 20.806 ng/ul
RT: 15.851 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:138 Resp: 234888
Ion Ratio Lower Upper
138 100
92 51.5 41.2 61.8
108 75.1 60.1 90.1









#66
4,6-Dinitro-2-methylphenol
Concen: 21.787 ng/ul
RT: 15.910 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:198 Resp: 288769
Ion Ratio Lower Upper
198 100
182 4.6 3.7 5.5
77 26.5 21.2 31.8



#67
N-Nitrosodiphenylamine
Concen: 22.094 ng/ul
RT: 16.034 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:169 Resp: 1323896
Ion Ratio Lower Upper
169 100
168 66.9 53.5 80.3
167 36.1 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 21.549 ng/ul
RT: 16.716 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:248 Resp: 452312
Ion Ratio Lower Upper
248 100
250 98.7 79.0 118.4
141 73.2 58.6 87.8



#69
Hexachlorobenzene
Concen: 21.620 ng/ul
RT: 16.839 min Scan# 2287
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:284 Resp: 536284
Ion Ratio Lower Upper
284 100
142 34.1 27.3 40.9
249 30.7 24.6 36.8





#70

Atrazine

Concen: 22.327 ng/ul

RT: 16.986 min Scan# 2312

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:200 Resp: 494501

Ion Ratio Lower Upper

200 100

173 26.8 21.4 32.2

215 47.6 38.1 57.1



#71

Pentachlorophenol

Concen: 21.234 ng/ul

RT: 17.186 min Scan# 2346

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:266 Resp: 329458


Ion Ratio Lower Upper

266 100

264 63.7 51.0 76.4

268 64.7 51.8 77.6





#72

Phenanthrene

Concen: 21.961 ng/ul

RT: 17.581 min Scan# 2413

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:178 Resp: 2436952

Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 19.5 15.6 23.4



#74

Anthracene

Concen: 22.236 ng/ul

RT: 17.675 min Scan# 2429

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:178 Resp: 2495170


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 19.0 15.2 22.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.683 ng/uL
RT: 13.587 min Scan# 1734
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47




Instrument :
BNA_P
ClientSampleId :
SSTD020613


Tgt Ion:216 Resp: 626494
Ion Ratio Lower Upper
216 100
214 77.8 62.2 93.4
218 47.7 38.2 57.2



#76
Pentachlorobenzene
Concen: 21.704 ng/uL
RT: 15.104 min Scan# 1992
Delta R.T. 0.000 min
Lab File: BP015062.D
Acq: 11 May 2023 12:47

Tgt Ion:250 Resp: 629659
Ion Ratio Lower Upper
250 100
248 62.2 49.8 74.6
252 64.5 51.6 77.4









#77
 Carbazole
 Concen: 22.914 ng/ul
 RT: 17.939 min Scan# 2474
 Delta R.T. 0.000 min
 Lab File: BP015062.D
 Acq: 11 May 2023 12:47

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020613

Tgt Ion:167 Resp: 2254668




Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.4	26.2
139	12.8	10.2	15.4




#78
 Di-n-butylphthalate
 Concen: 22.157 ng/ul
 RT: 18.469 min Scan# 2564
 Delta R.T. 0.000 min
 Lab File: BP015062.D
 Acq: 11 May 2023 12:47

Tgt Ion:149 Resp: 2807568

Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.3	10.9
104	4.9	3.9	5.9





#80

Fluoranthene

Concen: 21.821 ng/ul

RT: 19.533 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:202 Resp: 2822914

Ion Ratio Lower Upper

202 100

101 11.9 9.5 14.3

100 9.1 7.3 10.9



#82

Pyrene

Concen: 21.962 ng/ul

RT: 19.886 min Scan# 2805

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:202 Resp: 2933992


Ion Ratio Lower Upper

202 100

101 14.0 11.2 16.8

100 11.3 9.0 13.6





#83

Butylbenzylphthalate

Concen: 21.927 ng/ul

RT: 20.739 min Scan# 2950

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:149 Resp: 1298664

Ion Ratio Lower Upper

149 100

91 71.8 57.4 86.2

206 20.8 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 22.090 ng/ul

RT: 21.522 min Scan# 3083

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:252 Resp: 887846


Ion Ratio Lower Upper

252 100

254 63.5 50.8 76.2

126 12.3 9.8 14.8





#85

Benzo(a)anthracene

Concen: 21.664 ng/ul

RT: 21.604 min Scan# 3097

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:228 Resp: 2741357

Ion Ratio Lower Upper

228 100

229 20.1 16.1 24.1

226 27.4 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.034 ng/ul

RT: 21.498 min Scan# 3079

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:149 Resp: 1878783


Ion Ratio Lower Upper

149 100

167 27.2 21.8 32.6

279 4.0 3.2 4.8





#87

Chrysene

Concen: 21.702 ng/ul

RT: 21.657 min Scan# 3106

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:228 Resp: 2625177

Ion Ratio Lower Upper

228 100

226 31.3 25.0 37.6

229 20.1 16.1 24.1



#89

Di-n-octyl phthalate

Concen: 21.844 ng/ul




RT: 22.480 min Scan# 3246


Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Tgt Ion:149 Resp: 3230113





#90

Benzo(b)fluoranthene

Concen: 21.150 ng/ul

RT: 23.404 min Scan# 3403

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:252 Resp: 2894907

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.2

125 10.4 8.3 12.5



#91

Benzo(k)fluoranthene

Concen: 21.394 ng/ul

RT: 23.451 min Scan# 3411

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:252 Resp: 2811859


Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3

125 10.6 8.5 12.7





#93

Benzo(a)pyrene

Concen: 21.263 ng/ul

RT: 24.086 min Scan# 3519

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:252 Resp: 2572446

Ion Ratio Lower Upper

252 100

253 21.5 17.2 25.8

125 11.4 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.401 ng/ul

RT: 26.933 min Scan# 4003

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47




Tgt Ion:276 Resp: 3352487


Ion Ratio Lower Upper

276 100

138 25.1 20.1 30.1

277 24.8 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 21.489 ng/ul

RT: 26.951 min Scan# 4006

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Instrument :

BNA_P

ClientSampleId :

SSTD020613





Tgt Ion:278 Resp: 2763269

Ion Ratio Lower Upper

278 100

139 16.6 13.3 19.9

279 23.9 19.1 28.7



#96

Benzo(g,h,i)perylene

Concen: 21.400 ng/ul

RT: 27.780 min Scan# 4147

Delta R.T. 0.000 min

Lab File: BP015062.D

Acq: 11 May 2023 12:47

Tgt Ion:276 Resp: 2733689

Ion Ratio Lower Upper

276 100

138 22.1 17.7 26.5

277 23.5 18.8 28.2

